



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions

SDG No.: Q1211

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134464

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TOC	mg/L	0.2	0.5000	J	0.19	1	01/15/2025
Sample ID: CCB1 TOC	mg/L	0.26	0.5000	J	0.19	1	01/29/2025
Sample ID: CCB2 TOC	mg/L	0.26	0.5000	J	0.19	1	01/29/2025

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions	SDG No.: Q1211
Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169	RunNo.: LB134466

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	01/29/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	01/29/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	01/29/2025

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions

SDG No.: Q1211

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134473

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	01/20/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/20/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	01/20/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/20/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	01/20/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	01/20/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	01/20/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	01/29/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/29/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	01/29/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/29/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	01/29/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	01/29/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	01/29/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	01/29/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/29/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	01/29/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/29/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	01/29/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	01/29/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	01/29/2025

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions

SDG No.: Q1211

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134532

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/03/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/03/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/03/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/03/2025
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	02/03/2025