

### Laboratory Control Sample Summary

**Client:** Weston Solutions

**SDG No.:** Q1211

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**Run No.:** LB134464

| Analyte   | Units       | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|-------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB134464BSW |            |        |                 |            |                 |                     |               |
| TOC       | mg/L        | 10         | 9.80   |                 | 98         | 1               | 90-110              | 01/29/2025    |

### Laboratory Control Sample Summary

|                 |                                                 |                 |          |
|-----------------|-------------------------------------------------|-----------------|----------|
| <b>Client:</b>  | Weston Solutions                                | <b>SDG No.:</b> | Q1211    |
| <b>Project:</b> | Ft Meade Tipton Airfield Parcel RI - PO 0111169 | <b>Run No.:</b> | LB134466 |

| Analyte             | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|---------------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID           | lb134466BS |            |        |                 |            |                 |                     |               |
| Hexavalent Chromium | mg/L       | 0.5        | 0.51   |                 | 102        | 1               | 90-111              | 01/29/2025    |

### Laboratory Control Sample Summary

|                 |                                                 |                 |          |
|-----------------|-------------------------------------------------|-----------------|----------|
| <b>Client:</b>  | Weston Solutions                                | <b>SDG No.:</b> | Q1211    |
| <b>Project:</b> | Ft Meade Tipton Airfield Parcel RI - PO 0111169 | <b>Run No.:</b> | LB134473 |

| Analyte             | Units       | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|---------------------|-------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID           | LB134473BSW |            |        |                 |            |                 |                     |               |
| Bromide             | mg/L        | 10         | 9.90   |                 | 99         | 1               | 90-110              | 01/29/2025    |
| Chloride            | mg/L        | 3          | 3.00   |                 | 100        | 1               | 90-110              | 01/29/2025    |
| Fluoride            | mg/L        | 2          | 2.00   |                 | 100        | 1               | 90-110              | 01/29/2025    |
| Nitrite             | mg/L        | 3          | 3.00   |                 | 100        | 1               | 90-110              | 01/29/2025    |
| Nitrate             | mg/L        | 2.5        | 2.50   |                 | 100        | 1               | 90-110              | 01/29/2025    |
| Sulfate             | mg/L        | 15         | 14.7   |                 | 98         | 1               | 90-110              | 01/29/2025    |
| Orthophosphate as P | mg/L        | 5          | 5.00   |                 | 100        | 1               | 90-110              | 01/29/2025    |

### Laboratory Control Sample Summary

|                 |                                                 |                 |          |
|-----------------|-------------------------------------------------|-----------------|----------|
| <b>Client:</b>  | Weston Solutions                                | <b>SDG No.:</b> | Q1211    |
| <b>Project:</b> | Ft Meade Tipton Airfield Parcel RI - PO 0111169 | <b>Run No.:</b> | LB134521 |

| Analyte        | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|----------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID      | LB134521BS |            |        |                 |            |                 |                     |               |
| Oil and Grease | mg/L       | 20.0       | 16.9   |                 | 84         | 1               | 78-114              | 02/01/2025    |

### Laboratory Control Sample Summary

**Client:** Weston Solutions

**SDG No.:** Q1211

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**Run No.:** LB134532

| Analyte      | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|--------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID    | PB166477BS |            |        |                 |            |                 |                     |               |
| Ammonia as N | mg/L       | 1          | 0.97   |                 | 97         | 1               | 90-110              | 02/03/2025    |