

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Weston Solutions

SDG No.: Q1211

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166596BL		WATER		Batch Number:		PB166596		Prep Date:	02/06/2025	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	02/06/2025	12:49	LB134599

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Weston Solutions

SDG No.: Q1211

Instrument: P8

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166378BL	WATER			Batch Number:	PB166378			Prep Date:	01/30/2025	
	Aluminum	10.0	<10.0	U	5.00	10.0	P	02/06/2025	18:21	LB134616
	Antimony	1.00	<1.00	U	0.13	1.00	P	02/06/2025	18:21	LB134616
	Arsenic	0.50	<0.50	U	0.13	0.50	P	02/06/2025	18:21	LB134616
	Barium	5.00	<5.00	U	0.63	5.00	P	02/06/2025	18:21	LB134616
	Beryllium	0.080	<0.50	J	0.13	0.50	P	02/06/2025	18:21	LB134616
	Cadmium	0.50	<0.50	U	0.25	0.50	P	02/06/2025	18:21	LB134616
	Calcium	250	<250	U	95.0	250	P	02/06/2025	18:21	LB134616
	Chromium	1.00	<1.00	U	0.38	1.00	P	02/06/2025	18:21	LB134616
	Cobalt	0.50	<0.50	U	0.13	0.50	P	02/06/2025	18:21	LB134616
	Copper	1.00	<1.00	U	0.75	1.00	P	02/06/2025	18:21	LB134616
	Iron	25.0	<25.0	U	12.5	25.0	P	02/06/2025	18:21	LB134616
	Lead	0.070	<0.50	J	0.38	0.50	P	02/06/2025	18:21	LB134616
	Magnesium	250	<250	U	95.0	250	P	02/06/2025	18:21	LB134616
	Manganese	0.50	<0.50	U	0.38	0.50	P	02/06/2025	18:21	LB134616
	Nickel	0.14	<0.50	J	0.13	0.50	P	02/06/2025	18:21	LB134616
	Potassium	250	<250	U	95.0	250	P	02/06/2025	18:21	LB134616
	Selenium	2.50	<2.50	U	2.25	2.50	P	02/06/2025	18:21	LB134616
	Silver	0.50	<0.50	U	0.25	0.50	P	02/06/2025	18:21	LB134616
	Sodium	250	<250	U	95.0	250	P	02/06/2025	18:21	LB134616
	Thallium	0.50	<0.50	U	0.25	0.50	P	02/06/2025	18:21	LB134616
	Vanadium	2.50	<2.50	U	0.13	2.50	P	02/06/2025	18:21	LB134616
	Zinc	0.45	<2.50	J	0.75	2.50	P	02/06/2025	18:21	LB134616