

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions **SDG No.:** Q1211
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1211 **SAS No.:** Q1211

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166378BS							
Aluminum	ug/L	5000	4680		94	84 - 117	P
Antimony	ug/L	250	256		102	85 - 117	P
Arsenic	ug/L	250	253		101	84 - 116	P
Barium	ug/L	1300	1280		98	86 - 114	P
Beryllium	ug/L	250	243		97	83 - 121	P
Cadmium	ug/L	250	260		104	87 - 115	P
Calcium	ug/L	25000	24400		98	87 - 118	P
Chromium	ug/L	250	256		102	85 - 116	P
Cobalt	ug/L	250	257		103	86 - 115	P
Copper	ug/L	2500	2500		100	85 - 118	P
Iron	ug/L	12500	12600		101	87 - 118	P
Lead	ug/L	1300	1260		97	88 - 115	P
Magnesium	ug/L	25000	25200		101	83 - 118	P
Manganese	ug/L	2500	2500		100	87 - 115	P
Nickel	ug/L	250	240		96	85 - 117	P
Potassium	ug/L	12500	12100		97	87 - 115	P
Selenium	ug/L	250	251		100	80 - 120	P
Silver	ug/L	250	260		104	85 - 116	P
Sodium	ug/L	25000	25400		102	85 - 117	P
Thallium	ug/L	250	253		101	82 - 116	P
Vanadium	ug/L	250	252		101	86 - 115	P
Zinc	ug/L	2500	2450		98	83 - 119	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions

SDG No.: Q1211

Contract: WEST04

Lab Code: CHEM

Case No.: Q1211

SAS No.: Q1211

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166596BS Mercury	ug/L	4.0	3.81		95	82 - 119	CV