

WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE

Lab Name: Chemtech **Client:** Weston Solutions
Lab Code: CHEM **Cas No:** Q1211 **SAS No :** Q1211 **SDG No:** Q1211
Matrix Spike - EPA Sample No : PB166364BS **Datafile:** FE052160.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
DRO	200	0	182	91	78-117