

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013125\
 Data File : BF141334.D
 Acq On : 31 Jan 2025 16:26
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 31 18:34:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 29 17:41:25 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	121	0.00
2	1,4-Dioxane	40.000	37.986	5.0	122	-0.01
3	Pyridine	40.000	38.326	4.2	121	-0.02
4	n-Nitrosodimethylamine	40.000	38.269	4.3	121	-0.04
5 S	2-Fluorophenol	80.000	73.623	8.0	121	0.00
6	Aniline	40.000	38.018	5.0	119	0.00
7 S	Phenol-d6	80.000	72.667	9.2	120	-0.01
8	2-Chlorophenol	40.000	37.030	7.4	120	0.00
9	Benzaldehyde	40.000	39.268	1.8	127	0.00
10 C	Phenol	40.000	36.397	9.0	118	0.00
11	bis(2-Chloroethyl)ether	40.000	36.649	8.4	121	0.00
12	1,3-Dichlorobenzene	40.000	36.939	7.7	120	0.00
13 C	1,4-Dichlorobenzene	40.000	37.089	7.3	121	0.00
14	1,2-Dichlorobenzene	40.000	36.203	9.5	118	0.00
15	Benzyl Alcohol	40.000	37.479	6.3	118	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.050	7.4	120	0.00
17	2-Methylphenol	40.000	36.801	8.0	121	0.00
18	Hexachloroethane	40.000	36.943	7.6	120	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.368	9.1	121	-0.01
20	3+4-Methylphenols	40.000	35.761	10.6	117	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	121	0.00
22	Acetophenone	40.000	35.637	10.9	118	0.00
23 S	Nitrobenzene-d5	80.000	72.918	8.9	119	0.00
24	Nitrobenzene	40.000	36.727	8.2	118	-0.01
25	Isophorone	40.000	36.952	7.6	120	0.00
26 C	2-Nitrophenol	40.000	37.318	6.7	119	0.00
27	2,4-Dimethylphenol	40.000	37.535	6.2	120	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.489	8.8	119	0.00
29 C	2,4-Dichlorophenol	40.000	37.008	7.5	120	0.00
30	1,2,4-Trichlorobenzene	40.000	36.030	9.9	119	0.00
31	Naphthalene	40.000	35.989	10.0	118	0.00
32	Benzoic acid	40.000	42.441	-6.1	127	-0.03
33	4-Chloroaniline	40.000	35.481	11.3	114	0.00
34 C	Hexachlorobutadiene	40.000	36.308	9.2	119	0.00
35	Caprolactam	40.000	38.166	4.6	123	-0.02
36 C	4-Chloro-3-methylphenol	40.000	37.503	6.2	123	0.00
37	2-Methylnaphthalene	40.000	35.969	10.1	117	0.00
38	1-Methylnaphthalene	40.000	35.916	10.2	119	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.817	8.0	118	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.297	1.8	116	0.00
42 S	2,4,6-Tribromophenol	80.000	75.335	5.8	121	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.014	5.0	117	0.00
44	2,4,5-Trichlorophenol	40.000	37.124	7.2	120	0.00
45 S	2-Fluorobiphenyl	80.000	70.424	12.0	117	0.00
46	1,1'-Biphenyl	40.000	36.082	9.8	117	0.00
47	2-Chloronaphthalene	40.000	36.046	9.9	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013125\
 Data File : BF141334.D
 Acq On : 31 Jan 2025 16:26
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 31 18:34:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 29 17:41:25 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.746	5.6	120	0.00
49	Acenaphthylene	40.000	36.509	8.7	118	0.00
50	Dimethylphthalate	40.000	36.994	7.5	122	0.00
51	2,6-Dinitrotoluene	40.000	37.813	5.5	121	0.00
52 C	Acenaphthene	40.000	36.678	8.3	119	0.00
53	3-Nitroaniline	40.000	37.207	7.0	114	0.00
54 P	2,4-Dinitrophenol	40.000	41.956	-4.9	125	0.00
55	Dibenzofuran	40.000	35.957	10.1	116	0.00
56 P	4-Nitrophenol	40.000	39.808	0.5	121	0.00
57	2,4-Dinitrotoluene	40.000	38.519	3.7	122	0.00
58	Fluorene	40.000	35.892	10.3	117	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.450	3.9	122	0.00
60	Diethylphthalate	40.000	37.205	7.0	120	0.00
61	4-Chlorophenyl-phenylether	40.000	36.147	9.6	119	0.00
62	4-Nitroaniline	40.000	38.469	3.8	121	-0.01
63	Azobenzene	40.000	37.160	7.1	120	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	124	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.330	-10.8	129	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.253	6.9	119	0.00
67	4-Bromophenyl-phenylether	40.000	38.536	3.7	123	0.00
68	Hexachlorobenzene	40.000	37.954	5.1	122	0.00
69	Atrazine	40.000	34.836	12.9	109	0.00
70 C	Pentachlorophenol	40.000	42.456	-6.1	127	0.00
71	Phenanthrene	40.000	37.212	7.0	121	0.00
72	Anthracene	40.000	37.760	5.6	122	0.00
73	Carbazole	40.000	38.354	4.1	123	0.00
74	Di-n-butylphthalate	40.000	39.613	1.0	128	0.00
75 C	Fluoranthene	40.000	38.576	3.6	125	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	131	0.00
77	Benzydine	40.000	35.937	10.2	103	0.00
78	Pyrene	40.000	35.817	10.5	128	0.00
79 S	Terphenyl-d14	80.000	70.798	11.5	127	0.00
80	Butylbenzylphthalate	40.000	39.406	1.5	138	0.00
81	Benzo(a)anthracene	40.000	34.980	12.6	124	0.00
82	3,3'-Dichlorobenzidine	40.000	36.313	9.2	125	0.00
83	Chrysene	40.000	35.690	10.8	130	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.633	3.4	137	0.00
85 c	Di-n-octyl phthalate	40.000	39.037	2.4	137	0.00
86 I	Perylene-d12	20.000	20.000	0.0	119	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.527	6.2	114	0.00
88	Benzo(b)fluoranthene	40.000	38.817	3.0	129	0.00
89	Benzo(k)fluoranthene	40.000	34.632	13.4	106	0.00
90 C	Benzo(a)pyrene	40.000	37.270	6.8	116	0.00
91	Dibenzo(a,h)anthracene	40.000	37.536	6.2	114	-0.01
92	Benzo(g,h,i)perylene	40.000	38.355	4.1	118	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013125\
Data File : BF141334.D
Acq On : 31 Jan 2025 16:26
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 31 18:34:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 29 17:41:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0