

Surrogate Summary

SW-846

SDG No.: Q1211

Client: Weston Solutions

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166363BL	PB166363BL	2-Fluorophenol	150	119	79		19	119
		Phenol-d6	150	116	78		10	130
		Nitrobenzene-d5	100	81.7	82		44	120
		2-Fluorobiphenyl	100	77.3	77		44	119
		2,4,6-Tribromophenol	150	115	76		43	140
		Terphenyl-d14	100	78.3	78		50	134
PB166363BS	PB166363BS	2-Fluorophenol	150	117	78		19	119
		Phenol-d6	150	117	78		10	130
		Nitrobenzene-d5	100	81.2	81		44	120
		2-Fluorobiphenyl	100	77.1	77		44	119
		2,4,6-Tribromophenol	150	125	84		43	140
		Terphenyl-d14	100	82.3	82		50	134
PB166363BSD	PB166363BSD	2-Fluorophenol	150	118	78		19	119
		Phenol-d6	150	117	78		10	130
		Nitrobenzene-d5	100	82.1	82		44	120
		2-Fluorobiphenyl	100	77.8	78		44	119
		2,4,6-Tribromophenol	150	126	84		43	140
		Terphenyl-d14	100	83.2	83		50	134
Q1211-01	TAPHHA-MW01-012825-00-T4	2-Fluorophenol	150	57.4	38		19	119
		Phenol-d6	150	35.8	24		10	130
		Nitrobenzene-d5	100	85.4	85		44	120
		2-Fluorobiphenyl	100	80.4	80		44	119
		2,4,6-Tribromophenol	150	135	90		43	140
		Terphenyl-d14	100	68.7	69		50	134
Q1211-02	TAPIAL2-MW03-012825-00-T3	2-Fluorophenol	150	57.1	38		19	119
		Phenol-d6	150	35.2	23		10	130
		Nitrobenzene-d5	100	87.1	87		44	120
		2-Fluorobiphenyl	100	81.2	81		44	119
		2,4,6-Tribromophenol	150	134	89		43	140
		Terphenyl-d14	100	64.9	65		50	134