

Report of Analysis

Client:	Weston Solutions		Date Collected:	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	
Client Sample ID:	VN0130WBS01		SDG No.:	Q1211
Lab Sample ID:	VN0130WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085579.D	1		01/30/25 14:56	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	16.6		0.21	0.50	5.00	ug/L
74-87-3	Chloromethane	17.0		0.35	0.50	5.00	ug/L
75-01-4	Vinyl Chloride	17.2		0.34	0.75	5.00	ug/L
74-83-9	Bromomethane	18.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	18.8		0.56	0.50	5.00	ug/L
75-69-4	Trichlorofluoromethane	19.2		0.34	0.50	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.2		0.25	0.50	5.00	ug/L
75-35-4	1,1-Dichloroethene	18.7		0.26	0.75	5.00	ug/L
67-64-1	Acetone	110		1.40	3.80	25.0	ug/L
75-15-0	Carbon Disulfide	15.2		0.32	0.75	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.1		0.16	0.50	5.00	ug/L
79-20-9	Methyl Acetate	23.3		0.60	0.75	5.00	ug/L
75-09-2	Methylene Chloride	19.3		0.32	0.50	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.7		0.25	0.50	5.00	ug/L
75-34-3	1,1-Dichloroethane	19.7		0.23	0.50	5.00	ug/L
110-82-7	Cyclohexane	17.0		1.60	3.80	5.00	ug/L
78-93-3	2-Butanone	110		1.30	2.50	25.0	ug/L
56-23-5	Carbon Tetrachloride	19.4		0.25	0.50	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.0		0.25	0.50	5.00	ug/L
74-97-5	Bromochloromethane	18.9		0.18	0.50	5.00	ug/L
67-66-3	Chloroform	20.0		0.26	0.50	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.0		0.19	0.75	5.00	ug/L
108-87-2	Methylcyclohexane	16.9		0.19	0.50	5.00	ug/L
71-43-2	Benzene	19.4		0.16	0.50	5.00	ug/L
107-06-2	1,2-Dichloroethane	20.6		0.24	0.75	5.00	ug/L
79-01-6	Trichloroethene	19.0		0.32	0.75	5.00	ug/L
78-87-5	1,2-Dichloropropane	19.8		0.19	0.50	5.00	ug/L
75-27-4	Bromodichloromethane	21.0		0.24	0.50	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.75	2.50	25.0	ug/L
108-88-3	Toluene	20.5		0.18	0.50	5.00	ug/L

Report of Analysis

Client:	Weston Solutions		Date Collected:	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	
Client Sample ID:	VN0130WBS01		SDG No.:	Q1211
Lab Sample ID:	VN0130WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085579.D	1		01/30/25 14:56	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.0		0.21	0.50	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.2		0.18	0.50	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.4		0.21	0.50	5.00	ug/L
591-78-6	2-Hexanone	110		1.10	2.50	25.0	ug/L
124-48-1	Dibromochloromethane	21.1		0.18	0.50	5.00	ug/L
106-93-4	1,2-Dibromoethane	20.3		0.16	0.50	5.00	ug/L
127-18-4	Tetrachloroethene	20.2		0.25	0.50	5.00	ug/L
108-90-7	Chlorobenzene	20.1		0.13	0.50	5.00	ug/L
100-41-4	Ethyl Benzene	19.6		0.16	0.50	5.00	ug/L
179601-23-1	m/p-Xylenes	41.5		0.31	1.00	10.0	ug/L
95-47-6	o-Xylene	19.9		0.14	0.50	5.00	ug/L
100-42-5	Styrene	20.7		0.16	0.50	5.00	ug/L
75-25-2	Bromoform	21.4		0.21	0.50	5.00	ug/L
98-82-8	Isopropylbenzene	19.6		0.13	0.50	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.1		0.27	0.50	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.1		0.24	0.50	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.1		0.27	0.50	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.0		0.19	0.50	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	20.6		0.46	2.00	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.0		0.42	0.50	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.4		0.51	0.50	5.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.0		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	54.7		80 - 119		109%	SPK: 50
2037-26-5	Toluene-d8	54.9		89 - 112		110%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.1		85 - 114		110%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	185000	8.224				
540-36-3	1,4-Difluorobenzene	315000	9.1				
3114-55-4	Chlorobenzene-d5	278000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	135000	13.788				

Report of Analysis

Client:	Weston Solutions		Date Collected:	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	
Client Sample ID:	VN0130WBS01		SDG No.:	Q1211
Lab Sample ID:	VN0130WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085579.D	1		01/30/25 14:56	VN013025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products