

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WEST04
 Lab Code: CHEM Case No.: Q1211 SAS No.: Q1211 SDG NO.: Q1211
 Lab File ID: VN085591.D Date Analyzed: 01/31/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:04
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	190291	8.22	311950	9.10	284481	11.87
UPPER LIMIT	380582	8.724	623900	9.6	568962	12.365
LOWER LIMIT	95145.5	7.724	155975	8.6	142241	11.365
EPA SAMPLE NO.						
TAPIAL2-MW03-012825-00-T3	165654	8.22	310451	9.10	268445	11.87
VN0131WBL01	163540	8.22	304758	9.10	267449	11.87
VN0131WBS01	192673	8.22	320848	9.10	286465	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.