

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WEST04
 Lab Code: CHEM Case No.: Q1211 SAS No.: Q1211 SDG NO.: Q1211
 Lab File ID: VN085576.D Date Analyzed: 01/30/2025
 Instrument ID: MSVOA_N Time Analyzed: 13:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	143955	13.788				
UPPER LIMIT	287910	14.288				
LOWER LIMIT	71977.5	13.288				
EPA SAMPLE NO.						
TAPHHA-MW01-012825-00-T4	106487	13.79				
TAP-TB-02-012825-T4	105344	13.79				
VN0130WBL01	110956	13.79				
VN0130WBS01	135461	13.79				
VN0130WBSD01	127057	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.