

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WEST04
 Lab Code: CHEM Case No.: Q1211 SAS No.: Q1211 SDG NO.: Q1211
 Lab File ID: VN085591.D Date Analyzed: 01/31/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:04
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	145646	13.788				
UPPER LIMIT	291292	14.288				
LOWER LIMIT	72823	13.288				
EPA SAMPLE NO.						
TAPIAL2-MW03-012825-00-T3RE	102272	13.79				
VN0131WBL01	104737	13.79				
VN0131WBS01	135464	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.