

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q1212 SAS No.: Q1212 SDG No.: Q1212

Instrument ID: MSVOA_X Calibration Date/Time: 01/30/2025 09:23

Lab File ID: VX044769.D Init. Calib. Date(s): 01/16/2025 01/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:39 10:12

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.665	1.815	0.5	9.01	
Toluene	1.975	2.085	0.4	5.57	
Ethyl Benzene	2.156	2.256	0.1	4.64	
m/p-Xylenes	0.792	0.878	0.3	10.86	
o-Xylene	0.795	0.860	0.3	8.18	
1,2-Dichloroethane-d4	2.896	2.548	0.01	-12.02	
Toluene-d8	1.632	1.603	0.01	-1.78	
4-Bromofluorobenzene	0.583	0.559	0.2	-4.12	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.