

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1212

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VX044770.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0130WBS01	Benzene	20	20.7	ug/L	104			65	135	
	Toluene	20	20.5	ug/L	103			70	130	
	Ethyl Benzene	20	20.3	ug/L	102			60	140	
	m/p-Xylenes	40	42.7	ug/L	107			87	111	
	o-Xylene	20	21.2	ug/L	106			87	111	