

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: Q1212 SAS No.: Q1212 SDG NO.: Q1212
 Lab File ID: VX044769.D Date Analyzed: 01/30/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:23
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	29832	4.89	163814	6.75	150220	10.05
UPPER LIMIT	59664	5.385	327628	7.251	300440	10.549
LOWER LIMIT	14916	4.385	81907	6.251	75110	9.549
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (JAN)	27744	4.89	158226	6.76	142220	10.06
002-35TH-AVE (JAN)	25775	4.90	145121	6.76	129652	10.05
VX0130WBL01	24468	4.89	137072	6.76	123214	10.05
VX0130WBS01	25818	4.89	146320	6.75	136752	10.05
VX0130WBSD01	25350	4.89	141556	6.76	132262	10.05

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.