

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX011625

Review By	Mahesh Dadoda	Review On	1/17/2025 6:39:05 AM
Supervise By	Semsettin Yesilyurt	Supervise On	1/17/2025 6:41:46 AM
SubDirectory	VX011625	HP Acquire Method	HP Processing Method 624X011625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132565
Initial Calibration Stds	VP132567,VP132568,VP132569,VP132570,VP132571
CCC	VP132566,VP132573
Internal Standard/PEM	
ICV/I.BLK	VP132572
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044658.D	16 Jan 2025 08:02		 	Ok
2	VSTDIC005	VSTDIC005	VX044659.D	16 Jan 2025 08:39	V13516	 	Ok,M
3	VSTDICCC020	VSTDICCC020	VX044660.D	16 Jan 2025 09:02		 	Ok,M
4	VSTDIC050	VSTDIC050	VX044661.D	16 Jan 2025 09:25		 	Ok,M
5	VSTDIC100	VSTDIC100	VX044662.D	16 Jan 2025 09:49		 	Ok,M
6	VSTDIC150	VSTDIC150	VX044663.D	16 Jan 2025 10:12		 	Ok,M
7	IBLK	IBLK	VX044664.D	16 Jan 2025 10:35		 	Ok
8	VSTDICV020	ICVVX011625	VX044665.D	16 Jan 2025 11:45		 	Ok,M
9	VX0116WBS01	VX0116WBS01	VX044666.D	16 Jan 2025 12:15		 	Ok,M
10	VX0116WBSD01	VX0116WBSD01	VX044667.D	16 Jan 2025 12:42		 	Ok,M
11	VX0116WBL01	VX0116WBL01	VX044668.D	16 Jan 2025 13:05		 	Ok
12	Q1083-01	001-WILLETS-PT-BLVD	VX044669.D	16 Jan 2025 13:33	vial A pH<2	 	Ok
13	Q1083-02	002-35TH-AVE(DEC)	VX044670.D	16 Jan 2025 13:55	vial A pH<2	 	Ok
14	IBLK	IBLK	VX044671.D	16 Jan 2025 14:18		 	Ok
15	P4368-09 2.5PPB	MDL-WATER-03-QT4-2	VX044672.D	16 Jan 2025 14:46		 	Ok,M
16	VSTDCC020	VSTDCC020EC	VX044673.D	16 Jan 2025 16:04		 	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX013025

Review By	John Carlone	Review On	1/30/2025 12:31:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	1/31/2025 12:25:37 PM
SubDirectory	VX013025	HP Acquire Method	HP Processing Method 624X011625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132770		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132771,VP132772		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044768.D	30 Jan 2025 08:49		JC/MD	Ok
2	VSTDCCC020	VSTDCCC020	VX044769.D	30 Jan 2025 09:23	V13516	JC/MD	Ok,M
3	VX0130WBS01	VX0130WBS01	VX044770.D	30 Jan 2025 09:55		JC/MD	Ok,M
4	VX0130WBSD01	VX0130WBSD01	VX044771.D	30 Jan 2025 10:21		JC/MD	Ok,M
5	VX0130WBL01	VX0130WBL01	VX044772.D	30 Jan 2025 10:44		JC/MD	Ok
6	Q1212-01	001-WILLETS-PT-BLVD	VX044773.D	30 Jan 2025 11:15	vial A pH<2 Surrogate Fail	JC/MD	ReRun
7	Q1212-02	002-35TH-AVE(JAN)	VX044774.D	30 Jan 2025 11:37	vial A pH<2	JC/MD	Ok
8	Q1212-01	001-WILLETS-PT-BLVD	VX044775.D	30 Jan 2025 12:00	vial B pH<2	JC/MD	Ok
9	IBLK	IBLK	VX044776.D	30 Jan 2025 12:23		JC/MD	Ok
10	Q1214-01	PARAMUS-CONDENS	VX044777.D	30 Jan 2025 12:46	vial A pH<2 Surrogate Fail	JC/MD	ReRun
11	IBLK	IBLK	VX044778.D	30 Jan 2025 13:09		JC/MD	Ok
12	Q1214-01	PARAMUS-CONDENS	VX044779.D	30 Jan 2025 13:32	vial B pH<2	JC/MD	Ok
13	IBLK	IBLK	VX044780.D	30 Jan 2025 13:55		JC/MD	Ok
14	VSTDCCC020	VSTDCCC020EC	VX044781.D	30 Jan 2025 16:00		JC/MD	Ok,M

M : Manual Integration