284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

DATA PACKAGE

VOLATILE ORGANICS

PROJECT NAME : NYCDDC SANTWOBR BROOKLYN BRIDGE BBMCR

RU2 ENGINEERING, LLC

2 Melinda Drive

Monroe Township, NJ - 08831

Phone No: 732-261-2236

ORDER ID : Q1215

ATTENTION : Rutu Manani

Laboratory Certification ID # 20012

Table Of Contents for Q1215

1) TCLP VOLATILES DATA 2
2) Signature Page 4
3) Case Narrative 5
4) Qualifier Page 7
5) Conformance/Non Conformance 8
6) QA Checklist 10
7) Chronicle 11
8) Hit Summary 12
9) QC Data Summary For TCLP VOA 13
- 9.1) Deuterated Monitoring Compound Summary 14
- 9.2) LCS/LCSD Summary 15
- 9.3) Method Blank Summary 18
- 9.4) GS/MS Tune Summary 20
- 9.5) Internal Standard Area and RT Summary 24

10) Sample Data 28
- 10.1) JPP-29.1-012825 29
- 10.2) JPP-29.2-012825 38

11) Calibration Data Summary 47
- 11.1) Initial Calibration Data 48
  - 11.1.1) VN011425 48
  - 11.1.2) VX011525 417
- 11.2) Continued Calibration Data 786
  - 11.2.1) VN085591.D 786
  - 11.2.2) VX044802.D 845

12) QC Sample Data 904
- 12.1) Tune Raw Data 905
- 12.2) Method Blank Data 909
- 12.3) LCS Data 923
- 12.4) LCSD Data 1029

13) Manual Integration 1082
14) Analytical Runlogs 1088
15) Extraction Logs 1100
- 15.1) PB166357.pdf 1100
- 15.2) PB166357IC.pdf 1103

16) Standard Prep Logs 1104

17) Shipping Document 1243
17.1) Chain Of Custody 1244
17.2) Lab Certificate 1245
17.3) Internal COC 1246

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : Q1215

Project ID : NYCDDC SANTWOBR Brooklyn Bridge BBMCR

Client : RU2 Engineering, LLC

Lab Sample Number	Client Sample Number
Q1215-01	JPP-29.1-012825
Q1215-02	JPP-29.1-012825
Q1215-03	JPP-29.1-012825
Q1215-04	JPP-29.1-012825
Q1215-05	JPP-29.2-012825
Q1215-06	JPP-29.2-012825
Q1215-07	JPP-29.2-012825
Q1215-08	JPP-29.2-012825

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : APPROVED *By Nimisha Pandya, QA/QC Supervisor at 11:48 am, Feb 13, 2025* Date: 2/6/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

RU2 Engineering, LLC
Project Name: NYCDDC SANTWOBR Brooklyn Bridge BBMCR
Project # N/A
Chemtech Project # Q1215
Test Name: TCLP VOA

A. Number of Samples and Date of Receipt:
8 Solid samples were received on 01/29/2025.

B. Parameters
According to the Chain of Custody document, the following analyses were requested: Corrosivity, Diesel Range Organics, Gasoline Range Organics, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, Paint Filter, PCB, Pesticide-TCL, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL and VOCMS Group1. This data package contains results for TCLP VOA.

C. Analytical Techniques:
The analysis performed on instrument MSVOA_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868.The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UIThe analysis of TCLP VOA was based on method 8260D and TCLP extraction method was 1311.

D. QA/ QC Samples:
The Holding Times were met for all analysis.
The Surrogate recoveries met the acceptable criteria.
The Internal Standards Areas met the acceptable requirements.
The Retention Times were acceptable for all samples.
The RPD met criteria .
The Blank Spike met requirements for all samples .
The Blank Spike Duplicate met requirements for all samples .
The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements .
The Continuous Calibration met the requirements .
The Tuning criteria met requirements.

E. Additional Comments:
Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.
Trip Blank was not provided with this set of samples.

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:
Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____ **APPROVED**
By Nimisha Pandya, QA/QC Supervisor at 11:48 am, Feb 13, 2025

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

QA Control # A3040960

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1215 MATRIX: TCLP

METHOD: 8260D/1311

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all samples . The Blank Spike Duplicate met requirements for all samples .			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

REVIEWED
By Sohil Jodhani, QA/QC Director at 10:53 am, Feb 13, 2025

Date

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1215

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: SOHIL JODHANI **Date:** 02/06/2025

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q1215	**OrderDate:**	1/29/2025 11:20:00 AM
Client:	RU2 Engineering, LLC	**Project:**	NYCDDC SANTWOBR Brooklyn Bridge BBMCR
Contact:	Rutu Manani	**Location:**	E11,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1215-02	**JPP-29.1-012825**	**TCLP**			**01/28/25**			**01/29/25**
			TCLP VOA	8260D			01/31/25	
Q1215-06	**JPP-29.2-012825**	**TCLP**			**01/28/25**			**01/29/25**
			TCLP VOA	8260D			02/03/25	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1215

Client: RU2 Engineering, LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:								
				0				

Total Voc :

Total Concentration:

QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits Low	Limits High
Q1215-02	JPP-29.1-012825	1,2-Dichloroethane-d4	50	60.9	122	74	125
		Dibromofluoromethane	50	50.0	100	75	124
		Toluene-d8	50	51.9	104	86	113
		4-Bromofluorobenzene	50	41.7	83	77	121
Q1215-06	JPP-29.2-012825	1,2-Dichloroethane-d4	50	50.1	100	74	125
		Dibromofluoromethane	50	50.1	100	75	124
		Toluene-d8	50	55.2	110	86	113
		4-Bromofluorobenzene	50	52.5	105	77	121
VN0131WBL01	VN0131WBL01	1,2-Dichloroethane-d4	50	57.4	115	74	125
		Dibromofluoromethane	50	54.1	108	75	124
		Toluene-d8	50	49.7	99	86	113
		4-Bromofluorobenzene	50	45.4	91	77	121
VN0131WBS01	VN0131WBS01	1,2-Dichloroethane-d4	50	50.3	101	74	125
		Dibromofluoromethane	50	52.5	105	75	124
		Toluene-d8	50	53.6	107	86	113
		4-Bromofluorobenzene	50	50.8	102	77	121
VN0131WBSD0	VN0131WBSD01	1,2-Dichloroethane-d4	50	53.2	106	74	125
		Dibromofluoromethane	50	53.2	106	75	124
		Toluene-d8	50	52.1	104	86	113
		4-Bromofluorobenzene	50	53.2	106	77	121
VX0203WBL01	VX0203WBL01	1,2-Dichloroethane-d4	50	48.6	97	74	125
		Dibromofluoromethane	50	50.1	100	75	124
		Toluene-d8	50	55.2	110	86	113
		4-Bromofluorobenzene	50	52.9	106	77	121
VX0203WBS01	VX0203WBS01	1,2-Dichloroethane-d4	50	43.1	86	74	125
		Dibromofluoromethane	50	46.5	93	75	124
		Toluene-d8	50	50.2	100	86	113
		4-Bromofluorobenzene	50	47.5	95	77	121

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: SW8260D **Datafile :** VN085593.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	RPD
VN0131WBS01	Vinyl chloride	20	16.8	ug/L	84			65	117	
	1,1-Dichloroethene	20	17.0	ug/L	85			74	110	
	2-Butanone	100	97.0	ug/L	97			65	122	
	Carbon Tetrachloride	20	19.0	ug/L	95			77	113	
	Chloroform	20	19.2	ug/L	96			79	113	
	Benzene	20	18.9	ug/L	95			82	109	
	1,2-Dichloroethane	20	19.5	ug/L	98			80	115	
	Trichloroethene	20	18.3	ug/L	92			77	113	
	Tetrachloroethene	20	20.2	ug/L	101			67	123	
	Chlorobenzene	20	18.9	ug/L	95			82	109	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: SW8260D **Datafile :** VN085594.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	RPD
VN0131WBSD01	Vinyl chloride	20	17.2	ug/L	86	2		65	117	20
	1,1-Dichloroethene	20	18.1	ug/L	91	7		74	110	20
	2-Butanone	100	110	ug/L	110	13		65	122	20
	Carbon Tetrachloride	20	19.3	ug/L	97	2		77	113	20
	Chloroform	20	20.8	ug/L	104	8		79	113	20
	Benzene	20	19.4	ug/L	97	2		82	109	20
	1,2-Dichloroethane	20	21.3	ug/L	106	8		80	115	20
	Trichloroethene	20	18.9	ug/L	95	3		77	113	20
	Tetrachloroethene	20	20.6	ug/L	103	2		67	123	20
	Chlorobenzene	20	19.9	ug/L	100	5		82	109	20

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: SW8260D **Datafile :** VX044805.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	RPD
VX0203WBS01	Vinyl chloride	20	20.2	ug/L	101			65	117	
	1,1-Dichloroethene	20	16.8	ug/L	84			74	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	17.6	ug/L	88			77	113	
	Chloroform	20	19.3	ug/L	97			79	113	
	Benzene	20	20.7	ug/L	104			82	109	
	1,2-Dichloroethane	20	18.0	ug/L	90			80	115	
	Trichloroethene	20	19.4	ug/L	97			77	113	
	Tetrachloroethene	20	18.9	ug/L	95			67	123	
	Chlorobenzene	20	19.9	ug/L	100			82	109	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0131WBL01

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Lab File ID: VN085592.D Lab Sample ID: VN0131WBL01

Date Analyzed: 01/31/2025 Time Analyzed: 10:42

GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0131WBS01	VN0131WBS01	VN085593.D	01/31/2025
VN0131WBSD01	VN0131WBSD01	VN085594.D	01/31/2025
JPP-29.1-012825	Q1215-02	VN085614.D	01/31/2025

COMMENTS:

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0203WBL01

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Lab File ID: VX044804.D Lab Sample ID: VX0203WBL01

Date Analyzed: 02/03/2025 Time Analyzed: 10:44

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0203WBS01	VX0203WBS01	VX044805.D	02/03/2025
JPP-29.2-012825	Q1215-06	VX044808.D	02/03/2025

COMMENTS:

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Lab File ID: VN085437.D BFB Injection Date: 01/14/2025

Instrument ID: MSVOA_N BFB Injection Time: 14:22

GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.3
75	30.0 - 60.0% of mass 95	58
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.8
173	Less than 2.0% of mass 174	1.4 (1.8) 1
174	50.0 - 100.0% of mass 95	76
175	5.0 - 9.0% of mass 174	5.4 (7.1) 1
176	95.0 - 101.0% of mass 174	74.1 (97.4) 1
177	5.0 - 9.0% of mass 176	4.9 (6.7) 2

1-Value is % mass 174 2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN085438.D	01/14/2025	14:56
VSTDICCC050	VSTDICCC050	VN085439.D	01/14/2025	15:19
VSTDICC020	VSTDICC020	VN085440.D	01/14/2025	15:43
VSTDICC010	VSTDICC010	VN085441.D	01/14/2025	16:07
VSTDICC005	VSTDICC005	VN085442.D	01/14/2025	16:31
VSTDICC001	VSTDICC001	VN085443.D	01/14/2025	17:19

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Lab File ID: VN085590.D BFB Injection Date: 01/31/2025

Instrument ID: MSVOA_N BFB Injection Time: 09:30

GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.8
75	30.0 - 60.0% of mass 95	53.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.2
173	Less than 2.0% of mass 174	1 (1.4) 1
174	50.0 - 100.0% of mass 95	76.1
175	5.0 - 9.0% of mass 174	6.3 (8.3) 1
176	95.0 - 101.0% of mass 174	73.9 (97.1) 1
177	5.0 - 9.0% of mass 176	4.5 (6) 2

1-Value is % mass 174 2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN085591.D	01/31/2025	10:04
VN0131WBL01	VN0131WBL01	VN085592.D	01/31/2025	10:42
VN0131WBS01	VN0131WBS01	VN085593.D	01/31/2025	11:06
VN0131WBSD01	VN0131WBSD01	VN085594.D	01/31/2025	11:39
JPP-29.1-012825	Q1215-02	VN085614.D	01/31/2025	20:05

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Lab File ID: VX044647.D BFB Injection Date: 01/15/2025

Instrument ID: MSVOA_X BFB Injection Time: 09:30

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.6
75	30.0 - 60.0% of mass 95	50.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.6 (0.9) 1
174	50.0 - 100.0% of mass 95	67.1
175	5.0 - 9.0% of mass 174	4.9 (7.3) 1
176	95.0 - 101.0% of mass 174	64.4 (96) 1
177	5.0 - 9.0% of mass 176	4.3 (6.8) 2

1-Value is % mass 174 2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC001	VSTDICC001	VX044648.D	01/15/2025	10:31
VSTDICC005	VSTDICC005	VX044649.D	01/15/2025	11:17
VSTDICC020	VSTDICC020	VX044650.D	01/15/2025	11:40
VSTDICCC050	VSTDICCC050	VX044651.D	01/15/2025	12:02
VSTDICC100	VSTDICC100	VX044652.D	01/15/2025	13:14
VSTDICC150	VSTDICC150	VX044653.D	01/15/2025	13:37

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Lab File ID: VX044801.D BFB Injection Date: 02/03/2025

Instrument ID: MSVOA_X BFB Injection Time: 08:12

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.4
75	30.0 - 60.0% of mass 95	51.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.8 (1) 1
174	50.0 - 100.0% of mass 95	77.6
175	5.0 - 9.0% of mass 174	5.9 (7.6) 1
176	95.0 - 101.0% of mass 174	74.5 (96.1) 1
177	5.0 - 9.0% of mass 176	4.9 (6.6) 2

1-Value is % mass 174 2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX044802.D	02/03/2025	09:52
VX0203WBL01	VX0203WBL01	VX044804.D	02/03/2025	10:44
VX0203WBS01	VX0203WBS01	VX044805.D	02/03/2025	11:07
JPP-29.2-012825	Q1215-06	VX044808.D	02/03/2025	12:18

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Lab File ID: VN085591.D Date Analyzed: 01/31/2025

Instrument ID: MSVOA_N Time Analyzed: 10:04

GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	190291	8.22	311950	9.10	284481	11.87
UPPER LIMIT	380582	8.724	623900	9.6	568962	12.365
LOWER LIMIT	95145.5	7.724	155975	8.6	142241	11.365
EPA SAMPLE NO.						
JPP-29.1-012825	150642	8.22	317464	9.10	287869	11.87
VN0131WBL01	163540	8.22	304758	9.10	267449	11.87
VN0131WBS01	192673	8.22	320848	9.10	286465	11.87
VN0131WBSD01	170842	8.22	291614	9.10	260387	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Lab File ID: VN085591.D Date Analyzed: 01/31/2025

Instrument ID: MSVOA_N Time Analyzed: 10:04

GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	145646	13.788				
UPPER LIMIT	291292	14.288				
LOWER LIMIT	72823	13.288				
EPA SAMPLE NO.						
JPP-29.1-012825	92759	13.79				
VN0131WBL01	104737	13.79				
VN0131WBS01	135464	13.79				
VN0131WBSD01	125242	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Lab File ID: VX044802.D Date Analyzed: 02/03/2025

Instrument ID: MSVOA_X Time Analyzed: 09:52

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	162430	5.54	297715	6.75	252415	10.05
UPPER LIMIT	324860	6.038	595430	7.251	504830	10.549
LOWER LIMIT	81215	5.038	148858	6.251	126208	9.549
EPA SAMPLE NO.						
JPP-29.2-012825	118450	5.54	245754	6.76	219935	10.05
VX0203WBL01	124993	5.54	257688	6.76	232979	10.05
VX0203WBS01	151779	5.54	276067	6.76	242310	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Lab File ID: VX044802.D Date Analyzed: 02/03/2025

Instrument ID: MSVOA_X Time Analyzed: 09:52

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	101849	12.018				
UPPER LIMIT	203698	12.518				
LOWER LIMIT	50924.5	11.518				
EPA SAMPLE NO.						
JPP-29.2-012825	94615	12.02				
VX0203WBL01	100905	12.02				
VX0203WBS01	107943	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

 # Column used to flag values outside QC limits with an asterisk.
*** Values outside of QC limits.**

SAMPLE DATA

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/28/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/29/25
Client Sample ID:	JPP-29.1-012825	SDG No.:	Q1215
Lab Sample ID:	Q1215-02	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085614.D	1		01/31/25 20:05	VN013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5.00	U	0.34	5.00	ug/L
75-35-4	1,1-Dichloroethene	5.00	U	0.26	5.00	ug/L
78-93-3	2-Butanone	25.0	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	5.00	U	0.25	5.00	ug/L
67-66-3	Chloroform	5.00	U	0.26	5.00	ug/L
71-43-2	Benzene	5.00	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	5.00	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	5.00	U	0.32	5.00	ug/L
127-18-4	Tetrachloroethene	5.00	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	5.00	U	0.13	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	60.9		74 - 125	122%	SPK: 50
1868-53-7	Dibromofluoromethane	50.0		75 - 124	100%	SPK: 50
2037-26-5	Toluene-d8	51.9		86 - 113	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	41.7		77 - 121	83%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	151000	8.224			
540-36-3	1,4-Difluorobenzene	317000	9.1			
3114-55-4	Chlorobenzene-d5	288000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	92800	13.794			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085614.D
Acq On : 31 Jan 2025 20:05
Operator : JC\MD
Sample : Q1215-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-29.1-012825

Quant Time: Jan 31 22:35:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	150642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	317464	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	287869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	92759	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	148102	60.905	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 121.820%			
35) Dibromofluoromethane	8.165	113	110123	50.001	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.000%			
50) Toluene-d8	10.565	98	405836	51.863	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.720%			
62) 4-Bromofluorobenzene	12.847	95	111608	41.695	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 83.380%			
Target Compounds						Qvalue
16) Acetone	4.436	43	34560	43.987	ug/l	92
20) Methylene Chloride	5.271	84	2625	1.350	ug/l #	88
43) Isopropyl Acetate	8.688	43	211916	42.389	ug/l #	79


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085614.D
Acq On : 31 Jan 2025 20:05
Operator : JC\MD
Sample : Q1215-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-29.1-012825

Quant Time: Jan 31 22:35:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Instrument :
MSVOA_N
ClientSampleId :
JPP-29.1-012825

Tgt Ion:168 Resp: 150642

Ion	Ratio	Lower	Upper
168	100		
99	66.8	53.6	80.4







#16
Acetone
Concen: 43.987 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.017 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Tgt Ion: 43 Resp: 34560

Ion	Ratio	Lower	Upper
43	100		
58	25.5	23.8	35.6







#20
Methylene Chloride
Concen: 1.350 ug/l
RT: 5.271 min Scan# 56
Delta R.T. -0.000 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Instrument :
MSVOA_N
ClientSampleId :
JPP-29.1-012825

Tgt Ion: 84 Resp: 2625

Ion	Ratio	Lower	Upper
84	100		
49	120.8	104.1	156.1
51	43.6	31.0	46.4
86	47.7	51.9	77.9#







#33
1,2-Dichloroethane-d4
Concen: 60.905 ug/l
RT: 8.576 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Tgt Ion: 65 Resp: 148102

Ion	Ratio	Lower	Upper
65	100		
67	51.3	0.0	101.6







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 12
Delta R.T. 0.006 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Instrument :
MSVOA_N
ClientSampleId :
JPP-29.1-012825

Tgt Ion:114 Resp: 317464

Ion	Ratio	Lower	Upper
114	100		
63	21.5	0.0	47.6
88	15.7	0.0	32.6







#35
Dibromofluoromethane
Concen: 50.001 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Tgt Ion:113 Resp: 110123

Ion	Ratio	Lower	Upper
113	100		
111	103.4	82.7	124.1
192	17.8	14.3	21.5







#43
Isopropyl Acetate
Concen: 42.389 ug/l
RT: 8.688 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Instrument :
MSVOA_N
ClientSampleId :
JPP-29.1-012825

Tgt Ion: 43 Resp: 211916

Ion	Ratio	Lower	Upper
43	100		
61	12.9	20.7	31.1#
87	8.3	9.8	14.8#







#50
Toluene-d8
Concen: 51.863 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Tgt Ion: 98 Resp: 405836

Ion	Ratio	Lower	Upper
98	100		
100	63.5	52.2	78.4







#62
4-Bromofluorobenzene
Concen: 41.695 ug/l
RT: 12.847 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Instrument :
MSVOA_N
ClientSampleId :
JPP-29.1-012825

Tgt Ion: 95 Resp: 111608

Ion	Ratio	Lower	Upper
95	100		
174	74.5	0.0	145.0
176	70.5	0.0	142.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Tgt Ion:117 Resp: 287869

Ion	Ratio	Lower	Upper
117	100		
82	59.0	48.6	72.8
119	32.3	26.6	39.8






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN085614.D
Acq: 31 Jan 2025 20:05

Instrument :
MSVOA_N
ClientSampleId :
JPP-29.1-012825

Tgt Ion:152 Resp: 92759

Ion	Ratio	Lower	Upper
152	100		
115	62.6	31.1	93.3
150	159.4	0.0	343.6



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	01/28/25
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	01/29/25
Client Sample ID:	JPP-29.2-012825	SDG No.:	Q1215
Lab Sample ID:	Q1215-06	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044808.D	1		02/03/25 12:18	VX020325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5.00	U	0.34	5.00	ug/L
75-35-4	1,1-Dichloroethene	5.00	U	0.26	5.00	ug/L
78-93-3	2-Butanone	25.0	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	5.00	U	0.25	5.00	ug/L
67-66-3	Chloroform	5.00	U	0.26	5.00	ug/L
71-43-2	Benzene	5.00	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	5.00	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	5.00	U	0.32	5.00	ug/L
127-18-4	Tetrachloroethene	5.00	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	5.00	U	0.13	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.1		74 - 125	100%	SPK: 50
1868-53-7	Dibromofluoromethane	50.1		75 - 124	100%	SPK: 50
2037-26-5	Toluene-d8	55.2		86 - 113	110%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.5		77 - 121	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	118000	5.543			
540-36-3	1,4-Difluorobenzene	246000	6.756			
3114-55-4	Chlorobenzene-d5	220000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	94600	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044808.D
Acq On : 03 Feb 2025 12:18
Operator : JC/MD
Sample : Q1215-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-29.2-012825

Quant Time: Feb 04 02:07:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.543	168	118450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	245754	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	219935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	94615	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89647	50.141	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.280%			
35) Dibromofluoromethane	5.379	113	78267	50.131	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.260%			
50) Toluene-d8	8.646	98	300281	55.161	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.320%			
62) 4-Bromofluorobenzene	11.079	95	106536	52.497	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.000%			
Target Compounds						Qvalue
16) Acetone	2.379	43	31223	44.139	ug/l	97
18) Methyl Acetate	2.708	43	2547	1.167	ug/l	91
20) Methylene Chloride	2.788	84	2800	1.700	ug/l	85
43) Isopropyl Acetate	6.336	43	126423	29.694	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044808.D
Acq On : 03 Feb 2025 12:18
Operator : JC/MD
Sample : Q1215-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-29.2-012825

Quant Time: Feb 04 02:07:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.543 min Scan# 7
Delta R.T. -0.001 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Instrument :
MSVOA_X
ClientSampleId :
JPP-29.2-012825

Tgt Ion:168 Resp: 118450

Ion	Ratio	Lower	Upper
168	100		
99	59.3	51.1	76.7







#16
Acetone
Concen: 44.139 ug/l
RT: 2.379 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Tgt Ion: 43 Resp: 31223

Ion	Ratio	Lower	Upper
43	100		
58	32.4	24.6	37.0







#18
Methyl Acetate
Concen: 1.167 ug/l
RT: 2.708 min Scan# 2
Delta R.T. 0.006 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Instrument :
MSVOA_X
ClientSampleId :
JPP-29.2-012825

Tgt Ion: 43 Resp: 2547

Ion	Ratio	Lower	Upper
43	100		
74	20.5	19.8	29.8







#20
Methylene Chloride
Concen: 1.700 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.006 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Tgt Ion: 84 Resp: 2800

Ion	Ratio	Lower	Upper
84	100		
49	94.9	94.2	141.4
51	31.6	28.8	43.2
86	56.8	50.3	75.5







#33
1,2-Dichloroethane-d4
Concen: 50.141 ug/l
RT: 5.952 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Instrument :
MSVOA_X
ClientSampleId :
JPP-29.2-012825

Tgt Ion: 65 Resp: 89647

Ion	Ratio	Lower	Upper
65	100		
67	53.8	0.0	103.4







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.756 min Scan# 930
Delta R.T. -0.000 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Tgt Ion:114 Resp: 245754

Ion	Ratio	Lower	Upper
114	100		
63	21.0	0.0	43.0
88	16.0	0.0	35.4







#35
Dibromofluoromethane
Concen: 50.131 ug/l
RT: 5.379 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Instrument :
MSVOA_X
ClientSampleId :
JPP-29.2-012825

Tgt Ion:113 Resp: 78267

Ion	Ratio	Lower	Upper
113	100		
111	103.9	81.6	122.4
192	18.0	14.4	21.6







#43
Isopropyl Acetate
Concen: 29.694 ug/l
RT: 6.336 min Scan# 861
Delta R.T. -0.000 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Tgt Ion: 43 Resp: 126423

Ion	Ratio	Lower	Upper
43	100		
61	20.3	16.6	24.8
87	13.4	11.0	16.6







#50
Toluene-d8
Concen: 55.161 ug/l
RT: 8.646 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Instrument :
MSVOA_X
ClientSampleId :
JPP-29.2-012825

Tgt Ion: 98 Resp: 300281

Ion	Ratio	Lower	Upper
98	100		
100	65.9	53.2	79.8







#62
4-Bromofluorobenzene
Concen: 52.497 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Tgt Ion: 95 Resp: 106536

Ion	Ratio	Lower	Upper
95	100		
174	70.7	0.0	133.0
176	69.2	0.0	129.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Instrument :
MSVOA_X
ClientSampleId :
JPP-29.2-012825

Tgt Ion:117 Resp: 219935

Ion	Ratio	Lower	Upper
117	100		
82	58.1	47.6	71.4
119	32.3	25.4	38.2






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX044808.D
Acq: 03 Feb 2025 12:18

Tgt Ion:152 Resp: 94615

Ion	Ratio	Lower	Upper
152	100		
115	62.7	43.8	131.4
150	161.2	0.0	347.8



CALIBRATION SUMMARY

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG No.: Q1215

Instrument ID: MSVOA_N Calibration Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Calibration Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID: RRF100 = VN085438.D RRF050 = VN085439.D RRF020 = VN085440.D
RRF010 = VN085441.D RRF005 = VN085442.D RRF001 = VN085443.D

COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	$\overline{RRF}$	% RSD
Vinyl Chloride	0.697	0.686	0.727	0.711	0.781	0.819	0.737	7.1
1,1-Dichloroethene	0.548	0.533	0.556	0.526	0.559	0.497	0.537	4.3
2-Butanone	0.378	0.390	0.398	0.363	0.387	0.386	0.384	3.1
Carbon Tetrachloride	0.574	0.530	0.579	0.529	0.565	0.567	0.557	4
Chloroform	1.197	1.175	1.241	1.169	1.253	1.273	1.218	3.6
Benzene	1.551	1.449	1.527	1.376	1.474	1.400	1.463	4.7
1,2-Dichloroethane	0.569	0.547	0.575	0.522	0.574	0.517	0.551	4.8
Trichloroethene	0.362	0.324	0.352	0.310	0.343	0.352	0.341	5.8
Tetrachloroethene	0.351	0.322	0.365	0.338	0.346	0.323	0.341	4.9
Chlorobenzene	1.133	1.076	1.154	1.047	1.110	1.051	1.095	4
1,2-Dichloroethane-d4	0.774	0.831	0.754	0.762	0.914		0.807	8.3
Dibromofluoromethane	0.359	0.358	0.335	0.310	0.373		0.347	7.1
Toluene-d8	1.339	1.267	1.207	1.076	1.274		1.232	8.1
4-Bromofluorobenzene	0.475	0.449	0.410	0.357	0.417		0.422	10.6

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Response Factor Report MSVOA_N_NEW

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
Method File : 82N011425W.M
Title : SW846 8260
Last Update : Wed Jan 15 02:16:08 2025
Response Via : Initial Calibration

Calibration Files
1 =VN085443.D 5 =VN085442.D 10 =VN085441.D 20 =VN085440.D 50 =VN085439.D 100 =VN085438.D

		Compound	1	5	10	20	50	100	Avg	%RSD
1)	I	Pentafluorobenzene	----------ISTD----------							
2)	T	Dichlorodifluo...	0.714	0.708	0.667	0.681	0.629	0.664	0.677	4.59
3)	P	Chloromethane	0.839	0.779	0.693	0.727	0.680	0.680	0.733	8.77
4)	C	Vinyl Chloride	0.819	0.781	0.711	0.727	0.686	0.697	0.737	7.06#
5)	T	Bromomethane		0.525	0.437	0.454	0.417	0.392	0.445	11.26
6)	T	Chloroethane	0.542	0.505	0.429	0.468	0.424	0.435	0.467	10.27
7)	T	Trichlorofluor...	1.157	1.077	1.040	1.097	0.997	1.046	1.069	5.14
8)	T	Diethyl Ether	0.457	0.357	0.317	0.363	0.360	0.362	0.369	12.59
9)	T	1,1,2-Trichlor...	0.646	0.639	0.587	0.609	0.542	0.590	0.602	6.35
10)	T	Methyl Iodide		0.661	0.640	0.704	0.719	0.723	0.690	5.35
11)	T	Tert butyl alc...		0.097	0.092	0.095	0.093	0.085	0.092	4.80
12)	CM	1,1-Dichloroet...	0.497	0.559	0.526	0.556	0.533	0.548	0.537	4.31#
13)	T	Acrolein		0.101	0.131	0.124	0.140	0.136	0.126	12.09
14)	T	Allyl chloride	0.952	0.897	0.823	0.843	0.846	0.863	0.871	5.42
15)	T	Acrylonitrile	0.298	0.299	0.275	0.303	0.298	0.289	0.294	3.39
16)	T	Acetone	0.306	0.269	0.247	0.252	0.252	0.238	0.261	9.28
17)	T	Carbon Disulfide	1.978	1.719	1.537	1.647	1.477	1.555	1.652	10.95
18)	T	Methyl Acetate	0.871	0.810	0.779	0.758	0.790	0.751	0.793	5.54
19)	T	Methyl tert-bu...	1.545	1.685	1.664	1.853	1.873	1.834	1.742	7.53
20)	T	Methylene Chlo...	0.656	0.696	0.606	0.658	0.629	0.629	0.646	4.86
21)	T	trans-1,2-Dich...	0.632	0.569	0.539	0.574	0.555	0.571	0.573	5.50
22)	T	Diisopropyl ether	1.705	1.918	1.834	2.076	2.037	2.026	1.933	7.38
23)	T	Vinyl Acetate	1.110	1.270	1.279	1.466	1.495	1.495	1.353	11.67
24)	P	1,1-Dichloroet...	1.204	1.226	1.100	1.206	1.170	1.164	1.178	3.81
25)	T	2-Butanone	0.386	0.387	0.363	0.398	0.390	0.378	0.384	3.12
26)	T	2,2-Dichloropr...	0.930	0.986	0.922	0.985	0.936	0.958	0.953	2.96
27)	T	cis-1,2-Dichlo...	0.655	0.669	0.639	0.715	0.683	0.691	0.675	4.02
28)	T	Bromochloromet...	0.624	0.595	0.486	0.513	0.542	0.530	0.548	9.44
29)	T	Tetrahydrofuran	0.221	0.236	0.233	0.261	0.260	0.248	0.243	6.58
30)	C	Chloroform	1.273	1.253	1.169	1.241	1.175	1.197	1.218	3.59#
31)	T	Cyclohexane		1.198	1.026	1.033	0.881	0.984	1.024	11.18
32)	T	1,1,1-Trichlor...	1.102	1.148	1.000	1.091	1.016	1.053	1.068	5.24
33)	S	1,2-Dichloroet...		0.914	0.762	0.754	0.831	0.774	0.807	8.28
34)	I	1,4-Difluorobenzene	----------ISTD----------							
35)	S	Dibromofluorom...		0.373	0.310	0.335	0.358	0.359	0.347	7.10
36)	T	1,1-Dichloropr...	0.509	0.490	0.441	0.499	0.462	0.521	0.487	6.17
37)	T	Ethyl Acetate	0.504	0.520	0.452	0.488	0.488	0.496	0.491	4.66
38)	T	Carbon Tetrach...	0.567	0.565	0.529	0.579	0.530	0.574	0.557	3.96
39)	T	Methylcyclohexane	0.397	0.437	0.407	0.477	0.463	0.564	0.457	13.29
40)	TM	Benzene	1.400	1.474	1.376	1.527	1.449	1.551	1.463	4.70
41)	T	Methacrylonitrile	0.241	0.227	0.248	0.268	0.270	0.280	0.256	7.89
42)	TM	1,2-Dichloroet...	0.517	0.574	0.522	0.575	0.547	0.569	0.551	4.78
43)	T	Isopropyl Acetate	0.793	0.764	0.719	0.801	0.813	0.835	0.787	5.20
44)	TM	Trichloroethene	0.352	0.343	0.310	0.352	0.324	0.362	0.341	5.79
45)	C	1,2-Dichloropr...	0.371	0.388	0.334	0.388	0.371	0.390	0.374	5.69#
46)	T	Dibromomethane	0.292	0.254	0.254	0.277	0.267	0.275	0.270	5.50
47)	T	Bromodichlorom...	0.484	0.569	0.514	0.579	0.559	0.590	0.549	7.48
48)	T	Methyl methacr...	0.311	0.326	0.321	0.369	0.392	0.406	0.354	11.37
49)	T	1,4-Dioxane	0.006	0.006	0.006	0.006	0.006	0.006	0.006	6.72
50)	S	Toluene-d8		1.274	1.076	1.207	1.267	1.339	1.232	8.07
51)	T	4-Methyl-2-Pen...	0.380	0.443	0.432	0.495	0.492	0.499	0.457	10.33
52)	CM	Toluene	0.690	0.835	0.808	0.919	0.870	0.964	0.848	11.27#
53)	T	t-1,3-Dichloro...	0.416	0.527	0.481	0.544	0.551	0.594	0.519	11.98
54)	T	cis-1,3-Dichlo...	0.450	0.538	0.527	0.601	0.588	0.623	0.554	11.40
55)	T	1,1,2-Trichlor...	0.314	0.349	0.309	0.353	0.340	0.348	0.335	5.67

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
Method File : 82N011425W.M

56)	T	Ethyl methacry...	0.317	0.434	0.418	0.521	0.544	0.582	0.469	20.85
57)	T	1,3-Dichloropr...	0.514	0.584	0.555	0.618	0.602	0.627	0.583	7.31
58)	T	2-Chloroethyl ...	0.163	0.205	0.185	0.226	0.241	0.258	0.213	16.67
59)	T	2-Hexanone	0.261	0.302	0.298	0.353	0.357	0.358	0.321	12.60
60)	T	Dibromochlorom...	0.386	0.420	0.368	0.412	0.414	0.430	0.405	5.79
61)	T	1,2-Dibromoethane	0.334	0.321	0.309	0.356	0.334	0.349	0.334	5.16
62)	S	4-Bromofluorob...		0.417	0.357	0.410	0.449	0.475	0.422	10.57
63)	I	Chlorobenzene-d5	----------ISTD----------							
64)	T	Tetrachloroethene	0.323	0.346	0.338	0.365	0.322	0.351	0.341	4.87
65)	PM	Chlorobenzene	1.051	1.110	1.047	1.154	1.076	1.133	1.095	4.03
66)	T	1,1,1,2-Tetrac...	0.425	0.412	0.371	0.408	0.392	0.404	0.402	4.65
67)	C	Ethyl Benzene	1.430	1.709	1.685	1.940	1.867	2.072	1.784	12.67#
68)	T	m/p-Xylenes	0.492	0.616	0.615	0.750	0.707	0.775	0.659	16.01
69)	T	o-Xylene	0.482	0.582	0.584	0.713	0.681	0.738	0.630	15.47
70)	T	Styrene	0.742	0.929	0.956	1.186	1.173	1.271	1.043	19.21
71)	P	Bromoform	0.235	0.284	0.273	0.312	0.311	0.311	0.288	10.61
72)	I	1,4-Dichlorobenzen...	----------ISTD----------							
73)	T	Isopropylbenzene	2.766	3.157	3.272	3.681	3.448	3.922	3.375	12.06
74)	T	N-amyl acetate	1.364	1.394	1.405	1.642	1.622	1.673	1.517	9.43
75)	P	1,1,2,2-Tetrac...	1.314	1.228	1.157	1.187	1.145	1.121	1.192	5.89
76)	T	1,2,3-Trichlor...	1.104	0.965	0.957	1.118	0.893	1.061	1.016	8.96
77)	T	Bromobenzene	0.871	0.880	0.857	0.906	0.858	0.919	0.882	2.90
78)	T	n-propylbenzene	3.227	3.605	3.846	4.400	4.154	4.738	3.995	13.74
79)	T	2-Chlorotoluene	2.252	2.533	2.571	2.761	2.585	2.812	2.586	7.67
80)	T	1,3,5-Trimethy...	2.160	2.529	2.695	3.140	2.940	3.260	2.787	14.69
81)	T	trans-1,4-Dich...		0.335	0.347	0.386	0.390	0.419	0.376	9.04
82)	T	4-Chlorotoluene	2.144	2.536	2.512	2.820	2.588	2.847	2.574	9.91
83)	T	tert-Butylbenzene	1.885	2.189	2.236	2.570	2.439	2.726	2.341	12.85
84)	T	1,2,4-Trimethy...	1.888	2.542	2.710	3.190	3.016	3.323	2.778	18.89
85)	T	sec-Butylbenzene	2.318	3.026	3.147	3.643	3.421	3.912	3.244	17.17
86)	T	p-Isopropyltol...	1.777	2.368	2.541	2.942	2.872	3.286	2.631	20.04
87)	T	1,3-Dichlorobe...	1.526	1.656	1.574	1.701	1.565	1.720	1.624	4.91
88)	T	1,4-Dichlorobe...	1.767	1.743	1.607	1.713	1.562	1.706	1.683	4.78
89)	T	n-Butylbenzene	1.772	2.176	2.168	2.485	2.483	2.881	2.327	16.22
90)	T	Hexachloroethane	0.681	0.616	0.574	0.613	0.577	0.645	0.618	6.59
91)	T	1,2-Dichlorobe...	1.766	1.600	1.532	1.654	1.555	1.611	1.620	5.15
92)	T	1,2-Dibromo-3-...	0.202	0.228	0.212	0.224	0.222	0.218	0.218	4.31
93)	T	1,2,4-Trichlor...	0.658	0.717	0.704	0.799	0.781	0.858	0.753	9.69
94)	T	Hexachlorobuta...	0.413	0.441	0.395	0.396	0.367	0.386	0.400	6.31
95)	T	Naphthalene	2.115	1.987	1.958	2.348	2.482	2.588	2.246	11.79
96)	T	1,2,3-Trichlor...	0.750	0.693	0.732	0.792	0.786	0.817	0.762	5.96

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085438.D
Acq On : 14 Jan 2025 14:56
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jan 15 01:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.218	168	202613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	329363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	298938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	146196	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	313645	95.899	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 191.800%#			
35) Dibromofluoromethane	8.159	113	236159	103.353	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 206.700%#			
50) Toluene-d8	10.565	98	881956	108.636	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 217.280%#			
62) 4-Bromofluorobenzene	12.847	95	313161	112.765	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 225.520%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.124	85	269023	98.065	ug/l	99
3) Chloromethane	2.359	50	275415	92.724	ug/l	98
4) Vinyl Chloride	2.512	62	282361	94.576	ug/l	99
5) Bromomethane	2.942	94	158888	88.105	ug/l	98
6) Chloroethane	3.112	64	176111	93.043	ug/l	99
7) Trichlorofluoromethane	3.494	101	423967	97.870	ug/l	100
8) Diethyl Ether	3.953	74	146889	98.154	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.371	101	239104	97.997	ug/l	99
10) Methyl Iodide	4.589	142	292983	104.856	ug/l	95
11) Tert butyl alcohol	5.512	59	173044	462.061	ug/l	99
12) 1,1-Dichloroethene	4.336	96	222117	102.146	ug/l	99
13) Acrolein	4.171	56	275350	538.600	ug/l	99
14) Allyl chloride	5.018	41	349717	99.111	ug/l	98
15) Acrylonitrile	5.712	53	585370	492.095	ug/l	99
16) Acetone	4.418	43	483221	457.268	ug/l	96
17) Carbon Disulfide	4.712	76	630294	94.140	ug/l	100
18) Methyl Acetate	5.012	43	304226	94.674	ug/l	99
19) Methyl tert-butyl Ether	5.788	73	743008	105.230	ug/l	100
20) Methylene Chloride	5.271	84	254912	97.440	ug/l	98
21) trans-1,2-Dichloroethene	5.783	96	231503	99.618	ug/l	97
22) Diisopropyl ether	6.665	45	820871	104.812	ug/l	99
23) Vinyl Acetate	6.594	43	3029568	550.850	ug/l	99
24) 1,1-Dichloroethane	6.565	63	471831	98.803	ug/l	99
25) 2-Butanone	7.477	43	764979	491.912	ug/l	98
26) 2,2-Dichloropropane	7.482	77	388140	100.530	ug/l	100
27) cis-1,2-Dichloroethene	7.482	96	280197	102.367	ug/l	99
28) Bromochloromethane	7.806	49	214796	96.679	ug/l	100
29) Tetrahydrofuran	7.830	42	503459	510.635	ug/l	99
30) Chloroform	7.959	83	484941	98.253	ug/l	100
31) Cyclohexane	8.253	56	398641	96.041	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	426782	98.578	ug/l	98
36) 1,1-Dichloropropene	8.365	75	342961	106.945	ug/l	100
37) Ethyl Acetate	7.553	43	326802	100.961	ug/l	97
38) Carbon Tetrachloride	8.359	117	378390	103.065	ug/l	97
39) Methylcyclohexane	9.594	83	371734	123.354	ug/l	99
40) Benzene	8.600	78	1021393	106.001	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085438.D
Acq On : 14 Jan 2025 14:56
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jan 15 01:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc Units		Dev(Min)
41) Methacrylonitrile	7.777	41	184326	109.454 ug/l		96
42) 1,2-Dichloroethane	8.665	62	374938	103.314 ug/l		100
43) Isopropyl Acetate	8.682	43	549969	106.035 ug/l		100
44) Trichloroethene	9.347	130	238296	106.242 ug/l		96
45) 1,2-Dichloropropane	9.618	63	256964	104.405 ug/l		97
46) Dibromomethane	9.706	93	180990	101.913 ug/l		99
47) Bromodichloromethane	9.882	83	388499	107.377 ug/l		98
48) Methyl methacrylate	9.676	41	267689	114.685 ug/l		99
49) 1,4-Dioxane	9.688	88	83879	2133.992 ug/l		97
51) 4-Methyl-2-Pentanone	10.441	43	1643800	546.159 ug/l		100
52) Toluene	10.623	92	634743	113.689 ug/l		99
53) t-1,3-Dichloropropene	10.829	75	391238	114.424 ug/l		97
54) cis-1,3-Dichloropropene	10.306	75	410403	112.369 ug/l		99
55) 1,1,2-Trichloroethane	11.012	97	229522	103.879 ug/l		98
56) Ethyl methacrylate	10.870	69	383663	101.029 ug/l		99
57) 1,3-Dichloropropane	11.159	76	412886	107.469 ug/l		100
58) 2-Chloroethyl Vinyl ether	10.153	63	848636	605.243 ug/l		100
59) 2-Hexanone	11.194	43	1178831	556.643 ug/l		99
60) Dibromochloromethane	11.353	129	283361	106.227 ug/l		99
61) 1,2-Dibromoethane	11.465	107	230104	104.623 ug/l		98
64) Tetrachloroethene	11.100	164	209870	102.980 ug/l		98
65) Chlorobenzene	11.888	112	677679	103.482 ug/l		98
66) 1,1,1,2-Tetrachloroethane	11.959	131	241593	100.517 ug/l		99
67) Ethyl Benzene	11.959	91	1239080	116.169 ug/l		99
68) m/p-Xylenes	12.070	106	927244	235.219 ug/l		99
69) o-Xylene	12.394	106	441279	117.135 ug/l		100
70) Styrene	12.406	104	760136	121.919 ug/l		99
71) Bromoform	12.576	173	186124	108.227 ug/l	#	99
73) Isopropylbenzene	12.694	105	1146752	116.221 ug/l		100
74) N-amyl acetate	12.488	43	489217	110.308 ug/l		99
75) 1,1,2,2-Tetrachloroethane	12.935	83	327767	94.039 ug/l		100
76) 1,2,3-Trichloropropane	12.988	75	310115m	104.530 ug/l		
77) Bromobenzene	12.976	156	268616	104.200 ug/l		97
78) n-propylbenzene	13.029	91	1385353	118.600 ug/l		100
79) 2-Chlorotoluene	13.123	91	822258	108.767 ug/l		99
80) 1,3,5-Trimethylbenzene	13.170	105	953066	116.938 ug/l		100
81) trans-1,4-Dichloro-2-b...	12.735	75	122508	111.544 ug/l		87
82) 4-Chlorotoluene	13.217	91	832317	110.574 ug/l		99
83) tert-Butylbenzene	13.435	119	797069	116.457 ug/l		98
84) 1,2,4-Trimethylbenzene	13.476	105	971729	119.625 ug/l		100
85) sec-Butylbenzene	13.611	105	1143885	120.587 ug/l		100
86) p-Isopropyltoluene	13.723	119	960747	100.111 ug/l		100
87) 1,3-Dichlorobenzene	13.729	146	502960	105.943 ug/l		99
88) 1,4-Dichlorobenzene	13.811	146	498709	101.341 ug/l		97
89) n-Butylbenzene	14.053	91	842368	123.781 ug/l		98
90) Hexachloroethane	14.329	117	188541	104.405 ug/l		100
91) 1,2-Dichlorobenzene	14.100	146	470919	99.445 ug/l		99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	63846	100.313 ug/l		98
93) 1,2,4-Trichlorobenzene	15.388	180	250787	113.946 ug/l		99
94) Hexachlorobutadiene	15.500	225	112899	96.619 ug/l		99
95) Naphthalene	15.635	128	756816	115.234 ug/l		99
96) 1,2,3-Trichlorobenzene	15.835	180	238860	107.256 ug/l		100

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085438.D
Acq On : 14 Jan 2025 14:56
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085438.D
Acq On : 14 Jan 2025 14:56
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1


Quant Time: Jan 15 01:41:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:168 Resp: 202613


Ion	Ratio	Lower	Upper
168	100		
99	64.9	53.6	80.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#2
Dichlorodifluoromethane
Concen: 98.065 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 85 Resp: 269023

Ion	Ratio	Lower	Upper
85	100		
87	32.7	16.7	50.1






#3
Chloromethane
Concen: 92.724 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 50 Resp: 275415


Ion	Ratio	Lower	Upper
50	100		
52	32.4	27.0	40.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#4
Vinyl Chloride
Concen: 94.576 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 62 Resp: 282361

Ion	Ratio	Lower	Upper
62	100		
64	31.6	24.8	37.2






#5
Bromomethane
Concen: 88.105 ug/l
RT: 2.942 min Scan# 16
Delta R.T. -0.012 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 94 Resp: 158888


Ion	Ratio	Lower	Upper
94	100		
96	91.9	71.8	107.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#6
Chloroethane
Concen: 93.043 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 64 Resp: 176111

Ion	Ratio	Lower	Upper
64	100		
66	31.1	24.6	36.8






#7
Trichlorofluoromethane
Concen: 97.870 ug/l
RT: 3.494 min Scan# 26
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:101 Resp: 423967


Ion	Ratio	Lower	Upper
101	100		
103	64.4	51.4	77.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#8
Diethyl Ether
Concen: 98.154 ug/l
RT: 3.953 min Scan# 340
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 74 Resp: 146889

Ion	Ratio	Lower	Upper
74	100		
45	97.4	49.7	149.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 97.997 ug/l
RT: 4.371 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56


Tgt Ion:101 Resp: 239104


Ion	Ratio	Lower	Upper
101	100		
85	47.1	37.8	56.8
151	71.8	58.8	88.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025







#10
Methyl Iodide
Concen: 104.856 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:142 Resp: 292983

Ion	Ratio	Lower	Upper
142	100		
127	47.1	40.3	60.5
141	13.1	12.8	19.2






#11
Tert butyl alcohol
Concen: 462.061 ug/l
RT: 5.512 min Scan# 60
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 59 Resp: 173044


Ion	Ratio	Lower	Upper
59	100		
57	11.1	8.6	13.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#12
1,1-Dichloroethene
Concen: 102.146 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 96 Resp: 222117

Ion	Ratio	Lower	Upper
96	100		
61	175.3	141.0	211.4
98	62.7	49.0	73.6






#13
Acrolein
Concen: 538.600 ug/l
RT: 4.171 min Scan# 37
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 56 Resp: 275350


Ion	Ratio	Lower	Upper
56	100		
55	71.1	56.3	84.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#14
Allyl chloride
Concen: 99.111 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 41 Resp: 349717

Ion	Ratio	Lower	Upper
41	100		
39	83.3	64.4	96.6
76	37.9	30.5	45.7






#15
Acrylonitrile
Concen: 492.095 ug/l
RT: 5.712 min Scan# 63
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 53 Resp: 585370


Ion	Ratio	Lower	Upper
53	100		
52	83.0	65.5	98.3
51	37.1	29.8	44.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#16
Acetone
Concen: 457.268 ug/l
RT: 4.418 min Scan# 419
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 483221

Ion	Ratio	Lower	Upper
43	100		
58	27.3	23.8	35.6






#17
Carbon Disulfide
Concen: 94.140 ug/l
RT: 4.712 min Scan# 46
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 76 Resp: 630294


Ion	Ratio	Lower	Upper
76	100		
78	8.7	6.9	10.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#18
Methyl Acetate
Concen: 94.674 ug/l
RT: 5.012 min Scan# 520
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 304226

Ion	Ratio	Lower	Upper
43	100		
74	23.6	18.6	28.0






#19
Methyl tert-butyl Ether
Concen: 105.230 ug/l
RT: 5.788 min Scan# 65
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 73 Resp: 743008


Ion	Ratio	Lower	Upper
73	100		
57	22.3	17.8	26.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#20
Methylene Chloride
Concen: 97.440 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 84 Resp: 254912

Ion	Ratio	Lower	Upper
84	100		
49	128.5	104.1	156.1
51	37.8	31.0	46.4
86	61.1	51.9	77.9






#21
trans-1,2-Dichloroethene
Concen: 99.618 ug/l
RT: 5.783 min Scan# 65
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 96 Resp: 231503


Ion	Ratio	Lower	Upper
96	100		
61	148.3	115.7	173.5
98	65.0	49.8	74.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#22
Diisopropyl ether
Concen: 104.812 ug/l
RT: 6.665 min Scan# 801
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 45 Resp: 820871

Ion	Ratio	Lower	Upper
45	100		
43	55.7	45.0	67.6
87	27.2	21.8	32.8
59	11.6	10.2	15.2






#23
Vinyl Acetate
Concen: 550.850 ug/l
RT: 6.594 min Scan# 78
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 3029568


Ion	Ratio	Lower	Upper
43	100		
86	10.1	8.4	12.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#24
1,1-Dichloroethane
Concen: 98.803 ug/l
RT: 6.565 min Scan# 784
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 63 Resp: 471831

Ion	Ratio	Lower	Upper
63	100		
98	6.4	3.3	9.8
100	4.6	2.0	6.0






#25
2-Butanone
Concen: 491.912 ug/l
RT: 7.477 min Scan# 93
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 764979


Ion	Ratio	Lower	Upper
43	100		
72	26.3	20.2	30.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#26
2,2-Dichloropropane
Concen: 100.530 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 77 Resp: 388140

Ion	Ratio	Lower	Upper
77	100		
97	21.4	10.7	32.1






#27
cis-1,2-Dichloroethene
Concen: 102.367 ug/l
RT: 7.482 min Scan# 94
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 96 Resp: 280197


Ion	Ratio	Lower	Upper
96	100		
61	156.8	0.0	311.8
98	65.0	0.0	126.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#28
Bromochloromethane
Concen: 96.679 ug/l
RT: 7.806 min Scan# 995
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 49 Resp: 214796

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	4.4
130	68.4	55.0	82.4






#29
Tetrahydrofuran
Concen: 510.635 ug/l
RT: 7.830 min Scan# 99
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 42 Resp: 503459


Ion	Ratio	Lower	Upper
42	100		
72	43.1	34.2	51.2
71	40.9	32.5	48.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#30
Chloroform
Concen: 98.253 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 83 Resp: 484941

Ion	Ratio	Lower	Upper
83	100		
85	65.0	51.8	77.6






#31
Cyclohexane
Concen: 96.041 ug/l
RT: 8.253 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 56 Resp: 398641


Ion	Ratio	Lower	Upper
56	100		
69	29.2	22.2	33.4
84	86.4	66.4	99.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#32
1,1,1-Trichloroethane
Concen: 98.578 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 97 Resp: 426782

Ion	Ratio	Lower	Upper
97	100		
99	64.2	49.8	74.6
61	50.6	41.4	62.2






#33
1,2-Dichloroethane-d4
Concen: 95.899 ug/l
RT: 8.571 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 65 Resp: 313645


Ion	Ratio	Lower	Upper
65	100		
67	51.1	0.0	101.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.094 min Scan# 1214
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:114 Resp: 329363

Ion	Ratio	Lower	Upper
114	100		
63	21.2	0.0	47.6
88	16.2	0.0	32.6






#35
Dibromofluoromethane
Concen: 103.353 ug/l
RT: 8.159 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:113 Resp: 236159


Ion	Ratio	Lower	Upper
113	100		
111	102.3	82.7	124.1
192	17.4	14.3	21.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#36
1,1-Dichloropropene
Concen: 106.945 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 342961

Ion	Ratio	Lower	Upper
75	100		
110	32.6	16.5	49.5
77	30.6	24.4	36.6





#37
Ethyl Acetate
Concen: 100.961 ug/l
RT: 7.553 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56


Tgt Ion: 43 Resp: 326802


Ion	Ratio	Lower	Upper
43	100		
61	12.0	10.9	16.3
70	9.6	7.5	11.3

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025







#38
Carbon Tetrachloride
Concen: 103.065 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:117 Resp: 378390

Ion	Ratio	Lower	Upper
117	100		
119	98.1	75.4	113.2
121	30.2	24.6	37.0






#39
Methylcyclohexane
Concen: 123.354 ug/l
RT: 9.594 min Scan# 12
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 83 Resp: 371734


Ion	Ratio	Lower	Upper
83	100		
55	79.3	62.6	94.0
98	46.7	37.7	56.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#40
Benzene
Concen: 106.001 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 78 Resp: 1021393

Ion	Ratio	Lower	Upper
78	100		
77	23.9	19.0	28.6






#41
Methacrylonitrile
Concen: 109.454 ug/l
RT: 7.777 min Scan# 99
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 41 Resp: 184326


Ion	Ratio	Lower	Upper
41	100		
39	52.8	46.0	69.0
67	68.8	57.4	86.2
52	32.3	25.5	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#42
1,2-Dichloroethane
Concen: 103.314 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 62 Resp: 374938

Ion	Ratio	Lower	Upper
62	100		
98	8.4	0.0	17.0






#43
Isopropyl Acetate
Concen: 106.035 ug/l
RT: 8.682 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 549969


Ion	Ratio	Lower	Upper
43	100		
61	25.7	20.7	31.1
87	12.2	9.8	14.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#44
Trichloroethene
Concen: 106.242 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:130 Resp: 238296

Ion	Ratio	Lower	Upper
130	100		
95	101.5	0.0	195.8






#45
1,2-Dichloropropane
Concen: 104.405 ug/l
RT: 9.618 min Scan# 13
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 63 Resp: 256964


Ion	Ratio	Lower	Upper
63	100		
65	30.3	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#46
Dibromomethane
Concen: 101.913 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 93 Resp: 180990

Ion	Ratio	Lower	Upper
93	100		
95	81.6	64.7	97.1
174	87.5	69.0	103.4






#47
Bromodichloromethane
Concen: 107.377 ug/l
RT: 9.882 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 83 Resp: 388499


Ion	Ratio	Lower	Upper
83	100		
85	65.9	51.2	76.8
127	8.4	6.5	9.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#48
Methyl methacrylate
Concen: 114.685 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.001 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 41 Resp: 267689

Ion	Ratio	Lower	Upper
41	100		
69	80.9	64.7	97.1
39	63.1	49.0	73.6






#49
1,4-Dioxane
Concen: 2133.992 ug/l
RT: 9.688 min Scan# 13
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 88 Resp: 83879


Ion	Ratio	Lower	Upper
88	100		
43	31.3	26.6	39.8
58	72.1	59.5	89.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#50
Toluene-d8
Concen: 108.636 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 98 Resp: 881956

Ion	Ratio	Lower	Upper
98	100		
100	64.5	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 546.159 ug/l
RT: 10.441 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 1643800


Ion	Ratio	Lower	Upper
43	100		
58	38.4	30.5	45.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#52
Toluene
Concen: 113.689 ug/l
RT: 10.623 min Scan# 1474
Delta R.T. -0.001 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 92 Resp: 634743

Ion	Ratio	Lower	Upper
92	100		
91	173.1	139.2	208.8






#53
t-1,3-Dichloropropene
Concen: 114.424 ug/l
RT: 10.829 min Scan# 15
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 391238


Ion	Ratio	Lower	Upper
75	100		
77	31.0	23.5	35.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#54
cis-1,3-Dichloropropene
Concen: 112.369 ug/l
RT: 10.306 min Scan# 1420
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 410403

Ion	Ratio	Lower	Upper
75	100		
77	31.9	25.0	37.4
39	54.3	43.1	64.7






#55
1,1,2-Trichloroethane
Concen: 103.879 ug/l
RT: 11.012 min Scan# 15
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 97 Resp: 229522


Ion	Ratio	Lower	Upper
97	100		
83	88.4	69.0	103.6
85	56.3	44.6	66.8
99	62.9	48.6	72.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#56
Ethyl methacrylate
Concen: 101.029 ug/l
RT: 10.870 min Scan# 1516
Delta R.T. -0.001 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 69 Resp: 383663

Ion	Ratio	Lower	Upper
69	100		
41	69.1	54.6	82.0
39	40.1	32.4	48.6






#57
1,3-Dichloropropane
Concen: 107.469 ug/l
RT: 11.159 min Scan# 15
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 76 Resp: 412886


Ion	Ratio	Lower	Upper
76	100		
78	32.2	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#58
2-Chloroethyl Vinyl ether
Concen: 605.243 ug/l
RT: 10.153 min Scan# 1394
Delta R.T. -0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 63 Resp: 848636

Ion	Ratio	Lower	Upper
63	100		
106	26.8	21.6	32.4






#59
2-Hexanone
Concen: 556.643 ug/l
RT: 11.194 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 1178831


Ion	Ratio	Lower	Upper
43	100		
58	53.2	26.2	78.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#60
Dibromochloromethane
Concen: 106.227 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:129 Resp: 283361

Ion	Ratio	Lower	Upper
129	100		
127	78.1	38.6	115.8






#61
1,2-Dibromoethane
Concen: 104.623 ug/l
RT: 11.465 min Scan# 16
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:107 Resp: 230104


Ion	Ratio	Lower	Upper
107	100		
109	93.1	75.9	113.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#62
4-Bromofluorobenzene
Concen: 112.765 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 95 Resp: 313161

Ion	Ratio	Lower	Upper
95	100		
174	73.4	0.0	145.0
176	70.5	0.0	142.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 16
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:117 Resp: 298938


Ion	Ratio	Lower	Upper
117	100		
82	59.3	48.6	72.8
119	31.6	26.6	39.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#64
Tetrachloroethene
Concen: 102.980 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:164 Resp: 209870

Ion	Ratio	Lower	Upper
164	100		
166	126.8	103.4	155.2
129	98.3	79.2	118.8
131	94.7	77.1	115.7






#65
Chlorobenzene
Concen: 103.482 ug/l
RT: 11.888 min Scan# 16
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:112 Resp: 677679


Ion	Ratio	Lower	Upper
112	100		
114	32.5	25.3	37.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 100.517 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:131 Resp: 241593

Ion	Ratio	Lower	Upper
131	100		
133	94.3	47.4	142.3
119	64.9	33.1	99.5






#67
Ethyl Benzene
Concen: 116.169 ug/l
RT: 11.959 min Scan# 17
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 91 Resp: 1239080


Ion	Ratio	Lower	Upper
91	100		
106	30.3	23.8	35.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#68
m/p-Xylenes
Concen: 235.219 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:106 Resp: 927244

Ion	Ratio	Lower	Upper
106	100		
91	208.0	167.7	251.5






#69
o-Xylene
Concen: 117.135 ug/l
RT: 12.394 min Scan# 17
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:106 Resp: 441279


Ion	Ratio	Lower	Upper
106	100		
91	221.3	110.4	331.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#70
Styrene
Concen: 121.919 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:104 Resp: 760136

Ion	Ratio	Lower	Upper
104	100		
78	52.8	42.5	63.7
103	55.3	43.8	65.8






#71
Bromoform
Concen: 108.227 ug/l
RT: 12.576 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:173 Resp: 186124


Ion	Ratio	Lower	Upper
173	100		
175	48.4	24.4	73.2
254	0.1	0.0	0.0#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:152 Resp: 146196

Ion	Ratio	Lower	Upper
152	100		
115	62.7	31.1	93.3
150	193.1	0.0	343.6






#73
Isopropylbenzene
Concen: 116.221 ug/l
RT: 12.694 min Scan# 18
Delta R.T. 0.006 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:105 Resp: 1146752


Ion	Ratio	Lower	Upper
105	100		
120	25.7	12.8	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#74
N-amyl acetate
Concen: 110.308 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 43 Resp: 489217

Ion	Ratio	Lower	Upper
43	100		
70	41.9	34.0	51.0
55	25.7	21.4	32.2
61	23.5	19.0	28.4






#75
1,1,2,2-Tetrachloroethane
Concen: 94.039 ug/l
RT: 12.935 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 83 Resp: 327767


Ion	Ratio	Lower	Upper
83	100		
131	9.8	4.8	14.4
85	64.1	32.2	96.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#76
1,2,3-Trichloropropane
Concen: 104.530 ug/l m
RT: 12.988 min Scan# 1876
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 310115

Ion	Ratio	Lower	Upper
75	100		
77	192.6	109.7	329.2






#77
Bromobenzene
Concen: 104.200 ug/l
RT: 12.976 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:156 Resp: 268616


Ion	Ratio	Lower	Upper
156	100		
77	222.4	114.1	342.4
158	96.3	48.9	146.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#78
n-propylbenzene
Concen: 118.600 ug/l
RT: 13.029 min Scan# 1883
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 91 Resp: 1385353

Ion	Ratio	Lower	Upper
91	100		
120	21.8	10.9	32.6






#79
2-Chlorotoluene
Concen: 108.767 ug/l
RT: 13.123 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 91 Resp: 822258


Ion	Ratio	Lower	Upper
91	100		
126	31.8	15.7	47.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#80
1,3,5-Trimethylbenzene
Concen: 116.938 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:105 Resp: 953066

Ion	Ratio	Lower	Upper
105	100		
120	47.8	23.9	71.7






#81
trans-1,4-Dichloro-2-butene
Concen: 111.544 ug/l
RT: 12.735 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 122508


Ion	Ratio	Lower	Upper
75	100		
53	100.9	95.6	143.4
89	44.4	37.0	55.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#82
4-Chlorotoluene
Concen: 110.574 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 91 Resp: 832317

Ion	Ratio	Lower	Upper
91	100		
126	32.5	15.9	47.7






#83
tert-Butylbenzene
Concen: 116.457 ug/l
RT: 13.435 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:119 Resp: 797069


Ion	Ratio	Lower	Upper
119	100		
91	70.7	35.1	105.3
134	26.6	12.0	36.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#84
1,2,4-Trimethylbenzene
Concen: 119.625 ug/l
RT: 13.476 min Scan# 1959
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:105 Resp: 971729

Ion	Ratio	Lower	Upper
105	100		
120	43.4	21.6	65.0






#85
sec-Butylbenzene
Concen: 120.587 ug/l
RT: 13.611 min Scan# 19
Delta R.T. -0.001 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:105 Resp: 1143885


Ion	Ratio	Lower	Upper
105	100		
134	19.2	9.7	28.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#86
p-Isopropyltoluene
Concen: 100.111 ug/l
RT: 13.723 min Scan# 2001
Delta R.T. 0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:119 Resp: 960747

Ion	Ratio	Lower	Upper
119	100		
134	25.3	12.7	38.0
91	25.0	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 105.943 ug/l
RT: 13.729 min Scan# 20
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:146 Resp: 502960


Ion	Ratio	Lower	Upper
146	100		
111	42.8	21.4	64.3
148	63.6	32.3	96.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#88
1,4-Dichlorobenzene
Concen: 101.341 ug/l
RT: 13.811 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:146 Resp: 498709

Ion	Ratio	Lower	Upper
146	100		
111	40.1	21.3	63.7
148	63.1	32.4	97.0






#89
n-Butylbenzene
Concen: 123.781 ug/l
RT: 14.053 min Scan# 20
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 91 Resp: 842368


Ion	Ratio	Lower	Upper
91	100		
92	52.8	25.8	77.3
134	24.3	11.7	35.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#90
Hexachloroethane
Concen: 104.405 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:117 Resp: 188541

Ion	Ratio	Lower	Upper
117	100		
201	67.2	33.7	101.0






#91
1,2-Dichlorobenzene
Concen: 99.445 ug/l
RT: 14.100 min Scan# 20
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:146 Resp: 470919


Ion	Ratio	Lower	Upper
146	100		
111	43.9	21.7	65.1
148	64.0	31.4	94.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 100.313 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion: 75 Resp: 63846

Ion	Ratio	Lower	Upper
75	100		
155	73.5	36.4	109.2
157	93.0	45.4	136.1






#93
1,2,4-Trichlorobenzene
Concen: 113.946 ug/l
RT: 15.388 min Scan# 22
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:180 Resp: 250787


Ion	Ratio	Lower	Upper
180	100		
182	94.9	46.8	140.3
145	32.6	16.0	48.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#94
Hexachlorobutadiene
Concen: 96.619 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:225 Resp: 112899

Ion	Ratio	Lower	Upper
225	100		
223	61.8	30.7	92.1
227	63.6	30.9	92.5






#95
Naphthalene
Concen: 115.234 ug/l
RT: 15.635 min Scan# 23
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:128 Resp: 756816


Ion	Ratio	Lower	Upper
128	100		
127	13.1	10.6	16.0
129	10.8	8.8	13.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96
1,2,3-Trichlorobenzene
Concen: 107.256 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN085438.D
Acq: 14 Jan 2025 14:56

Tgt Ion:180 Resp: 238860

Ion	Ratio	Lower	Upper
180	100		
182	95.0	47.4	142.2
145	34.3	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085439.D
Acq On : 14 Jan 2025 15:19
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED
Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:42:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.218	168	199403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	339872	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	297366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	146624	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	165768	51.500	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.000%			
35) Dibromofluoromethane	8.159	113	121554	51.552	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.100%			
50) Toluene-d8	10.565	98	430698	51.411	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.820%			
62) 4-Bromofluorobenzene	12.847	95	152568	53.239	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 106.480%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.124	85	125504	46.486	ug/l	100
3) Chloromethane	2.359	50	135631	46.398	ug/l	100
4) Vinyl Chloride	2.512	62	136831	46.569	ug/l	100
5) Bromomethane	2.953	94	83199	46.877	ug/l	100
6) Chloroethane	3.118	64	84517	45.371	ug/l	100
7) Trichlorofluoromethane	3.501	101	198726	46.613	ug/l	100
8) Diethyl Ether	3.953	74	71764	48.726	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	108081	45.010	ug/l	100
10) Methyl Iodide	4.589	142	143383	52.142	ug/l	100
11) Tert butyl alcohol	5.512	59	92841	251.894	ug/l	100
12) 1,1-Dichloroethene	4.342	96	106247	49.647	ug/l	100
13) Acrolein	4.171	56	139234	276.734	ug/l	100
14) Allyl chloride	5.018	41	168654	48.567	ug/l	100
15) Acrylonitrile	5.712	53	296781	253.507	ug/l	100
16) Acetone	4.418	43	251271	241.603	ug/l	100
17) Carbon Disulfide	4.706	76	294604	44.710	ug/l	100
18) Methyl Acetate	5.018	43	157485	49.798	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	373561	53.758	ug/l	100
20) Methylene Chloride	5.271	84	125508	48.748	ug/l	100
21) trans-1,2-Dichloroethene	5.783	96	110684	48.395	ug/l	100
22) Diisopropyl ether	6.665	45	406253	52.707	ug/l	100
23) Vinyl Acetate	6.594	43	1491031	275.471	ug/l	100
24) 1,1-Dichloroethane	6.565	63	233336	49.648	ug/l	100
25) 2-Butanone	7.477	43	388457	253.815	ug/l	100
26) 2,2-Dichloropropane	7.483	77	186591	49.106	ug/l	100
27) cis-1,2-Dichloroethene	7.483	96	136229	50.571	ug/l	100
28) Bromochloromethane	7.806	49	108148	49.461	ug/l	100
29) Tetrahydrofuran	7.830	42	259683	267.624	ug/l	100
30) Chloroform	7.959	83	234271	48.229	ug/l	100
31) Cyclohexane	8.253	56	175588	42.984	ug/l	100
32) 1,1,1-Trichloroethane	8.165	97	202590	47.547	ug/l	100
36) 1,1-Dichloropropene	8.365	75	157031	47.453	ug/l	100
37) Ethyl Acetate	7.553	43	165866	49.658	ug/l	100
38) Carbon Tetrachloride	8.359	117	180199	47.564	ug/l	100
39) Methylcyclohexane	9.600	83	157234	50.562	ug/l	100
40) Benzene	8.600	78	492606	49.542	ug/l	100

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085439.D
Acq On : 14 Jan 2025 15:19
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED
Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:42:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	7.771	41	91711	52.775	ug/l		100
42) 1,2-Dichloroethane	8.665	62	186065	49.685	ug/l		100
43) Isopropyl Acetate	8.683	43	276228	51.611	ug/l		100
44) Trichloroethene	9.347	130	109953	47.506	ug/l		100
45) 1,2-Dichloropropane	9.618	63	126084	49.644	ug/l		100
46) Dibromomethane	9.706	93	90670	49.476	ug/l		100
47) Bromodichloromethane	9.882	83	189915	50.867	ug/l		100
48) Methyl methacrylate	9.677	41	133399	55.385	ug/l		100
49) 1,4-Dioxane	9.688	88	43048	1061.334	ug/l		100
51) 4-Methyl-2-Pentanone	10.441	43	836707	269.403	ug/l		100
52) Toluene	10.624	92	295792	51.341	ug/l		100
53) t-1,3-Dichloropropene	10.835	75	187439	53.125	ug/l		100
54) cis-1,3-Dichloropropene	10.306	75	199771	53.006	ug/l		100
55) 1,1,2-Trichloroethane	11.012	97	115462	50.641	ug/l		100
56) Ethyl methacrylate	10.871	69	184775	48.182	ug/l		100
57) 1,3-Dichloropropane	11.159	76	204590	51.606	ug/l		100
58) 2-Chloroethyl Vinyl ether	10.159	63	409594	283.088	ug/l		100
59) 2-Hexanone	11.188	43	606740	277.643	ug/l		100
60) Dibromochloromethane	11.353	129	140704	51.116	ug/l		100
61) 1,2-Dibromoethane	11.465	107	113541	50.028	ug/l		100
64) Tetrachloroethene	11.100	164	95879	47.295	ug/l		100
65) Chlorobenzene	11.888	112	319992	49.121	ug/l		100
66) 1,1,1,2-Tetrachloroethane	11.959	131	116488	48.722	ug/l		100
67) Ethyl Benzene	11.959	91	555322	52.339	ug/l		100
68) m/p-Xylenes	12.065	106	420702	107.286	ug/l		100
69) o-Xylene	12.394	106	202597	54.062	ug/l		100
70) Styrene	12.406	104	348720	56.227	ug/l		100
71) Bromoform	12.576	173	92366	53.993	ug/l	#	100
73) Isopropylbenzene	12.688	105	505619	51.094	ug/l		100
74) N-amyl acetate	12.488	43	237870	53.478	ug/l		100
75) 1,1,2,2-Tetrachloroethane	12.935	83	167856	48.019	ug/l		100
76) 1,2,3-Trichloropropane	12.988	75	130866m	43.982	ug/l		
77) Bromobenzene	12.976	156	125800	48.657	ug/l		100
78) n-propylbenzene	13.029	91	609022	51.986	ug/l		100
79) 2-Chlorotoluene	13.123	91	378981	49.985	ug/l		100
80) 1,3,5-Trimethylbenzene	13.170	105	431021	52.730	ug/l		100
81) trans-1,4-Dichloro-2-b...	12.735	75	57184	51.914	ug/l		100
82) 4-Chlorotoluene	13.218	91	379524	50.273	ug/l		100
83) tert-Butylbenzene	13.435	119	357563	52.090	ug/l		100
84) 1,2,4-Trimethylbenzene	13.476	105	442175	54.275	ug/l		100
85) sec-Butylbenzene	13.612	105	501561	52.719	ug/l		100
86) p-Isopropyltoluene	13.723	119	421052	49.240	ug/l		100
87) 1,3-Dichlorobenzene	13.729	146	229518	48.204	ug/l		100
88) 1,4-Dichlorobenzene	13.812	146	229084	46.416	ug/l		100
89) n-Butylbenzene	14.053	91	364111	53.348	ug/l		100
90) Hexachloroethane	14.329	117	84602	46.712	ug/l		100
91) 1,2-Dichlorobenzene	14.100	146	228024	48.012	ug/l		100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	32536	50.971	ug/l		100
93) 1,2,4-Trichlorobenzene	15.388	180	114468	51.857	ug/l		100
94) Hexachlorobutadiene	15.500	225	53846	45.947	ug/l		100
95) Naphthalene	15.635	128	363882	55.243	ug/l		100
96) 1,2,3-Trichlorobenzene	15.835	180	115251	51.601	ug/l		100

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085439.D
Acq On : 14 Jan 2025 15:19
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jan 15 01:42:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085439.D
Acq On : 14 Jan 2025 15:19
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Jan 15 01:42:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:168 Resp: 199403


Ion	Ratio	Lower	Upper
168	100		
99	67.0	53.6	80.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#2
Dichlorodifluoromethane
Concen: 46.486 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 85 Resp: 125504

Ion	Ratio	Lower	Upper
85	100		
87	33.4	16.7	50.1






#3
Chloromethane
Concen: 46.398 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 50 Resp: 135631


Ion	Ratio	Lower	Upper
50	100		
52	33.8	27.0	40.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#4
Vinyl Chloride
Concen: 46.569 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 62 Resp: 136831

Ion	Ratio	Lower	Upper
62	100		
64	31.0	24.8	37.2






#5
Bromomethane
Concen: 46.877 ug/l
RT: 2.953 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 94 Resp: 83199


Ion	Ratio	Lower	Upper
94	100		
96	89.7	71.8	107.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#6
Chloroethane
Concen: 45.371 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 64 Resp: 84517

Ion	Ratio	Lower	Upper
64	100		
66	30.7	24.6	36.8






#7
Trichlorofluoromethane
Concen: 46.613 ug/l
RT: 3.501 min Scan# 26
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:101 Resp: 198726


Ion	Ratio	Lower	Upper
101	100		
103	64.3	51.4	77.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#8
Diethyl Ether
Concen: 48.726 ug/l
RT: 3.953 min Scan# 340
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 74 Resp: 71764

Ion	Ratio	Lower	Upper
74	100		
45	99.4	49.7	149.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 45.010 ug/l
RT: 4.371 min Scan# 41
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19


Tgt Ion:101 Resp: 108081


Ion	Ratio	Lower	Upper
101	100		
85	47.3	37.8	56.8
151	73.5	58.8	88.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025







#10
Methyl Iodide
Concen: 52.142 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:142 Resp: 143383

Ion	Ratio	Lower	Upper
142	100		
127	50.4	40.3	60.5
141	16.0	12.8	19.2






#11
Tert butyl alcohol
Concen: 251.894 ug/l
RT: 5.512 min Scan# 60
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 59 Resp: 92841


Ion	Ratio	Lower	Upper
59	100		
57	10.8	8.6	13.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#12
1,1-Dichloroethene
Concen: 49.647 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 96 Resp: 106247

Ion	Ratio	Lower	Upper
96	100		
61	176.2	141.0	211.4
98	61.3	49.0	73.6






#13
Acrolein
Concen: 276.734 ug/l
RT: 4.171 min Scan# 37
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 56 Resp: 139234


Ion	Ratio	Lower	Upper
56	100		
55	70.4	56.3	84.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#14
Allyl chloride
Concen: 48.567 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 41 Resp: 168654

Ion	Ratio	Lower	Upper
41	100		
39	80.5	64.4	96.6
76	38.1	30.5	45.7






#15
Acrylonitrile
Concen: 253.507 ug/l
RT: 5.712 min Scan# 63
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 53 Resp: 296781


Ion	Ratio	Lower	Upper
53	100		
52	81.9	65.5	98.3
51	37.3	29.8	44.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#16
Acetone
Concen: 241.603 ug/l
RT: 4.418 min Scan# 419
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 251271

Ion	Ratio	Lower	Upper
43	100		
58	29.7	23.8	35.6






#17
Carbon Disulfide
Concen: 44.710 ug/l
RT: 4.706 min Scan# 46
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 76 Resp: 294604


Ion	Ratio	Lower	Upper
76	100		
78	8.6	6.9	10.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#18
Methyl Acetate
Concen: 49.798 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 157485

Ion	Ratio	Lower	Upper
43	100		
74	23.3	18.6	28.0






#19
Methyl tert-butyl Ether
Concen: 53.758 ug/l
RT: 5.789 min Scan# 65
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 73 Resp: 373561


Ion	Ratio	Lower	Upper
73	100		
57	22.3	17.8	26.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#20
Methylene Chloride
Concen: 48.748 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 84 Resp: 125508

Ion	Ratio	Lower	Upper
84	100		
49	130.1	104.1	156.1
51	38.7	31.0	46.4
86	64.9	51.9	77.9






#21
trans-1,2-Dichloroethene
Concen: 48.395 ug/l
RT: 5.783 min Scan# 65
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 96 Resp: 110684


Ion	Ratio	Lower	Upper
96	100		
61	144.6	115.7	173.5
98	62.3	49.8	74.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#22
Diisopropyl ether
Concen: 52.707 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 45 Resp: 406253

Ion	Ratio	Lower	Upper
45	100		
43	56.3	45.0	67.6
87	27.3	21.8	32.8
59	12.7	10.2	15.2






#23
Vinyl Acetate
Concen: 275.471 ug/l
RT: 6.594 min Scan# 78
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 1491031


Ion	Ratio	Lower	Upper
43	100		
86	10.5	8.4	12.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#24
1,1-Dichloroethane
Concen: 49.648 ug/l
RT: 6.565 min Scan# 784
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 63 Resp: 233336

Ion	Ratio	Lower	Upper
63	100		
98	6.5	3.3	9.8
100	4.0	2.0	6.0






#25
2-Butanone
Concen: 253.815 ug/l
RT: 7.477 min Scan# 93
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 388457


Ion	Ratio	Lower	Upper
43	100		
72	25.3	20.2	30.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#26
2,2-Dichloropropane
Concen: 49.106 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 77 Resp: 186591

Ion	Ratio	Lower	Upper
77	100		
97	21.4	10.7	32.1






#27
cis-1,2-Dichloroethene
Concen: 50.571 ug/l
RT: 7.483 min Scan# 94
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 96 Resp: 136229


Ion	Ratio	Lower	Upper
96	100		
61	155.9	0.0	311.8
98	63.0	0.0	126.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#28
Bromochloromethane
Concen: 49.461 ug/l
RT: 7.806 min Scan# 995
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 49 Resp: 108148

Ion	Ratio	Lower	Upper
49	100		
129	2.2	0.0	4.4
130	68.7	55.0	82.4






#29
Tetrahydrofuran
Concen: 267.624 ug/l
RT: 7.830 min Scan# 99
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 42 Resp: 259683


Ion	Ratio	Lower	Upper
42	100		
72	42.7	34.2	51.2
71	40.6	32.5	48.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#30
Chloroform
Concen: 48.229 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 83 Resp: 234271

Ion	Ratio	Lower	Upper
83	100		
85	64.7	51.8	77.6






#31
Cyclohexane
Concen: 42.984 ug/l
RT: 8.253 min Scan# 10
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 56 Resp: 175588


Ion	Ratio	Lower	Upper
56	100		
69	27.8	22.2	33.4
84	83.0	66.4	99.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#32
1,1,1-Trichloroethane
Concen: 47.547 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 97 Resp: 202590

Ion	Ratio	Lower	Upper
97	100		
99	62.2	49.8	74.6
61	51.8	41.4	62.2





#33
1,2-Dichloroethane-d4
Concen: 51.500 ug/l
RT: 8.571 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19


Tgt Ion: 65 Resp: 165768

Ion	Ratio	Lower	Upper
65	100		
67	50.8	0.0	101.6

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.094 min Scan# 1214
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:114 Resp: 339872


Ion	Ratio	Lower	Upper
114	100		
63	23.8	0.0	47.6
88	16.3	0.0	32.6



#35
Dibromofluoromethane
Concen: 51.552 ug/l
RT: 8.159 min Scan# 10
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:113 Resp: 121554


Ion	Ratio	Lower	Upper
113	100		
111	103.4	82.7	124.1
192	17.9	14.3	21.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#36
1,1-Dichloropropene
Concen: 47.453 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 75 Resp: 157031

Ion	Ratio	Lower	Upper
75	100		
110	33.0	16.5	49.5
77	30.5	24.4	36.6






#37
Ethyl Acetate
Concen: 49.658 ug/l
RT: 7.553 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 165866


Ion	Ratio	Lower	Upper
43	100		
61	13.6	10.9	16.3
70	9.4	7.5	11.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#38
Carbon Tetrachloride
Concen: 47.564 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:117 Resp: 180199

Ion	Ratio	Lower	Upper
117	100		
119	94.3	75.4	113.2
121	30.8	24.6	37.0






#39
Methylcyclohexane
Concen: 50.562 ug/l
RT: 9.600 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 83 Resp: 157234


Ion	Ratio	Lower	Upper
83	100		
55	78.3	62.6	94.0
98	47.1	37.7	56.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#40
Benzene
Concen: 49.542 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 78 Resp: 492606

Ion	Ratio	Lower	Upper
78	100		
77	23.8	19.0	28.6






#41
Methacrylonitrile
Concen: 52.775 ug/l
RT: 7.771 min Scan# 98
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 41 Resp: 91711


Ion	Ratio	Lower	Upper
41	100		
39	57.5	46.0	69.0
67	71.8	57.4	86.2
52	31.9	25.5	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#42
1,2-Dichloroethane
Concen: 49.685 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 62 Resp: 186065

Ion	Ratio	Lower	Upper
62	100		
98	8.5	0.0	17.0






#43
Isopropyl Acetate
Concen: 51.611 ug/l
RT: 8.683 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 276228


Ion	Ratio	Lower	Upper
43	100		
61	25.9	20.7	31.1
87	12.3	9.8	14.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#44
Trichloroethene
Concen: 47.506 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:130 Resp: 109953

Ion	Ratio	Lower	Upper
130	100		
95	97.9	0.0	195.8






#45
1,2-Dichloropropane
Concen: 49.644 ug/l
RT: 9.618 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 63 Resp: 126084


Ion	Ratio	Lower	Upper
63	100		
65	32.0	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#46
Dibromomethane
Concen: 49.476 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 93 Resp: 90670

Ion	Ratio	Lower	Upper
93	100		
95	80.9	64.7	97.1
174	86.2	69.0	103.4






#47
Bromodichloromethane
Concen: 50.867 ug/l
RT: 9.882 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 83 Resp: 189915


Ion	Ratio	Lower	Upper
83	100		
85	64.0	51.2	76.8
127	8.1	6.5	9.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#48
Methyl methacrylate
Concen: 55.385 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 41 Resp: 133399

Ion	Ratio	Lower	Upper
41	100		
69	80.9	64.7	97.1
39	61.3	49.0	73.6






#49
1,4-Dioxane
Concen: 1061.334 ug/l
RT: 9.688 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 88 Resp: 43048


Ion	Ratio	Lower	Upper
88	100		
43	33.2	26.6	39.8
58	74.4	59.5	89.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#50
Toluene-d8
Concen: 51.411 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 98 Resp: 430698

Ion	Ratio	Lower	Upper
98	100		
100	65.3	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 269.403 ug/l
RT: 10.441 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 836707


Ion	Ratio	Lower	Upper
43	100		
58	38.1	30.5	45.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#52
Toluene
Concen: 51.341 ug/l
RT: 10.624 min Scan# 1474
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 92 Resp: 295792

Ion	Ratio	Lower	Upper
92	100		
91	174.0	139.2	208.8






#53
t-1,3-Dichloropropene
Concen: 53.125 ug/l
RT: 10.835 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 75 Resp: 187439


Ion	Ratio	Lower	Upper
75	100		
77	29.4	23.5	35.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#54
cis-1,3-Dichloropropene
Concen: 53.006 ug/l
RT: 10.306 min Scan# 1420
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 75 Resp: 199771

Ion	Ratio	Lower	Upper
75	100		
77	31.2	25.0	37.4
39	53.9	43.1	64.7






#55
1,1,2-Trichloroethane
Concen: 50.641 ug/l
RT: 11.012 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 97 Resp: 115462


Ion	Ratio	Lower	Upper
97	100		
83	86.3	69.0	103.6
85	55.7	44.6	66.8
99	60.7	48.6	72.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#56
Ethyl methacrylate
Concen: 48.182 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 69 Resp: 184775

Ion	Ratio	Lower	Upper
69	100		
41	68.3	54.6	82.0
39	40.5	32.4	48.6






#57
1,3-Dichloropropane
Concen: 51.606 ug/l
RT: 11.159 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 76 Resp: 204590


Ion	Ratio	Lower	Upper
76	100		
78	32.0	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#58
2-Chloroethyl Vinyl ether
Concen: 283.088 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 63 Resp: 409594

Ion	Ratio	Lower	Upper
63	100		
106	27.0	21.6	32.4






#59
2-Hexanone
Concen: 277.643 ug/l
RT: 11.188 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 606740


Ion	Ratio	Lower	Upper
43	100		
58	52.4	26.2	78.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#60
Dibromochloromethane
Concen: 51.116 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:129 Resp: 140704

Ion	Ratio	Lower	Upper
129	100		
127	77.2	38.6	115.8






#61
1,2-Dibromoethane
Concen: 50.028 ug/l
RT: 11.465 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:107 Resp: 113541


Ion	Ratio	Lower	Upper
107	100		
109	94.9	75.9	113.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#62
4-Bromofluorobenzene
Concen: 53.239 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 95 Resp: 152568

Ion	Ratio	Lower	Upper
95	100		
174	72.5	0.0	145.0
176	71.2	0.0	142.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:117 Resp: 297366


Ion	Ratio	Lower	Upper
117	100		
82	60.7	48.6	72.8
119	33.2	26.6	39.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#64
Tetrachloroethene
Concen: 47.295 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:164 Resp: 95879

Ion	Ratio	Lower	Upper
164	100		
166	129.3	103.4	155.2
129	99.0	79.2	118.8
131	96.4	77.1	115.7






#65
Chlorobenzene
Concen: 49.121 ug/l
RT: 11.888 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:112 Resp: 319992


Ion	Ratio	Lower	Upper
112	100		
114	31.6	25.3	37.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 48.722 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:131 Resp: 116488

Ion	Ratio	Lower	Upper
131	100		
133	94.9	47.4	142.3
119	66.3	33.1	99.5






#67
Ethyl Benzene
Concen: 52.339 ug/l
RT: 11.959 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 91 Resp: 555322


Ion	Ratio	Lower	Upper
91	100		
106	29.8	23.8	35.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#68
m/p-Xylenes
Concen: 107.286 ug/l
RT: 12.065 min Scan# 1719
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:106 Resp: 420702

Ion	Ratio	Lower	Upper
106	100		
91	209.6	167.7	251.5





#69
o-Xylene
Concen: 54.062 ug/l
RT: 12.394 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19


Tgt Ion:106 Resp: 202597


Ion	Ratio	Lower	Upper
106	100		
91	220.8	110.4	331.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025







#70
Styrene
Concen: 56.227 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:104 Resp: 348720

Ion	Ratio	Lower	Upper
104	100		
78	53.1	42.5	63.7
103	54.8	43.8	65.8






#71
Bromoform
Concen: 53.993 ug/l
RT: 12.576 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:173 Resp: 92366


Ion	Ratio	Lower	Upper
173	100		
175	48.8	24.4	73.2
254	0.0	0.0	0.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:152 Resp: 146624

Ion	Ratio	Lower	Upper
152	100		
115	62.2	31.1	93.3
150	171.8	0.0	343.6






#73
Isopropylbenzene
Concen: 51.094 ug/l
RT: 12.688 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:105 Resp: 505619


Ion	Ratio	Lower	Upper
105	100		
120	25.5	12.8	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#74
N-amyl acetate
Concen: 53.478 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 43 Resp: 237870

Ion	Ratio	Lower	Upper
43	100		
70	42.5	34.0	51.0
55	26.8	21.4	32.2
61	23.7	19.0	28.4






#75
1,1,2,2-Tetrachloroethane
Concen: 48.019 ug/l
RT: 12.935 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 83 Resp: 167856


Ion	Ratio	Lower	Upper
83	100		
131	9.6	4.8	14.4
85	64.4	32.2	96.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#76
1,2,3-Trichloropropane
Concen: 43.982 ug/l m
RT: 12.988 min Scan# 1876
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 75 Resp: 130866

Ion	Ratio	Lower	Upper
75	100		
77	219.5	109.7	329.2






#77
Bromobenzene
Concen: 48.657 ug/l
RT: 12.976 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:156 Resp: 125800


Ion	Ratio	Lower	Upper
156	100		
77	228.3	114.1	342.4
158	97.9	48.9	146.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#78
n-propylbenzene
Concen: 51.986 ug/l
RT: 13.029 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 91 Resp: 609022

Ion	Ratio	Lower	Upper
91	100		
120	21.7	10.9	32.6






#79
2-Chlorotoluene
Concen: 49.985 ug/l
RT: 13.123 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 91 Resp: 378981


Ion	Ratio	Lower	Upper
91	100		
126	31.4	15.7	47.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#80
1,3,5-Trimethylbenzene
Concen: 52.730 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:105 Resp: 431021

Ion	Ratio	Lower	Upper
105	100		
120	47.8	23.9	71.7






#81
trans-1,4-Dichloro-2-butene
Concen: 51.914 ug/l
RT: 12.735 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 75 Resp: 57184


Ion	Ratio	Lower	Upper
75	100		
53	119.5	95.6	143.4
89	46.3	37.0	55.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#82
4-Chlorotoluene
Concen: 50.273 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 91 Resp: 379524

Ion	Ratio	Lower	Upper
91	100		
126	31.8	15.9	47.7






#83
tert-Butylbenzene
Concen: 52.090 ug/l
RT: 13.435 min Scan# 19
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:119 Resp: 357563


Ion	Ratio	Lower	Upper
119	100		
91	70.2	35.1	105.3
134	24.1	12.0	36.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#84
1,2,4-Trimethylbenzene
Concen: 54.275 ug/l
RT: 13.476 min Scan# 1959
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:105 Resp: 442175

Ion	Ratio	Lower	Upper
105	100		
120	43.3	21.6	65.0






#85
sec-Butylbenzene
Concen: 52.719 ug/l
RT: 13.612 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:105 Resp: 501561


Ion	Ratio	Lower	Upper
105	100		
134	19.3	9.7	28.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#86
p-Isopropyltoluene
Concen: 49.240 ug/l
RT: 13.723 min Scan# 2001
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:119 Resp: 421052

Ion	Ratio	Lower	Upper
119	100		
134	25.3	12.7	38.0
91	25.4	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 48.204 ug/l
RT: 13.729 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:146 Resp: 229518


Ion	Ratio	Lower	Upper
146	100		
111	42.9	21.4	64.3
148	64.6	32.3	96.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#88
1,4-Dichlorobenzene
Concen: 46.416 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:146 Resp: 229084

Ion	Ratio	Lower	Upper
146	100		
111	42.5	21.3	63.7
148	64.7	32.4	97.0






#89
n-Butylbenzene
Concen: 53.348 ug/l
RT: 14.053 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 91 Resp: 364111


Ion	Ratio	Lower	Upper
91	100		
92	51.5	25.8	77.3
134	23.4	11.7	35.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#90
Hexachloroethane
Concen: 46.712 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:117 Resp: 84602

Ion	Ratio	Lower	Upper
117	100		
201	67.3	33.7	101.0






#91
1,2-Dichlorobenzene
Concen: 48.012 ug/l
RT: 14.100 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:146 Resp: 228024


Ion	Ratio	Lower	Upper
146	100		
111	43.4	21.7	65.1
148	62.8	31.4	94.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 50.971 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion: 75 Resp: 32536

Ion	Ratio	Lower	Upper
75	100		
155	72.8	36.4	109.2
157	90.7	45.4	136.1






#93
1,2,4-Trichlorobenzene
Concen: 51.857 ug/l
RT: 15.388 min Scan# 22
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:180 Resp: 114468


Ion	Ratio	Lower	Upper
180	100		
182	93.5	46.8	140.3
145	32.0	16.0	48.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#94
Hexachlorobutadiene
Concen: 45.947 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:225 Resp: 53846

Ion	Ratio	Lower	Upper
225	100		
223	61.4	30.7	92.1
227	61.7	30.9	92.5






#95
Naphthalene
Concen: 55.243 ug/l
RT: 15.635 min Scan# 23
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:128 Resp: 363882


Ion	Ratio	Lower	Upper
128	100		
127	13.3	10.6	16.0
129	11.0	8.8	13.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96
1,2,3-Trichlorobenzene
Concen: 51.601 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN085439.D
Acq: 14 Jan 2025 15:19

Tgt Ion:180 Resp: 115251

Ion	Ratio	Lower	Upper
180	100		
182	94.8	47.4	142.2
145	33.8	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085440.D
Acq On : 14 Jan 2025 15:43
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jan 15 01:43:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	200796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	340593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	296298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143526	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	60593	18.694	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 37.380%#			
35) Dibromofluoromethane	8.165	113	45655	19.322	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 38.640%#			
50) Toluene-d8	10.559	98	164391	19.581	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 39.160%#			
62) 4-Bromofluorobenzene	12.847	95	55798	19.430	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 38.860%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.124	85	54666	20.107	ug/l	98
3) Chloromethane	2.359	50	58426	19.848	ug/l	100
4) Vinyl Chloride	2.512	62	58403	19.739	ug/l	99
5) Bromomethane	2.954	94	36440	20.389	ug/l	94
6) Chloroethane	3.118	64	37594	20.041	ug/l	93
7) Trichlorofluoromethane	3.501	101	88121	20.526	ug/l	99
8) Diethyl Ether	3.965	74	29138	19.647	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.371	101	48933	20.237	ug/l	99
10) Methyl Iodide	4.589	142	56567	20.428	ug/l	93
11) Tert butyl alcohol	5.506	59	37984	102.342	ug/l	100
12) 1,1-Dichloroethene	4.342	96	44670	20.728	ug/l	95
13) Acrolein	4.171	56	49649	97.995	ug/l	99
14) Allyl chloride	5.024	41	67705	19.362	ug/l	96
15) Acrylonitrile	5.712	53	121626	103.171	ug/l	99
16) Acetone	4.424	43	101065	96.502	ug/l	96
17) Carbon Disulfide	4.712	76	132300	19.939	ug/l	98
18) Methyl Acetate	5.018	43	60873	19.115	ug/l	99
19) Methyl tert-butyl Ether	5.795	73	148847	21.272	ug/l	100
20) Methylene Chloride	5.271	84	52813	20.370	ug/l	97
21) trans-1,2-Dichloroethene	5.789	96	46115	20.023	ug/l	98
22) Diisopropyl ether	6.671	45	166719	21.480	ug/l	97
23) Vinyl Acetate	6.594	43	588567m	107.984	ug/l	
24) 1,1-Dichloroethane	6.565	63	96875	20.469	ug/l #	96
25) 2-Butanone	7.477	43	159952	103.786	ug/l	98
26) 2,2-Dichloropropane	7.483	77	79145	20.684	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	57454	21.180	ug/l	98
28) Bromochloromethane	7.812	49	41181	18.703	ug/l	99
29) Tetrahydrofuran	7.836	42	104792	107.247	ug/l	99
30) Chloroform	7.959	83	99655	20.374	ug/l	99
31) Cyclohexane	8.253	56	82980	20.173	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	87659	20.431	ug/l	98
36) 1,1-Dichloropropene	8.371	75	67949	20.490	ug/l	98
37) Ethyl Acetate	7.553	43	66443	19.850	ug/l	96
38) Carbon Tetrachloride	8.359	117	78842	20.767	ug/l	96
39) Methylcyclohexane	9.594	83	65012	20.862	ug/l	96
40) Benzene	8.600	78	207985	20.873	ug/l	99

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085440.D
Acq On : 14 Jan 2025 15:43
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jan 15 01:43:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	7.771	41	36578	21.004	ug/l		99
42) 1,2-Dichloroethane	8.665	62	78402	20.891	ug/l		99
43) Isopropyl Acetate	8.683	43	109131	20.347	ug/l		98
44) Trichloroethene	9.347	130	48009	20.699	ug/l		96
45) 1,2-Dichloropropane	9.618	63	52833	20.758	ug/l		93
46) Dibromomethane	9.700	93	37721	20.540	ug/l		98
47) Bromodichloromethane	9.882	83	78918	21.093	ug/l		96
48) Methyl methacrylate	9.677	41	50224	20.808	ug/l		96
49) 1,4-Dioxane	9.688	88	17140	421.686	ug/l		95
51) 4-Methyl-2-Pentanone	10.441	43	336911	108.249	ug/l		100
52) Toluene	10.630	92	125172	21.680	ug/l		99
53) t-1,3-Dichloropropene	10.835	75	74129	20.965	ug/l		96
54) cis-1,3-Dichloropropene	10.306	75	81897	21.684	ug/l		99
55) 1,1,2-Trichloroethane	11.012	97	48039	21.025	ug/l		97
56) Ethyl methacrylate	10.871	69	70934	19.649	ug/l		99
57) 1,3-Dichloropropane	11.159	76	84210	21.196	ug/l		99
58) 2-Chloroethyl Vinyl ether	10.159	63	153681	105.991	ug/l		100
59) 2-Hexanone	11.194	43	240595	109.863	ug/l		98
60) Dibromochloromethane	11.353	129	56069	20.326	ug/l		96
61) 1,2-Dibromoethane	11.465	107	48465	21.309	ug/l		96
64) Tetrachloroethene	11.100	164	43243	21.408	ug/l		96
65) Chlorobenzene	11.888	112	136809	21.077	ug/l		98
66) 1,1,1,2-Tetrachloroethane	11.959	131	48370	20.304	ug/l		98
67) Ethyl Benzene	11.959	91	229978	21.753	ug/l		98
68) m/p-Xylenes	12.071	106	177755	45.494	ug/l		100
69) o-Xylene	12.394	106	84548	22.643	ug/l		98
70) Styrene	12.412	104	140581	22.749	ug/l		99
71) Bromoform	12.576	173	36961	21.683	ug/l	#	96
73) Isopropylbenzene	12.694	105	211354	21.819	ug/l		100
74) N-amyl acetate	12.494	43	94292	21.656	ug/l		98
75) 1,1,2,2-Tetrachloroethane	12.935	83	68151	19.917	ug/l		99
76) 1,2,3-Trichloropropane	12.988	75	64180m	22.036	ug/l		
77) Bromobenzene	12.976	156	52022	20.556	ug/l		99
78) n-propylbenzene	13.035	91	252610	22.028	ug/l		100
79) 2-Chlorotoluene	13.123	91	158511	21.358	ug/l		100
80) 1,3,5-Trimethylbenzene	13.171	105	180294	22.533	ug/l		100
81) trans-1,4-Dichloro-2-b...	12.735	75	22178m	20.569	ug/l		
82) 4-Chlorotoluene	13.218	91	161880	21.906	ug/l		100
83) tert-Butylbenzene	13.435	119	147541	21.958	ug/l		99
84) 1,2,4-Trimethylbenzene	13.482	105	183144	22.965	ug/l		99
85) sec-Butylbenzene	13.612	105	209133	22.457	ug/l		100
86) p-Isopropyltoluene	13.729	119	168888	21.661	ug/l		99
87) 1,3-Dichlorobenzene	13.729	146	97652	20.952	ug/l		100
88) 1,4-Dichlorobenzene	13.806	146	98321	20.351	ug/l		99
89) n-Butylbenzene	14.053	91	142673	21.355	ug/l		99
90) Hexachloroethane	14.329	117	35205	19.858	ug/l		98
91) 1,2-Dichlorobenzene	14.100	146	94954	20.425	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12859	20.580	ug/l		100
93) 1,2,4-Trichlorobenzene	15.388	180	45878	21.233	ug/l		99
94) Hexachlorobutadiene	15.500	225	22719	19.805	ug/l		95
95) Naphthalene	15.635	128	134793	20.906	ug/l		99
96) 1,2,3-Trichlorobenzene	15.835	180	45468	20.797	ug/l		97

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085440.D
Acq On : 14 Jan 2025 15:43
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:43:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085440.D
Acq On : 14 Jan 2025 15:43
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1


Quant Time: Jan 15 01:43:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:168 Resp: 200796


Ion	Ratio	Lower	Upper
168	100		
99	62.2	53.6	80.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#2
Dichlorodifluoromethane
Concen: 20.107 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 85 Resp: 54666

Ion	Ratio	Lower	Upper
85	100		
87	32.5	16.7	50.1






#3
Chloromethane
Concen: 19.848 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 50 Resp: 58426


Ion	Ratio	Lower	Upper
50	100		
52	34.0	27.0	40.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#4
Vinyl Chloride
Concen: 19.739 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 62 Resp: 58403

Ion	Ratio	Lower	Upper
62	100		
64	30.7	24.8	37.2






#5
Bromomethane
Concen: 20.389 ug/l
RT: 2.954 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 94 Resp: 36440


Ion	Ratio	Lower	Upper
94	100		
96	94.9	71.8	107.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#6
Chloroethane
Concen: 20.041 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 64 Resp: 37594

Ion	Ratio	Lower	Upper
64	100		
66	34.3	24.6	36.8






#7
Trichlorofluoromethane
Concen: 20.526 ug/l
RT: 3.501 min Scan# 26
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:101 Resp: 88121


Ion	Ratio	Lower	Upper
101	100		
103	65.3	51.4	77.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#8
Diethyl Ether
Concen: 19.647 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.012 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 74 Resp: 29138

Ion	Ratio	Lower	Upper
74	100		
45	101.8	49.7	149.1






#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.237 ug/l
RT: 4.371 min Scan# 41
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:101 Resp: 48933


Ion	Ratio	Lower	Upper
101	100		
85	46.6	37.8	56.8
151	72.1	58.8	88.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#10
Methyl Iodide
Concen: 20.428 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:142 Resp: 56567

Ion	Ratio	Lower	Upper
142	100		
127	44.6	40.3	60.5
141	14.8	12.8	19.2






#11
Tert butyl alcohol
Concen: 102.342 ug/l
RT: 5.506 min Scan# 60
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 59 Resp: 37984


Ion	Ratio	Lower	Upper
59	100		
57	10.7	8.6	13.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#12
1,1-Dichloroethene
Concen: 20.728 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 96 Resp: 44670

Ion	Ratio	Lower	Upper
96	100		
61	170.3	141.0	211.4
98	56.8	49.0	73.6






#13
Acrolein
Concen: 97.995 ug/l
RT: 4.171 min Scan# 37
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 56 Resp: 49649


Ion	Ratio	Lower	Upper
56	100		
55	71.4	56.3	84.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#14
Allyl chloride
Concen: 19.362 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 41 Resp: 67705

Ion	Ratio	Lower	Upper
41	100		
39	85.0	64.4	96.6
76	38.6	30.5	45.7






#15
Acrylonitrile
Concen: 103.171 ug/l
RT: 5.712 min Scan# 63
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 53 Resp: 121626


Ion	Ratio	Lower	Upper
53	100		
52	81.1	65.5	98.3
51	36.1	29.8	44.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#16
Acetone
Concen: 96.502 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 101065

Ion	Ratio	Lower	Upper
43	100		
58	27.7	23.8	35.6






#17
Carbon Disulfide
Concen: 19.939 ug/l
RT: 4.712 min Scan# 46
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 76 Resp: 132300


Ion	Ratio	Lower	Upper
76	100		
78	9.3	6.9	10.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#18
Methyl Acetate
Concen: 19.115 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 60873

Ion	Ratio	Lower	Upper
43	100		
74	24.0	18.6	28.0






#19
Methyl tert-butyl Ether
Concen: 21.272 ug/l
RT: 5.795 min Scan# 65
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 73 Resp: 148847


Ion	Ratio	Lower	Upper
73	100		
57	22.1	17.8	26.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#20
Methylene Chloride
Concen: 20.370 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 84 Resp: 52813

Ion	Ratio	Lower	Upper
84	100		
49	129.2	104.1	156.1
51	42.2	31.0	46.4
86	61.6	51.9	77.9






#21
trans-1,2-Dichloroethene
Concen: 20.023 ug/l
RT: 5.789 min Scan# 65
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 96 Resp: 46115


Ion	Ratio	Lower	Upper
96	100		
61	146.1	115.7	173.5
98	64.5	49.8	74.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#22
Diisopropyl ether
Concen: 21.480 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 45 Resp: 166719

Ion	Ratio	Lower	Upper
45	100		
43	54.3	45.0	67.6
87	25.3	21.8	32.8
59	10.8	10.2	15.2






#23
Vinyl Acetate
Concen: 107.984 ug/l m
RT: 6.594 min Scan# 78
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 588567


Ion	Ratio	Lower	Upper
43	100		
86	10.1	8.4	12.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#24
1,1-Dichloroethane
Concen: 20.469 ug/l
RT: 6.565 min Scan# 784
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 63 Resp: 96875

Ion	Ratio	Lower	Upper
63	100		
98	7.4	3.3	9.8
100	6.0	2.0	6.0#






#25
2-Butanone
Concen: 103.786 ug/l
RT: 7.477 min Scan# 93
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 159952


Ion	Ratio	Lower	Upper
43	100		
72	24.1	20.2	30.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#26
2,2-Dichloropropane
Concen: 20.684 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 77 Resp: 79145

Ion	Ratio	Lower	Upper
77	100		
97	22.0	10.7	32.1






#27
cis-1,2-Dichloroethene
Concen: 21.180 ug/l
RT: 7.483 min Scan# 94
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 96 Resp: 57454


Ion	Ratio	Lower	Upper
96	100		
61	153.0	0.0	311.8
98	63.7	0.0	126.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#28
Bromochloromethane
Concen: 18.703 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 49 Resp: 41181

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	4.4
130	69.3	55.0	82.4






#29
Tetrahydrofuran
Concen: 107.247 ug/l
RT: 7.836 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 42 Resp: 104792


Ion	Ratio	Lower	Upper
42	100		
72	42.6	34.2	51.2
71	41.4	32.5	48.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#30
Chloroform
Concen: 20.374 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 83 Resp: 99655

Ion	Ratio	Lower	Upper
83	100		
85	65.8	51.8	77.6






#31
Cyclohexane
Concen: 20.173 ug/l
RT: 8.253 min Scan# 10
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 56 Resp: 82980


Ion	Ratio	Lower	Upper
56	100		
69	31.2	22.2	33.4
84	80.8	66.4	99.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#32
1,1,1-Trichloroethane
Concen: 20.431 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 97 Resp: 87659

Ion	Ratio	Lower	Upper
97	100		
99	64.2	49.8	74.6
61	50.3	41.4	62.2






#33
1,2-Dichloroethane-d4
Concen: 18.694 ug/l
RT: 8.571 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 65 Resp: 60593


Ion	Ratio	Lower	Upper
65	100		
67	51.2	0.0	101.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.094 min Scan# 1214
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:114 Resp: 340593

Ion	Ratio	Lower	Upper
114	100		
63	21.6	0.0	47.6
88	16.0	0.0	32.6






#35
Dibromofluoromethane
Concen: 19.322 ug/l
RT: 8.165 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:113 Resp: 45655


Ion	Ratio	Lower	Upper
113	100		
111	100.6	82.7	124.1
192	17.9	14.3	21.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#36
1,1-Dichloropropene
Concen: 20.490 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 67949

Ion	Ratio	Lower	Upper
75	100		
110	31.8	16.5	49.5
77	31.1	24.4	36.6





#37
Ethyl Acetate
Concen: 19.850 ug/l
RT: 7.553 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43


Tgt Ion: 43 Resp: 66443


Ion	Ratio	Lower	Upper
43	100		
61	11.9	10.9	16.3
70	10.6	7.5	11.3

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025







#38
Carbon Tetrachloride
Concen: 20.767 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:117 Resp: 78842

Ion	Ratio	Lower	Upper
117	100		
119	98.9	75.4	113.2
121	28.9	24.6	37.0






#39
Methylcyclohexane
Concen: 20.862 ug/l
RT: 9.594 min Scan# 12
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 83 Resp: 65012


Ion	Ratio	Lower	Upper
83	100		
55	83.4	62.6	94.0
98	46.4	37.7	56.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#40
Benzene
Concen: 20.873 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 78 Resp: 207985

Ion	Ratio	Lower	Upper
78	100		
77	24.4	19.0	28.6






#41
Methacrylonitrile
Concen: 21.004 ug/l
RT: 7.771 min Scan# 98
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 41 Resp: 36578


Ion	Ratio	Lower	Upper
41	100		
39	56.6	46.0	69.0
67	70.3	57.4	86.2
52	32.4	25.5	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#42
1,2-Dichloroethane
Concen: 20.891 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 62 Resp: 78402

Ion	Ratio	Lower	Upper
62	100		
98	8.3	0.0	17.0






#43
Isopropyl Acetate
Concen: 20.347 ug/l
RT: 8.683 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 109131


Ion	Ratio	Lower	Upper
43	100		
61	27.2	20.7	31.1
87	12.3	9.8	14.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#44
Trichloroethene
Concen: 20.699 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:130 Resp: 48009

Ion	Ratio	Lower	Upper
130	100		
95	101.5	0.0	195.8






#45
1,2-Dichloropropane
Concen: 20.758 ug/l
RT: 9.618 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 63 Resp: 52833


Ion	Ratio	Lower	Upper
63	100		
65	28.3	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#46
Dibromomethane
Concen: 20.540 ug/l
RT: 9.700 min Scan# 1317
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 93 Resp: 37721

Ion	Ratio	Lower	Upper
93	100		
95	82.0	64.7	97.1
174	88.4	69.0	103.4






#47
Bromodichloromethane
Concen: 21.093 ug/l
RT: 9.882 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 83 Resp: 78918


Ion	Ratio	Lower	Upper
83	100		
85	67.8	51.2	76.8
127	8.6	6.5	9.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#48
Methyl methacrylate
Concen: 20.808 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 41 Resp: 50224

Ion	Ratio	Lower	Upper
41	100		
69	84.3	64.7	97.1
39	64.2	49.0	73.6






#49
1,4-Dioxane
Concen: 421.686 ug/l
RT: 9.688 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 88 Resp: 17140


Ion	Ratio	Lower	Upper
88	100		
43	34.9	26.6	39.8
58	69.0	59.5	89.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#50
Toluene-d8
Concen: 19.581 ug/l
RT: 10.559 min Scan# 1463
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 98 Resp: 164391

Ion	Ratio	Lower	Upper
98	100		
100	64.4	52.2	78.4





#51
4-Methyl-2-Pentanone
Concen: 108.249 ug/l
RT: 10.441 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43


Tgt Ion: 43 Resp: 336911

Ion	Ratio	Lower	Upper
43	100		
58	38.1	30.5	45.7

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#52
Toluene
Concen: 21.680 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 92 Resp: 125172


Ion	Ratio	Lower	Upper
92	100		
91	175.1	139.2	208.8



#53
t-1,3-Dichloropropene
Concen: 20.965 ug/l
RT: 10.835 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 74129


Ion	Ratio	Lower	Upper
75	100		
77	31.4	23.5	35.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#54
cis-1,3-Dichloropropene
Concen: 21.684 ug/l
RT: 10.306 min Scan# 1420
Delta R.T. -0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 81897

Ion	Ratio	Lower	Upper
75	100		
77	32.4	25.0	37.4
39	54.4	43.1	64.7






#55
1,1,2-Trichloroethane
Concen: 21.025 ug/l
RT: 11.012 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 97 Resp: 48039


Ion	Ratio	Lower	Upper
97	100		
83	84.1	69.0	103.6
85	54.6	44.6	66.8
99	64.8	48.6	72.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#56
Ethyl methacrylate
Concen: 19.649 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 69 Resp: 70934

Ion	Ratio	Lower	Upper
69	100		
41	67.6	54.6	82.0
39	41.7	32.4	48.6






#57
1,3-Dichloropropane
Concen: 21.196 ug/l
RT: 11.159 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 76 Resp: 84210


Ion	Ratio	Lower	Upper
76	100		
78	32.7	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#58
2-Chloroethyl Vinyl ether
Concen: 105.991 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 63 Resp: 153681

Ion	Ratio	Lower	Upper
63	100		
106	27.2	21.6	32.4






#59
2-Hexanone
Concen: 109.863 ug/l
RT: 11.194 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 240595


Ion	Ratio	Lower	Upper
43	100		
58	51.1	26.2	78.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#60
Dibromochloromethane
Concen: 20.326 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:129 Resp: 56069

Ion	Ratio	Lower	Upper
129	100		
127	80.4	38.6	115.8






#61
1,2-Dibromoethane
Concen: 21.309 ug/l
RT: 11.465 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:107 Resp: 48465


Ion	Ratio	Lower	Upper
107	100		
109	90.6	75.9	113.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#62
4-Bromofluorobenzene
Concen: 19.430 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 95 Resp: 55798

Ion	Ratio	Lower	Upper
95	100		
174	75.2	0.0	145.0
176	73.2	0.0	142.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:117 Resp: 296298


Ion	Ratio	Lower	Upper
117	100		
82	59.3	48.6	72.8
119	32.2	26.6	39.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#64
Tetrachloroethene
Concen: 21.408 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:164 Resp: 43243

Ion	Ratio	Lower	Upper
164	100		
166	130.6	103.4	155.2
129	107.4	79.2	118.8
131	94.7	77.1	115.7






#65
Chlorobenzene
Concen: 21.077 ug/l
RT: 11.888 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:112 Resp: 136809


Ion	Ratio	Lower	Upper
112	100		
114	32.9	25.3	37.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 20.304 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:131 Resp: 48370

Ion	Ratio	Lower	Upper
131	100		
133	96.9	47.4	142.3
119	67.2	33.1	99.5






#67
Ethyl Benzene
Concen: 21.753 ug/l
RT: 11.959 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 91 Resp: 229978


Ion	Ratio	Lower	Upper
91	100		
106	31.1	23.8	35.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#68
m/p-Xylenes
Concen: 45.494 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:106 Resp: 177755

Ion	Ratio	Lower	Upper
106	100		
91	209.3	167.7	251.5






#69
o-Xylene
Concen: 22.643 ug/l
RT: 12.394 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:106 Resp: 84548


Ion	Ratio	Lower	Upper
106	100		
91	217.0	110.4	331.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#70
Styrene
Concen: 22.749 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:104 Resp: 140581

Ion	Ratio	Lower	Upper
104	100		
78	54.4	42.5	63.7
103	54.8	43.8	65.8






#71
Bromoform
Concen: 21.683 ug/l
RT: 12.576 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:173 Resp: 36961


Ion	Ratio	Lower	Upper
173	100		
175	51.7	24.4	73.2
254	0.0	0.0	0.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:152 Resp: 143526

Ion	Ratio	Lower	Upper
152	100		
115	61.4	31.1	93.3
150	163.9	0.0	343.6






#73
Isopropylbenzene
Concen: 21.819 ug/l
RT: 12.694 min Scan# 18
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:105 Resp: 211354


Ion	Ratio	Lower	Upper
105	100		
120	25.5	12.8	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#74
N-amyl acetate
Concen: 21.656 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 43 Resp: 94292

Ion	Ratio	Lower	Upper
43	100		
70	41.4	34.0	51.0
55	27.9	21.4	32.2
61	23.1	19.0	28.4






#75
1,1,2,2-Tetrachloroethane
Concen: 19.917 ug/l
RT: 12.935 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 83 Resp: 68151


Ion	Ratio	Lower	Upper
83	100		
131	9.9	4.8	14.4
85	65.5	32.2	96.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#76
1,2,3-Trichloropropane
Concen: 22.036 ug/l m
RT: 12.988 min Scan# 1876
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 64180

Ion	Ratio	Lower	Upper
75	100		
77	186.8	109.7	329.2






#77
Bromobenzene
Concen: 20.556 ug/l
RT: 12.976 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:156 Resp: 52022


Ion	Ratio	Lower	Upper
156	100		
77	230.4	114.1	342.4
158	97.0	48.9	146.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#78
n-propylbenzene
Concen: 22.028 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 91 Resp: 252610

Ion	Ratio	Lower	Upper
91	100		
120	21.8	10.9	32.6






#79
2-Chlorotoluene
Concen: 21.358 ug/l
RT: 13.123 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 91 Resp: 158511


Ion	Ratio	Lower	Upper
91	100		
126	31.5	15.7	47.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#80
1,3,5-Trimethylbenzene
Concen: 22.533 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:105 Resp: 180294

Ion	Ratio	Lower	Upper
105	100		
120	47.9	23.9	71.7






#81
trans-1,4-Dichloro-2-butene
Concen: 20.569 ug/l m
RT: 12.735 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 22178


Ion	Ratio	Lower	Upper
75	100		
53	107.2	95.6	143.4
89	46.9	37.0	55.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#82
4-Chlorotoluene
Concen: 21.906 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 91 Resp: 161880

Ion	Ratio	Lower	Upper
91	100		
126	31.6	15.9	47.7






#83
tert-Butylbenzene
Concen: 21.958 ug/l
RT: 13.435 min Scan# 19
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:119 Resp: 147541


Ion	Ratio	Lower	Upper
119	100		
91	69.5	35.1	105.3
134	25.5	12.0	36.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#84
1,2,4-Trimethylbenzene
Concen: 22.965 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:105 Resp: 183144

Ion	Ratio	Lower	Upper
105	100		
120	42.8	21.6	65.0






#85
sec-Butylbenzene
Concen: 22.457 ug/l
RT: 13.612 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:105 Resp: 209133


Ion	Ratio	Lower	Upper
105	100		
134	19.2	9.7	28.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#86
p-Isopropyltoluene
Concen: 21.661 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:119 Resp: 168888

Ion	Ratio	Lower	Upper
119	100		
134	26.4	12.7	38.0
91	25.5	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 20.952 ug/l
RT: 13.729 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:146 Resp: 97652


Ion	Ratio	Lower	Upper
146	100		
111	42.9	21.4	64.3
148	64.3	32.3	96.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#88
1,4-Dichlorobenzene
Concen: 20.351 ug/l
RT: 13.806 min Scan# 2015
Delta R.T. -0.006 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:146 Resp: 98321

Ion	Ratio	Lower	Upper
146	100		
111	42.2	21.3	63.7
148	65.1	32.4	97.0






#89
n-Butylbenzene
Concen: 21.355 ug/l
RT: 14.053 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 91 Resp: 142673


Ion	Ratio	Lower	Upper
91	100		
92	51.0	25.8	77.3
134	23.0	11.7	35.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#90
Hexachloroethane
Concen: 19.858 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:117 Resp: 35205

Ion	Ratio	Lower	Upper
117	100		
201	68.8	33.7	101.0






#91
1,2-Dichlorobenzene
Concen: 20.425 ug/l
RT: 14.100 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:146 Resp: 94954


Ion	Ratio	Lower	Upper
146	100		
111	43.5	21.7	65.1
148	63.5	31.4	94.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 20.580 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion: 75 Resp: 12859

Ion	Ratio	Lower	Upper
75	100		
155	72.7	36.4	109.2
157	90.9	45.4	136.1






#93
1,2,4-Trichlorobenzene
Concen: 21.233 ug/l
RT: 15.388 min Scan# 22
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:180 Resp: 45878


Ion	Ratio	Lower	Upper
180	100		
182	94.2	46.8	140.3
145	31.4	16.0	48.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#94
Hexachlorobutadiene
Concen: 19.805 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:225 Resp: 22719

Ion	Ratio	Lower	Upper
225	100		
223	64.3	30.7	92.1
227	66.3	30.9	92.5





#95
Naphthalene
Concen: 20.906 ug/l
RT: 15.635 min Scan# 23
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43


Tgt Ion:128 Resp: 134793


Ion	Ratio	Lower	Upper
128	100		
127	13.3	10.6	16.0
129	10.6	8.8	13.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025






#96
1,2,3-Trichlorobenzene
Concen: 20.797 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN085440.D
Acq: 14 Jan 2025 15:43

Tgt Ion:180 Resp: 45468

Ion	Ratio	Lower	Upper
180	100		
182	97.9	47.4	142.2
145	35.7	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085441.D
Acq On : 14 Jan 2025 16:07
Operator : JC\MD
Sample : VSTDICC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jan 15 01:44:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	195889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	340403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	294508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	29856	9.442	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 18.880%#			
35) Dibromofluoromethane	8.165	113	21110	8.939	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 17.880%#			
50) Toluene-d8	10.565	98	73223	8.727	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 17.460%#			
62) 4-Bromofluorobenzene	12.847	95	24327	8.476	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 16.960%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.124	85	26124	9.850	ug/l	96
3) Chloromethane	2.359	50	27143	9.452	ug/l	99
4) Vinyl Chloride	2.512	62	27862	9.653	ug/l	100
5) Bromomethane	2.965	94	17139	9.830	ug/l	87
6) Chloroethane	3.130	64	16791	9.175	ug/l	96
7) Trichlorofluoromethane	3.501	101	40755	9.731	ug/l	98
8) Diethyl Ether	3.953	74	12415	8.581	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.371	101	22982	9.742	ug/l	98
10) Methyl Iodide	4.589	142	25085	9.286	ug/l	99
11) Tert butyl alcohol	5.512	59	17962	49.608	ug/l	97
12) 1,1-Dichloroethene	4.342	96	20615	9.806	ug/l	99
13) Acrolein	4.177	56	25568	51.729	ug/l	99
14) Allyl chloride	5.018	41	32240	9.451	ug/l	97
15) Acrylonitrile	5.712	53	53959	46.918	ug/l	98
16) Acetone	4.424	43	48475	47.446	ug/l	98
17) Carbon Disulfide	4.712	76	60201	9.300	ug/l #	94
18) Methyl Acetate	5.030	43	30500	9.817	ug/l	98
19) Methyl tert-butyl Ether	5.789	73	65207	9.552	ug/l	91
20) Methylene Chloride	5.277	84	23723	9.379	ug/l	96
21) trans-1,2-Dichloroethene	5.783	96	21127	9.403	ug/l	97
22) Diisopropyl ether	6.671	45	71857	9.490	ug/l	97
23) Vinyl Acetate	6.600	43	250575	47.125	ug/l	98
24) 1,1-Dichloroethane	6.565	63	43099	9.335	ug/l	95
25) 2-Butanone	7.477	43	71191	47.350	ug/l	99
26) 2,2-Dichloropropane	7.483	77	36105	9.672	ug/l	99
27) cis-1,2-Dichloroethene	7.489	96	25049	9.466	ug/l	97
28) Bromochloromethane	7.806	49	19037	8.863	ug/l	97
29) Tetrahydrofuran	7.836	42	45641	47.881	ug/l	98
30) Chloroform	7.965	83	45792	9.596	ug/l	100
31) Cyclohexane	8.253	56	40211	10.020	ug/l #	93
32) 1,1,1-Trichloroethane	8.165	97	39170	9.358	ug/l	99
36) 1,1-Dichloropropene	8.371	75	30015	9.056	ug/l	97
37) Ethyl Acetate	7.559	43	30760	9.195	ug/l	97
38) Carbon Tetrachloride	8.359	117	36015	9.492	ug/l	94
39) Methylcyclohexane	9.594	83	27721	8.900	ug/l	95
40) Benzene	8.600	78	93680	9.407	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085441.D
Acq On : 14 Jan 2025 16:07
Operator : JC\MD
Sample : VSTDICC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jan 15 01:44:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	7.777	41	16859	9.686	ug/l		96
42) 1,2-Dichloroethane	8.665	62	35569	9.483	ug/l		99
43) Isopropyl Acetate	8.683	43	48926	9.127	ug/l		99
44) Trichloroethene	9.347	130	21113	9.108	ug/l		95
45) 1,2-Dichloropropane	9.618	63	22747	8.942	ug/l		97
46) Dibromomethane	9.706	93	17260	9.404	ug/l		99
47) Bromodichloromethane	9.883	83	35020	9.365	ug/l		98
48) Methyl methacrylate	9.677	41	21873	9.067	ug/l		97
49) 1,4-Dioxane	9.694	88	7734	190.382	ug/l		98
51) 4-Methyl-2-Pentanone	10.441	43	146990	47.254	ug/l		98
52) Toluene	10.630	92	54976	9.527	ug/l		100
53) t-1,3-Dichloropropene	10.835	75	32780	9.276	ug/l		96
54) cis-1,3-Dichloropropene	10.312	75	35904	9.512	ug/l		98
55) 1,1,2-Trichloroethane	11.012	97	21060	9.222	ug/l		98
56) Ethyl methacrylate	10.871	69	28454	9.042	ug/l		95
57) 1,3-Dichloropropane	11.159	76	37768	9.512	ug/l		98
58) 2-Chloroethyl Vinyl ether	10.159	63	63110	43.550	ug/l		98
59) 2-Hexanone	11.194	43	101318	46.291	ug/l		98
60) Dibromochloromethane	11.353	129	25029	9.079	ug/l		98
61) 1,2-Dibromoethane	11.465	107	21055	9.263	ug/l		100
64) Tetrachloroethene	11.100	164	19934	9.928	ug/l		97
65) Chlorobenzene	11.888	112	61680	9.560	ug/l		93
66) 1,1,1,2-Tetrachloroethane	11.959	131	21853	9.229	ug/l		98
67) Ethyl Benzene	11.959	91	99248	9.445	ug/l		97
68) m/p-Xylenes	12.065	106	72431	18.650	ug/l		99
69) o-Xylene	12.394	106	34382	9.264	ug/l		99
70) Styrene	12.406	104	56285	9.163	ug/l		98
71) Bromoform	12.576	173	16071	9.485	ug/l	#	98
73) Isopropylbenzene	12.694	105	87003	9.697	ug/l		99
74) N-amyl acetate	12.494	43	37343	9.260	ug/l		97
75) 1,1,2,2-Tetrachloroethane	12.935	83	30770	9.709	ug/l		99
76) 1,2,3-Trichloropropane	12.994	75	25443m	9.432	ug/l		
77) Bromobenzene	12.976	156	22784	9.720	ug/l		97
78) n-propylbenzene	13.035	91	102262	9.628	ug/l		98
79) 2-Chlorotoluene	13.123	91	68346	9.943	ug/l		100
80) 1,3,5-Trimethylbenzene	13.171	105	71656	9.669	ug/l		99
81) trans-1,4-Dichloro-2-b...	12.735	75	9236	9.248	ug/l		100
82) 4-Chlorotoluene	13.218	91	66773	9.756	ug/l		97
83) tert-Butylbenzene	13.435	119	59455	9.553	ug/l		98
84) 1,2,4-Trimethylbenzene	13.476	105	72046	9.754	ug/l		99
85) sec-Butylbenzene	13.612	105	83662	9.699	ug/l		99
86) p-Isopropyltoluene	13.729	119	67556	9.674	ug/l		99
87) 1,3-Dichlorobenzene	13.729	146	41843	9.693	ug/l		99
88) 1,4-Dichlorobenzene	13.812	146	42728	9.549	ug/l		99
89) n-Butylbenzene	14.053	91	57632	9.314	ug/l		97
90) Hexachloroethane	14.329	117	15261	9.294	ug/l		97
91) 1,2-Dichlorobenzene	14.106	146	40726	9.458	ug/l		100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	5639	9.744	ug/l		95
93) 1,2,4-Trichlorobenzene	15.388	180	18720	9.354	ug/l		99
94) Hexachlorobutadiene	15.500	225	10491	9.874	ug/l		98
95) Naphthalene	15.635	128	52054	8.717	ug/l		99
96) 1,2,3-Trichlorobenzene	15.835	180	19450	9.605	ug/l		96

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085441.D
Acq On : 14 Jan 2025 16:07
Operator : JC\MD
Sample : VSTDICC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jan 15 01:44:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085441.D
Acq On : 14 Jan 2025 16:07
Operator : JC\MD
Sample : VSTDICC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1


Quant Time: Jan 15 01:44:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:168 Resp: 195889


Ion	Ratio	Lower	Upper
168	100		
99	61.7	53.6	80.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#2
Dichlorodifluoromethane
Concen: 9.850 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 85 Resp: 26124

Ion	Ratio	Lower	Upper
85	100		
87	35.8	16.7	50.1






#3
Chloromethane
Concen: 9.452 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 50 Resp: 27143


Ion	Ratio	Lower	Upper
50	100		
52	34.3	27.0	40.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#4
Vinyl Chloride
Concen: 9.653 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 62 Resp: 27862

Ion	Ratio	Lower	Upper
62	100		
64	31.2	24.8	37.2






#5
Bromomethane
Concen: 9.830 ug/l
RT: 2.965 min Scan# 17
Delta R.T. 0.012 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 94 Resp: 17139


Ion	Ratio	Lower	Upper
94	100		
96	102.1	71.8	107.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#6
Chloroethane
Concen: 9.175 ug/l
RT: 3.130 min Scan# 200
Delta R.T. 0.012 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 64 Resp: 16791

Ion	Ratio	Lower	Upper
64	100		
66	33.1	24.6	36.8






#7
Trichlorofluoromethane
Concen: 9.731 ug/l
RT: 3.501 min Scan# 26
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:101 Resp: 40755


Ion	Ratio	Lower	Upper
101	100		
103	65.9	51.4	77.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#8
Diethyl Ether
Concen: 8.581 ug/l
RT: 3.953 min Scan# 340
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 74 Resp: 12415

Ion	Ratio	Lower	Upper
74	100		
45	105.3	49.7	149.1






#9
1,1,2-Trichlorotrifluoroethane
Concen: 9.742 ug/l
RT: 4.371 min Scan# 41
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:101 Resp: 22982


Ion	Ratio	Lower	Upper
101	100		
85	45.8	37.8	56.8
151	75.2	58.8	88.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#10
Methyl Iodide
Concen: 9.286 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:142 Resp: 25085

Ion	Ratio	Lower	Upper
142	100		
127	50.9	40.3	60.5
141	15.0	12.8	19.2






#11
Tert butyl alcohol
Concen: 49.608 ug/l
RT: 5.512 min Scan# 60
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 59 Resp: 17962


Ion	Ratio	Lower	Upper
59	100		
57	11.9	8.6	13.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#12
1,1-Dichloroethene
Concen: 9.806 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 96 Resp: 20615

Ion	Ratio	Lower	Upper
96	100		
61	176.8	141.0	211.4
98	60.5	49.0	73.6






#13
Acrolein
Concen: 51.729 ug/l
RT: 4.177 min Scan# 37
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 56 Resp: 25568


Ion	Ratio	Lower	Upper
56	100		
55	70.9	56.3	84.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#14
Allyl chloride
Concen: 9.451 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 41 Resp: 32240

Ion	Ratio	Lower	Upper
41	100		
39	78.3	64.4	96.6
76	35.1	30.5	45.7






#15
Acrylonitrile
Concen: 46.918 ug/l
RT: 5.712 min Scan# 63
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 53 Resp: 53959


Ion	Ratio	Lower	Upper
53	100		
52	83.3	65.5	98.3
51	38.8	29.8	44.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#16
Acetone
Concen: 47.446 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 48475

Ion	Ratio	Lower	Upper
43	100		
58	30.8	23.8	35.6






#17
Carbon Disulfide
Concen: 9.300 ug/l
RT: 4.712 min Scan# 46
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 76 Resp: 60201


Ion	Ratio	Lower	Upper
76	100		
78	10.7	6.9	10.3#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#18
Methyl Acetate
Concen: 9.817 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 30500

Ion	Ratio	Lower	Upper
43	100		
74	24.3	18.6	28.0






#19
Methyl tert-butyl Ether
Concen: 9.552 ug/l
RT: 5.789 min Scan# 65
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 73 Resp: 65207


Ion	Ratio	Lower	Upper
73	100		
57	18.1	17.8	26.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#20
Methylene Chloride
Concen: 9.379 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 84 Resp: 23723

Ion	Ratio	Lower	Upper
84	100		
49	132.4	104.1	156.1
51	34.4	31.0	46.4
86	70.0	51.9	77.9






#21
trans-1,2-Dichloroethene
Concen: 9.403 ug/l
RT: 5.783 min Scan# 65
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 96 Resp: 21127


Ion	Ratio	Lower	Upper
96	100		
61	140.8	115.7	173.5
98	60.5	49.8	74.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#22
Diisopropyl ether
Concen: 9.490 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 45 Resp: 71857

Ion	Ratio	Lower	Upper
45	100		
43	53.2	45.0	67.6
87	28.8	21.8	32.8
59	13.2	10.2	15.2






#23
Vinyl Acetate
Concen: 47.125 ug/l
RT: 6.600 min Scan# 79
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 250575


Ion	Ratio	Lower	Upper
43	100		
86	9.8	8.4	12.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#24
1,1-Dichloroethane
Concen: 9.335 ug/l
RT: 6.565 min Scan# 784
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 63 Resp: 43099

Ion	Ratio	Lower	Upper
63	100		
98	8.8	3.3	9.8
100	5.0	2.0	6.0






#25
2-Butanone
Concen: 47.350 ug/l
RT: 7.477 min Scan# 93
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 71191


Ion	Ratio	Lower	Upper
43	100		
72	24.8	20.2	30.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#26
2,2-Dichloropropane
Concen: 9.672 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 77 Resp: 36105

Ion	Ratio	Lower	Upper
77	100		
97	20.8	10.7	32.1






#27
cis-1,2-Dichloroethene
Concen: 9.466 ug/l
RT: 7.489 min Scan# 94
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 96 Resp: 25049


Ion	Ratio	Lower	Upper
96	100		
61	151.8	0.0	311.8
98	64.7	0.0	126.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#28
Bromochloromethane
Concen: 8.863 ug/l
RT: 7.806 min Scan# 995
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 49 Resp: 19037

Ion	Ratio	Lower	Upper
49	100		
129	0.3	0.0	4.4
130	70.8	55.0	82.4






#29
Tetrahydrofuran
Concen: 47.881 ug/l
RT: 7.836 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 42 Resp: 45641


Ion	Ratio	Lower	Upper
42	100		
72	41.6	34.2	51.2
71	39.2	32.5	48.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#30
Chloroform
Concen: 9.596 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 83 Resp: 45792

Ion	Ratio	Lower	Upper
83	100		
85	64.9	51.8	77.6






#31
Cyclohexane
Concen: 10.020 ug/l
RT: 8.253 min Scan# 10
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 56 Resp: 40211


Ion	Ratio	Lower	Upper
56	100		
69	34.2	22.2	33.4#
84	87.2	66.4	99.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#32
1,1,1-Trichloroethane
Concen: 9.358 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 97 Resp: 39170

Ion	Ratio	Lower	Upper
97	100		
99	62.6	49.8	74.6
61	50.2	41.4	62.2






#33
1,2-Dichloroethane-d4
Concen: 9.442 ug/l
RT: 8.571 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 65 Resp: 29856


Ion	Ratio	Lower	Upper
65	100		
67	49.5	0.0	101.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:114 Resp: 340403

Ion	Ratio	Lower	Upper
114	100		
63	21.1	0.0	47.6
88	15.5	0.0	32.6






#35
Dibromofluoromethane
Concen: 8.939 ug/l
RT: 8.165 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:113 Resp: 21110


Ion	Ratio	Lower	Upper
113	100		
111	103.4	82.7	124.1
192	16.4	14.3	21.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#36
1,1-Dichloropropene
Concen: 9.056 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 30015

Ion	Ratio	Lower	Upper
75	100		
110	34.3	16.5	49.5
77	32.9	24.4	36.6






#37
Ethyl Acetate
Concen: 9.195 ug/l
RT: 7.559 min Scan# 95
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 30760


Ion	Ratio	Lower	Upper
43	100		
61	12.2	10.9	16.3
70	8.4	7.5	11.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#38
Carbon Tetrachloride
Concen: 9.492 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:117 Resp: 36015

Ion	Ratio	Lower	Upper
117	100		
119	101.2	75.4	113.2
121	29.1	24.6	37.0






#39
Methylcyclohexane
Concen: 8.900 ug/l
RT: 9.594 min Scan# 12
Delta R.T. -0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 83 Resp: 27721


Ion	Ratio	Lower	Upper
83	100		
55	83.8	62.6	94.0
98	45.8	37.7	56.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#40
Benzene
Concen: 9.407 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 78 Resp: 93680

Ion	Ratio	Lower	Upper
78	100		
77	22.0	19.0	28.6






#41
Methacrylonitrile
Concen: 9.686 ug/l
RT: 7.777 min Scan# 99
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 41 Resp: 16859


Ion	Ratio	Lower	Upper
41	100		
39	57.3	46.0	69.0
67	66.2	57.4	86.2
52	30.4	25.5	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#42
1,2-Dichloroethane
Concen: 9.483 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 62 Resp: 35569

Ion	Ratio	Lower	Upper
62	100		
98	8.2	0.0	17.0






#43
Isopropyl Acetate
Concen: 9.127 ug/l
RT: 8.683 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 48926


Ion	Ratio	Lower	Upper
43	100		
61	26.2	20.7	31.1
87	12.8	9.8	14.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#44
Trichloroethene
Concen: 9.108 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:130 Resp: 21113

Ion	Ratio	Lower	Upper
130	100		
95	92.9	0.0	195.8






#45
1,2-Dichloropropane
Concen: 8.942 ug/l
RT: 9.618 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 63 Resp: 22747


Ion	Ratio	Lower	Upper
63	100		
65	30.4	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#46
Dibromomethane
Concen: 9.404 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 93 Resp: 17260

Ion	Ratio	Lower	Upper
93	100		
95	81.5	64.7	97.1
174	87.8	69.0	103.4






#47
Bromodichloromethane
Concen: 9.365 ug/l
RT: 9.883 min Scan# 13
Delta R.T. 0.001 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 83 Resp: 35020


Ion	Ratio	Lower	Upper
83	100		
85	62.6	51.2	76.8
127	6.9	6.5	9.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#48
Methyl methacrylate
Concen: 9.067 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 41 Resp: 21873

Ion	Ratio	Lower	Upper
41	100		
69	80.0	64.7	97.1
39	65.2	49.0	73.6





#49
1,4-Dioxane
Concen: 190.382 ug/l
RT: 9.694 min Scan# 13
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07


Tgt Ion: 88 Resp: 7734


Ion	Ratio	Lower	Upper
88	100		
43	35.2	26.6	39.8
58	73.3	59.5	89.3

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025







#50
Toluene-d8
Concen: 8.727 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 98 Resp: 73223

Ion	Ratio	Lower	Upper
98	100		
100	64.5	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 47.254 ug/l
RT: 10.441 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 146990


Ion	Ratio	Lower	Upper
43	100		
58	37.1	30.5	45.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#52
Toluene
Concen: 9.527 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 92 Resp: 54976

Ion	Ratio	Lower	Upper
92	100		
91	173.7	139.2	208.8






#53
t-1,3-Dichloropropene
Concen: 9.276 ug/l
RT: 10.835 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 32780


Ion	Ratio	Lower	Upper
75	100		
77	27.2	23.5	35.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#54
cis-1,3-Dichloropropene
Concen: 9.512 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 35904

Ion	Ratio	Lower	Upper
75	100		
77	30.0	25.0	37.4
39	53.1	43.1	64.7






#55
1,1,2-Trichloroethane
Concen: 9.222 ug/l
RT: 11.012 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 97 Resp: 21060


Ion	Ratio	Lower	Upper
97	100		
83	88.3	69.0	103.6
85	57.3	44.6	66.8
99	60.2	48.6	72.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#56
Ethyl methacrylate
Concen: 9.042 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 69 Resp: 28454

Ion	Ratio	Lower	Upper
69	100		
41	72.1	54.6	82.0
39	43.4	32.4	48.6






#57
1,3-Dichloropropane
Concen: 9.512 ug/l
RT: 11.159 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 76 Resp: 37768


Ion	Ratio	Lower	Upper
76	100		
78	31.1	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#58
2-Chloroethyl Vinyl ether
Concen: 43.550 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 63 Resp: 63110

Ion	Ratio	Lower	Upper
63	100		
106	26.1	21.6	32.4





#59
2-Hexanone
Concen: 46.291 ug/l
RT: 11.194 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07


Tgt Ion: 43 Resp: 101318


Ion	Ratio	Lower	Upper
43	100		
58	50.9	26.2	78.6

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025







#60
Dibromochloromethane
Concen: 9.079 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:129 Resp: 25029

Ion	Ratio	Lower	Upper
129	100		
127	79.1	38.6	115.8






#61
1,2-Dibromoethane
Concen: 9.263 ug/l
RT: 11.465 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:107 Resp: 21055


Ion	Ratio	Lower	Upper
107	100		
109	94.6	75.9	113.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#62
4-Bromofluorobenzene
Concen: 8.476 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 95 Resp: 24327

Ion	Ratio	Lower	Upper
95	100		
174	73.9	0.0	145.0
176	72.1	0.0	142.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:117 Resp: 294508


Ion	Ratio	Lower	Upper
117	100		
82	58.7	48.6	72.8
119	32.2	26.6	39.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#64
Tetrachloroethene
Concen: 9.928 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:164 Resp: 19934

Ion	Ratio	Lower	Upper
164	100		
166	132.2	103.4	155.2
129	98.3	79.2	118.8
131	90.5	77.1	115.7






#65
Chlorobenzene
Concen: 9.560 ug/l
RT: 11.888 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:112 Resp: 61680


Ion	Ratio	Lower	Upper
112	100		
114	27.6	25.3	37.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 9.229 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:131 Resp: 21853

Ion	Ratio	Lower	Upper
131	100		
133	98.5	47.4	142.3
119	66.0	33.1	99.5






#67
Ethyl Benzene
Concen: 9.445 ug/l
RT: 11.959 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 91 Resp: 99248


Ion	Ratio	Lower	Upper
91	100		
106	28.2	23.8	35.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#68
m/p-Xylenes
Concen: 18.650 ug/l
RT: 12.065 min Scan# 1719
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:106 Resp: 72431

Ion	Ratio	Lower	Upper
106	100		
91	211.7	167.7	251.5






#69
o-Xylene
Concen: 9.264 ug/l
RT: 12.394 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:106 Resp: 34382


Ion	Ratio	Lower	Upper
106	100		
91	219.5	110.4	331.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#70
Styrene
Concen: 9.163 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:104 Resp: 56285

Ion	Ratio	Lower	Upper
104	100		
78	53.8	42.5	63.7
103	56.8	43.8	65.8






#71
Bromoform
Concen: 9.485 ug/l
RT: 12.576 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:173 Resp: 16071


Ion	Ratio	Lower	Upper
173	100		
175	50.0	24.4	73.2
254	0.0	0.0	0.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:152 Resp: 132934

Ion	Ratio	Lower	Upper
152	100		
115	62.3	31.1	93.3
150	159.8	0.0	343.6






#73
Isopropylbenzene
Concen: 9.697 ug/l
RT: 12.694 min Scan# 18
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:105 Resp: 87003


Ion	Ratio	Lower	Upper
105	100		
120	26.1	12.8	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#74
N-amyl acetate
Concen: 9.260 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 43 Resp: 37343

Ion	Ratio	Lower	Upper
43	100		
70	41.4	34.0	51.0
55	28.5	21.4	32.2
61	22.0	19.0	28.4






#75
1,1,2,2-Tetrachloroethane
Concen: 9.709 ug/l
RT: 12.935 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 83 Resp: 30770


Ion	Ratio	Lower	Upper
83	100		
131	9.6	4.8	14.4
85	63.6	32.2	96.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#76
1,2,3-Trichloropropane
Concen: 9.432 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 25443

Ion	Ratio	Lower	Upper
75	100		
77	198.5	109.7	329.2






#77
Bromobenzene
Concen: 9.720 ug/l
RT: 12.976 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:156 Resp: 22784


Ion	Ratio	Lower	Upper
156	100		
77	221.6	114.1	342.4
158	98.7	48.9	146.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#78
n-propylbenzene
Concen: 9.628 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 91 Resp: 102262

Ion	Ratio	Lower	Upper
91	100		
120	20.9	10.9	32.6






#79
2-Chlorotoluene
Concen: 9.943 ug/l
RT: 13.123 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 91 Resp: 68346


Ion	Ratio	Lower	Upper
91	100		
126	31.5	15.7	47.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#80
1,3,5-Trimethylbenzene
Concen: 9.669 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:105 Resp: 71656

Ion	Ratio	Lower	Upper
105	100		
120	46.8	23.9	71.7






#81
trans-1,4-Dichloro-2-butene
Concen: 9.248 ug/l
RT: 12.735 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 9236


Ion	Ratio	Lower	Upper
75	100		
53	119.6	95.6	143.4
89	46.6	37.0	55.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#82
4-Chlorotoluene
Concen: 9.756 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 91 Resp: 66773

Ion	Ratio	Lower	Upper
91	100		
126	33.6	15.9	47.7






#83
tert-Butylbenzene
Concen: 9.553 ug/l
RT: 13.435 min Scan# 19
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:119 Resp: 59455


Ion	Ratio	Lower	Upper
119	100		
91	70.0	35.1	105.3
134	27.8	12.0	36.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#84
1,2,4-Trimethylbenzene
Concen: 9.754 ug/l
RT: 13.476 min Scan# 1959
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:105 Resp: 72046

Ion	Ratio	Lower	Upper
105	100		
120	43.8	21.6	65.0






#85
sec-Butylbenzene
Concen: 9.699 ug/l
RT: 13.612 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:105 Resp: 83662


Ion	Ratio	Lower	Upper
105	100		
134	18.9	9.7	28.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#86
p-Isopropyltoluene
Concen: 9.674 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:119 Resp: 67556

Ion	Ratio	Lower	Upper
119	100		
134	25.2	12.7	38.0
91	25.0	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 9.693 ug/l
RT: 13.729 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:146 Resp: 41843


Ion	Ratio	Lower	Upper
146	100		
111	43.1	21.4	64.3
148	63.0	32.3	96.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#88
1,4-Dichlorobenzene
Concen: 9.549 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:146 Resp: 42728

Ion	Ratio	Lower	Upper
146	100		
111	44.0	21.3	63.7
148	64.6	32.4	97.0






#89
n-Butylbenzene
Concen: 9.314 ug/l
RT: 14.053 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 91 Resp: 57632


Ion	Ratio	Lower	Upper
91	100		
92	49.2	25.8	77.3
134	22.8	11.7	35.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#90
Hexachloroethane
Concen: 9.294 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:117 Resp: 15261

Ion	Ratio	Lower	Upper
117	100		
201	65.0	33.7	101.0






#91
1,2-Dichlorobenzene
Concen: 9.458 ug/l
RT: 14.106 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:146 Resp: 40726


Ion	Ratio	Lower	Upper
146	100		
111	42.9	21.7	65.1
148	62.9	31.4	94.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 9.744 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion: 75 Resp: 5639

Ion	Ratio	Lower	Upper
75	100		
155	67.9	36.4	109.2
157	94.1	45.4	136.1






#93
1,2,4-Trichlorobenzene
Concen: 9.354 ug/l
RT: 15.388 min Scan# 22
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:180 Resp: 18720


Ion	Ratio	Lower	Upper
180	100		
182	92.4	46.8	140.3
145	31.7	16.0	48.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#94
Hexachlorobutadiene
Concen: 9.874 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:225 Resp: 10491

Ion	Ratio	Lower	Upper
225	100		
223	63.3	30.7	92.1
227	62.1	30.9	92.5





#95
Naphthalene
Concen: 8.717 ug/l
RT: 15.635 min Scan# 23
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07


Tgt Ion:128 Resp: 52054


Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.6	16.0
129	11.2	8.8	13.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025






#96
1,2,3-Trichlorobenzene
Concen: 9.605 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN085441.D
Acq: 14 Jan 2025 16:07

Tgt Ion:180 Resp: 19450

Ion	Ratio	Lower	Upper
180	100		
182	90.3	47.4	142.2
145	34.7	16.9	50.7



ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085442.D
Acq On : 14 Jan 2025 16:31
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED
Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:45:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) Pentafluorobenzene	8.218	168	186556	50.000	ug/l		0.00
34) 1,4-Difluorobenzene	9.100	114	324413	50.000	ug/l		0.00
63) Chlorobenzene-d5	11.865	117	282203	50.000	ug/l		0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126086	50.000	ug/l		0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4	8.577	65	17046	5.660	ug/l		0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 11.320%#				
35) Dibromofluoromethane	8.165	113	12102	5.377	ug/l		0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 10.760%#				
50) Toluene-d8	10.565	98	41328	5.168	ug/l		0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 10.340%#				
62) 4-Bromofluorobenzene	12.847	95	13520	4.943	ug/l		0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 9.880%#				
Target Compounds							Qvalue
2) Dichlorodifluoromethane	2.124	85	13199	5.225	ug/l		97
3) Chloromethane	2.359	50	14526	5.311	ug/l		96
4) Vinyl Chloride	2.512	62	14564	5.298	ug/l		98
5) Bromomethane	2.953	94	9788	5.895	ug/l		98
6) Chloroethane	3.124	64	9430	5.411	ug/l		99
7) Trichlorofluoromethane	3.501	101	20098	5.039	ug/l		92
8) Diethyl Ether	3.959	74	6652	4.828	ug/l		100
9) 1,1,2-Trichlorotrifluo...	4.365	101	11917	5.305	ug/l		97
10) Methyl Iodide	4.583	142	12331	4.793	ug/l		99
11) Tert butyl alcohol	5.518	59	9075	26.318	ug/l	#	94
12) 1,1-Dichloroethene	4.342	96	10429	5.209	ug/l		96
13) Acrolein	4.171	56	9430	20.033	ug/l		97
14) Allyl chloride	5.030	41	16741	5.153	ug/l		97
15) Acrylonitrile	5.718	53	27869	25.445	ug/l		97
16) Acetone	4.430	43	25122	25.819	ug/l		97
17) Carbon Disulfide	4.706	76	32068	5.202	ug/l		96
18) Methyl Acetate	5.024	43	15106	5.106	ug/l		97
19) Methyl tert-butyl Ether	5.800	73	31439	4.836	ug/l		94
20) Methylene Chloride	5.271	84	12981	5.389	ug/l	#	83
21) trans-1,2-Dichloroethene	5.783	96	10614	4.960	ug/l	#	78
22) Diisopropyl ether	6.665	45	35785	4.962	ug/l	#	96
23) Vinyl Acetate	6.600	43	118439m	23.389	ug/l		
24) 1,1-Dichloroethane	6.565	63	22865	5.200	ug/l		99
25) 2-Butanone	7.483	43	36127	25.231	ug/l		99
26) 2,2-Dichloropropane	7.488	77	18398	5.175	ug/l		100
27) cis-1,2-Dichloroethene	7.488	96	12477	4.951	ug/l		99
28) Bromochloromethane	7.806	49	11096	5.424	ug/l		99
29) Tetrahydrofuran	7.841	42	21983	24.215	ug/l		99
30) Chloroform	7.959	83	23384	5.146	ug/l		98
31) Cyclohexane	8.247	56	22343	5.846	ug/l	#	92
32) 1,1,1-Trichloroethane	8.159	97	21416	5.372	ug/l		91
36) 1,1-Dichloropropene	8.371	75	15886	5.029	ug/l		96
37) Ethyl Acetate	7.553	43	16879m	5.294	ug/l		
38) Carbon Tetrachloride	8.359	117	18317	5.065	ug/l		96
39) Methylcyclohexane	9.600	83	14175	4.776	ug/l		98
40) Benzene	8.600	78	47833	5.040	ug/l		99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085442.D
Acq On : 14 Jan 2025 16:31
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED
Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:45:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	7.777	41	7380	4.449	ug/l		96
42) 1,2-Dichloroethane	8.671	62	18628	5.211	ug/l		98
43) Isopropyl Acetate	8.688	43	24792	4.853	ug/l		98
44) Trichloroethene	9.347	130	11133	5.039	ug/l		86
45) 1,2-Dichloropropane	9.618	63	12591	5.194	ug/l		92
46) Dibromomethane	9.700	93	8230	4.705	ug/l		91
47) Bromodichloromethane	9.888	83	18456	5.179	ug/l	#	96
48) Methyl methacrylate	9.677	41	10574	4.599	ug/l		95
49) 1,4-Dioxane	9.694	88	3619	93.477	ug/l		97
51) 4-Methyl-2-Pentanone	10.441	43	71889	24.250	ug/l		98
52) Toluene	10.629	92	27098	4.928	ug/l		98
53) t-1,3-Dichloropropene	10.829	75	17099	5.077	ug/l		95
54) cis-1,3-Dichloropropene	10.306	75	17440	4.848	ug/l		97
55) 1,1,2-Trichloroethane	11.012	97	11311	5.197	ug/l		96
56) Ethyl methacrylate	10.871	69	14094	5.627	ug/l		94
57) 1,3-Dichloropropane	11.159	76	18945	5.006	ug/l		100
58) 2-Chloroethyl Vinyl ether	10.159	63	33210	24.047	ug/l		97
59) 2-Hexanone	11.194	43	49021	23.501	ug/l		94
60) Dibromochloromethane	11.353	129	13635	5.190	ug/l		96
61) 1,2-Dibromoethane	11.471	107	10419	4.810	ug/l		98
64) Tetrachloroethene	11.100	164	9756	5.071	ug/l		91
65) Chlorobenzene	11.888	112	31312	5.065	ug/l		98
66) 1,1,1,2-Tetrachloroethane	11.959	131	11620	5.121	ug/l		98
67) Ethyl Benzene	11.965	91	48216	4.789	ug/l		98
68) m/p-Xylenes	12.065	106	34771	9.344	ug/l		99
69) o-Xylene	12.394	106	16419	4.617	ug/l		98
70) Styrene	12.412	104	26218	4.454	ug/l		99
71) Bromoform	12.576	173	8017	4.938	ug/l	#	99
73) Isopropylbenzene	12.694	105	39811	4.678	ug/l		99
74) N-amyl acetate	12.488	43	17577	4.595	ug/l		95
75) 1,1,2,2-Tetrachloroethane	12.935	83	15483	5.151	ug/l		97
76) 1,2,3-Trichloropropane	12.994	75	12173m	4.758	ug/l		
77) Bromobenzene	12.982	156	11091	4.989	ug/l		99
78) n-propylbenzene	13.035	91	45450	4.512	ug/l		99
79) 2-Chlorotoluene	13.123	91	31932	4.898	ug/l		98
80) 1,3,5-Trimethylbenzene	13.170	105	31891	4.537	ug/l		100
81) trans-1,4-Dichloro-2-b...	12.729	75	4229	4.465	ug/l		90
82) 4-Chlorotoluene	13.218	91	31976	4.926	ug/l		100
83) tert-Butylbenzene	13.435	119	27597	4.675	ug/l		99
84) 1,2,4-Trimethylbenzene	13.482	105	32054	4.575	ug/l		98
85) sec-Butylbenzene	13.612	105	38148	4.663	ug/l		98
86) p-Isopropyltoluene	13.723	119	29856	4.584	ug/l		98
87) 1,3-Dichlorobenzene	13.729	146	20876	5.099	ug/l		98
88) 1,4-Dichlorobenzene	13.812	146	21982	5.179	ug/l		97
89) n-Butylbenzene	14.053	91	27437	4.675	ug/l		96
90) Hexachloroethane	14.329	117	7767	4.987	ug/l		100
91) 1,2-Dichlorobenzene	14.106	146	20177	4.940	ug/l		100
92) 1,2-Dibromo-3-Chloropr...	14.729	75	2872	5.232	ug/l		90
93) 1,2,4-Trichlorobenzene	15.388	180	9037	4.761	ug/l		94
94) Hexachlorobutadiene	15.500	225	5563	5.520	ug/l		98
95) Naphthalene	15.641	128	25047	4.422	ug/l		96
96) 1,2,3-Trichlorobenzene	15.841	180	8737	4.549	ug/l		99

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085442.D
Acq On : 14 Jan 2025 16:31
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:45:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085442.D
Acq On : 14 Jan 2025 16:31
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1


Quant Time: Jan 15 01:45:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:168 Resp: 186556


Ion	Ratio	Lower	Upper
168	100		
99	65.9	53.6	80.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#2
Dichlorodifluoromethane
Concen: 5.225 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 85 Resp: 13199

Ion	Ratio	Lower	Upper
85	100		
87	31.5	16.7	50.1






#3
Chloromethane
Concen: 5.311 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 50 Resp: 14526


Ion	Ratio	Lower	Upper
50	100		
52	31.6	27.0	40.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#4
Vinyl Chloride
Concen: 5.298 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 62 Resp: 14564

Ion	Ratio	Lower	Upper
62	100		
64	29.9	24.8	37.2






#5
Bromomethane
Concen: 5.895 ug/l
RT: 2.953 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 94 Resp: 9788


Ion	Ratio	Lower	Upper
94	100		
96	87.7	71.8	107.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#6
Chloroethane
Concen: 5.411 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 64 Resp: 9430

Ion	Ratio	Lower	Upper
64	100		
66	31.4	24.6	36.8






#7
Trichlorofluoromethane
Concen: 5.039 ug/l
RT: 3.501 min Scan# 26
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:101 Resp: 20098


Ion	Ratio	Lower	Upper
101	100		
103	70.8	51.4	77.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#8
Diethyl Ether
Concen: 4.828 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 74 Resp: 6652

Ion	Ratio	Lower	Upper
74	100		
45	99.5	49.7	149.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 5.305 ug/l
RT: 4.365 min Scan# 41
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31


Tgt Ion:101 Resp: 11917


Ion	Ratio	Lower	Upper
101	100		
85	46.5	37.8	56.8
151	76.7	58.8	88.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025







#10
Methyl Iodide
Concen: 4.793 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:142 Resp: 12331

Ion	Ratio	Lower	Upper
142	100		
127	50.9	40.3	60.5
141	15.0	12.8	19.2





#11
Tert butyl alcohol
Concen: 26.318 ug/l
RT: 5.518 min Scan# 60
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31


Tgt Ion: 59 Resp: 9075

Ion	Ratio	Lower	Upper
59	100		
57	8.5	8.6	13.0#

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#12
1,1-Dichloroethene
Concen: 5.209 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 96 Resp: 10429


Ion	Ratio	Lower	Upper
96	100		
61	172.7	141.0	211.4
98	69.0	49.0	73.6



#13
Acrolein
Concen: 20.033 ug/l
RT: 4.171 min Scan# 37
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 56 Resp: 9430


Ion	Ratio	Lower	Upper
56	100		
55	73.2	56.3	84.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#14
Allyl chloride
Concen: 5.153 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 41 Resp: 16741

Ion	Ratio	Lower	Upper
41	100		
39	77.4	64.4	96.6
76	36.6	30.5	45.7






#15
Acrylonitrile
Concen: 25.445 ug/l
RT: 5.718 min Scan# 64
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 53 Resp: 27869


Ion	Ratio	Lower	Upper
53	100		
52	85.5	65.5	98.3
51	38.5	29.8	44.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#16
Acetone
Concen: 25.819 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.012 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 25122

Ion	Ratio	Lower	Upper
43	100		
58	27.9	23.8	35.6






#17
Carbon Disulfide
Concen: 5.202 ug/l
RT: 4.706 min Scan# 46
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 76 Resp: 32068


Ion	Ratio	Lower	Upper
76	100		
78	9.9	6.9	10.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#18
Methyl Acetate
Concen: 5.106 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 15106

Ion	Ratio	Lower	Upper
43	100		
74	22.0	18.6	28.0






#19
Methyl tert-butyl Ether
Concen: 4.836 ug/l
RT: 5.800 min Scan# 65
Delta R.T. 0.012 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 73 Resp: 31439


Ion	Ratio	Lower	Upper
73	100		
57	25.2	17.8	26.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#20
Methylene Chloride
Concen: 5.389 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 84 Resp: 12981

Ion	Ratio	Lower	Upper
84	100		
49	114.5	104.1	156.1
51	32.1	31.0	46.4
86	42.9	51.9	77.9#






#21
trans-1,2-Dichloroethene
Concen: 4.960 ug/l
RT: 5.783 min Scan# 65
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 96 Resp: 10614


Ion	Ratio	Lower	Upper
96	100		
61	118.1	115.7	173.5
98	44.7	49.8	74.8#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#22
Diisopropyl ether
Concen: 4.962 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 45 Resp: 35785

Ion	Ratio	Lower	Upper
45	100		
43	58.3	45.0	67.6
87	24.2	21.8	32.8
59	9.7	10.2	15.2#






#23
Vinyl Acetate
Concen: 23.389 ug/l m
RT: 6.600 min Scan# 79
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 118439


Ion	Ratio	Lower	Upper
43	100		
86	9.0	8.4	12.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#24
1,1-Dichloroethane
Concen: 5.200 ug/l
RT: 6.565 min Scan# 784
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 63 Resp: 22865

Ion	Ratio	Lower	Upper
63	100		
98	6.8	3.3	9.8
100	4.0	2.0	6.0






#25
2-Butanone
Concen: 25.231 ug/l
RT: 7.483 min Scan# 94
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 36127


Ion	Ratio	Lower	Upper
43	100		
72	24.9	20.2	30.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#26
2,2-Dichloropropane
Concen: 5.175 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 77 Resp: 18398

Ion	Ratio	Lower	Upper
77	100		
97	21.2	10.7	32.1






#27
cis-1,2-Dichloroethene
Concen: 4.951 ug/l
RT: 7.488 min Scan# 94
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 96 Resp: 12477


Ion	Ratio	Lower	Upper
96	100		
61	154.5	0.0	311.8
98	62.8	0.0	126.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#28
Bromochloromethane
Concen: 5.424 ug/l
RT: 7.806 min Scan# 995
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 49 Resp: 11096

Ion	Ratio	Lower	Upper
49	100		
129	0.7	0.0	4.4
130	69.5	55.0	82.4






#29
Tetrahydrofuran
Concen: 24.215 ug/l
RT: 7.841 min Scan# 10
Delta R.T. 0.012 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 42 Resp: 21983


Ion	Ratio	Lower	Upper
42	100		
72	41.5	34.2	51.2
71	41.0	32.5	48.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#30
Chloroform
Concen: 5.146 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 83 Resp: 23384

Ion	Ratio	Lower	Upper
83	100		
85	63.4	51.8	77.6






#31
Cyclohexane
Concen: 5.846 ug/l
RT: 8.247 min Scan# 10
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 56 Resp: 22343


Ion	Ratio	Lower	Upper
56	100		
69	42.3	22.2	33.4#
84	84.3	66.4	99.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#32
1,1,1-Trichloroethane
Concen: 5.372 ug/l
RT: 8.159 min Scan# 1055
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 97 Resp: 21416

Ion	Ratio	Lower	Upper
97	100		
99	50.5	49.8	74.6
61	50.1	41.4	62.2






#33
1,2-Dichloroethane-d4
Concen: 5.660 ug/l
RT: 8.577 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 65 Resp: 17046


Ion	Ratio	Lower	Upper
65	100		
67	49.2	0.0	101.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:114 Resp: 324413

Ion	Ratio	Lower	Upper
114	100		
63	23.3	0.0	47.6
88	16.0	0.0	32.6






#35
Dibromofluoromethane
Concen: 5.377 ug/l
RT: 8.165 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:113 Resp: 12102


Ion	Ratio	Lower	Upper
113	100		
111	107.4	82.7	124.1
192	19.7	14.3	21.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#36
1,1-Dichloropropene
Concen: 5.029 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 75 Resp: 15886

Ion	Ratio	Lower	Upper
75	100		
110	30.3	16.5	49.5
77	29.2	24.4	36.6






#37
Ethyl Acetate
Concen: 5.294 ug/l m
RT: 7.553 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 16879


Ion	Ratio	Lower	Upper
43	100		
61	12.3	10.9	16.3
70	8.4	7.5	11.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#38
Carbon Tetrachloride
Concen: 5.065 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:117 Resp: 18317

Ion	Ratio	Lower	Upper
117	100		
119	91.1	75.4	113.2
121	27.3	24.6	37.0






#39
Methylcyclohexane
Concen: 4.776 ug/l
RT: 9.600 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 83 Resp: 14175


Ion	Ratio	Lower	Upper
83	100		
55	76.5	62.6	94.0
98	47.7	37.7	56.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#40
Benzene
Concen: 5.040 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 78 Resp: 47833

Ion	Ratio	Lower	Upper
78	100		
77	23.2	19.0	28.6






#41
Methacrylonitrile
Concen: 4.449 ug/l
RT: 7.777 min Scan# 99
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 41 Resp: 7380


Ion	Ratio	Lower	Upper
41	100		
39	60.5	46.0	69.0
67	74.6	57.4	86.2
52	35.3	25.5	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#42
1,2-Dichloroethane
Concen: 5.211 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 62 Resp: 18628

Ion	Ratio	Lower	Upper
62	100		
98	7.8	0.0	17.0






#43
Isopropyl Acetate
Concen: 4.853 ug/l
RT: 8.688 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 24792


Ion	Ratio	Lower	Upper
43	100		
61	27.0	20.7	31.1
87	12.3	9.8	14.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#44
Trichloroethene
Concen: 5.039 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:130 Resp: 11133

Ion	Ratio	Lower	Upper
130	100		
95	112.2	0.0	195.8






#45
1,2-Dichloropropane
Concen: 5.194 ug/l
RT: 9.618 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 63 Resp: 12591


Ion	Ratio	Lower	Upper
63	100		
65	27.8	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#46
Dibromomethane
Concen: 4.705 ug/l
RT: 9.700 min Scan# 1317
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 93 Resp: 8230

Ion	Ratio	Lower	Upper
93	100		
95	88.7	64.7	97.1
174	95.0	69.0	103.4






#47
Bromodichloromethane
Concen: 5.179 ug/l
RT: 9.888 min Scan# 13
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 83 Resp: 18456


Ion	Ratio	Lower	Upper
83	100		
85	67.0	51.2	76.8
127	6.1	6.5	9.7#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#48
Methyl methacrylate
Concen: 4.599 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 41 Resp: 10574

Ion	Ratio	Lower	Upper
41	100		
69	84.9	64.7	97.1
39	65.3	49.0	73.6






#49
1,4-Dioxane
Concen: 93.477 ug/l
RT: 9.694 min Scan# 13
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 88 Resp: 3619


Ion	Ratio	Lower	Upper
88	100		
43	32.3	26.6	39.8
58	77.9	59.5	89.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#50
Toluene-d8
Concen: 5.168 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 98 Resp: 41328

Ion	Ratio	Lower	Upper
98	100		
100	60.9	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 24.250 ug/l
RT: 10.441 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 71889


Ion	Ratio	Lower	Upper
43	100		
58	36.9	30.5	45.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#52
Toluene
Concen: 4.928 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 92 Resp: 27098

Ion	Ratio	Lower	Upper
92	100		
91	176.4	139.2	208.8






#53
t-1,3-Dichloropropene
Concen: 5.077 ug/l
RT: 10.829 min Scan# 15
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 75 Resp: 17099


Ion	Ratio	Lower	Upper
75	100		
77	32.2	23.5	35.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#54
cis-1,3-Dichloropropene
Concen: 4.848 ug/l
RT: 10.306 min Scan# 1420
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 75 Resp: 17440

Ion	Ratio	Lower	Upper
75	100		
77	27.5	25.0	37.4
39	53.2	43.1	64.7






#55
1,1,2-Trichloroethane
Concen: 5.197 ug/l
RT: 11.012 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 97 Resp: 11311


Ion	Ratio	Lower	Upper
97	100		
83	82.7	69.0	103.6
85	51.3	44.6	66.8
99	62.4	48.6	72.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#56
Ethyl methacrylate
Concen: 5.627 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 69 Resp: 14094

Ion	Ratio	Lower	Upper
69	100		
41	73.8	54.6	82.0
39	42.6	32.4	48.6






#57
1,3-Dichloropropane
Concen: 5.006 ug/l
RT: 11.159 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 76 Resp: 18945


Ion	Ratio	Lower	Upper
76	100		
78	32.2	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#58
2-Chloroethyl Vinyl ether
Concen: 24.047 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 63 Resp: 33210

Ion	Ratio	Lower	Upper
63	100		
106	25.2	21.6	32.4






#59
2-Hexanone
Concen: 23.501 ug/l
RT: 11.194 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 49021


Ion	Ratio	Lower	Upper
43	100		
58	48.3	26.2	78.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#60
Dibromochloromethane
Concen: 5.190 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:129 Resp: 13635

Ion	Ratio	Lower	Upper
129	100		
127	73.9	38.6	115.8






#61
1,2-Dibromoethane
Concen: 4.810 ug/l
RT: 11.471 min Scan# 16
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:107 Resp: 10419


Ion	Ratio	Lower	Upper
107	100		
109	96.8	75.9	113.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#62
4-Bromofluorobenzene
Concen: 4.943 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 95 Resp: 13520

Ion	Ratio	Lower	Upper
95	100		
174	73.2	0.0	145.0
176	68.0	0.0	142.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:117 Resp: 282203


Ion	Ratio	Lower	Upper
117	100		
82	58.3	48.6	72.8
119	32.1	26.6	39.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#64
Tetrachloroethene
Concen: 5.071 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:164 Resp: 9756

Ion	Ratio	Lower	Upper
164	100		
166	130.6	103.4	155.2
129	111.5	79.2	118.8
131	110.5	77.1	115.7






#65
Chlorobenzene
Concen: 5.065 ug/l
RT: 11.888 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:112 Resp: 31312


Ion	Ratio	Lower	Upper
112	100		
114	30.7	25.3	37.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 5.121 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:131 Resp: 11620

Ion	Ratio	Lower	Upper
131	100		
133	92.7	47.4	142.3
119	65.0	33.1	99.5






#67
Ethyl Benzene
Concen: 4.789 ug/l
RT: 11.965 min Scan# 17
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 91 Resp: 48216


Ion	Ratio	Lower	Upper
91	100		
106	28.9	23.8	35.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#68
m/p-Xylenes
Concen: 9.344 ug/l
RT: 12.065 min Scan# 1719
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:106 Resp: 34771

Ion	Ratio	Lower	Upper
106	100		
91	207.9	167.7	251.5






#69
o-Xylene
Concen: 4.617 ug/l
RT: 12.394 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:106 Resp: 16419


Ion	Ratio	Lower	Upper
106	100		
91	224.0	110.4	331.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#70
Styrene
Concen: 4.454 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:104 Resp: 26218

Ion	Ratio	Lower	Upper
104	100		
78	52.6	42.5	63.7
103	54.2	43.8	65.8






#71
Bromoform
Concen: 4.938 ug/l
RT: 12.576 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:173 Resp: 8017


Ion	Ratio	Lower	Upper
173	100		
175	48.3	24.4	73.2
254	0.0	0.0	0.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:152 Resp: 126086

Ion	Ratio	Lower	Upper
152	100		
115	62.0	31.1	93.3
150	160.7	0.0	343.6






#73
Isopropylbenzene
Concen: 4.678 ug/l
RT: 12.694 min Scan# 18
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:105 Resp: 39811


Ion	Ratio	Lower	Upper
105	100		
120	26.1	12.8	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#74
N-amyl acetate
Concen: 4.595 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 43 Resp: 17577

Ion	Ratio	Lower	Upper
43	100		
70	39.5	34.0	51.0
55	30.8	21.4	32.2
61	22.8	19.0	28.4






#75
1,1,2,2-Tetrachloroethane
Concen: 5.151 ug/l
RT: 12.935 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 83 Resp: 15483


Ion	Ratio	Lower	Upper
83	100		
131	9.5	4.8	14.4
85	67.5	32.2	96.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#76
1,2,3-Trichloropropane
Concen: 4.758 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 75 Resp: 12173

Ion	Ratio	Lower	Upper
75	100		
77	209.1	109.7	329.2






#77
Bromobenzene
Concen: 4.989 ug/l
RT: 12.982 min Scan# 18
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:156 Resp: 11091


Ion	Ratio	Lower	Upper
156	100		
77	229.5	114.1	342.4
158	100.8	48.9	146.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#78
n-propylbenzene
Concen: 4.512 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 91 Resp: 45450

Ion	Ratio	Lower	Upper
91	100		
120	21.1	10.9	32.6






#79
2-Chlorotoluene
Concen: 4.898 ug/l
RT: 13.123 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 91 Resp: 31932


Ion	Ratio	Lower	Upper
91	100		
126	32.6	15.7	47.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#80
1,3,5-Trimethylbenzene
Concen: 4.537 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:105 Resp: 31891

Ion	Ratio	Lower	Upper
105	100		
120	48.1	23.9	71.7






#81
trans-1,4-Dichloro-2-butene
Concen: 4.465 ug/l
RT: 12.729 min Scan# 18
Delta R.T. -0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 75 Resp: 4229


Ion	Ratio	Lower	Upper
75	100		
53	132.6	95.6	143.4
89	49.7	37.0	55.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#82
4-Chlorotoluene
Concen: 4.926 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 91 Resp: 31976

Ion	Ratio	Lower	Upper
91	100		
126	31.9	15.9	47.7






#83
tert-Butylbenzene
Concen: 4.675 ug/l
RT: 13.435 min Scan# 19
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:119 Resp: 27597


Ion	Ratio	Lower	Upper
119	100		
91	71.1	35.1	105.3
134	24.8	12.0	36.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#84
1,2,4-Trimethylbenzene
Concen: 4.575 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:105 Resp: 32054

Ion	Ratio	Lower	Upper
105	100		
120	42.0	21.6	65.0






#85
sec-Butylbenzene
Concen: 4.663 ug/l
RT: 13.612 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:105 Resp: 38148


Ion	Ratio	Lower	Upper
105	100		
134	20.1	9.7	28.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#86
p-Isopropyltoluene
Concen: 4.584 ug/l
RT: 13.723 min Scan# 2001
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:119 Resp: 29856

Ion	Ratio	Lower	Upper
119	100		
134	26.5	12.7	38.0
91	26.5	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 5.099 ug/l
RT: 13.729 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:146 Resp: 20876


Ion	Ratio	Lower	Upper
146	100		
111	42.3	21.4	64.3
148	62.7	32.3	96.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#88
1,4-Dichlorobenzene
Concen: 5.179 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:146 Resp: 21982

Ion	Ratio	Lower	Upper
146	100		
111	47.1	21.3	63.7
148	65.5	32.4	97.0






#89
n-Butylbenzene
Concen: 4.675 ug/l
RT: 14.053 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 91 Resp: 27437


Ion	Ratio	Lower	Upper
91	100		
92	48.7	25.8	77.3
134	21.4	11.7	35.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#90
Hexachloroethane
Concen: 4.987 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:117 Resp: 7767

Ion	Ratio	Lower	Upper
117	100		
201	66.9	33.7	101.0






#91
1,2-Dichlorobenzene
Concen: 4.940 ug/l
RT: 14.106 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:146 Resp: 20177


Ion	Ratio	Lower	Upper
146	100		
111	44.0	21.7	65.1
148	62.8	31.4	94.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 5.232 ug/l
RT: 14.729 min Scan# 2172
Delta R.T. 0.012 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion: 75 Resp: 2872

Ion	Ratio	Lower	Upper
75	100		
155	64.2	36.4	109.2
157	99.8	45.4	136.1






#93
1,2,4-Trichlorobenzene
Concen: 4.761 ug/l
RT: 15.388 min Scan# 22
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:180 Resp: 9037


Ion	Ratio	Lower	Upper
180	100		
182	100.5	46.8	140.3
145	33.0	16.0	48.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#94
Hexachlorobutadiene
Concen: 5.520 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:225 Resp: 5563

Ion	Ratio	Lower	Upper
225	100		
223	59.2	30.7	92.1
227	60.1	30.9	92.5






#95
Naphthalene
Concen: 4.422 ug/l
RT: 15.641 min Scan# 23
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:128 Resp: 25047


Ion	Ratio	Lower	Upper
128	100		
127	14.7	10.6	16.0
129	9.1	8.8	13.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96
1,2,3-Trichlorobenzene
Concen: 4.549 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN085442.D
Acq: 14 Jan 2025 16:31

Tgt Ion:180 Resp: 8737

Ion	Ratio	Lower	Upper
180	100		
182	95.6	47.4	142.2
145	35.5	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085443.D
Acq On : 14 Jan 2025 17:19
Operator : JC\MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED
Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 01:46:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) Pentafluorobenzene	8.224	168	166880	50.000	ug/l		0.00
34) 1,4-Difluorobenzene	9.100	114	302938	50.000	ug/l		0.00
63) Chlorobenzene-d5	11.865	117	256851	50.000	ug/l		0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	99539	50.000	ug/l		0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l		
Spiked Amount 50.000	Range 74 - 125		Recovery =	0.000%#			
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l		
Spiked Amount 50.000	Range 75 - 124		Recovery =	0.000%#			
50) Toluene-d8	0.000	98	0d	0.000	ug/l		
Spiked Amount 50.000	Range 86 - 113		Recovery =	0.000%#			
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l		
Spiked Amount 50.000	Range 77 - 121		Recovery =	0.000%#			
Target Compounds							Qvalue
2) Dichlorodifluoromethane	2.124	85	2382	1.054	ug/l		91
3) Chloromethane	2.359	50	2801	1.145	ug/l		93
4) Vinyl Chloride	2.512	62	2732	1.111	ug/l	#	81
6) Chloroethane	3.124	64	1809	1.160	ug/l		94
7) Trichlorofluoromethane	3.512	101	3860	1.082	ug/l		88
8) Diethyl Ether	3.953	74	1526	1.238	ug/l		81
9) 1,1,2-Trichlorotrifluo...	4.359	101	2156m	1.073	ug/l		
12) 1,1-Dichloroethene	4.330	96	1660	0.927	ug/l	#	67
14) Allyl chloride	5.024	41	3179	1.094	ug/l	#	85
15) Acrylonitrile	5.718	53	4967	5.070	ug/l	#	70
16) Acetone	4.436	43	5102	5.862	ug/l		92
17) Carbon Disulfide	4.706	76	6601	1.197	ug/l	#	87
18) Methyl Acetate	5.030	43	2908	1.099	ug/l	#	80
19) Methyl tert-butyl Ether	5.789	73	5156	0.887	ug/l	#	87
20) Methylene Chloride	5.265	84	2190	1.016	ug/l	#	87
21) trans-1,2-Dichloroethene	5.794	96	2110	1.102	ug/l	#	66
22) Diisopropyl ether	6.677	45	5691	0.882	ug/l	#	88
23) Vinyl Acetate	6.606	43	18526m	4.090	ug/l		
24) 1,1-Dichloroethane	6.583	63	4020m	1.022	ug/l		
25) 2-Butanone	7.483	43	6448m	5.034	ug/l		
26) 2,2-Dichloropropane	7.488	77	3104	0.976	ug/l		92
27) cis-1,2-Dichloroethene	7.477	96	2185	0.969	ug/l		92
28) Bromochloromethane	7.812	49	2082	1.138	ug/l	#	89
29) Tetrahydrofuran	7.835	42	3693	4.548	ug/l		98
30) Chloroform	7.965	83	4250	1.045	ug/l	#	80
32) 1,1,1-Trichloroethane	8.165	97	3678	1.031	ug/l	#	56
36) 1,1-Dichloropropene	8.371	75	3084	1.046	ug/l	#	84
37) Ethyl Acetate	7.577	43	3056m	1.026	ug/l		
38) Carbon Tetrachloride	8.359	117	3436	1.018	ug/l		90
39) Methylcyclohexane	9.600	83	2403	0.867	ug/l	#	87
40) Benzene	8.606	78	8480	0.957	ug/l		91
41) Methacrylonitrile	7.771	41	1458	0.941	ug/l	#	76
42) 1,2-Dichloroethane	8.665	62	3131	0.938	ug/l		91
43) Isopropyl Acetate	8.682	43	4803	1.007	ug/l	#	87
44) Trichloroethene	9.347	130	2133	1.034	ug/l		71
45) 1,2-Dichloropropane	9.618	63	2246	0.992	ug/l		99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085443.D
Acq On : 14 Jan 2025 17:19
Operator : JC\MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jan 15 01:46:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc Units		Dev(Min)
46) Dibromomethane	9.706	93	1769	1.083 ug/l		92
47) Bromodichloromethane	9.882	83	2935	0.882 ug/l		90
48) Methyl methacrylate	9.677	41	1886	0.878 ug/l		96
49) 1,4-Dioxane	9.688	88	673	18.616 ug/l	#	47
51) 4-Methyl-2-Pentanone	10.441	43	11523	4.163 ug/l		93
52) Toluene	10.618	92	4180	0.814 ug/l		93
53) t-1,3-Dichloropropene	10.835	75	2522	0.802 ug/l		92
54) cis-1,3-Dichloropropene	10.312	75	2725	0.811 ug/l	#	83
55) 1,1,2-Trichloroethane	11.018	97	1901	0.935 ug/l		91
56) Ethyl methacrylate	10.871	69	1923	2.471 ug/l	#	88
57) 1,3-Dichloropropane	11.165	76	3113	0.881 ug/l		95
58) 2-Chloroethyl Vinyl ether	10.153	63	4929	3.822 ug/l		91
59) 2-Hexanone	11.194	43	7905	4.058 ug/l		87
60) Dibromochloromethane	11.359	129	2339	0.953 ug/l		98
61) 1,2-Dibromoethane	11.465	107	2022	1.000 ug/l		90
64) Tetrachloroethene	11.106	164	1658	0.947 ug/l	#	83
65) Chlorobenzene	11.888	112	5401	0.960 ug/l		93
66) 1,1,1,2-Tetrachloroethane	11.959	131	2185	1.058 ug/l	#	64
67) Ethyl Benzene	11.959	91	7347	0.802 ug/l		94
68) m/p-Xylenes	12.070	106	5059	1.494 ug/l		100
69) o-Xylene	12.394	106	2478	0.766 ug/l		97
70) Styrene	12.406	104	3812	0.712 ug/l		99
71) Bromoform	12.582	173	1208	0.818 ug/l	#	88
73) Isopropylbenzene	12.688	105	5506	0.820 ug/l		98
74) N-amyl acetate	12.494	43	2716	0.899 ug/l	#	87
75) 1,1,2,2-Tetrachloroethane	12.935	83	2616	1.102 ug/l	#	96
76) 1,2,3-Trichloropropane	12.982	75	2198m	1.088 ug/l		
77) Bromobenzene	12.976	156	1733	0.987 ug/l		97
78) n-propylbenzene	13.029	91	6424	0.808 ug/l		93
79) 2-Chlorotoluene	13.117	91	4483	0.871 ug/l		97
80) 1,3,5-Trimethylbenzene	13.165	105	4301	0.775 ug/l		90
82) 4-Chlorotoluene	13.223	91	4268	0.833 ug/l		99
83) tert-Butylbenzene	13.435	119	3753	0.805 ug/l		96
84) 1,2,4-Trimethylbenzene	13.482	105	3758	0.679 ug/l		93
85) sec-Butylbenzene	13.612	105	4614	0.714 ug/l		99
86) p-Isopropyltoluene	13.723	119	3538	0.716 ug/l		98
87) 1,3-Dichlorobenzene	13.723	146	3038	0.940 ug/l		95
88) 1,4-Dichlorobenzene	13.806	146	3517m	1.050 ug/l		
89) n-Butylbenzene	14.053	91	3527	0.761 ug/l		93
90) Hexachloroethane	14.341	117	1355	1.102 ug/l		99
91) 1,2-Dichlorobenzene	14.106	146	3515	1.090 ug/l		96
92) 1,2-Dibromo-3-Chloropr...	14.711	75	402	0.928 ug/l		74
93) 1,2,4-Trichlorobenzene	15.394	180	1310	0.874 ug/l	#	80
94) Hexachlorobutadiene	15.500	225	822	1.033 ug/l		89
95) Naphthalene	15.641	128	4210	0.941 ug/l	#	90
96) 1,2,3-Trichlorobenzene	15.835	180	1494	0.985 ug/l	#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085443.D
Acq On : 14 Jan 2025 17:19
Operator : JC\MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1


Quant Time: Jan 15 01:46:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 01:28:51 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:168 Resp: 166880


Ion	Ratio	Lower	Upper
168	100		
99	68.6	53.6	80.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#2
Dichlorodifluoromethane
Concen: 1.054 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 85 Resp: 2382

Ion	Ratio	Lower	Upper
85	100		
87	28.5	16.7	50.1






#3
Chloromethane
Concen: 1.145 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 50 Resp: 2801


Ion	Ratio	Lower	Upper
50	100		
52	29.7	27.0	40.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#4
Vinyl Chloride
Concen: 1.111 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 62 Resp: 2732

Ion	Ratio	Lower	Upper
62	100		
64	41.4	24.8	37.2#






#6
Chloroethane
Concen: 1.160 ug/l
RT: 3.124 min Scan# 19
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 64 Resp: 1809


Ion	Ratio	Lower	Upper
64	100		
66	34.1	24.6	36.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#7
Trichlorofluoromethane
Concen: 1.082 ug/l
RT: 3.512 min Scan# 265
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:101 Resp: 3860

Ion	Ratio	Lower	Upper
101	100		
103	55.0	51.4	77.2






#8
Diethyl Ether
Concen: 1.238 ug/l
RT: 3.953 min Scan# 34
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 74 Resp: 1526


Ion	Ratio	Lower	Upper
74	100		
45	80.9	49.7	149.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.073 ug/l m
RT: 4.359 min Scan# 409
Delta R.T. -0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:101 Resp: 2156

Ion	Ratio	Lower	Upper
101	100		
85	35.3	37.8	56.8#
151	66.4	58.8	88.2






#12
1,1-Dichloroethene
Concen: 0.927 ug/l
RT: 4.330 min Scan# 40
Delta R.T. -0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 96 Resp: 1660


Ion	Ratio	Lower	Upper
96	100		
61	122.1	141.0	211.4#
98	48.1	49.0	73.6#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#14
Allyl chloride
Concen: 1.094 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 41 Resp: 3179

Ion	Ratio	Lower	Upper
41	100		
39	73.1	64.4	96.6
76	21.4	30.5	45.7#






#15
Acrylonitrile
Concen: 5.070 ug/l
RT: 5.718 min Scan# 64
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 53 Resp: 4967


Ion	Ratio	Lower	Upper
53	100		
52	45.1	65.5	98.3#
51	33.3	29.8	44.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#16
Acetone
Concen: 5.862 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.018 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 5102

Ion	Ratio	Lower	Upper
43	100		
58	33.8	23.8	35.6






#17
Carbon Disulfide
Concen: 1.197 ug/l
RT: 4.706 min Scan# 46
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 76 Resp: 6601


Ion	Ratio	Lower	Upper
76	100		
78	13.2	6.9	10.3#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#18
Methyl Acetate
Concen: 1.099 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 2908

Ion	Ratio	Lower	Upper
43	100		
74	13.8	18.6	28.0#






#19
Methyl tert-butyl Ether
Concen: 0.887 ug/l
RT: 5.789 min Scan# 65
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 73 Resp: 5156


Ion	Ratio	Lower	Upper
73	100		
57	28.4	17.8	26.8#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#20
Methylene Chloride
Concen: 1.016 ug/l
RT: 5.265 min Scan# 563
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 84 Resp: 2190

Ion	Ratio	Lower	Upper
84	100		
49	115.6	104.1	156.1
51	36.9	31.0	46.4
86	48.9	51.9	77.9#






#21
trans-1,2-Dichloroethene
Concen: 1.102 ug/l
RT: 5.794 min Scan# 65
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 96 Resp: 2110


Ion	Ratio	Lower	Upper
96	100		
61	98.7	115.7	173.5#
98	41.8	49.8	74.8#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#22
Diisopropyl ether
Concen: 0.882 ug/l
RT: 6.677 min Scan# 803
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 45 Resp: 5691

Ion	Ratio	Lower	Upper
45	100		
43	65.9	45.0	67.6
87	35.0	21.8	32.8#
59	11.3	10.2	15.2






#23
Vinyl Acetate
Concen: 4.090 ug/l m
RT: 6.606 min Scan# 79
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 18526


Ion	Ratio	Lower	Upper
43	100		
86	9.8	8.4	12.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#24
1,1-Dichloroethane
Concen: 1.022 ug/l m
RT: 6.583 min Scan# 787
Delta R.T. 0.018 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 63 Resp: 4020

Ion	Ratio	Lower	Upper
63	100		
98	0.0	3.3	9.8#
100	0.0	2.0	6.0#






#25
2-Butanone
Concen: 5.034 ug/l m
RT: 7.483 min Scan# 94
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 6448


Ion	Ratio	Lower	Upper
43	100		
72	18.0	20.2	30.4#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#26
2,2-Dichloropropane
Concen: 0.976 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 77 Resp: 3104

Ion	Ratio	Lower	Upper
77	100		
97	17.7	10.7	32.1






#27
cis-1,2-Dichloroethene
Concen: 0.969 ug/l
RT: 7.477 min Scan# 93
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 96 Resp: 2185


Ion	Ratio	Lower	Upper
96	100		
61	166.1	0.0	311.8
98	68.7	0.0	126.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#28
Bromochloromethane
Concen: 1.138 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 49 Resp: 2082

Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	4.4
130	78.0	55.0	82.4






#29
Tetrahydrofuran
Concen: 4.548 ug/l
RT: 7.835 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 42 Resp: 3693


Ion	Ratio	Lower	Upper
42	100		
72	43.1	34.2	51.2
71	37.9	32.5	48.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#30
Chloroform
Concen: 1.045 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 83 Resp: 4250

Ion	Ratio	Lower	Upper
83	100		
85	80.2	51.8	77.6#






#32
1,1,1-Trichloroethane
Concen: 1.031 ug/l
RT: 8.165 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 97 Resp: 3678


Ion	Ratio	Lower	Upper
97	100		
99	0.0	49.8	74.6#
61	52.2	41.4	62.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#33
1,2-Dichloroethane-d4
Concen: 0.000 ug/l
Expected RT: 8.57 min

Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 65


Sig	Exp Ratio
65	100
67	50.8



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 12
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:114 Resp: 302938


Ion	Ratio	Lower	Upper
114	100		
63	22.3	0.0	47.6
88	16.8	0.0	32.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#35
Dibromofluoromethane
Concen: 0.000 ug/l
Expected RT: 8.16 min

Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 113


Sig	Exp Ratio
113	100
111	103.4
192	17.9



#36
1,1-Dichloropropene
Concen: 1.046 ug/l
RT: 8.371 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 75 Resp: 3084


Ion	Ratio	Lower	Upper
75	100		
110	30.1	16.5	49.5
77	15.9	24.4	36.6#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#37
Ethyl Acetate
Concen: 1.026 ug/l m
RT: 7.577 min Scan# 956
Delta R.T. 0.023 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 3056

Ion	Ratio	Lower	Upper
43	100		
61	4.3	10.9	16.3#
70	0.0	7.5	11.3#






#38
Carbon Tetrachloride
Concen: 1.018 ug/l
RT: 8.359 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:117 Resp: 3436


Ion	Ratio	Lower	Upper
117	100		
119	106.6	75.4	113.2
121	31.0	24.6	37.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#39
Methylcyclohexane
Concen: 0.867 ug/l
RT: 9.600 min Scan# 1300
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 83 Resp: 2403

Ion	Ratio	Lower	Upper
83	100		
55	72.7	62.6	94.0
98	31.8	37.7	56.5#






#40
Benzene
Concen: 0.957 ug/l
RT: 8.606 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 78 Resp: 8480


Ion	Ratio	Lower	Upper
78	100		
77	28.3	19.0	28.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#41
Methacrylonitrile
Concen: 0.941 ug/l
RT: 7.771 min Scan# 989
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 41 Resp: 1458

Ion	Ratio	Lower	Upper
41	100		
39	37.6	46.0	69.0#
67	53.8	57.4	86.2#
52	18.5	25.5	38.3#






#42
1,2-Dichloroethane
Concen: 0.938 ug/l
RT: 8.665 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 62 Resp: 3131


Ion	Ratio	Lower	Upper
62	100		
98	5.4	0.0	17.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#43
Isopropyl Acetate
Concen: 1.007 ug/l
RT: 8.682 min Scan# 1144
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 4803

Ion	Ratio	Lower	Upper
43	100		
61	18.9	20.7	31.1#
87	7.3	9.8	14.8#






#44
Trichloroethene
Concen: 1.034 ug/l
RT: 9.347 min Scan# 12
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:130 Resp: 2133


Ion	Ratio	Lower	Upper
130	100		
95	126.1	0.0	195.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#45
1,2-Dichloropropane
Concen: 0.992 ug/l
RT: 9.618 min Scan# 1303
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 63 Resp: 2246

Ion	Ratio	Lower	Upper
63	100		
65	32.6	25.6	38.4






#46
Dibromomethane
Concen: 1.083 ug/l
RT: 9.706 min Scan# 13
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 93 Resp: 1769


Ion	Ratio	Lower	Upper
93	100		
95	86.5	64.7	97.1
174	77.4	69.0	103.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#47
Bromodichloromethane
Concen: 0.882 ug/l
RT: 9.882 min Scan# 1348
Delta R.T. 0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 83 Resp: 2935

Ion	Ratio	Lower	Upper
83	100		
85	72.9	51.2	76.8
127	9.0	6.5	9.7






#48
Methyl methacrylate
Concen: 0.878 ug/l
RT: 9.677 min Scan# 13
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 41 Resp: 1886


Ion	Ratio	Lower	Upper
41	100		
69	76.2	64.7	97.1
39	62.8	49.0	73.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#49
1,4-Dioxane
Concen: 18.616 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 88 Resp: 673

Ion	Ratio	Lower	Upper
88	100		
43	0.0	26.6	39.8#
58	31.1	59.5	89.3#





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 10.56 min

Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion: 98


Sig	Exp Ratio
98	100
100	65.3

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025








#51
4-Methyl-2-Pentanone
Concen: 4.163 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion: 43 Resp: 11523

Ion	Ratio	Lower	Upper
43	100		
58	33.6	30.5	45.7










#52
Toluene
Concen: 0.814 ug/l
RT: 10.618 min Scan# 14
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 92 Resp: 4180


Ion	Ratio	Lower	Upper
92	100		
91	183.9	139.2	208.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#53
t-1,3-Dichloropropene
Concen: 0.802 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 75 Resp: 2522

Ion	Ratio	Lower	Upper
75	100		
77	33.5	23.5	35.3






#54
cis-1,3-Dichloropropene
Concen: 0.811 ug/l
RT: 10.312 min Scan# 14
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 75 Resp: 2725


Ion	Ratio	Lower	Upper
75	100		
77	39.0	25.0	37.4#
39	67.6	43.1	64.7#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#55
1,1,2-Trichloroethane
Concen: 0.935 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 97 Resp: 1901

Ion	Ratio	Lower	Upper
97	100		
83	101.5	69.0	103.6
85	60.3	44.6	66.8
99	59.9	48.6	72.8






#56
Ethyl methacrylate
Concen: 2.471 ug/l
RT: 10.871 min Scan# 15
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 69 Resp: 1923


Ion	Ratio	Lower	Upper
69	100		
41	74.4	54.6	82.0
39	52.4	32.4	48.6#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#57
1,3-Dichloropropane
Concen: 0.881 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 76 Resp: 3113

Ion	Ratio	Lower	Upper
76	100		
78	34.9	25.6	38.4






#58
2-Chloroethyl Vinyl ether
Concen: 3.822 ug/l
RT: 10.153 min Scan# 13
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 63 Resp: 4929


Ion	Ratio	Lower	Upper
63	100		
106	31.7	21.6	32.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#59
2-Hexanone
Concen: 4.058 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 7905

Ion	Ratio	Lower	Upper
43	100		
58	42.9	26.2	78.6






#60
Dibromochloromethane
Concen: 0.953 ug/l
RT: 11.359 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:129 Resp: 2339


Ion	Ratio	Lower	Upper
129	100		
127	78.9	38.6	115.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#61
1,2-Dibromoethane
Concen: 1.000 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:107 Resp: 2022

Ion	Ratio	Lower	Upper
107	100		
109	84.8	75.9	113.9





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 12.85 min

Lab File: VN085443.D
Acq: 14 Jan 2025 17:19


Tgt Ion: 95

Sig	Exp Ratio
95	100
174	72.5
176	71.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:117 Resp: 256851

Ion	Ratio	Lower	Upper
117	100		
82	59.0	48.6	72.8
119	32.2	26.6	39.8






#64
Tetrachloroethene
Concen: 0.947 ug/l
RT: 11.106 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:164 Resp: 1658


Ion	Ratio	Lower	Upper
164	100		
166	152.1	103.4	155.2
129	128.3	79.2	118.8#
131	96.9	77.1	115.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#65
Chlorobenzene
Concen: 0.960 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:112 Resp: 5401

Ion	Ratio	Lower	Upper
112	100		
114	35.3	25.3	37.9






#66
1,1,1,2-Tetrachloroethane
Concen: 1.058 ug/l
RT: 11.959 min Scan# 17
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:131 Resp: 2185


Ion	Ratio	Lower	Upper
131	100		
133	97.7	47.4	142.3
119	0.0	33.1	99.5#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#67
Ethyl Benzene
Concen: 0.802 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 91 Resp: 7347

Ion	Ratio	Lower	Upper
91	100		
106	32.8	23.8	35.8






#68
m/p-Xylenes
Concen: 1.494 ug/l
RT: 12.070 min Scan# 17
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:106 Resp: 5059


Ion	Ratio	Lower	Upper
106	100		
91	209.3	167.7	251.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#69
o-Xylene
Concen: 0.766 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:106 Resp: 2478

Ion	Ratio	Lower	Upper
106	100		
91	225.4	110.4	331.2






#70
Styrene
Concen: 0.712 ug/l
RT: 12.406 min Scan# 17
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:104 Resp: 3812


Ion	Ratio	Lower	Upper
104	100		
78	52.9	42.5	63.7
103	53.7	43.8	65.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#71
Bromoform
Concen: 0.818 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:173 Resp: 1208

Ion	Ratio	Lower	Upper
173	100		
175	40.7	24.4	73.2
254	0.0	0.0	0.0






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 20
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:152 Resp: 99539


Ion	Ratio	Lower	Upper
152	100		
115	63.8	31.1	93.3
150	158.5	0.0	343.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#73
Isopropylbenzene
Concen: 0.820 ug/l
RT: 12.688 min Scan# 1825
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:105 Resp: 5506

Ion	Ratio	Lower	Upper
105	100		
120	26.7	12.8	38.3






#74
N-amyl acetate
Concen: 0.899 ug/l
RT: 12.494 min Scan# 17
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 43 Resp: 2716


Ion	Ratio	Lower	Upper
43	100		
70	30.4	34.0	51.0#
55	24.2	21.4	32.2
61	17.6	19.0	28.4#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#75
1,1,2,2-Tetrachloroethane
Concen: 1.102 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 83 Resp: 2616

Ion	Ratio	Lower	Upper
83	100		
131	2.4	4.8	14.4#
85	65.7	32.2	96.6






#76
1,2,3-Trichloropropane
Concen: 1.088 ug/l m
RT: 12.982 min Scan# 18
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 75 Resp: 2198


Ion	Ratio	Lower	Upper
75	100		
77	184.2	109.7	329.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#77
Bromobenzene
Concen: 0.987 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:156 Resp: 1733

Ion	Ratio	Lower	Upper
156	100		
77	233.6	114.1	342.4
158	94.9	48.9	146.8






#78
n-propylbenzene
Concen: 0.808 ug/l
RT: 13.029 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 91 Resp: 6424


Ion	Ratio	Lower	Upper
91	100		
120	18.2	10.9	32.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#79
2-Chlorotoluene
Concen: 0.871 ug/l
RT: 13.117 min Scan# 1898
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 91 Resp: 4483

Ion	Ratio	Lower	Upper
91	100		
126	32.9	15.7	47.1






#80
1,3,5-Trimethylbenzene
Concen: 0.775 ug/l
RT: 13.165 min Scan# 19
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:105 Resp: 4301


Ion	Ratio	Lower	Upper
105	100		
120	41.2	23.9	71.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#82
4-Chlorotoluene
Concen: 0.833 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 91 Resp: 4268

Ion	Ratio	Lower	Upper
91	100		
126	31.0	15.9	47.7






#83
tert-Butylbenzene
Concen: 0.805 ug/l
RT: 13.435 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:119 Resp: 3753


Ion	Ratio	Lower	Upper
119	100		
91	65.9	35.1	105.3
134	24.9	12.0	36.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#84
1,2,4-Trimethylbenzene
Concen: 0.679 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:105 Resp: 3758

Ion	Ratio	Lower	Upper
105	100		
120	47.7	21.6	65.0






#85
sec-Butylbenzene
Concen: 0.714 ug/l
RT: 13.612 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:105 Resp: 4614


Ion	Ratio	Lower	Upper
105	100		
134	19.9	9.7	28.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#86
p-Isopropyltoluene
Concen: 0.716 ug/l
RT: 13.723 min Scan# 2001
Delta R.T. 0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:119 Resp: 3538

Ion	Ratio	Lower	Upper
119	100		
134	25.6	12.7	38.0
91	23.2	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 0.940 ug/l
RT: 13.723 min Scan# 20
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:146 Resp: 3038


Ion	Ratio	Lower	Upper
146	100		
111	43.0	21.4	64.3
148	70.9	32.3	96.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#88
1,4-Dichlorobenzene
Concen: 1.050 ug/l m
RT: 13.806 min Scan# 2015
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:146 Resp: 3517

Ion	Ratio	Lower	Upper
146	100		
111	37.2	21.3	63.7
148	61.2	32.4	97.0






#89
n-Butylbenzene
Concen: 0.761 ug/l
RT: 14.053 min Scan# 20
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 91 Resp: 3527


Ion	Ratio	Lower	Upper
91	100		
92	57.2	25.8	77.3
134	25.5	11.7	35.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#90
Hexachloroethane
Concen: 1.102 ug/l
RT: 14.341 min Scan# 2106
Delta R.T. 0.012 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:117 Resp: 1355

Ion	Ratio	Lower	Upper
117	100		
201	66.5	33.7	101.0






#91
1,2-Dichlorobenzene
Concen: 1.090 ug/l
RT: 14.106 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:146 Resp: 3515


Ion	Ratio	Lower	Upper
146	100		
111	43.4	21.7	65.1
148	57.8	31.4	94.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 0.928 ug/l
RT: 14.711 min Scan# 2169
Delta R.T. -0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion: 75 Resp: 402

Ion	Ratio	Lower	Upper
75	100		
155	48.0	36.4	109.2
157	112.7	45.4	136.1






#93
1,2,4-Trichlorobenzene
Concen: 0.874 ug/l
RT: 15.394 min Scan# 22
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:180 Resp: 1310


Ion	Ratio	Lower	Upper
180	100		
182	106.3	46.8	140.3
145	9.7	16.0	48.0#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#94
Hexachlorobutadiene
Concen: 1.033 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:225 Resp: 822

Ion	Ratio	Lower	Upper
225	100		
223	50.7	30.7	92.1
227	55.7	30.9	92.5






#95
Naphthalene
Concen: 0.941 ug/l
RT: 15.641 min Scan# 23
Delta R.T. 0.006 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:128 Resp: 4210


Ion	Ratio	Lower	Upper
128	100		
127	11.1	10.6	16.0
129	5.3	8.8	13.2#



Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96
1,2,3-Trichlorobenzene
Concen: 0.985 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN085443.D
Acq: 14 Jan 2025 17:19

Tgt Ion:180 Resp: 1494

Ion	Ratio	Lower	Upper
180	100		
182	90.2	47.4	142.2
145	11.8	16.9	50.7#



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED
Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.218	168	219757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	370487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	320948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	158525	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	176594	49.782	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 99.560%			
35) Dibromofluoromethane	8.165	113	130487	50.768	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.540%			
50) Toluene-d8	10.565	98	477630	52.302	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.600%			
62) 4-Bromofluorobenzene	12.847	95	165249	52.899	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.800%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.124	85	130242	43.772	ug/l	100
3) Chloromethane	2.360	50	138152	42.883	ug/l	99
4) Vinyl Chloride	2.513	62	138316	42.715	ug/l	99
5) Bromomethane	2.954	94	80019	40.910	ug/l	91
6) Chloroethane	3.124	64	86837	42.299	ug/l	99
7) Trichlorofluoromethane	3.501	101	206526	43.956	ug/l	99
8) Diethyl Ether	3.959	74	69545	42.846	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	118310	44.706	ug/l	98
10) Methyl Iodide	4.595	142	140190	46.259	ug/l	98
11) Tert butyl alcohol	5.518	59	83933	206.633	ug/l	99
12) 1,1-Dichloroethene	4.342	96	106880	45.317	ug/l	98
13) Acrolein	4.177	56	135720	244.765	ug/l	99
14) Allyl chloride	5.024	41	170308	44.501	ug/l	99
15) Acrylonitrile	5.712	53	294062	227.920	ug/l	99
16) Acetone	4.418	43	245248	213.971	ug/l	98
17) Carbon Disulfide	4.712	76	300215	41.342	ug/l	100
18) Methyl Acetate	5.018	43	156938	45.028	ug/l	100
19) Methyl tert-butyl Ether	5.795	73	365819	47.768	ug/l	100
20) Methylene Chloride	5.271	84	122059	43.017	ug/l	98
21) trans-1,2-Dichloroethene	5.783	96	112597	44.672	ug/l	93
22) Diisopropyl ether	6.665	45	399206	46.996	ug/l	99
23) Vinyl Acetate	6.601	43	1471068	247.459	ug/l	100
24) 1,1-Dichloroethane	6.571	63	228489	44.114	ug/l	98
25) 2-Butanone	7.477	43	386125	228.924	ug/l	100
26) 2,2-Dichloropropane	7.489	77	199303	47.593	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	134920	45.446	ug/l	98
28) Bromochloromethane	7.812	49	106323	44.122	ug/l	100
29) Tetrahydrofuran	7.836	42	255650	239.065	ug/l	99
30) Chloroform	7.965	83	235923	44.071	ug/l	98
31) Cyclohexane	8.253	56	193034	42.878	ug/l	93
32) 1,1,1-Trichloroethane	8.165	97	208172	44.332	ug/l	99
36) 1,1-Dichloropropene	8.365	75	161834	44.863	ug/l	98
37) Ethyl Acetate	7.559	43	160656	44.123	ug/l	97
38) Carbon Tetrachloride	8.359	117	183592	44.456	ug/l	99
39) Methylcyclohexane	9.600	83	168431	49.687	ug/l	98
40) Benzene	8.600	78	494871	45.657	ug/l	99

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	7.777	41	89326	47.155	ug/l		99
42) 1,2-Dichloroethane	8.665	62	185663	45.481	ug/l		100
43) Isopropyl Acetate	8.683	43	265107	45.440	ug/l		99
44) Trichloroethene	9.347	130	110058	43.622	ug/l		96
45) 1,2-Dichloropropane	9.618	63	123876	44.744	ug/l		97
46) Dibromomethane	9.700	93	88325	44.214	ug/l		98
47) Bromodichloromethane	9.889	83	185808	45.655	ug/l		99
48) Methyl methacrylate	9.677	41	131142	49.948	ug/l		97
49) 1,4-Dioxane	9.689	88	35051	792.760	ug/l		98
51) 4-Methyl-2-Pentanone	10.441	43	832426	245.877	ug/l		99
52) Toluene	10.630	92	301911	48.073	ug/l		99
53) t-1,3-Dichloropropene	10.830	75	183680	47.757	ug/l		94
54) cis-1,3-Dichloropropene	10.312	75	196307	47.783	ug/l		98
55) 1,1,2-Trichloroethane	11.012	97	114757	46.173	ug/l		97
56) Ethyl methacrylate	10.871	69	182170	43.762	ug/l		98
57) 1,3-Dichloropropane	11.159	76	202583	46.877	ug/l		100
58) 2-Chloroethyl Vinyl ether	10.159	63	403949	256.116	ug/l		99
59) 2-Hexanone	11.194	43	598433	251.213	ug/l		100
60) Dibromochloromethane	11.353	129	137831	45.935	ug/l		99
61) 1,2-Dibromoethane	11.465	107	110926	44.837	ug/l		100
64) Tetrachloroethene	11.100	164	98211	44.886	ug/l		96
65) Chlorobenzene	11.888	112	322723	45.901	ug/l		98
66) 1,1,1,2-Tetrachloroethane	11.959	131	115379	44.712	ug/l		99
67) Ethyl Benzene	11.959	91	577059	50.391	ug/l		97
68) m/p-Xylenes	12.071	106	437367	103.341	ug/l		100
69) o-Xylene	12.394	106	210698	52.093	ug/l		99
70) Styrene	12.406	104	355459	53.103	ug/l		100
71) Bromoform	12.577	173	90693	49.119	ug/l	#	100
73) Isopropylbenzene	12.694	105	534428	49.951	ug/l		99
74) N-amyl acetate	12.488	43	234342	48.730	ug/l		100
75) 1,1,2,2-Tetrachloroethane	12.935	83	164882	43.627	ug/l		100
76) 1,2,3-Trichloropropane	12.988	75	134976m	41.891	ug/l		
77) Bromobenzene	12.977	156	128968	46.138	ug/l		97
78) n-propylbenzene	13.029	91	645977	51.001	ug/l		100
79) 2-Chlorotoluene	13.124	91	392404	47.870	ug/l		99
80) 1,3,5-Trimethylbenzene	13.171	105	456048	51.604	ug/l		100
81) trans-1,4-Dichloro-2-b...	12.735	75	58623m	49.226	ug/l		
82) 4-Chlorotoluene	13.218	91	398110	48.776	ug/l		100
83) tert-Butylbenzene	13.435	119	378478	50.997	ug/l		98
84) 1,2,4-Trimethylbenzene	13.477	105	457340	51.922	ug/l		100
85) sec-Butylbenzene	13.612	105	532578	51.777	ug/l		100
86) p-Isopropyltoluene	13.729	119	440284	47.795	ug/l		100
87) 1,3-Dichlorobenzene	13.729	146	236387	45.920	ug/l		99
88) 1,4-Dichlorobenzene	13.806	146	235319	44.102	ug/l		99
89) n-Butylbenzene	14.053	91	373227	50.578	ug/l		99
90) Hexachloroethane	14.329	117	88538	45.215	ug/l		100
91) 1,2-Dichlorobenzene	14.100	146	231646	45.113	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	14.718	75	31899	46.221	ug/l		97
93) 1,2,4-Trichlorobenzene	15.388	180	109157	45.739	ug/l		98
94) Hexachlorobutadiene	15.500	225	53358	42.112	ug/l		99
95) Naphthalene	15.641	128	335535	47.116	ug/l		100
96) 1,2,3-Trichlorobenzene	15.835	180	107583	44.551	ug/l		99

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1


Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 10
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:168 Resp: 219757


Ion	Ratio	Lower	Upper
168	100		
99	66.9	53.6	80.4



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#2
Dichlorodifluoromethane
Concen: 43.772 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 85 Resp: 130242

Ion	Ratio	Lower	Upper
85	100		
87	33.2	16.7	50.1






#3
Chloromethane
Concen: 42.883 ug/l
RT: 2.360 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 50 Resp: 138152


Ion	Ratio	Lower	Upper
50	100		
52	34.4	27.0	40.6



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#4
Vinyl Chloride
Concen: 42.715 ug/l
RT: 2.513 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 62 Resp: 138316

Ion	Ratio	Lower	Upper
62	100		
64	31.4	24.8	37.2






#5
Bromomethane
Concen: 40.910 ug/l
RT: 2.954 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 94 Resp: 80019


Ion	Ratio	Lower	Upper
94	100		
96	98.2	71.8	107.6



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#6
Chloroethane
Concen: 42.299 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 64 Resp: 86837

Ion	Ratio	Lower	Upper
64	100		
66	31.1	24.6	36.8






#7
Trichlorofluoromethane
Concen: 43.956 ug/l
RT: 3.501 min Scan# 26
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:101 Resp: 206526


Ion	Ratio	Lower	Upper
101	100		
103	65.1	51.4	77.2



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#8
Diethyl Ether
Concen: 42.846 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 74 Resp: 69545

Ion	Ratio	Lower	Upper
74	100		
45	99.0	49.7	149.1






#9
1,1,2-Trichlorotrifluoroethane
Concen: 44.706 ug/l
RT: 4.371 min Scan# 41
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:101 Resp: 118310


Ion	Ratio	Lower	Upper
101	100		
85	45.9	37.8	56.8
151	71.5	58.8	88.2



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#10
Methyl Iodide
Concen: 46.259 ug/l
RT: 4.595 min Scan# 449
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:142 Resp: 140190

Ion	Ratio	Lower	Upper
142	100		
127	49.2	40.3	60.5
141	13.9	12.8	19.2






#11
Tert butyl alcohol
Concen: 206.633 ug/l
RT: 5.518 min Scan# 60
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 59 Resp: 83933


Ion	Ratio	Lower	Upper
59	100		
57	10.4	8.6	13.0



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#12
1,1-Dichloroethene
Concen: 45.317 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 96 Resp: 106880

Ion	Ratio	Lower	Upper
96	100		
61	171.9	141.0	211.4
98	61.0	49.0	73.6






#13
Acrolein
Concen: 244.765 ug/l
RT: 4.177 min Scan# 37
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 56 Resp: 135720


Ion	Ratio	Lower	Upper
56	100		
55	70.9	56.3	84.5



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#14
Allyl chloride
Concen: 44.501 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 41 Resp: 170308

Ion	Ratio	Lower	Upper
41	100		
39	78.9	64.4	96.6
76	38.2	30.5	45.7






#15
Acrylonitrile
Concen: 227.920 ug/l
RT: 5.712 min Scan# 63
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 53 Resp: 294062


Ion	Ratio	Lower	Upper
53	100		
52	82.7	65.5	98.3
51	36.8	29.8	44.8



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#16
Acetone
Concen: 213.971 ug/l
RT: 4.418 min Scan# 419
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 245248

Ion	Ratio	Lower	Upper
43	100		
58	28.4	23.8	35.6






#17
Carbon Disulfide
Concen: 41.342 ug/l
RT: 4.712 min Scan# 46
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 76 Resp: 300215


Ion	Ratio	Lower	Upper
76	100		
78	8.7	6.9	10.3



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#18
Methyl Acetate
Concen: 45.028 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 156938

Ion	Ratio	Lower	Upper
43	100		
74	23.4	18.6	28.0






#19
Methyl tert-butyl Ether
Concen: 47.768 ug/l
RT: 5.795 min Scan# 65
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 73 Resp: 365819


Ion	Ratio	Lower	Upper
73	100		
57	22.3	17.8	26.8



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#20
Methylene Chloride
Concen: 43.017 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 84 Resp: 122059

Ion	Ratio	Lower	Upper
84	100		
49	131.3	104.1	156.1
51	41.0	31.0	46.4
86	63.3	51.9	77.9






#21
trans-1,2-Dichloroethene
Concen: 44.672 ug/l
RT: 5.783 min Scan# 65
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 96 Resp: 112597


Ion	Ratio	Lower	Upper
96	100		
61	134.3	115.7	173.5
98	59.1	49.8	74.8



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#22
Diisopropyl ether
Concen: 46.996 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 45 Resp: 399206

Ion	Ratio	Lower	Upper
45	100		
43	56.0	45.0	67.6
87	26.8	21.8	32.8
59	12.2	10.2	15.2






#23
Vinyl Acetate
Concen: 247.459 ug/l
RT: 6.601 min Scan# 79
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 1471068


Ion	Ratio	Lower	Upper
43	100		
86	10.7	8.4	12.6



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#24
1,1-Dichloroethane
Concen: 44.114 ug/l
RT: 6.571 min Scan# 785
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 63 Resp: 228489

Ion	Ratio	Lower	Upper
63	100		
98	7.5	3.3	9.8
100	4.2	2.0	6.0






#25
2-Butanone
Concen: 228.924 ug/l
RT: 7.477 min Scan# 93
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 386125


Ion	Ratio	Lower	Upper
43	100		
72	25.4	20.2	30.4



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#26
2,2-Dichloropropane
Concen: 47.593 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 77 Resp: 199303

Ion	Ratio	Lower	Upper
77	100		
97	20.9	10.7	32.1






#27
cis-1,2-Dichloroethene
Concen: 45.446 ug/l
RT: 7.483 min Scan# 94
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 96 Resp: 134920


Ion	Ratio	Lower	Upper
96	100		
61	158.0	0.0	311.8
98	64.1	0.0	126.0



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#28
Bromochloromethane
Concen: 44.122 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 49 Resp: 106323

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	4.4
130	68.6	55.0	82.4






#29
Tetrahydrofuran
Concen: 239.065 ug/l
RT: 7.836 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 42 Resp: 255650


Ion	Ratio	Lower	Upper
42	100		
72	42.5	34.2	51.2
71	39.9	32.5	48.7



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#30
Chloroform
Concen: 44.071 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 83 Resp: 235923

Ion	Ratio	Lower	Upper
83	100		
85	63.5	51.8	77.6






#31
Cyclohexane
Concen: 42.878 ug/l
RT: 8.253 min Scan# 10
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 56 Resp: 193034


Ion	Ratio	Lower	Upper
56	100		
69	30.5	22.2	33.4
84	89.4	66.4	99.6



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#32
1,1,1-Trichloroethane
Concen: 44.332 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 97 Resp: 208172

Ion	Ratio	Lower	Upper
97	100		
99	61.8	49.8	74.6
61	52.8	41.4	62.2






#33
1,2-Dichloroethane-d4
Concen: 49.782 ug/l
RT: 8.571 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 65 Resp: 176594


Ion	Ratio	Lower	Upper
65	100		
67	50.9	0.0	101.6



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.094 min Scan# 1214
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:114 Resp: 370487

Ion	Ratio	Lower	Upper
114	100		
63	21.6	0.0	47.6
88	15.8	0.0	32.6






#35
Dibromofluoromethane
Concen: 50.768 ug/l
RT: 8.165 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:113 Resp: 130487


Ion	Ratio	Lower	Upper
113	100		
111	102.5	82.7	124.1
192	17.7	14.3	21.5



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#36
1,1-Dichloropropene
Concen: 44.863 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 161834

Ion	Ratio	Lower	Upper
75	100		
110	32.4	16.5	49.5
77	31.8	24.4	36.6






#37
Ethyl Acetate
Concen: 44.123 ug/l
RT: 7.559 min Scan# 95
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 160656


Ion	Ratio	Lower	Upper
43	100		
61	11.9	10.9	16.3
70	10.2	7.5	11.3



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#38
Carbon Tetrachloride
Concen: 44.456 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:117 Resp: 183592

Ion	Ratio	Lower	Upper
117	100		
119	94.4	75.4	113.2
121	29.0	24.6	37.0






#39
Methylcyclohexane
Concen: 49.687 ug/l
RT: 9.600 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 83 Resp: 168431


Ion	Ratio	Lower	Upper
83	100		
55	80.9	62.6	94.0
98	46.9	37.7	56.5



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#40
Benzene
Concen: 45.657 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 78 Resp: 494871

Ion	Ratio	Lower	Upper
78	100		
77	23.5	19.0	28.6






#41
Methacrylonitrile
Concen: 47.155 ug/l
RT: 7.777 min Scan# 99
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 41 Resp: 89326


Ion	Ratio	Lower	Upper
41	100		
39	56.6	46.0	69.0
67	70.9	57.4	86.2
52	32.6	25.5	38.3



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#42
1,2-Dichloroethane
Concen: 45.481 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 62 Resp: 185663

Ion	Ratio	Lower	Upper
62	100		
98	8.3	0.0	17.0






#43
Isopropyl Acetate
Concen: 45.440 ug/l
RT: 8.683 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 265107


Ion	Ratio	Lower	Upper
43	100		
61	26.7	20.7	31.1
87	12.9	9.8	14.8



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#44
Trichloroethene
Concen: 43.622 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:130 Resp: 110058

Ion	Ratio	Lower	Upper
130	100		
95	101.9	0.0	195.8






#45
1,2-Dichloropropane
Concen: 44.744 ug/l
RT: 9.618 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 63 Resp: 123876


Ion	Ratio	Lower	Upper
63	100		
65	33.5	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#46
Dibromomethane
Concen: 44.214 ug/l
RT: 9.700 min Scan# 1317
Delta R.T. -0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 93 Resp: 88325

Ion	Ratio	Lower	Upper
93	100		
95	82.9	64.7	97.1
174	88.0	69.0	103.4






#47
Bromodichloromethane
Concen: 45.655 ug/l
RT: 9.889 min Scan# 13
Delta R.T. 0.007 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 83 Resp: 185808


Ion	Ratio	Lower	Upper
83	100		
85	63.5	51.2	76.8
127	7.8	6.5	9.7



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#48
Methyl methacrylate
Concen: 49.948 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 41 Resp: 131142

Ion	Ratio	Lower	Upper
41	100		
69	78.6	64.7	97.1
39	63.3	49.0	73.6






#49
1,4-Dioxane
Concen: 792.760 ug/l
RT: 9.689 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 88 Resp: 35051


Ion	Ratio	Lower	Upper
88	100		
43	33.8	26.6	39.8
58	76.1	59.5	89.3



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#50
Toluene-d8
Concen: 52.302 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 98 Resp: 477630

Ion	Ratio	Lower	Upper
98	100		
100	64.7	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 245.877 ug/l
RT: 10.441 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 832426


Ion	Ratio	Lower	Upper
43	100		
58	37.8	30.5	45.7



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#52
Toluene
Concen: 48.073 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 92 Resp: 301911

Ion	Ratio	Lower	Upper
92	100		
91	175.7	139.2	208.8






#53
t-1,3-Dichloropropene
Concen: 47.757 ug/l
RT: 10.830 min Scan# 15
Delta R.T. -0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 183680


Ion	Ratio	Lower	Upper
75	100		
77	32.6	23.5	35.3



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#54
cis-1,3-Dichloropropene
Concen: 47.783 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 196307

Ion	Ratio	Lower	Upper
75	100		
77	30.2	25.0	37.4
39	51.9	43.1	64.7






#55
1,1,2-Trichloroethane
Concen: 46.173 ug/l
RT: 11.012 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 97 Resp: 114757


Ion	Ratio	Lower	Upper
97	100		
83	83.7	69.0	103.6
85	53.2	44.6	66.8
99	62.8	48.6	72.8



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#56
Ethyl methacrylate
Concen: 43.762 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 69 Resp: 182170

Ion	Ratio	Lower	Upper
69	100		
41	70.4	54.6	82.0
39	40.2	32.4	48.6






#57
1,3-Dichloropropane
Concen: 46.877 ug/l
RT: 11.159 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 76 Resp: 202583


Ion	Ratio	Lower	Upper
76	100		
78	32.1	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#58
2-Chloroethyl Vinyl ether
Concen: 256.116 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 63 Resp: 403949

Ion	Ratio	Lower	Upper
63	100		
106	26.4	21.6	32.4






#59
2-Hexanone
Concen: 251.213 ug/l
RT: 11.194 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 598433


Ion	Ratio	Lower	Upper
43	100		
58	52.1	26.2	78.6



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#60
Dibromochloromethane
Concen: 45.935 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:129 Resp: 137831

Ion	Ratio	Lower	Upper
129	100		
127	78.1	38.6	115.8






#61
1,2-Dibromoethane
Concen: 44.837 ug/l
RT: 11.465 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:107 Resp: 110926


Ion	Ratio	Lower	Upper
107	100		
109	94.7	75.9	113.9



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#62
4-Bromofluorobenzene
Concen: 52.899 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 95 Resp: 165249

Ion	Ratio	Lower	Upper
95	100		
174	73.1	0.0	145.0
176	70.8	0.0	142.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:117 Resp: 320948


Ion	Ratio	Lower	Upper
117	100		
82	60.1	48.6	72.8
119	32.4	26.6	39.8



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#64
Tetrachloroethene
Concen: 44.886 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:164 Resp: 98211

Ion	Ratio	Lower	Upper
164	100		
166	124.4	103.4	155.2
129	104.8	79.2	118.8
131	93.1	77.1	115.7






#65
Chlorobenzene
Concen: 45.901 ug/l
RT: 11.888 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:112 Resp: 322723


Ion	Ratio	Lower	Upper
112	100		
114	32.4	25.3	37.9



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 44.712 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:131 Resp: 115379

Ion	Ratio	Lower	Upper
131	100		
133	95.8	47.4	142.3
119	65.2	33.1	99.5






#67
Ethyl Benzene
Concen: 50.391 ug/l
RT: 11.959 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 91 Resp: 577059


Ion	Ratio	Lower	Upper
91	100		
106	31.2	23.8	35.8



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#68
m/p-Xylenes
Concen: 103.341 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:106 Resp: 437367

Ion	Ratio	Lower	Upper
106	100		
91	210.1	167.7	251.5






#69
o-Xylene
Concen: 52.093 ug/l
RT: 12.394 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:106 Resp: 210698


Ion	Ratio	Lower	Upper
106	100		
91	219.9	110.4	331.2



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#70
Styrene
Concen: 53.103 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:104 Resp: 355459

Ion	Ratio	Lower	Upper
104	100		
78	52.8	42.5	63.7
103	54.5	43.8	65.8






#71
Bromoform
Concen: 49.119 ug/l
RT: 12.577 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:173 Resp: 90693


Ion	Ratio	Lower	Upper
173	100		
175	48.7	24.4	73.2
254	0.0	0.0	0.0



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:152 Resp: 158525

Ion	Ratio	Lower	Upper
152	100		
115	62.2	31.1	93.3
150	171.0	0.0	343.6






#73
Isopropylbenzene
Concen: 49.951 ug/l
RT: 12.694 min Scan# 18
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:105 Resp: 534428


Ion	Ratio	Lower	Upper
105	100		
120	25.2	12.8	38.3



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#74
N-amyl acetate
Concen: 48.730 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 43 Resp: 234342

Ion	Ratio	Lower	Upper
43	100		
70	42.2	34.0	51.0
55	26.5	21.4	32.2
61	23.8	19.0	28.4






#75
1,1,2,2-Tetrachloroethane
Concen: 43.627 ug/l
RT: 12.935 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 83 Resp: 164882


Ion	Ratio	Lower	Upper
83	100		
131	9.9	4.8	14.4
85	64.6	32.2	96.6



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#76
1,2,3-Trichloropropane
Concen: 41.891 ug/l m
RT: 12.988 min Scan# 1876
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 134976

Ion	Ratio	Lower	Upper
75	100		
77	213.9	109.7	329.2






#77
Bromobenzene
Concen: 46.138 ug/l
RT: 12.977 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:156 Resp: 128968


Ion	Ratio	Lower	Upper
156	100		
77	223.8	114.1	342.4
158	95.7	48.9	146.8



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#78
n-propylbenzene
Concen: 51.001 ug/l
RT: 13.029 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 91 Resp: 645977

Ion	Ratio	Lower	Upper
91	100		
120	21.7	10.9	32.6






#79
2-Chlorotoluene
Concen: 47.870 ug/l
RT: 13.124 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 91 Resp: 392404


Ion	Ratio	Lower	Upper
91	100		
126	31.8	15.7	47.1



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#80
1,3,5-Trimethylbenzene
Concen: 51.604 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:105 Resp: 456048

Ion	Ratio	Lower	Upper
105	100		
120	48.1	23.9	71.7






#81
trans-1,4-Dichloro-2-butene
Concen: 49.226 ug/l m
RT: 12.735 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 58623


Ion	Ratio	Lower	Upper
75	100		
53	118.0	95.6	143.4
89	43.3	37.0	55.6



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#82
4-Chlorotoluene
Concen: 48.776 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 91 Resp: 398110

Ion	Ratio	Lower	Upper
91	100		
126	31.6	15.9	47.7






#83
tert-Butylbenzene
Concen: 50.997 ug/l
RT: 13.435 min Scan# 19
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:119 Resp: 378478


Ion	Ratio	Lower	Upper
119	100		
91	70.6	35.1	105.3
134	26.7	12.0	36.1



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#84
1,2,4-Trimethylbenzene
Concen: 51.922 ug/l
RT: 13.477 min Scan# 1959
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:105 Resp: 457340

Ion	Ratio	Lower	Upper
105	100		
120	43.6	21.6	65.0






#85
sec-Butylbenzene
Concen: 51.777 ug/l
RT: 13.612 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:105 Resp: 532578


Ion	Ratio	Lower	Upper
105	100		
134	19.2	9.7	28.9



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#86
p-Isopropyltoluene
Concen: 47.795 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:119 Resp: 440284

Ion	Ratio	Lower	Upper
119	100		
134	25.3	12.7	38.0
91	25.0	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 45.920 ug/l
RT: 13.729 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:146 Resp: 236387


Ion	Ratio	Lower	Upper
146	100		
111	42.6	21.4	64.3
148	63.8	32.3	96.9



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#88
1,4-Dichlorobenzene
Concen: 44.102 ug/l
RT: 13.806 min Scan# 2015
Delta R.T. -0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:146 Resp: 235319

Ion	Ratio	Lower	Upper
146	100		
111	41.9	21.3	63.7
148	63.5	32.4	97.0






#89
n-Butylbenzene
Concen: 50.578 ug/l
RT: 14.053 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 91 Resp: 373227


Ion	Ratio	Lower	Upper
91	100		
92	51.9	25.8	77.3
134	23.8	11.7	35.1



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#90
Hexachloroethane
Concen: 45.215 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:117 Resp: 88538

Ion	Ratio	Lower	Upper
117	100		
201	67.2	33.7	101.0






#91
1,2-Dichlorobenzene
Concen: 45.113 ug/l
RT: 14.100 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:146 Resp: 231646


Ion	Ratio	Lower	Upper
146	100		
111	43.9	21.7	65.1
148	63.1	31.4	94.2



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 46.221 ug/l
RT: 14.718 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion: 75 Resp: 31899

Ion	Ratio	Lower	Upper
75	100		
155	69.0	36.4	109.2
157	92.3	45.4	136.1






#93
1,2,4-Trichlorobenzene
Concen: 45.739 ug/l
RT: 15.388 min Scan# 22
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:180 Resp: 109157


Ion	Ratio	Lower	Upper
180	100		
182	96.3	46.8	140.3
145	32.7	16.0	48.0



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025




#94
Hexachlorobutadiene
Concen: 42.112 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:225 Resp: 53358

Ion	Ratio	Lower	Upper
225	100		
223	59.8	30.7	92.1
227	61.2	30.9	92.5






#95
Naphthalene
Concen: 47.116 ug/l
RT: 15.641 min Scan# 23
Delta R.T. 0.006 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:128 Resp: 335535


Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.6	16.0
129	11.1	8.8	13.2



Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Mahesh Dadoda 01/15/2025



#96
1,2,3-Trichlorobenzene
Concen: 44.551 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN085445.D
Acq: 14 Jan 2025 18:06

Tgt Ion:180 Resp: 107583

Ion	Ratio	Lower	Upper
180	100		
182	95.6	47.4	142.2
145	34.3	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2	T	Dichlorodifluoromethane	0.677	0.593	12.4	104	0.00
3	P	Chloromethane	0.733	0.629	14.2	102	0.00
4	C	Vinyl Chloride	0.737	0.629	14.7#	101	0.00
5	T	Bromomethane	0.445	0.364	18.2	96	0.00
6	T	Chloroethane	0.467	0.395	15.4	103	0.00
7	T	Trichlorofluoromethane	1.069	0.940	12.1	104	0.00
8	T	Diethyl Ether	0.369	0.316	14.4	97	0.00
9	T	1,1,2-Trichlorotrifluoroeth	0.602	0.538	10.6	109	0.00
10	T	Methyl Iodide	0.690	0.638	7.5	98	0.00
11	T	Tert butyl alcohol	0.092	0.076	17.4	90	0.00
12	CM	1,1-Dichloroethene	0.537	0.486	9.5#	101	0.00
13	T	Acrolein	0.126	0.124	1.6	97	0.00
14	T	Allyl chloride	0.871	0.775	11.0	101	0.00
15	T	Acrylonitrile	0.294	0.268	8.8	99	0.00
16	T	Acetone	0.261	0.223	14.6	98	0.00
17	T	Carbon Disulfide	1.652	1.366	17.3	102	0.00
18	T	Methyl Acetate	0.793	0.714	10.0	100	0.00
19	T	Methyl tert-butyl Ether	1.742	1.665	4.4	98	0.00
20	T	Methylene Chloride	0.646	0.555	14.1	97	0.00
21	T	trans-1,2-Dichloroethene	0.573	0.512	10.6	102	0.00
22	T	Diisopropyl ether	1.933	1.817	6.0	98	0.00
23	T	Vinyl Acetate	1.353	1.339	1.0	99	0.00
24	P	1,1-Dichloroethane	1.178	1.040	11.7	98	0.00
25	T	2-Butanone	0.384	0.351	8.6	99	0.00
26	T	2,2-Dichloropropane	0.953	0.907	4.8	107	0.00
27	T	cis-1,2-Dichloroethene	0.675	0.614	9.0	99	0.00
28	T	Bromochloromethane	0.548	0.484	11.7	98	0.00
29	T	Tetrahydrofuran	0.243	0.233	4.1	98	0.00
30	C	Chloroform	1.218	1.074	11.8#	101	0.00
31	T	Cyclohexane	1.024	0.878	14.3	110	0.00
32	T	1,1,1-Trichloroethane	1.068	0.947	11.3	103	0.00
33	S	1,2-Dichloroethane-d4	0.807	0.804	0.4	107	0.00
34	I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35	S	Dibromofluoromethane	0.347	0.352	-1.4	107	0.00
36	T	1,1-Dichloropropene	0.487	0.437	10.3	103	0.00
37	T	Ethyl Acetate	0.491	0.434	11.6	97	0.00
38	T	Carbon Tetrachloride	0.557	0.496	11.0	102	0.00
39	T	Methylcyclohexane	0.457	0.455	0.4	107	0.00
40	TM	Benzene	1.463	1.336	8.7	100	0.00
41	T	Methacrylonitrile	0.256	0.241	5.9	97	0.00
42	TM	1,2-Dichloroethane	0.551	0.501	9.1	100	0.00
43	T	Isopropyl Acetate	0.787	0.716	9.0	96	0.00
44	TM	Trichloroethene	0.341	0.297	12.9	100	0.00
45	C	1,2-Dichloropropane	0.374	0.334	10.7#	98	0.00
46	T	Dibromomethane	0.270	0.238	11.9	97	0.00
47	T	Bromodichloromethane	0.549	0.502	8.6	98	0.00
48	T	Methyl methacrylate	0.354	0.354	0.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	0.006	0.005	16.7	81	0.00
50	S	Toluene-d8	1.232	1.289	-4.6	111	0.00
51	T	4-Methyl-2-Pentanone	0.457	0.449	1.8	99	0.00
52	CM	Toluene	0.848	0.815	3.9#	102	0.00
53	T	t-1,3-Dichloropropene	0.519	0.496	4.4	98	0.00
54	T	cis-1,3-Dichloropropene	0.554	0.530	4.3	98	0.00
55	T	1,1,2-Trichloroethane	0.335	0.310	7.5	99	0.00
56	T	Ethyl methacrylate	0.469	0.492	-4.9	99	0.00
57	T	1,3-Dichloropropane	0.583	0.547	6.2	99	0.00
58	T	2-Chloroethyl Vinyl ether	0.213	0.218	-2.3	99	0.00
59	T	2-Hexanone	0.321	0.323	-0.6	99	0.00
60	T	Dibromochloromethane	0.405	0.372	8.1	98	0.00
61	T	1,2-Dibromoethane	0.334	0.299	10.5	98	0.00
62	S	4-Bromofluorobenzene	0.422	0.446	-5.7	108	0.00
63	I	Chlorobenzene-d5	1.000	1.000	0.0	108	0.00
64	T	Tetrachloroethene	0.341	0.306	10.3	102	0.00
65	PM	Chlorobenzene	1.095	1.006	8.1	101	0.00
66	T	1,1,1,2-Tetrachloroethane	0.402	0.359	10.7	99	0.00
67	C	Ethyl Benzene	1.784	1.798	-0.8#	104	0.00
68	T	m/p-Xylenes	0.659	0.681	-3.3	104	0.00
69	T	o-Xylene	0.630	0.656	-4.1	104	0.00
70	T	Styrene	1.043	1.108	-6.2	102	0.00
71	P	Bromoform	0.288	0.283	1.7	98	0.00
72	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73	T	Isopropylbenzene	3.375	3.371	0.1	106	0.00
74	T	N-amyl acetate	1.517	1.478	2.6	99	0.00
75	P	1,1,2,2-Tetrachloroethane	1.192	1.040	12.8	98	0.00
76	T	1,2,3-Trichloropropane	1.016	0.851	16.2	103	0.00
77	T	Bromobenzene	0.882	0.814	7.7	103	0.00
78	T	n-propylbenzene	3.995	4.075	-2.0	106	0.00
79	T	2-Chlorotoluene	2.586	2.475	4.3	104	0.00
80	T	1,3,5-Trimethylbenzene	2.787	2.877	-3.2	106	0.00
81	T	trans-1,4-Dichloro-2-butene	0.376	0.370	1.6	103	0.00
82	T	4-Chlorotoluene	2.574	2.511	2.4	105	0.00
83	T	tert-Butylbenzene	2.341	2.387	-2.0	106	0.00
84	T	1,2,4-Trimethylbenzene	2.778	2.885	-3.9	103	0.00
85	T	sec-Butylbenzene	3.244	3.360	-3.6	106	0.00
86	T	p-Isopropyltoluene	2.631	2.777	-5.5	105	0.00
87	T	1,3-Dichlorobenzene	1.624	1.491	8.2	103	0.00
88	T	1,4-Dichlorobenzene	1.683	1.484	11.8	103	0.00
89	T	n-Butylbenzene	2.327	2.354	-1.2	103	0.00
90	T	Hexachloroethane	0.618	0.559	9.5	105	0.00
91	T	1,2-Dichlorobenzene	1.620	1.461	9.8	102	0.00
92	T	1,2-Dibromo-3-Chloropropane	0.218	0.201	7.8	98	0.00
93	T	1,2,4-Trichlorobenzene	0.753	0.689	8.5	95	0.00
94	T	Hexachlorobutadiene	0.400	0.337	15.8	99	0.00
95	T	Naphthalene	2.246	2.117	5.7	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.762	0.679	10.9	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	50.000	50.000	0.0	110	0.00
2	T	Dichlorodifluoromethane	50.000	43.772	12.5	104	0.00
3	P	Chloromethane	50.000	42.883	14.2	102	0.00
4	C	Vinyl Chloride	50.000	42.715	14.6#	101	0.00
5	T	Bromomethane	50.000	40.910	18.2	96	0.00
6	T	Chloroethane	50.000	42.299	15.4	103	0.00
7	T	Trichlorofluoromethane	50.000	43.956	12.1	104	0.00
8	T	Diethyl Ether	50.000	42.846	14.3	97	0.00
9	T	1,1,2-Trichlorotrifluoroeth	50.000	44.706	10.6	109	0.00
10	T	Methyl Iodide	50.000	46.259	7.5	98	0.00
11	T	Tert butyl alcohol	250.000	206.633	17.3	90	0.00
12	CM	1,1-Dichloroethene	50.000	45.317	9.4#	101	0.00
13	T	Acrolein	250.000	244.765	2.1	97	0.00
14	T	Allyl chloride	50.000	44.501	11.0	101	0.00
15	T	Acrylonitrile	250.000	227.920	8.8	99	0.00
16	T	Acetone	250.000	213.971	14.4	98	0.00
17	T	Carbon Disulfide	50.000	41.342	17.3	102	0.00
18	T	Methyl Acetate	50.000	45.028	9.9	100	0.00
19	T	Methyl tert-butyl Ether	50.000	47.768	4.5	98	0.00
20	T	Methylene Chloride	50.000	43.017	14.0	97	0.00
21	T	trans-1,2-Dichloroethene	50.000	44.672	10.7	102	0.00
22	T	Diisopropyl ether	50.000	46.996	6.0	98	0.00
23	T	Vinyl Acetate	250.000	247.459	1.0	99	0.00
24	P	1,1-Dichloroethane	50.000	44.114	11.8	98	0.00
25	T	2-Butanone	250.000	228.924	8.4	99	0.00
26	T	2,2-Dichloropropane	50.000	47.593	4.8	107	0.00
27	T	cis-1,2-Dichloroethene	50.000	45.446	9.1	99	0.00
28	T	Bromochloromethane	50.000	44.122	11.8	98	0.00
29	T	Tetrahydrofuran	250.000	239.065	4.4	98	0.00
30	C	Chloroform	50.000	44.071	11.9#	101	0.00
31	T	Cyclohexane	50.000	42.878	14.2	110	0.00
32	T	1,1,1-Trichloroethane	50.000	44.332	11.3	103	0.00
33	S	1,2-Dichloroethane-d4	50.000	49.782	0.4	107	0.00
34	I	1,4-Difluorobenzene	50.000	50.000	0.0	109	0.00
35	S	Dibromofluoromethane	50.000	50.768	-1.5	107	0.00
36	T	1,1-Dichloropropene	50.000	44.863	10.3	103	0.00
37	T	Ethyl Acetate	50.000	44.123	11.8	97	0.00
38	T	Carbon Tetrachloride	50.000	44.456	11.1	102	0.00
39	T	Methylcyclohexane	50.000	49.687	0.6	107	0.00
40	TM	Benzene	50.000	45.657	8.7	100	0.00
41	T	Methacrylonitrile	50.000	47.155	5.7	97	0.00
42	TM	1,2-Dichloroethane	50.000	45.481	9.0	100	0.00
43	T	Isopropyl Acetate	50.000	45.440	9.1	96	0.00
44	TM	Trichloroethene	50.000	43.622	12.8	100	0.00
45	C	1,2-Dichloropropane	50.000	44.744	10.5#	98	0.00
46	T	Dibromomethane	50.000	44.214	11.6	97	0.00
47	T	Bromodichloromethane	50.000	45.655	8.7	98	0.00
48	T	Methyl methacrylate	50.000	49.948	0.1	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	1000.000	792.760	20.7	81	0.00
50	S	Toluene-d8	50.000	52.302	-4.6	111	0.00
51	T	4-Methyl-2-Pentanone	250.000	245.877	1.6	99	0.00
52	CM	Toluene	50.000	48.073	3.9#	102	0.00
53	T	t-1,3-Dichloropropene	50.000	47.757	4.5	98	0.00
54	T	cis-1,3-Dichloropropene	50.000	47.783	4.4	98	0.00
55	T	1,1,2-Trichloroethane	50.000	46.173	7.7	99	0.00
56	T	Ethyl methacrylate	50.000	43.762	12.5	99	0.00
57	T	1,3-Dichloropropane	50.000	46.877	6.2	99	0.00
58	T	2-Chloroethyl Vinyl ether	250.000	256.116	-2.4	99	0.00
59	T	2-Hexanone	250.000	251.213	-0.5	99	0.00
60	T	Dibromochloromethane	50.000	45.935	8.1	98	0.00
61	T	1,2-Dibromoethane	50.000	44.837	10.3	98	0.00
62	S	4-Bromofluorobenzene	50.000	52.899	-5.8	108	0.00
63	I	Chlorobenzene-d5	50.000	50.000	0.0	108	0.00
64	T	Tetrachloroethene	50.000	44.886	10.2	102	0.00
65	PM	Chlorobenzene	50.000	45.901	8.2	101	0.00
66	T	1,1,1,2-Tetrachloroethane	50.000	44.712	10.6	99	0.00
67	C	Ethyl Benzene	50.000	50.391	-0.8#	104	0.00
68	T	m/p-Xylenes	100.000	103.341	-3.3	104	0.00
69	T	o-Xylene	50.000	52.093	-4.2	104	0.00
70	T	Styrene	50.000	53.103	-6.2	102	0.00
71	P	Bromoform	50.000	49.119	1.8	98	0.00
72	I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	108	0.00
73	T	Isopropylbenzene	50.000	49.951	0.1	106	0.00
74	T	N-amyl acetate	50.000	48.730	2.5	99	0.00
75	P	1,1,2,2-Tetrachloroethane	50.000	43.627	12.7	98	0.00
76	T	1,2,3-Trichloropropane	50.000	41.891	16.2	103	0.00
77	T	Bromobenzene	50.000	46.138	7.7	103	0.00
78	T	n-propylbenzene	50.000	51.001	-2.0	106	0.00
79	T	2-Chlorotoluene	50.000	47.870	4.3	104	0.00
80	T	1,3,5-Trimethylbenzene	50.000	51.604	-3.2	106	0.00
81	T	trans-1,4-Dichloro-2-butene	50.000	49.226	1.5	103	0.00
82	T	4-Chlorotoluene	50.000	48.776	2.4	105	0.00
83	T	tert-Butylbenzene	50.000	50.997	-2.0	106	0.00
84	T	1,2,4-Trimethylbenzene	50.000	51.922	-3.8	103	0.00
85	T	sec-Butylbenzene	50.000	51.777	-3.6	106	0.00
86	T	p-Isopropyltoluene	50.000	47.795	4.4	105	0.00
87	T	1,3-Dichlorobenzene	50.000	45.920	8.2	103	0.00
88	T	1,4-Dichlorobenzene	50.000	44.102	11.8	103	0.00
89	T	n-Butylbenzene	50.000	50.578	-1.2	103	0.00
90	T	Hexachloroethane	50.000	45.215	9.6	105	0.00
91	T	1,2-Dichlorobenzene	50.000	45.113	9.8	102	0.00
92	T	1,2-Dibromo-3-Chloropropane	50.000	46.221	7.6	98	0.00
93	T	1,2,4-Trichlorobenzene	50.000	45.739	8.5	95	0.00
94	T	Hexachlorobutadiene	50.000	42.112	15.8	99	0.00
95	T	Naphthalene	50.000	47.116	5.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085445.D
Acq On : 14 Jan 2025 18:06
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN011425

Quant Time: Jan 15 02:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	44.551	10.9	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG No.: Q1215

Instrument ID: MSVOA_X Calibration Date(s): 01/15/2025 01/15/2025

Heated Purge: (Y/N) N Calibration Time(s): 10:31 13:37

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID: RRF001 = VX044648.D RRF005 = VX044649.D RRF020 = VX044650.D
RRF050 = VX044651.D RRF100 = VX044652.D RRF150 = VX044653.D

COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	$\overline{RRF}$	% RSD
Vinyl Chloride	0.801	0.681	0.681	0.702	0.699	0.749	0.719	6.6
1,1-Dichloroethene	0.638	0.546	0.572	0.566	0.571	0.618	0.585	6
2-Butanone	0.451	0.444	0.444	0.426	0.449	0.450	0.444	2.1
Carbon Tetrachloride	0.585	0.503	0.468	0.460	0.465	0.502	0.497	9.4
Chloroform	1.261	1.226	1.185	1.142	1.095	1.206	1.186	5
Benzene	1.517	1.376	1.364	1.306	1.274	1.384	1.370	6.1
1,2-Dichloroethane	0.552	0.494	0.491	0.477	0.473	0.513	0.500	5.8
Trichloroethene	0.338	0.319	0.324	0.316	0.324	0.355	0.329	4.5
Tetrachloroethene	0.346	0.309	0.299	0.296	0.308	0.326	0.314	5.9
Chlorobenzene	1.178	1.085	1.036	1.008	0.998	1.067	1.062	6.2
1,2-Dichloroethane-d4		0.863	0.691	0.745	0.689	0.785	0.755	9.6
Dibromofluoromethane		0.355	0.289	0.318	0.293	0.333	0.318	8.8
Toluene-d8		1.270	1.051	1.111	1.001	1.104	1.108	9.1
4-Bromofluorobenzene		0.445	0.394	0.421	0.383	0.421	0.413	6

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Response Factor Report MSVOA_X2

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
Method File : 82X011525W.M
Title : SW846 8260
Last Update : Thu Jan 16 01:16:09 2025
Response Via : Initial Calibration

Calibration Files
1 =VX044648.D 5 =VX044649.D 20 =VX044650.D 50 =VX044651.D 100 =VX044652.D 150 =VX044653.D

		Compound	1	5	20	50	100	150	Avg	%RSD
1)	I	Pentafluorobenzene	----------ISTD----------							
2)	T	Dichlorodifluo...	0.596	0.606	0.641	0.679	0.751	0.763	0.673	10.67
3)	P	Chloromethane	0.826	0.712	0.723	0.710	0.682	0.713	0.728	6.88
4)	C	Vinyl Chloride	0.801	0.681	0.681	0.702	0.699	0.749	0.719	6.61#
5)	T	Bromomethane		0.170	0.166	0.171	0.169	0.183	0.172	3.73
6)	T	Chloroethane	0.165	0.123	0.113	0.117	0.124	0.149	0.132	15.70
7)	T	Trichlorofluor...	0.859	0.749	0.755	0.765	0.808	0.830	0.795	5.66
8)	T	Diethyl Ether	0.394	0.320	0.304	0.292	0.292	0.321	0.321	11.92
9)	T	1,1,2-Trichlor...	0.678	0.563	0.562	0.554	0.598	0.629	0.597	8.15
10)	T	Methyl Iodide		0.818	0.832	0.808	0.798	0.822	0.815	1.59
11)	T	Tert butyl alc...		0.211	0.165	0.151	0.152	0.162	0.168	14.66
12)	CM	1,1-Dichloroet...	0.638	0.546	0.572	0.566	0.571	0.618	0.585	5.99#
13)	T	Acrolein		0.113	0.066	0.068	0.068	0.080	0.079	25.32
14)	T	Allyl chloride	1.432	1.117	1.078	1.049	1.051	1.112	1.140	12.83
15)	T	Acrylonitrile	0.350	0.337	0.325	0.313	0.329	0.326	0.330	3.83
16)	T	Acetone	0.369	0.277	0.274	0.267	0.309	0.295	0.299	12.72
17)	T	Carbon Disulfide	1.690	1.556	1.571	1.544	1.552	1.673	1.597	4.11
18)	T	Methyl Acetate	1.041	0.832	0.826	0.880	0.942	1.006	0.921	9.78
19)	T	Methyl tert-bu...	2.328	2.120	2.091	2.032	1.998	2.155	2.121	5.49
20)	T	Methylene Chlo...	0.828	0.699	0.682	0.650	0.633	0.679	0.695	9.95
21)	T	trans-1,2-Dich...	0.738	0.607	0.590	0.568	0.559	0.605	0.611	10.69
22)	T	Diisopropyl ether	2.109	2.041	1.991	1.901	1.854	1.993	1.981	4.67
23)	T	Vinyl Acetate	1.848	1.805	1.757	1.650	1.637	1.708	1.734	4.88
24)	P	1,1-Dichloroet...	1.251	1.164	1.161	1.126	1.111	1.238	1.175	4.89
25)	T	2-Butanone	0.451	0.444	0.444	0.426	0.449	0.450	0.444	2.09
26)	T	2,2-Dichloropr...	1.362	1.129	1.094	1.054	1.048	1.145	1.139	10.20
27)	T	cis-1,2-Dichlo...	0.940	0.735	0.746	0.717	0.704	0.771	0.769	11.33
28)	T	Bromochloromet...	0.638	0.554	0.521	0.531	0.527	0.523	0.549	8.21
29)	T	Tetrahydrofuran	0.319	0.284	0.280	0.268	0.280	0.287	0.286	6.09
30)	C	Chloroform	1.261	1.226	1.185	1.142	1.095	1.206	1.186	5.02#
31)	T	Cyclohexane		0.938	0.956	0.931	0.935	0.991	0.950	2.62
32)	T	1,1,1-Trichlor...	1.393	1.036	1.037	0.997	0.991	1.087	1.090	13.97
33)	S	1,2-Dichloroet...		0.863	0.691	0.745	0.689	0.785	0.755	9.62
34)	I	1,4-Difluorobenzene	----------ISTD----------							
35)	S	Dibromofluorom...		0.355	0.289	0.318	0.293	0.333	0.318	8.75
36)	T	1,1-Dichloropr...	0.545	0.420	0.422	0.408	0.415	0.444	0.442	11.69
37)	T	Ethyl Acetate	0.420	0.485	0.472	0.463	0.487	0.498	0.471	5.94
38)	T	Carbon Tetrach...	0.585	0.503	0.468	0.460	0.465	0.502	0.497	9.44
39)	T	Methylcyclohexane	0.662	0.555	0.556	0.551	0.583	0.614	0.587	7.49
40)	TM	Benzene	1.517	1.376	1.364	1.306	1.274	1.384	1.370	6.12
41)	T	Methacrylonitrile	0.310	0.277	0.257	0.254	0.273	0.274	0.274	7.24
42)	TM	1,2-Dichloroet...	0.552	0.494	0.491	0.477	0.473	0.513	0.500	5.79
43)	T	Isopropyl Acetate	0.981	0.845	0.850	0.815	0.831	0.875	0.866	6.88
44)	TM	Trichloroethene	0.338	0.319	0.324	0.316	0.324	0.355	0.329	4.46
45)	C	1,2-Dichloropr...	0.367	0.359	0.335	0.329	0.326	0.354	0.345	5.01#
46)	T	Dibromomethane	0.300	0.271	0.249	0.246	0.243	0.266	0.263	8.19
47)	T	Bromodichlorom...	0.634	0.530	0.518	0.508	0.499	0.542	0.538	9.13
48)	T	Methyl methacr...	0.389	0.390	0.408	0.399	0.403	0.423	0.402	3.21
49)	T	1,4-Dioxane	0.010	0.008	0.008	0.008	0.008	0.008	0.008	7.10
50)	S	Toluene-d8		1.270	1.051	1.111	1.001	1.104	1.108	9.12
51)	T	4-Methyl-2-Pen...	0.549	0.489	0.481	0.449	0.445	0.458	0.479	8.09
52)	CM	Toluene	0.941	0.864	0.849	0.805	0.769	0.804	0.839	7.25#
53)	T	t-1,3-Dichloro...	0.578	0.551	0.564	0.546	0.537	0.563	0.556	2.65
54)	T	cis-1,3-Dichlo...	0.684	0.589	0.603	0.574	0.567	0.603	0.603	7.02
55)	T	1,1,2-Trichlor...	0.401	0.342	0.341	0.319	0.312	0.321	0.339	9.61
56)	T	Ethyl methacry...	0.641	0.576	0.598	0.564	0.547	0.574	0.583	5.61

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
Method File : 82X011525W.M

57)	T	1,3-Dichloropr...	0.652	0.594	0.581	0.555	0.527	0.557	0.578	7.46
58)	T	2-Chloroethyl ...	0.377	0.281	0.252	0.273	0.265	0.271	0.286	15.86
59)	T	2-Hexanone	0.397	0.358	0.357	0.330	0.331	0.341	0.352	7.18
60)	T	Dibromochlorom...	0.449	0.407	0.399	0.381	0.369	0.386	0.398	7.06
61)	T	1,2-Dibromoethane	0.411	0.361	0.356	0.337	0.328	0.341	0.356	8.36
62)	S	4-Bromofluorob...		0.445	0.394	0.421	0.383	0.421	0.413	5.96
63)	I	Chlorobenzene-d5	----------ISTD----------							
64)	T	Tetrachloroethene	0.346	0.309	0.299	0.296	0.308	0.326	0.314	5.92
65)	PM	Chlorobenzene	1.178	1.085	1.036	1.008	0.998	1.067	1.062	6.18
66)	T	1,1,1,2-Tetrac...	0.420	0.398	0.376	0.367	0.361	0.394	0.386	5.72
67)	C	Ethyl Benzene	2.026	1.839	1.827	1.779	1.777	1.871	1.853	4.96#
68)	T	m/p-Xylenes	0.771	0.697	0.686	0.656	0.646	0.669	0.688	6.55
69)	T	o-Xylene	0.787	0.713	0.690	0.665	0.644	0.674	0.695	7.27
70)	T	Styrene	1.172	1.146	1.133	1.089	1.069	1.128	1.123	3.37
71)	P	Bromoform	0.333	0.315	0.302	0.302	0.304	0.331	0.314	4.58
72)	I	1,4-Dichlorobenzen...	----------ISTD----------							
73)	T	Isopropylbenzene	4.514	4.259	4.043	3.917	3.790	3.985	4.085	6.40
74)	T	N-amyl acetate	2.228	2.047	1.969	1.900	1.882	2.055	2.013	6.32
75)	P	1,1,2,2-Tetrac...	1.618	1.510	1.398	1.302	1.267	1.370	1.411	9.35
76)	T	1,2,3-Trichlor...	1.321	1.369	1.173	1.102	1.140	1.176	1.213	8.76
77)	T	Bromobenzene	1.085	0.985	0.928	0.903	0.876	0.952	0.955	7.79
78)	T	n-propylbenzene	4.902	4.366	4.481	4.392	4.346	4.541	4.505	4.63
79)	T	2-Chlorotoluene	3.385	2.964	2.860	2.763	2.657	2.831	2.910	8.73
80)	T	1,3,5-Trimethy...	3.707	3.522	3.312	3.187	3.057	3.167	3.325	7.38
81)	T	trans-1,4-Dich...		0.535	0.517	0.512	0.522	0.567	0.531	4.15
82)	T	4-Chlorotoluene	3.509	3.220	3.182	3.075	2.978	3.161	3.188	5.64
83)	T	tert-Butylbenzene	4.017	3.693	3.410	3.304	3.137	3.300	3.477	9.27
84)	T	1,2,4-Trimethy...	3.967	3.380	3.346	3.211	3.097	3.273	3.379	9.04
85)	T	sec-Butylbenzene	4.515	4.037	3.977	3.895	3.855	4.009	4.048	5.90
86)	T	p-Isopropyltol...	3.663	3.337	3.277	3.222	3.161	3.303	3.327	5.29
87)	T	1,3-Dichlorobe...	1.733	1.641	1.619	1.613	1.608	1.720	1.656	3.38
88)	T	1,4-Dichlorobe...	1.911	1.664	1.606	1.577	1.579	1.705	1.674	7.58
89)	T	n-Butylbenzene	2.992	2.713	2.770	2.806	2.878	2.986	2.857	4.02
90)	T	Hexachloroethane	0.957	0.755	0.704	0.690	0.678	0.736	0.753	13.77
91)	T	1,2-Dichlorobe...	1.919	1.756	1.665	1.626	1.548	1.687	1.700	7.48
92)	T	1,2-Dibromo-3-...	0.407	0.355	0.320	0.310	0.322	0.356	0.345	10.33
93)	T	1,2,4-Trichlor...	0.956	0.943	0.945	1.031	1.061	1.163	1.016	8.54
94)	T	Hexachlorobuta...	0.419	0.384	0.373	0.393	0.401	0.426	0.399	5.14
95)	T	Naphthalene	4.049	3.759	3.797	3.814	3.860	4.244	3.920	4.80
96)	T	1,2,3-Trichlor...	1.058	0.993	1.006	1.055	1.055	1.156	1.054	5.44

(#) = Out of Range

Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044648.D
Acq On : 15 Jan 2025 10:31
Operator : JC/MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jan 16 00:58:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED
Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) Pentafluorobenzene	5.538	168	214233	50.000	ug/l		0.00
34) 1,4-Difluorobenzene	6.751	114	389452	50.000	ug/l		0.00
63) Chlorobenzene-d5	10.049	117	329917	50.000	ug/l		0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	135288	50.000	ug/l		0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l		
Spiked Amount 50.000	Range 74 - 125		Recovery =	0.000%#			
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l		
Spiked Amount 50.000	Range 75 - 124		Recovery =	0.000%#			
50) Toluene-d8	0.000	98	0d	0.000	ug/l		
Spiked Amount 50.000	Range 86 - 113		Recovery =	0.000%#			
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l		
Spiked Amount 50.000	Range 77 - 121		Recovery =	0.000%#			
Target Compounds							Qvalue
2) Dichlorodifluoromethane	1.167	85	2555	0.886	ug/l		97
3) Chloromethane	1.295	50	3539	1.135	ug/l		96
4) Vinyl Chloride	1.374	62	3434	1.115	ug/l	#	64
6) Chloroethane	1.660	64	708	1.253	ug/l	#	83
7) Trichlorofluoromethane	1.862	101	3682	1.082	ug/l		89
8) Diethyl Ether	2.130	74	1689	1.229	ug/l		87
9) 1,1,2-Trichlorotrifluo...	2.313	101	2905	1.135	ug/l	#	88
12) 1,1-Dichloroethene	2.313	96	2735	1.090	ug/l		82
14) Allyl chloride	2.654	41	6137	1.257	ug/l		88
15) Acrylonitrile	3.069	53	7502	5.306	ug/l		95
16) Acetone	2.386	43	7914	6.186	ug/l		97
17) Carbon Disulfide	2.496	76	7240	1.058	ug/l		96
18) Methyl Acetate	2.703	43	4460	1.130	ug/l		97
19) Methyl tert-butyl Ether	3.111	73	9974	1.098	ug/l		96
20) Methylene Chloride	2.782	84	3548	1.191	ug/l		94
21) trans-1,2-Dichloroethene	3.087	96	3164	1.208	ug/l		94
22) Diisopropyl ether	3.758	45	9038	1.065	ug/l	#	76
23) Vinyl Acetate	3.721	43	39593	5.328	ug/l		98
24) 1,1-Dichloroethane	3.599	63	5359	1.064	ug/l		98
25) 2-Butanone	4.574	43	9672	5.085	ug/l		94
26) 2,2-Dichloropropane	4.459	77	5837	1.196	ug/l		95
27) cis-1,2-Dichloroethene	4.483	96	4029	1.223	ug/l		87
28) Bromochloromethane	4.891	49	2733	1.162	ug/l	#	95
29) Tetrahydrofuran	5.019	42	6842	5.576	ug/l		98
30) Chloroform	5.087	83	5401	1.063	ug/l		85
32) 1,1,1-Trichloroethane	5.379	97	5968	1.278	ug/l	#	47
36) 1,1-Dichloropropene	5.678	75	4243	1.232	ug/l	#	92
37) Ethyl Acetate	4.727	43	3269	0.891	ug/l	#	88
38) Carbon Tetrachloride	5.666	117	4558	1.177	ug/l		85
39) Methylcyclohexane	7.367	83	5156	1.128	ug/l		98
40) Benzene	6.025	78	11813	1.107	ug/l		97
41) Methacrylonitrile	4.928	41	2415m	1.146	ug/l		
42) 1,2-Dichloroethane	6.092	62	4299	1.104	ug/l		91
43) Isopropyl Acetate	6.355	43	7640	1.132	ug/l		96
44) Trichloroethene	7.123	130	2633	1.026	ug/l		91
45) 1,2-Dichloropropane	7.434	63	2857	1.064	ug/l		99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044648.D
Acq On : 15 Jan 2025 10:31
Operator : JC/MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Quant Time: Jan 16 00:58:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
46) Dibromomethane	7.580	93	2337	1.143	ug/l		94
47) Bromodichloromethane	7.818	83	4937	1.177	ug/l		89
48) Methyl methacrylate	7.702	41	3029	0.967	ug/l		91
49) 1,4-Dioxane	7.696	88	1485	22.727	ug/l	#	87
51) 4-Methyl-2-Pentanone	8.580	43	21388	5.736	ug/l		100
52) Toluene	8.714	92	7332	1.122	ug/l		93
53) t-1,3-Dichloropropene	8.982	75	4499	1.038	ug/l		99
54) cis-1,3-Dichloropropene	8.373	75	5328	1.134	ug/l	#	87
55) 1,1,2-Trichloroethane	9.153	97	3124	1.182	ug/l		91
56) Ethyl methacrylate	9.122	69	4994	1.099	ug/l		89
57) 1,3-Dichloropropane	9.311	76	5078	1.128	ug/l		92
58) 2-Chloroethyl Vinyl ether	8.245	63	14681	6.581	ug/l		98
59) 2-Hexanone	9.439	43	15480	5.643	ug/l		98
60) Dibromochloromethane	9.519	129	3495	1.127	ug/l		100
61) 1,2-Dibromoethane	9.610	107	3200	1.155	ug/l		96
64) Tetrachloroethene	9.269	164	2280	1.101	ug/l		87
65) Chlorobenzene	10.073	112	7770	1.109	ug/l		99
66) 1,1,1,2-Tetrachloroethane	10.159	131	2769	1.087	ug/l	#	62
67) Ethyl Benzene	10.195	91	13366	1.093	ug/l		93
68) m/p-Xylenes	10.299	106	10180	2.244	ug/l		100
69) o-Xylene	10.640	106	5193	1.132	ug/l		97
70) Styrene	10.659	104	7732	1.044	ug/l		93
71) Bromoform	10.799	173	2194	1.058	ug/l	#	93
73) Isopropylbenzene	10.964	105	12214	1.105	ug/l		97
74) N-amyl acetate	10.848	43	6028	1.107	ug/l		97
75) 1,1,2,2-Tetrachloroethane	11.214	83	4377	1.147	ug/l		91
76) 1,2,3-Trichloropropane	11.244	75	3574m	1.119	ug/l		
77) Bromobenzene	11.201	156	2937	1.137	ug/l		93
78) n-propylbenzene	11.305	91	13264	1.088	ug/l		98
79) 2-Chlorotoluene	11.366	91	9159	1.163	ug/l		98
80) 1,3,5-Trimethylbenzene	11.451	105	10031	1.115	ug/l		97
82) 4-Chlorotoluene	11.457	91	9494	1.101	ug/l		93
83) tert-Butylbenzene	11.713	119	10869	1.155	ug/l		96
84) 1,2,4-Trimethylbenzene	11.750	105	10735	1.174	ug/l		94
85) sec-Butylbenzene	11.890	105	12217	1.115	ug/l		99
86) p-Isopropyltoluene	12.012	119	9910	1.101	ug/l		90
87) 1,3-Dichlorobenzene	11.969	146	4688	1.046	ug/l		97
88) 1,4-Dichlorobenzene	12.037	146	5172m	1.163	ug/l		
89) n-Butylbenzene	12.335	91	8095	1.047	ug/l		94
90) Hexachloroethane	12.536	117	2589	1.270	ug/l		92
91) 1,2-Dichlorobenzene	12.341	146	5192	1.128	ug/l		98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1100	1.178	ug/l		97
93) 1,2,4-Trichlorobenzene	13.591	180	2587	0.941	ug/l		92
94) Hexachlorobutadiene	13.725	225	1133	1.049	ug/l		91
95) Naphthalene	13.780	128	10956	1.033	ug/l		99
96) 1,2,3-Trichlorobenzene	13.963	180	2863	1.004	ug/l		96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044648.D
Acq On : 15 Jan 2025 10:31
Operator : JC/MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1


Quant Time: Jan 16 00:58:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:168 Resp: 21423


Ion	Ratio	Lower	Upper
168	100		
99	66.1	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#2
Dichlorodifluoromethane
Concen: 0.886 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 85 Resp: 2555

Ion	Ratio	Lower	Upper
85	100		
87	31.8	16.8	50.4






#3
Chloromethane
Concen: 1.135 ug/l
RT: 1.295 min Scan# 34
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 50 Resp: 3539


Ion	Ratio	Lower	Upper
50	100		
52	31.5	27.0	40.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#4
Vinyl Chloride
Concen: 1.115 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 62 Resp: 3434

Ion	Ratio	Lower	Upper
62	100		
64	53.6	26.6	39.8#






#6
Chloroethane
Concen: 1.253 ug/l
RT: 1.660 min Scan# 9
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 64 Resp: 70


Ion	Ratio	Lower	Upper
64	100		
66	23.4	26.6	39.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#7
Trichlorofluoromethane
Concen: 1.082 ug/l
RT: 1.862 min Scan# 127
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:101 Resp: 3682

Ion	Ratio	Lower	Upper
101	100		
103	57.0	52.5	78.7






#8
Diethyl Ether
Concen: 1.229 ug/l
RT: 2.130 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 74 Resp: 168


Ion	Ratio	Lower	Upper
74	100		
45	106.7	47.1	141.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.135 ug/l
RT: 2.313 min Scan# 201
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:101 Resp: 2905

Ion	Ratio	Lower	Upper
101	100		
85	62.7	35.3	52.9#
151	71.2	58.2	87.4






#12
1,1-Dichloroethene
Concen: 1.090 ug/l
RT: 2.313 min Scan# 2
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 96 Resp: 273


Ion	Ratio	Lower	Upper
96	100		
61	131.3	131.1	196.7
98	63.5	51.2	76.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#14
Allyl chloride
Concen: 1.257 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 41 Resp: 6137

Ion	Ratio	Lower	Upper
41	100		
39	58.2	54.9	82.3
76	28.7	28.4	42.6






#15
Acrylonitrile
Concen: 5.306 ug/l
RT: 3.069 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 53 Resp: 750


Ion	Ratio	Lower	Upper
53	100		
52	86.7	66.0	99.0
51	40.4	28.9	43.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#16
Acetone
Concen: 6.186 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 43 Resp: 7914

Ion	Ratio	Lower	Upper
43	100		
58	32.5	24.6	37.0






#17
Carbon Disulfide
Concen: 1.058 ug/l
RT: 2.496 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 76 Resp: 724


Ion	Ratio	Lower	Upper
76	100		
78	10.2	7.1	10.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#18
Methyl Acetate
Concen: 1.130 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 43 Resp: 4460

Ion	Ratio	Lower	Upper
43	100		
74	26.5	19.8	29.8






#19
Methyl tert-butyl Ether
Concen: 1.098 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 73 Resp: 997


Ion	Ratio	Lower	Upper
73	100		
57	23.7	17.4	26.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#20
Methylene Chloride
Concen: 1.191 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 84 Resp: 3548

Ion	Ratio	Lower	Upper
84	100		
49	108.9	94.2	141.4
51	34.2	28.8	43.2
86	64.5	50.3	75.5






#21
trans-1,2-Dichloroethene
Concen: 1.208 ug/l
RT: 3.087 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 96 Resp: 316


Ion	Ratio	Lower	Upper
96	100		
61	131.8	111.0	166.6
98	58.2	50.2	75.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#22
Diisopropyl ether
Concen: 1.065 ug/l
RT: 3.758 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 45 Resp: 9038

Ion	Ratio	Lower	Upper
45	100		
43	106.2	61.0	91.6#
87	25.7	22.7	34.1
59	11.8	9.4	14.2






#23
Vinyl Acetate
Concen: 5.328 ug/l
RT: 3.721 min Scan# 4
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 43 Resp: 3959


Ion	Ratio	Lower	Upper
43	100		
86	10.3	8.9	13.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#24
1,1-Dichloroethane
Concen: 1.064 ug/l
RT: 3.599 min Scan# 412
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 63 Resp: 5359

Ion	Ratio	Lower	Upper
63	100		
98	8.4	3.6	10.9
100	5.2	2.5	7.5






#25
2-Butanone
Concen: 5.085 ug/l
RT: 4.574 min Scan# 5
Delta R.T. 0.018 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 43 Resp: 967


Ion	Ratio	Lower	Upper
43	100		
72	30.4	21.7	32.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#26
2,2-Dichloropropane
Concen: 1.196 ug/l
RT: 4.459 min Scan# 553
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 77 Resp: 5837

Ion	Ratio	Lower	Upper
77	100		
97	24.2	10.9	32.6






#27
cis-1,2-Dichloroethene
Concen: 1.223 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 96 Resp: 402


Ion	Ratio	Lower	Upper
96	100		
61	124.2	0.0	296.0
98	64.2	0.0	130.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#28
Bromochloromethane
Concen: 1.162 ug/l
RT: 4.891 min Scan# 624
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 49 Resp: 2733

Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	4.4
130	74.9	63.4	95.2






#29
Tetrahydrofuran
Concen: 5.576 ug/l
RT: 5.019 min Scan# 64
Delta R.T. 0.018 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 42 Resp: 684


Ion	Ratio	Lower	Upper
42	100		
72	48.2	38.2	57.4
71	46.5	35.3	52.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#30
Chloroform
Concen: 1.063 ug/l
RT: 5.087 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 83 Resp: 5401

Ion	Ratio	Lower	Upper
83	100		
85	75.4	51.2	76.8






#32
1,1,1-Trichloroethane
Concen: 1.278 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 97 Resp: 596


Ion	Ratio	Lower	Upper
97	100		
99	0.0	51.6	77.4
61	37.6	35.6	53.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#33
1,2-Dichloroethane-d4
Concen: 0.000 ug/l
Expected RT: 5.95 min

Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 65

Sig	Exp Ratio
65	100
67	51.7



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.751 min Scan# 9
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31


Tgt Ion:114 Resp: 38945


Ion	Ratio	Lower	Upper
114	100		
63	21.8	0.0	43.0
88	17.4	0.0	35.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#35
Dibromofluoromethane
Concen: 0.000 ug/l
Expected RT: 5.38 min

Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 113

Sig	Exp Ratio
113	100
111	102.0
192	18.0






#36
1,1-Dichloropropene
Concen: 1.232 ug/l
RT: 5.678 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 75 Resp: 424


Ion	Ratio	Lower	Upper
75	100		
110	34.5	17.9	53.8
77	22.1	24.5	36.7#



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#37
Ethyl Acetate
Concen: 0.891 ug/l
RT: 4.727 min Scan# 597
Delta R.T. 0.019 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 43 Resp: 3269

Ion	Ratio	Lower	Upper
43	100		
61	7.9	10.9	16.3#
70	16.0	10.1	15.1#






#38
Carbon Tetrachloride
Concen: 1.177 ug/l
RT: 5.666 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:117 Resp: 455


Ion	Ratio	Lower	Upper
117	100		
119	76.4	76.0	114.0
121	32.5	24.6	36.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#39
Methylcyclohexane
Concen: 1.128 ug/l
RT: 7.367 min Scan# 1030
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 83 Resp: 5156

Ion	Ratio	Lower	Upper
83	100		
55	76.8	60.5	90.7
98	46.4	39.0	58.4






#40
Benzene
Concen: 1.107 ug/l
RT: 6.025 min Scan# 8
Delta R.T. -0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 78 Resp: 1181


Ion	Ratio	Lower	Upper
78	100		
77	25.1	18.7	28.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#41
Methacrylonitrile
Concen: 1.146 ug/l m
RT: 4.928 min Scan# 630
Delta R.T. 0.012 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 41 Resp: 2415

Ion	Ratio	Lower	Upper
41	100		
39	28.4	44.4	66.6#
67	60.1	60.6	90.8#
52	24.1	24.3	36.5#






#42
1,2-Dichloroethane
Concen: 1.104 ug/l
RT: 6.092 min Scan# 8
Delta R.T. 0.012 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 62 Resp: 429


Ion	Ratio	Lower	Upper
62	100		
98	13.1	0.0	19.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#43
Isopropyl Acetate
Concen: 1.132 ug/l
RT: 6.355 min Scan# 864
Delta R.T. 0.018 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 43 Resp: 7640

Ion	Ratio	Lower	Upper
43	100		
61	18.9	16.6	24.8
87	12.6	11.0	16.6






#44
Trichloroethene
Concen: 1.026 ug/l
RT: 7.123 min Scan# 9
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:130 Resp: 263


Ion	Ratio	Lower	Upper
130	100		
95	95.2	0.0	208.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#45
1,2-Dichloropropane
Concen: 1.064 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 63 Resp: 2857

Ion	Ratio	Lower	Upper
63	100		
65	29.8	24.2	36.2






#46
Dibromomethane
Concen: 1.143 ug/l
RT: 7.580 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 93 Resp: 233


Ion	Ratio	Lower	Upper
93	100		
95	81.0	66.4	99.6
174	81.3	72.0	108.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#47
Bromodichloromethane
Concen: 1.177 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 83 Resp: 4937

Ion	Ratio	Lower	Upper
83	100		
85	53.1	50.1	75.1
127	8.8	6.5	9.7






#48
Methyl methacrylate
Concen: 0.967 ug/l
RT: 7.702 min Scan# 1
Delta R.T. 0.012 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 41 Resp: 302


Ion	Ratio	Lower	Upper
41	100		
69	91.2	66.8	100.2
39	51.7	47.4	71.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#49
1,4-Dioxane
Concen: 22.727 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.031 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 88 Resp: 1485

Ion	Ratio	Lower	Upper
88	100		
43	33.9	27.0	40.4
58	85.0	55.8	83.8#





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 8.65 min

Lab File: VX044648.D
Acq: 15 Jan 2025 10:31


Tgt Ion: 98


Sig	Exp Ratio
98	100
100	66.5

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025








#51
4-Methyl-2-Pentanone
Concen: 5.736 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31


Tgt Ion: 43 Resp: 21388

Ion	Ratio	Lower	Upper
43	100		
58	40.0	31.9	47.9










#52
Toluene
Concen: 1.122 ug/l
RT: 8.714 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 92 Resp: 733


Ion	Ratio	Lower	Upper
92	100		
91	178.6	135.7	203.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#53
t-1,3-Dichloropropene
Concen: 1.038 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 75 Resp: 4499

Ion	Ratio	Lower	Upper
75	100		
77	31.1	24.4	36.6






#54
cis-1,3-Dichloropropene
Concen: 1.134 ug/l
RT: 8.373 min Scan# 1
Delta R.T. 0.012 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 75 Resp: 532


Ion	Ratio	Lower	Upper
75	100		
77	23.0	25.1	37.7#
39	42.1	40.3	60.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#55
1,1,2-Trichloroethane
Concen: 1.182 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 97 Resp: 3124

Ion	Ratio	Lower	Upper
97	100		
83	97.1	68.8	103.2
85	63.4	43.7	65.5
99	61.0	49.5	74.3






#56
Ethyl methacrylate
Concen: 1.099 ug/l
RT: 9.122 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 69 Resp: 499


Ion	Ratio	Lower	Upper
69	100		
41	56.4	53.2	79.8
39	41.7	30.2	45.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#57
1,3-Dichloropropane
Concen: 1.128 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 76 Resp: 5078

Ion	Ratio	Lower	Upper
76	100		
78	36.9	26.0	39.0





#58
2-Chloroethyl Vinyl ether
Concen: 6.581 ug/l
RT: 8.245 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31


Tgt Ion: 63 Resp: 1468


Ion	Ratio	Lower	Upper
63	100		
106	29.8	23.0	34.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#59
2-Hexanone
Concen: 5.643 ug/l
RT: 9.439 min Scan# 1370
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 43 Resp: 15480

Ion	Ratio	Lower	Upper
43	100		
58	54.4	27.8	83.3






#60
Dibromochloromethane
Concen: 1.127 ug/l
RT: 9.519 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:129 Resp: 349


Ion	Ratio	Lower	Upper
129	100		
127	77.3	38.5	115.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#61
1,2-Dibromoethane
Concen: 1.155 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:107 Resp: 3200

Ion	Ratio	Lower	Upper
107	100		
109	98.2	75.3	112.9





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 11.08 min

Lab File: VX044648.D
Acq: 15 Jan 2025 10:31


Tgt Ion: 95


Sig	Exp Ratio
95	100
174	66.5
176	64.7


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001


Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025











#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:117 Resp: 329917

Ion	Ratio	Lower	Upper
117	100		
82	59.0	47.6	71.4
119	31.7	25.4	38.2





#64
Tetrachloroethene
Concen: 1.101 ug/l
RT: 9.269 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31


Tgt Ion:164 Resp: 228

Ion	Ratio	Lower	Upper
164	100		
166	114.6	102.0	153.0
129	90.7	81.1	121.7
131	82.7	79.0	118.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#65
Chlorobenzene
Concen: 1.109 ug/l
RT: 10.073 min Scan# 1474
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:112 Resp: 7770

Ion	Ratio	Lower	Upper
112	100		
114	32.2	25.3	37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 1.087 ug/l
RT: 10.159 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31


Tgt Ion:131 Resp: 276

Ion	Ratio	Lower	Upper
131	100		
133	100.8	47.8	143.3
119	0.0	33.4	100.2

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#67
Ethyl Benzene
Concen: 1.093 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 91 Resp: 13366

Ion	Ratio	Lower	Upper
91	100		
106	27.6	25.0	37.6



#68
m/p-Xylenes
Concen: 2.244 ug/l
RT: 10.299 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31


Tgt Ion:106 Resp: 1018


Ion	Ratio	Lower	Upper
106	100		
91	207.4	166.4	249.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#69
o-Xylene
Concen: 1.132 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:106 Resp: 5193

Ion	Ratio	Lower	Upper
106	100		
91	213.5	108.7	326.3






#70
Styrene
Concen: 1.044 ug/l
RT: 10.659 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:104 Resp: 773


Ion	Ratio	Lower	Upper
104	100		
78	61.6	43.0	64.6
103	56.1	43.4	65.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#71
Bromoform
Concen: 1.058 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:173 Resp: 2194

Ion	Ratio	Lower	Upper
173	100		
175	51.7	23.6	70.8
254	0.0	0.1	0.1#






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:152 Resp: 13528


Ion	Ratio	Lower	Upper
152	100		
115	64.1	43.8	131.4
150	159.9	0.0	347.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#73
Isopropylbenzene
Concen: 1.105 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:105 Resp: 12214

Ion	Ratio	Lower	Upper
105	100		
120	27.8	13.2	39.6






#74
N-amyl acetate
Concen: 1.107 ug/l
RT: 10.848 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 43 Resp: 602


Ion	Ratio	Lower	Upper
43	100		
70	41.6	35.2	52.8
55	25.6	19.0	28.4
61	25.5	20.0	30.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#75
1,1,2,2-Tetrachloroethane
Concen: 1.147 ug/l
RT: 11.214 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 83 Resp: 4377

Ion	Ratio	Lower	Upper
83	100		
131	11.7	5.0	14.9
85	72.0	32.0	96.2





#76
1,2,3-Trichloropropane
Concen: 1.119 ug/l m
RT: 11.244 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31


Tgt Ion: 75 Resp: 357


Ion	Ratio	Lower	Upper
75	100		
77	56.8	23.4	70.2

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#77
Bromobenzene
Concen: 1.137 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:156 Resp: 2937

Ion	Ratio	Lower	Upper
156	100		
77	170.8	90.3	271.0
158	89.3	48.4	145.3






#78
n-propylbenzene
Concen: 1.088 ug/l
RT: 11.305 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 91 Resp: 1326


Ion	Ratio	Lower	Upper
91	100		
120	21.9	11.4	34.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#79
2-Chlorotoluene
Concen: 1.163 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 91 Resp: 9159

Ion	Ratio	Lower	Upper
91	100		
126	33.2	16.2	48.5






#80
1,3,5-Trimethylbenzene
Concen: 1.115 ug/l
RT: 11.451 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:105 Resp: 1003


Ion	Ratio	Lower	Upper
105	100		
120	49.2	23.7	71.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#82
4-Chlorotoluene
Concen: 1.101 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 91 Resp: 9494

Ion	Ratio	Lower	Upper
91	100		
126	31.3	13.9	41.7






#83
tert-Butylbenzene
Concen: 1.155 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:119 Resp: 1086


Ion	Ratio	Lower	Upper
119	100		
91	63.4	30.0	90.0
134	23.5	11.2	33.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#84
1,2,4-Trimethylbenzene
Concen: 1.174 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:105 Resp: 10735

Ion	Ratio	Lower	Upper
105	100		
120	40.5	22.2	66.6






#85
sec-Butylbenzene
Concen: 1.115 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:105 Resp: 1221


Ion	Ratio	Lower	Upper
105	100		
134	19.3	9.8	29.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#86
p-Isopropyltoluene
Concen: 1.101 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:119 Resp: 9910

Ion	Ratio	Lower	Upper
119	100		
134	26.5	12.8	38.4
91	34.4	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 1.046 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:146 Resp: 468


Ion	Ratio	Lower	Upper
146	100		
111	44.8	21.9	65.7
148	67.2	31.8	95.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#88
1,4-Dichlorobenzene
Concen: 1.163 ug/l m
RT: 12.037 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:146 Resp: 5172

Ion	Ratio	Lower	Upper
146	100		
111	40.6	21.5	64.5
148	60.9	32.1	96.5






#89
n-Butylbenzene
Concen: 1.047 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 91 Resp: 809


Ion	Ratio	Lower	Upper
91	100		
92	50.5	27.4	82.0
134	22.8	12.8	38.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#90
Hexachloroethane
Concen: 1.270 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:117 Resp: 2589

Ion	Ratio	Lower	Upper
117	100		
201	58.4	32.3	96.8






#91
1,2-Dichlorobenzene
Concen: 1.128 ug/l
RT: 12.341 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:146 Resp: 519


Ion	Ratio	Lower	Upper
146	100		
111	47.0	22.4	67.2
148	62.3	31.6	94.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 1.178 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion: 75 Resp: 1100

Ion	Ratio	Lower	Upper
75	100		
155	75.9	39.0	116.9
157	95.6	49.9	149.7






#93
1,2,4-Trichlorobenzene
Concen: 0.941 ug/l
RT: 13.591 min Scan# 2
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:180 Resp: 258


Ion	Ratio	Lower	Upper
180	100		
182	101.8	47.4	142.2
145	38.4	16.7	50.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#94
Hexachlorobutadiene
Concen: 1.049 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:225 Resp: 1133

Ion	Ratio	Lower	Upper
225	100		
223	56.4	31.9	95.7
227	57.7	31.8	95.4






#95
Naphthalene
Concen: 1.033 ug/l
RT: 13.780 min Scan# 2
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:128 Resp: 1095


Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.3	15.5
129	11.1	8.7	13.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#96
1,2,3-Trichlorobenzene
Concen: 1.004 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX044648.D
Acq: 15 Jan 2025 10:31

Tgt Ion:180 Resp: 2863

Ion	Ratio	Lower	Upper
180	100		
182	92.5	47.2	141.6
145	40.9	17.5	52.6



Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044649.D
Acq On : 15 Jan 2025 11:17
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jan 16 00:59:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED
Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	194627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	357706	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	300165	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	123491	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	16795	5.717	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 11.440%#			
35) Dibromofluoromethane	5.379	113	12712	5.594	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 11.180%#			
50) Toluene-d8	8.647	98	45425	5.733	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 11.460%#			
62) 4-Bromofluorobenzene	11.079	95	15909	5.386	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 10.780%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	11790	4.502	ug/l	96
3) Chloromethane	1.301	50	13859	4.893	ug/l	99
4) Vinyl Chloride	1.374	62	13246	4.734	ug/l	97
5) Bromomethane	1.593	94	3303	4.943	ug/l	98
6) Chloroethane	1.660	64	2387	4.652	ug/l #	83
7) Trichlorofluoromethane	1.868	101	14571	4.711	ug/l	95
8) Diethyl Ether	2.136	74	6231	4.992	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	10967	4.717	ug/l	99
10) Methyl Iodide	2.441	142	15920	5.016	ug/l	99
11) Tert butyl alcohol	2.989	59	20543m	31.009	ug/l	
12) 1,1-Dichloroethene	2.307	96	10636	4.667	ug/l	91
13) Acrolein	2.239	56	11045	35.838	ug/l	95
14) Allyl chloride	2.654	41	21748	4.901	ug/l	97
15) Acrylonitrile	3.069	53	32783	25.521	ug/l	98
16) Acetone	2.386	43	26927	23.167	ug/l	97
17) Carbon Disulfide	2.502	76	30278	4.869	ug/l	97
18) Methyl Acetate	2.703	43	16192	4.516	ug/l	95
19) Methyl tert-butyl Ether	3.117	73	41254	4.998	ug/l	100
20) Methylene Chloride	2.782	84	13610	5.028	ug/l	97
21) trans-1,2-Dichloroethene	3.087	96	11813	4.965	ug/l	98
22) Diisopropyl ether	3.764	45	39716	5.149	ug/l	92
23) Vinyl Acetate	3.721	43	175668	26.022	ug/l	99
24) 1,1-Dichloroethane	3.605	63	22655	4.953	ug/l	98
25) 2-Butanone	4.568	43	43210	25.004	ug/l	96
26) 2,2-Dichloropropane	4.465	77	21972	4.957	ug/l	98
27) cis-1,2-Dichloroethene	4.483	96	14297	4.777	ug/l	97
28) Bromochloromethane	4.891	49	10785	5.046	ug/l	100
29) Tetrahydrofuran	5.007	42	27602	24.762	ug/l	98
30) Chloroform	5.086	83	23860	5.169	ug/l	93
31) Cyclohexane	5.458	56	18252	4.934	ug/l	94
32) 1,1,1-Trichloroethane	5.373	97	20169	4.753	ug/l	99
36) 1,1-Dichloropropene	5.690	75	15040	4.754	ug/l	94
37) Ethyl Acetate	4.721	43	17349	5.150	ug/l	95
38) Carbon Tetrachloride	5.666	117	17983	5.054	ug/l	92
39) Methylcyclohexane	7.373	83	19848	4.727	ug/l	95
40) Benzene	6.038	78	49235	5.023	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044649.D
Acq On : 15 Jan 2025 11:17
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jan 16 00:59:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.934	41	9920	5.124	ug/l		98
42) 1,2-Dichloroethane	6.086	62	17678	4.941	ug/l		97
43) Isopropyl Acetate	6.342	43	30226	4.877	ug/l		99
44) Trichloroethene	7.123	130	11427	4.849	ug/l		97
45) 1,2-Dichloropropane	7.428	63	12829	5.199	ug/l		97
46) Dibromomethane	7.580	93	9684	5.156	ug/l		94
47) Bromodichloromethane	7.818	83	18964	4.923	ug/l		94
48) Methyl methacrylate	7.696	41	13933	4.844	ug/l		97
49) 1,4-Dioxane	7.677	88	5674	94.542	ug/l		96
51) 4-Methyl-2-Pentanone	8.574	43	87503	25.551	ug/l		98
52) Toluene	8.714	92	30892	5.149	ug/l		97
53) t-1,3-Dichloropropene	8.976	75	19714	4.954	ug/l		98
54) cis-1,3-Dichloropropene	8.366	75	21054	4.879	ug/l		96
55) 1,1,2-Trichloroethane	9.153	97	12227	5.037	ug/l		99
56) Ethyl methacrylate	9.116	69	20596	4.935	ug/l		96
57) 1,3-Dichloropropane	9.305	76	21253	5.142	ug/l		99
58) 2-Chloroethyl Vinyl ether	8.238	63	50319	24.557	ug/l		100
59) 2-Hexanone	9.433	43	63985	25.395	ug/l		100
60) Dibromochloromethane	9.519	129	14553	5.108	ug/l		99
61) 1,2-Dibromoethane	9.610	107	12916	5.077	ug/l		97
64) Tetrachloroethene	9.269	164	9274	4.920	ug/l		93
65) Chlorobenzene	10.073	112	32572	5.109	ug/l		97
66) 1,1,1,2-Tetrachloroethane	10.159	131	11955	5.160	ug/l		98
67) Ethyl Benzene	10.189	91	55208	4.963	ug/l		98
68) m/p-Xylenes	10.299	106	41816	10.129	ug/l		99
69) o-Xylene	10.640	106	21392	5.124	ug/l		96
70) Styrene	10.653	104	34407	5.105	ug/l		98
71) Bromoform	10.799	173	9452	5.012	ug/l	#	97
73) Isopropylbenzene	10.963	105	52599	5.214	ug/l		100
74) N-amyl acetate	10.842	43	25282	5.084	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.213	83	18645	5.351	ug/l		98
76) 1,2,3-Trichloropropane	11.238	75	16903m	5.796	ug/l		
77) Bromobenzene	11.201	156	12158	5.156	ug/l		96
78) n-propylbenzene	11.305	91	53913	4.846	ug/l		99
79) 2-Chlorotoluene	11.366	91	36602	5.093	ug/l		98
80) 1,3,5-Trimethylbenzene	11.451	105	43488	5.295	ug/l		100
81) trans-1,4-Dichloro-2-b...	11.018	75	6607	5.042	ug/l		95
82) 4-Chlorotoluene	11.451	91	39763	5.051	ug/l		99
83) tert-Butylbenzene	11.713	119	45606	5.311	ug/l		98
84) 1,2,4-Trimethylbenzene	11.750	105	41737	5.001	ug/l		99
85) sec-Butylbenzene	11.890	105	49858	4.987	ug/l		99
86) p-Isopropyltoluene	12.006	119	41214	5.016	ug/l		99
87) 1,3-Dichlorobenzene	11.969	146	20268	4.956	ug/l		100
88) 1,4-Dichlorobenzene	12.043	146	20555	5.062	ug/l		93
89) n-Butylbenzene	12.329	91	33505	4.747	ug/l		98
90) Hexachloroethane	12.536	117	9325	5.012	ug/l		97
91) 1,2-Dichlorobenzene	12.335	146	21690	5.165	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	4384	5.145	ug/l		96
93) 1,2,4-Trichlorobenzene	13.591	180	11649	4.640	ug/l		99
94) Hexachlorobutadiene	13.725	225	4738	4.804	ug/l		99
95) Naphthalene	13.774	128	46418	4.794	ug/l		99
96) 1,2,3-Trichlorobenzene	13.957	180	12258	4.710	ug/l		96

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044649.D
Acq On : 15 Jan 2025 11:17
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Quant Time: Jan 16 00:59:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044649.D
Acq On : 15 Jan 2025 11:17
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1


Quant Time: Jan 16 00:59:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:168 Resp: 19462


Ion	Ratio	Lower	Upper
168	100		
99	65.0	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#2
Dichlorodifluoromethane
Concen: 4.502 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 85 Resp: 11790

Ion	Ratio	Lower	Upper
85	100		
87	31.2	16.8	50.4






#3
Chloromethane
Concen: 4.893 ug/l
RT: 1.301 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 50 Resp: 1385


Ion	Ratio	Lower	Upper
50	100		
52	33.0	27.0	40.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#4
Vinyl Chloride
Concen: 4.734 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 62 Resp: 13246

Ion	Ratio	Lower	Upper
62	100		
64	35.1	26.6	39.8






#5
Bromomethane
Concen: 4.943 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 94 Resp: 330


Ion	Ratio	Lower	Upper
94	100		
96	92.5	72.8	109.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#6
Chloroethane
Concen: 4.652 ug/l
RT: 1.660 min Scan# 94
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 64 Resp: 2387

Ion	Ratio	Lower	Upper
64	100		
66	42.6	26.6	39.8#






#7
Trichlorofluoromethane
Concen: 4.711 ug/l
RT: 1.868 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:101 Resp: 1457


Ion	Ratio	Lower	Upper
101	100		
103	69.2	52.5	78.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#8
Diethyl Ether
Concen: 4.992 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 74 Resp: 6231

Ion	Ratio	Lower	Upper
74	100		
45	93.9	47.1	141.3






#9
1,1,2-Trichlorotrifluoroethane
Concen: 4.717 ug/l
RT: 2.319 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:101 Resp: 1096


Ion	Ratio	Lower	Upper
101	100		
85	44.0	35.3	52.9
151	71.6	58.2	87.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#10
Methyl Iodide
Concen: 5.016 ug/l
RT: 2.441 min Scan# 222
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:142 Resp: 15920

Ion	Ratio	Lower	Upper
142	100		
127	43.8	35.0	52.4
141	13.9	11.7	17.5






#11
Tert butyl alcohol
Concen: 31.009 ug/l m
RT: 2.989 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 59 Resp: 2054


Ion	Ratio	Lower	Upper
59	100		
57	5.9	3.9	5.9#



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#12
1,1-Dichloroethene
Concen: 4.667 ug/l
RT: 2.307 min Scan# 200
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 96 Resp: 10636

Ion	Ratio	Lower	Upper
96	100		
61	176.8	131.1	196.7
98	70.5	51.2	76.8






#13
Acrolein
Concen: 35.838 ug/l
RT: 2.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 56 Resp: 1104


Ion	Ratio	Lower	Upper
56	100		
55	72.7	55.1	82.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#14
Allyl chloride
Concen: 4.901 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 41 Resp: 21748

Ion	Ratio	Lower	Upper
41	100		
39	65.4	54.9	82.3
76	34.4	28.4	42.6






#15
Acrylonitrile
Concen: 25.521 ug/l
RT: 3.069 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 53 Resp: 3278


Ion	Ratio	Lower	Upper
53	100		
52	80.5	66.0	99.0
51	35.8	28.9	43.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#16
Acetone
Concen: 23.167 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 43 Resp: 26927

Ion	Ratio	Lower	Upper
43	100		
58	29.2	24.6	37.0






#17
Carbon Disulfide
Concen: 4.869 ug/l
RT: 2.502 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 76 Resp: 3027


Ion	Ratio	Lower	Upper
76	100		
78	7.7	7.1	10.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#18
Methyl Acetate
Concen: 4.516 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 43 Resp: 16192

Ion	Ratio	Lower	Upper
43	100		
74	27.3	19.8	29.8






#19
Methyl tert-butyl Ether
Concen: 4.998 ug/l
RT: 3.117 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 73 Resp: 4125


Ion	Ratio	Lower	Upper
73	100		
57	21.6	17.4	26.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#20
Methylene Chloride
Concen: 5.028 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 84 Resp: 13610

Ion	Ratio	Lower	Upper
84	100		
49	114.2	94.2	141.4
51	33.5	28.8	43.2
86	60.9	50.3	75.5






#21
trans-1,2-Dichloroethene
Concen: 4.965 ug/l
RT: 3.087 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 96 Resp: 1181


Ion	Ratio	Lower	Upper
96	100		
61	136.3	111.0	166.6
98	62.3	50.2	75.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#22
Diisopropyl ether
Concen: 5.149 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 45 Resp: 39716

Ion	Ratio	Lower	Upper
45	100		
43	85.7	61.0	91.6
87	29.1	22.7	34.1
59	10.7	9.4	14.2






#23
Vinyl Acetate
Concen: 26.022 ug/l
RT: 3.721 min Scan# 4
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 43 Resp: 17566


Ion	Ratio	Lower	Upper
43	100		
86	11.3	8.9	13.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#24
1,1-Dichloroethane
Concen: 4.953 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 63 Resp: 22655

Ion	Ratio	Lower	Upper
63	100		
98	8.0	3.6	10.9
100	4.4	2.5	7.5






#25
2-Butanone
Concen: 25.004 ug/l
RT: 4.568 min Scan# 5
Delta R.T. 0.012 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 43 Resp: 4321


Ion	Ratio	Lower	Upper
43	100		
72	25.0	21.7	32.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#26
2,2-Dichloropropane
Concen: 4.957 ug/l
RT: 4.465 min Scan# 554
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 77 Resp: 21972

Ion	Ratio	Lower	Upper
77	100		
97	22.5	10.9	32.6






#27
cis-1,2-Dichloroethene
Concen: 4.777 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 96 Resp: 1429


Ion	Ratio	Lower	Upper
96	100		
61	149.1	0.0	296.0
98	70.1	0.0	130.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#28
Bromochloromethane
Concen: 5.046 ug/l
RT: 4.891 min Scan# 624
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 49 Resp: 10785

Ion	Ratio	Lower	Upper
49	100		
129	2.4	0.0	4.4
130	79.7	63.4	95.2






#29
Tetrahydrofuran
Concen: 24.762 ug/l
RT: 5.007 min Scan# 64
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 42 Resp: 2760


Ion	Ratio	Lower	Upper
42	100		
72	48.8	38.2	57.4
71	46.1	35.3	52.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#30
Chloroform
Concen: 5.169 ug/l
RT: 5.086 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 83 Resp: 23860

Ion	Ratio	Lower	Upper
83	100		
85	69.4	51.2	76.8






#31
Cyclohexane
Concen: 4.934 ug/l
RT: 5.458 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 56 Resp: 1825


Ion	Ratio	Lower	Upper
56	100		
69	30.5	26.8	40.2
84	83.4	71.0	106.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#32
1,1,1-Trichloroethane
Concen: 4.753 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 97 Resp: 20169

Ion	Ratio	Lower	Upper
97	100		
99	64.4	51.6	77.4
61	45.9	35.6	53.4






#33
1,2-Dichloroethane-d4
Concen: 5.717 ug/l
RT: 5.958 min Scan# 7
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 65 Resp: 1679


Ion	Ratio	Lower	Upper
65	100		
67	56.9	0.0	103.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.751 min Scan# 929
Delta R.T. -0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:114 Resp: 357706

Ion	Ratio	Lower	Upper
114	100		
63	21.0	0.0	43.0
88	17.4	0.0	35.4






#35
Dibromofluoromethane
Concen: 5.594 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:113 Resp: 1271


Ion	Ratio	Lower	Upper
113	100		
111	105.1	81.6	122.4
192	17.9	14.4	21.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#36
1,1-Dichloropropene
Concen: 4.754 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 75 Resp: 15040

Ion	Ratio	Lower	Upper
75	100		
110	34.8	17.9	53.8
77	36.6	24.5	36.7






#37
Ethyl Acetate
Concen: 5.150 ug/l
RT: 4.721 min Scan# 5
Delta R.T. 0.012 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 43 Resp: 1734


Ion	Ratio	Lower	Upper
43	100		
61	16.3	10.9	16.3
70	14.1	10.1	15.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#38
Carbon Tetrachloride
Concen: 5.054 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:117 Resp: 17983

Ion	Ratio	Lower	Upper
117	100		
119	85.8	76.0	114.0
121	31.6	24.6	36.8






#39
Methylcyclohexane
Concen: 4.727 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 83 Resp: 1984


Ion	Ratio	Lower	Upper
83	100		
55	72.0	60.5	90.7
98	44.2	39.0	58.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#40
Benzene
Concen: 5.023 ug/l
RT: 6.038 min Scan# 812
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 78 Resp: 49235

Ion	Ratio	Lower	Upper
78	100		
77	24.2	18.7	28.1






#41
Methacrylonitrile
Concen: 5.124 ug/l
RT: 4.934 min Scan# 6
Delta R.T. 0.018 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 41 Resp: 992


Ion	Ratio	Lower	Upper
41	100		
39	52.9	44.4	66.6
67	74.3	60.6	90.8
52	31.0	24.3	36.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#42
1,2-Dichloroethane
Concen: 4.941 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 62 Resp: 17678

Ion	Ratio	Lower	Upper
62	100		
98	10.9	0.0	19.4






#43
Isopropyl Acetate
Concen: 4.877 ug/l
RT: 6.342 min Scan# 8
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 43 Resp: 3022


Ion	Ratio	Lower	Upper
43	100		
61	21.0	16.6	24.8
87	14.7	11.0	16.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#44
Trichloroethene
Concen: 4.849 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:130 Resp: 11427

Ion	Ratio	Lower	Upper
130	100		
95	100.7	0.0	208.2






#45
1,2-Dichloropropane
Concen: 5.199 ug/l
RT: 7.428 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 63 Resp: 1282


Ion	Ratio	Lower	Upper
63	100		
65	32.0	24.2	36.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#46
Dibromomethane
Concen: 5.156 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 93 Resp: 9684

Ion	Ratio	Lower	Upper
93	100		
95	77.5	66.4	99.6
174	84.6	72.0	108.0






#47
Bromodichloromethane
Concen: 4.923 ug/l
RT: 7.818 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 83 Resp: 1896


Ion	Ratio	Lower	Upper
83	100		
85	57.7	50.1	75.1
127	7.5	6.5	9.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#48
Methyl methacrylate
Concen: 4.844 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 41 Resp: 13933

Ion	Ratio	Lower	Upper
41	100		
69	86.9	66.8	100.2
39	60.6	47.4	71.0






#49
1,4-Dioxane
Concen: 94.542 ug/l
RT: 7.677 min Scan# 1
Delta R.T. 0.012 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 88 Resp: 567


Ion	Ratio	Lower	Upper
88	100		
43	34.3	27.0	40.4
58	74.9	55.8	83.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#50
Toluene-d8
Concen: 5.733 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 98 Resp: 45425

Ion	Ratio	Lower	Upper
98	100		
100	64.1	53.2	79.8






#51
4-Methyl-2-Pentanone
Concen: 25.551 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 43 Resp: 8750


Ion	Ratio	Lower	Upper
43	100		
58	38.4	31.9	47.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#52
Toluene
Concen: 5.149 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 92 Resp: 30892

Ion	Ratio	Lower	Upper
92	100		
91	173.3	135.7	203.5






#53
t-1,3-Dichloropropene
Concen: 4.954 ug/l
RT: 8.976 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 75 Resp: 1971


Ion	Ratio	Lower	Upper
75	100		
77	29.4	24.4	36.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#54
cis-1,3-Dichloropropene
Concen: 4.879 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 75 Resp: 21054

Ion	Ratio	Lower	Upper
75	100		
77	34.6	25.1	37.7
39	52.2	40.3	60.5






#55
1,1,2-Trichloroethane
Concen: 5.037 ug/l
RT: 9.153 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 97 Resp: 1222


Ion	Ratio	Lower	Upper
97	100		
83	84.6	68.8	103.2
85	54.6	43.7	65.5
99	61.7	49.5	74.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#56
Ethyl methacrylate
Concen: 4.935 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 69 Resp: 20596

Ion	Ratio	Lower	Upper
69	100		
41	68.9	53.2	79.8
39	33.8	30.2	45.2






#57
1,3-Dichloropropane
Concen: 5.142 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 76 Resp: 2125


Ion	Ratio	Lower	Upper
76	100		
78	32.2	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#58
2-Chloroethyl Vinyl ether
Concen: 24.557 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 63 Resp: 50319

Ion	Ratio	Lower	Upper
63	100		
106	28.7	23.0	34.6






#59
2-Hexanone
Concen: 25.395 ug/l
RT: 9.433 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 43 Resp: 6398


Ion	Ratio	Lower	Upper
43	100		
58	55.5	27.8	83.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#60
Dibromochloromethane
Concen: 5.108 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:129 Resp: 14553

Ion	Ratio	Lower	Upper
129	100		
127	78.0	38.5	115.5






#61
1,2-Dibromoethane
Concen: 5.077 ug/l
RT: 9.610 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:107 Resp: 1291


Ion	Ratio	Lower	Upper
107	100		
109	97.2	75.3	112.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#62
4-Bromofluorobenzene
Concen: 5.386 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 95 Resp: 15909

Ion	Ratio	Lower	Upper
95	100		
174	64.6	0.0	133.0
176	61.6	0.0	129.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:117 Resp: 30016


Ion	Ratio	Lower	Upper
117	100		
82	58.3	47.6	71.4
119	31.9	25.4	38.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#64
Tetrachloroethene
Concen: 4.920 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:164 Resp: 9274

Ion	Ratio	Lower	Upper
164	100		
166	119.3	102.0	153.0
129	106.0	81.1	121.7
131	90.1	79.0	118.6






#65
Chlorobenzene
Concen: 5.109 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:112 Resp: 3257


Ion	Ratio	Lower	Upper
112	100		
114	29.8	25.3	37.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 5.160 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:131 Resp: 11955

Ion	Ratio	Lower	Upper
131	100		
133	91.8	47.8	143.3
119	67.1	33.4	100.2






#67
Ethyl Benzene
Concen: 4.963 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 91 Resp: 5520


Ion	Ratio	Lower	Upper
91	100		
106	30.0	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#68
m/p-Xylenes
Concen: 10.129 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:106 Resp: 41816

Ion	Ratio	Lower	Upper
106	100		
91	206.3	166.4	249.6






#69
o-Xylene
Concen: 5.124 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:106 Resp: 2139


Ion	Ratio	Lower	Upper
106	100		
91	223.1	108.7	326.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#70
Styrene
Concen: 5.105 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:104 Resp: 34407

Ion	Ratio	Lower	Upper
104	100		
78	52.8	43.0	64.6
103	52.7	43.4	65.2






#71
Bromoform
Concen: 5.012 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:173 Resp: 945


Ion	Ratio	Lower	Upper
173	100		
175	49.3	23.6	70.8
254	0.0	0.1	0.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:152 Resp: 123491

Ion	Ratio	Lower	Upper
152	100		
115	66.3	43.8	131.4
150	158.9	0.0	347.8






#73
Isopropylbenzene
Concen: 5.214 ug/l
RT: 10.963 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:105 Resp: 5259


Ion	Ratio	Lower	Upper
105	100		
120	26.3	13.2	39.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#74
N-amyl acetate
Concen: 5.084 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 43 Resp: 25282

Ion	Ratio	Lower	Upper
43	100		
70	43.3	35.2	52.8
55	25.0	19.0	28.4
61	25.5	20.0	30.0






#75
1,1,2,2-Tetrachloroethane
Concen: 5.351 ug/l
RT: 11.213 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 83 Resp: 1864


Ion	Ratio	Lower	Upper
83	100		
131	10.8	5.0	14.9
85	65.7	32.0	96.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#76
1,2,3-Trichloropropane
Concen: 5.796 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 75 Resp: 16903

Ion	Ratio	Lower	Upper
75	100		
77	44.4	23.4	70.2






#77
Bromobenzene
Concen: 5.156 ug/l
RT: 11.201 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:156 Resp: 1215


Ion	Ratio	Lower	Upper
156	100		
77	172.0	90.3	271.0
158	96.4	48.4	145.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#78
n-propylbenzene
Concen: 4.846 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 91 Resp: 53913

Ion	Ratio	Lower	Upper
91	100		
120	23.3	11.4	34.1






#79
2-Chlorotoluene
Concen: 5.093 ug/l
RT: 11.366 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 91 Resp: 3660


Ion	Ratio	Lower	Upper
91	100		
126	33.5	16.2	48.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#80
1,3,5-Trimethylbenzene
Concen: 5.295 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:105 Resp: 43488

Ion	Ratio	Lower	Upper
105	100		
120	47.3	23.7	71.1





#81
trans-1,4-Dichloro-2-butene
Concen: 5.042 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17


Tgt Ion: 75 Resp: 660


Ion	Ratio	Lower	Upper
75	100		
53	100.4	74.3	111.5
89	46.8	37.4	56.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#82
4-Chlorotoluene
Concen: 5.051 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 91 Resp: 39763

Ion	Ratio	Lower	Upper
91	100		
126	28.1	13.9	41.7






#83
tert-Butylbenzene
Concen: 5.311 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:119 Resp: 4560


Ion	Ratio	Lower	Upper
119	100		
91	58.6	30.0	90.0
134	22.8	11.2	33.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#84
1,2,4-Trimethylbenzene
Concen: 5.001 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:105 Resp: 41737

Ion	Ratio	Lower	Upper
105	100		
120	43.7	22.2	66.6






#85
sec-Butylbenzene
Concen: 4.987 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:105 Resp: 4985


Ion	Ratio	Lower	Upper
105	100		
134	19.4	9.8	29.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#86
p-Isopropyltoluene
Concen: 5.016 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:119 Resp: 41214

Ion	Ratio	Lower	Upper
119	100		
134	25.9	12.8	38.4
91	25.8	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 4.956 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:146 Resp: 2026


Ion	Ratio	Lower	Upper
146	100		
111	43.3	21.9	65.7
148	63.7	31.8	95.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#88
1,4-Dichlorobenzene
Concen: 5.062 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:146 Resp: 20555

Ion	Ratio	Lower	Upper
146	100		
111	51.4	21.5	64.5
148	66.7	32.1	96.5






#89
n-Butylbenzene
Concen: 4.747 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 91 Resp: 3350


Ion	Ratio	Lower	Upper
91	100		
92	53.5	27.4	82.0
134	24.2	12.8	38.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#90
Hexachloroethane
Concen: 5.012 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:117 Resp: 9325

Ion	Ratio	Lower	Upper
117	100		
201	66.5	32.3	96.8





#91
1,2-Dichlorobenzene
Concen: 5.165 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17


Tgt Ion:146 Resp: 2169

Ion	Ratio	Lower	Upper
146	100		
111	44.3	22.4	67.2
148	62.5	31.6	94.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 5.145 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion: 75 Resp: 4384


Ion	Ratio	Lower	Upper
75	100		
155	76.0	39.0	116.9
157	94.7	49.9	149.7



#93
1,2,4-Trichlorobenzene
Concen: 4.640 ug/l
RT: 13.591 min Scan# 2
Delta R.T. 0.006 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:180 Resp: 1164


Ion	Ratio	Lower	Upper
180	100		
182	95.7	47.4	142.2
145	35.1	16.7	50.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#94
Hexachlorobutadiene
Concen: 4.804 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:225 Resp: 4738

Ion	Ratio	Lower	Upper
225	100		
223	62.7	31.9	95.7
227	64.2	31.8	95.4






#95
Naphthalene
Concen: 4.794 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:128 Resp: 4641


Ion	Ratio	Lower	Upper
128	100		
127	13.5	10.3	15.5
129	10.9	8.7	13.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#96
1,2,3-Trichlorobenzene
Concen: 4.710 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX044649.D
Acq: 15 Jan 2025 11:17

Tgt Ion:180 Resp: 12258

Ion	Ratio	Lower	Upper
180	100		
182	90.2	47.2	141.6
145	36.9	17.5	52.6



Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044650.D
Acq On : 15 Jan 2025 11:40
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jan 16 00:59:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	189658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	351528	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	305422	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	130861	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	52421	18.311	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 36.620%#			
35) Dibromofluoromethane	5.379	113	40585	18.173	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 36.340%#			
50) Toluene-d8	8.647	98	147843	18.987	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 37.980%#			
62) 4-Bromofluorobenzene	11.079	95	55407	19.087	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 38.180%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	48623	19.052	ug/l	97
3) Chloromethane	1.295	50	54827	19.866	ug/l	97
4) Vinyl Chloride	1.374	62	51689	18.955	ug/l	98
5) Bromomethane	1.593	94	12588	19.333	ug/l	98
6) Chloroethane	1.660	64	8576	17.150	ug/l	98
7) Trichlorofluoromethane	1.868	101	57313	19.017	ug/l	100
8) Diethyl Ether	2.136	74	23029	18.934	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.319	101	42608	18.807	ug/l	98
10) Methyl Iodide	2.441	142	63102	20.401	ug/l	99
11) Tert butyl alcohol	2.977	59	62614m	96.990	ug/l	
12) 1,1-Dichloroethene	2.307	96	43408	19.547	ug/l	98
13) Acrolein	2.233	56	24912	82.949	ug/l	93
14) Allyl chloride	2.654	41	81766	18.911	ug/l	100
15) Acrylonitrile	3.063	53	123315	98.515	ug/l	98
16) Acetone	2.386	43	103836	91.678	ug/l	98
17) Carbon Disulfide	2.502	76	119145	19.662	ug/l	100
18) Methyl Acetate	2.703	43	62684	17.939	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	158601	19.718	ug/l	98
20) Methylene Chloride	2.782	84	51758	19.623	ug/l	98
21) trans-1,2-Dichloroethene	3.081	96	44786	19.317	ug/l	95
22) Diisopropyl ether	3.758	45	151033	20.095	ug/l	94
23) Vinyl Acetate	3.715	43	666641	101.337	ug/l	100
24) 1,1-Dichloroethane	3.605	63	88060	19.758	ug/l	100
25) 2-Butanone	4.556	43	168309	99.946	ug/l	99
26) 2,2-Dichloropropane	4.465	77	82977	19.209	ug/l	97
27) cis-1,2-Dichloroethene	4.483	96	56594	19.406	ug/l	99
28) Bromochloromethane	4.891	49	39526	18.978	ug/l	98
29) Tetrahydrofuran	5.001	42	106241	97.807	ug/l	100
30) Chloroform	5.087	83	89891	19.985	ug/l	99
31) Cyclohexane	5.458	56	72527	20.120	ug/l	97
32) 1,1,1-Trichloroethane	5.373	97	78644	19.019	ug/l	100
36) 1,1-Dichloropropene	5.684	75	59303	19.076	ug/l	99
37) Ethyl Acetate	4.715	43	66324	20.034	ug/l	98
38) Carbon Tetrachloride	5.666	117	65838	18.829	ug/l	95
39) Methylcyclohexane	7.373	83	78248	18.961	ug/l	97
40) Benzene	6.032	78	191801	19.911	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044650.D
Acq On : 15 Jan 2025 11:40
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jan 16 00:59:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.916	41	36180	19.016	ug/l		98
42) 1,2-Dichloroethane	6.080	62	69034	19.635	ug/l		99
43) Isopropyl Acetate	6.336	43	119568	19.633	ug/l		99
44) Trichloroethene	7.117	130	45597	19.689	ug/l		94
45) 1,2-Dichloropropane	7.428	63	47111	19.429	ug/l		100
46) Dibromomethane	7.580	93	35056	18.991	ug/l		98
47) Bromodichloromethane	7.818	83	72837	19.241	ug/l		94
48) Methyl methacrylate	7.690	41	57413	20.310	ug/l		99
49) 1,4-Dioxane	7.665	88	23433	397.311	ug/l		97
51) 4-Methyl-2-Pentanone	8.574	43	338304	100.520	ug/l		99
52) Toluene	8.714	92	119410	20.253	ug/l		100
53) t-1,3-Dichloropropene	8.976	75	79260	20.266	ug/l		98
54) cis-1,3-Dichloropropene	8.366	75	84851	20.008	ug/l		99
55) 1,1,2-Trichloroethane	9.147	97	47930	20.093	ug/l		97
56) Ethyl methacrylate	9.116	69	84067	20.496	ug/l		98
57) 1,3-Dichloropropane	9.305	76	81718	20.117	ug/l		99
58) 2-Chloroethyl Vinyl ether	8.238	63	176999	87.899	ug/l		99
59) 2-Hexanone	9.427	43	250792	101.288	ug/l		99
60) Dibromochloromethane	9.519	129	56038	20.015	ug/l		99
61) 1,2-Dibromoethane	9.604	107	50060	20.025	ug/l		99
64) Tetrachloroethene	9.269	164	36566	19.066	ug/l		96
65) Chlorobenzene	10.073	112	126540	19.506	ug/l		98
66) 1,1,1,2-Tetrachloroethane	10.159	131	45959	19.493	ug/l		99
67) Ethyl Benzene	10.189	91	223249	19.723	ug/l		99
68) m/p-Xylenes	10.299	106	167689	39.921	ug/l		100
69) o-Xylene	10.640	106	84280	19.840	ug/l		99
70) Styrene	10.653	104	138465	20.189	ug/l		99
71) Bromoform	10.799	173	36839	19.198	ug/l	#	98
73) Isopropylbenzene	10.957	105	211633	19.797	ug/l		99
74) N-amyl acetate	10.842	43	103040	19.555	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.207	83	73190	19.823	ug/l		98
76) 1,2,3-Trichloropropane	11.238	75	61409m	19.870	ug/l		
77) Bromobenzene	11.195	156	48555	19.432	ug/l		100
78) n-propylbenzene	11.299	91	234567	19.896	ug/l		100
79) 2-Chlorotoluene	11.360	91	149688	19.655	ug/l		99
80) 1,3,5-Trimethylbenzene	11.451	105	173348	19.919	ug/l		98
81) trans-1,4-Dichloro-2-b...	11.018	75	27061	19.486	ug/l		98
82) 4-Chlorotoluene	11.451	91	166586	19.968	ug/l		100
83) tert-Butylbenzene	11.713	119	178504	19.617	ug/l		100
84) 1,2,4-Trimethylbenzene	11.750	105	175164	19.807	ug/l		99
85) sec-Butylbenzene	11.890	105	208186	19.649	ug/l		100
86) p-Isopropyltoluene	12.006	119	171507	19.697	ug/l		99
87) 1,3-Dichlorobenzene	11.969	146	84759	19.560	ug/l		100
88) 1,4-Dichlorobenzene	12.037	146	84055	19.536	ug/l		100
89) n-Butylbenzene	12.329	91	145015	19.391	ug/l		99
90) Hexachloroethane	12.536	117	36858	18.693	ug/l		99
91) 1,2-Dichlorobenzene	12.335	146	87147	19.582	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	16770	18.572	ug/l		99
93) 1,2,4-Trichlorobenzene	13.585	180	49471	18.595	ug/l		98
94) Hexachlorobutadiene	13.719	225	19513	18.672	ug/l		99
95) Naphthalene	13.774	128	198769	19.372	ug/l		99
96) 1,2,3-Trichlorobenzene	13.957	180	52651	19.093	ug/l		99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044650.D
Acq On : 15 Jan 2025 11:40
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Quant Time: Jan 16 00:59:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044650.D
Acq On : 15 Jan 2025 11:40
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Jan 16 00:59:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:168 Resp: 18965


Ion	Ratio	Lower	Upper
168	100		
99	67.7	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#2
Dichlorodifluoromethane
Concen: 19.052 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 85 Resp: 48623

Ion	Ratio	Lower	Upper
85	100		
87	32.1	16.8	50.4






#3
Chloromethane
Concen: 19.866 ug/l
RT: 1.295 min Scan# 34
Delta R.T. -0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 50 Resp: 5482


Ion	Ratio	Lower	Upper
50	100		
52	31.9	27.0	40.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#4
Vinyl Chloride
Concen: 18.955 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 62 Resp: 51689

Ion	Ratio	Lower	Upper
62	100		
64	34.3	26.6	39.8





#5
Bromomethane
Concen: 19.333 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40


Tgt Ion: 94 Resp: 1258

Ion	Ratio	Lower	Upper
94	100		
96	89.5	72.8	109.2

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#6
Chloroethane
Concen: 17.150 ug/l
RT: 1.660 min Scan# 94
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 64 Resp: 8576


Ion	Ratio	Lower	Upper
64	100		
66	32.3	26.6	39.8



#7
Trichlorofluoromethane
Concen: 19.017 ug/l
RT: 1.868 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:101 Resp: 5731


Ion	Ratio	Lower	Upper
101	100		
103	65.9	52.5	78.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#8
Diethyl Ether
Concen: 18.934 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 74 Resp: 23029

Ion	Ratio	Lower	Upper
74	100		
45	95.7	47.1	141.3






#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.807 ug/l
RT: 2.319 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:101 Resp: 4260


Ion	Ratio	Lower	Upper
101	100		
85	43.3	35.3	52.9
151	70.4	58.2	87.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#10
Methyl Iodide
Concen: 20.401 ug/l
RT: 2.441 min Scan# 222
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:142 Resp: 63102

Ion	Ratio	Lower	Upper
142	100		
127	44.6	35.0	52.4
141	14.2	11.7	17.5






#11
Tert butyl alcohol
Concen: 96.990 ug/l m
RT: 2.977 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 59 Resp: 6261


Ion	Ratio	Lower	Upper
59	100		
57	6.9	3.9	5.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#12
1,1-Dichloroethene
Concen: 19.547 ug/l
RT: 2.307 min Scan# 200
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 96 Resp: 43408

Ion	Ratio	Lower	Upper
96	100		
61	159.8	131.1	196.7
98	63.9	51.2	76.8






#13
Acrolein
Concen: 82.949 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 56 Resp: 2491


Ion	Ratio	Lower	Upper
56	100		
55	74.5	55.1	82.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#14
Allyl chloride
Concen: 18.911 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 41 Resp: 81766

Ion	Ratio	Lower	Upper
41	100		
39	68.5	54.9	82.3
76	35.8	28.4	42.6






#15
Acrylonitrile
Concen: 98.515 ug/l
RT: 3.063 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 53 Resp: 12331


Ion	Ratio	Lower	Upper
53	100		
52	84.1	66.0	99.0
51	36.7	28.9	43.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#16
Acetone
Concen: 91.678 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 43 Resp: 103836

Ion	Ratio	Lower	Upper
43	100		
58	31.9	24.6	37.0






#17
Carbon Disulfide
Concen: 19.662 ug/l
RT: 2.502 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 76 Resp: 11914


Ion	Ratio	Lower	Upper
76	100		
78	8.8	7.1	10.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#18
Methyl Acetate
Concen: 17.939 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 43 Resp: 62684

Ion	Ratio	Lower	Upper
43	100		
74	24.3	19.8	29.8






#19
Methyl tert-butyl Ether
Concen: 19.718 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 73 Resp: 15860


Ion	Ratio	Lower	Upper
73	100		
57	22.5	17.4	26.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#20
Methylene Chloride
Concen: 19.623 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 84 Resp: 51758

Ion	Ratio	Lower	Upper
84	100		
49	116.0	94.2	141.4
51	37.1	28.8	43.2
86	65.2	50.3	75.5






#21
trans-1,2-Dichloroethene
Concen: 19.317 ug/l
RT: 3.081 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 96 Resp: 4478


Ion	Ratio	Lower	Upper
96	100		
61	146.2	111.0	166.6
98	61.0	50.2	75.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#22
Diisopropyl ether
Concen: 20.095 ug/l
RT: 3.758 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 45 Resp: 151033

Ion	Ratio	Lower	Upper
45	100		
43	83.3	61.0	91.6
87	29.1	22.7	34.1
59	11.7	9.4	14.2





#23
Vinyl Acetate
Concen: 101.337 ug/l
RT: 3.715 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40


Tgt Ion: 43 Resp: 66664


Ion	Ratio	Lower	Upper
43	100		
86	11.3	8.9	13.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#24
1,1-Dichloroethane
Concen: 19.758 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 63 Resp: 88060

Ion	Ratio	Lower	Upper
63	100		
98	7.4	3.6	10.9
100	4.8	2.5	7.5






#25
2-Butanone
Concen: 99.946 ug/l
RT: 4.556 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 43 Resp: 16830


Ion	Ratio	Lower	Upper
43	100		
72	27.5	21.7	32.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#26
2,2-Dichloropropane
Concen: 19.209 ug/l
RT: 4.465 min Scan# 554
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 77 Resp: 82977

Ion	Ratio	Lower	Upper
77	100		
97	23.3	10.9	32.6






#27
cis-1,2-Dichloroethene
Concen: 19.406 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 96 Resp: 5659


Ion	Ratio	Lower	Upper
96	100		
61	145.5	0.0	296.0
98	65.0	0.0	130.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#28
Bromochloromethane
Concen: 18.978 ug/l
RT: 4.891 min Scan# 624
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 49 Resp: 39526

Ion	Ratio	Lower	Upper
49	100		
129	2.1	0.0	4.4
130	80.8	63.4	95.2






#29
Tetrahydrofuran
Concen: 97.807 ug/l
RT: 5.001 min Scan# 64
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 42 Resp: 10624


Ion	Ratio	Lower	Upper
42	100		
72	48.2	38.2	57.4
71	44.0	35.3	52.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#30
Chloroform
Concen: 19.985 ug/l
RT: 5.087 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 83 Resp: 89891

Ion	Ratio	Lower	Upper
83	100		
85	65.1	51.2	76.8






#31
Cyclohexane
Concen: 20.120 ug/l
RT: 5.458 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 56 Resp: 7252


Ion	Ratio	Lower	Upper
56	100		
69	32.5	26.8	40.2
84	85.4	71.0	106.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#32
1,1,1-Trichloroethane
Concen: 19.019 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 97 Resp: 78644

Ion	Ratio	Lower	Upper
97	100		
99	64.7	51.6	77.4
61	44.6	35.6	53.4






#33
1,2-Dichloroethane-d4
Concen: 18.311 ug/l
RT: 5.952 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 65 Resp: 5242


Ion	Ratio	Lower	Upper
65	100		
67	51.5	0.0	103.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:114 Resp: 351528

Ion	Ratio	Lower	Upper
114	100		
63	21.2	0.0	43.0
88	17.4	0.0	35.4






#35
Dibromofluoromethane
Concen: 18.173 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:113 Resp: 4058


Ion	Ratio	Lower	Upper
113	100		
111	104.7	81.6	122.4
192	18.0	14.4	21.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#36
1,1-Dichloropropene
Concen: 19.076 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 75 Resp: 59303

Ion	Ratio	Lower	Upper
75	100		
110	35.8	17.9	53.8
77	29.8	24.5	36.7






#37
Ethyl Acetate
Concen: 20.034 ug/l
RT: 4.715 min Scan# 5
Delta R.T. 0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 43 Resp: 6632


Ion	Ratio	Lower	Upper
43	100		
61	15.0	10.9	16.3
70	12.0	10.1	15.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#38
Carbon Tetrachloride
Concen: 18.829 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:117 Resp: 65838

Ion	Ratio	Lower	Upper
117	100		
119	89.1	76.0	114.0
121	29.7	24.6	36.8






#39
Methylcyclohexane
Concen: 18.961 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 83 Resp: 7824


Ion	Ratio	Lower	Upper
83	100		
55	79.3	60.5	90.7
98	49.2	39.0	58.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#40
Benzene
Concen: 19.911 ug/l
RT: 6.032 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 78 Resp: 191801

Ion	Ratio	Lower	Upper
78	100		
77	23.7	18.7	28.1






#41
Methacrylonitrile
Concen: 19.016 ug/l
RT: 4.916 min Scan# 6
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 41 Resp: 3618


Ion	Ratio	Lower	Upper
41	100		
39	58.8	44.4	66.6
67	75.9	60.6	90.8
52	31.8	24.3	36.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#42
1,2-Dichloroethane
Concen: 19.635 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 62 Resp: 69034

Ion	Ratio	Lower	Upper
62	100		
98	10.2	0.0	19.4






#43
Isopropyl Acetate
Concen: 19.633 ug/l
RT: 6.336 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 43 Resp: 11956


Ion	Ratio	Lower	Upper
43	100		
61	21.1	16.6	24.8
87	13.5	11.0	16.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#44
Trichloroethene
Concen: 19.689 ug/l
RT: 7.117 min Scan# 989
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:130 Resp: 45597

Ion	Ratio	Lower	Upper
130	100		
95	97.7	0.0	208.2






#45
1,2-Dichloropropane
Concen: 19.429 ug/l
RT: 7.428 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 63 Resp: 4711


Ion	Ratio	Lower	Upper
63	100		
65	30.4	24.2	36.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#46
Dibromomethane
Concen: 18.991 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 93 Resp: 35056

Ion	Ratio	Lower	Upper
93	100		
95	82.3	66.4	99.6
174	92.1	72.0	108.0






#47
Bromodichloromethane
Concen: 19.241 ug/l
RT: 7.818 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 83 Resp: 7283


Ion	Ratio	Lower	Upper
83	100		
85	67.6	50.1	75.1
127	7.9	6.5	9.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#48
Methyl methacrylate
Concen: 20.310 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 41 Resp: 57413

Ion	Ratio	Lower	Upper
41	100		
69	84.9	66.8	100.2
39	60.0	47.4	71.0





#49
1,4-Dioxane
Concen: 397.311 ug/l
RT: 7.665 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40


Tgt Ion: 88 Resp: 2343

Ion	Ratio	Lower	Upper
88	100		
43	32.8	27.0	40.4
58	72.7	55.8	83.8

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#50
Toluene-d8
Concen: 18.987 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 98 Resp: 147843


Ion	Ratio	Lower	Upper
98	100		
100	65.5	53.2	79.8



#51
4-Methyl-2-Pentanone
Concen: 100.520 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 43 Resp: 33830


Ion	Ratio	Lower	Upper
43	100		
58	40.4	31.9	47.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#52
Toluene
Concen: 20.253 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 92 Resp: 119410

Ion	Ratio	Lower	Upper
92	100		
91	170.2	135.7	203.5






#53
t-1,3-Dichloropropene
Concen: 20.266 ug/l
RT: 8.976 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 75 Resp: 7926


Ion	Ratio	Lower	Upper
75	100		
77	31.6	24.4	36.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#54
cis-1,3-Dichloropropene
Concen: 20.008 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 75 Resp: 84851

Ion	Ratio	Lower	Upper
75	100		
77	32.4	25.1	37.7
39	50.6	40.3	60.5






#55
1,1,2-Trichloroethane
Concen: 20.093 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 97 Resp: 4793


Ion	Ratio	Lower	Upper
97	100		
83	81.9	68.8	103.2
85	55.0	43.7	65.5
99	63.5	49.5	74.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#56
Ethyl methacrylate
Concen: 20.496 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 69 Resp: 84067

Ion	Ratio	Lower	Upper
69	100		
41	65.2	53.2	79.8
39	36.9	30.2	45.2






#57
1,3-Dichloropropane
Concen: 20.117 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 76 Resp: 8171


Ion	Ratio	Lower	Upper
76	100		
78	31.9	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#58
2-Chloroethyl Vinyl ether
Concen: 87.899 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 63 Resp: 176999

Ion	Ratio	Lower	Upper
63	100		
106	28.4	23.0	34.6





#59
2-Hexanone
Concen: 101.288 ug/l
RT: 9.427 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40


Tgt Ion: 43 Resp: 25079


Ion	Ratio	Lower	Upper
43	100		
58	55.0	27.8	83.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#60
Dibromochloromethane
Concen: 20.015 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:129 Resp: 56038

Ion	Ratio	Lower	Upper
129	100		
127	77.5	38.5	115.5






#61
1,2-Dibromoethane
Concen: 20.025 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:107 Resp: 5006


Ion	Ratio	Lower	Upper
107	100		
109	93.4	75.3	112.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#62
4-Bromofluorobenzene
Concen: 19.087 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 95 Resp: 55407

Ion	Ratio	Lower	Upper
95	100		
174	66.7	0.0	133.0
176	65.3	0.0	129.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:117 Resp: 30542


Ion	Ratio	Lower	Upper
117	100		
82	59.4	47.6	71.4
119	31.9	25.4	38.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#64
Tetrachloroethene
Concen: 19.066 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:164 Resp: 36566

Ion	Ratio	Lower	Upper
164	100		
166	121.3	102.0	153.0
129	99.4	81.1	121.7
131	94.7	79.0	118.6






#65
Chlorobenzene
Concen: 19.506 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:112 Resp: 12654


Ion	Ratio	Lower	Upper
112	100		
114	32.6	25.3	37.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 19.493 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:131 Resp: 45959

Ion	Ratio	Lower	Upper
131	100		
133	96.3	47.8	143.3
119	67.1	33.4	100.2






#67
Ethyl Benzene
Concen: 19.723 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 91 Resp: 22324


Ion	Ratio	Lower	Upper
91	100		
106	31.0	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#68
m/p-Xylenes
Concen: 39.921 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:106 Resp: 167689

Ion	Ratio	Lower	Upper
106	100		
91	207.5	166.4	249.6






#69
o-Xylene
Concen: 19.840 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:106 Resp: 8428


Ion	Ratio	Lower	Upper
106	100		
91	219.4	108.7	326.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#70
Styrene
Concen: 20.189 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:104 Resp: 138465

Ion	Ratio	Lower	Upper
104	100		
78	52.7	43.0	64.6
103	55.0	43.4	65.2






#71
Bromoform
Concen: 19.198 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:173 Resp: 3683


Ion	Ratio	Lower	Upper
173	100		
175	48.8	23.6	70.8
254	0.1	0.1	0.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:152 Resp: 130861

Ion	Ratio	Lower	Upper
152	100		
115	73.2	43.8	131.4
150	161.8	0.0	347.8






#73
Isopropylbenzene
Concen: 19.797 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:105 Resp: 21163


Ion	Ratio	Lower	Upper
105	100		
120	26.1	13.2	39.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#74
N-amyl acetate
Concen: 19.555 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 43 Resp: 103040

Ion	Ratio	Lower	Upper
43	100		
70	44.6	35.2	52.8
55	23.6	19.0	28.4
61	25.3	20.0	30.0






#75
1,1,2,2-Tetrachloroethane
Concen: 19.823 ug/l
RT: 11.207 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 83 Resp: 7319


Ion	Ratio	Lower	Upper
83	100		
131	9.6	5.0	14.9
85	62.5	32.0	96.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#76
1,2,3-Trichloropropane
Concen: 19.870 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 75 Resp: 61409

Ion	Ratio	Lower	Upper
75	100		
77	46.7	23.4	70.2






#77
Bromobenzene
Concen: 19.432 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:156 Resp: 4855


Ion	Ratio	Lower	Upper
156	100		
77	181.1	90.3	271.0
158	97.5	48.4	145.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#78
n-propylbenzene
Concen: 19.896 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 91 Resp: 234567

Ion	Ratio	Lower	Upper
91	100		
120	22.5	11.4	34.1






#79
2-Chlorotoluene
Concen: 19.655 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 91 Resp: 14968


Ion	Ratio	Lower	Upper
91	100		
126	31.8	16.2	48.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#80
1,3,5-Trimethylbenzene
Concen: 19.919 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:105 Resp: 173348

Ion	Ratio	Lower	Upper
105	100		
120	48.5	23.7	71.1






#81
trans-1,4-Dichloro-2-butene
Concen: 19.486 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 75 Resp: 2706


Ion	Ratio	Lower	Upper
75	100		
53	93.5	74.3	111.5
89	48.8	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#82
4-Chlorotoluene
Concen: 19.968 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 91 Resp: 166586

Ion	Ratio	Lower	Upper
91	100		
126	27.8	13.9	41.7






#83
tert-Butylbenzene
Concen: 19.617 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:119 Resp: 17850


Ion	Ratio	Lower	Upper
119	100		
91	59.8	30.0	90.0
134	22.3	11.2	33.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#84
1,2,4-Trimethylbenzene
Concen: 19.807 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:105 Resp: 175164

Ion	Ratio	Lower	Upper
105	100		
120	44.9	22.2	66.6






#85
sec-Butylbenzene
Concen: 19.649 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:105 Resp: 20818


Ion	Ratio	Lower	Upper
105	100		
134	19.5	9.8	29.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#86
p-Isopropyltoluene
Concen: 19.697 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:119 Resp: 171507

Ion	Ratio	Lower	Upper
119	100		
134	26.0	12.8	38.4
91	25.0	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 19.560 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:146 Resp: 8475


Ion	Ratio	Lower	Upper
146	100		
111	43.7	21.9	65.7
148	63.9	31.8	95.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#88
1,4-Dichlorobenzene
Concen: 19.536 ug/l
RT: 12.037 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:146 Resp: 84055

Ion	Ratio	Lower	Upper
146	100		
111	42.8	21.5	64.5
148	64.1	32.1	96.5






#89
n-Butylbenzene
Concen: 19.391 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 91 Resp: 14501


Ion	Ratio	Lower	Upper
91	100		
92	54.4	27.4	82.0
134	25.0	12.8	38.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#90
Hexachloroethane
Concen: 18.693 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:117 Resp: 36858

Ion	Ratio	Lower	Upper
117	100		
201	64.9	32.3	96.8






#91
1,2-Dichlorobenzene
Concen: 19.582 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:146 Resp: 8714


Ion	Ratio	Lower	Upper
146	100		
111	45.1	22.4	67.2
148	64.1	31.6	94.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 18.572 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion: 75 Resp: 16770

Ion	Ratio	Lower	Upper
75	100		
155	79.6	39.0	116.9
157	99.0	49.9	149.7






#93
1,2,4-Trichlorobenzene
Concen: 18.595 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:180 Resp: 4947


Ion	Ratio	Lower	Upper
180	100		
182	97.1	47.4	142.2
145	34.5	16.7	50.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#94
Hexachlorobutadiene
Concen: 18.672 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:225 Resp: 19513

Ion	Ratio	Lower	Upper
225	100		
223	62.8	31.9	95.7
227	64.4	31.8	95.4






#95
Naphthalene
Concen: 19.372 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:128 Resp: 19876


Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.3	15.5
129	11.1	8.7	13.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#96
1,2,3-Trichlorobenzene
Concen: 19.093 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX044650.D
Acq: 15 Jan 2025 11:40

Tgt Ion:180 Resp: 52651

Ion	Ratio	Lower	Upper
180	100		
182	95.9	47.2	141.6
145	35.2	17.5	52.6



Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044651.D
Acq On : 15 Jan 2025 12:02
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jan 16 01:00:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED
Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	191095	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	351701	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	301141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	127832	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	142316	49.339	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 98.680%			
35) Dibromofluoromethane	5.379	113	111739	50.011	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.020%			
50) Toluene-d8	8.647	98	390660	50.145	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.300%			
62) 4-Bromofluorobenzene	11.079	95	148223	51.037	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 102.080%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	129814	50.482	ug/l	100
3) Chloromethane	1.300	50	135661	48.786	ug/l	100
4) Vinyl Chloride	1.374	62	134180	48.836	ug/l	100
5) Bromomethane	1.593	94	32585	49.668	ug/l	100
6) Chloroethane	1.660	64	22356	44.371	ug/l	100
7) Trichlorofluoromethane	1.867	101	146280	48.172	ug/l	100
8) Diethyl Ether	2.136	74	55843	45.569	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	105786	46.343	ug/l	100
10) Methyl Iodide	2.440	142	154347	49.525	ug/l	100
11) Tert butyl alcohol	2.983	59	144614m	222.325	ug/l	
12) 1,1-Dichloroethene	2.306	96	108167	48.342	ug/l	100
13) Acrolein	2.239	56	64703	213.821	ug/l	100
14) Allyl chloride	2.654	41	200446	46.011	ug/l	100
15) Acrylonitrile	3.062	53	298629	236.777	ug/l	100
16) Acetone	2.386	43	255513	223.898	ug/l	100
17) Carbon Disulfide	2.501	76	295098	48.333	ug/l	100
18) Methyl Acetate	2.703	43	168213	47.777	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	388306	47.913	ug/l	100
20) Methylene Chloride	2.782	84	124271	46.761	ug/l	100
21) trans-1,2-Dichloroethene	3.081	96	108489	46.441	ug/l	100
22) Diisopropyl ether	3.757	45	363224	47.964	ug/l	100
23) Vinyl Acetate	3.715	43	1576614	237.862	ug/l	100
24) 1,1-Dichloroethane	3.599	63	215178	47.915	ug/l	100
25) 2-Butanone	4.556	43	407104	239.931	ug/l	100
26) 2,2-Dichloropropane	4.465	77	201388	46.271	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	137091	46.654	ug/l	100
28) Bromochloromethane	4.891	49	101453	48.346	ug/l	100
29) Tetrahydrofuran	5.001	42	256051	233.952	ug/l	100
30) Chloroform	5.086	83	218193	48.144	ug/l	100
31) Cyclohexane	5.458	56	177904	48.981	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	190478	45.718	ug/l	100
36) 1,1-Dichloropropene	5.684	75	143445	46.120	ug/l	100
37) Ethyl Acetate	4.708	43	162853	49.168	ug/l	100
38) Carbon Tetrachloride	5.672	117	161882	46.273	ug/l	100
39) Methylcyclohexane	7.373	83	193759	46.929	ug/l	100
40) Benzene	6.031	78	459433	47.670	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044651.D
Acq On : 15 Jan 2025 12:02
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jan 16 01:00:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
41) Methacrylonitrile	4.916	41	89498	47.017 ug/l	100
42) 1,2-Dichloroethane	6.080	62	167878	47.725 ug/l	100
43) Isopropyl Acetate	6.336	43	286693	47.053 ug/l	100
44) Trichloroethene	7.116	130	111112	47.956 ug/l	100
45) 1,2-Dichloropropane	7.427	63	115621	47.660 ug/l	100
46) Dibromomethane	7.574	93	86455	46.813 ug/l	100
47) Bromodichloromethane	7.818	83	178531	47.138 ug/l	100
48) Methyl methacrylate	7.690	41	140450	49.659 ug/l	100
49) 1,4-Dioxane	7.665	88	55679	943.583 ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	790152	234.662 ug/l	100
52) Toluene	8.714	92	283156	48.001 ug/l	100
53) t-1,3-Dichloropropene	8.976	75	191876	49.036 ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	201804	47.563 ug/l	100
55) 1,1,2-Trichloroethane	9.147	97	112274	47.043 ug/l	100
56) Ethyl methacrylate	9.116	69	198422	48.353 ug/l	100
57) 1,3-Dichloropropane	9.305	76	195182	48.025 ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	479521	238.018 ug/l	100
59) 2-Hexanone	9.433	43	580029	234.142 ug/l	100
60) Dibromochloromethane	9.518	129	133941	47.816 ug/l	100
61) 1,2-Dibromoethane	9.604	107	118382	47.331 ug/l	100
64) Tetrachloroethene	9.269	164	89149	47.145 ug/l	100
65) Chlorobenzene	10.073	112	303658	47.474 ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	110395	47.490 ug/l	100
67) Ethyl Benzene	10.189	91	535714	48.000 ug/l	100
68) m/p-Xylenes	10.299	106	395084	95.393 ug/l	100
69) o-Xylene	10.640	106	200396	47.845 ug/l	100
70) Styrene	10.652	104	327844	48.481 ug/l	100
71) Bromoform	10.799	173	90837	48.011 ug/l	100
73) Isopropylbenzene	10.957	105	500669	47.943 ug/l	100
74) N-amyl acetate	10.841	43	242893	47.188 ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.213	83	166468	46.155 ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	140834m	46.648 ug/l	
77) Bromobenzene	11.195	156	115377	47.268 ug/l	100
78) n-propylbenzene	11.299	91	561396	48.746 ug/l	100
79) 2-Chlorotoluene	11.360	91	353178	47.474 ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	407362	47.918 ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	65476	48.266 ug/l	100
82) 4-Chlorotoluene	11.451	91	393120	48.238 ug/l	100
83) tert-Butylbenzene	11.713	119	422348	47.513 ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	410477	47.515 ug/l	100
85) sec-Butylbenzene	11.890	105	497964	48.113 ug/l	100
86) p-Isopropyltoluene	12.006	119	411853	48.421 ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	206233	48.720 ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	201574	47.960 ug/l	100
89) n-Butylbenzene	12.329	91	358698	49.100 ug/l	100
90) Hexachloroethane	12.536	117	88230	45.808 ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	207903	47.824 ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	39659	44.962 ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	131779	50.708 ug/l	100
94) Hexachlorobutadiene	13.725	225	50255	49.230 ug/l	100
95) Naphthalene	13.774	128	487525	48.639 ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	134816	50.047 ug/l	100

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044651.D
Acq On : 15 Jan 2025 12:02
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Quant Time: Jan 16 01:00:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044651.D
Acq On : 15 Jan 2025 12:02
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1


Quant Time: Jan 16 01:00:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:168 Resp: 19109


Ion	Ratio	Lower	Upper
168	100		
99	63.9	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#2
Dichlorodifluoromethane
Concen: 50.482 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 85 Resp: 129814

Ion	Ratio	Lower	Upper
85	100		
87	33.6	16.8	50.4






#3
Chloromethane
Concen: 48.786 ug/l
RT: 1.300 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 50 Resp: 13566


Ion	Ratio	Lower	Upper
50	100		
52	33.7	27.0	40.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#4
Vinyl Chloride
Concen: 48.836 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 62 Resp: 134180

Ion	Ratio	Lower	Upper
62	100		
64	33.2	26.6	39.8






#5
Bromomethane
Concen: 49.668 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 94 Resp: 3258


Ion	Ratio	Lower	Upper
94	100		
96	91.0	72.8	109.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#6
Chloroethane
Concen: 44.371 ug/l
RT: 1.660 min Scan# 94
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 64 Resp: 22356

Ion	Ratio	Lower	Upper
64	100		
66	33.2	26.6	39.8






#7
Trichlorofluoromethane
Concen: 48.172 ug/l
RT: 1.867 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:101 Resp: 14628


Ion	Ratio	Lower	Upper
101	100		
103	65.6	52.5	78.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#8
Diethyl Ether
Concen: 45.569 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 74 Resp: 55843

Ion	Ratio	Lower	Upper
74	100		
45	94.2	47.1	141.3






#9
1,1,2-Trichlorotrifluoroethane
Concen: 46.343 ug/l
RT: 2.319 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:101 Resp: 10578


Ion	Ratio	Lower	Upper
101	100		
85	44.1	35.3	52.9
151	72.8	58.2	87.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#10
Methyl Iodide
Concen: 49.525 ug/l
RT: 2.440 min Scan# 222
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:142 Resp: 154347

Ion	Ratio	Lower	Upper
142	100		
127	43.7	35.0	52.4
141	14.6	11.7	17.5






#11
Tert butyl alcohol
Concen: 222.325 ug/l m
RT: 2.983 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 59 Resp: 14461


Ion	Ratio	Lower	Upper
59	100		
57	4.9	3.9	5.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#12
1,1-Dichloroethene
Concen: 48.342 ug/l
RT: 2.306 min Scan# 200
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 96 Resp: 108167

Ion	Ratio	Lower	Upper
96	100		
61	163.9	131.1	196.7
98	64.0	51.2	76.8






#13
Acrolein
Concen: 213.821 ug/l
RT: 2.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 56 Resp: 6470


Ion	Ratio	Lower	Upper
56	100		
55	68.9	55.1	82.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#14
Allyl chloride
Concen: 46.011 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 41 Resp: 200446

Ion	Ratio	Lower	Upper
41	100		
39	68.6	54.9	82.3
76	35.5	28.4	42.6






#15
Acrylonitrile
Concen: 236.777 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 53 Resp: 29862


Ion	Ratio	Lower	Upper
53	100		
52	82.5	66.0	99.0
51	36.1	28.9	43.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#16
Acetone
Concen: 223.898 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 43 Resp: 255513

Ion	Ratio	Lower	Upper
43	100		
58	30.8	24.6	37.0






#17
Carbon Disulfide
Concen: 48.333 ug/l
RT: 2.501 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 76 Resp: 29509


Ion	Ratio	Lower	Upper
76	100		
78	8.9	7.1	10.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#18
Methyl Acetate
Concen: 47.777 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 43 Resp: 168213

Ion	Ratio	Lower	Upper
43	100		
74	24.8	19.8	29.8






#19
Methyl tert-butyl Ether
Concen: 47.913 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 73 Resp: 38830


Ion	Ratio	Lower	Upper
73	100		
57	21.7	17.4	26.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#20
Methylene Chloride
Concen: 46.761 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 84 Resp: 124271

Ion	Ratio	Lower	Upper
84	100		
49	117.8	94.2	141.4
51	36.0	28.8	43.2
86	62.9	50.3	75.5






#21
trans-1,2-Dichloroethene
Concen: 46.441 ug/l
RT: 3.081 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 96 Resp: 10848


Ion	Ratio	Lower	Upper
96	100		
61	138.8	111.0	166.6
98	62.7	50.2	75.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#22
Diisopropyl ether
Concen: 47.964 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 45 Resp: 363224

Ion	Ratio	Lower	Upper
45	100		
43	76.3	61.0	91.6
87	28.4	22.7	34.1
59	11.8	9.4	14.2





#23
Vinyl Acetate
Concen: 237.862 ug/l
RT: 3.715 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02


Tgt Ion: 43 Resp: 157661


Ion	Ratio	Lower	Upper
43	100		
86	11.1	8.9	13.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#24
1,1-Dichloroethane
Concen: 47.915 ug/l
RT: 3.599 min Scan# 412
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 63 Resp: 215178

Ion	Ratio	Lower	Upper
63	100		
98	7.3	3.6	10.9
100	5.0	2.5	7.5






#25
2-Butanone
Concen: 239.931 ug/l
RT: 4.556 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 43 Resp: 40710


Ion	Ratio	Lower	Upper
43	100		
72	27.1	21.7	32.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#26
2,2-Dichloropropane
Concen: 46.271 ug/l
RT: 4.465 min Scan# 554
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 77 Resp: 201388

Ion	Ratio	Lower	Upper
77	100		
97	21.7	10.9	32.6






#27
cis-1,2-Dichloroethene
Concen: 46.654 ug/l
RT: 4.483 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 96 Resp: 13709


Ion	Ratio	Lower	Upper
96	100		
61	148.0	0.0	296.0
98	65.0	0.0	130.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#28
Bromochloromethane
Concen: 48.346 ug/l
RT: 4.891 min Scan# 624
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 49 Resp: 101453

Ion	Ratio	Lower	Upper
49	100		
129	2.2	0.0	4.4
130	79.3	63.4	95.2






#29
Tetrahydrofuran
Concen: 233.952 ug/l
RT: 5.001 min Scan# 64
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 42 Resp: 25605


Ion	Ratio	Lower	Upper
42	100		
72	47.8	38.2	57.4
71	44.1	35.3	52.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#30
Chloroform
Concen: 48.144 ug/l
RT: 5.086 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 83 Resp: 218193

Ion	Ratio	Lower	Upper
83	100		
85	64.0	51.2	76.8






#31
Cyclohexane
Concen: 48.981 ug/l
RT: 5.458 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 56 Resp: 17790


Ion	Ratio	Lower	Upper
56	100		
69	33.5	26.8	40.2
84	88.7	71.0	106.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#32
1,1,1-Trichloroethane
Concen: 45.718 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 97 Resp: 190478

Ion	Ratio	Lower	Upper
97	100		
99	64.5	51.6	77.4
61	44.5	35.6	53.4






#33
1,2-Dichloroethane-d4
Concen: 49.339 ug/l
RT: 5.952 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 65 Resp: 14231


Ion	Ratio	Lower	Upper
65	100		
67	51.7	0.0	103.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:114 Resp: 351701

Ion	Ratio	Lower	Upper
114	100		
63	21.5	0.0	43.0
88	17.7	0.0	35.4






#35
Dibromofluoromethane
Concen: 50.011 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:113 Resp: 11173


Ion	Ratio	Lower	Upper
113	100		
111	102.0	81.6	122.4
192	18.0	14.4	21.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#36
1,1-Dichloropropene
Concen: 46.120 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 75 Resp: 143445

Ion	Ratio	Lower	Upper
75	100		
110	35.9	17.9	53.8
77	30.6	24.5	36.7






#37
Ethyl Acetate
Concen: 49.168 ug/l
RT: 4.708 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 43 Resp: 16285


Ion	Ratio	Lower	Upper
43	100		
61	13.6	10.9	16.3
70	12.6	10.1	15.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#38
Carbon Tetrachloride
Concen: 46.273 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:117 Resp: 161882

Ion	Ratio	Lower	Upper
117	100		
119	95.0	76.0	114.0
121	30.7	24.6	36.8






#39
Methylcyclohexane
Concen: 46.929 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 83 Resp: 19375


Ion	Ratio	Lower	Upper
83	100		
55	75.6	60.5	90.7
98	48.7	39.0	58.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#40
Benzene
Concen: 47.670 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 78 Resp: 459433

Ion	Ratio	Lower	Upper
78	100		
77	23.4	18.7	28.1






#41
Methacrylonitrile
Concen: 47.017 ug/l
RT: 4.916 min Scan# 6
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 41 Resp: 8949


Ion	Ratio	Lower	Upper
41	100		
39	55.5	44.4	66.6
67	75.7	60.6	90.8
52	30.4	24.3	36.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#42
1,2-Dichloroethane
Concen: 47.725 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 62 Resp: 167878

Ion	Ratio	Lower	Upper
62	100		
98	9.7	0.0	19.4





#43
Isopropyl Acetate
Concen: 47.053 ug/l
RT: 6.336 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02


Tgt Ion: 43 Resp: 28669


Ion	Ratio	Lower	Upper
43	100		
61	20.7	16.6	24.8
87	13.8	11.0	16.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#44
Trichloroethene
Concen: 47.956 ug/l
RT: 7.116 min Scan# 989
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:130 Resp: 111112

Ion	Ratio	Lower	Upper
130	100		
95	104.1	0.0	208.2






#45
1,2-Dichloropropane
Concen: 47.660 ug/l
RT: 7.427 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 63 Resp: 11562


Ion	Ratio	Lower	Upper
63	100		
65	30.2	24.2	36.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#46
Dibromomethane
Concen: 46.813 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 93 Resp: 86455

Ion	Ratio	Lower	Upper
93	100		
95	83.0	66.4	99.6
174	90.0	72.0	108.0






#47
Bromodichloromethane
Concen: 47.138 ug/l
RT: 7.818 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 83 Resp: 17853


Ion	Ratio	Lower	Upper
83	100		
85	62.6	50.1	75.1
127	8.1	6.5	9.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#48
Methyl methacrylate
Concen: 49.659 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 41 Resp: 140450

Ion	Ratio	Lower	Upper
41	100		
69	83.5	66.8	100.2
39	59.2	47.4	71.0






#49
1,4-Dioxane
Concen: 943.583 ug/l
RT: 7.665 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 88 Resp: 5567


Ion	Ratio	Lower	Upper
88	100		
43	33.7	27.0	40.4
58	69.8	55.8	83.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#50
Toluene-d8
Concen: 50.145 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 98 Resp: 390660

Ion	Ratio	Lower	Upper
98	100		
100	66.5	53.2	79.8






#51
4-Methyl-2-Pentanone
Concen: 234.662 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 43 Resp: 79015


Ion	Ratio	Lower	Upper
43	100		
58	39.9	31.9	47.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#52
Toluene
Concen: 48.001 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 92 Resp: 283156

Ion	Ratio	Lower	Upper
92	100		
91	169.6	135.7	203.5






#53
t-1,3-Dichloropropene
Concen: 49.036 ug/l
RT: 8.976 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 75 Resp: 19187


Ion	Ratio	Lower	Upper
75	100		
77	30.5	24.4	36.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#54
cis-1,3-Dichloropropene
Concen: 47.563 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 75 Resp: 201804

Ion	Ratio	Lower	Upper
75	100		
77	31.4	25.1	37.7
39	50.4	40.3	60.5






#55
1,1,2-Trichloroethane
Concen: 47.043 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 97 Resp: 11227


Ion	Ratio	Lower	Upper
97	100		
83	86.0	68.8	103.2
85	54.6	43.7	65.5
99	61.9	49.5	74.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#56
Ethyl methacrylate
Concen: 48.353 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 69 Resp: 198422

Ion	Ratio	Lower	Upper
69	100		
41	66.5	53.2	79.8
39	37.7	30.2	45.2






#57
1,3-Dichloropropane
Concen: 48.025 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 76 Resp: 19518


Ion	Ratio	Lower	Upper
76	100		
78	32.5	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#58
2-Chloroethyl Vinyl ether
Concen: 238.018 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 63 Resp: 479521

Ion	Ratio	Lower	Upper
63	100		
106	28.8	23.0	34.6






#59
2-Hexanone
Concen: 234.142 ug/l
RT: 9.433 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 43 Resp: 58002


Ion	Ratio	Lower	Upper
43	100		
58	55.5	27.8	83.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#60
Dibromochloromethane
Concen: 47.816 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:129 Resp: 133941

Ion	Ratio	Lower	Upper
129	100		
127	77.0	38.5	115.5






#61
1,2-Dibromoethane
Concen: 47.331 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:107 Resp: 11838


Ion	Ratio	Lower	Upper
107	100		
109	94.1	75.3	112.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#62
4-Bromofluorobenzene
Concen: 51.037 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 95 Resp: 148223

Ion	Ratio	Lower	Upper
95	100		
174	66.5	0.0	133.0
176	64.7	0.0	129.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:117 Resp: 30114


Ion	Ratio	Lower	Upper
117	100		
82	59.5	47.6	71.4
119	31.8	25.4	38.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#64
Tetrachloroethene
Concen: 47.145 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:164 Resp: 89149

Ion	Ratio	Lower	Upper
164	100		
166	127.5	102.0	153.0
129	101.4	81.1	121.7
131	98.8	79.0	118.6






#65
Chlorobenzene
Concen: 47.474 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:112 Resp: 30365


Ion	Ratio	Lower	Upper
112	100		
114	31.6	25.3	37.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 47.490 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:131 Resp: 110395

Ion	Ratio	Lower	Upper
131	100		
133	95.5	47.8	143.3
119	66.8	33.4	100.2






#67
Ethyl Benzene
Concen: 48.000 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 91 Resp: 53571


Ion	Ratio	Lower	Upper
91	100		
106	31.3	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#68
m/p-Xylenes
Concen: 95.393 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:106 Resp: 395084

Ion	Ratio	Lower	Upper
106	100		
91	208.0	166.4	249.6





#69
o-Xylene
Concen: 47.845 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02


Tgt Ion:106 Resp: 20039

Ion	Ratio	Lower	Upper
106	100		
91	217.5	108.7	326.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#70
Styrene
Concen: 48.481 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:104 Resp: 327844


Ion	Ratio	Lower	Upper
104	100		
78	53.8	43.0	64.6
103	54.3	43.4	65.2



#71
Bromoform
Concen: 48.011 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:173 Resp: 9083


Ion	Ratio	Lower	Upper
173	100		
175	47.2	23.6	70.8
254	0.1	0.1	0.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:152 Resp: 127832

Ion	Ratio	Lower	Upper
152	100		
115	87.6	43.8	131.4
150	173.9	0.0	347.8






#73
Isopropylbenzene
Concen: 47.943 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:105 Resp: 50066


Ion	Ratio	Lower	Upper
105	100		
120	26.4	13.2	39.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#74
N-amyl acetate
Concen: 47.188 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 43 Resp: 242893

Ion	Ratio	Lower	Upper
43	100		
70	44.0	35.2	52.8
55	23.7	19.0	28.4
61	25.0	20.0	30.0






#75
1,1,2,2-Tetrachloroethane
Concen: 46.155 ug/l
RT: 11.213 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 83 Resp: 16646


Ion	Ratio	Lower	Upper
83	100		
131	9.9	5.0	14.9
85	64.1	32.0	96.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#76
1,2,3-Trichloropropane
Concen: 46.648 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 75 Resp: 140834

Ion	Ratio	Lower	Upper
75	100		
77	46.8	23.4	70.2






#77
Bromobenzene
Concen: 47.268 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:156 Resp: 11537


Ion	Ratio	Lower	Upper
156	100		
77	180.7	90.3	271.0
158	96.9	48.4	145.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#78
n-propylbenzene
Concen: 48.746 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 91 Resp: 561396

Ion	Ratio	Lower	Upper
91	100		
120	22.7	11.4	34.1






#79
2-Chlorotoluene
Concen: 47.474 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 91 Resp: 35317


Ion	Ratio	Lower	Upper
91	100		
126	32.3	16.2	48.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#80
1,3,5-Trimethylbenzene
Concen: 47.918 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:105 Resp: 407362

Ion	Ratio	Lower	Upper
105	100		
120	47.4	23.7	71.1






#81
trans-1,4-Dichloro-2-butene
Concen: 48.266 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 75 Resp: 6547


Ion	Ratio	Lower	Upper
75	100		
53	92.9	74.3	111.5
89	46.7	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#82
4-Chlorotoluene
Concen: 48.238 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 91 Resp: 393120

Ion	Ratio	Lower	Upper
91	100		
126	27.8	13.9	41.7






#83
tert-Butylbenzene
Concen: 47.513 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:119 Resp: 42234


Ion	Ratio	Lower	Upper
119	100		
91	60.0	30.0	90.0
134	22.3	11.2	33.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#84
1,2,4-Trimethylbenzene
Concen: 47.515 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:105 Resp: 410477

Ion	Ratio	Lower	Upper
105	100		
120	44.4	22.2	66.6






#85
sec-Butylbenzene
Concen: 48.113 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:105 Resp: 49796


Ion	Ratio	Lower	Upper
105	100		
134	19.7	9.8	29.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#86
p-Isopropyltoluene
Concen: 48.421 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:119 Resp: 411853

Ion	Ratio	Lower	Upper
119	100		
134	25.6	12.8	38.4
91	25.4	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 48.720 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:146 Resp: 20623


Ion	Ratio	Lower	Upper
146	100		
111	43.8	21.9	65.7
148	63.6	31.8	95.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#88
1,4-Dichlorobenzene
Concen: 47.960 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:146 Resp: 201574

Ion	Ratio	Lower	Upper
146	100		
111	43.0	21.5	64.5
148	64.3	32.1	96.5






#89
n-Butylbenzene
Concen: 49.100 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 91 Resp: 35869


Ion	Ratio	Lower	Upper
91	100		
92	54.7	27.4	82.0
134	25.5	12.8	38.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#90
Hexachloroethane
Concen: 45.808 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:117 Resp: 88230

Ion	Ratio	Lower	Upper
117	100		
201	64.5	32.3	96.8






#91
1,2-Dichlorobenzene
Concen: 47.824 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:146 Resp: 20790


Ion	Ratio	Lower	Upper
146	100		
111	44.8	22.4	67.2
148	63.2	31.6	94.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 44.962 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion: 75 Resp: 39659

Ion	Ratio	Lower	Upper
75	100		
155	77.9	39.0	116.9
157	99.8	49.9	149.7






#93
1,2,4-Trichlorobenzene
Concen: 50.708 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:180 Resp: 13177


Ion	Ratio	Lower	Upper
180	100		
182	94.8	47.4	142.2
145	33.4	16.7	50.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#94
Hexachlorobutadiene
Concen: 49.230 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:225 Resp: 50255

Ion	Ratio	Lower	Upper
225	100		
223	63.8	31.9	95.7
227	63.6	31.8	95.4






#95
Naphthalene
Concen: 48.639 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:128 Resp: 48752


Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.3	15.5
129	10.9	8.7	13.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#96
1,2,3-Trichlorobenzene
Concen: 50.047 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX044651.D
Acq: 15 Jan 2025 12:02

Tgt Ion:180 Resp: 134816

Ion	Ratio	Lower	Upper
180	100		
182	94.4	47.2	141.6
145	35.1	17.5	52.6



Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044652.D
Acq On : 15 Jan 2025 13:14
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jan 16 01:01:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.538	168	170573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	308450	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	255886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	110019	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	235193	91.349	ug/l	-0.01
Spiked Amount 50.000	Range 74 - 125		Recovery = 182.700%#			
35) Dibromofluoromethane	5.373	113	181010	92.374	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 184.740%#			
50) Toluene-d8	8.641	98	617688	90.404	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 180.800%#			
62) 4-Bromofluorobenzene	11.079	95	236119	92.702	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 185.400%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	256287	111.657	ug/l	99
3) Chloromethane	1.301	50	232745	93.768	ug/l	100
4) Vinyl Chloride	1.374	62	238376	97.198	ug/l	100
5) Bromomethane	1.593	94	57815	98.728	ug/l	94
6) Chloroethane	1.660	64	42244	93.931	ug/l	97
7) Trichlorofluoromethane	1.861	101	275756	101.736	ug/l	99
8) Diethyl Ether	2.130	74	99765	91.205	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.313	101	203976	100.109	ug/l	99
10) Methyl Iodide	2.441	142	272297	97.884	ug/l	99
11) Tert butyl alcohol	3.026	59	259620m	447.152	ug/l	
12) 1,1-Dichloroethene	2.306	96	194870	97.570	ug/l	98
13) Acrolein	2.233	56	116830	432.534	ug/l	98
14) Allyl chloride	2.654	41	358388	92.162	ug/l	100
15) Acrylonitrile	3.062	53	560841	498.179	ug/l	100
16) Acetone	2.386	43	527741	518.081	ug/l	100
17) Carbon Disulfide	2.495	76	529418	97.144	ug/l	99
18) Methyl Acetate	2.697	43	321291	102.235	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	681776	94.245	ug/l	99
20) Methylene Chloride	2.776	84	216105	91.100	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	190588	91.400	ug/l	99
22) Diisopropyl ether	3.751	45	632562	93.579	ug/l	100
23) Vinyl Acetate	3.715	43	2791754	471.863	ug/l	100
24) 1,1-Dichloroethane	3.599	63	378924	94.530	ug/l	99
25) 2-Butanone	4.550	43	765500	505.435	ug/l	100
26) 2,2-Dichloropropane	4.458	77	357666	92.064	ug/l	99
27) cis-1,2-Dichloroethene	4.471	96	240149	91.559	ug/l	99
28) Bromochloromethane	4.885	49	179896	96.041	ug/l	100
29) Tetrahydrofuran	4.995	42	477227	488.500	ug/l	100
30) Chloroform	5.080	83	373681	92.373	ug/l	99
31) Cyclohexane	5.452	56	319131	98.436	ug/l	98
32) 1,1,1-Trichloroethane	5.367	97	338119	90.918	ug/l	99
36) 1,1-Dichloropropene	5.678	75	255714	93.746	ug/l	99
37) Ethyl Acetate	4.702	43	300700	103.517	ug/l	99
38) Carbon Tetrachloride	5.666	117	287048	93.555	ug/l	99
39) Methylcyclohexane	7.367	83	359758	99.352	ug/l	98
40) Benzene	6.025	78	785626	92.946	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044652.D
Acq On : 15 Jan 2025 13:14
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jan 16 01:01:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.910	41	168481	100.920	ug/l		98
42) 1,2-Dichloroethane	6.074	62	292036	94.663	ug/l		99
43) Isopropyl Acetate	6.330	43	512598	95.925	ug/l		99
44) Trichloroethene	7.117	130	199570	98.212	ug/l		96
45) 1,2-Dichloropropane	7.421	63	200877	94.413	ug/l		99
46) Dibromomethane	7.568	93	150112	92.680	ug/l		98
47) Bromodichloromethane	7.812	83	307990	92.723	ug/l		99
48) Methyl methacrylate	7.684	41	248667	100.250	ug/l		100
49) 1,4-Dioxane	7.665	88	101384	1959.056	ug/l		99
51) 4-Methyl-2-Pentanone	8.574	43	1372888	464.897	ug/l		100
52) Toluene	8.708	92	474392	91.697	ug/l		100
53) t-1,3-Dichloropropene	8.970	75	330982	96.446	ug/l		98
54) cis-1,3-Dichloropropene	8.360	75	349516	93.927	ug/l		98
55) 1,1,2-Trichloroethane	9.147	97	192326	91.885	ug/l		98
56) Ethyl methacrylate	9.110	69	337626	93.812	ug/l		99
57) 1,3-Dichloropropane	9.305	76	325142	91.220	ug/l		100
58) 2-Chloroethyl Vinyl ether	8.238	63	817528	462.693	ug/l		99
59) 2-Hexanone	9.427	43	1019940	469.454	ug/l		99
60) Dibromochloromethane	9.519	129	227418	92.570	ug/l		99
61) 1,2-Dibromoethane	9.604	107	202247	92.200	ug/l		100
64) Tetrachloroethene	9.269	164	157748	98.176	ug/l		98
65) Chlorobenzene	10.073	112	510916	94.003	ug/l		99
66) 1,1,1,2-Tetrachloroethane	10.159	131	184816	93.565	ug/l		100
67) Ethyl Benzene	10.189	91	909208	95.872	ug/l		99
68) m/p-Xylenes	10.299	106	661488	187.963	ug/l		100
69) o-Xylene	10.640	106	329336	92.536	ug/l		99
70) Styrene	10.653	104	547016	95.197	ug/l		100
71) Bromoform	10.799	173	155379	96.648	ug/l	#	99
73) Isopropylbenzene	10.957	105	833874	92.779	ug/l		100
74) N-amyl acetate	10.842	43	414046	93.463	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.213	83	278679	89.776	ug/l		100
76) 1,2,3-Trichloropropane	11.238	75	250918m	96.568	ug/l		
77) Bromobenzene	11.195	156	192700	91.729	ug/l		99
78) n-propylbenzene	11.299	91	956202	96.470	ug/l		100
79) 2-Chlorotoluene	11.360	91	584682	91.317	ug/l		100
80) 1,3,5-Trimethylbenzene	11.451	105	672586	91.926	ug/l		100
81) trans-1,4-Dichloro-2-b...	11.018	75	114832	98.354	ug/l		99
82) 4-Chlorotoluene	11.451	91	655250	93.421	ug/l		100
83) tert-Butylbenzene	11.713	119	690320	90.233	ug/l		100
84) 1,2,4-Trimethylbenzene	11.750	105	681398	91.646	ug/l		100
85) sec-Butylbenzene	11.890	105	848327	95.235	ug/l		100
86) p-Isopropyltoluene	12.006	119	695433	94.999	ug/l		99
87) 1,3-Dichlorobenzene	11.969	146	353809	97.115	ug/l		99
88) 1,4-Dichlorobenzene	12.036	146	347517	96.070	ug/l		99
89) n-Butylbenzene	12.329	91	633246	100.715	ug/l		99
90) Hexachloroethane	12.536	117	149253	90.037	ug/l		97
91) 1,2-Dichlorobenzene	12.335	146	340683	91.056	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	70862	93.344	ug/l		99
93) 1,2,4-Trichlorobenzene	13.585	180	233417	104.359	ug/l		99
94) Hexachlorobutadiene	13.725	225	88214	100.406	ug/l		100
95) Naphthalene	13.774	128	849403	98.464	ug/l		100
96) 1,2,3-Trichlorobenzene	13.957	180	232152	100.133	ug/l		99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044652.D
Acq On : 15 Jan 2025 13:14
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Quant Time: Jan 16 01:01:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044652.D
Acq On : 15 Jan 2025 13:14
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1


Quant Time: Jan 16 01:01:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:168 Resp: 17057


Ion	Ratio	Lower	Upper
168	100		
99	64.8	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#2
Dichlorodifluoromethane
Concen: 111.657 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 85 Resp: 256287

Ion	Ratio	Lower	Upper
85	100		
87	32.8	16.8	50.4






#3
Chloromethane
Concen: 93.768 ug/l
RT: 1.301 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 50 Resp: 23274


Ion	Ratio	Lower	Upper
50	100		
52	33.5	27.0	40.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#4
Vinyl Chloride
Concen: 97.198 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 62 Resp: 238376

Ion	Ratio	Lower	Upper
62	100		
64	33.3	26.6	39.8






#5
Bromomethane
Concen: 98.728 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 94 Resp: 5781


Ion	Ratio	Lower	Upper
94	100		
96	97.0	72.8	109.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#6
Chloroethane
Concen: 93.931 ug/l
RT: 1.660 min Scan# 94
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 64 Resp: 42244

Ion	Ratio	Lower	Upper
64	100		
66	34.8	26.6	39.8






#7
Trichlorofluoromethane
Concen: 101.736 ug/l
RT: 1.861 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:101 Resp: 27575


Ion	Ratio	Lower	Upper
101	100		
103	64.5	52.5	78.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#8
Diethyl Ether
Concen: 91.205 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 74 Resp: 99765

Ion	Ratio	Lower	Upper
74	100		
45	97.0	47.1	141.3






#9
1,1,2-Trichlorotrifluoroethane
Concen: 100.109 ug/l
RT: 2.313 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:101 Resp: 20397


Ion	Ratio	Lower	Upper
101	100		
85	43.8	35.3	52.9
151	72.1	58.2	87.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#10
Methyl Iodide
Concen: 97.884 ug/l
RT: 2.441 min Scan# 222
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:142 Resp: 272297

Ion	Ratio	Lower	Upper
142	100		
127	44.3	35.0	52.4
141	14.1	11.7	17.5






#11
Tert butyl alcohol
Concen: 447.152 ug/l m
RT: 3.026 min Scan# 3
Delta R.T. 0.043 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 59 Resp: 25962


Ion	Ratio	Lower	Upper
59	100		
57	5.0	3.9	5.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#12
1,1-Dichloroethene
Concen: 97.570 ug/l
RT: 2.306 min Scan# 200
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 96 Resp: 194870

Ion	Ratio	Lower	Upper
96	100		
61	161.3	131.1	196.7
98	64.6	51.2	76.8






#13
Acrolein
Concen: 432.534 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 56 Resp: 11683


Ion	Ratio	Lower	Upper
56	100		
55	70.2	55.1	82.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#14
Allyl chloride
Concen: 92.162 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 41 Resp: 358388

Ion	Ratio	Lower	Upper
41	100		
39	68.7	54.9	82.3
76	35.1	28.4	42.6






#15
Acrylonitrile
Concen: 498.179 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 53 Resp: 56084


Ion	Ratio	Lower	Upper
53	100		
52	82.7	66.0	99.0
51	36.1	28.9	43.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#16
Acetone
Concen: 518.081 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 43 Resp: 527741

Ion	Ratio	Lower	Upper
43	100		
58	31.0	24.6	37.0






#17
Carbon Disulfide
Concen: 97.144 ug/l
RT: 2.495 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 76 Resp: 52941


Ion	Ratio	Lower	Upper
76	100		
78	9.4	7.1	10.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#18
Methyl Acetate
Concen: 102.235 ug/l
RT: 2.697 min Scan# 264
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 43 Resp: 321291

Ion	Ratio	Lower	Upper
43	100		
74	25.1	19.8	29.8






#19
Methyl tert-butyl Ether
Concen: 94.245 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 73 Resp: 68177


Ion	Ratio	Lower	Upper
73	100		
57	21.1	17.4	26.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#20
Methylene Chloride
Concen: 91.100 ug/l
RT: 2.776 min Scan# 277
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 84 Resp: 216105

Ion	Ratio	Lower	Upper
84	100		
49	121.9	94.2	141.4
51	38.4	28.8	43.2
86	64.3	50.3	75.5






#21
trans-1,2-Dichloroethene
Concen: 91.400 ug/l
RT: 3.081 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 96 Resp: 19058


Ion	Ratio	Lower	Upper
96	100		
61	137.2	111.0	166.6
98	63.6	50.2	75.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#22
Diisopropyl ether
Concen: 93.579 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 45 Resp: 632562

Ion	Ratio	Lower	Upper
45	100		
43	76.7	61.0	91.6
87	28.0	22.7	34.1
59	11.8	9.4	14.2





#23
Vinyl Acetate
Concen: 471.863 ug/l
RT: 3.715 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14


Tgt Ion: 43 Resp: 279175


Ion	Ratio	Lower	Upper
43	100		
86	11.1	8.9	13.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#24
1,1-Dichloroethane
Concen: 94.530 ug/l
RT: 3.599 min Scan# 412
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 63 Resp: 378924

Ion	Ratio	Lower	Upper
63	100		
98	7.6	3.6	10.9
100	4.4	2.5	7.5





#25
2-Butanone
Concen: 505.435 ug/l
RT: 4.550 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14


Tgt Ion: 43 Resp: 76550


Ion	Ratio	Lower	Upper
43	100		
72	26.9	21.7	32.5

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#26
2,2-Dichloropropane
Concen: 92.064 ug/l
RT: 4.458 min Scan# 553
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 77 Resp: 357666

Ion	Ratio	Lower	Upper
77	100		
97	22.0	10.9	32.6






#27
cis-1,2-Dichloroethene
Concen: 91.559 ug/l
RT: 4.471 min Scan# 5
Delta R.T. -0.012 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 96 Resp: 24014


Ion	Ratio	Lower	Upper
96	100		
61	147.5	0.0	296.0
98	64.4	0.0	130.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#28
Bromochloromethane
Concen: 96.041 ug/l
RT: 4.885 min Scan# 623
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 49 Resp: 179896

Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	4.4
130	79.6	63.4	95.2






#29
Tetrahydrofuran
Concen: 488.500 ug/l
RT: 4.995 min Scan# 64
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 42 Resp: 47722


Ion	Ratio	Lower	Upper
42	100		
72	47.6	38.2	57.4
71	43.9	35.3	52.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#30
Chloroform
Concen: 92.373 ug/l
RT: 5.080 min Scan# 655
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 83 Resp: 373681

Ion	Ratio	Lower	Upper
83	100		
85	65.1	51.2	76.8





#31
Cyclohexane
Concen: 98.436 ug/l
RT: 5.452 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14


Tgt Ion: 56 Resp: 31913


Ion	Ratio	Lower	Upper
56	100		
69	32.0	26.8	40.2
84	90.6	71.0	106.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#32
1,1,1-Trichloroethane
Concen: 90.918 ug/l
RT: 5.367 min Scan# 702
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 97 Resp: 338119

Ion	Ratio	Lower	Upper
97	100		
99	63.8	51.6	77.4
61	43.6	35.6	53.4






#33
1,2-Dichloroethane-d4
Concen: 91.349 ug/l
RT: 5.940 min Scan# 7
Delta R.T. -0.012 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 65 Resp: 23519


Ion	Ratio	Lower	Upper
65	100		
67	50.9	0.0	103.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.751 min Scan# 929
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:114 Resp: 308450

Ion	Ratio	Lower	Upper
114	100		
63	21.4	0.0	43.0
88	17.8	0.0	35.4






#35
Dibromofluoromethane
Concen: 92.374 ug/l
RT: 5.373 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:113 Resp: 18101


Ion	Ratio	Lower	Upper
113	100		
111	103.1	81.6	122.4
192	17.8	14.4	21.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#36
1,1-Dichloropropene
Concen: 93.746 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 75 Resp: 255714

Ion	Ratio	Lower	Upper
75	100		
110	35.5	17.9	53.8
77	30.9	24.5	36.7






#37
Ethyl Acetate
Concen: 103.517 ug/l
RT: 4.702 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 43 Resp: 30070


Ion	Ratio	Lower	Upper
43	100		
61	13.6	10.9	16.3
70	12.0	10.1	15.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#38
Carbon Tetrachloride
Concen: 93.555 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:117 Resp: 287048

Ion	Ratio	Lower	Upper
117	100		
119	94.5	76.0	114.0
121	31.7	24.6	36.8






#39
Methylcyclohexane
Concen: 99.352 ug/l
RT: 7.367 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 83 Resp: 35975


Ion	Ratio	Lower	Upper
83	100		
55	78.5	60.5	90.7
98	49.1	39.0	58.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#40
Benzene
Concen: 92.946 ug/l
RT: 6.025 min Scan# 810
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 78 Resp: 785626

Ion	Ratio	Lower	Upper
78	100		
77	23.5	18.7	28.1






#41
Methacrylonitrile
Concen: 100.920 ug/l
RT: 4.910 min Scan# 6
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 41 Resp: 16848


Ion	Ratio	Lower	Upper
41	100		
39	55.5	44.4	66.6
67	73.0	60.6	90.8
52	29.6	24.3	36.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#42
1,2-Dichloroethane
Concen: 94.663 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 62 Resp: 292036

Ion	Ratio	Lower	Upper
62	100		
98	9.2	0.0	19.4






#43
Isopropyl Acetate
Concen: 95.925 ug/l
RT: 6.330 min Scan# 8
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 43 Resp: 51259


Ion	Ratio	Lower	Upper
43	100		
61	21.1	16.6	24.8
87	14.0	11.0	16.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#44
Trichloroethene
Concen: 98.212 ug/l
RT: 7.117 min Scan# 989
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:130 Resp: 199570

Ion	Ratio	Lower	Upper
130	100		
95	99.7	0.0	208.2






#45
1,2-Dichloropropane
Concen: 94.413 ug/l
RT: 7.421 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 63 Resp: 20087


Ion	Ratio	Lower	Upper
63	100		
65	30.9	24.2	36.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#46
Dibromomethane
Concen: 92.680 ug/l
RT: 7.568 min Scan# 1063
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 93 Resp: 150112

Ion	Ratio	Lower	Upper
93	100		
95	83.5	66.4	99.6
174	92.9	72.0	108.0






#47
Bromodichloromethane
Concen: 92.723 ug/l
RT: 7.812 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 83 Resp: 30799


Ion	Ratio	Lower	Upper
83	100		
85	63.3	50.1	75.1
127	8.0	6.5	9.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#48
Methyl methacrylate
Concen: 100.250 ug/l
RT: 7.684 min Scan# 1082
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 41 Resp: 248667

Ion	Ratio	Lower	Upper
41	100		
69	83.6	66.8	100.2
39	59.8	47.4	71.0






#49
1,4-Dioxane
Concen: 1959.056 ug/l
RT: 7.665 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 88 Resp: 10138


Ion	Ratio	Lower	Upper
88	100		
43	33.1	27.0	40.4
58	71.0	55.8	83.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#50
Toluene-d8
Concen: 90.404 ug/l
RT: 8.641 min Scan# 1239
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 98 Resp: 617688

Ion	Ratio	Lower	Upper
98	100		
100	66.8	53.2	79.8






#51
4-Methyl-2-Pentanone
Concen: 464.897 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 43 Resp: 137288


Ion	Ratio	Lower	Upper
43	100		
58	40.0	31.9	47.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#52
Toluene
Concen: 91.697 ug/l
RT: 8.708 min Scan# 1250
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 92 Resp: 474392

Ion	Ratio	Lower	Upper
92	100		
91	169.3	135.7	203.5






#53
t-1,3-Dichloropropene
Concen: 96.446 ug/l
RT: 8.970 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 75 Resp: 33098


Ion	Ratio	Lower	Upper
75	100		
77	31.5	24.4	36.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#54
cis-1,3-Dichloropropene
Concen: 93.927 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 75 Resp: 349516

Ion	Ratio	Lower	Upper
75	100		
77	32.1	25.1	37.7
39	52.5	40.3	60.5






#55
1,1,2-Trichloroethane
Concen: 91.885 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 97 Resp: 19232


Ion	Ratio	Lower	Upper
97	100		
83	83.1	68.8	103.2
85	53.4	43.7	65.5
99	61.5	49.5	74.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#56
Ethyl methacrylate
Concen: 93.812 ug/l
RT: 9.110 min Scan# 1316
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 69 Resp: 337626

Ion	Ratio	Lower	Upper
69	100		
41	66.7	53.2	79.8
39	38.4	30.2	45.2






#57
1,3-Dichloropropane
Concen: 91.220 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 76 Resp: 32514


Ion	Ratio	Lower	Upper
76	100		
78	32.7	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#58
2-Chloroethyl Vinyl ether
Concen: 462.693 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 63 Resp: 817528

Ion	Ratio	Lower	Upper
63	100		
106	29.3	23.0	34.6






#59
2-Hexanone
Concen: 469.454 ug/l
RT: 9.427 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 43 Resp: 101994


Ion	Ratio	Lower	Upper
43	100		
58	55.0	27.8	83.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#60
Dibromochloromethane
Concen: 92.570 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:129 Resp: 227418

Ion	Ratio	Lower	Upper
129	100		
127	77.5	38.5	115.5






#61
1,2-Dibromoethane
Concen: 92.200 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:107 Resp: 20224


Ion	Ratio	Lower	Upper
107	100		
109	93.7	75.3	112.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#62
4-Bromofluorobenzene
Concen: 92.702 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 95 Resp: 236119

Ion	Ratio	Lower	Upper
95	100		
174	66.9	0.0	133.0
176	64.0	0.0	129.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:117 Resp: 25588


Ion	Ratio	Lower	Upper
117	100		
82	58.3	47.6	71.4
119	31.3	25.4	38.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#64
Tetrachloroethene
Concen: 98.176 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:164 Resp: 157748

Ion	Ratio	Lower	Upper
164	100		
166	128.4	102.0	153.0
129	98.9	81.1	121.7
131	96.5	79.0	118.6






#65
Chlorobenzene
Concen: 94.003 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:112 Resp: 51091


Ion	Ratio	Lower	Upper
112	100		
114	32.0	25.3	37.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 93.565 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:131 Resp: 184816

Ion	Ratio	Lower	Upper
131	100		
133	96.0	47.8	143.3
119	67.1	33.4	100.2






#67
Ethyl Benzene
Concen: 95.872 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 91 Resp: 90920


Ion	Ratio	Lower	Upper
91	100		
106	30.7	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#68
m/p-Xylenes
Concen: 187.963 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:106 Resp: 661488

Ion	Ratio	Lower	Upper
106	100		
91	207.9	166.4	249.6






#69
o-Xylene
Concen: 92.536 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:106 Resp: 32933


Ion	Ratio	Lower	Upper
106	100		
91	218.7	108.7	326.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#70
Styrene
Concen: 95.197 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:104 Resp: 547016

Ion	Ratio	Lower	Upper
104	100		
78	53.9	43.0	64.6
103	54.1	43.4	65.2






#71
Bromoform
Concen: 96.648 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:173 Resp: 15537


Ion	Ratio	Lower	Upper
173	100		
175	48.0	23.6	70.8
254	0.1	0.1	0.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:152 Resp: 110019

Ion	Ratio	Lower	Upper
152	100		
115	110.6	43.8	131.4
150	186.5	0.0	347.8






#73
Isopropylbenzene
Concen: 92.779 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:105 Resp: 83387


Ion	Ratio	Lower	Upper
105	100		
120	26.4	13.2	39.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#74
N-amyl acetate
Concen: 93.463 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 43 Resp: 414046

Ion	Ratio	Lower	Upper
43	100		
70	43.5	35.2	52.8
55	23.3	19.0	28.4
61	25.1	20.0	30.0





#75
1,1,2,2-Tetrachloroethane
Concen: 89.776 ug/l
RT: 11.213 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14


Tgt Ion: 83 Resp: 27867


Ion	Ratio	Lower	Upper
83	100		
131	9.9	5.0	14.9
85	64.5	32.0	96.2

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#76
1,2,3-Trichloropropane
Concen: 96.568 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 75 Resp: 250918

Ion	Ratio	Lower	Upper
75	100		
77	43.9	23.4	70.2






#77
Bromobenzene
Concen: 91.729 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:156 Resp: 19270


Ion	Ratio	Lower	Upper
156	100		
77	181.5	90.3	271.0
158	97.9	48.4	145.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#78
n-propylbenzene
Concen: 96.470 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 91 Resp: 956202

Ion	Ratio	Lower	Upper
91	100		
120	22.7	11.4	34.1






#79
2-Chlorotoluene
Concen: 91.317 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 91 Resp: 58468


Ion	Ratio	Lower	Upper
91	100		
126	32.6	16.2	48.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#80
1,3,5-Trimethylbenzene
Concen: 91.926 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:105 Resp: 672586

Ion	Ratio	Lower	Upper
105	100		
120	47.2	23.7	71.1





#81
trans-1,4-Dichloro-2-butene
Concen: 98.354 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14


Tgt Ion: 75 Resp: 11483


Ion	Ratio	Lower	Upper
75	100		
53	93.6	74.3	111.5
89	47.4	37.4	56.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#82
4-Chlorotoluene
Concen: 93.421 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 91 Resp: 655250

Ion	Ratio	Lower	Upper
91	100		
126	27.9	13.9	41.7






#83
tert-Butylbenzene
Concen: 90.233 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:119 Resp: 69032


Ion	Ratio	Lower	Upper
119	100		
91	59.9	30.0	90.0
134	22.4	11.2	33.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#84
1,2,4-Trimethylbenzene
Concen: 91.646 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:105 Resp: 681398

Ion	Ratio	Lower	Upper
105	100		
120	44.4	22.2	66.6






#85
sec-Butylbenzene
Concen: 95.235 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:105 Resp: 84832


Ion	Ratio	Lower	Upper
105	100		
134	19.5	9.8	29.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#86
p-Isopropyltoluene
Concen: 94.999 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:119 Resp: 695433

Ion	Ratio	Lower	Upper
119	100		
134	26.1	12.8	38.4
91	25.7	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 97.115 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:146 Resp: 353809


Ion	Ratio	Lower	Upper
146	100		
111	43.2	21.9	65.7
148	63.3	31.8	95.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#88
1,4-Dichlorobenzene
Concen: 96.070 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:146 Resp: 347517

Ion	Ratio	Lower	Upper
146	100		
111	42.1	21.5	64.5
148	63.9	32.1	96.5






#89
n-Butylbenzene
Concen: 100.715 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 91 Resp: 63324


Ion	Ratio	Lower	Upper
91	100		
92	55.3	27.4	82.0
134	25.1	12.8	38.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#90
Hexachloroethane
Concen: 90.037 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:117 Resp: 149253

Ion	Ratio	Lower	Upper
117	100		
201	66.5	32.3	96.8






#91
1,2-Dichlorobenzene
Concen: 91.056 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:146 Resp: 34068


Ion	Ratio	Lower	Upper
146	100		
111	45.7	22.4	67.2
148	64.4	31.6	94.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 93.344 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion: 75 Resp: 70862

Ion	Ratio	Lower	Upper
75	100		
155	78.9	39.0	116.9
157	99.1	49.9	149.7






#93
1,2,4-Trichlorobenzene
Concen: 104.359 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:180 Resp: 23341


Ion	Ratio	Lower	Upper
180	100		
182	93.9	47.4	142.2
145	33.7	16.7	50.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#94
Hexachlorobutadiene
Concen: 100.406 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:225 Resp: 88214

Ion	Ratio	Lower	Upper
225	100		
223	63.6	31.9	95.7
227	63.4	31.8	95.4






#95
Naphthalene
Concen: 98.464 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:128 Resp: 84940


Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.3	15.5
129	10.9	8.7	13.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#96
1,2,3-Trichlorobenzene
Concen: 100.133 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX044652.D
Acq: 15 Jan 2025 13:14

Tgt Ion:180 Resp: 232152

Ion	Ratio	Lower	Upper
180	100		
182	96.0	47.2	141.6
145	35.5	17.5	52.6



Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044653.D
Acq On : 15 Jan 2025 13:37
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jan 16 01:02:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED
Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	160078	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	290219	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	234362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	100996	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	377212	156.114	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 312.220%#			
35) Dibromofluoromethane	5.379	113	289994	157.288	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 314.580%#			
50) Toluene-d8	8.647	98	961550	149.572	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 299.140%#			
62) 4-Bromofluorobenzene	11.079	95	366922	153.105	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 306.200%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	366593	170.185	ug/l	98
3) Chloromethane	1.300	50	342210	146.909	ug/l	97
4) Vinyl Chloride	1.374	62	359721	156.293	ug/l	99
5) Bromomethane	1.593	94	87716	159.609	ug/l	97
6) Chloroethane	1.660	64	71666	169.800	ug/l	98
7) Trichlorofluoromethane	1.855	101	398544	156.676	ug/l	97
8) Diethyl Ether	2.136	74	154291	150.299	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.313	101	302048	157.960	ug/l	99
10) Methyl Iodide	2.441	142	394541	151.126	ug/l	99
11) Tert butyl alcohol	3.062	59	387839m	711.783	ug/l	
12) 1,1-Dichloroethene	2.306	96	296987	158.449	ug/l	99
13) Acrolein	2.239	56	193280	762.485	ug/l	99
14) Allyl chloride	2.654	41	534150	146.366	ug/l	100
15) Acrylonitrile	3.068	53	783976	742.040	ug/l	99
16) Acetone	2.392	43	708100	740.713	ug/l	100
17) Carbon Disulfide	2.495	76	803379	157.079	ug/l	99
18) Methyl Acetate	2.703	43	483139	163.814	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	1034710	152.410	ug/l	99
20) Methylene Chloride	2.782	84	325957	146.417	ug/l	99
21) trans-1,2-Dichloroethene	3.081	96	290671	148.536	ug/l	98
22) Diisopropyl ether	3.763	45	957045	150.864	ug/l	91
23) Vinyl Acetate	3.721	43	4101581	738.701	ug/l	100
24) 1,1-Dichloroethane	3.599	63	594429	158.014	ug/l	99
25) 2-Butanone	4.562	43	1079716	759.640	ug/l	100
26) 2,2-Dichloropropane	4.465	77	550106	150.882	ug/l	99
27) cis-1,2-Dichloroethene	4.477	96	370177	150.386	ug/l	99
28) Bromochloromethane	4.891	49	251238	142.922	ug/l	99
29) Tetrahydrofuran	5.007	42	689928	752.528	ug/l	99
30) Chloroform	5.086	83	579322	152.595	ug/l	99
31) Cyclohexane	5.458	56	476106	156.483	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	522062	149.582	ug/l	99
36) 1,1-Dichloropropene	5.684	75	386304	150.517	ug/l	99
37) Ethyl Acetate	4.715	43	433924	158.764	ug/l	99
38) Carbon Tetrachloride	5.666	117	437442	151.528	ug/l	99
39) Methylcyclohexane	7.373	83	534976	157.021	ug/l	99
40) Benzene	6.031	78	1205057	151.523	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044653.D
Acq On : 15 Jan 2025 13:37
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jan 16 01:02:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
41) Methacrylonitrile	4.916	41	238601	151.900 ug/l	98
42) 1,2-Dichloroethane	6.080	62	446380	153.782 ug/l	98
43) Isopropyl Acetate	6.336	43	761866	151.528 ug/l	100
44) Trichloroethene	7.117	130	309226	161.735 ug/l	96
45) 1,2-Dichloropropane	7.421	63	308641	154.175 ug/l	99
46) Dibromomethane	7.574	93	231674	152.021 ug/l	99
47) Bromodichloromethane	7.818	83	471682	150.924 ug/l	99
48) Methyl methacrylate	7.690	41	368626	157.946 ug/l	99
49) 1,4-Dioxane	7.671	88	146350	3005.585 ug/l	99
51) 4-Methyl-2-Pentanone	8.574	43	1994498	717.818 ug/l	100
52) Toluene	8.714	92	699585	143.719 ug/l	100
53) t-1,3-Dichloropropene	8.976	75	490414	151.880 ug/l	96
54) cis-1,3-Dichloropropene	8.360	75	524796	149.891 ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	279505	141.923 ug/l	99
56) Ethyl methacrylate	9.116	69	499842	147.609 ug/l	100
57) 1,3-Dichloropropane	9.305	76	485347	144.720 ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	1178391	708.825 ug/l	99
59) 2-Hexanone	9.433	43	1482861	725.401 ug/l	99
60) Dibromochloromethane	9.519	129	335937	145.333 ug/l	99
61) 1,2-Dibromoethane	9.604	107	296982	143.893 ug/l	100
64) Tetrachloroethene	9.269	164	229004	155.612 ug/l	98
65) Chlorobenzene	10.079	112	750161	150.697 ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	276991	153.108 ug/l	99
67) Ethyl Benzene	10.189	91	1315242	151.424 ug/l	99
68) m/p-Xylenes	10.299	106	941355	292.054 ug/l	100
69) o-Xylene	10.640	106	473907	145.386 ug/l	98
70) Styrene	10.652	104	792889	150.660 ug/l	100
71) Bromoform	10.799	173	232464	157.877 ug/l	99
73) Isopropylbenzene	10.957	105	1207427	146.343 ug/l	100
74) N-amyl acetate	10.841	43	622491	153.069 ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	415068	145.660 ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	356222m	149.343 ug/l	
77) Bromobenzene	11.195	156	288579	149.641 ug/l	100
78) n-propylbenzene	11.305	91	1375979	151.223 ug/l	100
79) 2-Chlorotoluene	11.366	91	857617	145.911 ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	959600	142.870 ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	171786	160.280 ug/l	99
82) 4-Chlorotoluene	11.451	91	957835	148.761 ug/l	100
83) tert-Butylbenzene	11.713	119	999760	142.356 ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	991595	145.281 ug/l	100
85) sec-Butylbenzene	11.890	105	1214662	148.543 ug/l	100
86) p-Isopropyltoluene	12.006	119	1000632	148.902 ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	521112	155.816 ug/l	100
88) 1,4-Dichlorobenzene	12.042	146	516467	155.532 ug/l	99
89) n-Butylbenzene	12.329	91	904589	156.724 ug/l	99
90) Hexachloroethane	12.536	117	222866	146.455 ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	511280	148.860 ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	107815	154.709 ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	352298	171.582 ug/l	99
94) Hexachlorobutadiene	13.725	225	129218	160.216 ug/l	99
95) Naphthalene	13.774	128	1285749	162.361 ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	350134	164.515 ug/l	99

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044653.D
Acq On : 15 Jan 2025 13:37
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Quant Time: Jan 16 01:02:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044653.D
Acq On : 15 Jan 2025 13:37
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1


Quant Time: Jan 16 01:02:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 00:57:36 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:168 Resp: 16007


Ion	Ratio	Lower	Upper
168	100		
99	65.5	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#2
Dichlorodifluoromethane
Concen: 170.185 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 85 Resp: 366593

Ion	Ratio	Lower	Upper
85	100		
87	32.5	16.8	50.4






#3
Chloromethane
Concen: 146.909 ug/l
RT: 1.300 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 50 Resp: 34221


Ion	Ratio	Lower	Upper
50	100		
52	32.2	27.0	40.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#4
Vinyl Chloride
Concen: 156.293 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 62 Resp: 359721

Ion	Ratio	Lower	Upper
62	100		
64	32.5	26.6	39.8






#5
Bromomethane
Concen: 159.609 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 94 Resp: 8771


Ion	Ratio	Lower	Upper
94	100		
96	93.4	72.8	109.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#6
Chloroethane
Concen: 169.800 ug/l
RT: 1.660 min Scan# 94
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 64 Resp: 71666

Ion	Ratio	Lower	Upper
64	100		
66	31.8	26.6	39.8






#7
Trichlorofluoromethane
Concen: 156.676 ug/l
RT: 1.855 min Scan# 1
Delta R.T. -0.012 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:101 Resp: 39854


Ion	Ratio	Lower	Upper
101	100		
103	63.3	52.5	78.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#8
Diethyl Ether
Concen: 150.299 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 74 Resp: 154291

Ion	Ratio	Lower	Upper
74	100		
45	98.3	47.1	141.3





#9
1,1,2-Trichlorotrifluoroethane
Concen: 157.960 ug/l
RT: 2.313 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37


Tgt Ion:101 Resp: 30204


Ion	Ratio	Lower	Upper
101	100		
85	43.8	35.3	52.9
151	72.0	58.2	87.4

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#10
Methyl Iodide
Concen: 151.126 ug/l
RT: 2.441 min Scan# 222
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:142 Resp: 394541

Ion	Ratio	Lower	Upper
142	100		
127	44.4	35.0	52.4
141	14.2	11.7	17.5






#11
Tert butyl alcohol
Concen: 711.783 ug/l m
RT: 3.062 min Scan# 3
Delta R.T. 0.079 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 59 Resp: 38783


Ion	Ratio	Lower	Upper
59	100		
57	4.5	3.9	5.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#12
1,1-Dichloroethene
Concen: 158.449 ug/l
RT: 2.306 min Scan# 200
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 96 Resp: 296987

Ion	Ratio	Lower	Upper
96	100		
61	161.3	131.1	196.7
98	64.0	51.2	76.8






#13
Acrolein
Concen: 762.485 ug/l
RT: 2.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 56 Resp: 19328


Ion	Ratio	Lower	Upper
56	100		
55	69.9	55.1	82.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#14
Allyl chloride
Concen: 146.366 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 41 Resp: 534150

Ion	Ratio	Lower	Upper
41	100		
39	68.7	54.9	82.3
76	34.9	28.4	42.6






#15
Acrylonitrile
Concen: 742.040 ug/l
RT: 3.068 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 53 Resp: 78397


Ion	Ratio	Lower	Upper
53	100		
52	83.1	66.0	99.0
51	36.2	28.9	43.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#16
Acetone
Concen: 740.713 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 43 Resp: 708100

Ion	Ratio	Lower	Upper
43	100		
58	30.7	24.6	37.0





#17
Carbon Disulfide
Concen: 157.079 ug/l
RT: 2.495 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37


Tgt Ion: 76 Resp: 80337


Ion	Ratio	Lower	Upper
76	100		
78	9.1	7.1	10.7

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#18
Methyl Acetate
Concen: 163.814 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 43 Resp: 483139

Ion	Ratio	Lower	Upper
43	100		
74	24.7	19.8	29.8






#19
Methyl tert-butyl Ether
Concen: 152.410 ug/l
RT: 3.117 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 73 Resp: 103471


Ion	Ratio	Lower	Upper
73	100		
57	21.2	17.4	26.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#20
Methylene Chloride
Concen: 146.417 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 84 Resp: 325957

Ion	Ratio	Lower	Upper
84	100		
49	119.1	94.2	141.4
51	36.7	28.8	43.2
86	64.5	50.3	75.5






#21
trans-1,2-Dichloroethene
Concen: 148.536 ug/l
RT: 3.081 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 96 Resp: 29067


Ion	Ratio	Lower	Upper
96	100		
61	141.3	111.0	166.6
98	64.2	50.2	75.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#22
Diisopropyl ether
Concen: 150.864 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 45 Resp: 957045

Ion	Ratio	Lower	Upper
45	100		
43	64.8	61.0	91.6
87	29.6	22.7	34.1
59	11.7	9.4	14.2





#23
Vinyl Acetate
Concen: 738.701 ug/l
RT: 3.721 min Scan# 4
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37


Tgt Ion: 43 Resp: 410158


Ion	Ratio	Lower	Upper
43	100		
86	11.3	8.9	13.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#24
1,1-Dichloroethane
Concen: 158.014 ug/l
RT: 3.599 min Scan# 412
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 63 Resp: 594429

Ion	Ratio	Lower	Upper
63	100		
98	7.0	3.6	10.9
100	4.8	2.5	7.5






#25
2-Butanone
Concen: 759.640 ug/l
RT: 4.562 min Scan# 5
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 43 Resp: 107971


Ion	Ratio	Lower	Upper
43	100		
72	27.1	21.7	32.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#26
2,2-Dichloropropane
Concen: 150.882 ug/l
RT: 4.465 min Scan# 554
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 77 Resp: 550106

Ion	Ratio	Lower	Upper
77	100		
97	22.0	10.9	32.6






#27
cis-1,2-Dichloroethene
Concen: 150.386 ug/l
RT: 4.477 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 96 Resp: 37017


Ion	Ratio	Lower	Upper
96	100		
61	147.0	0.0	296.0
98	64.4	0.0	130.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#28
Bromochloromethane
Concen: 142.922 ug/l
RT: 4.891 min Scan# 624
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 49 Resp: 251238

Ion	Ratio	Lower	Upper
49	100		
129	2.2	0.0	4.4
130	80.2	63.4	95.2






#29
Tetrahydrofuran
Concen: 752.528 ug/l
RT: 5.007 min Scan# 64
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 42 Resp: 68992


Ion	Ratio	Lower	Upper
42	100		
72	47.0	38.2	57.4
71	43.6	35.3	52.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#30
Chloroform
Concen: 152.595 ug/l
RT: 5.086 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 83 Resp: 579322

Ion	Ratio	Lower	Upper
83	100		
85	64.5	51.2	76.8






#31
Cyclohexane
Concen: 156.483 ug/l
RT: 5.458 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 56 Resp: 47610


Ion	Ratio	Lower	Upper
56	100		
69	32.1	26.8	40.2
84	89.3	71.0	106.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#32
1,1,1-Trichloroethane
Concen: 149.582 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 97 Resp: 522062

Ion	Ratio	Lower	Upper
97	100		
99	64.1	51.6	77.4
61	43.9	35.6	53.4






#33
1,2-Dichloroethane-d4
Concen: 156.114 ug/l
RT: 5.946 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 65 Resp: 37721


Ion	Ratio	Lower	Upper
65	100		
67	51.4	0.0	103.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.751 min Scan# 929
Delta R.T. -0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:114 Resp: 290219

Ion	Ratio	Lower	Upper
114	100		
63	21.6	0.0	43.0
88	17.8	0.0	35.4






#35
Dibromofluoromethane
Concen: 157.288 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:113 Resp: 28999


Ion	Ratio	Lower	Upper
113	100		
111	103.2	81.6	122.4
192	17.9	14.4	21.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#36
1,1-Dichloropropene
Concen: 150.517 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 75 Resp: 386304

Ion	Ratio	Lower	Upper
75	100		
110	35.5	17.9	53.8
77	31.1	24.5	36.7






#37
Ethyl Acetate
Concen: 158.764 ug/l
RT: 4.715 min Scan# 5
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 43 Resp: 43392


Ion	Ratio	Lower	Upper
43	100		
61	14.0	10.9	16.3
70	12.3	10.1	15.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#38
Carbon Tetrachloride
Concen: 151.528 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:117 Resp: 437442

Ion	Ratio	Lower	Upper
117	100		
119	95.8	76.0	114.0
121	30.3	24.6	36.8






#39
Methylcyclohexane
Concen: 157.021 ug/l
RT: 7.373 min Scan# 1(
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 83 Resp: 53497(


Ion	Ratio	Lower	Upper
83	100		
55	76.6	60.5	90.7
98	49.0	39.0	58.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#40
Benzene
Concen: 151.523 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 78 Resp: 1205057

Ion	Ratio	Lower	Upper
78	100		
77	24.0	18.7	28.1






#41
Methacrylonitrile
Concen: 151.900 ug/l
RT: 4.916 min Scan# 6
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 41 Resp: 23860


Ion	Ratio	Lower	Upper
41	100		
39	57.9	44.4	66.6
67	74.8	60.6	90.8
52	30.6	24.3	36.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#42
1,2-Dichloroethane
Concen: 153.782 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 62 Resp: 446380

Ion	Ratio	Lower	Upper
62	100		
98	9.1	0.0	19.4






#43
Isopropyl Acetate
Concen: 151.528 ug/l
RT: 6.336 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 43 Resp: 76186


Ion	Ratio	Lower	Upper
43	100		
61	20.8	16.6	24.8
87	14.0	11.0	16.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#44
Trichloroethene
Concen: 161.735 ug/l
RT: 7.117 min Scan# 989
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:130 Resp: 309226

Ion	Ratio	Lower	Upper
130	100		
95	99.7	0.0	208.2






#45
1,2-Dichloropropane
Concen: 154.175 ug/l
RT: 7.421 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 63 Resp: 30864


Ion	Ratio	Lower	Upper
63	100		
65	29.7	24.2	36.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#46
Dibromomethane
Concen: 152.021 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 93 Resp: 231674

Ion	Ratio	Lower	Upper
93	100		
95	82.6	66.4	99.6
174	91.2	72.0	108.0






#47
Bromodichloromethane
Concen: 150.924 ug/l
RT: 7.818 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 83 Resp: 47168


Ion	Ratio	Lower	Upper
83	100		
85	63.8	50.1	75.1
127	8.3	6.5	9.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#48
Methyl methacrylate
Concen: 157.946 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 41 Resp: 368626

Ion	Ratio	Lower	Upper
41	100		
69	82.4	66.8	100.2
39	59.1	47.4	71.0






#49
1,4-Dioxane
Concen: 3005.585 ug/l
RT: 7.671 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 88 Resp: 14635


Ion	Ratio	Lower	Upper
88	100		
43	33.0	27.0	40.4
58	70.7	55.8	83.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#50
Toluene-d8
Concen: 149.572 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 98 Resp: 961550

Ion	Ratio	Lower	Upper
98	100		
100	66.1	53.2	79.8






#51
4-Methyl-2-Pentanone
Concen: 717.818 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 43 Resp: 199449


Ion	Ratio	Lower	Upper
43	100		
58	39.7	31.9	47.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#52
Toluene
Concen: 143.719 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 92 Resp: 699585

Ion	Ratio	Lower	Upper
92	100		
91	169.9	135.7	203.5






#53
t-1,3-Dichloropropene
Concen: 151.880 ug/l
RT: 8.976 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 75 Resp: 49041


Ion	Ratio	Lower	Upper
75	100		
77	32.5	24.4	36.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#54
cis-1,3-Dichloropropene
Concen: 149.891 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 75 Resp: 524796

Ion	Ratio	Lower	Upper
75	100		
77	31.2	25.1	37.7
39	52.0	40.3	60.5






#55
1,1,2-Trichloroethane
Concen: 141.923 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 97 Resp: 27950


Ion	Ratio	Lower	Upper
97	100		
83	86.2	68.8	103.2
85	56.0	43.7	65.5
99	62.6	49.5	74.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#56
Ethyl methacrylate
Concen: 147.609 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 69 Resp: 499842

Ion	Ratio	Lower	Upper
69	100		
41	66.9	53.2	79.8
39	38.0	30.2	45.2






#57
1,3-Dichloropropane
Concen: 144.720 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 76 Resp: 48534


Ion	Ratio	Lower	Upper
76	100		
78	32.1	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#58
2-Chloroethyl Vinyl ether
Concen: 708.825 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 63 Resp: 1178391

Ion	Ratio	Lower	Upper
63	100		
106	29.1	23.0	34.6






#59
2-Hexanone
Concen: 725.401 ug/l
RT: 9.433 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 43 Resp: 148286


Ion	Ratio	Lower	Upper
43	100		
58	55.1	27.8	83.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#60
Dibromochloromethane
Concen: 145.333 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:129 Resp: 335937

Ion	Ratio	Lower	Upper
129	100		
127	77.9	38.5	115.5






#61
1,2-Dibromoethane
Concen: 143.893 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:107 Resp: 29698


Ion	Ratio	Lower	Upper
107	100		
109	94.6	75.3	112.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#62
4-Bromofluorobenzene
Concen: 153.105 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 95 Resp: 366922

Ion	Ratio	Lower	Upper
95	100		
174	67.5	0.0	133.0
176	64.4	0.0	129.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:117 Resp: 23436


Ion	Ratio	Lower	Upper
117	100		
82	59.6	47.6	71.4
119	31.7	25.4	38.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#64
Tetrachloroethene
Concen: 155.612 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:164 Resp: 229004

Ion	Ratio	Lower	Upper
164	100		
166	124.0	102.0	153.0
129	101.7	81.1	121.7
131	97.5	79.0	118.6






#65
Chlorobenzene
Concen: 150.697 ug/l
RT: 10.079 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:112 Resp: 75016


Ion	Ratio	Lower	Upper
112	100		
114	32.5	25.3	37.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 153.108 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:131 Resp: 276991

Ion	Ratio	Lower	Upper
131	100		
133	95.0	47.8	143.3
119	65.5	33.4	100.2






#67
Ethyl Benzene
Concen: 151.424 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 91 Resp: 131524


Ion	Ratio	Lower	Upper
91	100		
106	30.5	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#68
m/p-Xylenes
Concen: 292.054 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:106 Resp: 941355

Ion	Ratio	Lower	Upper
106	100		
91	207.8	166.4	249.6






#69
o-Xylene
Concen: 145.386 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:106 Resp: 47390


Ion	Ratio	Lower	Upper
106	100		
91	219.9	108.7	326.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#70
Styrene
Concen: 150.660 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:104 Resp: 792889

Ion	Ratio	Lower	Upper
104	100		
78	54.0	43.0	64.6
103	54.3	43.4	65.2






#71
Bromoform
Concen: 157.877 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:173 Resp: 23246


Ion	Ratio	Lower	Upper
173	100		
175	47.9	23.6	70.8
254	0.1	0.1	0.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:152 Resp: 100996

Ion	Ratio	Lower	Upper
152	100		
115	137.3	43.8	131.4#
150	206.7	0.0	347.8






#73
Isopropylbenzene
Concen: 146.343 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:105 Resp: 120742


Ion	Ratio	Lower	Upper
105	100		
120	26.2	13.2	39.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#74
N-amyl acetate
Concen: 153.069 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 43 Resp: 622491

Ion	Ratio	Lower	Upper
43	100		
70	43.3	35.2	52.8
55	23.3	19.0	28.4
61	24.7	20.0	30.0





#75
1,1,2,2-Tetrachloroethane
Concen: 145.660 ug/l
RT: 11.213 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37


Tgt Ion: 83 Resp: 41506

Ion	Ratio	Lower	Upper
83	100		
131	10.0	5.0	14.9
85	64.5	32.0	96.2

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#76
1,2,3-Trichloropropane
Concen: 149.343 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 75 Resp: 356222


Ion	Ratio	Lower	Upper
75	100		
77	45.8	23.4	70.2



#77
Bromobenzene
Concen: 149.641 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:156 Resp: 28857


Ion	Ratio	Lower	Upper
156	100		
77	180.4	90.3	271.0
158	97.5	48.4	145.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#78
n-propylbenzene
Concen: 151.223 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 91 Resp: 1375979

Ion	Ratio	Lower	Upper
91	100		
120	22.7	11.4	34.1






#79
2-Chlorotoluene
Concen: 145.911 ug/l
RT: 11.366 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 91 Resp: 85761


Ion	Ratio	Lower	Upper
91	100		
126	32.0	16.2	48.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#80
1,3,5-Trimethylbenzene
Concen: 142.870 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:105 Resp: 959600

Ion	Ratio	Lower	Upper
105	100		
120	47.3	23.7	71.1





#81
trans-1,4-Dichloro-2-butene
Concen: 160.280 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37


Tgt Ion: 75 Resp: 17178


Ion	Ratio	Lower	Upper
75	100		
53	92.3	74.3	111.5
89	47.0	37.4	56.0

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#82
4-Chlorotoluene
Concen: 148.761 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 91 Resp: 957835

Ion	Ratio	Lower	Upper
91	100		
126	27.8	13.9	41.7






#83
tert-Butylbenzene
Concen: 142.356 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:119 Resp: 99976


Ion	Ratio	Lower	Upper
119	100		
91	61.4	30.0	90.0
134	22.5	11.2	33.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#84
1,2,4-Trimethylbenzene
Concen: 145.281 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:105 Resp: 991595

Ion	Ratio	Lower	Upper
105	100		
120	44.6	22.2	66.6






#85
sec-Butylbenzene
Concen: 148.543 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:105 Resp: 121466


Ion	Ratio	Lower	Upper
105	100		
134	19.9	9.8	29.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#86
p-Isopropyltoluene
Concen: 148.902 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:119 Resp: 1000632

Ion	Ratio	Lower	Upper
119	100		
134	25.8	12.8	38.4
91	25.9	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 155.816 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:146 Resp: 52111


Ion	Ratio	Lower	Upper
146	100		
111	43.7	21.9	65.7
148	64.0	31.8	95.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#88
1,4-Dichlorobenzene
Concen: 155.532 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. 0.006 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:146 Resp: 516467

Ion	Ratio	Lower	Upper
146	100		
111	42.2	21.5	64.5
148	63.1	32.1	96.5






#89
n-Butylbenzene
Concen: 156.724 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 91 Resp: 90458


Ion	Ratio	Lower	Upper
91	100		
92	55.2	27.4	82.0
134	25.3	12.8	38.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#90
Hexachloroethane
Concen: 146.455 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:117 Resp: 222866

Ion	Ratio	Lower	Upper
117	100		
201	66.1	32.3	96.8






#91
1,2-Dichlorobenzene
Concen: 148.860 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:146 Resp: 51128


Ion	Ratio	Lower	Upper
146	100		
111	45.3	22.4	67.2
148	63.5	31.6	94.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 154.709 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion: 75 Resp: 107815

Ion	Ratio	Lower	Upper
75	100		
155	79.1	39.0	116.9
157	98.6	49.9	149.7






#93
1,2,4-Trichlorobenzene
Concen: 171.582 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:180 Resp: 35229


Ion	Ratio	Lower	Upper
180	100		
182	93.9	47.4	142.2
145	33.4	16.7	50.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#94
Hexachlorobutadiene
Concen: 160.216 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:225 Resp: 129218

Ion	Ratio	Lower	Upper
225	100		
223	63.5	31.9	95.7
227	64.4	31.8	95.4






#95
Naphthalene
Concen: 162.361 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:128 Resp: 128574


Ion	Ratio	Lower	Upper
128	100		
127	12.8	10.3	15.5
129	10.9	8.7	13.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 01/15/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#96
1,2,3-Trichlorobenzene
Concen: 164.515 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX044653.D
Acq: 15 Jan 2025 13:37

Tgt Ion:180 Resp: 350134

Ion	Ratio	Lower	Upper
180	100		
182	95.6	47.2	141.6
145	35.9	17.5	52.6



Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044655.D
Acq On : 15 Jan 2025 15:15
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jan 16 01:21:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED
Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.537	168	179405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	323521	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	266872	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	115489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	141798	52.363	ug/l	-0.01
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.720%			
35) Dibromofluoromethane	5.373	113	110333	53.683	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.360%			
50) Toluene-d8	8.641	98	380624	53.113	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.220%			
62) 4-Bromofluorobenzene	11.079	95	140770	52.693	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.380%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	136518	56.549	ug/l	99
3) Chloromethane	1.300	50	124691	47.763	ug/l	99
4) Vinyl Chloride	1.374	62	128558	49.839	ug/l	98
5) Bromomethane	1.593	94	30815	50.031	ug/l	98
6) Chloroethane	1.660	64	23473	49.624	ug/l	98
7) Trichlorofluoromethane	1.861	101	147644	51.789	ug/l	98
8) Diethyl Ether	2.130	74	54624	47.479	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.312	101	112018	52.271	ug/l	100
10) Methyl Iodide	2.441	142	150245	51.350	ug/l	99
11) Tert butyl alcohol	2.983	59	150021m	248.503	ug/l	
12) 1,1-Dichloroethene	2.306	96	107643	51.243	ug/l	97
13) Acrolein	2.233	56	65602	230.919	ug/l	98
14) Allyl chloride	2.654	41	197270	48.232	ug/l	100
15) Acrylonitrile	3.062	53	307790	259.942	ug/l	99
16) Acetone	2.386	43	285777	266.735	ug/l	100
17) Carbon Disulfide	2.495	76	291339	50.827	ug/l	100
18) Methyl Acetate	2.703	43	177037	53.560	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	383394	50.389	ug/l	99
20) Methylene Chloride	2.782	84	119950	48.076	ug/l	98
21) trans-1,2-Dichloroethene	3.081	96	107127	48.846	ug/l	97
22) Diisopropyl ether	3.757	45	360169	50.659	ug/l	93
23) Vinyl Acetate	3.715	43	1583340	254.442	ug/l	99
24) 1,1-Dichloroethane	3.599	63	211436	50.150	ug/l	99
25) 2-Butanone	4.550	43	423083	265.596	ug/l	99
26) 2,2-Dichloropropane	4.458	77	198981	48.697	ug/l	98
27) cis-1,2-Dichloroethene	4.477	96	134711	48.831	ug/l	100
28) Bromochloromethane	4.885	49	97608	49.545	ug/l	99
29) Tetrahydrofuran	4.995	42	263635	256.578	ug/l	99
30) Chloroform	5.080	83	214034	50.304	ug/l	99
31) Cyclohexane	5.452	56	180309	52.878	ug/l	97
32) 1,1,1-Trichloroethane	5.367	97	191114	48.859	ug/l	99
36) 1,1-Dichloropropene	5.678	75	143891	50.294	ug/l	99
37) Ethyl Acetate	4.702	43	166128	54.526	ug/l	99
38) Carbon Tetrachloride	5.659	117	161829	50.287	ug/l	100
39) Methylcyclohexane	7.366	83	199470	52.520	ug/l	98
40) Benzene	6.025	78	448684	50.610	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044655.D
Acq On : 15 Jan 2025 15:15
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jan 16 01:21:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.910	41	91151	51.342	ug/l		100
42) 1,2-Dichloroethane	6.074	62	164707	50.902	ug/l		100
43) Isopropyl Acetate	6.330	43	285121	50.871	ug/l		100
44) Trichloroethene	7.116	130	110588	51.887	ug/l		99
45) 1,2-Dichloropropane	7.421	63	113836	51.011	ug/l		97
46) Dibromomethane	7.574	93	84838	49.939	ug/l		100
47) Bromodichloromethane	7.811	83	174297	50.029	ug/l		99
48) Methyl methacrylate	7.683	41	137603	52.890	ug/l		99
49) 1,4-Dioxane	7.665	88	53851	992.095	ug/l		97
51) 4-Methyl-2-Pentanone	8.567	43	779855	251.778	ug/l		100
52) Toluene	8.714	92	271071	49.955	ug/l		98
53) t-1,3-Dichloropropene	8.970	75	184653	51.300	ug/l		99
54) cis-1,3-Dichloropropene	8.360	75	199240	51.049	ug/l		99
55) 1,1,2-Trichloroethane	9.147	97	109957	50.085	ug/l		99
56) Ethyl methacrylate	9.110	69	190986	50.595	ug/l		99
57) 1,3-Dichloropropane	9.305	76	187407	50.129	ug/l		99
58) 2-Chloroethyl Vinyl ether	8.232	63	466737	251.852	ug/l		100
59) 2-Hexanone	9.427	43	576549	253.010	ug/l		100
60) Dibromochloromethane	9.518	129	128727	49.957	ug/l		100
61) 1,2-Dibromoethane	9.604	107	114330	49.693	ug/l		100
64) Tetrachloroethene	9.269	164	89460	53.384	ug/l		97
65) Chlorobenzene	10.073	112	287202	50.667	ug/l		98
66) 1,1,1,2-Tetrachloroethane	10.159	131	105329	51.129	ug/l		99
67) Ethyl Benzene	10.189	91	511560	51.721	ug/l		98
68) m/p-Xylenes	10.299	106	376743	102.645	ug/l		100
69) o-Xylene	10.640	106	186995	50.378	ug/l		99
70) Styrene	10.652	104	308018	51.398	ug/l		100
71) Bromoform	10.799	173	86512	51.597	ug/l	#	99
73) Isopropylbenzene	10.957	105	474051	50.246	ug/l		100
74) N-amyl acetate	10.841	43	233338	50.177	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.213	83	159673	49.002	ug/l		98
76) 1,2,3-Trichloropropane	11.238	75	140149m	50.004	ug/l		
77) Bromobenzene	11.195	156	109171	49.506	ug/l		99
78) n-propylbenzene	11.299	91	538048	51.712	ug/l		100
79) 2-Chlorotoluene	11.360	91	333891	49.678	ug/l		100
80) 1,3,5-Trimethylbenzene	11.445	105	382880	49.852	ug/l		100
81) trans-1,4-Dichloro-2-b...	11.018	75	63708	51.981	ug/l		100
82) 4-Chlorotoluene	11.451	91	373655	50.750	ug/l		100
83) tert-Butylbenzene	11.713	119	397191	49.459	ug/l		99
84) 1,2,4-Trimethylbenzene	11.750	105	388010	49.714	ug/l		100
85) sec-Butylbenzene	11.890	105	477595	51.076	ug/l		100
86) p-Isopropyltoluene	12.006	119	392123	51.028	ug/l		99
87) 1,3-Dichlorobenzene	11.969	146	196862	51.476	ug/l		99
88) 1,4-Dichlorobenzene	12.036	146	193343	50.011	ug/l		99
89) n-Butylbenzene	12.329	91	350704	53.136	ug/l		99
90) Hexachloroethane	12.536	117	84427	48.518	ug/l		100
91) 1,2-Dichlorobenzene	12.335	146	194138	49.430	ug/l		98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	39583	49.672	ug/l		99
93) 1,2,4-Trichlorobenzene	13.585	180	125286	53.362	ug/l		99
94) Hexachlorobutadiene	13.725	225	48702	52.807	ug/l		99
95) Naphthalene	13.774	128	479024	52.899	ug/l		100
96) 1,2,3-Trichlorobenzene	13.957	180	128038	52.611	ug/l		100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044655.D
Acq On : 15 Jan 2025 15:15
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025

Quant Time: Jan 16 01:21:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044655.D
Acq On : 15 Jan 2025 15:15
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1


Quant Time: Jan 16 01:21:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.537 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:168 Resp: 17940


Ion	Ratio	Lower	Upper
168	100		
99	65.9	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#2
Dichlorodifluoromethane
Concen: 56.549 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 85 Resp: 136518

Ion	Ratio	Lower	Upper
85	100		
87	33.0	16.8	50.4





#3
Chloromethane
Concen: 47.763 ug/l
RT: 1.300 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15


Tgt Ion: 50 Resp: 12469

Ion	Ratio	Lower	Upper
50	100		
52	33.3	27.0	40.4

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#4
Vinyl Chloride
Concen: 49.839 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 62 Resp: 128558


Ion	Ratio	Lower	Upper
62	100		
64	32.2	26.6	39.8



#5
Bromomethane
Concen: 50.031 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 94 Resp: 3081


Ion	Ratio	Lower	Upper
94	100		
96	92.8	72.8	109.2



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#6
Chloroethane
Concen: 49.624 ug/l
RT: 1.660 min Scan# 94
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 64 Resp: 23473

Ion	Ratio	Lower	Upper
64	100		
66	32.2	26.6	39.8






#7
Trichlorofluoromethane
Concen: 51.789 ug/l
RT: 1.861 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:101 Resp: 14764


Ion	Ratio	Lower	Upper
101	100		
103	63.7	52.5	78.7



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#8
Diethyl Ether
Concen: 47.479 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 74 Resp: 54624

Ion	Ratio	Lower	Upper
74	100		
45	98.4	47.1	141.3





#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.271 ug/l
RT: 2.312 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15


Tgt Ion:101 Resp: 11201


Ion	Ratio	Lower	Upper
101	100		
85	43.6	35.3	52.9
151	72.6	58.2	87.4

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#10
Methyl Iodide
Concen: 51.350 ug/l
RT: 2.441 min Scan# 222
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:142 Resp: 150245

Ion	Ratio	Lower	Upper
142	100		
127	43.6	35.0	52.4
141	13.8	11.7	17.5






#11
Tert butyl alcohol
Concen: 248.503 ug/l m
RT: 2.983 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 59 Resp: 15002


Ion	Ratio	Lower	Upper
59	100		
57	5.2	3.9	5.9



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#12
1,1-Dichloroethene
Concen: 51.243 ug/l
RT: 2.306 min Scan# 200
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 96 Resp: 107643

Ion	Ratio	Lower	Upper
96	100		
61	158.8	131.1	196.7
98	64.3	51.2	76.8






#13
Acrolein
Concen: 230.919 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 56 Resp: 6560


Ion	Ratio	Lower	Upper
56	100		
55	70.6	55.1	82.7



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#14
Allyl chloride
Concen: 48.232 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 41 Resp: 197270

Ion	Ratio	Lower	Upper
41	100		
39	68.5	54.9	82.3
76	35.6	28.4	42.6






#15
Acrylonitrile
Concen: 259.942 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 53 Resp: 30779


Ion	Ratio	Lower	Upper
53	100		
52	83.2	66.0	99.0
51	36.2	28.9	43.3



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#16
Acetone
Concen: 266.735 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 43 Resp: 285777

Ion	Ratio	Lower	Upper
43	100		
58	31.0	24.6	37.0





#17
Carbon Disulfide
Concen: 50.827 ug/l
RT: 2.495 min Scan# 2
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15


Tgt Ion: 76 Resp: 29133


Ion	Ratio	Lower	Upper
76	100		
78	9.0	7.1	10.7

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#18
Methyl Acetate
Concen: 53.560 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 43 Resp: 177037

Ion	Ratio	Lower	Upper
43	100		
74	25.4	19.8	29.8






#19
Methyl tert-butyl Ether
Concen: 50.389 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 73 Resp: 38339


Ion	Ratio	Lower	Upper
73	100		
57	21.1	17.4	26.0



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#20
Methylene Chloride
Concen: 48.076 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 84 Resp: 119950

Ion	Ratio	Lower	Upper
84	100		
49	119.5	94.2	141.4
51	37.6	28.8	43.2
86	65.8	50.3	75.5






#21
trans-1,2-Dichloroethene
Concen: 48.846 ug/l
RT: 3.081 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 96 Resp: 10712


Ion	Ratio	Lower	Upper
96	100		
61	141.2	111.0	166.6
98	65.6	50.2	75.2



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#22
Diisopropyl ether
Concen: 50.659 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 45 Resp: 360169

Ion	Ratio	Lower	Upper
45	100		
43	67.3	61.0	91.6
87	29.5	22.7	34.1
59	11.6	9.4	14.2





#23
Vinyl Acetate
Concen: 254.442 ug/l
RT: 3.715 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15


Tgt Ion: 43 Resp: 158334


Ion	Ratio	Lower	Upper
43	100		
86	11.5	8.9	13.3

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#24
1,1-Dichloroethane
Concen: 50.150 ug/l
RT: 3.599 min Scan# 412
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 63 Resp: 211436

Ion	Ratio	Lower	Upper
63	100		
98	7.2	3.6	10.9
100	4.4	2.5	7.5





#25
2-Butanone
Concen: 265.596 ug/l
RT: 4.550 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15


Tgt Ion: 43 Resp: 42308


Ion	Ratio	Lower	Upper
43	100		
72	26.8	21.7	32.5

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#26
2,2-Dichloropropane
Concen: 48.697 ug/l
RT: 4.458 min Scan# 553
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 77 Resp: 198981

Ion	Ratio	Lower	Upper
77	100		
97	22.6	10.9	32.6






#27
cis-1,2-Dichloroethene
Concen: 48.831 ug/l
RT: 4.477 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 96 Resp: 13471


Ion	Ratio	Lower	Upper
96	100		
61	148.4	0.0	296.0
98	64.7	0.0	130.0



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#28
Bromochloromethane
Concen: 49.545 ug/l
RT: 4.885 min Scan# 623
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 49 Resp: 97608

Ion	Ratio	Lower	Upper
49	100		
129	2.2	0.0	4.4
130	80.4	63.4	95.2






#29
Tetrahydrofuran
Concen: 256.578 ug/l
RT: 4.995 min Scan# 64
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 42 Resp: 26363


Ion	Ratio	Lower	Upper
42	100		
72	48.3	38.2	57.4
71	44.3	35.3	52.9



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#30
Chloroform
Concen: 50.304 ug/l
RT: 5.080 min Scan# 655
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 83 Resp: 214034

Ion	Ratio	Lower	Upper
83	100		
85	63.5	51.2	76.8






#31
Cyclohexane
Concen: 52.878 ug/l
RT: 5.452 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 56 Resp: 18030


Ion	Ratio	Lower	Upper
56	100		
69	31.2	26.8	40.2
84	86.0	71.0	106.4



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#32
1,1,1-Trichloroethane
Concen: 48.859 ug/l
RT: 5.367 min Scan# 702
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 97 Resp: 191114

Ion	Ratio	Lower	Upper
97	100		
99	63.9	51.6	77.4
61	43.9	35.6	53.4






#33
1,2-Dichloroethane-d4
Concen: 52.363 ug/l
RT: 5.940 min Scan# 7
Delta R.T. -0.012 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 65 Resp: 14179


Ion	Ratio	Lower	Upper
65	100		
67	52.0	0.0	103.4



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.751 min Scan# 929
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:114 Resp: 323521

Ion	Ratio	Lower	Upper
114	100		
63	21.3	0.0	43.0
88	16.8	0.0	35.4






#35
Dibromofluoromethane
Concen: 53.683 ug/l
RT: 5.373 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:113 Resp: 11033


Ion	Ratio	Lower	Upper
113	100		
111	104.5	81.6	122.4
192	17.7	14.4	21.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#36
1,1-Dichloropropene
Concen: 50.294 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 75 Resp: 143891

Ion	Ratio	Lower	Upper
75	100		
110	35.5	17.9	53.8
77	30.2	24.5	36.7






#37
Ethyl Acetate
Concen: 54.526 ug/l
RT: 4.702 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 43 Resp: 16612


Ion	Ratio	Lower	Upper
43	100		
61	13.7	10.9	16.3
70	12.2	10.1	15.1



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#38
Carbon Tetrachloride
Concen: 50.287 ug/l
RT: 5.659 min Scan# 750
Delta R.T. -0.012 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:117 Resp: 161829

Ion	Ratio	Lower	Upper
117	100		
119	94.5	76.0	114.0
121	30.8	24.6	36.8






#39
Methylcyclohexane
Concen: 52.520 ug/l
RT: 7.366 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 83 Resp: 19947


Ion	Ratio	Lower	Upper
83	100		
55	77.8	60.5	90.7
98	48.9	39.0	58.4



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#40
Benzene
Concen: 50.610 ug/l
RT: 6.025 min Scan# 810
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 78 Resp: 448684

Ion	Ratio	Lower	Upper
78	100		
77	23.2	18.7	28.1






#41
Methacrylonitrile
Concen: 51.342 ug/l
RT: 4.910 min Scan# 6
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 41 Resp: 9115


Ion	Ratio	Lower	Upper
41	100		
39	56.2	44.4	66.6
67	75.5	60.6	90.8
52	30.4	24.3	36.5



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#42
1,2-Dichloroethane
Concen: 50.902 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 62 Resp: 164707

Ion	Ratio	Lower	Upper
62	100		
98	9.7	0.0	19.4





#43
Isopropyl Acetate
Concen: 50.871 ug/l
RT: 6.330 min Scan# 8
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15


Tgt Ion: 43 Resp: 28512


Ion	Ratio	Lower	Upper
43	100		
61	20.9	16.6	24.8
87	13.8	11.0	16.6

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#44
Trichloroethene
Concen: 51.887 ug/l
RT: 7.116 min Scan# 989
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:130 Resp: 110588

Ion	Ratio	Lower	Upper
130	100		
95	103.0	0.0	208.2






#45
1,2-Dichloropropane
Concen: 51.011 ug/l
RT: 7.421 min Scan# 1(
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 63 Resp: 11383(


Ion	Ratio	Lower	Upper
63	100		
65	31.6	24.2	36.2



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#46
Dibromomethane
Concen: 49.939 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 93 Resp: 84838

Ion	Ratio	Lower	Upper
93	100		
95	83.4	66.4	99.6
174	90.0	72.0	108.0






#47
Bromodichloromethane
Concen: 50.029 ug/l
RT: 7.811 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 83 Resp: 17429


Ion	Ratio	Lower	Upper
83	100		
85	61.7	50.1	75.1
127	7.7	6.5	9.7



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#48
Methyl methacrylate
Concen: 52.890 ug/l
RT: 7.683 min Scan# 1082
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 41 Resp: 137603

Ion	Ratio	Lower	Upper
41	100		
69	84.2	66.8	100.2
39	59.6	47.4	71.0






#49
1,4-Dioxane
Concen: 992.095 ug/l
RT: 7.665 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 88 Resp: 5385


Ion	Ratio	Lower	Upper
88	100		
43	35.2	27.0	40.4
58	72.4	55.8	83.8



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#50
Toluene-d8
Concen: 53.113 ug/l
RT: 8.641 min Scan# 1239
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 98 Resp: 380624

Ion	Ratio	Lower	Upper
98	100		
100	66.9	53.2	79.8





#51
4-Methyl-2-Pentanone
Concen: 251.778 ug/l
RT: 8.567 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15


Tgt Ion: 43 Resp: 77985

Ion	Ratio	Lower	Upper
43	100		
58	39.9	31.9	47.9

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#52
Toluene
Concen: 49.955 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 92 Resp: 271071


Ion	Ratio	Lower	Upper
92	100		
91	172.7	135.7	203.5



#53
t-1,3-Dichloropropene
Concen: 51.300 ug/l
RT: 8.970 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 75 Resp: 18465


Ion	Ratio	Lower	Upper
75	100		
77	31.3	24.4	36.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#54
cis-1,3-Dichloropropene
Concen: 51.049 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 75 Resp: 199240

Ion	Ratio	Lower	Upper
75	100		
77	31.3	25.1	37.7
39	51.9	40.3	60.5






#55
1,1,2-Trichloroethane
Concen: 50.085 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 97 Resp: 10995


Ion	Ratio	Lower	Upper
97	100		
83	85.6	68.8	103.2
85	56.2	43.7	65.5
99	63.6	49.5	74.3



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#56
Ethyl methacrylate
Concen: 50.595 ug/l
RT: 9.110 min Scan# 1316
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 69 Resp: 190986

Ion	Ratio	Lower	Upper
69	100		
41	66.0	53.2	79.8
39	38.4	30.2	45.2






#57
1,3-Dichloropropane
Concen: 50.129 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 76 Resp: 18740


Ion	Ratio	Lower	Upper
76	100		
78	32.9	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#58
2-Chloroethyl Vinyl ether
Concen: 251.852 ug/l
RT: 8.232 min Scan# 1172
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 63 Resp: 466737

Ion	Ratio	Lower	Upper
63	100		
106	28.9	23.0	34.6






#59
2-Hexanone
Concen: 253.010 ug/l
RT: 9.427 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 43 Resp: 576549


Ion	Ratio	Lower	Upper
43	100		
58	55.4	27.8	83.3



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#60
Dibromochloromethane
Concen: 49.957 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:129 Resp: 128727

Ion	Ratio	Lower	Upper
129	100		
127	77.2	38.5	115.5






#61
1,2-Dibromoethane
Concen: 49.693 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:107 Resp: 11433


Ion	Ratio	Lower	Upper
107	100		
109	94.5	75.3	112.9



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#62
4-Bromofluorobenzene
Concen: 52.693 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 95 Resp: 140770

Ion	Ratio	Lower	Upper
95	100		
174	67.6	0.0	133.0
176	67.1	0.0	129.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:117 Resp: 26687


Ion	Ratio	Lower	Upper
117	100		
82	59.1	47.6	71.4
119	32.1	25.4	38.2



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#64
Tetrachloroethene
Concen: 53.384 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:164 Resp: 89460

Ion	Ratio	Lower	Upper
164	100		
166	125.5	102.0	153.0
129	98.0	81.1	121.7
131	94.8	79.0	118.6






#65
Chlorobenzene
Concen: 50.667 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:112 Resp: 28720


Ion	Ratio	Lower	Upper
112	100		
114	32.9	25.3	37.9



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 51.129 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:131 Resp: 105329

Ion	Ratio	Lower	Upper
131	100		
133	94.8	47.8	143.3
119	66.5	33.4	100.2






#67
Ethyl Benzene
Concen: 51.721 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 91 Resp: 51156


Ion	Ratio	Lower	Upper
91	100		
106	30.3	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#68
m/p-Xylenes
Concen: 102.645 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:106 Resp: 376743

Ion	Ratio	Lower	Upper
106	100		
91	207.7	166.4	249.6






#69
o-Xylene
Concen: 50.378 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:106 Resp: 18699


Ion	Ratio	Lower	Upper
106	100		
91	219.3	108.7	326.3



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#70
Styrene
Concen: 51.398 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:104 Resp: 308018

Ion	Ratio	Lower	Upper
104	100		
78	53.4	43.0	64.6
103	54.4	43.4	65.2






#71
Bromoform
Concen: 51.597 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:173 Resp: 8651


Ion	Ratio	Lower	Upper
173	100		
175	47.9	23.6	70.8
254	0.0	0.1	0.1#



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:152 Resp: 115489

Ion	Ratio	Lower	Upper
152	100		
115	87.0	43.8	131.4
150	171.8	0.0	347.8






#73
Isopropylbenzene
Concen: 50.246 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:105 Resp: 47405


Ion	Ratio	Lower	Upper
105	100		
120	26.2	13.2	39.6



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#74
N-amyl acetate
Concen: 50.177 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 43 Resp: 233338

Ion	Ratio	Lower	Upper
43	100		
70	43.2	35.2	52.8
55	23.6	19.0	28.4
61	25.1	20.0	30.0






#75
1,1,2,2-Tetrachloroethane
Concen: 49.002 ug/l
RT: 11.213 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 83 Resp: 15967


Ion	Ratio	Lower	Upper
83	100		
131	9.9	5.0	14.9
85	66.2	32.0	96.2



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#76
1,2,3-Trichloropropane
Concen: 50.004 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 75 Resp: 140149

Ion	Ratio	Lower	Upper
75	100		
77	44.9	23.4	70.2






#77
Bromobenzene
Concen: 49.506 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:156 Resp: 10917


Ion	Ratio	Lower	Upper
156	100		
77	178.2	90.3	271.0
158	97.0	48.4	145.3



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#78
n-propylbenzene
Concen: 51.712 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 91 Resp: 538048

Ion	Ratio	Lower	Upper
91	100		
120	22.7	11.4	34.1





#79
2-Chlorotoluene
Concen: 49.678 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15


Tgt Ion: 91 Resp: 33389


Ion	Ratio	Lower	Upper
91	100		
126	32.3	16.2	48.5

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025







#80
1,3,5-Trimethylbenzene
Concen: 49.852 ug/l
RT: 11.445 min Scan# 1699
Delta R.T. -0.006 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:105 Resp: 382880

Ion	Ratio	Lower	Upper
105	100		
120	47.6	23.7	71.1






#81
trans-1,4-Dichloro-2-butene
Concen: 51.981 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 75 Resp: 6370


Ion	Ratio	Lower	Upper
75	100		
53	93.1	74.3	111.5
89	47.1	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#82
4-Chlorotoluene
Concen: 50.750 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 91 Resp: 373655

Ion	Ratio	Lower	Upper
91	100		
126	28.1	13.9	41.7






#83
tert-Butylbenzene
Concen: 49.459 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:119 Resp: 39719


Ion	Ratio	Lower	Upper
119	100		
91	60.8	30.0	90.0
134	22.4	11.2	33.5



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#84
1,2,4-Trimethylbenzene
Concen: 49.714 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:105 Resp: 388010

Ion	Ratio	Lower	Upper
105	100		
120	44.2	22.2	66.6






#85
sec-Butylbenzene
Concen: 51.076 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:105 Resp: 47759


Ion	Ratio	Lower	Upper
105	100		
134	19.8	9.8	29.5



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#86
p-Isopropyltoluene
Concen: 51.028 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:119 Resp: 392123

Ion	Ratio	Lower	Upper
119	100		
134	26.0	12.8	38.4
91	25.6	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 51.476 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:146 Resp: 19686


Ion	Ratio	Lower	Upper
146	100		
111	43.4	21.9	65.7
148	62.9	31.8	95.4



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#88
1,4-Dichlorobenzene
Concen: 50.011 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:146 Resp: 193343

Ion	Ratio	Lower	Upper
146	100		
111	42.3	21.5	64.5
148	63.7	32.1	96.5






#89
n-Butylbenzene
Concen: 53.136 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 91 Resp: 35070


Ion	Ratio	Lower	Upper
91	100		
92	55.0	27.4	82.0
134	25.1	12.8	38.3



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#90
Hexachloroethane
Concen: 48.518 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:117 Resp: 84427

Ion	Ratio	Lower	Upper
117	100		
201	64.4	32.3	96.8






#91
1,2-Dichlorobenzene
Concen: 49.430 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:146 Resp: 19413


Ion	Ratio	Lower	Upper
146	100		
111	45.4	22.4	67.2
148	64.9	31.6	94.8



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 49.672 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion: 75 Resp: 39583

Ion	Ratio	Lower	Upper
75	100		
155	76.9	39.0	116.9
157	99.4	49.9	149.7






#93
1,2,4-Trichlorobenzene
Concen: 53.362 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:180 Resp: 12528


Ion	Ratio	Lower	Upper
180	100		
182	94.5	47.4	142.2
145	34.3	16.7	50.1



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025




#94
Hexachlorobutadiene
Concen: 52.807 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:225 Resp: 48702

Ion	Ratio	Lower	Upper
225	100		
223	62.9	31.9	95.7
227	63.0	31.8	95.4






#95
Naphthalene
Concen: 52.899 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:128 Resp: 47902


Ion	Ratio	Lower	Upper
128	100		
127	12.7	10.3	15.5
129	10.9	8.7	13.1



Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 01/16/2025
Supervised By :Semsettin Yesilyurt 01/16/2025



#96
1,2,3-Trichlorobenzene
Concen: 52.611 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX044655.D
Acq: 15 Jan 2025 15:15

Tgt Ion:180 Resp: 128038

Ion	Ratio	Lower	Upper
180	100		
182	93.8	47.2	141.6
145	35.2	17.5	52.6



Evaluate Continuing Calibration Report

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044655.D
Acq On : 15 Jan 2025 15:15
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Quant Time: Jan 16 01:21:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2	T	Dichlorodifluoromethane	0.673	0.761	-13.1	105	0.00
3	P	Chloromethane	0.728	0.695	4.5	92	0.00
4	C	Vinyl Chloride	0.719	0.717	0.3#	96	0.00
5	T	Bromomethane	0.172	0.172	0.0	95	0.00
6	T	Chloroethane	0.132	0.131	0.8	105	0.00
7	T	Trichlorofluoromethane	0.795	0.823	-3.5	101	0.00
8	T	Diethyl Ether	0.321	0.304	5.3	98	0.00
9	T	1,1,2-Trichlorotrifluoroeth	0.597	0.624	-4.5	106	0.00
10	T	Methyl Iodide	0.815	0.837	-2.7	97	0.00
11	T	Tert butyl alcohol	0.168	0.167	0.6	104	0.00
12	CM	1,1-Dichloroethene	0.585	0.600	-2.6#	100	0.00
13	T	Acrolein	0.079	0.073	7.6	101	0.00
14	T	Allyl chloride	1.140	1.100	3.5	98	0.00
15	T	Acrylonitrile	0.330	0.343	-3.9	103	0.00
16	T	Acetone	0.299	0.319	-6.7	112	0.00
17	T	Carbon Disulfide	1.597	1.624	-1.7	99	0.00
18	T	Methyl Acetate	0.921	0.987	-7.2	105	0.00
19	T	Methyl tert-butyl Ether	2.121	2.137	-0.8	99	0.00
20	T	Methylene Chloride	0.695	0.669	3.7	97	0.00
21	T	trans-1,2-Dichloroethene	0.611	0.597	2.3	99	0.00
22	T	Diisopropyl ether	1.981	2.008	-1.4	99	0.00
23	T	Vinyl Acetate	1.734	1.765	-1.8	100	0.00
24	P	1,1-Dichloroethane	1.175	1.179	-0.3	98	0.00
25	T	2-Butanone	0.444	0.472	-6.3	104	0.00
26	T	2,2-Dichloropropane	1.139	1.109	2.6	99	0.00
27	T	cis-1,2-Dichloroethene	0.769	0.751	2.3	98	0.00
28	T	Bromochloromethane	0.549	0.544	0.9	96	0.00
29	T	Tetrahydrofuran	0.286	0.294	-2.8	103	0.00
30	C	Chloroform	1.186	1.193	-0.6#	98	0.00
31	T	Cyclohexane	0.950	1.005	-5.8	101	0.00
32	T	1,1,1-Trichloroethane	1.090	1.065	2.3	100	0.00
33	S	1,2-Dichloroethane-d4	0.755	0.790	-4.6	100	-0.01
34	I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35	S	Dibromofluoromethane	0.318	0.341	-7.2	99	0.00
36	T	1,1-Dichloropropene	0.442	0.445	-0.7	100	0.00
37	T	Ethyl Acetate	0.471	0.513	-8.9	102	0.00
38	T	Carbon Tetrachloride	0.497	0.500	-0.6	100	-0.01
39	T	Methylcyclohexane	0.587	0.617	-5.1	103	0.00
40	TM	Benzene	1.370	1.387	-1.2	98	0.00
41	T	Methacrylonitrile	0.274	0.282	-2.9	102	0.00
42	TM	1,2-Dichloroethane	0.500	0.509	-1.8	98	0.00
43	T	Isopropyl Acetate	0.866	0.881	-1.7	99	0.00
44	TM	Trichloroethene	0.329	0.342	-4.0	100	0.00
45	C	1,2-Dichloropropane	0.345	0.352	-2.0#	98	0.00
46	T	Dibromomethane	0.263	0.262	0.4	98	0.00
47	T	Bromodichloromethane	0.538	0.539	-0.2	98	0.00
48	T	Methyl methacrylate	0.402	0.425	-5.7	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044655.D
Acq On : 15 Jan 2025 15:15
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Quant Time: Jan 16 01:21:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	0.008	0.008	0.0	97	0.00
50	S	Toluene-d8	1.108	1.177	-6.2	97	0.00
51	T	4-Methyl-2-Pentanone	0.479	0.482	-0.6	99	0.00
52	CM	Toluene	0.839	0.838	0.1#	96	0.00
53	T	t-1,3-Dichloropropene	0.556	0.571	-2.7	96	0.00
54	T	cis-1,3-Dichloropropene	0.603	0.616	-2.2	99	0.00
55	T	1,1,2-Trichloroethane	0.339	0.340	-0.3	98	0.00
56	T	Ethyl methacrylate	0.583	0.590	-1.2	96	0.00
57	T	1,3-Dichloropropane	0.578	0.579	-0.2	96	0.00
58	T	2-Chloroethyl Vinyl ether	0.286	0.289	-1.0	97	0.00
59	T	2-Hexanone	0.352	0.356	-1.1	99	0.00
60	T	Dibromochloromethane	0.398	0.398	0.0	96	0.00
61	T	1,2-Dibromoethane	0.356	0.353	0.8	97	0.00
62	S	4-Bromofluorobenzene	0.413	0.435	-5.3	95	0.00
63	I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64	T	Tetrachloroethene	0.314	0.335	-6.7	100	0.00
65	PM	Chlorobenzene	1.062	1.076	-1.3	95	0.00
66	T	1,1,1,2-Tetrachloroethane	0.386	0.395	-2.3	95	0.00
67	C	Ethyl Benzene	1.853	1.917	-3.5#	95	0.00
68	T	m/p-Xylenes	0.688	0.706	-2.6	95	0.00
69	T	o-Xylene	0.695	0.701	-0.9	93	0.00
70	T	Styrene	1.123	1.154	-2.8	94	0.00
71	P	Bromoform	0.314	0.324	-3.2	95	0.00
72	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73	T	Isopropylbenzene	4.085	4.105	-0.5	95	0.00
74	T	N-amyl acetate	2.013	2.020	-0.3	96	0.00
75	P	1,1,2,2-Tetrachloroethane	1.411	1.383	2.0	96	0.00
76	T	1,2,3-Trichloropropane	1.213	1.214	-0.1	100	0.00
77	T	Bromobenzene	0.955	0.945	1.0	95	0.00
78	T	n-propylbenzene	4.505	4.659	-3.4	96	0.00
79	T	2-Chlorotoluene	2.910	2.891	0.7	95	0.00
80	T	1,3,5-Trimethylbenzene	3.325	3.315	0.3	94	0.00
81	T	trans-1,4-Dichloro-2-butene	0.531	0.552	-4.0	97	0.00
82	T	4-Chlorotoluene	3.188	3.235	-1.5	95	0.00
83	T	tert-Butylbenzene	3.477	3.439	1.1	94	0.00
84	T	1,2,4-Trimethylbenzene	3.379	3.360	0.6	95	0.00
85	T	sec-Butylbenzene	4.048	4.135	-2.1	96	0.00
86	T	p-Isopropyltoluene	3.327	3.395	-2.0	95	0.00
87	T	1,3-Dichlorobenzene	1.656	1.705	-3.0	95	0.00
88	T	1,4-Dichlorobenzene	1.674	1.674	0.0	96	0.00
89	T	n-Butylbenzene	2.857	3.037	-6.3	98	0.00
90	T	Hexachloroethane	0.753	0.731	2.9	96	0.00
91	T	1,2-Dichlorobenzene	1.700	1.681	1.1	93	0.00
92	T	1,2-Dibromo-3-Chloropropane	0.345	0.343	0.6	100	0.00
93	T	1,2,4-Trichlorobenzene	1.016	1.085	-6.8	95	0.00
94	T	Hexachlorobutadiene	0.399	0.422	-5.8	97	0.00
95	T	Naphthalene	3.920	4.148	-5.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044655.D
Acq On : 15 Jan 2025 15:15
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Quant Time: Jan 16 01:21:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.054	1.109	-5.2	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Evaluate Continuing Calibration Report

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044655.D
Acq On : 15 Jan 2025 15:15
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Quant Time: Jan 16 01:21:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	50.000	50.000	0.0	94	0.00
2	T	Dichlorodifluoromethane	50.000	56.549	-13.1	105	0.00
3	P	Chloromethane	50.000	47.763	4.5	92	0.00
4	C	Vinyl Chloride	50.000	49.839	0.3#	96	0.00
5	T	Bromomethane	50.000	50.031	-0.1	95	0.00
6	T	Chloroethane	50.000	49.624	0.8	105	0.00
7	T	Trichlorofluoromethane	50.000	51.789	-3.6	101	0.00
8	T	Diethyl Ether	50.000	47.479	5.0	98	0.00
9	T	1,1,2-Trichlorotrifluoroeth	50.000	52.271	-4.5	106	0.00
10	T	Methyl Iodide	50.000	51.350	-2.7	97	0.00
11	T	Tert butyl alcohol	250.000	248.503	0.6	104	0.00
12	CM	1,1-Dichloroethene	50.000	51.243	-2.5#	100	0.00
13	T	Acrolein	250.000	230.919	7.6	101	0.00
14	T	Allyl chloride	50.000	48.232	3.5	98	0.00
15	T	Acrylonitrile	250.000	259.942	-4.0	103	0.00
16	T	Acetone	250.000	266.735	-6.7	112	0.00
17	T	Carbon Disulfide	50.000	50.827	-1.7	99	0.00
18	T	Methyl Acetate	50.000	53.560	-7.1	105	0.00
19	T	Methyl tert-butyl Ether	50.000	50.389	-0.8	99	0.00
20	T	Methylene Chloride	50.000	48.076	3.8	97	0.00
21	T	trans-1,2-Dichloroethene	50.000	48.846	2.3	99	0.00
22	T	Diisopropyl ether	50.000	50.659	-1.3	99	0.00
23	T	Vinyl Acetate	250.000	254.442	-1.8	100	0.00
24	P	1,1-Dichloroethane	50.000	50.150	-0.3	98	0.00
25	T	2-Butanone	250.000	265.596	-6.2	104	0.00
26	T	2,2-Dichloropropane	50.000	48.697	2.6	99	0.00
27	T	cis-1,2-Dichloroethene	50.000	48.831	2.3	98	0.00
28	T	Bromochloromethane	50.000	49.545	0.9	96	0.00
29	T	Tetrahydrofuran	250.000	256.578	-2.6	103	0.00
30	C	Chloroform	50.000	50.304	-0.6#	98	0.00
31	T	Cyclohexane	50.000	52.878	-5.8	101	0.00
32	T	1,1,1-Trichloroethane	50.000	48.859	2.3	100	0.00
33	S	1,2-Dichloroethane-d4	50.000	52.363	-4.7	100	-0.01
34	I	1,4-Difluorobenzene	50.000	50.000	0.0	92	0.00
35	S	Dibromofluoromethane	50.000	53.683	-7.4	99	0.00
36	T	1,1-Dichloropropene	50.000	50.294	-0.6	100	0.00
37	T	Ethyl Acetate	50.000	54.526	-9.1	102	0.00
38	T	Carbon Tetrachloride	50.000	50.287	-0.6	100	-0.01
39	T	Methylcyclohexane	50.000	52.520	-5.0	103	0.00
40	TM	Benzene	50.000	50.610	-1.2	98	0.00
41	T	Methacrylonitrile	50.000	51.342	-2.7	102	0.00
42	TM	1,2-Dichloroethane	50.000	50.902	-1.8	98	0.00
43	T	Isopropyl Acetate	50.000	50.871	-1.7	99	0.00
44	TM	Trichloroethene	50.000	51.887	-3.8	100	0.00
45	C	1,2-Dichloropropane	50.000	51.011	-2.0#	98	0.00
46	T	Dibromomethane	50.000	49.939	0.1	98	0.00
47	T	Bromodichloromethane	50.000	50.029	-0.1	98	0.00
48	T	Methyl methacrylate	50.000	52.890	-5.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044655.D
Acq On : 15 Jan 2025 15:15
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Quant Time: Jan 16 01:21:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	1000.000	992.095	0.8	97	0.00
50	S	Toluene-d8	50.000	53.113	-6.2	97	0.00
51	T	4-Methyl-2-Pentanone	250.000	251.778	-0.7	99	0.00
52	CM	Toluene	50.000	49.955	0.1#	96	0.00
53	T	t-1,3-Dichloropropene	50.000	51.300	-2.6	96	0.00
54	T	cis-1,3-Dichloropropene	50.000	51.049	-2.1	99	0.00
55	T	1,1,2-Trichloroethane	50.000	50.085	-0.2	98	0.00
56	T	Ethyl methacrylate	50.000	50.595	-1.2	96	0.00
57	T	1,3-Dichloropropane	50.000	50.129	-0.3	96	0.00
58	T	2-Chloroethyl Vinyl ether	250.000	251.852	-0.7	97	0.00
59	T	2-Hexanone	250.000	253.010	-1.2	99	0.00
60	T	Dibromochloromethane	50.000	49.957	0.1	96	0.00
61	T	1,2-Dibromoethane	50.000	49.693	0.6	97	0.00
62	S	4-Bromofluorobenzene	50.000	52.693	-5.4	95	0.00
63	I	Chlorobenzene-d5	50.000	50.000	0.0	89	0.00
64	T	Tetrachloroethene	50.000	53.384	-6.8	100	0.00
65	PM	Chlorobenzene	50.000	50.667	-1.3	95	0.00
66	T	1,1,1,2-Tetrachloroethane	50.000	51.129	-2.3	95	0.00
67	C	Ethyl Benzene	50.000	51.721	-3.4#	95	0.00
68	T	m/p-Xylenes	100.000	102.645	-2.6	95	0.00
69	T	o-Xylene	50.000	50.378	-0.8	93	0.00
70	T	Styrene	50.000	51.398	-2.8	94	0.00
71	P	Bromoform	50.000	51.597	-3.2	95	0.00
72	I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	90	0.00
73	T	Isopropylbenzene	50.000	50.246	-0.5	95	0.00
74	T	N-amyl acetate	50.000	50.177	-0.4	96	0.00
75	P	1,1,2,2-Tetrachloroethane	50.000	49.002	2.0	96	0.00
76	T	1,2,3-Trichloropropane	50.000	50.004	-0.0	100	0.00
77	T	Bromobenzene	50.000	49.506	1.0	95	0.00
78	T	n-propylbenzene	50.000	51.712	-3.4	96	0.00
79	T	2-Chlorotoluene	50.000	49.678	0.6	95	0.00
80	T	1,3,5-Trimethylbenzene	50.000	49.852	0.3	94	0.00
81	T	trans-1,4-Dichloro-2-butene	50.000	51.981	-4.0	97	0.00
82	T	4-Chlorotoluene	50.000	50.750	-1.5	95	0.00
83	T	tert-Butylbenzene	50.000	49.459	1.1	94	0.00
84	T	1,2,4-Trimethylbenzene	50.000	49.714	0.6	95	0.00
85	T	sec-Butylbenzene	50.000	51.076	-2.2	96	0.00
86	T	p-Isopropyltoluene	50.000	51.028	-2.1	95	0.00
87	T	1,3-Dichlorobenzene	50.000	51.476	-3.0	95	0.00
88	T	1,4-Dichlorobenzene	50.000	50.011	-0.0	96	0.00
89	T	n-Butylbenzene	50.000	53.136	-6.3	98	0.00
90	T	Hexachloroethane	50.000	48.518	3.0	96	0.00
91	T	1,2-Dichlorobenzene	50.000	49.430	1.1	93	0.00
92	T	1,2-Dibromo-3-Chloropropane	50.000	49.672	0.7	100	0.00
93	T	1,2,4-Trichlorobenzene	50.000	53.362	-6.7	95	0.00
94	T	Hexachlorobutadiene	50.000	52.807	-5.6	97	0.00
95	T	Naphthalene	50.000	52.899	-5.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044655.D
Acq On : 15 Jan 2025 15:15
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011525

Quant Time: Jan 16 01:21:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.611	-5.2	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG No.: Q1215

Instrument ID: MSVOA_N Calibration Date/Time: 01/31/2025 10:04

Lab File ID: VN085591.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.737	0.631		-14.38	20
1,1-Dichloroethene	0.537	0.497		-7.45	20
2-Butanone	0.384	0.389		1.3	20
Carbon Tetrachloride	0.557	0.576		3.41	20
Chloroform	1.218	1.221		0.25	20
Benzene	1.463	1.491		1.91	20
1,2-Dichloroethane	0.551	0.571		3.63	20
Trichloroethene	0.341	0.337		-1.17	20
Tetrachloroethene	0.341	0.357		4.69	20
Chlorobenzene	1.095	1.091	0.3	-0.37	20
1,2-Dichloroethane-d4	0.807	0.801		-0.74	20
Dibromofluoromethane	0.347	0.374		7.78	20
Toluene-d8	1.232	1.332		8.12	20
4-Bromofluorobenzene	0.422	0.473		12.09	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085591.D
Acq On : 31 Jan 2025 10:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jan 31 22:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) Pentafluorobenzene	8.224	168	190291	50.000 ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	311950	50.000 ug/l	0.00
63) Chlorobenzene-d5	11.865	117	284481	50.000 ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	145646	50.000 ug/l	0.00
System Monitoring Compounds					
33) 1,2-Dichloroethane-d4	8.577	65	152459	49.634 ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 99.260%		
35) Dibromofluoromethane	8.165	113	116708	53.927 ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.860%		
50) Toluene-d8	10.565	98	415377	54.021 ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 108.040%		
62) 4-Bromofluorobenzene	12.847	95	147501	56.078 ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 112.160%		
Target Compounds					Qvalue
2) Dichlorodifluoromethane	2.124	85	103770	40.276 ug/l	99
3) Chloromethane	2.359	50	110719	39.689 ug/l	99
4) Vinyl Chloride	2.512	62	120110	42.836 ug/l	97
5) Bromomethane	2.948	94	72429	42.763 ug/l	98
6) Chloroethane	3.112	64	81085	45.613 ug/l	99
7) Trichlorofluoromethane	3.501	101	191046	46.958 ug/l	99
8) Diethyl Ether	3.953	74	61086	43.462 ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.365	101	110162	48.073 ug/l	99
10) Methyl Iodide	4.589	142	122885	46.827 ug/l	96
11) Tert butyl alcohol	5.506	59	102478	291.355 ug/l	100
12) 1,1-Dichloroethene	4.336	96	94663	46.352 ug/l	98
13) Acrolein	4.183	56	90056	187.561 ug/l	97
14) Allyl chloride	5.024	41	143367	43.262 ug/l	98
15) Acrylonitrile	5.712	53	279267	249.970 ug/l	99
16) Acetone	4.424	43	238111	239.913 ug/l	99
17) Carbon Disulfide	4.706	76	231929	36.884 ug/l	100
18) Methyl Acetate	5.018	43	163080	54.036 ug/l	98
19) Methyl tert-butyl Ether	5.789	73	343910	51.861 ug/l	97
20) Methylene Chloride	5.271	84	117510	47.827 ug/l	98
21) trans-1,2-Dichloroethene	5.783	96	102553	46.987 ug/l	100
22) Diisopropyl ether	6.665	45	379410	51.582 ug/l	97
23) Vinyl Acetate	6.600	43	1337620	259.853 ug/l	100
24) 1,1-Dichloroethane	6.565	63	220300	49.119 ug/l	98
25) 2-Butanone	7.477	43	370535	253.698 ug/l	100
26) 2,2-Dichloropropane	7.489	77	199084	54.902 ug/l	100
27) cis-1,2-Dichloroethene	7.483	96	126248	49.110 ug/l	98
28) Bromochloromethane	7.806	49	98995	47.443 ug/l	97
29) Tetrahydrofuran	7.836	42	246967	266.707 ug/l	99
30) Chloroform	7.959	83	232358	50.126 ug/l	100
31) Cyclohexane	8.253	56	161397	41.402 ug/l	93
32) 1,1,1-Trichloroethane	8.165	97	201948	49.666 ug/l	97
36) 1,1-Dichloropropene	8.365	75	148837	49.002 ug/l	99
37) Ethyl Acetate	7.553	43	151771	49.505 ug/l	97
38) Carbon Tetrachloride	8.359	117	179763	51.696 ug/l	98
39) Methylcyclohexane	9.600	83	136843	47.944 ug/l	98
40) Benzene	8.600	78	465224	50.976 ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085591.D
Acq On : 31 Jan 2025 10:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jan 31 22:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	7.771	41	85378	53.528	ug/l		97
42) 1,2-Dichloroethane	8.665	62	178224	51.851	ug/l		99
43) Isopropyl Acetate	8.683	43	256966	52.309	ug/l		99
44) Trichloroethene	9.347	130	105127	49.486	ug/l		97
45) 1,2-Dichloropropane	9.618	63	120938	51.880	ug/l		98
46) Dibromomethane	9.706	93	85588	50.884	ug/l		97
47) Bromodichloromethane	9.882	83	185704	54.191	ug/l		99
48) Methyl methacrylate	9.677	41	122530	55.425	ug/l		99
49) 1,4-Dioxane	9.688	88	40300	1082.516	ug/l		99
51) 4-Methyl-2-Pentanone	10.441	43	813239	285.284	ug/l		99
52) Toluene	10.630	92	283889	53.686	ug/l		99
53) t-1,3-Dichloropropene	10.835	75	175773	54.277	ug/l		95
54) cis-1,3-Dichloropropene	10.312	75	188416	54.468	ug/l		99
55) 1,1,2-Trichloroethane	11.012	97	113662	54.314	ug/l		98
56) Ethyl methacrylate	10.871	69	168851	47.979	ug/l		99
57) 1,3-Dichloropropane	11.159	76	196125	53.898	ug/l		100
58) 2-Chloroethyl Vinyl ether	10.159	63	369534	278.261	ug/l		100
59) 2-Hexanone	11.194	43	586256	292.282	ug/l		100
60) Dibromochloromethane	11.359	129	139102	55.058	ug/l		100
61) 1,2-Dibromoethane	11.465	107	107453	51.584	ug/l		99
64) Tetrachloroethene	11.100	164	101650	52.413	ug/l		98
65) Chlorobenzene	11.888	112	310267	49.786	ug/l		99
66) 1,1,1,2-Tetrachloroethane	11.959	131	116111	50.764	ug/l		97
67) Ethyl Benzene	11.959	91	539837	53.184	ug/l		100
68) m/p-Xylenes	12.065	106	417283	111.234	ug/l		99
69) o-Xylene	12.394	106	196000	54.671	ug/l		99
70) Styrene	12.406	104	344358	58.039	ug/l		99
71) Bromoform	12.576	173	90428	55.254	ug/l	#	99
73) Isopropylbenzene	12.694	105	498892	50.753	ug/l		99
74) N-amyl acetate	12.488	43	218240	49.394	ug/l		99
75) 1,1,2,2-Tetrachloroethane	12.935	83	164103	47.260	ug/l		100
76) 1,2,3-Trichloropropane	12.988	75	156230m	52.775	ug/l		
77) Bromobenzene	12.976	156	124316	48.406	ug/l		99
78) n-propylbenzene	13.035	91	617823	53.092	ug/l		100
79) 2-Chlorotoluene	13.123	91	376644	50.010	ug/l		98
80) 1,3,5-Trimethylbenzene	13.171	105	435734	53.665	ug/l		100
81) trans-1,4-Dichloro-2-b...	12.735	75	58483	53.451	ug/l		96
82) 4-Chlorotoluene	13.218	91	387664	51.696	ug/l		99
83) tert-Butylbenzene	13.435	119	353864	51.897	ug/l		99
84) 1,2,4-Trimethylbenzene	13.482	105	442366	54.663	ug/l		99
85) sec-Butylbenzene	13.612	105	515181	54.515	ug/l		100
86) p-Isopropyltoluene	13.723	119	426531	50.108	ug/l		100
87) 1,3-Dichlorobenzene	13.729	146	241411	51.042	ug/l		98
88) 1,4-Dichlorobenzene	13.812	146	233405	47.611	ug/l		99
89) n-Butylbenzene	14.053	91	364737	53.798	ug/l		100
90) Hexachloroethane	14.329	117	86465	48.061	ug/l		99
91) 1,2-Dichlorobenzene	14.100	146	232755	49.337	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	30775	48.536	ug/l		98
93) 1,2,4-Trichlorobenzene	15.388	180	106670	48.649	ug/l		97
94) Hexachlorobutadiene	15.500	225	56304	48.367	ug/l		99
95) Naphthalene	15.635	128	315500	48.220	ug/l		99
96) 1,2,3-Trichlorobenzene	15.835	180	107609	48.503	ug/l		99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085591.D
Acq On : 31 Jan 2025 10:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jan 31 22:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085591.D
Acq On : 31 Jan 2025 10:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1


Quant Time: Jan 31 22:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:168 Resp: 190291


Ion	Ratio	Lower	Upper
168	100		
99	59.3	53.6	80.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#2
Dichlorodifluoromethane
Concen: 40.276 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 85 Resp: 103770

Ion	Ratio	Lower	Upper
85	100		
87	32.9	16.7	50.1






#3
Chloromethane
Concen: 39.689 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 50 Resp: 110719


Ion	Ratio	Lower	Upper
50	100		
52	34.1	27.0	40.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#4
Vinyl Chloride
Concen: 42.836 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 62 Resp: 120110

Ion	Ratio	Lower	Upper
62	100		
64	32.8	24.8	37.2






#5
Bromomethane
Concen: 42.763 ug/l
RT: 2.948 min Scan# 16
Delta R.T. -0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 94 Resp: 72429


Ion	Ratio	Lower	Upper
94	100		
96	87.7	71.8	107.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#6
Chloroethane
Concen: 45.613 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 64 Resp: 81085

Ion	Ratio	Lower	Upper
64	100		
66	30.1	24.6	36.8






#7
Trichlorofluoromethane
Concen: 46.958 ug/l
RT: 3.501 min Scan# 26
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:101 Resp: 191046


Ion	Ratio	Lower	Upper
101	100		
103	63.8	51.4	77.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#8
Diethyl Ether
Concen: 43.462 ug/l
RT: 3.953 min Scan# 340
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 74 Resp: 61086

Ion	Ratio	Lower	Upper
74	100		
45	101.0	49.7	149.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.073 ug/l
RT: 4.365 min Scan# 41
Delta R.T. -0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04


Tgt Ion:101 Resp: 110162


Ion	Ratio	Lower	Upper
101	100		
85	45.3	37.8	56.8
151	73.0	58.8	88.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025







#10
Methyl Iodide
Concen: 46.827 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:142 Resp: 122885

Ion	Ratio	Lower	Upper
142	100		
127	53.7	40.3	60.5
141	14.5	12.8	19.2






#11
Tert butyl alcohol
Concen: 291.355 ug/l
RT: 5.506 min Scan# 60
Delta R.T. -0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 59 Resp: 102478


Ion	Ratio	Lower	Upper
59	100		
57	10.7	8.6	13.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#12
1,1-Dichloroethene
Concen: 46.352 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 96 Resp: 94663

Ion	Ratio	Lower	Upper
96	100		
61	172.6	141.0	211.4
98	60.2	49.0	73.6






#13
Acrolein
Concen: 187.561 ug/l
RT: 4.183 min Scan# 37
Delta R.T. 0.012 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 56 Resp: 90056


Ion	Ratio	Lower	Upper
56	100		
55	73.2	56.3	84.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#14
Allyl chloride
Concen: 43.262 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 41 Resp: 143367

Ion	Ratio	Lower	Upper
41	100		
39	82.9	64.4	96.6
76	37.9	30.5	45.7






#15
Acrylonitrile
Concen: 249.970 ug/l
RT: 5.712 min Scan# 63
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 53 Resp: 279267


Ion	Ratio	Lower	Upper
53	100		
52	82.7	65.5	98.3
51	38.1	29.8	44.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#16
Acetone
Concen: 239.913 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 43 Resp: 238111

Ion	Ratio	Lower	Upper
43	100		
58	29.0	23.8	35.6






#17
Carbon Disulfide
Concen: 36.884 ug/l
RT: 4.706 min Scan# 46
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 76 Resp: 231929


Ion	Ratio	Lower	Upper
76	100		
78	8.5	6.9	10.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#18
Methyl Acetate
Concen: 54.036 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 43 Resp: 163080

Ion	Ratio	Lower	Upper
43	100		
74	22.6	18.6	28.0






#19
Methyl tert-butyl Ether
Concen: 51.861 ug/l
RT: 5.789 min Scan# 65
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 73 Resp: 343910


Ion	Ratio	Lower	Upper
73	100		
57	20.9	17.8	26.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#20
Methylene Chloride
Concen: 47.827 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 84 Resp: 117510

Ion	Ratio	Lower	Upper
84	100		
49	133.1	104.1	156.1
51	38.9	31.0	46.4
86	62.8	51.9	77.9






#21
trans-1,2-Dichloroethene
Concen: 46.987 ug/l
RT: 5.783 min Scan# 65
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 96 Resp: 102553


Ion	Ratio	Lower	Upper
96	100		
61	144.3	115.7	173.5
98	62.9	49.8	74.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#22
Diisopropyl ether
Concen: 51.582 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 45 Resp: 379410

Ion	Ratio	Lower	Upper
45	100		
43	55.1	45.0	67.6
87	25.2	21.8	32.8
59	11.4	10.2	15.2






#23
Vinyl Acetate
Concen: 259.853 ug/l
RT: 6.600 min Scan# 79
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 43 Resp: 1337620


Ion	Ratio	Lower	Upper
43	100		
86	10.4	8.4	12.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#24
1,1-Dichloroethane
Concen: 49.119 ug/l
RT: 6.565 min Scan# 784
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 63 Resp: 220300

Ion	Ratio	Lower	Upper
63	100		
98	7.1	3.3	9.8
100	4.5	2.0	6.0






#25
2-Butanone
Concen: 253.698 ug/l
RT: 7.477 min Scan# 93
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 43 Resp: 370535


Ion	Ratio	Lower	Upper
43	100		
72	25.4	20.2	30.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#26
2,2-Dichloropropane
Concen: 54.902 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 77 Resp: 199084

Ion	Ratio	Lower	Upper
77	100		
97	21.2	10.7	32.1






#27
cis-1,2-Dichloroethene
Concen: 49.110 ug/l
RT: 7.483 min Scan# 94
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 96 Resp: 126248


Ion	Ratio	Lower	Upper
96	100		
61	158.7	0.0	311.8
98	64.9	0.0	126.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#28
Bromochloromethane
Concen: 47.443 ug/l
RT: 7.806 min Scan# 995
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 49 Resp: 98995

Ion	Ratio	Lower	Upper
49	100		
129	1.7	0.0	4.4
130	70.9	55.0	82.4






#29
Tetrahydrofuran
Concen: 266.707 ug/l
RT: 7.836 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 42 Resp: 246967


Ion	Ratio	Lower	Upper
42	100		
72	42.1	34.2	51.2
71	39.9	32.5	48.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#30
Chloroform
Concen: 50.126 ug/l
RT: 7.959 min Scan# 1021
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 83 Resp: 232358

Ion	Ratio	Lower	Upper
83	100		
85	64.6	51.8	77.6






#31
Cyclohexane
Concen: 41.402 ug/l
RT: 8.253 min Scan# 10
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 56 Resp: 161397


Ion	Ratio	Lower	Upper
56	100		
69	31.7	22.2	33.4
84	89.1	66.4	99.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#32
1,1,1-Trichloroethane
Concen: 49.666 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 97 Resp: 201948

Ion	Ratio	Lower	Upper
97	100		
99	65.3	49.8	74.6
61	52.3	41.4	62.2






#33
1,2-Dichloroethane-d4
Concen: 49.634 ug/l
RT: 8.577 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 65 Resp: 152459


Ion	Ratio	Lower	Upper
65	100		
67	52.4	0.0	101.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:114 Resp: 311950

Ion	Ratio	Lower	Upper
114	100		
63	22.1	0.0	47.6
88	15.7	0.0	32.6






#35
Dibromofluoromethane
Concen: 53.927 ug/l
RT: 8.165 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:113 Resp: 116708


Ion	Ratio	Lower	Upper
113	100		
111	103.4	82.7	124.1
192	17.8	14.3	21.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#36
1,1-Dichloropropene
Concen: 49.002 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 75 Resp: 148837

Ion	Ratio	Lower	Upper
75	100		
110	33.3	16.5	49.5
77	31.1	24.4	36.6





#37
Ethyl Acetate
Concen: 49.505 ug/l
RT: 7.553 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04


Tgt Ion: 43 Resp: 151771


Ion	Ratio	Lower	Upper
43	100		
61	12.0	10.9	16.3
70	10.2	7.5	11.3

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025







#38
Carbon Tetrachloride
Concen: 51.696 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:117 Resp: 179763

Ion	Ratio	Lower	Upper
117	100		
119	92.4	75.4	113.2
121	28.7	24.6	37.0






#39
Methylcyclohexane
Concen: 47.944 ug/l
RT: 9.600 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 83 Resp: 136843


Ion	Ratio	Lower	Upper
83	100		
55	80.3	62.6	94.0
98	45.5	37.7	56.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#40
Benzene
Concen: 50.976 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 78 Resp: 465224

Ion	Ratio	Lower	Upper
78	100		
77	24.0	19.0	28.6






#41
Methacrylonitrile
Concen: 53.528 ug/l
RT: 7.771 min Scan# 98
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 41 Resp: 85378


Ion	Ratio	Lower	Upper
41	100		
39	56.2	46.0	69.0
67	67.9	57.4	86.2
52	32.0	25.5	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#42
1,2-Dichloroethane
Concen: 51.851 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 62 Resp: 178224

Ion	Ratio	Lower	Upper
62	100		
98	8.1	0.0	17.0






#43
Isopropyl Acetate
Concen: 52.309 ug/l
RT: 8.683 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 43 Resp: 256966


Ion	Ratio	Lower	Upper
43	100		
61	25.5	20.7	31.1
87	12.2	9.8	14.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#44
Trichloroethene
Concen: 49.486 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:130 Resp: 105127

Ion	Ratio	Lower	Upper
130	100		
95	94.6	0.0	195.8






#45
1,2-Dichloropropane
Concen: 51.880 ug/l
RT: 9.618 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 63 Resp: 120938


Ion	Ratio	Lower	Upper
63	100		
65	30.6	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#46
Dibromomethane
Concen: 50.884 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 93 Resp: 85588

Ion	Ratio	Lower	Upper
93	100		
95	82.6	64.7	97.1
174	89.5	69.0	103.4






#47
Bromodichloromethane
Concen: 54.191 ug/l
RT: 9.882 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 83 Resp: 185704


Ion	Ratio	Lower	Upper
83	100		
85	64.9	51.2	76.8
127	8.4	6.5	9.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#48
Methyl methacrylate
Concen: 55.425 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 41 Resp: 122530

Ion	Ratio	Lower	Upper
41	100		
69	80.3	64.7	97.1
39	63.2	49.0	73.6






#49
1,4-Dioxane
Concen: 1082.516 ug/l
RT: 9.688 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 88 Resp: 40300


Ion	Ratio	Lower	Upper
88	100		
43	31.2	26.6	39.8
58	74.6	59.5	89.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#50
Toluene-d8
Concen: 54.021 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 98 Resp: 415377

Ion	Ratio	Lower	Upper
98	100		
100	64.1	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 285.284 ug/l
RT: 10.441 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 43 Resp: 813239


Ion	Ratio	Lower	Upper
43	100		
58	37.5	30.5	45.7



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#52
Toluene
Concen: 53.686 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 92 Resp: 283889

Ion	Ratio	Lower	Upper
92	100		
91	175.0	139.2	208.8






#53
t-1,3-Dichloropropene
Concen: 54.277 ug/l
RT: 10.835 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 75 Resp: 175773


Ion	Ratio	Lower	Upper
75	100		
77	32.1	23.5	35.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#54
cis-1,3-Dichloropropene
Concen: 54.468 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 75 Resp: 188416

Ion	Ratio	Lower	Upper
75	100		
77	30.7	25.0	37.4
39	52.8	43.1	64.7






#55
1,1,2-Trichloroethane
Concen: 54.314 ug/l
RT: 11.012 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 97 Resp: 113662


Ion	Ratio	Lower	Upper
97	100		
83	84.5	69.0	103.6
85	53.0	44.6	66.8
99	61.6	48.6	72.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#56
Ethyl methacrylate
Concen: 47.979 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 69 Resp: 168851

Ion	Ratio	Lower	Upper
69	100		
41	67.5	54.6	82.0
39	40.9	32.4	48.6






#57
1,3-Dichloropropane
Concen: 53.898 ug/l
RT: 11.159 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 76 Resp: 196125


Ion	Ratio	Lower	Upper
76	100		
78	32.1	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#58
2-Chloroethyl Vinyl ether
Concen: 278.261 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 63 Resp: 369534

Ion	Ratio	Lower	Upper
63	100		
106	27.2	21.6	32.4






#59
2-Hexanone
Concen: 292.282 ug/l
RT: 11.194 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 43 Resp: 586256


Ion	Ratio	Lower	Upper
43	100		
58	52.2	26.2	78.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#60
Dibromochloromethane
Concen: 55.058 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:129 Resp: 139102

Ion	Ratio	Lower	Upper
129	100		
127	76.9	38.6	115.8






#61
1,2-Dibromoethane
Concen: 51.584 ug/l
RT: 11.465 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:107 Resp: 107453


Ion	Ratio	Lower	Upper
107	100		
109	96.1	75.9	113.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#62
4-Bromofluorobenzene
Concen: 56.078 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 95 Resp: 147501

Ion	Ratio	Lower	Upper
95	100		
174	74.1	0.0	145.0
176	71.4	0.0	142.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:117 Resp: 284481


Ion	Ratio	Lower	Upper
117	100		
82	57.5	48.6	72.8
119	32.1	26.6	39.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#64
Tetrachloroethene
Concen: 52.413 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:164 Resp: 101650

Ion	Ratio	Lower	Upper
164	100		
166	125.4	103.4	155.2
129	97.3	79.2	118.8
131	94.6	77.1	115.7






#65
Chlorobenzene
Concen: 49.786 ug/l
RT: 11.888 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:112 Resp: 310267


Ion	Ratio	Lower	Upper
112	100		
114	32.0	25.3	37.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 50.764 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:131 Resp: 116111

Ion	Ratio	Lower	Upper
131	100		
133	98.4	47.4	142.3
119	65.3	33.1	99.5






#67
Ethyl Benzene
Concen: 53.184 ug/l
RT: 11.959 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 91 Resp: 539837


Ion	Ratio	Lower	Upper
91	100		
106	30.0	23.8	35.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#68
m/p-Xylenes
Concen: 111.234 ug/l
RT: 12.065 min Scan# 1719
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:106 Resp: 417283

Ion	Ratio	Lower	Upper
106	100		
91	208.0	167.7	251.5






#69
o-Xylene
Concen: 54.671 ug/l
RT: 12.394 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:106 Resp: 196000


Ion	Ratio	Lower	Upper
106	100		
91	219.3	110.4	331.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#70
Styrene
Concen: 58.039 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:104 Resp: 344358

Ion	Ratio	Lower	Upper
104	100		
78	52.0	42.5	63.7
103	55.0	43.8	65.8





#71
Bromoform
Concen: 55.254 ug/l
RT: 12.576 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04


Tgt Ion:173 Resp: 90428


Ion	Ratio	Lower	Upper
173	100		
175	48.1	24.4	73.2
254	0.0	0.0	0.0

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025







#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:152 Resp: 145646

Ion	Ratio	Lower	Upper
152	100		
115	62.5	31.1	93.3
150	176.8	0.0	343.6






#73
Isopropylbenzene
Concen: 50.753 ug/l
RT: 12.694 min Scan# 18
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:105 Resp: 498892


Ion	Ratio	Lower	Upper
105	100		
120	26.1	12.8	38.3



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#74
N-amyl acetate
Concen: 49.394 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 43 Resp: 218240

Ion	Ratio	Lower	Upper
43	100		
70	42.7	34.0	51.0
55	27.7	21.4	32.2
61	23.8	19.0	28.4






#75
1,1,2,2-Tetrachloroethane
Concen: 47.260 ug/l
RT: 12.935 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 83 Resp: 164103


Ion	Ratio	Lower	Upper
83	100		
131	10.1	4.8	14.4
85	64.2	32.2	96.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#76
1,2,3-Trichloropropane
Concen: 52.775 ug/l m
RT: 12.988 min Scan# 1876
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 75 Resp: 156230

Ion	Ratio	Lower	Upper
75	100		
77	183.3	109.7	329.2






#77
Bromobenzene
Concen: 48.406 ug/l
RT: 12.976 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:156 Resp: 124316


Ion	Ratio	Lower	Upper
156	100		
77	230.3	114.1	342.4
158	98.4	48.9	146.8



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#78
n-propylbenzene
Concen: 53.092 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 91 Resp: 617823

Ion	Ratio	Lower	Upper
91	100		
120	21.8	10.9	32.6






#79
2-Chlorotoluene
Concen: 50.010 ug/l
RT: 13.123 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 91 Resp: 376644


Ion	Ratio	Lower	Upper
91	100		
126	32.6	15.7	47.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#80
1,3,5-Trimethylbenzene
Concen: 53.665 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:105 Resp: 435734

Ion	Ratio	Lower	Upper
105	100		
120	47.8	23.9	71.7






#81
trans-1,4-Dichloro-2-butene
Concen: 53.451 ug/l
RT: 12.735 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 75 Resp: 58483


Ion	Ratio	Lower	Upper
75	100		
53	115.3	95.6	143.4
89	43.3	37.0	55.6



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#82
4-Chlorotoluene
Concen: 51.696 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 91 Resp: 387664

Ion	Ratio	Lower	Upper
91	100		
126	32.3	15.9	47.7






#83
tert-Butylbenzene
Concen: 51.897 ug/l
RT: 13.435 min Scan# 19
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:119 Resp: 353864


Ion	Ratio	Lower	Upper
119	100		
91	69.2	35.1	105.3
134	24.2	12.0	36.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#84
1,2,4-Trimethylbenzene
Concen: 54.663 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:105 Resp: 442366

Ion	Ratio	Lower	Upper
105	100		
120	43.9	21.6	65.0






#85
sec-Butylbenzene
Concen: 54.515 ug/l
RT: 13.612 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:105 Resp: 515181


Ion	Ratio	Lower	Upper
105	100		
134	19.4	9.7	28.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#86
p-Isopropyltoluene
Concen: 50.108 ug/l
RT: 13.723 min Scan# 2001
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:119 Resp: 426531

Ion	Ratio	Lower	Upper
119	100		
134	25.6	12.7	38.0
91	25.4	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 51.042 ug/l
RT: 13.729 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:146 Resp: 241411


Ion	Ratio	Lower	Upper
146	100		
111	42.5	21.4	64.3
148	62.9	32.3	96.9



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#88
1,4-Dichlorobenzene
Concen: 47.611 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:146 Resp: 233405

Ion	Ratio	Lower	Upper
146	100		
111	41.5	21.3	63.7
148	64.9	32.4	97.0






#89
n-Butylbenzene
Concen: 53.798 ug/l
RT: 14.053 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 91 Resp: 364737


Ion	Ratio	Lower	Upper
91	100		
92	51.6	25.8	77.3
134	24.1	11.7	35.1



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#90
Hexachloroethane
Concen: 48.061 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:117 Resp: 86465

Ion	Ratio	Lower	Upper
117	100		
201	68.0	33.7	101.0






#91
1,2-Dichlorobenzene
Concen: 49.337 ug/l
RT: 14.100 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:146 Resp: 232755


Ion	Ratio	Lower	Upper
146	100		
111	43.5	21.7	65.1
148	63.8	31.4	94.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 48.536 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion: 75 Resp: 30775

Ion	Ratio	Lower	Upper
75	100		
155	71.5	36.4	109.2
157	92.5	45.4	136.1






#93
1,2,4-Trichlorobenzene
Concen: 48.649 ug/l
RT: 15.388 min Scan# 22
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:180 Resp: 106670


Ion	Ratio	Lower	Upper
180	100		
182	96.6	46.8	140.3
145	32.6	16.0	48.0



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#94
Hexachlorobutadiene
Concen: 48.367 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:225 Resp: 56304

Ion	Ratio	Lower	Upper
225	100		
223	63.0	30.7	92.1
227	61.7	30.9	92.5






#95
Naphthalene
Concen: 48.220 ug/l
RT: 15.635 min Scan# 23
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:128 Resp: 315500


Ion	Ratio	Lower	Upper
128	100		
127	13.4	10.6	16.0
129	10.7	8.8	13.2



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025



#96
1,2,3-Trichlorobenzene
Concen: 48.503 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN085591.D
Acq: 31 Jan 2025 10:04

Tgt Ion:180 Resp: 107609

Ion	Ratio	Lower	Upper
180	100		
182	95.4	47.4	142.2
145	34.0	16.9	50.7



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085591.D
Acq On : 31 Jan 2025 10:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 22:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2	T	Dichlorodifluoromethane	0.677	0.545	19.5	83	0.00
3	P	Chloromethane	0.733	0.582	20.6	82	0.00
4	C	Vinyl Chloride	0.737	0.631	14.4#	88	0.00
5	T	Bromomethane	0.445	0.381	14.4	87	0.00
6	T	Chloroethane	0.467	0.426	8.8	96	0.00
7	T	Trichlorofluoromethane	1.069	1.004	6.1	96	0.00
8	T	Diethyl Ether	0.369	0.321	13.0	85	0.00
9	T	1,1,2-Trichlorotrifluoroeth	0.602	0.579	3.8	102	0.00
10	T	Methyl Iodide	0.690	0.646	6.4	86	0.00
11	T	Tert butyl alcohol	0.092	0.108	-17.4	110	0.00
12	CM	1,1-Dichloroethene	0.537	0.497	7.4#	89	0.00
13	T	Acrolein	0.126	0.095	24.6	65	0.01
14	T	Allyl chloride	0.871	0.753	13.5	85	0.00
15	T	Acrylonitrile	0.294	0.294	0.0	94	0.00
16	T	Acetone	0.261	0.250	4.2	95	0.00
17	T	Carbon Disulfide	1.652	1.219	26.2#	79	0.00
18	T	Methyl Acetate	0.793	0.857	-8.1	104	0.00
19	T	Methyl tert-butyl Ether	1.742	1.807	-3.7	92	0.00
20	T	Methylene Chloride	0.646	0.618	4.3	94	0.00
21	T	trans-1,2-Dichloroethene	0.573	0.539	5.9	93	0.00
22	T	Diisopropyl ether	1.933	1.994	-3.2	93	0.00
23	T	Vinyl Acetate	1.353	1.406	-3.9	90	0.00
24	P	1,1-Dichloroethane	1.178	1.158	1.7	94	0.00
25	T	2-Butanone	0.384	0.389	-1.3	95	0.00
26	T	2,2-Dichloropropane	0.953	1.046	-9.8	107	0.00
27	T	cis-1,2-Dichloroethene	0.675	0.663	1.8	93	0.00
28	T	Bromochloromethane	0.548	0.520	5.1	92	0.00
29	T	Tetrahydrofuran	0.243	0.260	-7.0	95	0.00
30	C	Chloroform	1.218	1.221	-0.2#	99	0.00
31	T	Cyclohexane	1.024	0.848	17.2	92	0.00
32	T	1,1,1-Trichloroethane	1.068	1.061	0.7	100	0.00
33	S	1,2-Dichloroethane-d4	0.807	0.801	0.7	92	0.00
34	I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35	S	Dibromofluoromethane	0.347	0.374	-7.8	96	0.00
36	T	1,1-Dichloropropene	0.487	0.477	2.1	95	0.00
37	T	Ethyl Acetate	0.491	0.487	0.8	92	0.00
38	T	Carbon Tetrachloride	0.557	0.576	-3.4	100	0.00
39	T	Methylcyclohexane	0.457	0.439	3.9	87	0.00
40	TM	Benzene	1.463	1.491	-1.9	94	0.00
41	T	Methacrylonitrile	0.256	0.274	-7.0	93	0.00
42	TM	1,2-Dichloroethane	0.551	0.571	-3.6	96	0.00
43	T	Isopropyl Acetate	0.787	0.824	-4.7	93	0.00
44	TM	Trichloroethene	0.341	0.337	1.2	96	0.00
45	C	1,2-Dichloropropane	0.374	0.388	-3.7#	96	0.00
46	T	Dibromomethane	0.270	0.274	-1.5	94	0.00
47	T	Bromodichloromethane	0.549	0.595	-8.4	98	0.00
48	T	Methyl methacrylate	0.354	0.393	-11.0	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085591.D
Acq On : 31 Jan 2025 10:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 22:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	0.006	0.006	0.0	94	0.00
50	S	Toluene-d8	1.232	1.332	-8.1	96	0.00
51	T	4-Methyl-2-Pentanone	0.457	0.521	-14.0	97	0.00
52	CM	Toluene	0.848	0.910	-7.3#	96	0.00
53	T	t-1,3-Dichloropropene	0.519	0.563	-8.5	94	0.00
54	T	cis-1,3-Dichloropropene	0.554	0.604	-9.0	94	0.00
55	T	1,1,2-Trichloroethane	0.335	0.364	-8.7	98	0.00
56	T	Ethyl methacrylate	0.469	0.541	-15.4	91	0.00
57	T	1,3-Dichloropropane	0.583	0.629	-7.9	96	0.00
58	T	2-Chloroethyl Vinyl ether	0.213	0.237	-11.3	90	0.00
59	T	2-Hexanone	0.321	0.376	-17.1	97	0.00
60	T	Dibromochloromethane	0.405	0.446	-10.1	99	0.00
61	T	1,2-Dibromoethane	0.334	0.344	-3.0	95	0.00
62	S	4-Bromofluorobenzene	0.422	0.473	-12.1	97	0.00
63	I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64	T	Tetrachloroethene	0.341	0.357	-4.7	106	0.00
65	PM	Chlorobenzene	1.095	1.091	0.4	97	0.00
66	T	1,1,1,2-Tetrachloroethane	0.402	0.408	-1.5	100	0.00
67	C	Ethyl Benzene	1.784	1.898	-6.4#	97	0.00
68	T	m/p-Xylenes	0.659	0.733	-11.2	99	0.00
69	T	o-Xylene	0.630	0.689	-9.4	97	0.00
70	T	Styrene	1.043	1.210	-16.0	99	0.00
71	P	Bromoform	0.288	0.318	-10.4	98	0.00
72	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73	T	Isopropylbenzene	3.375	3.425	-1.5	99	0.00
74	T	N-amyl acetate	1.517	1.498	1.3	92	0.00
75	P	1,1,2,2-Tetrachloroethane	1.192	1.127	5.5	98	0.00
76	T	1,2,3-Trichloropropane	1.016	1.073	-5.6	119	0.00
77	T	Bromobenzene	0.882	0.854	3.2	99	0.00
78	T	n-propylbenzene	3.995	4.242	-6.2	101	0.00
79	T	2-Chlorotoluene	2.586	2.586	0.0	99	0.00
80	T	1,3,5-Trimethylbenzene	2.787	2.992	-7.4	101	0.00
81	T	trans-1,4-Dichloro-2-butene	0.376	0.402	-6.9	102	0.00
82	T	4-Chlorotoluene	2.574	2.662	-3.4	102	0.00
83	T	tert-Butylbenzene	2.341	2.430	-3.8	99	0.00
84	T	1,2,4-Trimethylbenzene	2.778	3.037	-9.3	100	0.00
85	T	sec-Butylbenzene	3.244	3.537	-9.0	103	0.00
86	T	p-Isopropyltoluene	2.631	2.929	-11.3	101	0.00
87	T	1,3-Dichlorobenzene	1.624	1.658	-2.1	105	0.00
88	T	1,4-Dichlorobenzene	1.683	1.603	4.8	102	0.00
89	T	n-Butylbenzene	2.327	2.504	-7.6	100	0.00
90	T	Hexachloroethane	0.618	0.594	3.9	102	0.00
91	T	1,2-Dichlorobenzene	1.620	1.598	1.4	102	0.00
92	T	1,2-Dibromo-3-Chloropropane	0.218	0.211	3.2	95	0.00
93	T	1,2,4-Trichlorobenzene	0.753	0.732	2.8	93	0.00
94	T	Hexachlorobutadiene	0.400	0.387	3.3	105	0.00
95	T	Naphthalene	2.246	2.166	3.6	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085591.D
Acq On : 31 Jan 2025 10:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 22:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.762	0.739	3.0	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085591.D
Acq On : 31 Jan 2025 10:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 22:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2	T	Dichlorodifluoromethane	50.000	40.276	19.4	83	0.00
3	P	Chloromethane	50.000	39.689	20.6	82	0.00
4	C	Vinyl Chloride	50.000	42.836	14.3#	88	0.00
5	T	Bromomethane	50.000	42.763	14.5	87	0.00
6	T	Chloroethane	50.000	45.613	8.8	96	0.00
7	T	Trichlorofluoromethane	50.000	46.958	6.1	96	0.00
8	T	Diethyl Ether	50.000	43.462	13.1	85	0.00
9	T	1,1,2-Trichlorotrifluoroeth	50.000	48.073	3.9	102	0.00
10	T	Methyl Iodide	50.000	46.827	6.3	86	0.00
11	T	Tert butyl alcohol	250.000	291.355	-16.5	110	0.00
12	CM	1,1-Dichloroethene	50.000	46.352	7.3#	89	0.00
13	T	Acrolein	250.000	187.561	25.0	65	0.01
14	T	Allyl chloride	50.000	43.262	13.5	85	0.00
15	T	Acrylonitrile	250.000	249.970	0.0	94	0.00
16	T	Acetone	250.000	239.913	4.0	95	0.00
17	T	Carbon Disulfide	50.000	36.884	26.2#	79	0.00
18	T	Methyl Acetate	50.000	54.036	-8.1	104	0.00
19	T	Methyl tert-butyl Ether	50.000	51.861	-3.7	92	0.00
20	T	Methylene Chloride	50.000	47.827	4.3	94	0.00
21	T	trans-1,2-Dichloroethene	50.000	46.987	6.0	93	0.00
22	T	Diisopropyl ether	50.000	51.582	-3.2	93	0.00
23	T	Vinyl Acetate	250.000	259.853	-3.9	90	0.00
24	P	1,1-Dichloroethane	50.000	49.119	1.8	94	0.00
25	T	2-Butanone	250.000	253.698	-1.5	95	0.00
26	T	2,2-Dichloropropane	50.000	54.902	-9.8	107	0.00
27	T	cis-1,2-Dichloroethene	50.000	49.110	1.8	93	0.00
28	T	Bromochloromethane	50.000	47.443	5.1	92	0.00
29	T	Tetrahydrofuran	250.000	266.707	-6.7	95	0.00
30	C	Chloroform	50.000	50.126	-0.3#	99	0.00
31	T	Cyclohexane	50.000	41.402	17.2	92	0.00
32	T	1,1,1-Trichloroethane	50.000	49.666	0.7	100	0.00
33	S	1,2-Dichloroethane-d4	50.000	49.634	0.7	92	0.00
34	I	1,4-Difluorobenzene	50.000	50.000	0.0	92	0.00
35	S	Dibromofluoromethane	50.000	53.927	-7.9	96	0.00
36	T	1,1-Dichloropropene	50.000	49.002	2.0	95	0.00
37	T	Ethyl Acetate	50.000	49.505	1.0	92	0.00
38	T	Carbon Tetrachloride	50.000	51.696	-3.4	100	0.00
39	T	Methylcyclohexane	50.000	47.944	4.1	87	0.00
40	TM	Benzene	50.000	50.976	-2.0	94	0.00
41	T	Methacrylonitrile	50.000	53.528	-7.1	93	0.00
42	TM	1,2-Dichloroethane	50.000	51.851	-3.7	96	0.00
43	T	Isopropyl Acetate	50.000	52.309	-4.6	93	0.00
44	TM	Trichloroethene	50.000	49.486	1.0	96	0.00
45	C	1,2-Dichloropropane	50.000	51.880	-3.8#	96	0.00
46	T	Dibromomethane	50.000	50.884	-1.8	94	0.00
47	T	Bromodichloromethane	50.000	54.191	-8.4	98	0.00
48	T	Methyl methacrylate	50.000	55.425	-10.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085591.D
Acq On : 31 Jan 2025 10:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 22:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	1000.000	1082.516	-8.3	94	0.00
50	S	Toluene-d8	50.000	54.021	-8.0	96	0.00
51	T	4-Methyl-2-Pentanone	250.000	285.284	-14.1	97	0.00
52	CM	Toluene	50.000	53.686	-7.4#	96	0.00
53	T	t-1,3-Dichloropropene	50.000	54.277	-8.6	94	0.00
54	T	cis-1,3-Dichloropropene	50.000	54.468	-8.9	94	0.00
55	T	1,1,2-Trichloroethane	50.000	54.314	-8.6	98	0.00
56	T	Ethyl methacrylate	50.000	47.979	4.0	91	0.00
57	T	1,3-Dichloropropane	50.000	53.898	-7.8	96	0.00
58	T	2-Chloroethyl Vinyl ether	250.000	278.261	-11.3	90	0.00
59	T	2-Hexanone	250.000	292.282	-16.9	97	0.00
60	T	Dibromochloromethane	50.000	55.058	-10.1	99	0.00
61	T	1,2-Dibromoethane	50.000	51.584	-3.2	95	0.00
62	S	4-Bromofluorobenzene	50.000	56.078	-12.2	97	0.00
63	I	Chlorobenzene-d5	50.000	50.000	0.0	96	0.00
64	T	Tetrachloroethene	50.000	52.413	-4.8	106	0.00
65	PM	Chlorobenzene	50.000	49.786	0.4	97	0.00
66	T	1,1,1,2-Tetrachloroethane	50.000	50.764	-1.5	100	0.00
67	C	Ethyl Benzene	50.000	53.184	-6.4#	97	0.00
68	T	m/p-Xylenes	100.000	111.234	-11.2	99	0.00
69	T	o-Xylene	50.000	54.671	-9.3	97	0.00
70	T	Styrene	50.000	58.039	-16.1	99	0.00
71	P	Bromoform	50.000	55.254	-10.5	98	0.00
72	I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	99	0.00
73	T	Isopropylbenzene	50.000	50.753	-1.5	99	0.00
74	T	N-amyl acetate	50.000	49.394	1.2	92	0.00
75	P	1,1,2,2-Tetrachloroethane	50.000	47.260	5.5	98	0.00
76	T	1,2,3-Trichloropropane	50.000	52.775	-5.5	119	0.00
77	T	Bromobenzene	50.000	48.406	3.2	99	0.00
78	T	n-propylbenzene	50.000	53.092	-6.2	101	0.00
79	T	2-Chlorotoluene	50.000	50.010	-0.0	99	0.00
80	T	1,3,5-Trimethylbenzene	50.000	53.665	-7.3	101	0.00
81	T	trans-1,4-Dichloro-2-butene	50.000	53.451	-6.9	102	0.00
82	T	4-Chlorotoluene	50.000	51.696	-3.4	102	0.00
83	T	tert-Butylbenzene	50.000	51.897	-3.8	99	0.00
84	T	1,2,4-Trimethylbenzene	50.000	54.663	-9.3	100	0.00
85	T	sec-Butylbenzene	50.000	54.515	-9.0	103	0.00
86	T	p-Isopropyltoluene	50.000	50.108	-0.2	101	0.00
87	T	1,3-Dichlorobenzene	50.000	51.042	-2.1	105	0.00
88	T	1,4-Dichlorobenzene	50.000	47.611	4.8	102	0.00
89	T	n-Butylbenzene	50.000	53.798	-7.6	100	0.00
90	T	Hexachloroethane	50.000	48.061	3.9	102	0.00
91	T	1,2-Dichlorobenzene	50.000	49.337	1.3	102	0.00
92	T	1,2-Dibromo-3-Chloropropane	50.000	48.536	2.9	95	0.00
93	T	1,2,4-Trichlorobenzene	50.000	48.649	2.7	93	0.00
94	T	Hexachlorobutadiene	50.000	48.367	3.3	105	0.00
95	T	Naphthalene	50.000	48.220	3.6	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085591.D
Acq On : 31 Jan 2025 10:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 22:23:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.503	3.0	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG No.: Q1215

Instrument ID: MSVOA_X Calibration Date/Time: 02/03/2025 09:52

Lab File ID: VX044802.D Init. Calib. Date(s): 01/15/2025 01/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:31 13:37

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.719	0.766		6.54	20
1,1-Dichloroethene	0.585	0.636		8.72	20
2-Butanone	0.444	0.459		3.38	20
Carbon Tetrachloride	0.497	0.454		-8.65	20
Chloroform	1.186	1.162		-2.02	20
Benzene	1.370	1.440		5.11	20
1,2-Dichloroethane	0.500	0.454		-9.2	20
Trichloroethene	0.329	0.337		2.43	20
Tetrachloroethene	0.314	0.312		-0.64	20
Chlorobenzene	1.062	1.101	0.3	3.67	20
1,2-Dichloroethane-d4	0.755	0.726		-3.84	20
Dibromofluoromethane	0.318	0.333		4.72	20
Toluene-d8	1.108	1.249		12.73	20
4-Bromofluorobenzene	0.413	0.437		5.81	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044802.D
Acq On : 03 Feb 2025 09:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Feb 04 02:00:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.538	168	162430	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	297715	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	252415	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	101849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	117965	48.114	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 96.220%			
35) Dibromofluoromethane	5.379	113	99232	52.467	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.940%			
50) Toluene-d8	8.641	98	371824	56.382	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 112.760%			
62) 4-Bromofluorobenzene	11.079	95	130114	52.925	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.860%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	104350	47.741	ug/l	100
3) Chloromethane	1.301	50	122617	51.877	ug/l	100
4) Vinyl Chloride	1.374	62	124437	53.283	ug/l	97
5) Bromomethane	1.593	94	32108	57.578	ug/l	100
6) Chloroethane	1.660	64	22426m	52.365	ug/l	
7) Trichlorofluoromethane	1.868	101	155099	60.090	ug/l	99
8) Diethyl Ether	2.130	74	51512	49.453	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.319	101	103805	53.500	ug/l	98
10) Methyl Iodide	2.441	142	132750	50.112	ug/l	97
11) Tert butyl alcohol	2.977	59	111502m	204.000	ug/l	
12) 1,1-Dichloroethene	2.307	96	103312	54.321	ug/l	98
13) Acrolein	2.233	56	55610	216.204	ug/l	98
14) Allyl chloride	2.654	41	182859	49.381	ug/l	97
15) Acrylonitrile	3.062	53	284374	265.265	ug/l	99
16) Acetone	2.380	43	235232	242.503	ug/l	97
17) Carbon Disulfide	2.502	76	274133	52.823	ug/l	99
18) Methyl Acetate	2.697	43	160361	53.585	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	337910	49.052	ug/l	97
20) Methylene Chloride	2.782	84	113693	50.330	ug/l	95
21) trans-1,2-Dichloroethene	3.081	96	100823	50.775	ug/l	99
22) Diisopropyl ether	3.751	45	353913	54.981	ug/l	97
23) Vinyl Acetate	3.715	43	1458707	258.911	ug/l	100
24) 1,1-Dichloroethane	3.599	63	198603	52.029	ug/l	100
25) 2-Butanone	4.550	43	373119	258.709	ug/l	98
26) 2,2-Dichloropropane	4.465	77	168163	45.456	ug/l	96
27) cis-1,2-Dichloroethene	4.477	96	125972	50.436	ug/l	97
28) Bromochloromethane	4.885	49	90867	50.943	ug/l	94
29) Tetrahydrofuran	4.995	42	240754	258.796	ug/l	96
30) Chloroform	5.080	83	188807	49.012	ug/l	99
31) Cyclohexane	5.458	56	177264	57.418	ug/l	93
32) 1,1,1-Trichloroethane	5.367	97	160997	45.461	ug/l	99
36) 1,1-Dichloropropene	5.678	75	133384	50.662	ug/l	99
37) Ethyl Acetate	4.702	43	146622	52.295	ug/l	96
38) Carbon Tetrachloride	5.666	117	135277	45.680	ug/l	97
39) Methylcyclohexane	7.373	83	187791	53.731	ug/l	99
40) Benzene	6.025	78	428598	52.535	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044802.D
Acq On : 03 Feb 2025 09:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Feb 04 02:00:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.910	41	82519	50.508	ug/l		99
42) 1,2-Dichloroethane	6.074	62	135051	45.355	ug/l		100
43) Isopropyl Acetate	6.330	43	253694	49.187	ug/l		99
44) Trichloroethene	7.117	130	100231	51.104	ug/l		95
45) 1,2-Dichloropropane	7.421	63	109264	53.206	ug/l		97
46) Dibromomethane	7.574	93	73965	47.313	ug/l		99
47) Bromodichloromethane	7.812	83	149483	46.626	ug/l		99
48) Methyl methacrylate	7.684	41	119095	49.744	ug/l		98
49) 1,4-Dioxane	7.659	88	45154	903.978	ug/l		100
51) 4-Methyl-2-Pentanone	8.568	43	734014	257.519	ug/l		99
52) Toluene	8.714	92	261673	52.403	ug/l		99
53) t-1,3-Dichloropropene	8.970	75	159515	48.158	ug/l		98
54) cis-1,3-Dichloropropene	8.360	75	175713	48.923	ug/l		98
55) 1,1,2-Trichloroethane	9.147	97	100638	49.814	ug/l		97
56) Ethyl methacrylate	9.110	69	170918	49.203	ug/l		99
57) 1,3-Dichloropropane	9.305	76	172698	50.198	ug/l		99
58) 2-Chloroethyl Vinyl ether	8.238	63	427737	250.814	ug/l		97
59) 2-Hexanone	9.427	43	531025	253.231	ug/l		100
60) Dibromochloromethane	9.513	129	111967	47.219	ug/l		99
61) 1,2-Dibromoethane	9.604	107	102485	48.405	ug/l		99
64) Tetrachloroethene	9.269	164	78776	49.701	ug/l		98
65) Chlorobenzene	10.073	112	277930	51.839	ug/l		99
66) 1,1,1,2-Tetrachloroethane	10.159	131	93479	47.975	ug/l		99
67) Ethyl Benzene	10.189	91	496981	53.125	ug/l		100
68) m/p-Xylenes	10.299	106	371687	107.068	ug/l		97
69) o-Xylene	10.640	106	181303	51.643	ug/l		99
70) Styrene	10.653	104	302980	53.453	ug/l		97
71) Bromoform	10.799	173	73161	46.133	ug/l	#	100
73) Isopropylbenzene	10.957	105	462864	55.630	ug/l		100
74) N-amyl acetate	10.842	43	214304	52.255	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.207	83	151416	52.692	ug/l		99
76) 1,2,3-Trichloropropane	11.238	75	126793m	51.297	ug/l		
77) Bromobenzene	11.195	156	103391	53.164	ug/l		98
78) n-propylbenzene	11.299	91	543610	59.244	ug/l		99
79) 2-Chlorotoluene	11.360	91	318475	53.730	ug/l		99
80) 1,3,5-Trimethylbenzene	11.445	105	388082	57.296	ug/l		98
81) trans-1,4-Dichloro-2-b...	11.012	75	54681	50.591	ug/l		97
82) 4-Chlorotoluene	11.451	91	375241	57.791	ug/l		99
83) tert-Butylbenzene	11.713	119	355525	50.199	ug/l		97
84) 1,2,4-Trimethylbenzene	11.750	105	351562	51.077	ug/l		98
85) sec-Butylbenzene	11.890	105	430027	52.148	ug/l		100
86) p-Isopropyltoluene	12.006	119	364463	53.781	ug/l		98
87) 1,3-Dichlorobenzene	11.963	146	179777	53.304	ug/l		99
88) 1,4-Dichlorobenzene	12.036	146	176282	51.704	ug/l		100
89) n-Butylbenzene	12.329	91	341907	58.741	ug/l		99
90) Hexachloroethane	12.536	117	78195	50.955	ug/l		99
91) 1,2-Dichlorobenzene	12.329	146	183157	52.880	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	29314	41.712	ug/l		94
93) 1,2,4-Trichlorobenzene	13.585	180	120382	58.139	ug/l		99
94) Hexachlorobutadiene	13.719	225	47155	57.977	ug/l		98
95) Naphthalene	13.774	128	435692	54.557	ug/l		100
96) 1,2,3-Trichlorobenzene	13.957	180	119840	55.837	ug/l		100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044802.D
Acq On : 03 Feb 2025 09:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Feb 04 02:00:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044802.D
Acq On : 03 Feb 2025 09:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1


Quant Time: Feb 04 02:00:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:168 Resp: 16243


Ion	Ratio	Lower	Upper
168	100		
99	61.0	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#2
Dichlorodifluoromethane
Concen: 47.741 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 85 Resp: 104350

Ion	Ratio	Lower	Upper
85	100		
87	33.7	16.8	50.4






#3
Chloromethane
Concen: 51.877 ug/l
RT: 1.301 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 50 Resp: 12261


Ion	Ratio	Lower	Upper
50	100		
52	33.6	27.0	40.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#4
Vinyl Chloride
Concen: 53.283 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 62 Resp: 124437

Ion	Ratio	Lower	Upper
62	100		
64	31.7	26.6	39.8






#5
Bromomethane
Concen: 57.578 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 94 Resp: 3210


Ion	Ratio	Lower	Upper
94	100		
96	90.7	72.8	109.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#6
Chloroethane
Concen: 52.365 ug/l m
RT: 1.660 min Scan# 94
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 64 Resp: 22426

Ion	Ratio	Lower	Upper
64	100		
66	31.6	26.6	39.8






#7
Trichlorofluoromethane
Concen: 60.090 ug/l
RT: 1.868 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:101 Resp: 15509


Ion	Ratio	Lower	Upper
101	100		
103	66.3	52.5	78.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#8
Diethyl Ether
Concen: 49.453 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 74 Resp: 51512

Ion	Ratio	Lower	Upper
74	100		
45	99.5	47.1	141.3






#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.500 ug/l
RT: 2.319 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:101 Resp: 10380


Ion	Ratio	Lower	Upper
101	100		
85	43.7	35.3	52.9
151	70.1	58.2	87.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#10
Methyl Iodide
Concen: 50.112 ug/l
RT: 2.441 min Scan# 222
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:142 Resp: 132750

Ion	Ratio	Lower	Upper
142	100		
127	41.4	35.0	52.4
141	14.1	11.7	17.5






#11
Tert butyl alcohol
Concen: 204.000 ug/l m
RT: 2.977 min Scan# 3
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 59 Resp: 11150


Ion	Ratio	Lower	Upper
59	100		
57	3.9	3.9	5.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#12
1,1-Dichloroethene
Concen: 54.321 ug/l
RT: 2.307 min Scan# 200
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 96 Resp: 103312

Ion	Ratio	Lower	Upper
96	100		
61	159.8	131.1	196.7
98	63.2	51.2	76.8






#13
Acrolein
Concen: 216.204 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 56 Resp: 5561


Ion	Ratio	Lower	Upper
56	100		
55	70.3	55.1	82.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#14
Allyl chloride
Concen: 49.381 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 41 Resp: 182859

Ion	Ratio	Lower	Upper
41	100		
39	64.6	54.9	82.3
76	35.8	28.4	42.6






#15
Acrylonitrile
Concen: 265.265 ug/l
RT: 3.062 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 53 Resp: 28437


Ion	Ratio	Lower	Upper
53	100		
52	82.9	66.0	99.0
51	35.6	28.9	43.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#16
Acetone
Concen: 242.503 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 43 Resp: 235232

Ion	Ratio	Lower	Upper
43	100		
58	32.3	24.6	37.0






#17
Carbon Disulfide
Concen: 52.823 ug/l
RT: 2.502 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 76 Resp: 27413


Ion	Ratio	Lower	Upper
76	100		
78	8.7	7.1	10.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#18
Methyl Acetate
Concen: 53.585 ug/l
RT: 2.697 min Scan# 264
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 43 Resp: 160361

Ion	Ratio	Lower	Upper
43	100		
74	24.1	19.8	29.8






#19
Methyl tert-butyl Ether
Concen: 49.052 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 73 Resp: 33791


Ion	Ratio	Lower	Upper
73	100		
57	23.0	17.4	26.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#20
Methylene Chloride
Concen: 50.330 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 84 Resp: 113693

Ion	Ratio	Lower	Upper
84	100		
49	123.8	94.2	141.4
51	39.5	28.8	43.2
86	64.6	50.3	75.5





#21
trans-1,2-Dichloroethene
Concen: 50.775 ug/l
RT: 3.081 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52


Tgt Ion: 96 Resp: 10082


Ion	Ratio	Lower	Upper
96	100		
61	140.6	111.0	166.6
98	62.6	50.2	75.2

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#22
Diisopropyl ether
Concen: 54.981 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 45 Resp: 353913

Ion	Ratio	Lower	Upper
45	100		
43	78.6	61.0	91.6
87	26.4	22.7	34.1
59	11.2	9.4	14.2





#23
Vinyl Acetate
Concen: 258.911 ug/l
RT: 3.715 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52


Tgt Ion: 43 Resp: 145870


Ion	Ratio	Lower	Upper
43	100		
86	10.9	8.9	13.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#24
1,1-Dichloroethane
Concen: 52.029 ug/l
RT: 3.599 min Scan# 412
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 63 Resp: 198603

Ion	Ratio	Lower	Upper
63	100		
98	7.1	3.6	10.9
100	4.9	2.5	7.5






#25
2-Butanone
Concen: 258.709 ug/l
RT: 4.550 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 43 Resp: 37311


Ion	Ratio	Lower	Upper
43	100		
72	26.1	21.7	32.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#26
2,2-Dichloropropane
Concen: 45.456 ug/l
RT: 4.465 min Scan# 554
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 77 Resp: 168163

Ion	Ratio	Lower	Upper
77	100		
97	23.5	10.9	32.6






#27
cis-1,2-Dichloroethene
Concen: 50.436 ug/l
RT: 4.477 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 96 Resp: 12597


Ion	Ratio	Lower	Upper
96	100		
61	143.2	0.0	296.0
98	63.3	0.0	130.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#28
Bromochloromethane
Concen: 50.943 ug/l
RT: 4.885 min Scan# 623
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 49 Resp: 90867

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	4.4
130	73.7	63.4	95.2





#29
Tetrahydrofuran
Concen: 258.796 ug/l
RT: 4.995 min Scan# 64
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52


Tgt Ion: 42 Resp: 240754


Ion	Ratio	Lower	Upper
42	100		
72	45.8	38.2	57.4
71	41.3	35.3	52.9

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#30
Chloroform
Concen: 49.012 ug/l
RT: 5.080 min Scan# 655
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 83 Resp: 188807

Ion	Ratio	Lower	Upper
83	100		
85	63.1	51.2	76.8






#31
Cyclohexane
Concen: 57.418 ug/l
RT: 5.458 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 56 Resp: 17726


Ion	Ratio	Lower	Upper
56	100		
69	29.9	26.8	40.2
84	82.5	71.0	106.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#32
1,1,1-Trichloroethane
Concen: 45.461 ug/l
RT: 5.367 min Scan# 702
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 97 Resp: 160997

Ion	Ratio	Lower	Upper
97	100		
99	63.7	51.6	77.4
61	44.8	35.6	53.4






#33
1,2-Dichloroethane-d4
Concen: 48.114 ug/l
RT: 5.946 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 65 Resp: 11796


Ion	Ratio	Lower	Upper
65	100		
67	53.8	0.0	103.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.751 min Scan# 929
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:114 Resp: 297715

Ion	Ratio	Lower	Upper
114	100		
63	21.5	0.0	43.0
88	16.5	0.0	35.4






#35
Dibromofluoromethane
Concen: 52.467 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:113 Resp: 9923


Ion	Ratio	Lower	Upper
113	100		
111	103.8	81.6	122.4
192	17.7	14.4	21.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#36
1,1-Dichloropropene
Concen: 50.662 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 75 Resp: 133384

Ion	Ratio	Lower	Upper
75	100		
110	34.8	17.9	53.8
77	30.7	24.5	36.7






#37
Ethyl Acetate
Concen: 52.295 ug/l
RT: 4.702 min Scan# 5
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 43 Resp: 14662


Ion	Ratio	Lower	Upper
43	100		
61	14.6	10.9	16.3
70	10.7	10.1	15.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#38
Carbon Tetrachloride
Concen: 45.680 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:117 Resp: 135277

Ion	Ratio	Lower	Upper
117	100		
119	91.9	76.0	114.0
121	29.2	24.6	36.8






#39
Methylcyclohexane
Concen: 53.731 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 83 Resp: 18779


Ion	Ratio	Lower	Upper
83	100		
55	74.7	60.5	90.7
98	49.5	39.0	58.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#40
Benzene
Concen: 52.535 ug/l
RT: 6.025 min Scan# 810
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 78 Resp: 428598

Ion	Ratio	Lower	Upper
78	100		
77	23.4	18.7	28.1






#41
Methacrylonitrile
Concen: 50.508 ug/l
RT: 4.910 min Scan# 6
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 41 Resp: 8251


Ion	Ratio	Lower	Upper
41	100		
39	55.4	44.4	66.6
67	74.9	60.6	90.8
52	29.9	24.3	36.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#42
1,2-Dichloroethane
Concen: 45.355 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 62 Resp: 135051

Ion	Ratio	Lower	Upper
62	100		
98	9.8	0.0	19.4






#43
Isopropyl Acetate
Concen: 49.187 ug/l
RT: 6.330 min Scan# 8
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 43 Resp: 25369


Ion	Ratio	Lower	Upper
43	100		
61	20.5	16.6	24.8
87	13.0	11.0	16.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#44
Trichloroethene
Concen: 51.104 ug/l
RT: 7.117 min Scan# 989
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:130 Resp: 100231

Ion	Ratio	Lower	Upper
130	100		
95	99.0	0.0	208.2






#45
1,2-Dichloropropane
Concen: 53.206 ug/l
RT: 7.421 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 63 Resp: 10926


Ion	Ratio	Lower	Upper
63	100		
65	28.6	24.2	36.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#46
Dibromomethane
Concen: 47.313 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 93 Resp: 73965

Ion	Ratio	Lower	Upper
93	100		
95	84.2	66.4	99.6
174	89.9	72.0	108.0






#47
Bromodichloromethane
Concen: 46.626 ug/l
RT: 7.812 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 83 Resp: 14948


Ion	Ratio	Lower	Upper
83	100		
85	63.6	50.1	75.1
127	7.9	6.5	9.7



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#48
Methyl methacrylate
Concen: 49.744 ug/l
RT: 7.684 min Scan# 1082
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 41 Resp: 119095

Ion	Ratio	Lower	Upper
41	100		
69	85.8	66.8	100.2
39	58.5	47.4	71.0






#49
1,4-Dioxane
Concen: 903.978 ug/l
RT: 7.659 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 88 Resp: 4515


Ion	Ratio	Lower	Upper
88	100		
43	33.9	27.0	40.4
58	69.5	55.8	83.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#50
Toluene-d8
Concen: 56.382 ug/l
RT: 8.641 min Scan# 1239
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 98 Resp: 371824

Ion	Ratio	Lower	Upper
98	100		
100	66.0	53.2	79.8






#51
4-Methyl-2-Pentanone
Concen: 257.519 ug/l
RT: 8.568 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 43 Resp: 73401


Ion	Ratio	Lower	Upper
43	100		
58	39.5	31.9	47.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#52
Toluene
Concen: 52.403 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 92 Resp: 261673

Ion	Ratio	Lower	Upper
92	100		
91	170.5	135.7	203.5






#53
t-1,3-Dichloropropene
Concen: 48.158 ug/l
RT: 8.970 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 75 Resp: 15951


Ion	Ratio	Lower	Upper
75	100		
77	31.5	24.4	36.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#54
cis-1,3-Dichloropropene
Concen: 48.923 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 75 Resp: 175713

Ion	Ratio	Lower	Upper
75	100		
77	31.4	25.1	37.7
39	48.2	40.3	60.5






#55
1,1,2-Trichloroethane
Concen: 49.814 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 97 Resp: 10063


Ion	Ratio	Lower	Upper
97	100		
83	87.8	68.8	103.2
85	57.4	43.7	65.5
99	63.6	49.5	74.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#56
Ethyl methacrylate
Concen: 49.203 ug/l
RT: 9.110 min Scan# 1316
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 69 Resp: 170918

Ion	Ratio	Lower	Upper
69	100		
41	67.3	53.2	79.8
39	37.4	30.2	45.2






#57
1,3-Dichloropropane
Concen: 50.198 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 76 Resp: 17269


Ion	Ratio	Lower	Upper
76	100		
78	32.1	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#58
2-Chloroethyl Vinyl ether
Concen: 250.814 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 63 Resp: 427737

Ion	Ratio	Lower	Upper
63	100		
106	27.2	23.0	34.6





#59
2-Hexanone
Concen: 253.231 ug/l
RT: 9.427 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52


Tgt Ion: 43 Resp: 53102


Ion	Ratio	Lower	Upper
43	100		
58	55.3	27.8	83.3

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#60
Dibromochloromethane
Concen: 47.219 ug/l
RT: 9.513 min Scan# 1382
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:129 Resp: 111967

Ion	Ratio	Lower	Upper
129	100		
127	77.4	38.5	115.5






#61
1,2-Dibromoethane
Concen: 48.405 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:107 Resp: 10248


Ion	Ratio	Lower	Upper
107	100		
109	94.9	75.3	112.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#62
4-Bromofluorobenzene
Concen: 52.925 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 95 Resp: 130114

Ion	Ratio	Lower	Upper
95	100		
174	71.5	0.0	133.0
176	69.4	0.0	129.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:117 Resp: 25241


Ion	Ratio	Lower	Upper
117	100		
82	58.6	47.6	71.4
119	32.3	25.4	38.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#64
Tetrachloroethene
Concen: 49.701 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:164 Resp: 78776

Ion	Ratio	Lower	Upper
164	100		
166	130.9	102.0	153.0
129	101.2	81.1	121.7
131	96.8	79.0	118.6






#65
Chlorobenzene
Concen: 51.839 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:112 Resp: 27793


Ion	Ratio	Lower	Upper
112	100		
114	30.8	25.3	37.9



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 47.975 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:131 Resp: 93479

Ion	Ratio	Lower	Upper
131	100		
133	94.4	47.8	143.3
119	68.3	33.4	100.2





#67
Ethyl Benzene
Concen: 53.125 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52


Tgt Ion: 91 Resp: 49698


Ion	Ratio	Lower	Upper
91	100		
106	31.2	25.0	37.6

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#68
m/p-Xylenes
Concen: 107.068 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:106 Resp: 371687

Ion	Ratio	Lower	Upper
106	100		
91	203.5	166.4	249.6






#69
o-Xylene
Concen: 51.643 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:106 Resp: 18130


Ion	Ratio	Lower	Upper
106	100		
91	215.9	108.7	326.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#70
Styrene
Concen: 53.453 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:104 Resp: 302980

Ion	Ratio	Lower	Upper
104	100		
78	50.3	43.0	64.6
103	54.2	43.4	65.2






#71
Bromoform
Concen: 46.133 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:173 Resp: 7316


Ion	Ratio	Lower	Upper
173	100		
175	47.3	23.6	70.8
254	0.0	0.1	0.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:152 Resp: 101849

Ion	Ratio	Lower	Upper
152	100		
115	93.5	43.8	131.4
150	174.1	0.0	347.8






#73
Isopropylbenzene
Concen: 55.630 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:105 Resp: 46286


Ion	Ratio	Lower	Upper
105	100		
120	26.1	13.2	39.6



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#74
N-amyl acetate
Concen: 52.255 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 43 Resp: 214304

Ion	Ratio	Lower	Upper
43	100		
70	43.0	35.2	52.8
55	23.1	19.0	28.4
61	24.3	20.0	30.0






#75
1,1,2,2-Tetrachloroethane
Concen: 52.692 ug/l
RT: 11.207 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 83 Resp: 15141


Ion	Ratio	Lower	Upper
83	100		
131	9.0	5.0	14.9
85	63.5	32.0	96.2



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#76
1,2,3-Trichloropropane
Concen: 51.297 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 75 Resp: 126793

Ion	Ratio	Lower	Upper
75	100		
77	42.9	23.4	70.2






#77
Bromobenzene
Concen: 53.164 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:156 Resp: 10339


Ion	Ratio	Lower	Upper
156	100		
77	176.5	90.3	271.0
158	97.8	48.4	145.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#78
n-propylbenzene
Concen: 59.244 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 91 Resp: 543610

Ion	Ratio	Lower	Upper
91	100		
120	22.3	11.4	34.1






#79
2-Chlorotoluene
Concen: 53.730 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 91 Resp: 31847


Ion	Ratio	Lower	Upper
91	100		
126	32.7	16.2	48.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#80
1,3,5-Trimethylbenzene
Concen: 57.296 ug/l
RT: 11.445 min Scan# 1699
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:105 Resp: 388082

Ion	Ratio	Lower	Upper
105	100		
120	46.4	23.7	71.1






#81
trans-1,4-Dichloro-2-butene
Concen: 50.591 ug/l
RT: 11.012 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 75 Resp: 5468


Ion	Ratio	Lower	Upper
75	100		
53	89.9	74.3	111.5
89	44.7	37.4	56.0



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#82
4-Chlorotoluene
Concen: 57.791 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 91 Resp: 375241

Ion	Ratio	Lower	Upper
91	100		
126	27.5	13.9	41.7






#83
tert-Butylbenzene
Concen: 50.199 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:119 Resp: 35552


Ion	Ratio	Lower	Upper
119	100		
91	62.7	30.0	90.0
134	22.7	11.2	33.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#84
1,2,4-Trimethylbenzene
Concen: 51.077 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:105 Resp: 351562

Ion	Ratio	Lower	Upper
105	100		
120	43.2	22.2	66.6






#85
sec-Butylbenzene
Concen: 52.148 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:105 Resp: 43002


Ion	Ratio	Lower	Upper
105	100		
134	19.9	9.8	29.5



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#86
p-Isopropyltoluene
Concen: 53.781 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:119 Resp: 364463

Ion	Ratio	Lower	Upper
119	100		
134	25.5	12.8	38.4
91	27.0	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 53.304 ug/l
RT: 11.963 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:146 Resp: 17977


Ion	Ratio	Lower	Upper
146	100		
111	43.2	21.9	65.7
148	63.2	31.8	95.4



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#88
1,4-Dichlorobenzene
Concen: 51.704 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:146 Resp: 176282

Ion	Ratio	Lower	Upper
146	100		
111	43.0	21.5	64.5
148	63.7	32.1	96.5






#89
n-Butylbenzene
Concen: 58.741 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 91 Resp: 34190


Ion	Ratio	Lower	Upper
91	100		
92	55.0	27.4	82.0
134	24.8	12.8	38.3



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#90
Hexachloroethane
Concen: 50.955 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:117 Resp: 78195

Ion	Ratio	Lower	Upper
117	100		
201	64.9	32.3	96.8






#91
1,2-Dichlorobenzene
Concen: 52.880 ug/l
RT: 12.329 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:146 Resp: 18315


Ion	Ratio	Lower	Upper
146	100		
111	45.7	22.4	67.2
148	63.1	31.6	94.8



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 41.712 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion: 75 Resp: 29314

Ion	Ratio	Lower	Upper
75	100		
155	83.1	39.0	116.9
157	105.1	49.9	149.7






#93
1,2,4-Trichlorobenzene
Concen: 58.139 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:180 Resp: 12038


Ion	Ratio	Lower	Upper
180	100		
182	94.3	47.4	142.2
145	32.2	16.7	50.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#94
Hexachlorobutadiene
Concen: 57.977 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:225 Resp: 47155

Ion	Ratio	Lower	Upper
225	100		
223	61.6	31.9	95.7
227	63.2	31.8	95.4






#95
Naphthalene
Concen: 54.557 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:128 Resp: 43569


Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.3	15.5
129	11.0	8.7	13.1



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025



#96
1,2,3-Trichlorobenzene
Concen: 55.837 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX044802.D
Acq: 03 Feb 2025 09:52

Tgt Ion:180 Resp: 119840

Ion	Ratio	Lower	Upper
180	100		
182	94.3	47.2	141.6
145	34.7	17.5	52.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044802.D
Acq On : 03 Feb 2025 09:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 04 02:00:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2	T	Dichlorodifluoromethane	0.673	0.642	4.6	80	0.00
3	P	Chloromethane	0.728	0.755	-3.7	90	0.00
4	C	Vinyl Chloride	0.719	0.766	-6.5#	93	0.00
5	T	Bromomethane	0.172	0.198	-15.1	99	0.00
6	T	Chloroethane	0.132	0.138	-4.5	100	0.00
7	T	Trichlorofluoromethane	0.795	0.955	-20.1	106	0.00
8	T	Diethyl Ether	0.321	0.317	1.2	92	0.00
9	T	1,1,2-Trichlorotrifluoroeth	0.597	0.639	-7.0	98	0.00
10	T	Methyl Iodide	0.815	0.817	-0.2	86	0.00
11	T	Tert butyl alcohol	0.168	0.137	18.5	77	0.00
12	CM	1,1-Dichloroethene	0.585	0.636	-8.7#	96	0.00
13	T	Acrolein	0.079	0.068	13.9	86	0.00
14	T	Allyl chloride	1.140	1.126	1.2	91	0.00
15	T	Acrylonitrile	0.330	0.350	-6.1	95	0.00
16	T	Acetone	0.299	0.290	3.0	92	0.00
17	T	Carbon Disulfide	1.597	1.688	-5.7	93	0.00
18	T	Methyl Acetate	0.921	0.987	-7.2	95	0.00
19	T	Methyl tert-butyl Ether	2.121	2.080	1.9	87	0.00
20	T	Methylene Chloride	0.695	0.700	-0.7	91	0.00
21	T	trans-1,2-Dichloroethene	0.611	0.621	-1.6	93	0.00
22	T	Diisopropyl ether	1.981	2.179	-10.0	97	0.00
23	T	Vinyl Acetate	1.734	1.796	-3.6	93	0.00
24	P	1,1-Dichloroethane	1.175	1.223	-4.1	92	0.00
25	T	2-Butanone	0.444	0.459	-3.4	92	0.00
26	T	2,2-Dichloropropane	1.139	1.035	9.1	84	0.00
27	T	cis-1,2-Dichloroethene	0.769	0.776	-0.9	92	0.00
28	T	Bromochloromethane	0.549	0.559	-1.8	90	0.00
29	T	Tetrahydrofuran	0.286	0.296	-3.5	94	0.00
30	C	Chloroform	1.186	1.162	2.0#	87	0.00
31	T	Cyclohexane	0.950	1.091	-14.8	100	0.00
32	T	1,1,1-Trichloroethane	1.090	0.991	9.1	85	0.00
33	S	1,2-Dichloroethane-d4	0.755	0.726	3.8	83	0.00
34	I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35	S	Dibromofluoromethane	0.318	0.333	-4.7	89	0.00
36	T	1,1-Dichloropropene	0.442	0.448	-1.4	93	0.00
37	T	Ethyl Acetate	0.471	0.492	-4.5	90	0.00
38	T	Carbon Tetrachloride	0.497	0.454	8.7	84	0.00
39	T	Methylcyclohexane	0.587	0.631	-7.5	97	0.00
40	TM	Benzene	1.370	1.440	-5.1	93	0.00
41	T	Methacrylonitrile	0.274	0.277	-1.1	92	0.00
42	TM	1,2-Dichloroethane	0.500	0.454	9.2	80	0.00
43	T	Isopropyl Acetate	0.866	0.852	1.6	88	0.00
44	TM	Trichloroethene	0.329	0.337	-2.4	90	0.00
45	C	1,2-Dichloropropane	0.345	0.367	-6.4#	95	0.00
46	T	Dibromomethane	0.263	0.248	5.7	86	0.00
47	T	Bromodichloromethane	0.538	0.502	6.7	84	0.00
48	T	Methyl methacrylate	0.402	0.400	0.5	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044802.D
Acq On : 03 Feb 2025 09:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 04 02:00:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	0.008	0.008	0.0	81	0.00
50	S	Toluene-d8	1.108	1.249	-12.7	95	0.00
51	T	4-Methyl-2-Pentanone	0.479	0.493	-2.9	93	0.00
52	CM	Toluene	0.839	0.879	-4.8#	92	0.00
53	T	t-1,3-Dichloropropene	0.556	0.536	3.6	83	0.00
54	T	cis-1,3-Dichloropropene	0.603	0.590	2.2	87	0.00
55	T	1,1,2-Trichloroethane	0.339	0.338	0.3	90	0.00
56	T	Ethyl methacrylate	0.583	0.574	1.5	86	0.00
57	T	1,3-Dichloropropane	0.578	0.580	-0.3	88	0.00
58	T	2-Chloroethyl Vinyl ether	0.286	0.287	-0.3	89	0.00
59	T	2-Hexanone	0.352	0.357	-1.4	92	0.00
60	T	Dibromochloromethane	0.398	0.376	5.5	84	0.00
61	T	1,2-Dibromoethane	0.356	0.344	3.4	87	0.00
62	S	4-Bromofluorobenzene	0.413	0.437	-5.8	88	0.00
63	I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64	T	Tetrachloroethene	0.314	0.312	0.6	88	0.00
65	PM	Chlorobenzene	1.062	1.101	-3.7	92	0.00
66	T	1,1,1,2-Tetrachloroethane	0.386	0.370	4.1	85	0.00
67	C	Ethyl Benzene	1.853	1.969	-6.3#	93	0.00
68	T	m/p-Xylenes	0.688	0.736	-7.0	94	0.00
69	T	o-Xylene	0.695	0.718	-3.3	90	0.00
70	T	Styrene	1.123	1.200	-6.9	92	0.00
71	P	Bromoform	0.314	0.290	7.6	81	0.00
72	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73	T	Isopropylbenzene	4.085	4.545	-11.3	92	0.00
74	T	N-amyl acetate	2.013	2.104	-4.5	88	0.00
75	P	1,1,2,2-Tetrachloroethane	1.411	1.487	-5.4	91	0.00
76	T	1,2,3-Trichloropropane	1.213	1.245	-2.6	90	0.00
77	T	Bromobenzene	0.955	1.015	-6.3	90	0.00
78	T	n-propylbenzene	4.505	5.337	-18.5	97	0.00
79	T	2-Chlorotoluene	2.910	3.127	-7.5	90	0.00
80	T	1,3,5-Trimethylbenzene	3.325	3.810	-14.6	95	0.00
81	T	trans-1,4-Dichloro-2-butene	0.531	0.537	-1.1	84	0.00
82	T	4-Chlorotoluene	3.188	3.684	-15.6	95	0.00
83	T	tert-Butylbenzene	3.477	3.491	-0.4	84	0.00
84	T	1,2,4-Trimethylbenzene	3.379	3.452	-2.2	86	0.00
85	T	sec-Butylbenzene	4.048	4.222	-4.3	86	0.00
86	T	p-Isopropyltoluene	3.327	3.578	-7.5	88	0.00
87	T	1,3-Dichlorobenzene	1.656	1.765	-6.6	87	0.00
88	T	1,4-Dichlorobenzene	1.674	1.731	-3.4	87	0.00
89	T	n-Butylbenzene	2.857	3.357	-17.5	95	0.00
90	T	Hexachloroethane	0.753	0.768	-2.0	89	0.00
91	T	1,2-Dichlorobenzene	1.700	1.798	-5.8	88	0.00
92	T	1,2-Dibromo-3-Chloropropane	0.345	0.288	16.5	74	0.00
93	T	1,2,4-Trichlorobenzene	1.016	1.182	-16.3	91	0.00
94	T	Hexachlorobutadiene	0.399	0.463	-16.0	94	0.00
95	T	Naphthalene	3.920	4.278	-9.1	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044802.D
Acq On : 03 Feb 2025 09:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 04 02:00:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.054	1.177	-11.7	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044802.D
Acq On : 03 Feb 2025 09:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 04 02:00:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Pentafluorobenzene	50.000	50.000	0.0	85	0.00
2	T	Dichlorodifluoromethane	50.000	47.741	4.5	80	0.00
3	P	Chloromethane	50.000	51.877	-3.8	90	0.00
4	C	Vinyl Chloride	50.000	53.283	-6.6#	93	0.00
5	T	Bromomethane	50.000	57.578	-15.2	99	0.00
6	T	Chloroethane	50.000	52.365	-4.7	100	0.00
7	T	Trichlorofluoromethane	50.000	60.090	-20.2	106	0.00
8	T	Diethyl Ether	50.000	49.453	1.1	92	0.00
9	T	1,1,2-Trichlorotrifluoroeth	50.000	53.500	-7.0	98	0.00
10	T	Methyl Iodide	50.000	50.112	-0.2	86	0.00
11	T	Tert butyl alcohol	250.000	204.000	18.4	77	0.00
12	CM	1,1-Dichloroethene	50.000	54.321	-8.6#	96	0.00
13	T	Acrolein	250.000	216.204	13.5	86	0.00
14	T	Allyl chloride	50.000	49.381	1.2	91	0.00
15	T	Acrylonitrile	250.000	265.265	-6.1	95	0.00
16	T	Acetone	250.000	242.503	3.0	92	0.00
17	T	Carbon Disulfide	50.000	52.823	-5.6	93	0.00
18	T	Methyl Acetate	50.000	53.585	-7.2	95	0.00
19	T	Methyl tert-butyl Ether	50.000	49.052	1.9	87	0.00
20	T	Methylene Chloride	50.000	50.330	-0.7	91	0.00
21	T	trans-1,2-Dichloroethene	50.000	50.775	-1.5	93	0.00
22	T	Diisopropyl ether	50.000	54.981	-10.0	97	0.00
23	T	Vinyl Acetate	250.000	258.911	-3.6	93	0.00
24	P	1,1-Dichloroethane	50.000	52.029	-4.1	92	0.00
25	T	2-Butanone	250.000	258.709	-3.5	92	0.00
26	T	2,2-Dichloropropane	50.000	45.456	9.1	84	0.00
27	T	cis-1,2-Dichloroethene	50.000	50.436	-0.9	92	0.00
28	T	Bromochloromethane	50.000	50.943	-1.9	90	0.00
29	T	Tetrahydrofuran	250.000	258.796	-3.5	94	0.00
30	C	Chloroform	50.000	49.012	2.0#	87	0.00
31	T	Cyclohexane	50.000	57.418	-14.8	100	0.00
32	T	1,1,1-Trichloroethane	50.000	45.461	9.1	85	0.00
33	S	1,2-Dichloroethane-d4	50.000	48.114	3.8	83	0.00
34	I	1,4-Difluorobenzene	50.000	50.000	0.0	85	0.00
35	S	Dibromofluoromethane	50.000	52.467	-4.9	89	0.00
36	T	1,1-Dichloropropene	50.000	50.662	-1.3	93	0.00
37	T	Ethyl Acetate	50.000	52.295	-4.6	90	0.00
38	T	Carbon Tetrachloride	50.000	45.680	8.6	84	0.00
39	T	Methylcyclohexane	50.000	53.731	-7.5	97	0.00
40	TM	Benzene	50.000	52.535	-5.1	93	0.00
41	T	Methacrylonitrile	50.000	50.508	-1.0	92	0.00
42	TM	1,2-Dichloroethane	50.000	45.355	9.3	80	0.00
43	T	Isopropyl Acetate	50.000	49.187	1.6	88	0.00
44	TM	Trichloroethene	50.000	51.104	-2.2	90	0.00
45	C	1,2-Dichloropropane	50.000	53.206	-6.4#	95	0.00
46	T	Dibromomethane	50.000	47.313	5.4	86	0.00
47	T	Bromodichloromethane	50.000	46.626	6.7	84	0.00
48	T	Methyl methacrylate	50.000	49.744	0.5	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044802.D
Acq On : 03 Feb 2025 09:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 04 02:00:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49	T	1,4-Dioxane	1000.000	903.978	9.6	81	0.00
50	S	Toluene-d8	50.000	56.382	-12.8	95	0.00
51	T	4-Methyl-2-Pentanone	250.000	257.519	-3.0	93	0.00
52	CM	Toluene	50.000	52.403	-4.8#	92	0.00
53	T	t-1,3-Dichloropropene	50.000	48.158	3.7	83	0.00
54	T	cis-1,3-Dichloropropene	50.000	48.923	2.2	87	0.00
55	T	1,1,2-Trichloroethane	50.000	49.814	0.4	90	0.00
56	T	Ethyl methacrylate	50.000	49.203	1.6	86	0.00
57	T	1,3-Dichloropropane	50.000	50.198	-0.4	88	0.00
58	T	2-Chloroethyl Vinyl ether	250.000	250.814	-0.3	89	0.00
59	T	2-Hexanone	250.000	253.231	-1.3	92	0.00
60	T	Dibromochloromethane	50.000	47.219	5.6	84	0.00
61	T	1,2-Dibromoethane	50.000	48.405	3.2	87	0.00
62	S	4-Bromofluorobenzene	50.000	52.925	-5.8	88	0.00
63	I	Chlorobenzene-d5	50.000	50.000	0.0	84	0.00
64	T	Tetrachloroethene	50.000	49.701	0.6	88	0.00
65	PM	Chlorobenzene	50.000	51.839	-3.7	92	0.00
66	T	1,1,1,2-Tetrachloroethane	50.000	47.975	4.0	85	0.00
67	C	Ethyl Benzene	50.000	53.125	-6.3#	93	0.00
68	T	m/p-Xylenes	100.000	107.068	-7.1	94	0.00
69	T	o-Xylene	50.000	51.643	-3.3	90	0.00
70	T	Styrene	50.000	53.453	-6.9	92	0.00
71	P	Bromoform	50.000	46.133	7.7	81	0.00
72	I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	80	0.00
73	T	Isopropylbenzene	50.000	55.630	-11.3	92	0.00
74	T	N-amyl acetate	50.000	52.255	-4.5	88	0.00
75	P	1,1,2,2-Tetrachloroethane	50.000	52.692	-5.4	91	0.00
76	T	1,2,3-Trichloropropane	50.000	51.297	-2.6	90	0.00
77	T	Bromobenzene	50.000	53.164	-6.3	90	0.00
78	T	n-propylbenzene	50.000	59.244	-18.5	97	0.00
79	T	2-Chlorotoluene	50.000	53.730	-7.5	90	0.00
80	T	1,3,5-Trimethylbenzene	50.000	57.296	-14.6	95	0.00
81	T	trans-1,4-Dichloro-2-butene	50.000	50.591	-1.2	84	0.00
82	T	4-Chlorotoluene	50.000	57.791	-15.6	95	0.00
83	T	tert-Butylbenzene	50.000	50.199	-0.4	84	0.00
84	T	1,2,4-Trimethylbenzene	50.000	51.077	-2.2	86	0.00
85	T	sec-Butylbenzene	50.000	52.148	-4.3	86	0.00
86	T	p-Isopropyltoluene	50.000	53.781	-7.6	88	0.00
87	T	1,3-Dichlorobenzene	50.000	53.304	-6.6	87	0.00
88	T	1,4-Dichlorobenzene	50.000	51.704	-3.4	87	0.00
89	T	n-Butylbenzene	50.000	58.741	-17.5	95	0.00
90	T	Hexachloroethane	50.000	50.955	-1.9	89	0.00
91	T	1,2-Dichlorobenzene	50.000	52.880	-5.8	88	0.00
92	T	1,2-Dibromo-3-Chloropropane	50.000	41.712	16.6	74	0.00
93	T	1,2,4-Trichlorobenzene	50.000	58.139	-16.3	91	0.00
94	T	Hexachlorobutadiene	50.000	57.977	-16.0	94	0.00
95	T	Naphthalene	50.000	54.557	-9.1	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044802.D
Acq On : 03 Feb 2025 09:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 04 02:00:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	55.837	-11.7	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6


QC SAMPLE DATA


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011425\
Data File : VN085437.D
Acq On : 14 Jan 2025 14:22
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Title : SW846 8260
Last Update : Wed Jan 15 02:16:08 2025





AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	4864	PASS
75	95	30	60	58.0	12679	PASS
95	95	100	100	100.0	21856	PASS
96	95	5	9	7.8	1703	PASS
173	174	0.00	2	1.8	296	PASS
174	95	50	100	76.0	16613	PASS
175	174	5	9	7.1	1179	PASS
176	174	95	101	97.4	16185	PASS
177	176	5	9	6.7	1079	PASS


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085590.D
Acq On : 31 Jan 2025 09:30
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Title : SW846 8260
Last Update : Thu Jan 16 01:16:09 2025





AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.8	5647	PASS
75	95	30	60	53.5	12707	PASS
95	95	100	100	100.0	23757	PASS
96	95	5	9	7.2	1718	PASS
173	174	0.00	2	1.4	245	PASS
174	95	50	100	76.1	18083	PASS
175	174	5	9	8.3	1508	PASS
176	174	95	101	97.1	17560	PASS
177	176	5	9	6.0	1060	PASS


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011525\
Data File : VX044647.D
Acq On : 15 Jan 2025 09:30
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Title : SW846 8260
Last Update : Thu Jan 16 01:16:09 2025





AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.6	10879	PASS
75	95	30	60	50.2	29285	PASS
95	95	100	100	100.0	58391	PASS
96	95	5	9	6.5	3805	PASS
173	174	0.00	2	0.9	350	PASS
174	95	50	100	67.1	39163	PASS
175	174	5	9	7.3	2842	PASS
176	174	95	101	96.0	37587	PASS
177	176	5	9	6.8	2538	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044801.D
Acq On : 03 Feb 2025 08:12
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Title : SW846 8260
Last Update : Thu Jan 16 01:16:09 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	11987	PASS
75	95	30	60	51.1	33397	PASS
95	95	100	100	100.0	65304	PASS
96	95	5	9	6.8	4458	PASS
173	174	0.00	2	1.0	492	PASS
174	95	50	100	77.6	50659	PASS
175	174	5	9	7.6	3833	PASS
176	174	95	101	96.1	48672	PASS
177	176	5	9	6.6	3220	PASS

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	VN0131WBL01	SDG No.:	Q1215
Lab Sample ID:	VN0131WBL01	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085592.D	1		01/31/25 10:42	VN013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5.00	U	0.34	5.00	ug/L
75-35-4	1,1-Dichloroethene	5.00	U	0.26	5.00	ug/L
78-93-3	2-Butanone	25.0	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	5.00	U	0.25	5.00	ug/L
67-66-3	Chloroform	5.00	U	0.26	5.00	ug/L
71-43-2	Benzene	5.00	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	5.00	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	5.00	U	0.32	5.00	ug/L
127-18-4	Tetrachloroethene	5.00	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	5.00	U	0.13	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	57.3		74 - 125	115%	SPK: 50
1868-53-7	Dibromofluoromethane	54.1		75 - 124	108%	SPK: 50
2037-26-5	Toluene-d8	49.7		86 - 113	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.4		77 - 121	91%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	164000	8.224			
540-36-3	1,4-Difluorobenzene	305000	9.1			
3114-55-4	Chlorobenzene-d5	267000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	105000	13.788			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085592.D
Acq On : 31 Jan 2025 10:42
Operator : JC\MD
Sample : VN0131WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01

Quant Time: Jan 31 22:24:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	163540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	267449	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	104737	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	151392	57.348	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.700%			
35) Dibromofluoromethane	8.165	113	114315	54.068	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.140%			
50) Toluene-d8	10.565	98	373656	49.741	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.480%			
62) 4-Bromofluorobenzene	12.847	95	116594	45.373	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 90.740%			
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085592.D
Acq On : 31 Jan 2025 10:42
Operator : JC\MD
Sample : VN0131WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01

Quant Time: Jan 31 22:24:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085592.D
Acq: 31 Jan 2025 10:42

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01

Tgt Ion:168 Resp: 163540

Ion	Ratio	Lower	Upper
168	100		
99	64.9	53.6	80.4







#33
1,2-Dichloroethane-d4
Concen: 57.348 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN085592.D
Acq: 31 Jan 2025 10:42

Tgt Ion: 65 Resp: 151392

Ion	Ratio	Lower	Upper
65	100		
67	50.2	0.0	101.6







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 12
Delta R.T. 0.006 min
Lab File: VN085592.D
Acq: 31 Jan 2025 10:42

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01

Tgt Ion:114 Resp: 304758

Ion	Ratio	Lower	Upper
114	100		
63	21.2	0.0	47.6
88	16.0	0.0	32.6







#35
Dibromofluoromethane
Concen: 54.068 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN085592.D
Acq: 31 Jan 2025 10:42

Tgt Ion:113 Resp: 114315

Ion	Ratio	Lower	Upper
113	100		
111	103.3	82.7	124.1
192	18.0	14.3	21.5







#50
Toluene-d8
Concen: 49.741 ug/l
RT: 10.565 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN085592.D
Acq: 31 Jan 2025 10:42

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01

Tgt Ion: 98 Resp: 373656

Ion	Ratio	Lower	Upper
98	100		
100	63.7	52.2	78.4







#62
4-Bromofluorobenzene
Concen: 45.373 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085592.D
Acq: 31 Jan 2025 10:42

Tgt Ion: 95 Resp: 116594

Ion	Ratio	Lower	Upper
95	100		
174	75.3	0.0	145.0
176	72.4	0.0	142.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085592.D
Acq: 31 Jan 2025 10:42

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01

Tgt Ion:117 Resp: 267449

Ion	Ratio	Lower	Upper
117	100		
82	59.4	48.6	72.8
119	32.2	26.6	39.8






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085592.D
Acq: 31 Jan 2025 10:42

Tgt Ion:152 Resp: 104737

Ion	Ratio	Lower	Upper
152	100		
115	62.0	31.1	93.3
150	158.0	0.0	343.6



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	VX0203WBL01	SDG No.:	Q1215
Lab Sample ID:	VX0203WBL01	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044804.D	1		02/03/25 10:44	VX020325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5.00	U	0.34	5.00	ug/L
75-35-4	1,1-Dichloroethene	5.00	U	0.26	5.00	ug/L
78-93-3	2-Butanone	25.0	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	5.00	U	0.25	5.00	ug/L
67-66-3	Chloroform	5.00	U	0.26	5.00	ug/L
71-43-2	Benzene	5.00	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	5.00	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	5.00	U	0.32	5.00	ug/L
127-18-4	Tetrachloroethene	5.00	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	5.00	U	0.13	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.6		74 - 125	97%	SPK: 50
1868-53-7	Dibromofluoromethane	50.1		75 - 124	100%	SPK: 50
2037-26-5	Toluene-d8	55.2		86 - 113	110%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.9		77 - 121	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	125000	5.544			
540-36-3	1,4-Difluorobenzene	258000	6.757			
3114-55-4	Chlorobenzene-d5	233000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	101000	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044804.D
Acq On : 03 Feb 2025 10:44
Operator : JC/MD
Sample : VX0203WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBL01

Quant Time: Feb 04 02:05:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	124993	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	257688	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	232979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	100905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	91752	48.632	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 97.260%			
35) Dibromofluoromethane	5.379	113	82089	50.144	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.280%			
50) Toluene-d8	8.647	98	315174	55.215	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.440%			
62) 4-Bromofluorobenzene	11.079	95	112520	52.878	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.760%			
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044804.D
Acq On : 03 Feb 2025 10:44
Operator : JC/MD
Sample : VX0203WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBL01

Quant Time: Feb 04 02:05:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX044804.D
Acq: 03 Feb 2025 10:44

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBL01

Tgt Ion:168 Resp: 124993

Ion	Ratio	Lower	Upper
168	100		
99	64.0	51.1	76.7







#33
1,2-Dichloroethane-d4
Concen: 48.632 ug/l
RT: 5.946 min Scan# 797
Delta R.T. -0.006 min
Lab File: VX044804.D
Acq: 03 Feb 2025 10:44

Tgt Ion: 65 Resp: 91752

Ion	Ratio	Lower	Upper
65	100		
67	54.6	0.0	103.4







#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 9
Delta R.T. 0.000 min
Lab File: VX044804.D
Acq: 03 Feb 2025 10:44

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBL01

Tgt Ion:114 Resp: 257688

Ion	Ratio	Lower	Upper
114	100		
63	21.3	0.0	43.0
88	16.7	0.0	35.4







#35
Dibromofluoromethane
Concen: 50.144 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX044804.D
Acq: 03 Feb 2025 10:44

Tgt Ion:113 Resp: 82089

Ion	Ratio	Lower	Upper
113	100		
111	102.2	81.6	122.4
192	17.7	14.4	21.6







#50
Toluene-d8
Concen: 55.215 ug/l
RT: 8.647 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044804.D
Acq: 03 Feb 2025 10:44

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBL01

Tgt Ion: 98 Resp: 315174

Ion	Ratio	Lower	Upper
98	100		
100	65.7	53.2	79.8







#62
4-Bromofluorobenzene
Concen: 52.878 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044804.D
Acq: 03 Feb 2025 10:44

Tgt Ion: 95 Resp: 112520

Ion	Ratio	Lower	Upper
95	100		
174	71.4	0.0	133.0
176	68.5	0.0	129.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044804.D
Acq: 03 Feb 2025 10:44

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBL01

Tgt Ion:117 Resp: 232979

Ion	Ratio	Lower	Upper
117	100		
82	57.8	47.6	71.4
119	31.9	25.4	38.2






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX044804.D
Acq: 03 Feb 2025 10:44

Tgt Ion:152 Resp: 100905

Ion	Ratio	Lower	Upper
152	100		
115	61.2	43.8	131.4
150	157.6	0.0	347.8



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	VN0131WBS01	SDG No.:	Q1215
Lab Sample ID:	VN0131WBS01	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085593.D	1		01/31/25 11:06	VN013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	16.8		0.34	5.00	ug/L
75-35-4	1,1-Dichloroethene	17.0		0.26	5.00	ug/L
78-93-3	2-Butanone	97.0		1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	19.0		0.25	5.00	ug/L
67-66-3	Chloroform	19.2		0.26	5.00	ug/L
71-43-2	Benzene	18.9		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	19.5		0.24	5.00	ug/L
79-01-6	Trichloroethene	18.3		0.32	5.00	ug/L
127-18-4	Tetrachloroethene	20.2		0.25	5.00	ug/L
108-90-7	Chlorobenzene	18.9		0.13	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.3		74 - 125	101%	SPK: 50
1868-53-7	Dibromofluoromethane	52.5		75 - 124	105%	SPK: 50
2037-26-5	Toluene-d8	53.6		86 - 113	107%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.8		77 - 121	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	193000	8.224			
540-36-3	1,4-Difluorobenzene	321000	9.1			
3114-55-4	Chlorobenzene-d5	286000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	135000	13.788			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085593.D
Acq On : 31 Jan 2025 11:06
Operator : JC\MD
Sample : VN0131WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jan 31 22:24:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	192673	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	320848	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	286465	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	135464	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	156371	50.278	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.560%			
35) Dibromofluoromethane	8.165	113	116861	52.501	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.000%			
50) Toluene-d8	10.565	98	423886	53.598	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.200%			
62) 4-Bromofluorobenzene	12.847	95	137397	50.788	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.580%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.124	85	41293	15.829	ug/l	99
3) Chloromethane	2.359	50	43303	15.331	ug/l	97
4) Vinyl Chloride	2.512	62	47683	16.795	ug/l	89
5) Bromomethane	2.959	94	30192	17.605	ug/l	89
6) Chloroethane	3.124	64	31734	17.631	ug/l	90
7) Trichlorofluoromethane	3.495	101	74659	18.124	ug/l	91
8) Diethyl Ether	3.959	74	23279	16.358	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	43198	18.618	ug/l	99
10) Methyl Iodide	4.589	142	46104	17.352	ug/l	94
11) Tert butyl alcohol	5.512	59	35618	100.014	ug/l #	85
12) 1,1-Dichloroethene	4.342	96	35063	16.956	ug/l	99
13) Acrolein	4.183	56	30327	62.382	ug/l	95
14) Allyl chloride	5.018	41	54768	16.322	ug/l	99
15) Acrylonitrile	5.718	53	109503	96.803	ug/l	99
16) Acetone	4.424	43	96448	95.976	ug/l	94
17) Carbon Disulfide	4.712	76	89788	14.103	ug/l	97
18) Methyl Acetate	5.018	43	65797	21.532	ug/l	100
19) Methyl tert-butyl Ether	5.794	73	126200	18.795	ug/l	97
20) Methylene Chloride	5.271	84	45794	18.408	ug/l	98
21) trans-1,2-Dichloroethene	5.789	96	38369	17.362	ug/l	93
22) Diisopropyl ether	6.671	45	143433	19.259	ug/l	97
23) Vinyl Acetate	6.600	43	475426	91.217	ug/l	100
24) 1,1-Dichloroethane	6.571	63	86655	19.082	ug/l	100
25) 2-Butanone	7.483	43	143410	96.976	ug/l	98
26) 2,2-Dichloropropane	7.483	77	75360	20.525	ug/l	100
27) cis-1,2-Dichloroethene	7.488	96	45788	17.591	ug/l	96
28) Bromochloromethane	7.812	49	41206	19.503	ug/l	100
29) Tetrahydrofuran	7.835	42	91268	97.344	ug/l	99
30) Chloroform	7.965	83	89934	19.161	ug/l	95
31) Cyclohexane	8.253	56	60554	15.341	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	78856	19.154	ug/l	98
36) 1,1-Dichloropropene	8.371	75	56124	17.965	ug/l	97
37) Ethyl Acetate	7.559	43	59547	18.884	ug/l	99
38) Carbon Tetrachloride	8.359	117	68090	19.038	ug/l	98
39) Methylcyclohexane	9.600	83	47355	16.131	ug/l	95
40) Benzene	8.606	78	176989	18.856	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085593.D
Acq On : 31 Jan 2025 11:06
Operator : JC\MD
Sample : VN0131WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jan 31 22:24:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
41) Methacrylonitrile	7.771	41	33484	20.411 ug/l	95
42) 1,2-Dichloroethane	8.671	62	69061	19.535 ug/l	100
43) Isopropyl Acetate	8.682	43	102099	20.207 ug/l	99
44) Trichloroethene	9.347	130	39952	18.285 ug/l	98
45) 1,2-Dichloropropane	9.618	63	45593	19.016 ug/l	99
46) Dibromomethane	9.706	93	33666	19.460 ug/l	99
47) Bromodichloromethane	9.888	83	71425	20.265 ug/l	96
48) Methyl methacrylate	9.677	41	45383	19.959 ug/l	98
49) 1,4-Dioxane	9.694	88	14866	388.248 ug/l #	82
51) 4-Methyl-2-Pentanone	10.441	43	312333	106.528 ug/l	99
52) Toluene	10.629	92	106928	19.660 ug/l	96
53) t-1,3-Dichloropropene	10.835	75	65874	19.777 ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	68879	19.360 ug/l	97
55) 1,1,2-Trichloroethane	11.012	97	44068	20.474 ug/l	97
56) Ethyl methacrylate	10.871	69	59018	17.580 ug/l	97
57) 1,3-Dichloropropane	11.159	76	75438	20.157 ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	103394	75.697 ug/l	98
59) 2-Hexanone	11.194	43	223307	108.244 ug/l	97
60) Dibromochloromethane	11.353	129	51998	20.010 ug/l	98
61) 1,2-Dibromoethane	11.471	107	41782	19.501 ug/l	99
64) Tetrachloroethene	11.100	164	39360	20.154 ug/l	97
65) Chlorobenzene	11.888	112	118461	18.877 ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	45402	19.712 ug/l	99
67) Ethyl Benzene	11.959	91	189996	18.588 ug/l	100
68) m/p-Xylenes	12.070	106	150545	39.852 ug/l	97
69) o-Xylene	12.394	106	68176	18.885 ug/l	99
70) Styrene	12.406	104	113181	18.944 ug/l	97
71) Bromoform	12.576	173	33682	20.438 ug/l #	100
73) Isopropylbenzene	12.694	105	176447	19.299 ug/l	99
74) N-amyl acetate	12.488	43	77354	18.824 ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	64490	19.969 ug/l	99
76) 1,2,3-Trichloropropane	12.988	75	52234m	18.971 ug/l	
77) Bromobenzene	12.976	156	45326	18.976 ug/l	98
78) n-propylbenzene	13.035	91	211911	19.579 ug/l	100
79) 2-Chlorotoluene	13.123	91	135337	19.320 ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	151009	19.996 ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	19703	19.361 ug/l	99
82) 4-Chlorotoluene	13.217	91	134471	19.280 ug/l	99
83) tert-Butylbenzene	13.435	119	118897	18.748 ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	150071	19.938 ug/l	100
85) sec-Butylbenzene	13.612	105	169773	19.315 ug/l	99
86) p-Isopropyltoluene	13.729	119	135685	18.592 ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	83123	18.896 ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	85651	18.785 ug/l	97
89) n-Butylbenzene	14.053	91	111994	17.761 ug/l	100
90) Hexachloroethane	14.329	117	31211	18.652 ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	79551	18.130 ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.717	75	11200	18.991 ug/l	99
93) 1,2,4-Trichlorobenzene	15.388	180	34537	16.935 ug/l	97
94) Hexachlorobutadiene	15.500	225	19804	18.291 ug/l	99
95) Naphthalene	15.641	128	96159	15.801 ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	36004	17.448 ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085593.D
Acq On : 31 Jan 2025 11:06
Operator : JC\MD
Sample : VN0131WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jan 31 22:24:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085593.D
Acq On : 31 Jan 2025 11:06
Operator : JC\MD
Sample : VN0131WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Jan 31 22:24:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:168 Resp: 192673


Ion	Ratio	Lower	Upper
168	100		
99	67.3	53.6	80.4



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#2
Dichlorodifluoromethane
Concen: 15.829 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 85 Resp: 41293

Ion	Ratio	Lower	Upper
85	100		
87	32.7	16.7	50.1






#3
Chloromethane
Concen: 15.331 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 50 Resp: 43303


Ion	Ratio	Lower	Upper
50	100		
52	32.3	27.0	40.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#4
Vinyl Chloride
Concen: 16.795 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 62 Resp: 47683

Ion	Ratio	Lower	Upper
62	100		
64	36.9	24.8	37.2






#5
Bromomethane
Concen: 17.605 ug/l
RT: 2.959 min Scan# 17
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 94 Resp: 30192


Ion	Ratio	Lower	Upper
94	100		
96	100.5	71.8	107.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#6
Chloroethane
Concen: 17.631 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 64 Resp: 31734

Ion	Ratio	Lower	Upper
64	100		
66	36.3	24.6	36.8






#7
Trichlorofluoromethane
Concen: 18.124 ug/l
RT: 3.495 min Scan# 26
Delta R.T. -0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:101 Resp: 74659


Ion	Ratio	Lower	Upper
101	100		
103	57.6	51.4	77.2



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#8
Diethyl Ether
Concen: 16.358 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 74 Resp: 23279

Ion	Ratio	Lower	Upper
74	100		
45	99.4	49.7	149.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.618 ug/l
RT: 4.365 min Scan# 41
Delta R.T. -0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06


Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01


Tgt Ion:101 Resp: 43198

Ion	Ratio	Lower	Upper
101	100		
85	46.3	37.8	56.8
151	73.3	58.8	88.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025







#10
Methyl Iodide
Concen: 17.352 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:142 Resp: 46104

Ion	Ratio	Lower	Upper
142	100		
127	44.9	40.3	60.5
141	15.5	12.8	19.2






#11
Tert butyl alcohol
Concen: 100.014 ug/l
RT: 5.512 min Scan# 60
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 59 Resp: 35618


Ion	Ratio	Lower	Upper
59	100		
57	5.2	8.6	13.0#



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#12
1,1-Dichloroethene
Concen: 16.956 ug/l
RT: 4.342 min Scan# 406
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 96 Resp: 35063

Ion	Ratio	Lower	Upper
96	100		
61	175.0	141.0	211.4
98	58.9	49.0	73.6






#13
Acrolein
Concen: 62.382 ug/l
RT: 4.183 min Scan# 37
Delta R.T. 0.012 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 56 Resp: 30327


Ion	Ratio	Lower	Upper
56	100		
55	74.7	56.3	84.5



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#14
Allyl chloride
Concen: 16.322 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 41 Resp: 54768

Ion	Ratio	Lower	Upper
41	100		
39	81.5	64.4	96.6
76	38.4	30.5	45.7






#15
Acrylonitrile
Concen: 96.803 ug/l
RT: 5.718 min Scan# 64
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 53 Resp: 109503


Ion	Ratio	Lower	Upper
53	100		
52	82.5	65.5	98.3
51	38.2	29.8	44.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#16
Acetone
Concen: 95.976 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 43 Resp: 96448

Ion	Ratio	Lower	Upper
43	100		
58	33.0	23.8	35.6






#17
Carbon Disulfide
Concen: 14.103 ug/l
RT: 4.712 min Scan# 46
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 76 Resp: 89788


Ion	Ratio	Lower	Upper
76	100		
78	9.5	6.9	10.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#18
Methyl Acetate
Concen: 21.532 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 43 Resp: 65797

Ion	Ratio	Lower	Upper
43	100		
74	23.2	18.6	28.0






#19
Methyl tert-butyl Ether
Concen: 18.795 ug/l
RT: 5.794 min Scan# 65
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 73 Resp: 126200


Ion	Ratio	Lower	Upper
73	100		
57	21.1	17.8	26.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#20
Methylene Chloride
Concen: 18.408 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 84 Resp: 45794

Ion	Ratio	Lower	Upper
84	100		
49	132.9	104.1	156.1
51	39.6	31.0	46.4
86	62.5	51.9	77.9






#21
trans-1,2-Dichloroethene
Concen: 17.362 ug/l
RT: 5.789 min Scan# 65
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 96 Resp: 38369


Ion	Ratio	Lower	Upper
96	100		
61	132.2	115.7	173.5
98	62.6	49.8	74.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#22
Diisopropyl ether
Concen: 19.259 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 45 Resp: 143433

Ion	Ratio	Lower	Upper
45	100		
43	55.2	45.0	67.6
87	24.3	21.8	32.8
59	12.0	10.2	15.2






#23
Vinyl Acetate
Concen: 91.217 ug/l
RT: 6.600 min Scan# 79
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 43 Resp: 475426


Ion	Ratio	Lower	Upper
43	100		
86	10.4	8.4	12.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#24
1,1-Dichloroethane
Concen: 19.082 ug/l
RT: 6.571 min Scan# 785
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 63 Resp: 86655

Ion	Ratio	Lower	Upper
63	100		
98	6.5	3.3	9.8
100	4.0	2.0	6.0






#25
2-Butanone
Concen: 96.976 ug/l
RT: 7.483 min Scan# 94
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 43 Resp: 143410


Ion	Ratio	Lower	Upper
43	100		
72	24.1	20.2	30.4



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#26
2,2-Dichloropropane
Concen: 20.525 ug/l
RT: 7.483 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 77 Resp: 75360

Ion	Ratio	Lower	Upper
77	100		
97	21.5	10.7	32.1






#27
cis-1,2-Dichloroethene
Concen: 17.591 ug/l
RT: 7.488 min Scan# 94
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 96 Resp: 45788


Ion	Ratio	Lower	Upper
96	100		
61	160.8	0.0	311.8
98	66.1	0.0	126.0



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#28
Bromochloromethane
Concen: 19.503 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 49 Resp: 41206

Ion	Ratio	Lower	Upper
49	100		
129	1.5	0.0	4.4
130	68.5	55.0	82.4






#29
Tetrahydrofuran
Concen: 97.344 ug/l
RT: 7.835 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 42 Resp: 91268


Ion	Ratio	Lower	Upper
42	100		
72	43.8	34.2	51.2
71	40.2	32.5	48.7



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#30
Chloroform
Concen: 19.161 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 83 Resp: 89934

Ion	Ratio	Lower	Upper
83	100		
85	60.9	51.8	77.6






#31
Cyclohexane
Concen: 15.341 ug/l
RT: 8.253 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 56 Resp: 60554


Ion	Ratio	Lower	Upper
56	100		
69	32.7	22.2	33.4
84	83.5	66.4	99.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#32
1,1,1-Trichloroethane
Concen: 19.154 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 97 Resp: 78856

Ion	Ratio	Lower	Upper
97	100		
99	63.1	49.8	74.6
61	49.2	41.4	62.2






#33
1,2-Dichloroethane-d4
Concen: 50.278 ug/l
RT: 8.577 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 65 Resp: 156371


Ion	Ratio	Lower	Upper
65	100		
67	50.6	0.0	101.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:114 Resp: 320848

Ion	Ratio	Lower	Upper
114	100		
63	22.0	0.0	47.6
88	16.6	0.0	32.6






#35
Dibromofluoromethane
Concen: 52.501 ug/l
RT: 8.165 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:113 Resp: 116861


Ion	Ratio	Lower	Upper
113	100		
111	102.4	82.7	124.1
192	17.3	14.3	21.5



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#36
1,1-Dichloropropene
Concen: 17.965 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 75 Resp: 56124

Ion	Ratio	Lower	Upper
75	100		
110	30.9	16.5	49.5
77	29.6	24.4	36.6






#37
Ethyl Acetate
Concen: 18.884 ug/l
RT: 7.559 min Scan# 95
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 43 Resp: 59547


Ion	Ratio	Lower	Upper
43	100		
61	13.3	10.9	16.3
70	10.1	7.5	11.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#38
Carbon Tetrachloride
Concen: 19.038 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:117 Resp: 68090

Ion	Ratio	Lower	Upper
117	100		
119	97.1	75.4	113.2
121	30.9	24.6	37.0






#39
Methylcyclohexane
Concen: 16.131 ug/l
RT: 9.600 min Scan# 13
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 83 Resp: 47355


Ion	Ratio	Lower	Upper
83	100		
55	83.5	62.6	94.0
98	49.0	37.7	56.5



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#40
Benzene
Concen: 18.856 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 78 Resp: 176989

Ion	Ratio	Lower	Upper
78	100		
77	25.8	19.0	28.6






#41
Methacrylonitrile
Concen: 20.411 ug/l
RT: 7.771 min Scan# 98
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 41 Resp: 33484


Ion	Ratio	Lower	Upper
41	100		
39	54.3	46.0	69.0
67	66.4	57.4	86.2
52	30.3	25.5	38.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#42
1,2-Dichloroethane
Concen: 19.535 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 62 Resp: 69061

Ion	Ratio	Lower	Upper
62	100		
98	8.5	0.0	17.0






#43
Isopropyl Acetate
Concen: 20.207 ug/l
RT: 8.682 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 43 Resp: 102099


Ion	Ratio	Lower	Upper
43	100		
61	25.5	20.7	31.1
87	11.7	9.8	14.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#44
Trichloroethene
Concen: 18.285 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:130 Resp: 39952

Ion	Ratio	Lower	Upper
130	100		
95	100.3	0.0	195.8






#45
1,2-Dichloropropane
Concen: 19.016 ug/l
RT: 9.618 min Scan# 13
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 63 Resp: 45593


Ion	Ratio	Lower	Upper
63	100		
65	32.4	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#46
Dibromomethane
Concen: 19.460 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 93 Resp: 33666

Ion	Ratio	Lower	Upper
93	100		
95	79.8	64.7	97.1
174	86.6	69.0	103.4






#47
Bromodichloromethane
Concen: 20.265 ug/l
RT: 9.888 min Scan# 13
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 83 Resp: 71425


Ion	Ratio	Lower	Upper
83	100		
85	60.1	51.2	76.8
127	8.2	6.5	9.7



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#48
Methyl methacrylate
Concen: 19.959 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 41 Resp: 45383

Ion	Ratio	Lower	Upper
41	100		
69	78.7	64.7	97.1
39	61.0	49.0	73.6






#49
1,4-Dioxane
Concen: 388.248 ug/l
RT: 9.694 min Scan# 13
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 88 Resp: 14866


Ion	Ratio	Lower	Upper
88	100		
43	0.0	26.6	39.8#
58	73.9	59.5	89.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#50
Toluene-d8
Concen: 53.598 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 98 Resp: 423886

Ion	Ratio	Lower	Upper
98	100		
100	63.7	52.2	78.4






#51
4-Methyl-2-Pentanone
Concen: 106.528 ug/l
RT: 10.441 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 43 Resp: 312333


Ion	Ratio	Lower	Upper
43	100		
58	37.7	30.5	45.7



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#52
Toluene
Concen: 19.660 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 92 Resp: 106928

Ion	Ratio	Lower	Upper
92	100		
91	179.2	139.2	208.8






#53
t-1,3-Dichloropropene
Concen: 19.777 ug/l
RT: 10.835 min Scan# 15
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 75 Resp: 65874


Ion	Ratio	Lower	Upper
75	100		
77	31.6	23.5	35.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#54
cis-1,3-Dichloropropene
Concen: 19.360 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 75 Resp: 68879

Ion	Ratio	Lower	Upper
75	100		
77	33.4	25.0	37.4
39	56.2	43.1	64.7






#55
1,1,2-Trichloroethane
Concen: 20.474 ug/l
RT: 11.012 min Scan# 15
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 97 Resp: 44068


Ion	Ratio	Lower	Upper
97	100		
83	87.9	69.0	103.6
85	54.8	44.6	66.8
99	64.8	48.6	72.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#56
Ethyl methacrylate
Concen: 17.580 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 69 Resp: 59018

Ion	Ratio	Lower	Upper
69	100		
41	69.1	54.6	82.0
39	43.7	32.4	48.6






#57
1,3-Dichloropropane
Concen: 20.157 ug/l
RT: 11.159 min Scan# 15
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 76 Resp: 75438


Ion	Ratio	Lower	Upper
76	100		
78	32.2	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#58
2-Chloroethyl Vinyl ether
Concen: 75.697 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 63 Resp: 103394

Ion	Ratio	Lower	Upper
63	100		
106	28.0	21.6	32.4






#59
2-Hexanone
Concen: 108.244 ug/l
RT: 11.194 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 43 Resp: 223307


Ion	Ratio	Lower	Upper
43	100		
58	50.5	26.2	78.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#60
Dibromochloromethane
Concen: 20.010 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:129 Resp: 51998

Ion	Ratio	Lower	Upper
129	100		
127	79.0	38.6	115.8






#61
1,2-Dibromoethane
Concen: 19.501 ug/l
RT: 11.471 min Scan# 16
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:107 Resp: 41782


Ion	Ratio	Lower	Upper
107	100		
109	93.5	75.9	113.9



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#62
4-Bromofluorobenzene
Concen: 50.788 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 95 Resp: 137397

Ion	Ratio	Lower	Upper
95	100		
174	74.6	0.0	145.0
176	71.6	0.0	142.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 16
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:117 Resp: 286465


Ion	Ratio	Lower	Upper
117	100		
82	59.5	48.6	72.8
119	32.7	26.6	39.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#64
Tetrachloroethene
Concen: 20.154 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:164 Resp: 39360

Ion	Ratio	Lower	Upper
164	100		
166	124.0	103.4	155.2
129	98.8	79.2	118.8
131	91.9	77.1	115.7






#65
Chlorobenzene
Concen: 18.877 ug/l
RT: 11.888 min Scan# 16
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:112 Resp: 118461


Ion	Ratio	Lower	Upper
112	100		
114	33.5	25.3	37.9



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 19.712 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:131 Resp: 45402

Ion	Ratio	Lower	Upper
131	100		
133	95.1	47.4	142.3
119	65.0	33.1	99.5






#67
Ethyl Benzene
Concen: 18.588 ug/l
RT: 11.959 min Scan# 17
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 91 Resp: 189996


Ion	Ratio	Lower	Upper
91	100		
106	30.0	23.8	35.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#68
m/p-Xylenes
Concen: 39.852 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:106 Resp: 150545

Ion	Ratio	Lower	Upper
106	100		
91	205.0	167.7	251.5






#69
o-Xylene
Concen: 18.885 ug/l
RT: 12.394 min Scan# 17
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:106 Resp: 68176


Ion	Ratio	Lower	Upper
106	100		
91	219.8	110.4	331.2



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#70
Styrene
Concen: 18.944 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:104 Resp: 113181

Ion	Ratio	Lower	Upper
104	100		
78	54.9	42.5	63.7
103	57.4	43.8	65.8





#71
Bromoform
Concen: 20.438 ug/l
RT: 12.576 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06


Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01


Tgt Ion:173 Resp: 33682

Ion	Ratio	Lower	Upper
173	100		
175	49.1	24.4	73.2
254	0.0	0.0	0.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025







#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:152 Resp: 135464

Ion	Ratio	Lower	Upper
152	100		
115	61.4	31.1	93.3
150	165.4	0.0	343.6






#73
Isopropylbenzene
Concen: 19.299 ug/l
RT: 12.694 min Scan# 18
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:105 Resp: 176447


Ion	Ratio	Lower	Upper
105	100		
120	26.1	12.8	38.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#74
N-amyl acetate
Concen: 18.824 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 43 Resp: 77354

Ion	Ratio	Lower	Upper
43	100		
70	41.8	34.0	51.0
55	28.4	21.4	32.2
61	22.8	19.0	28.4






#75
1,1,2,2-Tetrachloroethane
Concen: 19.969 ug/l
RT: 12.935 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 83 Resp: 64490


Ion	Ratio	Lower	Upper
83	100		
131	9.5	4.8	14.4
85	63.5	32.2	96.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#76
1,2,3-Trichloropropane
Concen: 18.971 ug/l m
RT: 12.988 min Scan# 1876
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 75 Resp: 52234

Ion	Ratio	Lower	Upper
75	100		
77	201.2	109.7	329.2






#77
Bromobenzene
Concen: 18.976 ug/l
RT: 12.976 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:156 Resp: 45326


Ion	Ratio	Lower	Upper
156	100		
77	231.9	114.1	342.4
158	98.3	48.9	146.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#78
n-propylbenzene
Concen: 19.579 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 91 Resp: 211911

Ion	Ratio	Lower	Upper
91	100		
120	21.6	10.9	32.6






#79
2-Chlorotoluene
Concen: 19.320 ug/l
RT: 13.123 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 91 Resp: 135337


Ion	Ratio	Lower	Upper
91	100		
126	31.0	15.7	47.1



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#80
1,3,5-Trimethylbenzene
Concen: 19.996 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:105 Resp: 151009

Ion	Ratio	Lower	Upper
105	100		
120	46.4	23.9	71.7






#81
trans-1,4-Dichloro-2-butene
Concen: 19.361 ug/l
RT: 12.735 min Scan# 18
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 75 Resp: 19703


Ion	Ratio	Lower	Upper
75	100		
53	118.8	95.6	143.4
89	45.1	37.0	55.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#82
4-Chlorotoluene
Concen: 19.280 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 91 Resp: 134471

Ion	Ratio	Lower	Upper
91	100		
126	32.3	15.9	47.7






#83
tert-Butylbenzene
Concen: 18.748 ug/l
RT: 13.435 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:119 Resp: 118897


Ion	Ratio	Lower	Upper
119	100		
91	70.9	35.1	105.3
134	25.0	12.0	36.1



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#84
1,2,4-Trimethylbenzene
Concen: 19.938 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:105 Resp: 150071

Ion	Ratio	Lower	Upper
105	100		
120	43.1	21.6	65.0






#85
sec-Butylbenzene
Concen: 19.315 ug/l
RT: 13.612 min Scan# 19
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:105 Resp: 169773


Ion	Ratio	Lower	Upper
105	100		
134	19.6	9.7	28.9



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#86
p-Isopropyltoluene
Concen: 18.592 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:119 Resp: 135685

Ion	Ratio	Lower	Upper
119	100		
134	25.3	12.7	38.0
91	27.1	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 18.896 ug/l
RT: 13.729 min Scan# 20
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:146 Resp: 83123


Ion	Ratio	Lower	Upper
146	100		
111	43.9	21.4	64.3
148	64.7	32.3	96.9



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#88
1,4-Dichlorobenzene
Concen: 18.785 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:146 Resp: 85651

Ion	Ratio	Lower	Upper
146	100		
111	41.0	21.3	63.7
148	62.2	32.4	97.0






#89
n-Butylbenzene
Concen: 17.761 ug/l
RT: 14.053 min Scan# 20
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 91 Resp: 111994


Ion	Ratio	Lower	Upper
91	100		
92	51.5	25.8	77.3
134	23.1	11.7	35.1



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#90
Hexachloroethane
Concen: 18.652 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:117 Resp: 31211

Ion	Ratio	Lower	Upper
117	100		
201	66.4	33.7	101.0






#91
1,2-Dichlorobenzene
Concen: 18.130 ug/l
RT: 14.106 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:146 Resp: 79551


Ion	Ratio	Lower	Upper
146	100		
111	45.1	21.7	65.1
148	64.8	31.4	94.2



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 18.991 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion: 75 Resp: 11200

Ion	Ratio	Lower	Upper
75	100		
155	70.5	36.4	109.2
157	91.0	45.4	136.1






#93
1,2,4-Trichlorobenzene
Concen: 16.935 ug/l
RT: 15.388 min Scan# 22
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:180 Resp: 34537


Ion	Ratio	Lower	Upper
180	100		
182	96.1	46.8	140.3
145	33.5	16.0	48.0



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#94
Hexachlorobutadiene
Concen: 18.291 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:225 Resp: 19804

Ion	Ratio	Lower	Upper
225	100		
223	63.0	30.7	92.1
227	61.8	30.9	92.5






#95
Naphthalene
Concen: 15.801 ug/l
RT: 15.641 min Scan# 23
Delta R.T. 0.006 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:128 Resp: 96159


Ion	Ratio	Lower	Upper
128	100		
127	13.9	10.6	16.0
129	10.9	8.8	13.2



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025



#96
1,2,3-Trichlorobenzene
Concen: 17.448 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN085593.D
Acq: 31 Jan 2025 11:06

Tgt Ion:180 Resp: 36004

Ion	Ratio	Lower	Upper
180	100		
182	92.0	47.4	142.2
145	34.7	16.9	50.7



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	VX0203WBS01	SDG No.:	Q1215
Lab Sample ID:	VX0203WBS01	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044805.D	1		02/03/25 11:07	VX020325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	20.2		0.34	5.00	ug/L
75-35-4	1,1-Dichloroethene	16.8		0.26	5.00	ug/L
78-93-3	2-Butanone	110		1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	17.6		0.25	5.00	ug/L
67-66-3	Chloroform	19.3		0.26	5.00	ug/L
71-43-2	Benzene	20.7		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	18.0		0.24	5.00	ug/L
79-01-6	Trichloroethene	19.4		0.32	5.00	ug/L
127-18-4	Tetrachloroethene	18.9		0.25	5.00	ug/L
108-90-7	Chlorobenzene	19.9		0.13	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	43.1		74 - 125	86%	SPK: 50
1868-53-7	Dibromofluoromethane	46.5		75 - 124	93%	SPK: 50
2037-26-5	Toluene-d8	50.2		86 - 113	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.5		77 - 121	95%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	152000	5.544			
540-36-3	1,4-Difluorobenzene	276000	6.757			
3114-55-4	Chlorobenzene-d5	242000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	108000	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044805.D
Acq On : 03 Feb 2025 11:07
Operator : JC/MD
Sample : VX0203WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Feb 04 02:05:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.544	168	151779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	276067	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	242310	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	107943	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	98701	43.082	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 86.160%			
35) Dibromofluoromethane	5.379	113	81631	46.545	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.080%			
50) Toluene-d8	8.647	98	307152	50.228	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.460%			
62) 4-Bromofluorobenzene	11.079	95	108190	47.458	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 94.920%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	37035	18.133	ug/l	98
3) Chloromethane	1.307	50	45886	20.776	ug/l	100
4) Vinyl Chloride	1.374	62	43980	20.153	ug/l	98
5) Bromomethane	1.593	94	12016	23.060	ug/l	95
6) Chloroethane	1.666	64	17283	43.188	ug/l	100
7) Trichlorofluoromethane	1.874	101	55799	23.135	ug/l	99
8) Diethyl Ether	2.130	74	16397	16.846	ug/l	89
9) 1,1,2-Trichlorotrifluo...	2.319	101	33617	18.542	ug/l	97
10) Methyl Iodide	2.447	142	43114	17.417	ug/l	92
11) Tert butyl alcohol	2.971	59	42368	82.955	ug/l #	82
12) 1,1-Dichloroethene	2.313	96	29904	16.827	ug/l	92
13) Acrolein	2.233	56	13741	57.172	ug/l	92
14) Allyl chloride	2.654	41	59191	17.106	ug/l	96
15) Acrylonitrile	3.063	53	87810	87.657	ug/l	99
16) Acetone	2.380	43	79139	87.310	ug/l	94
17) Carbon Disulfide	2.502	76	78503	16.188	ug/l	100
18) Methyl Acetate	2.703	43	51979	18.588	ug/l	96
19) Methyl tert-butyl Ether	3.111	73	115173	17.892	ug/l	100
20) Methylene Chloride	2.782	84	35521	16.828	ug/l	98
21) trans-1,2-Dichloroethene	3.087	96	31659	17.063	ug/l	97
22) Diisopropyl ether	3.758	45	113346	18.844	ug/l	95
23) Vinyl Acetate	3.715	43	475314	90.285	ug/l	99
24) 1,1-Dichloroethane	3.605	63	65529	18.372	ug/l	98
25) 2-Butanone	4.556	43	142936	106.062	ug/l	99
26) 2,2-Dichloropropane	4.465	77	60119	17.391	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	44993	19.278	ug/l	98
28) Bromochloromethane	4.891	49	35017	21.009	ug/l	93
29) Tetrahydrofuran	5.001	42	94460	108.664	ug/l	96
30) Chloroform	5.087	83	69627	19.343	ug/l	98
31) Cyclohexane	5.465	56	63642	22.061	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	58095	17.556	ug/l	100
36) 1,1-Dichloropropene	5.690	75	47190	19.329	ug/l	99
37) Ethyl Acetate	4.715	43	54836	21.092	ug/l	98
38) Carbon Tetrachloride	5.666	117	48299	17.588	ug/l	98
39) Methylcyclohexane	7.373	83	66006	20.367	ug/l	100
40) Benzene	6.032	78	156314	20.662	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044805.D
Acq On : 03 Feb 2025 11:07
Operator : JC/MD
Sample : VX0203WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Feb 04 02:05:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
41) Methacrylonitrile	4.916	41	30477	20.117	ug/l		99
42) 1,2-Dichloroethane	6.080	62	49626	17.973	ug/l		100
43) Isopropyl Acetate	6.336	43	92685	19.379	ug/l		99
44) Trichloroethene	7.123	130	35209	19.359	ug/l		98
45) 1,2-Dichloropropane	7.428	63	40268	21.146	ug/l		96
46) Dibromomethane	7.580	93	27392	18.896	ug/l		98
47) Bromodichloromethane	7.818	83	53846	18.112	ug/l		94
48) Methyl methacrylate	7.690	41	44083	19.857	ug/l		99
49) 1,4-Dioxane	7.659	88	19327	417.265	ug/l		99
51) 4-Methyl-2-Pentanone	8.574	43	290951	110.081	ug/l		98
52) Toluene	8.714	92	95820	20.694	ug/l		98
53) t-1,3-Dichloropropene	8.976	75	56705	18.462	ug/l		94
54) cis-1,3-Dichloropropene	8.366	75	62067	18.636	ug/l		97
55) 1,1,2-Trichloroethane	9.147	97	38045	20.308	ug/l		97
56) Ethyl methacrylate	9.116	69	64208	19.933	ug/l		99
57) 1,3-Dichloropropane	9.305	76	64638	20.262	ug/l		100
58) 2-Chloroethyl Vinyl ether	8.238	63	139078	87.947	ug/l		97
59) 2-Hexanone	9.427	43	213385	109.737	ug/l		99
60) Dibromochloromethane	9.519	129	40586	18.458	ug/l		99
61) 1,2-Dibromoethane	9.604	107	37936	19.323	ug/l		100
64) Tetrachloroethene	9.269	164	28697	18.860	ug/l		97
65) Chlorobenzene	10.073	112	102598	19.934	ug/l		99
66) 1,1,1,2-Tetrachloroethane	10.159	131	34132	18.248	ug/l		99
67) Ethyl Benzene	10.189	91	180008	20.045	ug/l		100
68) m/p-Xylenes	10.299	106	135490	40.657	ug/l		96
69) o-Xylene	10.640	106	69022	20.480	ug/l		96
70) Styrene	10.653	104	113346	20.831	ug/l		97
71) Bromoform	10.799	173	26269	17.255	ug/l	#	97
73) Isopropylbenzene	10.957	105	172841	19.601	ug/l		99
74) N-amyl acetate	10.842	43	80964	18.628	ug/l		99
75) 1,1,2,2-Tetrachloroethane	11.213	83	59880	19.661	ug/l		100
76) 1,2,3-Trichloropropane	11.238	75	47698m	18.208	ug/l		
77) Bromobenzene	11.195	156	39304	19.069	ug/l		96
78) n-propylbenzene	11.299	91	198314	20.392	ug/l		99
79) 2-Chlorotoluene	11.360	91	119970	19.097	ug/l		98
80) 1,3,5-Trimethylbenzene	11.451	105	141828	19.757	ug/l		99
81) trans-1,4-Dichloro-2-b...	11.018	75	19878	17.353	ug/l		98
82) 4-Chlorotoluene	11.451	91	133356	19.379	ug/l		96
83) tert-Butylbenzene	11.713	119	145335	19.362	ug/l		96
84) 1,2,4-Trimethylbenzene	11.750	105	140872	19.311	ug/l		98
85) sec-Butylbenzene	11.890	105	176864	20.237	ug/l		99
86) p-Isopropyltoluene	12.006	119	143197	19.937	ug/l		99
87) 1,3-Dichlorobenzene	11.969	146	70582	19.746	ug/l		99
88) 1,4-Dichlorobenzene	12.037	146	70484	19.506	ug/l		100
89) n-Butylbenzene	12.329	91	125028	20.268	ug/l		99
90) Hexachloroethane	12.536	117	28043	17.242	ug/l		97
91) 1,2-Dichlorobenzene	12.335	146	72245	19.681	ug/l		98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	11160	14.983	ug/l		96
93) 1,2,4-Trichlorobenzene	13.585	180	41539	18.929	ug/l		98
94) Hexachlorobutadiene	13.719	225	15697	18.210	ug/l		99
95) Naphthalene	13.774	128	162454	19.194	ug/l		99
96) 1,2,3-Trichlorobenzene	13.957	180	41958	18.446	ug/l		99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044805.D
Acq On : 03 Feb 2025 11:07
Operator : JC/MD
Sample : VX0203WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Feb 04 02:05:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020325\
Data File : VX044805.D
Acq On : 03 Feb 2025 11:07
Operator : JC/MD
Sample : VX0203WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Quant Time: Feb 04 02:05:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 7
Delta R.T. -0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:168 Resp: 15177


Ion	Ratio	Lower	Upper
168	100		
99	60.9	51.1	76.7



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#2
Dichlorodifluoromethane
Concen: 18.133 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 85 Resp: 37035

Ion	Ratio	Lower	Upper
85	100		
87	32.5	16.8	50.4






#3
Chloromethane
Concen: 20.776 ug/l
RT: 1.307 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 50 Resp: 4588


Ion	Ratio	Lower	Upper
50	100		
52	34.0	27.0	40.4



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#4
Vinyl Chloride
Concen: 20.153 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 62 Resp: 43980

Ion	Ratio	Lower	Upper
62	100		
64	31.9	26.6	39.8






#5
Bromomethane
Concen: 23.060 ug/l
RT: 1.593 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 94 Resp: 1201


Ion	Ratio	Lower	Upper
94	100		
96	86.2	72.8	109.2



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#6
Chloroethane
Concen: 43.188 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 64 Resp: 17283

Ion	Ratio	Lower	Upper
64	100		
66	32.9	26.6	39.8






#7
Trichlorofluoromethane
Concen: 23.135 ug/l
RT: 1.874 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:101 Resp: 5579


Ion	Ratio	Lower	Upper
101	100		
103	64.9	52.5	78.7



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#8
Diethyl Ether
Concen: 16.846 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 74 Resp: 16397

Ion	Ratio	Lower	Upper
74	100		
45	105.2	47.1	141.3





#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.542 ug/l
RT: 2.319 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07


Tgt Ion:101 Resp: 3361


Ion	Ratio	Lower	Upper
101	100		
85	44.6	35.3	52.9
151	69.8	58.2	87.4

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#10
Methyl Iodide
Concen: 17.417 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:142 Resp: 43114

Ion	Ratio	Lower	Upper
142	100		
127	50.6	35.0	52.4
141	14.7	11.7	17.5





#11
Tert butyl alcohol
Concen: 82.955 ug/l
RT: 2.971 min Scan# 3
Delta R.T. -0.012 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07


Tgt Ion: 59 Resp: 4236


Ion	Ratio	Lower	Upper
59	100		
57	10.8	3.9	5.9

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#12
1,1-Dichloroethene
Concen: 16.827 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 96 Resp: 29904

Ion	Ratio	Lower	Upper
96	100		
61	175.6	131.1	196.7
98	68.5	51.2	76.8






#13
Acrolein
Concen: 57.172 ug/l
RT: 2.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 56 Resp: 1374


Ion	Ratio	Lower	Upper
56	100		
55	75.5	55.1	82.7



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#14
Allyl chloride
Concen: 17.106 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 41 Resp: 59191

Ion	Ratio	Lower	Upper
41	100		
39	71.7	54.9	82.3
76	32.1	28.4	42.6






#15
Acrylonitrile
Concen: 87.657 ug/l
RT: 3.063 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 53 Resp: 8781


Ion	Ratio	Lower	Upper
53	100		
52	82.6	66.0	99.0
51	37.8	28.9	43.3



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#16
Acetone
Concen: 87.310 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 43 Resp: 79139

Ion	Ratio	Lower	Upper
43	100		
58	27.8	24.6	37.0






#17
Carbon Disulfide
Concen: 16.188 ug/l
RT: 2.502 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 76 Resp: 7850


Ion	Ratio	Lower	Upper
76	100		
78	8.8	7.1	10.7



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#18
Methyl Acetate
Concen: 18.588 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 43 Resp: 51979

Ion	Ratio	Lower	Upper
43	100		
74	22.6	19.8	29.8






#19
Methyl tert-butyl Ether
Concen: 17.892 ug/l
RT: 3.111 min Scan# 3
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 73 Resp: 11517


Ion	Ratio	Lower	Upper
73	100		
57	21.6	17.4	26.0



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#20
Methylene Chloride
Concen: 16.828 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 84 Resp: 35521

Ion	Ratio	Lower	Upper
84	100		
49	118.2	94.2	141.4
51	37.2	28.8	43.2
86	60.0	50.3	75.5






#21
trans-1,2-Dichloroethene
Concen: 17.063 ug/l
RT: 3.087 min Scan# 3
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 96 Resp: 3165


Ion	Ratio	Lower	Upper
96	100		
61	144.0	111.0	166.6
98	62.6	50.2	75.2



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#22
Diisopropyl ether
Concen: 18.844 ug/l
RT: 3.758 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 45 Resp: 113346

Ion	Ratio	Lower	Upper
45	100		
43	82.3	61.0	91.6
87	27.9	22.7	34.1
59	10.7	9.4	14.2





#23
Vinyl Acetate
Concen: 90.285 ug/l
RT: 3.715 min Scan# 4
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07


Tgt Ion: 43 Resp: 47531


Ion	Ratio	Lower	Upper
43	100		
86	10.7	8.9	13.3

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#24
1,1-Dichloroethane
Concen: 18.372 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 63 Resp: 65529

Ion	Ratio	Lower	Upper
63	100		
98	6.6	3.6	10.9
100	4.4	2.5	7.5






#25
2-Butanone
Concen: 106.062 ug/l
RT: 4.556 min Scan# 5
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 43 Resp: 14293


Ion	Ratio	Lower	Upper
43	100		
72	26.5	21.7	32.5



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#26
2,2-Dichloropropane
Concen: 17.391 ug/l
RT: 4.465 min Scan# 554
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 77 Resp: 60119

Ion	Ratio	Lower	Upper
77	100		
97	24.0	10.9	32.6






#27
cis-1,2-Dichloroethene
Concen: 19.278 ug/l
RT: 4.489 min Scan# 5
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 96 Resp: 4499


Ion	Ratio	Lower	Upper
96	100		
61	146.2	0.0	296.0
98	63.4	0.0	130.0



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#28
Bromochloromethane
Concen: 21.009 ug/l
RT: 4.891 min Scan# 624
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 49 Resp: 35017

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	4.4
130	73.0	63.4	95.2





#29
Tetrahydrofuran
Concen: 108.664 ug/l
RT: 5.001 min Scan# 64
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07


Tgt Ion: 42 Resp: 9446


Ion	Ratio	Lower	Upper
42	100		
72	45.8	38.2	57.4
71	40.7	35.3	52.9

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#30
Chloroform
Concen: 19.343 ug/l
RT: 5.087 min Scan# 656
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 83 Resp: 69627

Ion	Ratio	Lower	Upper
83	100		
85	62.2	51.2	76.8






#31
Cyclohexane
Concen: 22.061 ug/l
RT: 5.465 min Scan# 7
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 56 Resp: 6364


Ion	Ratio	Lower	Upper
56	100		
69	29.9	26.8	40.2
84	85.0	71.0	106.4



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#32
1,1,1-Trichloroethane
Concen: 17.556 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 97 Resp: 58095

Ion	Ratio	Lower	Upper
97	100		
99	64.5	51.6	77.4
61	44.8	35.6	53.4






#33
1,2-Dichloroethane-d4
Concen: 43.082 ug/l
RT: 5.946 min Scan# 7
Delta R.T. -0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 65 Resp: 9870


Ion	Ratio	Lower	Upper
65	100		
67	53.7	0.0	103.4



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:114 Resp: 276067

Ion	Ratio	Lower	Upper
114	100		
63	20.2	0.0	43.0
88	16.4	0.0	35.4






#35
Dibromofluoromethane
Concen: 46.545 ug/l
RT: 5.379 min Scan# 7
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:113 Resp: 8163


Ion	Ratio	Lower	Upper
113	100		
111	104.4	81.6	122.4
192	18.0	14.4	21.6



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#36
1,1-Dichloropropene
Concen: 19.329 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 75 Resp: 47190

Ion	Ratio	Lower	Upper
75	100		
110	35.4	17.9	53.8
77	31.5	24.5	36.7






#37
Ethyl Acetate
Concen: 21.092 ug/l
RT: 4.715 min Scan# 5
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 43 Resp: 5483


Ion	Ratio	Lower	Upper
43	100		
61	14.4	10.9	16.3
70	11.8	10.1	15.1



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#38
Carbon Tetrachloride
Concen: 17.588 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:117 Resp: 48299

Ion	Ratio	Lower	Upper
117	100		
119	92.3	76.0	114.0
121	30.5	24.6	36.8





#39
Methylcyclohexane
Concen: 20.367 ug/l
RT: 7.373 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07


Tgt Ion: 83 Resp: 6600


Ion	Ratio	Lower	Upper
83	100		
55	75.6	60.5	90.7
98	48.7	39.0	58.4

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#40
Benzene
Concen: 20.662 ug/l
RT: 6.032 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 78 Resp: 156314

Ion	Ratio	Lower	Upper
78	100		
77	25.4	18.7	28.1






#41
Methacrylonitrile
Concen: 20.117 ug/l
RT: 4.916 min Scan# 6
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 41 Resp: 3047


Ion	Ratio	Lower	Upper
41	100		
39	54.6	44.4	66.6
67	77.1	60.6	90.8
52	31.2	24.3	36.5



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#42
1,2-Dichloroethane
Concen: 17.973 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 62 Resp: 49626

Ion	Ratio	Lower	Upper
62	100		
98	9.7	0.0	19.4






#43
Isopropyl Acetate
Concen: 19.379 ug/l
RT: 6.336 min Scan# 8
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 43 Resp: 9268


Ion	Ratio	Lower	Upper
43	100		
61	21.3	16.6	24.8
87	13.2	11.0	16.6



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#44
Trichloroethene
Concen: 19.359 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:130 Resp: 35209

Ion	Ratio	Lower	Upper
130	100		
95	106.3	0.0	208.2






#45
1,2-Dichloropropane
Concen: 21.146 ug/l
RT: 7.428 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 63 Resp: 4026


Ion	Ratio	Lower	Upper
63	100		
65	32.1	24.2	36.2



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#46
Dibromomethane
Concen: 18.896 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 93 Resp: 27392

Ion	Ratio	Lower	Upper
93	100		
95	81.1	66.4	99.6
174	88.6	72.0	108.0






#47
Bromodichloromethane
Concen: 18.112 ug/l
RT: 7.818 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 83 Resp: 5384


Ion	Ratio	Lower	Upper
83	100		
85	67.8	50.1	75.1
127	8.5	6.5	9.7



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#48
Methyl methacrylate
Concen: 19.857 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 41 Resp: 44083

Ion	Ratio	Lower	Upper
41	100		
69	84.4	66.8	100.2
39	58.0	47.4	71.0





#49
1,4-Dioxane
Concen: 417.265 ug/l
RT: 7.659 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07


Tgt Ion: 88 Resp: 1932


Ion	Ratio	Lower	Upper
88	100		
43	32.5	27.0	40.4
58	69.9	55.8	83.8

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#50
Toluene-d8
Concen: 50.228 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 98 Resp: 307152

Ion	Ratio	Lower	Upper
98	100		
100	67.3	53.2	79.8






#51
4-Methyl-2-Pentanone
Concen: 110.081 ug/l
RT: 8.574 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 43 Resp: 29095


Ion	Ratio	Lower	Upper
43	100		
58	38.8	31.9	47.9



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#52
Toluene
Concen: 20.694 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 92 Resp: 95820

Ion	Ratio	Lower	Upper
92	100		
91	172.6	135.7	203.5






#53
t-1,3-Dichloropropene
Concen: 18.462 ug/l
RT: 8.976 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 75 Resp: 5670


Ion	Ratio	Lower	Upper
75	100		
77	33.8	24.4	36.6



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#54
cis-1,3-Dichloropropene
Concen: 18.636 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 75 Resp: 62067

Ion	Ratio	Lower	Upper
75	100		
77	32.5	25.1	37.7
39	47.5	40.3	60.5






#55
1,1,2-Trichloroethane
Concen: 20.308 ug/l
RT: 9.147 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 97 Resp: 3804


Ion	Ratio	Lower	Upper
97	100		
83	89.1	68.8	103.2
85	58.4	43.7	65.5
99	63.4	49.5	74.3



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#56
Ethyl methacrylate
Concen: 19.933 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 69 Resp: 64208

Ion	Ratio	Lower	Upper
69	100		
41	66.4	53.2	79.8
39	36.1	30.2	45.2






#57
1,3-Dichloropropane
Concen: 20.262 ug/l
RT: 9.305 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 76 Resp: 6463


Ion	Ratio	Lower	Upper
76	100		
78	32.4	26.0	39.0



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#58
2-Chloroethyl Vinyl ether
Concen: 87.947 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 63 Resp: 139078

Ion	Ratio	Lower	Upper
63	100		
106	27.0	23.0	34.6






#59
2-Hexanone
Concen: 109.737 ug/l
RT: 9.427 min Scan# 1
Delta R.T. -0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 43 Resp: 21338


Ion	Ratio	Lower	Upper
43	100		
58	54.4	27.8	83.3



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#60
Dibromochloromethane
Concen: 18.458 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:129 Resp: 40586

Ion	Ratio	Lower	Upper
129	100		
127	77.5	38.5	115.5






#61
1,2-Dibromoethane
Concen: 19.323 ug/l
RT: 9.604 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:107 Resp: 3793


Ion	Ratio	Lower	Upper
107	100		
109	94.4	75.3	112.9



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#62
4-Bromofluorobenzene
Concen: 47.458 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 95 Resp: 108190

Ion	Ratio	Lower	Upper
95	100		
174	72.2	0.0	133.0
176	70.0	0.0	129.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:117 Resp: 24231


Ion	Ratio	Lower	Upper
117	100		
82	59.7	47.6	71.4
119	31.5	25.4	38.2



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#64
Tetrachloroethene
Concen: 18.860 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:164 Resp: 28697

Ion	Ratio	Lower	Upper
164	100		
166	125.9	102.0	153.0
129	97.4	81.1	121.7
131	95.4	79.0	118.6






#65
Chlorobenzene
Concen: 19.934 ug/l
RT: 10.073 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:112 Resp: 10259


Ion	Ratio	Lower	Upper
112	100		
114	31.2	25.3	37.9



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 18.248 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:131 Resp: 34132

Ion	Ratio	Lower	Upper
131	100		
133	94.3	47.8	143.3
119	68.1	33.4	100.2






#67
Ethyl Benzene
Concen: 20.045 ug/l
RT: 10.189 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 91 Resp: 18000


Ion	Ratio	Lower	Upper
91	100		
106	31.5	25.0	37.6



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#68
m/p-Xylenes
Concen: 40.657 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:106 Resp: 135490

Ion	Ratio	Lower	Upper
106	100		
91	202.4	166.4	249.6






#69
o-Xylene
Concen: 20.480 ug/l
RT: 10.640 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:106 Resp: 6902


Ion	Ratio	Lower	Upper
106	100		
91	211.9	108.7	326.3



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#70
Styrene
Concen: 20.831 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:104 Resp: 113346

Ion	Ratio	Lower	Upper
104	100		
78	49.2	43.0	64.6
103	54.4	43.4	65.2






#71
Bromoform
Concen: 17.255 ug/l
RT: 10.799 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:173 Resp: 2626


Ion	Ratio	Lower	Upper
173	100		
175	49.4	23.6	70.8
254	0.0	0.1	0.1



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:152 Resp: 107943

Ion	Ratio	Lower	Upper
152	100		
115	71.0	43.8	131.4
150	163.8	0.0	347.8






#73
Isopropylbenzene
Concen: 19.601 ug/l
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:105 Resp: 17284


Ion	Ratio	Lower	Upper
105	100		
120	26.1	13.2	39.6



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#74
N-amyl acetate
Concen: 18.628 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 43 Resp: 80964

Ion	Ratio	Lower	Upper
43	100		
70	43.6	35.2	52.8
55	23.2	19.0	28.4
61	24.0	20.0	30.0






#75
1,1,2,2-Tetrachloroethane
Concen: 19.661 ug/l
RT: 11.213 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 83 Resp: 5988


Ion	Ratio	Lower	Upper
83	100		
131	9.4	5.0	14.9
85	64.0	32.0	96.2



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#76
1,2,3-Trichloropropane
Concen: 18.208 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 75 Resp: 47698

Ion	Ratio	Lower	Upper
75	100		
77	44.2	23.4	70.2






#77
Bromobenzene
Concen: 19.069 ug/l
RT: 11.195 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:156 Resp: 3930


Ion	Ratio	Lower	Upper
156	100		
77	171.0	90.3	271.0
158	97.0	48.4	145.3



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#78
n-propylbenzene
Concen: 20.392 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 91 Resp: 198314

Ion	Ratio	Lower	Upper
91	100		
120	22.2	11.4	34.1






#79
2-Chlorotoluene
Concen: 19.097 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 91 Resp: 11997


Ion	Ratio	Lower	Upper
91	100		
126	33.2	16.2	48.5



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#80
1,3,5-Trimethylbenzene
Concen: 19.757 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:105 Resp: 141828

Ion	Ratio	Lower	Upper
105	100		
120	47.8	23.7	71.1





#81
trans-1,4-Dichloro-2-butene
Concen: 17.353 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07


Tgt Ion: 75 Resp: 1987


Ion	Ratio	Lower	Upper
75	100		
53	93.6	74.3	111.5
89	44.2	37.4	56.0

Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025







#82
4-Chlorotoluene
Concen: 19.379 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 91 Resp: 133356

Ion	Ratio	Lower	Upper
91	100		
126	30.0	13.9	41.7






#83
tert-Butylbenzene
Concen: 19.362 ug/l
RT: 11.713 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:119 Resp: 14533


Ion	Ratio	Lower	Upper
119	100		
91	56.7	30.0	90.0
134	23.3	11.2	33.5



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#84
1,2,4-Trimethylbenzene
Concen: 19.311 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:105 Resp: 140872

Ion	Ratio	Lower	Upper
105	100		
120	45.4	22.2	66.6






#85
sec-Butylbenzene
Concen: 20.237 ug/l
RT: 11.890 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:105 Resp: 17686


Ion	Ratio	Lower	Upper
105	100		
134	19.1	9.8	29.5



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#86
p-Isopropyltoluene
Concen: 19.937 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:119 Resp: 143197

Ion	Ratio	Lower	Upper
119	100		
134	26.1	12.8	38.4
91	24.6	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 19.746 ug/l
RT: 11.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:146 Resp: 7058


Ion	Ratio	Lower	Upper
146	100		
111	42.1	21.9	65.7
148	64.0	31.8	95.4



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#88
1,4-Dichlorobenzene
Concen: 19.506 ug/l
RT: 12.037 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:146 Resp: 70484

Ion	Ratio	Lower	Upper
146	100		
111	42.8	21.5	64.5
148	63.9	32.1	96.5






#89
n-Butylbenzene
Concen: 20.268 ug/l
RT: 12.329 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 91 Resp: 12502


Ion	Ratio	Lower	Upper
91	100		
92	54.7	27.4	82.0
134	24.2	12.8	38.3



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#90
Hexachloroethane
Concen: 17.242 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:117 Resp: 28043

Ion	Ratio	Lower	Upper
117	100		
201	61.9	32.3	96.8






#91
1,2-Dichlorobenzene
Concen: 19.681 ug/l
RT: 12.335 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:146 Resp: 7224


Ion	Ratio	Lower	Upper
146	100		
111	43.9	22.4	67.2
148	61.8	31.6	94.8



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 14.983 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion: 75 Resp: 11160

Ion	Ratio	Lower	Upper
75	100		
155	80.5	39.0	116.9
157	104.2	49.9	149.7






#93
1,2,4-Trichlorobenzene
Concen: 18.929 ug/l
RT: 13.585 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:180 Resp: 4153


Ion	Ratio	Lower	Upper
180	100		
182	93.0	47.4	142.2
145	31.9	16.7	50.1



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025




#94
Hexachlorobutadiene
Concen: 18.210 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:225 Resp: 15697

Ion	Ratio	Lower	Upper
225	100		
223	62.9	31.9	95.7
227	64.7	31.8	95.4






#95
Naphthalene
Concen: 19.194 ug/l
RT: 13.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:128 Resp: 16245


Ion	Ratio	Lower	Upper
128	100		
127	12.5	10.3	15.5
129	10.6	8.7	13.1



Instrument :
MSVOA_X
ClientSampleId :
VX0203WBS01

Manual Integrations
APPROVED


Reviewed By :John Carlone 02/04/2025
Supervised By :Mahesh Dadoda 02/04/2025



#96
1,2,3-Trichlorobenzene
Concen: 18.446 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX044805.D
Acq: 03 Feb 2025 11:07

Tgt Ion:180 Resp: 41958

Ion	Ratio	Lower	Upper
180	100		
182	94.9	47.2	141.6
145	34.0	17.5	52.6



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	RU2 Engineering, LLC	Date Collected:	
Project:	NYCDDC SANTWOBR Brooklyn Bridge BBMCR	Date Received:	
Client Sample ID:	VN0131WBSD01	SDG No.:	Q1215
Lab Sample ID:	VN0131WBSD01	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085594.D	1		01/31/25 11:39	VN013125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	17.2		0.34	5.00	ug/L
75-35-4	1,1-Dichloroethene	18.1		0.26	5.00	ug/L
78-93-3	2-Butanone	110		1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	19.3		0.25	5.00	ug/L
67-66-3	Chloroform	20.8		0.26	5.00	ug/L
71-43-2	Benzene	19.4		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	21.3		0.24	5.00	ug/L
79-01-6	Trichloroethene	18.9		0.32	5.00	ug/L
127-18-4	Tetrachloroethene	20.6		0.25	5.00	ug/L
108-90-7	Chlorobenzene	19.9		0.13	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.2		74 - 125	106%	SPK: 50
1868-53-7	Dibromofluoromethane	53.2		75 - 124	106%	SPK: 50
2037-26-5	Toluene-d8	52.1		86 - 113	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.2		77 - 121	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	171000	8.224			
540-36-3	1,4-Difluorobenzene	292000	9.1			
3114-55-4	Chlorobenzene-d5	260000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	125000	13.788			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085594.D
Acq On : 31 Jan 2025 11:39
Operator : JC\MD
Sample : VN0131WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jan 31 22:25:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	170842	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	291614	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	260387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	125242	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	146607	53.162	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.320%			
35) Dibromofluoromethane	8.165	113	107584	53.178	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.360%			
50) Toluene-d8	10.565	98	374675	52.125	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.260%			
62) 4-Bromofluorobenzene	12.847	95	130847	53.215	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 106.440%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.124	85	38099	16.471	ug/l	100
3) Chloromethane	2.359	50	41483	16.563	ug/l	96
4) Vinyl Chloride	2.512	62	43379	17.232	ug/l	97
5) Bromomethane	2.959	94	27138	17.847	ug/l	92
6) Chloroethane	3.130	64	29949	18.765	ug/l	94
7) Trichlorofluoromethane	3.501	101	67983	18.612	ug/l	99
8) Diethyl Ether	3.965	74	23399	18.543	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.365	101	40081	19.482	ug/l	100
10) Methyl Iodide	4.589	142	46082	19.559	ug/l	98
11) Tert butyl alcohol	5.518	59	36797	116.527	ug/l	100
12) 1,1-Dichloroethene	4.348	96	33138	18.073	ug/l	90
13) Acrolein	4.177	56	32299	74.928	ug/l	100
14) Allyl chloride	5.024	41	52879	17.773	ug/l	98
15) Acrylonitrile	5.712	53	112978	112.638	ug/l	100
16) Acetone	4.430	43	98420	110.454	ug/l	99
17) Carbon Disulfide	4.706	76	83044	14.710	ug/l	96
18) Methyl Acetate	5.018	43	66636	24.593	ug/l	100
19) Methyl tert-butyl Ether	5.794	73	127597	21.432	ug/l	100
20) Methylene Chloride	5.271	84	43643	19.785	ug/l	97
21) trans-1,2-Dichloroethene	5.783	96	35923	18.333	ug/l	96
22) Diisopropyl ether	6.671	45	141569	21.438	ug/l	99
23) Vinyl Acetate	6.600	43	494432	106.985	ug/l	99
24) 1,1-Dichloroethane	6.565	63	83257	20.676	ug/l	100
25) 2-Butanone	7.477	43	146709	111.884	ug/l	97
26) 2,2-Dichloropropane	7.489	77	69887	21.467	ug/l	100
27) cis-1,2-Dichloroethene	7.489	96	46577	20.181	ug/l	98
28) Bromochloromethane	7.812	49	40972	21.871	ug/l	99
29) Tetrahydrofuran	7.836	42	94383	113.531	ug/l	99
30) Chloroform	7.965	83	86579	20.804	ug/l	97
31) Cyclohexane	8.253	56	58088	16.597	ug/l	97
32) 1,1,1-Trichloroethane	8.171	97	72352	19.820	ug/l	94
36) 1,1-Dichloropropene	8.371	75	51402	18.103	ug/l	99
37) Ethyl Acetate	7.559	43	62903	21.949	ug/l	97
38) Carbon Tetrachloride	8.359	117	62612	19.262	ug/l	98
39) Methylcyclohexane	9.600	83	44653	16.736	ug/l	99
40) Benzene	8.600	78	165782	19.432	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085594.D
Acq On : 31 Jan 2025 11:39
Operator : JC\MD
Sample : VN0131WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jan 31 22:25:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
41) Methacrylonitrile	7.777	41	32075	21.512 ug/l	96
42) 1,2-Dichloroethane	8.671	62	68576	21.342 ug/l	98
43) Isopropyl Acetate	8.688	43	101965	22.204 ug/l	99
44) Trichloroethene	9.347	130	37559	18.913 ug/l	90
45) 1,2-Dichloropropane	9.618	63	44141	20.256 ug/l	100
46) Dibromomethane	9.712	93	32921	20.937 ug/l	99
47) Bromodichloromethane	9.882	83	69434	21.675 ug/l #	98
48) Methyl methacrylate	9.677	41	44798	21.677 ug/l	99
49) 1,4-Dioxane	9.694	88	14929	428.980 ug/l #	92
51) 4-Methyl-2-Pentanone	10.441	43	312143	117.136 ug/l	99
52) Toluene	10.629	92	102996	20.836 ug/l	98
53) t-1,3-Dichloropropene	10.835	75	63224	20.885 ug/l	97
54) cis-1,3-Dichloropropene	10.312	75	69645	21.537 ug/l	97
55) 1,1,2-Trichloroethane	11.012	97	43870	22.425 ug/l	97
56) Ethyl methacrylate	10.871	69	60017	19.440 ug/l	96
57) 1,3-Dichloropropane	11.159	76	73415	21.583 ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	110235	88.796 ug/l	100
59) 2-Hexanone	11.194	43	225189	120.099 ug/l	99
60) Dibromochloromethane	11.359	129	51298	21.720 ug/l	98
61) 1,2-Dibromoethane	11.465	107	41427	21.274 ug/l	98
64) Tetrachloroethene	11.100	164	36529	20.578 ug/l	99
65) Chlorobenzene	11.888	112	113226	19.850 ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	42230	20.171 ug/l	99
67) Ethyl Benzene	11.965	91	179492	19.320 ug/l	99
68) m/p-Xylenes	12.071	106	139208	40.542 ug/l	99
69) o-Xylene	12.394	106	64375	19.618 ug/l	98
70) Styrene	12.406	104	111627	20.555 ug/l	99
71) Bromoform	12.576	173	33846	22.594 ug/l #	99
73) Isopropylbenzene	12.694	105	166281	19.672 ug/l	100
74) N-amyl acetate	12.488	43	78287	20.605 ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	63591	21.297 ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	52049m	20.447 ug/l	
77) Bromobenzene	12.976	156	44454	20.129 ug/l	98
78) n-propylbenzene	13.035	91	202101	20.197 ug/l	99
79) 2-Chlorotoluene	13.123	91	128406	19.827 ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	140053	20.059 ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	22836	24.271 ug/l	86
82) 4-Chlorotoluene	13.218	91	129888	20.143 ug/l	99
83) tert-Butylbenzene	13.435	119	112298	19.153 ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	141674	20.359 ug/l	99
85) sec-Butylbenzene	13.618	105	162740	20.026 ug/l	99
86) p-Isopropyltoluene	13.729	119	130871	19.356 ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	80731	19.850 ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	79386	18.832 ug/l	98
89) n-Butylbenzene	14.053	91	105667	18.125 ug/l	99
90) Hexachloroethane	14.329	117	30543	19.743 ug/l	96
91) 1,2-Dichlorobenzene	14.106	146	78528	19.357 ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12261	22.487 ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	34814	18.464 ug/l	96
94) Hexachlorobutadiene	15.500	225	18346	18.327 ug/l	97
95) Naphthalene	15.641	128	101833	18.099 ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	36018	18.879 ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085594.D
Acq On : 31 Jan 2025 11:39
Operator : JC\MD
Sample : VN0131WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jan 31 22:25:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085594.D
Acq On : 31 Jan 2025 11:39
Operator : JC\MD
Sample : VN0131WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1


Quant Time: Jan 31 22:25:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:168 Resp: 170842


Ion	Ratio	Lower	Upper
168	100		
99	64.6	53.6	80.4



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#2
Dichlorodifluoromethane
Concen: 16.471 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 85 Resp: 38099

Ion	Ratio	Lower	Upper
85	100		
87	33.3	16.7	50.1






#3
Chloromethane
Concen: 16.563 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 50 Resp: 41483


Ion	Ratio	Lower	Upper
50	100		
52	31.5	27.0	40.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#4
Vinyl Chloride
Concen: 17.232 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 62 Resp: 43379

Ion	Ratio	Lower	Upper
62	100		
64	32.9	24.8	37.2






#5
Bromomethane
Concen: 17.847 ug/l
RT: 2.959 min Scan# 17
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 94 Resp: 27138


Ion	Ratio	Lower	Upper
94	100		
96	97.0	71.8	107.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#6
Chloroethane
Concen: 18.765 ug/l
RT: 3.130 min Scan# 200
Delta R.T. 0.012 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 64 Resp: 29949

Ion	Ratio	Lower	Upper
64	100		
66	27.6	24.6	36.8






#7
Trichlorofluoromethane
Concen: 18.612 ug/l
RT: 3.501 min Scan# 26
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:101 Resp: 67983


Ion	Ratio	Lower	Upper
101	100		
103	65.4	51.4	77.2



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#8
Diethyl Ether
Concen: 18.543 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.012 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 74 Resp: 23399

Ion	Ratio	Lower	Upper
74	100		
45	104.0	49.7	149.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.482 ug/l
RT: 4.365 min Scan# 41
Delta R.T. -0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39


Tgt Ion:101 Resp: 40081


Ion	Ratio	Lower	Upper
101	100		
85	47.0	37.8	56.8
151	73.2	58.8	88.2

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025







#10
Methyl Iodide
Concen: 19.559 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:142 Resp: 46082

Ion	Ratio	Lower	Upper
142	100		
127	52.3	40.3	60.5
141	15.3	12.8	19.2






#11
Tert butyl alcohol
Concen: 116.527 ug/l
RT: 5.518 min Scan# 60
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 59 Resp: 36797


Ion	Ratio	Lower	Upper
59	100		
57	10.8	8.6	13.0



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#12
1,1-Dichloroethene
Concen: 18.073 ug/l
RT: 4.348 min Scan# 407
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 96 Resp: 33138

Ion	Ratio	Lower	Upper
96	100		
61	160.5	141.0	211.4
98	65.6	49.0	73.6






#13
Acrolein
Concen: 74.928 ug/l
RT: 4.177 min Scan# 37
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 56 Resp: 32299


Ion	Ratio	Lower	Upper
56	100		
55	70.6	56.3	84.5



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#14
Allyl chloride
Concen: 17.773 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 41 Resp: 52879

Ion	Ratio	Lower	Upper
41	100		
39	78.7	64.4	96.6
76	37.1	30.5	45.7






#15
Acrylonitrile
Concen: 112.638 ug/l
RT: 5.712 min Scan# 63
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 53 Resp: 112978


Ion	Ratio	Lower	Upper
53	100		
52	81.6	65.5	98.3
51	37.0	29.8	44.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#16
Acetone
Concen: 110.454 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.012 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 43 Resp: 98420

Ion	Ratio	Lower	Upper
43	100		
58	28.9	23.8	35.6






#17
Carbon Disulfide
Concen: 14.710 ug/l
RT: 4.706 min Scan# 46
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 76 Resp: 83044


Ion	Ratio	Lower	Upper
76	100		
78	10.0	6.9	10.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#18
Methyl Acetate
Concen: 24.593 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 43 Resp: 66636

Ion	Ratio	Lower	Upper
43	100		
74	23.3	18.6	28.0






#19
Methyl tert-butyl Ether
Concen: 21.432 ug/l
RT: 5.794 min Scan# 65
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 73 Resp: 127597


Ion	Ratio	Lower	Upper
73	100		
57	22.4	17.8	26.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#20
Methylene Chloride
Concen: 19.785 ug/l
RT: 5.271 min Scan# 564
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 84 Resp: 43643

Ion	Ratio	Lower	Upper
84	100		
49	129.6	104.1	156.1
51	39.2	31.0	46.4
86	58.8	51.9	77.9






#21
trans-1,2-Dichloroethene
Concen: 18.333 ug/l
RT: 5.783 min Scan# 65
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 96 Resp: 35923


Ion	Ratio	Lower	Upper
96	100		
61	139.3	115.7	173.5
98	59.9	49.8	74.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#22
Diisopropyl ether
Concen: 21.438 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 45 Resp: 141569

Ion	Ratio	Lower	Upper
45	100		
43	55.5	45.0	67.6
87	27.2	21.8	32.8
59	13.8	10.2	15.2






#23
Vinyl Acetate
Concen: 106.985 ug/l
RT: 6.600 min Scan# 79
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 43 Resp: 494432


Ion	Ratio	Lower	Upper
43	100		
86	10.2	8.4	12.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#24
1,1-Dichloroethane
Concen: 20.676 ug/l
RT: 6.565 min Scan# 784
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 63 Resp: 83257

Ion	Ratio	Lower	Upper
63	100		
98	6.5	3.3	9.8
100	3.7	2.0	6.0






#25
2-Butanone
Concen: 111.884 ug/l
RT: 7.477 min Scan# 93
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 43 Resp: 146709


Ion	Ratio	Lower	Upper
43	100		
72	23.9	20.2	30.4



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#26
2,2-Dichloropropane
Concen: 21.467 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 77 Resp: 69887

Ion	Ratio	Lower	Upper
77	100		
97	21.4	10.7	32.1






#27
cis-1,2-Dichloroethene
Concen: 20.181 ug/l
RT: 7.489 min Scan# 94
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 96 Resp: 46577


Ion	Ratio	Lower	Upper
96	100		
61	153.5	0.0	311.8
98	63.8	0.0	126.0



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#28
Bromochloromethane
Concen: 21.871 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 49 Resp: 40972

Ion	Ratio	Lower	Upper
49	100		
129	1.7	0.0	4.4
130	69.8	55.0	82.4






#29
Tetrahydrofuran
Concen: 113.531 ug/l
RT: 7.836 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 42 Resp: 94383


Ion	Ratio	Lower	Upper
42	100		
72	43.8	34.2	51.2
71	40.9	32.5	48.7



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#30
Chloroform
Concen: 20.804 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 83 Resp: 86579

Ion	Ratio	Lower	Upper
83	100		
85	62.4	51.8	77.6






#31
Cyclohexane
Concen: 16.597 ug/l
RT: 8.253 min Scan# 10
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 56 Resp: 58088


Ion	Ratio	Lower	Upper
56	100		
69	28.8	22.2	33.4
84	85.9	66.4	99.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#32
1,1,1-Trichloroethane
Concen: 19.820 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 97 Resp: 72352

Ion	Ratio	Lower	Upper
97	100		
99	56.3	49.8	74.6
61	49.2	41.4	62.2






#33
1,2-Dichloroethane-d4
Concen: 53.162 ug/l
RT: 8.577 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 65 Resp: 146607


Ion	Ratio	Lower	Upper
65	100		
67	51.2	0.0	101.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:114 Resp: 291614

Ion	Ratio	Lower	Upper
114	100		
63	22.9	0.0	47.6
88	15.3	0.0	32.6






#35
Dibromofluoromethane
Concen: 53.178 ug/l
RT: 8.165 min Scan# 10
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:113 Resp: 107584


Ion	Ratio	Lower	Upper
113	100		
111	102.1	82.7	124.1
192	17.8	14.3	21.5



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#36
1,1-Dichloropropene
Concen: 18.103 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 75 Resp: 51402

Ion	Ratio	Lower	Upper
75	100		
110	32.1	16.5	49.5
77	31.2	24.4	36.6






#37
Ethyl Acetate
Concen: 21.949 ug/l
RT: 7.559 min Scan# 95
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 43 Resp: 62903


Ion	Ratio	Lower	Upper
43	100		
61	11.6	10.9	16.3
70	9.8	7.5	11.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#38
Carbon Tetrachloride
Concen: 19.262 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:117 Resp: 62612

Ion	Ratio	Lower	Upper
117	100		
119	96.0	75.4	113.2
121	32.4	24.6	37.0





#39
Methylcyclohexane
Concen: 16.736 ug/l
RT: 9.600 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39


Tgt Ion: 83 Resp: 44653


Ion	Ratio	Lower	Upper
83	100		
55	78.4	62.6	94.0
98	48.9	37.7	56.5

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025







#40
Benzene
Concen: 19.432 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 78 Resp: 165782

Ion	Ratio	Lower	Upper
78	100		
77	24.7	19.0	28.6






#41
Methacrylonitrile
Concen: 21.512 ug/l
RT: 7.777 min Scan# 99
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 41 Resp: 32075


Ion	Ratio	Lower	Upper
41	100		
39	61.1	46.0	69.0
67	74.3	57.4	86.2
52	34.2	25.5	38.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#42
1,2-Dichloroethane
Concen: 21.342 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 62 Resp: 68576

Ion	Ratio	Lower	Upper
62	100		
98	7.8	0.0	17.0






#43
Isopropyl Acetate
Concen: 22.204 ug/l
RT: 8.688 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 43 Resp: 101965


Ion	Ratio	Lower	Upper
43	100		
61	25.5	20.7	31.1
87	11.9	9.8	14.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#44
Trichloroethene
Concen: 18.913 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:130 Resp: 37559

Ion	Ratio	Lower	Upper
130	100		
95	108.3	0.0	195.8






#45
1,2-Dichloropropane
Concen: 20.256 ug/l
RT: 9.618 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 63 Resp: 44141


Ion	Ratio	Lower	Upper
63	100		
65	32.1	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#46
Dibromomethane
Concen: 20.937 ug/l
RT: 9.712 min Scan# 1319
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 93 Resp: 32921

Ion	Ratio	Lower	Upper
93	100		
95	80.0	64.7	97.1
174	88.0	69.0	103.4






#47
Bromodichloromethane
Concen: 21.675 ug/l
RT: 9.882 min Scan# 13
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 83 Resp: 69434


Ion	Ratio	Lower	Upper
83	100		
85	62.9	51.2	76.8
127	10.3	6.5	9.7#



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#48
Methyl methacrylate
Concen: 21.677 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 41 Resp: 44798

Ion	Ratio	Lower	Upper
41	100		
69	81.4	64.7	97.1
39	62.3	49.0	73.6





#49
1,4-Dioxane
Concen: 428.980 ug/l
RT: 9.694 min Scan# 13
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39


Tgt Ion: 88 Resp: 14929


Ion	Ratio	Lower	Upper
88	100		
43	41.2	26.6	39.8#
58	79.4	59.5	89.3

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025







#50
Toluene-d8
Concen: 52.125 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 98 Resp: 374675

Ion	Ratio	Lower	Upper
98	100		
100	64.3	52.2	78.4





#51
4-Methyl-2-Pentanone
Concen: 117.136 ug/l
RT: 10.441 min Scan# 14
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39


Tgt Ion: 43 Resp: 312143


Ion	Ratio	Lower	Upper
43	100		
58	37.7	30.5	45.7

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025







#52
Toluene
Concen: 20.836 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.005 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 92 Resp: 102996

Ion	Ratio	Lower	Upper
92	100		
91	171.7	139.2	208.8






#53
t-1,3-Dichloropropene
Concen: 20.885 ug/l
RT: 10.835 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 75 Resp: 63224


Ion	Ratio	Lower	Upper
75	100		
77	31.0	23.5	35.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#54
cis-1,3-Dichloropropene
Concen: 21.537 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 75 Resp: 69645

Ion	Ratio	Lower	Upper
75	100		
77	27.8	25.0	37.4
39	54.8	43.1	64.7






#55
1,1,2-Trichloroethane
Concen: 22.425 ug/l
RT: 11.012 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 97 Resp: 43870


Ion	Ratio	Lower	Upper
97	100		
83	82.6	69.0	103.6
85	56.8	44.6	66.8
99	58.4	48.6	72.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#56
Ethyl methacrylate
Concen: 19.440 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 69 Resp: 60017

Ion	Ratio	Lower	Upper
69	100		
41	70.8	54.6	82.0
39	44.1	32.4	48.6






#57
1,3-Dichloropropane
Concen: 21.583 ug/l
RT: 11.159 min Scan# 15
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 76 Resp: 73415


Ion	Ratio	Lower	Upper
76	100		
78	32.9	25.6	38.4



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#58
2-Chloroethyl Vinyl ether
Concen: 88.796 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 63 Resp: 110235

Ion	Ratio	Lower	Upper
63	100		
106	26.9	21.6	32.4






#59
2-Hexanone
Concen: 120.099 ug/l
RT: 11.194 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 43 Resp: 225189


Ion	Ratio	Lower	Upper
43	100		
58	51.7	26.2	78.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#60
Dibromochloromethane
Concen: 21.720 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:129 Resp: 51298

Ion	Ratio	Lower	Upper
129	100		
127	75.9	38.6	115.8






#61
1,2-Dibromoethane
Concen: 21.274 ug/l
RT: 11.465 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:107 Resp: 41427


Ion	Ratio	Lower	Upper
107	100		
109	93.2	75.9	113.9



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#62
4-Bromofluorobenzene
Concen: 53.215 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 95 Resp: 130847

Ion	Ratio	Lower	Upper
95	100		
174	74.2	0.0	145.0
176	70.6	0.0	142.4






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:117 Resp: 260387


Ion	Ratio	Lower	Upper
117	100		
82	58.4	48.6	72.8
119	32.8	26.6	39.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#64
Tetrachloroethene
Concen: 20.578 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:164 Resp: 36529

Ion	Ratio	Lower	Upper
164	100		
166	127.0	103.4	155.2
129	100.0	79.2	118.8
131	95.9	77.1	115.7






#65
Chlorobenzene
Concen: 19.850 ug/l
RT: 11.888 min Scan# 16
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:112 Resp: 113226


Ion	Ratio	Lower	Upper
112	100		
114	31.8	25.3	37.9



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 20.171 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:131 Resp: 42230

Ion	Ratio	Lower	Upper
131	100		
133	97.0	47.4	142.3
119	66.6	33.1	99.5






#67
Ethyl Benzene
Concen: 19.320 ug/l
RT: 11.965 min Scan# 17
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 91 Resp: 179492


Ion	Ratio	Lower	Upper
91	100		
106	30.5	23.8	35.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#68
m/p-Xylenes
Concen: 40.542 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:106 Resp: 139208

Ion	Ratio	Lower	Upper
106	100		
91	207.4	167.7	251.5






#69
o-Xylene
Concen: 19.618 ug/l
RT: 12.394 min Scan# 17
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:106 Resp: 64375


Ion	Ratio	Lower	Upper
106	100		
91	224.4	110.4	331.2



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#70
Styrene
Concen: 20.555 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:104 Resp: 111627

Ion	Ratio	Lower	Upper
104	100		
78	51.3	42.5	63.7
103	54.9	43.8	65.8






#71
Bromoform
Concen: 22.594 ug/l
RT: 12.576 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:173 Resp: 33846


Ion	Ratio	Lower	Upper
173	100		
175	49.7	24.4	73.2
254	0.0	0.0	0.0



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:152 Resp: 125242

Ion	Ratio	Lower	Upper
152	100		
115	61.7	31.1	93.3
150	163.6	0.0	343.6






#73
Isopropylbenzene
Concen: 19.672 ug/l
RT: 12.694 min Scan# 18
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:105 Resp: 166281


Ion	Ratio	Lower	Upper
105	100		
120	25.3	12.8	38.3



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#74
N-amyl acetate
Concen: 20.605 ug/l
RT: 12.488 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 43 Resp: 78287

Ion	Ratio	Lower	Upper
43	100		
70	42.5	34.0	51.0
55	27.2	21.4	32.2
61	22.4	19.0	28.4






#75
1,1,2,2-Tetrachloroethane
Concen: 21.297 ug/l
RT: 12.935 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 83 Resp: 63591


Ion	Ratio	Lower	Upper
83	100		
131	9.5	4.8	14.4
85	64.4	32.2	96.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#76
1,2,3-Trichloropropane
Concen: 20.447 ug/l m
RT: 12.988 min Scan# 1876
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 75 Resp: 52049

Ion	Ratio	Lower	Upper
75	100		
77	198.1	109.7	329.2






#77
Bromobenzene
Concen: 20.129 ug/l
RT: 12.976 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:156 Resp: 44454


Ion	Ratio	Lower	Upper
156	100		
77	232.0	114.1	342.4
158	96.8	48.9	146.8



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#78
n-propylbenzene
Concen: 20.197 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 91 Resp: 202101

Ion	Ratio	Lower	Upper
91	100		
120	22.0	10.9	32.6






#79
2-Chlorotoluene
Concen: 19.827 ug/l
RT: 13.123 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 91 Resp: 128406


Ion	Ratio	Lower	Upper
91	100		
126	31.8	15.7	47.1



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#80
1,3,5-Trimethylbenzene
Concen: 20.059 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:105 Resp: 140053

Ion	Ratio	Lower	Upper
105	100		
120	48.4	23.9	71.7






#81
trans-1,4-Dichloro-2-butene
Concen: 24.271 ug/l
RT: 12.735 min Scan# 18
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 75 Resp: 22836


Ion	Ratio	Lower	Upper
75	100		
53	102.2	95.6	143.4
89	39.5	37.0	55.6



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#82
4-Chlorotoluene
Concen: 20.143 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 91 Resp: 129888

Ion	Ratio	Lower	Upper
91	100		
126	32.4	15.9	47.7






#83
tert-Butylbenzene
Concen: 19.153 ug/l
RT: 13.435 min Scan# 19
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:119 Resp: 112298


Ion	Ratio	Lower	Upper
119	100		
91	71.5	35.1	105.3
134	24.0	12.0	36.1



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#84
1,2,4-Trimethylbenzene
Concen: 20.359 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:105 Resp: 141674

Ion	Ratio	Lower	Upper
105	100		
120	44.1	21.6	65.0






#85
sec-Butylbenzene
Concen: 20.026 ug/l
RT: 13.618 min Scan# 19
Delta R.T. 0.005 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:105 Resp: 162740


Ion	Ratio	Lower	Upper
105	100		
134	18.9	9.7	28.9



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#86
p-Isopropyltoluene
Concen: 19.356 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:119 Resp: 130871

Ion	Ratio	Lower	Upper
119	100		
134	24.8	12.7	38.0
91	25.3	12.7	38.1






#87
1,3-Dichlorobenzene
Concen: 19.850 ug/l
RT: 13.729 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:146 Resp: 80731


Ion	Ratio	Lower	Upper
146	100		
111	42.7	21.4	64.3
148	63.6	32.3	96.9



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#88
1,4-Dichlorobenzene
Concen: 18.832 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:146 Resp: 79386

Ion	Ratio	Lower	Upper
146	100		
111	44.1	21.3	63.7
148	66.3	32.4	97.0






#89
n-Butylbenzene
Concen: 18.125 ug/l
RT: 14.053 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 91 Resp: 105667


Ion	Ratio	Lower	Upper
91	100		
92	51.0	25.8	77.3
134	23.4	11.7	35.1



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#90
Hexachloroethane
Concen: 19.743 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:117 Resp: 30543

Ion	Ratio	Lower	Upper
117	100		
201	64.0	33.7	101.0






#91
1,2-Dichlorobenzene
Concen: 19.357 ug/l
RT: 14.106 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:146 Resp: 78528


Ion	Ratio	Lower	Upper
146	100		
111	42.9	21.7	65.1
148	63.3	31.4	94.2



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 22.487 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion: 75 Resp: 12261

Ion	Ratio	Lower	Upper
75	100		
155	66.6	36.4	109.2
157	91.7	45.4	136.1






#93
1,2,4-Trichlorobenzene
Concen: 18.464 ug/l
RT: 15.394 min Scan# 22
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:180 Resp: 34814


Ion	Ratio	Lower	Upper
180	100		
182	97.6	46.8	140.3
145	33.8	16.0	48.0



Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025




#94
Hexachlorobutadiene
Concen: 18.327 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:225 Resp: 18346

Ion	Ratio	Lower	Upper
225	100		
223	62.5	30.7	92.1
227	64.8	30.9	92.5





#95
Naphthalene
Concen: 18.099 ug/l
RT: 15.641 min Scan# 23
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39


Tgt Ion:128 Resp: 101833


Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.6	16.0
129	10.6	8.8	13.2

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBSD01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/03/2025
Supervised By :Semsettin Yesilyurt 02/03/2025






#96
1,2,3-Trichlorobenzene
Concen: 18.879 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN085594.D
Acq: 31 Jan 2025 11:39

Tgt Ion:180 Resp: 36018

Ion	Ratio	Lower	Upper
180	100		
182	93.2	47.4	142.2
145	34.4	16.9	50.7



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VN011425	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC100	VN085438.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:09 AM	MMDadoda	1/15/2025 12:55:19 PM	Peak Integrated by Software
VSTDICCC050	VN085439.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:14 AM	MMDadoda	1/15/2025 12:55:22 PM	Peak Integrated by Software
VSTDICC020	VN085440.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:19 AM	MMDadoda	1/15/2025 12:55:25 PM	Peak Integrated by Software
VSTDICC020	VN085440.D	trans-1,4-Dichloro-2-but ene	JOHN	1/15/2025 9:31:19 AM	MMDadoda	1/15/2025 12:55:25 PM	Peak Integrated by Software
VSTDICC020	VN085440.D	Vinyl Acetate	JOHN	1/15/2025 9:31:19 AM	MMDadoda	1/15/2025 12:55:25 PM	Peak Integrated by Software
VSTDICC010	VN085441.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:23 AM	MMDadoda	1/15/2025 12:55:29 PM	Peak Integrated by Software
VSTDICC005	VN085442.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:28 AM	MMDadoda	1/15/2025 12:55:34 PM	Peak Integrated by Software
VSTDICC005	VN085442.D	Ethyl Acetate	JOHN	1/15/2025 9:31:28 AM	MMDadoda	1/15/2025 12:55:34 PM	Peak Integrated by Software
VSTDICC005	VN085442.D	Vinyl Acetate	JOHN	1/15/2025 9:31:28 AM	MMDadoda	1/15/2025 12:55:34 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	1,1,2-Trichlorotrifluoroet hane	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	1,1-Dichloroethane	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	1,4-Dichlorobenzene	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VN011425	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VN085443.D	2-Butanone	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	Ethyl Acetate	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICC001	VN085443.D	Vinyl Acetate	JOHN	1/15/2025 9:31:32 AM	MMDadoda	1/15/2025 12:55:39 PM	Peak Integrated by Software
VSTDICV050	VN085445.D	1,2,3-Trichloropropane	JOHN	1/15/2025 9:31:37 AM	MMDadoda	1/15/2025 12:55:43 PM	Peak Integrated by Software
VSTDICV050	VN085445.D	trans-1,4-Dichloro-2-but ene	JOHN	1/15/2025 9:31:37 AM	MMDadoda	1/15/2025 12:55:43 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VN013125	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN085591.D	1,2,3-Trichloropropane	MMDADODA	2/3/2025 7:21:19 AM	SAM	2/3/2025 7:23:17 AM	Peak Integrated by Software
VN0131WBS01	VN085593.D	1,2,3-Trichloropropane	MMDADODA	2/3/2025 7:21:22 AM	SAM	2/3/2025 7:23:19 AM	Peak Integrated by Software
VN0131WBSD01	VN085594.D	1,2,3-Trichloropropane	MMDADODA	2/3/2025 7:21:23 AM	SAM	2/3/2025 7:23:21 AM	Peak Integrated by Software
VSTDCCC050	VN085615.D	1,2,3-Trichloropropane	MMDADODA	2/3/2025 7:21:24 AM	SAM	2/3/2025 7:23:22 AM	Peak Integrated by Software
VSTDCCC050	VN085615.D	Isopropyl Acetate	MMDADODA	2/3/2025 7:21:24 AM	SAM	2/3/2025 7:23:22 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VX011525	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VX044648.D	1,2,3-Trichloropropane	JOHN	1/15/2025 3:49:52 PM	SAM	1/16/2025 7:37:32 AM	Peak Integrated by Software
VSTDICC001	VX044648.D	1,4-Dichlorobenzene	JOHN	1/15/2025 3:49:52 PM	SAM	1/16/2025 7:37:32 AM	Peak Integrated by Software
VSTDICC001	VX044648.D	Methacrylonitrile	JOHN	1/15/2025 3:49:52 PM	SAM	1/16/2025 7:37:32 AM	Peak Integrated by Software
VSTDICC005	VX044649.D	1,2,3-Trichloropropane	JOHN	1/15/2025 3:49:56 PM	SAM	1/16/2025 7:37:34 AM	Peak Integrated by Software
VSTDICC005	VX044649.D	Tert butyl alcohol	JOHN	1/15/2025 3:49:56 PM	SAM	1/16/2025 7:37:34 AM	Peak Integrated by Software
VSTDICC020	VX044650.D	1,2,3-Trichloropropane	JOHN	1/15/2025 3:50:00 PM	SAM	1/16/2025 7:37:35 AM	Peak Integrated by Software
VSTDICC020	VX044650.D	Tert butyl alcohol	JOHN	1/15/2025 3:50:00 PM	SAM	1/16/2025 7:37:35 AM	Peak Integrated by Software
VSTDICCC050	VX044651.D	1,2,3-Trichloropropane	JOHN	1/15/2025 3:50:04 PM	SAM	1/16/2025 7:37:43 AM	Peak Integrated by Software
VSTDICCC050	VX044651.D	Tert butyl alcohol	JOHN	1/15/2025 3:50:04 PM	SAM	1/16/2025 7:37:43 AM	Peak Integrated by Software
VSTDICC100	VX044652.D	1,2,3-Trichloropropane	JOHN	1/15/2025 3:50:09 PM	SAM	1/16/2025 7:37:45 AM	Peak Integrated by Software
VSTDICC100	VX044652.D	Tert butyl alcohol	JOHN	1/15/2025 3:50:09 PM	SAM	1/16/2025 7:37:45 AM	Peak Integrated by Software
VSTDICC150	VX044653.D	1,2,3-Trichloropropane	JOHN	1/15/2025 3:50:13 PM	SAM	1/16/2025 7:37:47 AM	Peak Integrated by Software
VSTDICC150	VX044653.D	Tert butyl alcohol	JOHN	1/15/2025 3:50:13 PM	SAM	1/16/2025 7:37:47 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VX011525	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICV050	VX044655.D	1,2,3-Trichloropropane	MMDADODA	1/16/2025 7:35:14 AM	SAM	1/16/2025 7:37:49 AM	Peak Integrated by Software
VSTDICV050	VX044655.D	Tert butyl alcohol	MMDADODA	1/16/2025 7:35:14 AM	SAM	1/16/2025 7:37:49 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VX020325	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX044802.D	1,2,3-Trichloropropane	JOHN	2/4/2025 9:48:52 AM	MMDadoda	2/4/2025 3:45:27 PM	Peak Integrated by Software
VSTDCCC050	VX044802.D	Chloroethane	JOHN	2/4/2025 9:48:52 AM	MMDadoda	2/4/2025 3:45:27 PM	Peak Integrated by Software
VSTDCCC050	VX044802.D	Tert butyl alcohol	JOHN	2/4/2025 9:48:52 AM	MMDadoda	2/4/2025 3:45:27 PM	Peak Integrated by Software
VX0203WBS01	VX044805.D	1,2,3-Trichloropropane	JOHN	2/4/2025 9:48:56 AM	MMDadoda	2/4/2025 3:45:29 PM	Peak Integrated by Software
VSTDCCC050	VX044826.D	1,2,3-Trichloropropane	JOHN	2/4/2025 9:49:15 AM	MMDadoda	2/4/2025 3:45:37 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QCBatch ID # VN011425

Review By	John Carlone	Review On	1/15/2025 9:31:51 AM				
Supervise By	Mahesh Dadoda	Supervise On	1/15/2025 12:55:51 PM				
SubDirectory	VN011425	HP Acquire Method	MSVOA_N	HP Processing Method	82N011425W.M		

STD. NAME	STD REF.#
Tune/Reschk	VP132529
Initial Calibration Stds	VP132530,VP132531,VP132532,VP132533,VP132534,VP132535
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP132544
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN085437.D	14 Jan 2025 14:22	JC\MD	Ok
2	VSTDICC100	VN085438.D	14 Jan 2025 14:56	JC\MD	Ok,M
3	VSTDICCC050	VN085439.D	14 Jan 2025 15:19	JC\MD	Ok,M
4	VSTDICC020	VN085440.D	14 Jan 2025 15:43	JC\MD	Ok,M
5	VSTDICC010	VN085441.D	14 Jan 2025 16:07	JC\MD	Ok,M
6	VSTDICC005	VN085442.D	14 Jan 2025 16:31	JC\MD	Ok,M
7	VSTDICC001	VN085443.D	14 Jan 2025 17:19	JC\MD	Ok,M
8	IBLK	VN085444.D	14 Jan 2025 17:42	JC\MD	Ok
9	VSTDICV050	VN085445.D	14 Jan 2025 18:06	JC\MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QCBatch ID # VN013125

Review By	Mahesh Dadoda	Review On	2/3/2025 7:21:29 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/3/2025 7:23:27 AM		
SubDirectory	VN013125	HP Acquire Method		HP Processing Method	82N011425W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132796,VP132802
Initial Calibration Stds	
CCC	VP132797,VP132898,VP132803,VP132804
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN085590.D	31 Jan 2025 09:30	JC\MD	Ok
2	VSTDCCC050	VN085591.D	31 Jan 2025 10:04	JC\MD	Ok,M
3	VN0131WBL01	VN085592.D	31 Jan 2025 10:42	JC\MD	Ok
4	VN0131WBS01	VN085593.D	31 Jan 2025 11:06	JC\MD	Ok,M
5	VN0131WBSD01	VN085594.D	31 Jan 2025 11:39	JC\MD	Ok,M
6	Q1211-02RE	VN085595.D	31 Jan 2025 12:03	JC\MD	Ok
7	IBLK	VN085596.D	31 Jan 2025 12:51	JC\MD	Ok
8	PB166357ZHE#02	VN085597.D	31 Jan 2025 13:15	JC\MD	Ok
9	PB166357ZHE#03	VN085598.D	31 Jan 2025 13:39	JC\MD	Ok
10	PB166357ZHE#04	VN085599.D	31 Jan 2025 14:03	JC\MD	Ok
11	PB166357ZHE#05	VN085600.D	31 Jan 2025 14:27	JC\MD	Ok
12	PB166357ZHE#06	VN085601.D	31 Jan 2025 14:51	JC\MD	Ok
13	PB166357ZHE#07	VN085602.D	31 Jan 2025 15:15	JC\MD	Ok
14	PB166357ZHE#08	VN085603.D	31 Jan 2025 15:39	JC\MD	Ok
15	PB166357ZHE#09	VN085604.D	31 Jan 2025 16:04	JC\MD	Ok
16	PB166357ZHE#10	VN085605.D	31 Jan 2025 16:28	JC\MD	Ok
17	PB166357ZHE#11	VN085606.D	31 Jan 2025 16:52	JC\MD	Ok
18	PB166357ZHE#12	VN085607.D	31 Jan 2025 17:16	JC\MD	Ok
19	PB166357ZHE#13	VN085608.D	31 Jan 2025 17:40	JC\MD	Ok
20	PB166357ZHE#14	VN085609.D	31 Jan 2025 18:04	JC\MD	Ok
21	PB166357ZHE#15	VN085610.D	31 Jan 2025 18:28	JC\MD	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QCBatch ID # VN013125

Review By	Mahesh Dadoda	Review On	2/3/2025 7:21:29 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/3/2025 7:23:27 AM		
SubDirectory	VN013125	HP Acquire Method		HP Processing Method	82N011425W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132796,VP132802
Initial Calibration Stds	
CCC	VP132797,VP132898,VP132803,VP132804
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	PB166357ZHE#16	VN085611.D	31 Jan 2025 18:53	JC\MD	Ok
23	PB166357ZHE#17	VN085612.D	31 Jan 2025 19:17	JC\MD	Ok
24	PB166357ZHE#01	VN085613.D	31 Jan 2025 19:41	JC\MD	Ok
25	Q1215-02	VN085614.D	31 Jan 2025 20:05	JC\MD	Ok
26	VSTDCCC050	VN085615.D	31 Jan 2025 20:29	JC\MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX011525

Review By	John Carlone	Review On	1/15/2025 4:32:35 PM		
Supervise By	Mahesh Dadoda	Supervise On	1/16/2025 7:35:39 AM		
SubDirectory	VX011525	HP Acquire Method		HP Processing Method	82X011525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132553
Initial Calibration Stds	VP132558,VP132559,VP132560,VP132561,VP132562,VP132563
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP132564
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX044647.D	15 Jan 2025 09:30	JC/MD	Ok
2	VSTDICC001	VX044648.D	15 Jan 2025 10:31	JC/MD	Ok,M
3	VSTDICC005	VX044649.D	15 Jan 2025 11:17	JC/MD	Ok,M
4	VSTDICC020	VX044650.D	15 Jan 2025 11:40	JC/MD	Ok,M
5	VSTDICCC050	VX044651.D	15 Jan 2025 12:02	JC/MD	Ok,M
6	VSTDICC100	VX044652.D	15 Jan 2025 13:14	JC/MD	Ok,M
7	VSTDICC150	VX044653.D	15 Jan 2025 13:37	JC/MD	Ok,M
8	IBLK	VX044654.D	15 Jan 2025 14:00	JC/MD	Ok
9	VSTDICV050	VX044655.D	15 Jan 2025 15:15	JC/MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX020325

Review By	John Carlone	Review On	2/4/2025 9:51:15 AM		
Supervise By	Mahesh Dadoda	Supervise On	2/4/2025 3:45:43 PM		
SubDirectory	VX020325	HP Acquire Method		HP Processing Method	82X011525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132831
Initial Calibration Stds	
CCC	VP132832,VP132833
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX044801.D	03 Feb 2025 08:12	JC/MD	Ok
2	VSTDCCC050	VX044802.D	03 Feb 2025 09:52	JC/MD	Ok,M
3	VX0203MBL01	VX044803.D	03 Feb 2025 10:21	JC/MD	Ok
4	VX0203WBL01	VX044804.D	03 Feb 2025 10:44	JC/MD	Ok
5	VX0203WBS01	VX044805.D	03 Feb 2025 11:07	JC/MD	Ok,M
6	VX0203WBSD01	VX044806.D	03 Feb 2025 11:32	JC/MD	Ok,M
7	PB166357TB	VX044807.D	03 Feb 2025 11:55	JC/MD	Ok
8	Q1215-06	VX044808.D	03 Feb 2025 12:18	JC/MD	Ok
9	Q1216-02	VX044809.D	03 Feb 2025 12:41	JC/MD	Ok
10	Q1216-06	VX044810.D	03 Feb 2025 13:04	JC/MD	Ok
11	Q1216-10	VX044811.D	03 Feb 2025 13:27	JC/MD	Ok
12	Q1221-02	VX044812.D	03 Feb 2025 13:50	JC/MD	Ok,M
13	Q1218-02	VX044813.D	03 Feb 2025 14:36	JC/MD	Ok
14	Q1239-07	VX044814.D	03 Feb 2025 14:59	JC/MD	Ok
15	VX0203MBS01	VX044815.D	03 Feb 2025 15:22	JC/MD	Ok,M
16	Q1232-06	VX044816.D	03 Feb 2025 15:45	JC/MD	Ok
17	Q1239-10	VX044817.D	03 Feb 2025 16:08	JC/MD	Ok
18	Q1232-14	VX044818.D	03 Feb 2025 16:32	JC/MD	Ok
19	Q1232-18	VX044819.D	03 Feb 2025 16:55	JC/MD	Ok
20	Q1235-02	VX044820.D	03 Feb 2025 17:18	JC/MD	Ok
21	Q1235-06	VX044821.D	03 Feb 2025 17:41	JC/MD	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX020325

Review By	John Carlone	Review On	2/4/2025 9:51:15 AM		
Supervise By	Mahesh Dadoda	Supervise On	2/4/2025 3:45:43 PM		
SubDirectory	VX020325	HP Acquire Method		HP Processing Method	82X011525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132831
Initial Calibration Stds	
CCC	VP132832,VP132833
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	Q1239-02	VX044822.D	03 Feb 2025 18:04	JC/MD	Ok
23	Q1239-05	VX044823.D	03 Feb 2025 18:27	JC/MD	Ok
24	Q1239-08	VX044824.D	03 Feb 2025 18:50	JC/MD	Ok
25	Q1239-11	VX044825.D	03 Feb 2025 19:14	JC/MD	Ok
26	VSTDCCC050	VX044826.D	03 Feb 2025 19:37	JC/MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QCBatch ID # VN011425

Review By	John Carlone	Review On	1/15/2025 9:31:51 AM		
Supervise By	Mahesh Dadoda	Supervise On	1/15/2025 12:55:51 PM		
SubDirectory	VN011425	HP Acquire Method	MSVOA_N	HP Processing Method	82N011425W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132529
Initial Calibration Stds	VP132530,VP132531,VP132532,VP132533,VP132534,VP132535
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP132544
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085437.D	14 Jan 2025 14:22		JC\MD	Ok
2	VSTDICC100	VSTDICC100	VN085438.D	14 Jan 2025 14:56		JC\MD	Ok,M
3	VSTDICCC050	VSTDICCC050	VN085439.D	14 Jan 2025 15:19	Comp.#56 is on Linear Regression	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN085440.D	14 Jan 2025 15:43	Comp.#86 is on Quadratic Regression	JC\MD	Ok,M
5	VSTDICC010	VSTDICC010	VN085441.D	14 Jan 2025 16:07		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN085442.D	14 Jan 2025 16:31		JC\MD	Ok,M
7	VSTDICC001	VSTDICC001	VN085443.D	14 Jan 2025 17:19		JC\MD	Ok,M
8	IBLK	IBLK	VN085444.D	14 Jan 2025 17:42		JC\MD	Ok
9	VSTDICV050	ICVVN011425	VN085445.D	14 Jan 2025 18:06		JC\MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QCBatch ID # VN013125

Review By	Mahesh Dadoda	Review On	2/3/2025 7:21:29 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/3/2025 7:23:27 AM		
SubDirectory	VN013125	HP Acquire Method		HP Processing Method	82N011425W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132796,VP132802
Initial Calibration Stds	
CCC	VP132797,VP132898,VP132803,VP132804
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085590.D	31 Jan 2025 09:30		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN085591.D	31 Jan 2025 10:04	CCAL Low for com.#17	JC\MD	Ok,M
3	VN0131WBL01	VN0131WBL01	VN085592.D	31 Jan 2025 10:42		JC\MD	Ok
4	VN0131WBS01	VN0131WBS01	VN085593.D	31 Jan 2025 11:06		JC\MD	Ok,M
5	VN0131WBSD01	VN0131WBSD01	VN085594.D	31 Jan 2025 11:39		JC\MD	Ok,M
6	Q1211-02RE	TAPIAL2-MW03-01282	VN085595.D	31 Jan 2025 12:03	vial B pH<2	JC\MD	Ok
7	IBLK	IBLK	VN085596.D	31 Jan 2025 12:51		JC\MD	Ok
8	PB166357ZHE#02	PB166357ZHE#02	VN085597.D	31 Jan 2025 13:15		JC\MD	Ok
9	PB166357ZHE#03	PB166357ZHE#03	VN085598.D	31 Jan 2025 13:39		JC\MD	Ok
10	PB166357ZHE#04	PB166357ZHE#04	VN085599.D	31 Jan 2025 14:03		JC\MD	Ok
11	PB166357ZHE#05	PB166357ZHE#05	VN085600.D	31 Jan 2025 14:27		JC\MD	Ok
12	PB166357ZHE#06	PB166357ZHE#06	VN085601.D	31 Jan 2025 14:51		JC\MD	Ok
13	PB166357ZHE#07	PB166357ZHE#07	VN085602.D	31 Jan 2025 15:15		JC\MD	Ok
14	PB166357ZHE#08	PB166357ZHE#08	VN085603.D	31 Jan 2025 15:39		JC\MD	Ok
15	PB166357ZHE#09	PB166357ZHE#09	VN085604.D	31 Jan 2025 16:04		JC\MD	Ok
16	PB166357ZHE#10	PB166357ZHE#10	VN085605.D	31 Jan 2025 16:28		JC\MD	Ok
17	PB166357ZHE#11	PB166357ZHE#11	VN085606.D	31 Jan 2025 16:52		JC\MD	Ok
18	PB166357ZHE#12	PB166357ZHE#12	VN085607.D	31 Jan 2025 17:16		JC\MD	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QCBatch ID # VN013125

Review By	Mahesh Dadoda	Review On	2/3/2025 7:21:29 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/3/2025 7:23:27 AM		
SubDirectory	VN013125	HP Acquire Method		HP Processing Method	82N011425W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132796,VP132802
Initial Calibration Stds	
CCC	VP132797,VP132898,VP132803,VP132804
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	PB166357ZHE#13	PB166357ZHE#13	VN085608.D	31 Jan 2025 17:40		JC\MD	Ok
20	PB166357ZHE#14	PB166357ZHE#14	VN085609.D	31 Jan 2025 18:04		JC\MD	Ok
21	PB166357ZHE#15	PB166357ZHE#15	VN085610.D	31 Jan 2025 18:28		JC\MD	Ok
22	PB166357ZHE#16	PB166357ZHE#16	VN085611.D	31 Jan 2025 18:53		JC\MD	Ok
23	PB166357ZHE#17	PB166357ZHE#17	VN085612.D	31 Jan 2025 19:17		JC\MD	Ok
24	PB166357ZHE#01	PB166357ZHE#01	VN085613.D	31 Jan 2025 19:41		JC\MD	Ok
25	Q1215-02	JPP-29.1-012825	VN085614.D	31 Jan 2025 20:05	vial A pH#5.0	JC\MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VN085615.D	31 Jan 2025 20:29		JC\MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX011525

Review By	John Carlone	Review On	1/15/2025 4:32:35 PM		
Supervise By	Mahesh Dadoda	Supervise On	1/16/2025 7:35:39 AM		
SubDirectory	VX011525	HP Acquire Method		HP Processing Method	82X011525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132553
Initial Calibration Stds	VP132558,VP132559,VP132560,VP132561,VP132562,VP132563
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP132564
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044647.D	15 Jan 2025 09:30		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX044648.D	15 Jan 2025 10:31		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX044649.D	15 Jan 2025 11:17		JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX044650.D	15 Jan 2025 11:40		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX044651.D	15 Jan 2025 12:02	Method failed for Acrolein	JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX044652.D	15 Jan 2025 13:14		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX044653.D	15 Jan 2025 13:37		JC/MD	Ok,M
8	IBLK	IBLK	VX044654.D	15 Jan 2025 14:00		JC/MD	Ok
9	VSTDICV050	ICVVX011525	VX044655.D	15 Jan 2025 15:15		JC/MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX020325

Review By	John Carlone	Review On	2/4/2025 9:51:15 AM		
Supervise By	Mahesh Dadoda	Supervise On	2/4/2025 3:45:43 PM		
SubDirectory	VX020325	HP Acquire Method		HP Processing Method	82X011525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132831
Initial Calibration Stds	
CCC	VP132832,VP132833
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044801.D	03 Feb 2025 08:12		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX044802.D	03 Feb 2025 09:52		JC/MD	Ok,M
3	VX0203MBL01	VX0203MBL01	VX044803.D	03 Feb 2025 10:21	V13516	JC/MD	Ok
4	VX0203WBL01	VX0203WBL01	VX044804.D	03 Feb 2025 10:44		JC/MD	Ok
5	VX0203WBS01	VX0203WBS01	VX044805.D	03 Feb 2025 11:07		JC/MD	Ok,M
6	VX0203WBSD01	VX0203WBSD01	VX044806.D	03 Feb 2025 11:32		JC/MD	Ok,M
7	PB166357TB	PB166357TB	VX044807.D	03 Feb 2025 11:55		JC/MD	Ok
8	Q1215-06	JPP-29.2-012825	VX044808.D	03 Feb 2025 12:18	vial A pH#5.0	JC/MD	Ok
9	Q1216-02	JPP-18.1-012825	VX044809.D	03 Feb 2025 12:41	vial A pH#5.0	JC/MD	Ok
10	Q1216-06	JPP-21.1-012825	VX044810.D	03 Feb 2025 13:04	vial A pH#5.0	JC/MD	Ok
11	Q1216-10	JPP-21.2-012825	VX044811.D	03 Feb 2025 13:27	vial A pH#5.0	JC/MD	Ok
12	Q1221-02	JPP-26.1-012825	VX044812.D	03 Feb 2025 13:50	vial A pH#5.0	JC/MD	Ok,M
13	Q1218-02	BELL-25-002	VX044813.D	03 Feb 2025 14:36	vial A pH#5.0	JC/MD	Ok
14	Q1239-07	RBR22266	VX044814.D	03 Feb 2025 14:59		JC/MD	Ok
15	VX0203MBS01	VX0203MBS01	VX044815.D	03 Feb 2025 15:22		JC/MD	Ok,M
16	Q1232-06	JPP-46.1-012925	VX044816.D	03 Feb 2025 15:45	vial A pH#5.0	JC/MD	Ok
17	Q1239-10	357	VX044817.D	03 Feb 2025 16:08		JC/MD	Ok
18	Q1232-14	JPP-42.2-012925	VX044818.D	03 Feb 2025 16:32	vial A pH#5.0	JC/MD	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QCBatch ID # VX020325

Review By	John Carlone	Review On	2/4/2025 9:51:15 AM		
Supervise By	Mahesh Dadoda	Supervise On	2/4/2025 3:45:43 PM		
SubDirectory	VX020325	HP Acquire Method		HP Processing Method	82X011525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132831
Initial Calibration Stds	
CCC	VP132832,VP132833
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q1232-18	JPP-51.1-012925	VX044819.D	03 Feb 2025 16:55	vial A pH#5.0	JC/MD	Ok
20	Q1235-02	JPP-51.2-012925	VX044820.D	03 Feb 2025 17:18	vial A pH#5.0	JC/MD	Ok
21	Q1235-06	JPP-16.1-012925	VX044821.D	03 Feb 2025 17:41	vial A pH#5.0	JC/MD	Ok
22	Q1239-02	286	VX044822.D	03 Feb 2025 18:04	vial A pH#5.0	JC/MD	Ok
23	Q1239-05	348	VX044823.D	03 Feb 2025 18:27	vial A pH#5.0	JC/MD	Ok
24	Q1239-08	RBR22266	VX044824.D	03 Feb 2025 18:50	vial A pH#5.0	JC/MD	Ok
25	Q1239-11	357	VX044825.D	03 Feb 2025 19:14	vial A pH#5.0	JC/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VX044826.D	03 Feb 2025 19:37		JC/MD	Ok,M

M : Manual Integration

Alliance TECHNICAL GROUP

TCLP EXTRACTION LOGPAGE

PB166357

SOP ID : M1311-TCLP-15

SDG No : N/A

Weigh By : JP

Balance ID : WC SC-7

pH Meter ID : WC PH METER-1

Extraction By : JP

Filter By : JP

Pippete ID : WC

Tumbler ID : ZHE-1 / ZHE-2

TCLP Filter ID : 50223706

Start Prep Date : 01/30/2025 **Time :** 14:00

End Prep Date : 01/31/2025 **Time :** 07:15

Combination Ratio : 20

ZHE Cleaning Batch : ~~N/A~~ VN013125, JP

Initial Room Temperature: 23 °C

Final Room Temperature: 22 °C

TCLP Technician Signature : JP

Supervisor By : IZ

Standared Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP110801
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
40ml VOA Vials	23237	N/A

Extraction Conformance/Non-Conformance Comments:

ALL ZHE samples are extracted and given as vial A & B. Leak checked after 10 mintues of tumbling. Tumbler ZHE-1/ ZHE-2 checked,30 rpm.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
01/31/25 09:40	JP / ICP Room	S.Y / VOC Lab
	Preparation Group	Analysis Group

Alliance TECHNICAL GROUP

TCLP EXTRACTION LOGPAGE

PB166357

Sample ID	ClientID	ZHE Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB166357TB	LEB357	N/A	N/A	500	N/A	N/A	N/A	4.94	N/A	ZHE-2
Q1215-02	JPP-29.1-012825	01	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q1215-06	JPP-29.2-012825	02	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q1216-02	JPP-18.1-012825	03	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q1216-06	JPP-21.1-012825	04	25.04	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q1216-10	JPP-21.2-012825	05	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q1216-14	JPP-26.1-012825	06	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q1216-18	JPP-26.2-012825	07	25.01	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q1218-02	BELL-25-002	08	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q1219-02	LAW-25-0015	09	25.01	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q1221-02	CHESTNUT-CONCRETE	10	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q1232-02	JPP-46.2-012925	11	25.04	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
Q1232-06	JPP-46.1-012925	12	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
Q1232-10	JPP-42.1-012925	13	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
Q1232-14	JPP-42.2-012925	14	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
Q1232-18	JPP-51.1-012925	15	25.01	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
Q1235-02	JPP-51.2-012925	16	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-2
Q1235-06	JPP-16.1-012925	17	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB166357TB	LEB357	N/A	N/A	N/A	N/A	N/A	N/A
Q1215-02	JPP-29.1-012825	N/A	N/A	N/A	N/A	100	N/A
Q1215-06	JPP-29.2-012825	N/A	N/A	N/A	N/A	100	N/A
Q1216-02	JPP-18.1-012825	N/A	N/A	N/A	N/A	100	N/A
Q1216-06	JPP-21.1-012825	N/A	N/A	N/A	N/A	100	N/A
Q1216-10	JPP-21.2-012825	N/A	N/A	N/A	N/A	100	N/A
Q1216-14	JPP-26.1-012825	N/A	N/A	N/A	N/A	100	N/A
Q1216-18	JPP-26.2-012825	N/A	N/A	N/A	N/A	100	N/A
Q1218-02	BELL-25-002	N/A	N/A	N/A	N/A	100	N/A
Q1219-02	LAW-25-0015	N/A	N/A	N/A	N/A	100	N/A
Q1221-02	CHESTNUT-CONCRETE	N/A	N/A	N/A	N/A	100	N/A
Q1232-02	JPP-46.2-012925	N/A	N/A	N/A	N/A	100	N/A
Q1232-06	JPP-46.1-012925	N/A	N/A	N/A	N/A	100	N/A
Q1232-10	JPP-42.1-012925	N/A	N/A	N/A	N/A	100	N/A
Q1232-14	JPP-42.2-012925	N/A	N/A	N/A	N/A	100	N/A
Q1232-18	JPP-51.1-012925	N/A	N/A	N/A	N/A	100	N/A
Q1235-02	JPP-51.2-012925	N/A	N/A	N/A	N/A	100	N/A
Q1235-06	JPP-16.1-012925	N/A	N/A	N/A	N/A	100	N/A

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TCLP ZHE Q1221 **WorkList ID :** 187276 **Department :** TCLP Extraction **Date :** 01-30-2025 08:09:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1215-02	JPP-29.1-012825	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/28/2025	1311 ZHE
Q1215-06	JPP-29.2-012825	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/28/2025	1311 ZHE
Q1216-02	JPP-18.1-012825	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/28/2025	1311 ZHE
Q1216-06	JPP-21.1-012825	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/28/2025	1311 ZHE
Q1216-10	JPP-21.2-012825	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/28/2025	1311 ZHE
Q1216-14	JPP-26.1-012825	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/28/2025	1311 ZHE
Q1216-18	JPP-26.2-012825	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/28/2025	1311 ZHE
Q1218-02	BELL-25-002	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/28/2025	1311 ZHE
Q1219-02	LAW-25-0015	Solid	TCLP ZHE Extraction	Cool 4 deg C	PSEG03	N41	01/29/2025	1311 ZHE
Q1221-02	CHESTNUT-CONCRETE	Solid	TCLP ZHE Extraction	Cool 4 deg C	PSEG03	N41	01/29/2025	1311 ZHE
Q1232-02	JPP-46.2-012925	Solid	TCLP ZHE Extraction	Cool 4 deg C	PSEG03	N41	01/29/2025	1311 ZHE
Q1232-06	JPP-46.1-012925	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/29/2025	1311 ZHE
Q1232-10	JPP-42.1-012925	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/29/2025	1311 ZHE
Q1232-14	JPP-42.2-012925	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/29/2025	1311 ZHE
Q1232-18	JPP-51.1-012925	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/29/2025	1311 ZHE
Q1235-02	JPP-51.2-012925	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/29/2025	1311 ZHE
Q1235-06	JPP-16.1-012925	Solid	TCLP ZHE Extraction	Cool 4 deg C	RUTW01	E11	01/29/2025	1311 ZHE

Date/Time 01/30/25 12:30

Raw Sample Received by: JP COC

Raw Sample Relinquished by: CR SM

Date/Time 01/30/25 16:00

Raw Sample Received by: CR SM

Raw Sample Relinquished by: JP COC

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : Q1215

Test : TCLP VOA

Prepbatch ID :

Sequence ID/Qc Batch ID: VN013125,VX020325,

Standard ID :
VP130430,VP131746,VP131767,VP132035,VP132096,VP132468,VP132613,VP132692,VP132711,VP132796,VP132797,VP132802,VP132803,VP132804,VP132819,VP132831,VP132832,VP132833,VP132898,

Chemical ID :
V13391,V13440,V13446,V13457,V13460,V13465,V13466,V13707,V13845,V13858,V14081,V14145,V14154,V14175,V14176,V14289,V14308,V14352,V14433,V14439,V14454,V14521,V14522,V14554,V14605,V14607,V14610,V14611,V14614,V14624,V14627,V14630,V14631,V14632,V14633,V14722,V14723,V14724,V14726,V14753,V14754,V14756,V14801,V14814,V14830,V14831,V14832,W3112,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
617	8260 Surrogate, 400PPM	VP130430	09/20/2024	02/28/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 09/26/2024

FROM 0.40000ml of V13707 + 24.60000ml of V14145 = Final Quantity: 25.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
247	8260 Internal Standard, 250PPM	VP131746	11/22/2024	05/18/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 11/23/2024

FROM 0.50000ml of V14289 + 49.50000ml of V14154 = Final Quantity: 50.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
218	BFB, 25PPM	VP131767	11/22/2024	05/18/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 11/27/2024

FROM 0.50000ml of V13391 + 49.50000ml of V14154 = Final Quantity: 50.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1810	8260 Working Std(2-CVE)-800ppm	VP132035	12/10/2024	06/10/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 12/12/2024

FROM 1.00000ml of V14630 + 1.00000ml of V14631 + 1.00000ml of V14632 + 1.00000ml of V14633 + 46.00000ml of V14614 = Final Quantity: 50.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
719	8260 Working STD (BCM)-First source, 400PPM	VP132096	12/12/2024	06/10/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 12/19/2024

FROM 1.00000ml of V13465 + 1.00000ml of V13466 + 1.50000ml of V13457 + 1.50000ml of V13460 + 20.00000ml of V14614 = Final Quantity: 25.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
51	8260 Working STD (Acrolein) -first source, 800PPM	VP132468	01/08/2025	02/07/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 01/17/2025

FROM 1.00000ml of V14832 + 1.50000ml of V14830 + 1.50000ml of V14831 + 21.00000ml of V14627 = Final Quantity: 25.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
257	8260 Calibration Working STD Mix-First source, 160PPM	VP132613	01/20/2025	02/28/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 01/29/2025

FROM 0.40000ml of V13446 + 1.00000ml of V14175 + 1.00000ml of V14176 + 1.00000ml of V14433 + 1.00000ml of V14439 + 1.00000ml of V14521 + 1.00000ml of V14522 + 1.00000ml of V14722 + 1.00000ml of V14754 + 1.00000ml of V14756 + 1.00000ml of V14801 + 1.00000ml of V14814 + 1.50000ml of V14723 + 1.50000ml of V14724 + 10.60000ml of V14624 = Final Quantity: 25.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1896	Trace internal standard 50 ppm	VP132692	01/27/2025	03/01/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 01/29/2025

FROM 0.20000ml of V14352 + 9.80000ml of V14624 = Final Quantity: 10.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1897	Trace surrogate mix 25 ppm	VP132711	01/27/2025	03/01/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 01/29/2025

FROM 0.50000ml of V14454 + 0.50000ml of V14611 + 1.50000ml of V14605 + 1.50000ml of V14607 + 1.50000ml of V14610 + 4.50000ml of V14624 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
589	BFB TUNE CHECK	VP132796	01/31/2025	02/01/2025	John Carlone	None	None	Semsettin Yesilyurt 02/03/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP131767 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
620	50 PPB CCC, 8260-Water	VP132797	01/31/2025	02/01/2025	John Carlone	None	None	Semsettin Yesilyurt 02/03/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP130430 + 0.00500ml of VP132096 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP132468 + 0.01250ml of VP132613 = Final Quantity: 40.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
589	BFB TUNE CHECK	VP132802	01/31/2025	02/01/2025	John Carlone	None	None	Semsettin Yesilyurt 02/03/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP131767 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
620	50 PPB CCC, 8260-Water	VP132803	01/31/2025	02/01/2025	John Carlone	None	None	Semsettin Yesilyurt 02/03/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP130430 + 0.00500ml of VP132096 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP132468 + 0.01250ml of VP132613 = Final Quantity: 40.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
620	50 PPB CCC, 8260-Water	VP132804	01/31/2025	02/01/2025	John Carlone	None	None	Semsettin Yesilyurt 02/03/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP130430 + 0.00500ml of VP132096 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP132468 + 0.01250ml of VP132613 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1721	SOM01.2 TRACE-Calibration Mix,25 PPM	VP132819	01/30/2025	03/08/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 02/06/2025

FROM 0.12500ml of V13440 + 0.12500ml of V13845 + 0.12500ml of V13858 + 0.12500ml of V14081 + 0.12500ml of V14308 + 0.12500ml of V14554 + 0.12500ml of V14753 + 0.50000ml of V14726 + 8.62500ml of V14624 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
589	BFB TUNE CHECK	VP132831	02/03/2025	02/04/2025	John Carlone	None	None	Semsettin Yesilyurt 02/05/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP131767 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
620	50 PPB CCC, 8260-Water	VP132832	02/03/2025	02/04/2025	John Carlone	None	None	Semsettin Yesilyurt 02/05/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP130430 + 0.00500ml of VP132096 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP132468 + 0.01250ml of VP132613 = Final Quantity: 40.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
620	50 PPB CCC, 8260-Water	VP132833	02/03/2025	02/04/2025	John Carlone	None	None	Semsettin Yesilyurt 02/05/2025

FROM 39.94450ml of W3112 + 0.00500ml of VP130430 + 0.00500ml of VP132096 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP132468 + 0.01250ml of VP132613 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1727	5 PPB CCC-CCV SOM01.2 Trace	VP132898	02/05/2025	02/06/2025	Amit Patel	None	None	Semsettin Yesilyurt 02/06/2025

FROM 39.98000ml of W3112 + 0.00400ml of VP132692 + 0.00800ml of VP132711 + 0.00800ml of VP132819 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30429 / 1,2,3-Trichloropropane Standard, 2,000 ug/ml	A0188973	07/30/2025	01/30/2025 / SAM	01/23/2023 / SAM	V13440

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0181905	02/28/2025	01/10/2025 / SAM	01/23/2023 / SAM	V13446

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13457

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13460

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13465

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13466

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	06/10/2025	06/10/2024 / SAM	04/12/2023 / SAM	V13707

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	31491 / 1,2,4-Trimethylbenzene 2000ppm	040821	07/30/2025	01/30/2025 / SAM	06/22/2023 / SAM	V13845

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	90319 / 1,3,5-Trimethylbenzene-2000 ug/mL	061923	07/30/2025	01/30/2025 / SAM	06/22/2023 / SAM	V13858

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	94159 / CLP SOM01.1 Volatiles	012323	07/30/2025	01/30/2025 / SAM	12/21/2023 / SAM	V14081

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	02/28/2025	08/29/2024 / SAM	02/06/2024 / SAM	V14145

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	05/18/2025	11/18/2024 / pedro	02/06/2024 / SAM	V14154

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	07/10/2025	01/10/2025 / SAM	02/20/2024 / SAM	V14175

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	07/10/2025	01/10/2025 / SAM	02/20/2024 / SAM	V14176

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	11/22/2025	11/22/2024 / SAM	04/15/2024 / SAM	V14289

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	90298 / Naphthalene, 2000 ug/ml	020123	07/30/2025	01/30/2025 / SAM	04/17/2024 / SAM	V14308

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30091 / VOA Mix, CLP method L/C Internal Std 2500uq/ml, PT&M, 1ml/ampul	A0209905	04/14/2025	10/14/2024 / SAM	05/03/2024 / SAM	V14352

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	07/10/2025	01/10/2025 / SAM	08/15/2024 / SAM	V14433

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	07/10/2025	01/10/2025 / SAM	08/15/2024 / SAM	V14439

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30624 / VOA Stock Standard, OLC 3.2 VOA non-ketone, deuterated monitoring compounds, 1mL, 500ug/mL, Methanol-d	A0211457	07/27/2025	01/27/2025 / SAM	08/15/2024 / SAM	V14454

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	07/10/2025	01/10/2025 / SAM	09/18/2024 / SAM	V14521

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	07/10/2025	01/10/2025 / SAM	09/18/2024 / SAM	V14522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	94559 / 1,3,5-Trichlorobenzene, 2000 ug/mL, in methanol	051421	07/30/2025	01/30/2025 / SAM	10/09/2024 / SAM	V14554

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30625 / VOA Stock Std, OLC 3.2 VOA Ketone Deuterated Monitoring Compounds, 1mL, 500ug/mL, d2O		07/13/2025	01/13/2025 / SAM	11/18/2024 / SAM	V14605

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30625 / VOA Stock Std, OLC 3.2 VOA Ketone Deuterated Monitoring Compounds, 1mL, 500ug/mL, d2O		07/13/2025	01/13/2025 / SAM	11/18/2024 / SAM	V14607

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30625 / VOA Stock Std, OLC 3.2 VOA Ketone Deuterated Monitoring Compounds, 1mL, 500ug/mL, d2O	A0219189	06/12/2025	12/12/2024 / SAM	11/22/2024 / SAM	V14610

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30625 / VOA Stock Std, OLC 3.2 VOA Ketone Deuterated Monitoring Compounds, 1mL, 500ug/mL, d2O	A0219189	06/12/2025	12/12/2024 / SAM	11/22/2024 / SAM	V14611

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	06/10/2025	12/10/2024 / SAM	11/26/2024 / SAM	V14614

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	07/13/2025	01/13/2025 / SAM	11/26/2024 / SAM	V14624

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	07/06/2025	01/06/2025 / SAM	11/26/2024 / SAM	V14627

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14630

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14631

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14632

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14633

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	07/10/2025	01/10/2025 / SAM	12/17/2024 / SAM	V14722

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	07/10/2025	01/10/2025 / SAM	12/17/2024 / SAM	V14723

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	07/10/2025	01/10/2025 / SAM	12/17/2024 / SAM	V14724

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	07/30/2025	01/30/2025 / SAM	12/17/2024 / SAM	V14726

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	07/30/2025	01/30/2025 / SAM	12/17/2024 / SAM	V14753

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	05/31/2031	01/10/2025 / SAM	12/17/2024 / SAM	V14754

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	07/10/2025	01/10/2025 / SAM	12/17/2024 / SAM	V14756

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE LOTS	A0220563	06/30/2026	01/10/2025 / SAM	01/08/2025 / SAM	V14801

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE LOTS	A0220471	07/10/2025	01/10/2025 / SAM	01/08/2025 / SAM	V14814

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	010725	02/07/2025	01/08/2025 / SAM	01/08/2025 / SAM	V14830

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	010725	02/07/2025	01/08/2025 / SAM	01/08/2025 / SAM	V14831

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	010725	02/07/2025	01/08/2025 / SAM	01/08/2025 / SAM	V14832

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 2310762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 2310762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

12 Baxter

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

James Ethier
Jamie Ethier
Vice President Global Quality

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

Dec 01/08/25

5 vial

V14823-to V14827 V14829

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 010725
Description: Acrolein

Expiration Date: 020725
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

Solvent(s): Water **Lot#** 072324Q


5E-05 Balance Uncertainty
0.001 Flask Uncertainty


Weight(s) shown below were combined and diluted to (mL): 10.0

	010725
Formulated By: Anthony Mahoney	DATE
	010725
Reviewed By: Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05178	5011.8	52.6	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.

PO Box 5585
Hamden, CT 06518-0585

Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN WATER**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr.	Emergency Telephone International	**1-352-323-3500**
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	**Use in ventilated area**	**H315**	**Causes skin and eye irritation.**
P302,332	**If on skin, wash with soap and water**	**P280**	**Use gloves, eye protection/face sheild**
		P305,351,338	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Water	CAS#: 7732-18-5	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water CAS#: 7732-18-5 TWA: 500 ppm

Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H2O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

	NA		0°C
Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	NA

Solubility in Water — Completely miscible

Appearance and Odor — CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability — Stable under recommended storage conditions.
Possibility of hazardous reactions — NA
Conditions to avoid — NA
Materials to avoid — NA
Hazardous decomposition products - No data available

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat — NA
LC50 Inhalation - Rat — NA
LD50 Dermal - Guinea pig — NA
Causes skin irritation.
Eye irritation

Section XII. ECOLOGICAL INFORMATION

LC50 — NA
EC50 — NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
Not dangerous goods
Proper shipping name: Water

IATA
Not dangerous goods
Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

Dec 01/08/25

5 vial

V14823-to V14827 V14829

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 010725
Description: Acrolein

Expiration Date: 020725
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

Solvent(s): Water **Lot#** 072324Q


5E-05 Balance Uncertainty
0.001 Flask Uncertainty


Weight(s) shown below were combined and diluted to (mL): 10.0

Formulated By: Anthony Mahoney	010725 DATE
Reviewed By: Pedro L. Rentas	010725 DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05178	5011.8	52.6	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN WATER**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr.	Emergency Telephone International	**1-352-323-3500**
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	**Use in ventilated area**	**H315**	**Causes skin and eye irritation.**
P302,332	**If on skin, wash with soap and water**	**P280**	**Use gloves, eye protection/face sheild**
		P305,351,338	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Water	CAS#: 7732-18-5	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water CAS#: 7732-18-5 TWA: 500 ppm

Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H2O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

	NA		0°C
Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	NA

Solubility in Water — Completely miscible

Appearance and Odor — CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability — Stable under recommended storage conditions.
Possibility of hazardous reactions — NA
Conditions to avoid — NA
Materials to avoid — NA
Hazardous decomposition products - No data available

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat — NA
LC50 Inhalation - Rat — NA
LD50 Dermal - Guinea pig — NA
Causes skin irritation.
Eye irritation

Section XII. ECOLOGICAL INFORMATION

LC50 — NA
EC50 — NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
Not dangerous goods
Proper shipping name: Water

IATA
Not dangerous goods
Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

Dec 01/08/25

5 vial

V14823-to V14827 V14829

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 010725
Description: Acrolein

Expiration Date: 020725
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

Solvent(s): Water **Lot#** 072324Q


5E-05 Balance Uncertainty
0.001 Flask Uncertainty


Weight(s) shown below were combined and diluted to (mL): 10.0

	010725
Formulated By: Anthony Mahoney	DATE
	010725
Reviewed By: Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05178	5011.8	52.6	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN WATER**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr.	Emergency Telephone International	**1-352-323-3500**
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	**Use in ventilated area**	**H315**	**Causes skin and eye irritation.**
P302,332	**If on skin, wash with soap and water**	**P280**	**Use gloves, eye protection/face sheild**
		P305,351,338	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Water	CAS#: 7732-18-5	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water CAS#: 7732-18-5 TWA: 500 ppm

Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H2O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

	NA		0°C
Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	NA

Solubility in Water — Completely miscible

Appearance and Odor — CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability — Stable under recommended storage conditions.
Possibility of hazardous reactions — NA
Conditions to avoid — NA
Materials to avoid — NA
Hazardous decomposition products - No data available

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat — NA
LC50 Inhalation - Rat — NA
LD50 Dermal - Guinea pig — NA
Causes skin irritation.
Eye irritation

Section XII. ECOLOGICAL INFORMATION

LC50 — NA
EC50 — NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
Not dangerous goods
Proper shipping name: Water

IATA
Not dangerous goods
Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

10 ug/mL

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 94159
Lot Number: 012323
Description: CLP SOM 01.1 Volatiles
42 components
Expiration Date: 012326
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): Varied
NIST Test ID#: 6UTB

Solvent: Methanol **Lot#** EF282-USQ1


5E-05 Balance Uncertainty
Volume(s) shown below were combined and diluted to (mL): 100.0 0.012 Flask Uncertainty

Formulated By:	Prashant Chauhan	012323 DATE
Reviewed By:	Pedro L. Rentas	012323 DATE

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(ug/mL)	Final Conc.(ug/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Benzene	93831	060616	0.10	10.00	0.042	20006.6	2000.6	18.7	71-43-2	1 ppm	orl-rat 4894mg/kg
2. Toluene	93831	060616	0.10	10.00	0.042	20002.8	2000.2	18.7	108-88-3	200 ppm	orl-rat 5000mg/kg
3. Ethyl benzene	93831	060616	0.10	10.00	0.042	20002.5	2000.2	18.7	100-41-4	100 ppm (435mg/m3/8H)	orl-rat >2000mg/kg
4. o-Xylene	93831	060616	0.10	10.00	0.042	20003.8	2000.3	18.7	95-47-6	100 ppm (435mg/m3/8H)	ipr-mus 1364mg/kg
5. m-Xylene	93831	060616	0.10	10.00	0.042	10003.3	1000.3	9.4	108-38-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg
6. p-Xylene	93831	060616	0.10	10.00	0.042	10005.2	1000.5	9.4	106-42-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg
7. Bromodichloromethane	35171	100220	0.05	5.00	0.017	40018.8	2000.8	15.9	75-27-4	N/A	orl-rat 916mg/kg
8. Dibromochloromethane	35171	100220	0.05	5.00	0.017	40007.7	2000.3	15.9	124-48-1	N/A	orl-rat 848mg/kg
9. cis-1,2-Dichloroethene	35171	100220	0.05	5.00	0.017	40012.4	2000.5	15.8	156-59-2	N/A	N/A
10. trans-1,2-Dichloroethene	35171	100220	0.05	5.00	0.017	40005.6	2000.2	15.9	156-60-5	N/A	orl-rat 1235mg/kg
11. Methylene chloride	35171	100220	0.05	5.00	0.017	40013.9	2000.6	15.8	75-09-2	500 ppm	orl-rat 820mg/kg
12. 1,1-Dichloroethene	32251	031821	0.10	10.00	0.042	20009.1	2000.8	18.7	75-35-4	1 ppm (4mg/m3/8H)	orl-rat 200mg/kg
13. Bromochloromethane	94170	010616	0.10	10.00	0.042	20017.5	2001.7	18.7	74-97-5	200 ppm (1050mg/m3/8H)	orl-rat 5000mg/kg
14. Bromoform	94170	010616	0.10	10.00	0.042	20010.4	2000.9	18.7	75-25-2	0.5 ppm (5mg/m3) (skin)	orl-rat 933mg/kg
15. Carbon tetrachloride	94170	010616	0.10	10.00	0.042	20006.0	2000.5	18.7	56-23-5	2 ppm (12.6mg/m3/8H)	orl-rat 2350mg/kg
16. Chloroform	94170	010616	0.10	10.00	0.042	20019.5	2001.8	18.7	67-66-3	50 ppm (240mg/m3) (CL)	orl-rat 908mg/kg
17. 1,1-Dichloroethane	94170	010616	0.10	10.00	0.042	20007.6	2000.7	18.7	75-34-3	100 ppm	orl-rat 725mg/kg
18. Tetrachloroethene	94170	010616	0.10	10.00	0.042	20015.7	2001.5	18.7	127-18-4	25 ppm (170mg/m3/8H)(final)	orl-rat 2629mg/kg
19. 1,1,1-Trichloroethane	94170	010616	0.10	10.00	0.042	20007.4	2000.6	18.7	71-55-6	350 ppm (1900mg/m3/8H)	orl-rat 10300mg/kg
20. 1,2-Dibromo-3-chloropropane	94171	010616	0.10	10.00	0.042	20015.3	2001.4	18.7	96-12-8	0.001 ppm	orl-rat 170mg/kg
21. 1,2-Dibromoethane	94171	010616	0.10	10.00	0.042	20017.3	2001.6	18.7	106-93-4	20 ppm (8H)	orl-rat 108mg/kg
22. 1,2-Dichloroethane	94171	010616	0.10	10.00	0.042	20005.4	2000.4	18.7	107-06-2	50 ppm (8H)	orl-rat 670mg/kg
23. 1,2-Dichloropropane	94171	010616	0.10	10.00	0.042	20006.4	2000.5	18.7	78-87-5	75 ppm (350mg/m3/8H)	orl-rat 1947mg/kg
24. cis-1,3-Dichloropropene	94171	010616	0.10	10.00	0.042	20016.0	2001.5	18.7	10061-01-5	N/A	N/A
25. trans-1,3-Dichloropropene	94171	010616	0.10	10.00	0.042	20011.4	2001.0	18.7	10061-02-6	N/A	N/A
26. 1,1,2,2-Tetrachloroethane	94171	010616	0.10	10.00	0.042	20014.3	2001.3	18.7	79-34-5	5 ppm (35mg/m3/8H)(skin)	orl-rat 800mg/kg
27. 1,1,2-Trichloroethane	94171	010616	0.10	10.00	0.042	20024.9	2002.4	18.8	79-00-5	10 ppm (45mg/m3/8H)(skin)	orl-rat 836mg/kg
28. Trichloroethene	94171	010616	0.10	10.00	0.042	20012.9	2001.2	18.7	79-01-6	50 ppm (270mg/m3/8H)	orl-mus 2402mg/kg
29. Chlorobenzene	99783	091118	0.10	10.00	0.042	20001.9	2000.1	18.7	108-90-7	75 ppm (350mg/m3/8H)	orl-rat 2290mg/kg
30. 1,2-Dichlorobenzene	99783	091118	0.10	10.00	0.042	20002.9	2000.2	18.7	95-50-1	50 ppm (300mg/m3) (CL)	orl-rat 500mg/kg
31. 1,3-Dichlorobenzene	99783	091118	0.10	10.00	0.042	20003.7	2000.3	18.7	541-73-1	N/A	ipr-mus 1062mg/kg
32. 1,4-Dichlorobenzene	99783	091118	0.10	10.00	0.042	20005.9	2000.5	18.7	106-46-7	75 ppm (450mg/m3/8H)	orl-rat 500mg/kg
33. Isopropylbenzene	99783	091118	0.10	10.00	0.042	20391.8	2039.1	19.0	98-82-8	50 ppm (245mg/m3/8H)	orl-rat 1400mg/kg
34. 1,2,3-Trichlorobenzene	99783	091118	0.10	10.00	0.042	20003.7	2000.3	18.7	87-61-6	N/A	ipr-mus 1390mg/kg
35. 1,2,4-Trichlorobenzene	99783	091118	0.10	10.00	0.042	20084.7	2008.4	18.8	120-82-1	5 ppm (CL) (40mg/m3)	orl-rat 756mg/kg
36. Styrene	32381	052120	0.10	10.00	0.042	20041.4	2004.0	18.7	100-42-5	100 ppm	orl-rat 5000mg/kg
37. Carbon disulphide	94173	010716	0.10	10.00	0.042	20001.9	2000.1	18.7	75-15-0	4 ppm (12mg/m3) (skin)	orl-rat 1200mg/kg
38. Cyclohexane	94173	010716	0.10	10.00	0.042	20002.0	2000.1	18.7	110-82-7	300 ppm (1050mg/m3/8H)	orl-rat 12705mg/kg
39. Methyl acetate	94173	010716	0.10	10.00	0.042	20002.4	2000.1	18.7	79-20-9	200 ppm (610mg/m3/8H)	orl-rbt 3705mg/kg
40. Methylcyclohexane	94173	010716	0.10	10.00	0.042	20001.7	2000.1	18.7	108-87-2	N/A	orl-mus 2250mg/kg
41. Methyl tert-butyl ether (MTBE)	94173	010716	0.10	10.00	0.042	20001.2	2000.0	18.7	1634-04-4	N/A	orl-rat 4g/kg
42. 1,1,2-Trichloro-1,2,2-trifluoroethane	94173	010716	0.10	10.00	0.042	20001.9	2000.1	18.7	76-13-1	1000 ppm (7600mg/m3/8H)	orl-rat 43g/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Method: GC6MSD1. **Detector:** Mass Selective Detector. **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1=10min.), Temp. 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector Temp.= 200°C, Detector Temp. = 220°C. **Analyst:** Gina McLane.



Retention Time (min.)	
1,1,2-Trichloro-1,2,2-trifluoroethane	8.99
1,1-Dichloroethene	9.83
Methyl acetate	11.55
Carbon disulphide	11.91
Methylene chloride	12.05
Methyl tert-butyl ether (MTBE)	12.62
trans-1,2-Dichloroethene	13.18
1,1-Dichloroethane	15.12
cis-1,2-Dichloroethene	17.67
Chloroform	18.29
Bromochloromethane	19.03
Cyclohexane	19.62
1,1,1-Trichloroethane	19.79
Carbon tetrachloride	20.76
Benzene	21.57
1,2-Dichloroethane	21.57
Trichloroethene	23.89
Methylcyclohexane	23.89
1,2-Dichloropropane	24.7
Bromodichloromethane	25.68
cis-1,3-Dichloropropene	27.8
Toluene	28.94
trans-1,3-Dichloropropene	29.79
1,1,2-Trichloroethane	30.41
Tetrachloroethene	31.42
Dibromochloromethane	32.47
1,2-Dibromoethane	33.25
Chlorobenzene	34.81
Ethylbenzene	34.94
m-Xylene	35.22
p-Xylene	35.22
o-Xylene	36.99
Styrene	37.19
Isopropylbenzene	38.38
Bromoform	38.73
1,1,2,2-Tetrachloroethane	39.26
1,3-Dichlorobenzene	43.98
1,4-Dichlorobenzene	44.43
1,2-Dichlorobenzene	45.97
impurity	48.67
1,2-Dibromo-3-chloropropane	49.27
1,2,4-Trichlorobenzene	52.79
1,2,3-Trichlorobenzene	55.22

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

10


CERTIFIED WEIGHT REPORT


Part Number:	**90298**	**Solvent(s):**	**Lot#**	
Lot Number:	020123	Methanol	EF282-US	
Description:	Naphthalene			
Expiration Date:	020128			
Recommended Storage:	Refrigerate (4 °C)			
Nominal Concentration (µg/mL):	2000			
NIST Test ID#:	6UTB	5E-05	Balance Uncertainty	
Volume(s) shown below were combined and diluted to (mL):	100.0	0.012	Flask Uncertainty	

Justin Dippold		020123
Formulated By:	Justin Dippold	DATE
Pedro Rentas		020123
Reviewed By:	Pedro L. Rentas	DATE

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(ug/mL)	Final Conc.(ug/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Naphthalene	32361	011623	0.10	10.00	0.042	20000.3	1999.9	18.7	91-20-3	10 ppm (50mg/m3/8H)	orl-rat 490mg/kg

Method: GC6MSD-1. **Detector:** MSD (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1=35°C (Time 1=10min.), Temp. 2=200°C (Time 2=8.75 min.), Rate=4°C/min., Injector Temp.=200°C, Detector Temp.=220°C. **Analyst:** Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

CERTIFIED WEIGHT REPORT

Part Number: 95317
Lot Number: 021624
Description: Universal VOA Megamix
69 components
Expiration Date: 021627
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 6UTB

Solvent(s): Methanol **Lot#** EG359-USQ12

Formulated By: Prashant Chauhan 021624 DATE

Reviewed By: Pedro L. Rentas 021624 DATE

Weight(s) shown below were combined and diluted to (mL): 100.0 5E-05 Balance Uncertainty 0.021 Flask Uncertainty


	Compound	(RM#) Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Initial Conc.(µg/mL)	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1.	Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	orl-rat 2460mg/kg
2.	Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	orl-rat 700mg/kg
3.	Carbon disulphide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	orl-rat 1200mg/kg
4.	cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1476-11-5	N/A	N/A
5.	trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	N/A	N/A
6.	Diethyl ether	(0153)	IK18CAS0000	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7.	Ethyl methacrylate	(0381)	06126PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	orl-rat 14800mg/kg
8.	Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	orl-rat 76mg/kg
9.	2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	orl-rat 2460mg/kg
10.	Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m3/8H)(skin)	orl-rat 120mg/kg
11.	Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	orl-rat 277mg/kg
12.	Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	orl-rat 7872mg/kg
13.	Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	orl-rat 780mg/kg
14.	2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-46-9	10 ppm (35mg/m3/8H)	orl-rat 720mg/kg
15.	Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	76-01-7	N/A	N/A
16.	1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	orl-rat 43g/kg
17.	Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	orl-rat 916mg/kg
18.	Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	orl-rat 848mg/kg
19.	cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
20.	trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	orl-rat 1235mg/kg
21.	Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	orl-rat 820mg/kg
22.	1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	orl-rat 200mg/kg
23.	Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	orl-rat 933mg/kg
24.	Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	orl-rat 2350mg/kg
25.	Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.9	20.5	67-66-3	50 ppm (240mg/m3) (CL)	orl-rat 908mg/kg
26.	Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	orl-rat 108mg/kg
27.	1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	orl-rat 725mg/kg
28.	2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
29.	Tetrachloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.6	127-18-4	25 ppm (170mg/m3/8H)(final)	orl-rat 2629mg/kg
30.	1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	orl-rat 10300mg/kg
31.	1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-8	0.001 ppm	orl-rat 170mg/kg
32.	1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	orl-rat 108mg/kg
33.	1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	orl-rat 670mg/kg
34.	1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	orl-rat 1947mg/kg
35.	1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	unr-mus 3600mg/kg
36.	1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	29.7	563-58-6	N/A	N/A
37.	cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
38.	trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.8	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
39.	Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	orl-rat 82mg/kg
40.	1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	630-20-6	N/A	orl-rat 670mg/kg
41.	1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	orl-rat 800mg/kg
42.	1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (46mg/m3/8H)(skin)	orl-rat 836mg/kg
43.	Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	orl-mus 2402mg/kg
44.	1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	orl-rat 149.6mg/kg
45.	Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	orl-rat 4894mg/kg
46.	Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-86-1	N/A	orl-rat 2699mg/kg
47.	n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
48.	Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	orl-rat >2000mg/kg
49.	p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	orl-rat 4750mg/kg
50.	Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	orl-rat 490mg/kg
51.	Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	orl-rat 5000mg/kg
52.	Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	orl-rat 5000mg/kg
53.	1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	N/A	ipr-mus 1390mg/kg
54.	1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	orl-rat 756mg/kg
55.	1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A	orl-rat 5g/kg
56.	1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-67-8	N/A	orl-rat 5000mg/kg
57.	m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg
58.	tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	98-06-6	N/A	N/A
59.	sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	22.9	135-98-8	N/A	orl-rat 2240mg/kg
60.	Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	orl-rat 2290mg/kg
61.	2-Chlorotoluene	35163	101923	0.05	5.00	40000.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	orl-rat 3900mg/kg
62.	4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	orl-rat 2100mg/kg
63.	1,2-Dichlorobenzene	35163	101923	0.05	5.00	40003.6	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	orl-rat 500mg/kg
64.	1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	ipr-mus 1062mg/kg
65.	1,4-Dichlorobenzene	35163	101923	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	106-46-7	75 ppm (450mg/m3/8H)	orl-rat 500mg/kg
66.	Isopropylbenzene	35163	101923	0.05	5.00	40000.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	98-82-8	50 ppm (245mg/m3/8H)	orl-rat 1400mg/kg
67.	n-Propylbenzene	35163	101923	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.7	23.0	103-65-1	N/A	orl-rat 6040mg/kg
68.	o-Xylene	35163	101923	0.05	5.00	40040.8	2000	NA	NA	0.017	NA	NA	2001.5	23.0	95-47-6	100 ppm (435mg/m3/8H)	ipr-mus 1364mg/kg
69.	p-Xylene	35163	101923	0.05	5.00	40000.6	2000	NA	NA	0.017	NA	NA	1999.5	22.9	106-42-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening sample, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty References: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Run 16, "P95317 L021624 [2000µg/mL in MeOH]"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.33
3	1,1-Dichloroethene	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.31
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Methacrylonitrile/Methyl acrylate/Chloroform	18.46
13	Isobutanol/1,1,1-Trichloroethane	19.91
14	1,1-Dichloropropene	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.49
17	Trichloroethene	23.58
18	1,2-Dichloropropane	24.28
19	Methyl methacrylate	24.52
20	Bromodichloromethane	25.13
21	Dibromomethane/2-Nitropropane	25.46
22	cis-1,3-Dichloropropene	27.02
23	Toluene	28.05
24	Ethyl methacrylate/trans-1,3-Dichloropropene	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethene/1,3-Dichloropropane	30.24
27	Dibromochloromethane	31.16
28	1,2-Dibromoethene	31.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.86
32	o-Xylene	35.22
33	Styrene	35.38
34	Isopropylbenzene/Bromoform	36.18
35	cis-1,4-Dichloro-2-Butene	36.80
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,3-Trichloropropane	37.77
38	n-Propylbenzene	37.93
39	trans-1,4-Dichloro-2-Butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.80
42	2-Chlorotoluene	38.82
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.53
52	1,2-Dichlorobenzene	43.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.46
55	1,2,4-Trichlorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.26
58	1,2,3-Trichlorobenzene	51.16

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	**1-352-323-3500**
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification In accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability Stable under recommended storage conditions.
Possibility of hazardous reactions Vapours may form explosive mixture with air.
Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

CERTIFIED WEIGHT REPORT

Part Number: 95317
Lot Number: 021624
Description: Universal VOA Megamix
69 components
Expiration Date: 021627
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 6UTB

Solvent(s): Methanol **Lot#** EG359-USQ12

Formulated By: Prashant Chauhan 021624 DATE

Reviewed By: Pedro L. Rentas 021624 DATE

Weight(s) shown below were combined and diluted to (mL): 100.0 5E-05 Balance Uncertainty 0.021 Flask Uncertainty


	Compound	(RM#) Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Initial Conc.(µg/mL)	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1.	Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	orl-rat 2460mg/kg
2.	Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	orl-rat 700mg/kg
3.	Carbon disulphide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	orl-rat 1200mg/kg
4.	cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1476-11-5	N/A	N/A
5.	trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	N/A	N/A
6.	Diethyl ether	(0153)	IK18CAS0000	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7.	Ethyl methacrylate	(0361)	06126PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	orl-rat 14800mg/kg
8.	Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	orl-rat 76mg/kg
9.	2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	orl-rat 2460mg/kg
10.	Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m3/8H)(skin)	orl-rat 120mg/kg
11.	Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	orl-rat 277mg/kg
12.	Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	orl-rat 7872mg/kg
13.	Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	orl-rat 780mg/kg
14.	2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-46-9	10 ppm (35mg/m3/8H)	orl-rat 720mg/kg
15.	Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	76-01-7	N/A	N/A
16.	1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	orl-rat 43g/kg
17.	Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	orl-rat 916mg/kg
18.	Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	orl-rat 848mg/kg
19.	cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
20.	trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	orl-rat 1235mg/kg
21.	Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	orl-rat 820mg/kg
22.	1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	orl-rat 200mg/kg
23.	Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	orl-rat 933mg/kg
24.	Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	orl-rat 2350mg/kg
25.	Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.9	20.5	67-66-3	50 ppm (240mg/m3) (CL)	orl-rat 908mg/kg
26.	Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	orl-rat 108mg/kg
27.	1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	orl-rat 725mg/kg
28.	2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
29.	Tetrachloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.6	127-18-4	25 ppm (170mg/m3/8H)(final)	orl-rat 2629mg/kg
30.	1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	orl-rat 10300mg/kg
31.	1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-8	0.001 ppm	orl-rat 170mg/kg
32.	1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	orl-rat 108mg/kg
33.	1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	orl-rat 670mg/kg
34.	1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	orl-rat 1947mg/kg
35.	1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	unr-mus 3600mg/kg
36.	1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	29.7	563-58-6	N/A	N/A
37.	cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
38.	trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.8	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
39.	Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	orl-rat 82mg/kg
40.	1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	630-20-6	N/A	orl-rat 670mg/kg
41.	1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	orl-rat 800mg/kg
42.	1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (46mg/m3/8H)(skin)	orl-rat 836mg/kg
43.	Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	orl-mus 2402mg/kg
44.	1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	orl-rat 149.6mg/kg
45.	Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	orl-rat 4894mg/kg
46.	Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-86-1	N/A	orl-rat 2699mg/kg
47.	n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
48.	Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	orl-rat >2000mg/kg
49.	p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	orl-rat 4750mg/kg
50.	Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	orl-rat 490mg/kg
51.	Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	orl-rat 5000mg/kg
52.	Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	orl-rat 5000mg/kg
53.	1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	N/A	ipr-mus 1390mg/kg
54.	1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	orl-rat 756mg/kg
55.	1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A	orl-rat 5g/kg
56.	1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-67-8	N/A	orl-rat 5000mg/kg
57.	m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg
58.	tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	98-06-6	N/A	N/A
59.	sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	22.9	135-98-8	N/A	orl-rat 2240mg/kg
60.	Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	orl-rat 2290mg/kg
61.	2-Chlorotoluene	35163	101923	0.05	5.00	40000.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	orl-rat 3900mg/kg
62.	4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	orl-rat 2100mg/kg
63.	1,2-Dichlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	orl-rat 500mg/kg
64.	1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	ipr-mus 1062mg/kg
65.	1,4-Dichlorobenzene	35163	101923	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	106-46-7	75 ppm (450mg/m3/8H)	orl-rat 500mg/kg
66.	Isopropylbenzene	35163	101923	0.05	5.00	40000.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	98-82-8	50 ppm (245mg/m3/8H)	orl-rat 1400mg/kg
67.	n-Propylbenzene	35163	101923	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.7	23.0	103-65-1	N/A	orl-rat 6040mg/kg
68.	o-Xylene	35163	101923	0.05	5.00	40040.8	2000	NA	NA	0.017	NA	NA	2001.5	23.0	95-47-6	100 ppm (435mg/m3/8H)	ipr-mus 1364mg/kg
69.	p-Xylene	35163	101923	0.05	5.00	40000.6	2000	NA	NA	0.017	NA	NA	1999.5	22.9	106-42-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening sample, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty References: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Run 16, "P95317 L021624 [2000µg/mL in MeOH]"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.33
3	1,1-Dichloroethene	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.31
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Methacrylonitrile/Methyl acrylate/Chloroform	18.46
13	Isobutanol/1,1,1-Trichloroethane	19.91
14	1,1-Dichloropropene	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.49
17	Trichloroethene	23.58
18	1,2-Dichloropropane	24.28
19	Methyl methacrylate	24.52
20	Bromodichloromethane	25.13
21	Dibromomethane/2-Nitropropane	25.46
22	cis-1,3-Dichloropropene	27.02
23	Toluene	28.05
24	Ethyl methacrylate/trans-1,3-Dichloropropene	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethene/1,3-Dichloropropane	30.24
27	Dibromochloromethane	31.16
28	1,2-Dibromoethene	31.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.86
32	o-Xylene	35.22
33	Styrene	35.38
34	Isopropylbenzene/Bromoform	36.18
35	cis-1,4-Dichloro-2-Butene	36.80
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,3-Trichloropropane	37.77
38	n-Propylbenzene	37.93
39	trans-1,4-Dichloro-2-Butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.80
42	2-Chlorotoluene	38.82
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.53
52	1,2-Dichlorobenzene	43.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.46
55	1,2,4-Trichlorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.26
58	1,2,3-Trichlorobenzene	51.16

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	**1-352-323-3500**
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification In accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability Stable under recommended storage conditions.
Possibility of hazardous reactions Vapours may form explosive mixture with air.
Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 31491
Lot Number: 040821
Description: 1,2,4-Trimethylbenzene


Solvent(s): Methanol **Lot#** DY186-US


Expiration Date: 040826
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 6UTB 5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 50.0 0.057 Flask Uncertainty

Formulated By:	Prashant Chauhan	040821 DATE
Reviewed By:	Pedro L. Rentas	040821 DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 1,2,4-Trimethylbenzene	475	WXBC9778V	2000	98.8	0.2	0.10127	0.10140	2002.5	9.5	95-63-6	N/A	orl-rat 5g/kg

Method GC6MSD-1: Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Temp. 1 = 35°C (10min.), Temp. 2 = 200°C (8.75 min.), Rate = 4°C/min., Injector Temp.= 200°C, Detector Temp. = 220°C. Analysis performed by Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

CERTIFIED WEIGHT REPORT

Part Number: 94559
Lot Number: 051421
Description: 1,3,5-Trichlorobenzene

Expiration Date: 051426
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 6UTB

Solvent(s): **Lot#**
Methanol DY186-US


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 100.0 0.012 Flask Uncertainty

		051421
Formulated By:	Benson Chan	DATE
		051421
Reviewed By:	Pedro L. Rentas	DATE


Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 1,3,5-Trichlorobenzene	409	STBH8643	2000	99.9	0.2	0.20021	0.20084	2006.3	8.1	108-70-3	N/A	orl-rat 800mg/kg

Method GC6MSD-1: Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Temp. 1=35°C (10min.), Temp. 2=200°C (8.75 min.), Rate=4°C/min., Injector Temp.=200°C, Detector Temp.=220°C. Analysis performed by Candice Warren.

TIC: [BSB3]70409.D



Scan 3238 (53.133 min): [BSB3]70409.D



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International Date Prepared/Revised	**1-352-323-3500** January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability Stable under recommended storage conditions.
Possibility of hazardous reactions Vapours may form explosive mixture with air.
Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredite
AR-1539 Certificate Numbe
https://Absolutestandards.cor

CERTIFIED WEIGHT REPORT

Part Number: 90319
Lot Number: 061923
Description: 1,3,5-Trimethylbenzene [Mesitylene]
Expiration Date: 061928
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 6UTB


Solvent:	Lot#
Methanol	EF282-US


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 50.0 0.001 Flask Uncertainty

Gabriel Helland		061923
Formulated By:	Gabriel Helland	DATE
Pedro Rentas		061923
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity (%)	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 1,3,5-Trimethylbenzene	301	TOOOF-IC	2000	97	0.2	0.10315	0.10341	2004.9	8.5	108-67-8	N/A	orl-rat 5000mg/kg

Method GC6MSD-1: Column: Vocol 60m X 0.25mm ID X 1.5µm film thickness). Temp. 1 = 35°C (10min.), Temp. 2 = 200°C (8.75 min.), Rate = 4°C/min., Injector Temp.= 200°C, Detector Temp. = 220°C. Analysis performed by Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 09/16/24 2 Vials

ANAB ISO 17034 Accredited
AR-1539 Certificate Numb
https://Absolutestandards.co...

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 091424
Description: Acrolein

Solvent(s): Water **Lot#** 072324Q


Expiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB


Weight(s) shown below were combined and diluted to (mL): 10.0 | 5E-05 Balance Uncertainty | 0.001 Flask Uncertainty

Justin Dippold	091424
Formulated By: Justin Dippold	DATE
Pedro Rentas	091424
Reviewed By: Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 09/16/24 2 Vials

ANAB ISO 17034 Accredited
AR-1539 Certificate Numb
https://Absolutestandards.co...

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 091424
Description: Acrolein

Expiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

Solvent(s): **Lot#**
Water 072324Q


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty


Justin Dippold 091424
Formulated By: Justin Dippold DATE

Pedro Rentas 091424
Reviewed By: Pedro L. Rentas DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	orl-rat 46mg/kg

Method: GC6MSD-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. **Analyst:** Pedro Rentas. **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately Long term storage is not recommended. Please contact our technical department if further information is required.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 12/6/24
20 vial
V14630 to
V14649

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB

Solvent(s): Methanol **Lot#** EJ143-US


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 50.0 0.001 Flask Uncertainty


Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE

Pedro Rentas 120524
Reviewed By: Pedro L. Rentas DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp.= 200°C, Detector B Temp. = 220°C. **Analyst:** Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International Date Prepared/Revised	**1-352-323-3500** January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability Stable under recommended storage conditions.
Possibility of hazardous reactions Vapours may form explosive mixture with air.
Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 12/6/24
20 vial
V14630 to
V14649

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB


Solvent(s): Methanol **Lot#** EJ143-US


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 50.0 0.001 Flask Uncertainty

Prashant Chauhan	120524
Formulated By: Prashant Chauhan	DATE
Pedro Rentas	120524
Reviewed By: Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp.= 200°C, Detector B Temp. = 220°C. **Analyst:** Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.

PO Box 5585
Hamden, CT 06518-0585

Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International Date Prepared/Revised	**1-352-323-3500** January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability — Stable under recommended storage conditions.
Possibility of hazardous reactions — Vapours may form explosive mixture with air.
Conditions to avoid — Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid — Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Exp 12/6/24
20 vial
V14630 to
V14649

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB

Solvent(s): Methanol **Lot#** EJ143-US


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 50.0 0.001 Flask Uncertainty


Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE

Pedro Rentas 120524
Reviewed By: Pedro L. Rentas DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp.= 200°C, Detector B Temp. = 220°C. **Analyst:** Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International Date Prepared/Revised	**1-352-323-3500** January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability — Stable under recommended storage conditions.
Possibility of hazardous reactions — Vapours may form explosive mixture with air.
Conditions to avoid — Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid — Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec 12/6/24
20 vial
V14630 to
V14649

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB

Solvent(s): Methanol **Lot#** EJ143-US


5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 50.0 0.001 Flask Uncertainty


Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE

Pedro Rentas 120524
Reviewed By: Pedro L. Rentas DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp.= 200°C, Detector B Temp. = 220°C. **Analyst:** Candice Warren.





- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards Inc.
PO Box 5585
Hamden, CT 06518-0585
Phone: 203-281-2917
FAX: 203-281-2922

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY **ANALYTICAL STANDARD DISSOLVED IN METHANOL**

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	**1-800-535-5053**
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International Date Prepared/Revised	**1-352-323-3500** January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	**Highly Flammable Liquid and Vapor**	**H301, 311, 331**	**Toxic if swallowed, skin contact, inhaled**
H370	**Cause damage to organs**	**H351**	**Suspected of causing cancer**
P271	**Use in ventilated area**	**P280**	**Use gloves, eye protection/face sheild**
P302,332	**If on skin, wash with soap and water**	**P305,351,338**	**If in eyes, remove contacts, rinse with water**



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance.Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol 67-56-1 TWA 200 ppm
Skin notation TWA 200 ppm
Potential for skin absorption , ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H2O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6

Solubility in Water COMPLETE

Appearance and Odor CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.

Section X. STABILITY AND REACTIVITY

Chemical stability — Stable under recommended storage conditions.
Possibility of hazardous reactions — Vapours may form explosive mixture with air.
Conditions to avoid — Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid — Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions. - Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50	15,400 mg/l - 96 h
EC50	24,500.00 mg/l - 48 h
EC100	10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30429 **Lot No.:** A0188973

Description : 1,2,3-Trichloropropane Standard

1,2,3-Trichloropropane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	1,2,3-Trichloropropane **CAS #** 96-18-4 **Purity** 99% (Lot 332900)	2,000.0 µg/mL	+/- 11.7371 µg/mL +/- 112.1494 µg/mL +/- 114.7730 µg/mL		Gravimetric Unstressed Stressed

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)


Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Daniel Wasson - Operations Tech I **Date Mixed:** 24-Aug-2022 **Balance:** 1127510105

Christie Mills - Operations Tech II - ARM QC **Date Passed:** 29-Aug-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k\sqrt{U^2_{gravimetric} + U^2_{homogeneity} + U^2_{storage\ stability} + U^2_{shipping\ stability}}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK®
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 **Lot No.:** A0181905

Description : tert-Butanol Standard

tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2025 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)		Expanded Uncertainty (95% C.L.; K=2)		
1	tert-Butanol (TBA)		50,126.0	µg/mL	+/- 293.4988	µg/mL	Gravimetric
	CAS # 75-65-0	(Lot SHBM7694)			+/- 1,073.7654	µg/mL	Unstressed
	Purity 99%				+/- 1,104.9494	µg/mL	Stressed

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)


Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 16-Feb-2022 **Balance:** B442140311

Merlina Cowan - Operations Tech I

Date Passed: 21-Feb-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k\sqrt{U^2_{gravimetric} + U^2_{homogeneity} + U^2_{storage\ stability} + U^2_{shipping\ stability}}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022 **Balance Serial #** B251644995

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 **Balance Serial #** B707717271

Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 **Balance Serial #** B707717271

Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 **Balance Serial #** B707717271

Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 **Balance Serial #** B707717271

Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0196865

Description : Custom 8260A/B Surrogate Mix

Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 µg/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 µg/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 µg/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 µg/mL	+/- 1,417.4648

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Apr-2023 **Balance:** 1127510105

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods . Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024 **Balance Serial #** B707717271

Dillan Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods . Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024 **Balance Serial #** B707717271

Dillan Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30091 **Lot No.:** A0209905

Description : L/C VOA Internal Standard Mix

L/C Internal Std 2500µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2029 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	2,508.0 µg/mL	+/- 142.0596
2	Chlorobenzene-d5	3114-55-4	PR-31132	99%	2,512.0 µg/mL	+/- 142.2862
3	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,512.0 µg/mL	+/- 142.2862

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 05-Apr-2024 **Balance Serial #** 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 08-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024 **Balance:** 1127510105

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 V.4

Certificate of Analysis

chromatographic plus

V14697-to-14726

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 V.4

Certificate of Analysis

chromatographic plus

V14697-to-14726

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 V.4

Certificate of Analysis

chromatographic plus

V14697-to-14726

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 V.4

Certificate of Analysis

chromatographic plus

V14697-to-14726

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30624 **Lot No.:** A0211457

Description : SOM 01.1 VOA DMC Non-Ketones Standard

SOM 01.1 VOA DMC Non-Ketones Standard 500µg/mL, Methanol-OD, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl Chloride-d3	6745-35-3	PR-26294	99%	515.3 µg/mL	+/- 42.5161
2	Chloroethane-d5	19199-91-8	PR-19060	99%	498.2 µg/mL	+/- 40.0866
3	1,1-Dichloroethylene-d2	22280-73-5	PR-21050	99%	503.0 µg/mL	+/- 28.2630
4	Chloroform-d	865-49-6	A0219685001	99%	503.0 µg/mL	+/- 28.2630
5	1,2-Dichloroethane-d4	17060-07-0	PR-33313	99%	503.0 µg/mL	+/- 28.2630
6	Benzene-d6	1076-43-3	PR-33510	99%	501.0 µg/mL	+/- 28.1506
7	1,2-Dichloropropane-d6	93952-08-0	Z-322	99%	503.0 µg/mL	+/- 28.2630
8	1,3-Dichloropropene-d4 (cis/ trans mixture) 58% cis Isomer; 42% trans Isomer	202656-23-3	Z-181	99%	504.0 µg/mL	+/- 28.3192
9	Toluene-d8	2037-26-5	PR-34141	99%	503.0 µg/mL	+/- 28.2630
10	1,1,2,2-Tetrachloroethane-d2	33685-54-0	F465P1	99%	502.0 µg/mL	+/- 28.2068
11	1,2-Dichlorobenzene-d4	2199-69-1	PR-32597	99%	503.0 µg/mL	+/- 28.2630

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol-OD
CAS # 1455-13-6
Purity 99%

Quality Confirmation Test

Column:
60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:
helium-constant pressure 30 psi

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
MSD

Split Vent:
25.0 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 15-May-2024 **Balance Serial #** 1128342314

Dillan Murphy - Operations Technician I

Date Passed: 17-May-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30625 **Lot No.:** A0216280

Description : OLC 3.2 VOA Deuterated Monitoring Compounds

OLC 3.2 VOA Ketone Deuterated Monitoring Compounds 500µg/mL, Deuterium Oxide, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Butanone-d5	24313-50-6	M-276	99%	504.0 µg/mL	+/- 17.5357
2	2-Hexanone-d5	4840-82-8	GH-242	99%	502.0 µg/mL	+/- 17.4661

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Deuterium oxide
CAS # 7789-20-0
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Richard Zimmerman - Operations Tech I

Date Mixed: 10-Sep-2024 **Balance Serial #** B251644995

Dillan Murphy - Operations Technician I

Date Passed: 12-Sep-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30625 **Lot No.:** A0216280

Description : OLC 3.2 VOA Deuterated Monitoring Compounds

OLC 3.2 VOA Ketone Deuterated Monitoring Compounds 500µg/mL, Deuterium Oxide, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Butanone-d5	24313-50-6	M-276	99%	504.0 µg/mL	+/- 17.5357
2	2-Hexanone-d5	4840-82-8	GH-242	99%	502.0 µg/mL	+/- 17.4661

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Deuterium oxide
CAS # 7789-20-0
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Richard Zimmerman - Operations Tech I

Date Mixed: 10-Sep-2024 **Balance Serial #** B251644995

Dillan Murphy - Operations Technician I

Date Passed: 12-Sep-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 Vial

Certificate of Analysis

chromatographic plus

V14727 to
V14756

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826

Description : 502.2 Calibration Mix #1

502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2031 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:
helium-constant flow 2.0 mL/min.

Temp. Program:
40°C (hold 6 min.) to 100°C
@ 6°C/min.


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
MSD

Split Vent:
Split ratio 10:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician **Date Mixed:** 23-Sep-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC **Date Passed:** 04-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 Vial

Certificate of Analysis

chromatographic plus

V14727 to
V14756

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826

Description : 502.2 Calibration Mix #1

502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2031 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:
helium-constant flow 2.0 mL/min.

Temp. Program:
40°C (hold 6 min.) to 100°C
@ 6°C/min.


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
MSD

Split Vent:
Split ratio 10:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician **Date Mixed:** 23-Sep-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC **Date Passed:** 04-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 Vial

Certificate of Analysis

chromatographic plus

V14727 to
V14756

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826

Description : 502.2 Calibration Mix #1

502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2031 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:
helium-constant flow 2.0 mL/min.

Temp. Program:
40°C (hold 6 min.) to 100°C
@ 6°C/min.


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
MSD

Split Vent:
Split ratio 10:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician **Date Mixed:** 23-Sep-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC **Date Passed:** 04-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30625 **Lot No.:** A0219189

Description : OLC 3.2 VOA Deuterated Monitoring Compounds

OLC 3.2 VOA Ketone Deuterated Monitoring Compounds 500µg/mL, Deuterium Oxide, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Butanone-d5	24313-50-6	HJ-279	99%	504.0 µg/mL	+/- 17.5357
2	2-Hexanone-d5	4840-82-8	I-500	99%	504.0 µg/mL	+/- 17.5357

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Deuterium oxide
CAS # 7789-20-0
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Aaron Enyart - Operations Tech I

Date Mixed: 15-Nov-2024 **Balance Serial #** B345965662

Dillan Murphy - Operations Technician I

Date Passed: 19-Nov-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30625 **Lot No.:** A0219189

Description : OLC 3.2 VOA Deuterated Monitoring Compounds

OLC 3.2 VOA Ketone Deuterated Monitoring Compounds 500µg/mL, Deuterium Oxide, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Butanone-d5	24313-50-6	HJ-279	99%	504.0 µg/mL	+/- 17.5357
2	2-Hexanone-d5	4840-82-8	I-500	99%	504.0 µg/mL	+/- 17.5357

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Deuterium oxide
CAS # 7789-20-0
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Aaron Enyart - Operations Tech I

Date Mixed: 15-Nov-2024 **Balance Serial #** B345965662

Dillan Murphy - Operations Technician I

Date Passed: 19-Nov-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Vinyl Rec 01/08/21

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471

Description : Custom Vinyl Acetate Standard

Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2026 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods . Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024 **Balance Serial #** 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

10 vial Rec 01/08/25

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14793 to 14802

ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-FL **Lot No.:** A0220563

Description : Custom Vinyl Acetate Standard

Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2026 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,060.0 µg/mL	+/- 278.5905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods . Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 30-Dec-2024 **Balance Serial #** B345965662

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

James T Ethier
Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

James T Ethier
Jamie Ethier
Vice President Global Quality

SHIPPING DOCUMENTS

CHEMTECH
CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. Q1215
QUOTE NO.
COC Number 2041067

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: RU2 Engineering LLC
ADDRESS: 2 Melinda Drive, Monroe Twp, NJ 08831
CITY: ZIP:
ATTENTION: Rutu Manani
PHONE: 609-409-4564 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: SANDTWOBR BMCR Project
PROJECT NO.: LOCATION: Brooklyn, NYC
PROJECT MANAGER: Rutu Manani
e-mail: Rmanani@Ru2eng.com
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: Same as Company address PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

DATA TURNAROUND INFORMATION

FAX (RUSH) Standard 10 days DAYS*
HARDCOPY (DATA PACKAGE): Standard 10 days DAYS*
EDD: Standard 10 days DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

- ☐ Level 1 (Results Only)
- ☐ Level 2 (Results + QC)
- ☐ Level 3 (Results + QC + Raw Data)
- ☐ Level 4 (QC + Full Raw Data)
- ☐ NJ Reduced
- ☐ NYS ASP A
- ☒ NYS ASP B
- ☐ Other
- ☐ EDD FORMAT

ANALYSIS

1 TCL VOC + TICs MTBE, TBA
2 TCLP VOCs
3 TPH GRO-DRO
4 TCL SVOCs + TICs
5 TAL Metals
6 Pesticides PCBs
7 RCRA characteristics
8 Paint filter
9 full TCLP

PRESERVATIVES

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HN03 ICE
C-H2SO4 F-OTHER

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	1	2	3	4	5	6	7	8	9
1.	JPP-29.1-012825	Soil		G	1/28/25	14:26	3	X	X	X						
2.	JPP-29.1-012825	Soil	C		1/28/25	14:30	7			X	X	X	X	X	X	X
3.	JPP-29.2-012825	Soil		G	1/28/25	15:15	3	X	X	X						
4.	JPP-29.2-012825	Soil	C		1/28/25	15:22	7			X	X	X	X	X	X	X
5.																
6.																
7.																
8.																
9.																
10.																

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
1. RA	1/29/2025	1. 1045 1-29-25
2.		2.
3.	1614 1-29-25	3.

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.8 °C
Comments: Preserve extra sample jar if additional analysis is Required

Page 2 of 2

CLIENT: ☐ Hand Delivered ☐ Other
CHEMTECH: ☐ Picked Up ☐ Field Sampling

Shipment Complete ☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY

CHEMTECH 284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1215 RUTW01
Order Date : 1/29/2025 11:20:00 AM
YG
02/03/25
Project Mgr :
Client Name : RU2 Engineering, LLC
Project Name : ~~SANDTWOBR BMCR Bro~~
NYCDDC SANTWOBR Brooklyn Bridge BBMCR
Report Type : NYS ASP B
Client Contact : Rutu Manani
Receive DateTime : 1/29/2025 4:14:00 PM
EDD Type : Excel NY
Invoice Name : RU2 Engineering, LLC
Purchase Order :
Hard Copy Date :
Invoice Contact : Rutu Manani
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1215-01	JPP-29.1-012825	Solid	01/28/2025	14:26					
					VOCMS Group1		8260D		10 Bus. Days
Q1215-05	JPP-29.2-012825	Solid	01/28/2025	15:15					
					VOCMS Group1		8260D		10 Bus. Days

Relinguished By :
Date / Time : 1-29-25 1700

Received By :
Date / Time : 1-29-25 17:00

Samples in Sm Frig @ 1700.

Storage Area : VOA Refridgerator Room

left stored rc