



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	RU2 Engineering, LLC				SDG No.:	Q1215			
Contract:	RUTW01	Lab Code:	CHEM		Case No.:	Q1215	SAS No.:	Q1215	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB81	Mercury	0.20	+/-0.20	U	0.20	CV	02/03/2025	16:01	LB134537

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1215

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1215

SAS No.: Q1215

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB10	Mercury	0.20	+/-0.20	U	0.20	CV	02/03/2025	16:06	LB134537
CCB11	Mercury	0.20	+/-0.20	U	0.20	CV	02/03/2025	16:36	LB134537
CCB12	Mercury	0.20	+/-0.20	U	0.20	CV	02/03/2025	17:03	LB134537
CCB13	Mercury	0.20	+/-0.20	U	0.20	CV	02/03/2025	17:26	LB134537

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1215

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1215

SAS No.: Q1215

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	02/10/2025	14:26	LB134659
	Barium	100	+/-100	U	100	P	02/10/2025	14:26	LB134659
	Cadmium	6.00	+/-6.00	U	6.00	P	02/10/2025	14:26	LB134659
	Chromium	10.0	+/-10.0	U	10.0	P	02/10/2025	14:26	LB134659
	Lead	12.0	+/-12.0	U	12.0	P	02/10/2025	14:26	LB134659
	Selenium	20.0	+/-20.0	U	20.0	P	02/10/2025	14:26	LB134659
	Silver	10.0	+/-10.0	U	10.0	P	02/10/2025	14:26	LB134659

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1215

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1215

SAS No.: Q1215

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	02/10/2025	15:03	LB134659
	Barium	100	+/-100	U	100	P	02/10/2025	15:03	LB134659
	Cadmium	6.00	+/-6.00	U	6.00	P	02/10/2025	15:03	LB134659
	Chromium	10.0	+/-10.0	U	10.0	P	02/10/2025	15:03	LB134659
	Lead	12.0	+/-12.0	U	12.0	P	02/10/2025	15:03	LB134659
	Selenium	20.0	+/-20.0	U	20.0	P	02/10/2025	15:03	LB134659
	Silver	10.0	+/-10.0	U	10.0	P	02/10/2025	15:03	LB134659
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	02/10/2025	16:08	LB134659
	Barium	100	+/-100	U	100	P	02/10/2025	16:08	LB134659
	Cadmium	6.00	+/-6.00	U	6.00	P	02/10/2025	16:08	LB134659
	Chromium	10.0	+/-10.0	U	10.0	P	02/10/2025	16:08	LB134659
	Lead	12.0	+/-12.0	U	12.0	P	02/10/2025	16:08	LB134659
	Selenium	20.0	+/-20.0	U	20.0	P	02/10/2025	16:08	LB134659
	Silver	10.0	+/-10.0	U	10.0	P	02/10/2025	16:08	LB134659
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	02/10/2025	17:06	LB134659
	Barium	100	+/-100	U	100	P	02/10/2025	17:06	LB134659
	Cadmium	6.00	+/-6.00	U	6.00	P	02/10/2025	17:06	LB134659
	Chromium	10.0	+/-10.0	U	10.0	P	02/10/2025	17:06	LB134659
	Lead	12.0	+/-12.0	U	12.0	P	02/10/2025	17:06	LB134659
	Selenium	20.0	+/-20.0	U	20.0	P	02/10/2025	17:06	LB134659
	Silver	10.0	+/-10.0	U	10.0	P	02/10/2025	17:06	LB134659
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	02/10/2025	18:00	LB134659
	Barium	100	+/-100	U	100	P	02/10/2025	18:00	LB134659
	Cadmium	6.00	+/-6.00	U	6.00	P	02/10/2025	18:00	LB134659
	Chromium	10.0	+/-10.0	U	10.0	P	02/10/2025	18:00	LB134659
	Lead	12.0	+/-12.0	U	12.0	P	02/10/2025	18:00	LB134659
	Selenium	20.0	+/-20.0	U	20.0	P	02/10/2025	18:00	LB134659
	Silver	10.0	+/-10.0	U	10.0	P	02/10/2025	18:00	LB134659
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	02/10/2025	18:52	LB134659
	Barium	100	+/-100	U	100	P	02/10/2025	18:52	LB134659
	Cadmium	6.00	+/-6.00	U	6.00	P	02/10/2025	18:52	LB134659
	Chromium	10.0	+/-10.0	U	10.0	P	02/10/2025	18:52	LB134659
	Lead	12.0	+/-12.0	U	12.0	P	02/10/2025	18:52	LB134659
	Selenium	20.0	+/-20.0	U	20.0	P	02/10/2025	18:52	LB134659
	Silver	10.0	+/-10.0	U	10.0	P	02/10/2025	18:52	LB134659
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	02/10/2025	19:45	LB134659
	Barium	100	+/-100	U	100	P	02/10/2025	19:45	LB134659
	Cadmium	6.00	+/-6.00	U	6.00	P	02/10/2025	19:45	LB134659
	Chromium	10.0	+/-10.0	U	10.0	P	02/10/2025	19:45	LB134659

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>RU2 Engineering, LLC</u>		SDG No.: <u>Q1215</u>							
Contract: <u>RUTW01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1215</u>	SAS No.: <u>Q1215</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	02/10/2025	19:45	LB134659
	Selenium	20.0	+/-20.0	U	20.0	P	02/10/2025	19:45	LB134659
	Silver	10.0	+/-10.0	U	10.0	P	02/10/2025	19:45	LB134659
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	02/10/2025	20:41	LB134659
	Barium	100	+/-100	U	100	P	02/10/2025	20:41	LB134659
	Cadmium	6.00	+/-6.00	U	6.00	P	02/10/2025	20:41	LB134659
	Chromium	10.0	+/-10.0	U	10.0	P	02/10/2025	20:41	LB134659
	Lead	12.0	+/-12.0	U	12.0	P	02/10/2025	20:41	LB134659
	Selenium	20.0	+/-20.0	U	20.0	P	02/10/2025	20:41	LB134659
	Silver	10.0	+/-10.0	U	10.0	P	02/10/2025	20:41	LB134659
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	02/10/2025	21:06	LB134659
	Barium	100	+/-100	U	100	P	02/10/2025	21:06	LB134659
	Cadmium	6.00	+/-6.00	U	6.00	P	02/10/2025	21:06	LB134659
	Chromium	10.0	+/-10.0	U	10.0	P	02/10/2025	21:06	LB134659
	Lead	12.0	+/-12.0	U	12.0	P	02/10/2025	21:06	LB134659
	Selenium	20.0	+/-20.0	U	20.0	P	02/10/2025	21:06	LB134659
	Silver	10.0	+/-10.0	U	10.0	P	02/10/2025	21:06	LB134659