

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>RU2 Engineering, LLC</u>	SDG No.: <u>Q1215</u>
Contract: <u>RUTW01</u>	Lab Code: <u>CHEM</u>
ICS Source: <u>EPA</u>	Case No.: <u>Q1215</u>
	SAS No.: <u>Q1215</u>
	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	1.27			-20	20	02/10/2025	14:35	LB134659
	Barium	1.84	6.0	31	-94	106	02/10/2025	14:35	LB134659
	Cadmium	3.51	1.0	351	-5	7	02/10/2025	14:35	LB134659
	Chromium	56.7	52.0	109	42	62	02/10/2025	14:35	LB134659
	Lead	8.37			-12	12	02/10/2025	14:35	LB134659
	Selenium	-6.12			-20	20	02/10/2025	14:35	LB134659
	Silver	2.79			-10	10	02/10/2025	14:35	LB134659
ICSAB01	Arsenic	106	104	102	88.4	120	02/10/2025	14:41	LB134659
	Barium	442	537	82	437	637	02/10/2025	14:41	LB134659
	Cadmium	963	972	99	826	1120	02/10/2025	14:41	LB134659
	Chromium	556	542	103	460	624	02/10/2025	14:41	LB134659
	Lead	55.7	49.0	114	37	61	02/10/2025	14:41	LB134659
	Selenium	44.9	46.0	98	26	66	02/10/2025	14:41	LB134659
	Silver	216	201	108	170	232	02/10/2025	14:41	LB134659