

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: RUTW01
ProjectID: NYCDDC SANTWOBR Brooklyn Bridge BBMCR
Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG No.: Q1215
DataFile: FG015247.D Analyst Name: YP\AJ Analyst Date: 01-30-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	69869948	139740	140262	0.372