

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1215 **SAS No.:** Q1215 **SDG NO.:** Q1215

Instrument ID: ECD_O **Calibration Date(s):** 01/21/2025 01/22/2025
Calibration Times: 17:36 01:50

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO108982.D	CF 750 =	PO108983.D			
CF 500 =		PO108984.D	CF 250 =	PO108985.D	CF 050 =	PO108986.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	233065045	244410019	248250902	263810212	271896440	252286524	6
Aroclor-1016-2	(2)	325363679	338972765	344308556	357996712	357782160	344884774	4
Aroclor-1016-3	(3)	223049536	234801713	240773660	257313424	264644780	244116623	7
Aroclor-1016-4	(4)	175974311	184579665	189100190	198987084	205504880	190829226	6
Aroclor-1016-5	(5)	190134930	199311205	207066494	222947156	224348000	208761557	7
Aroclor-1260-1	(1)	348800771	364882765	375462458	399932624	416699380	381155600	7
Aroclor-1260-2	(2)	430525248	448276513	461774946	488220872	519105560	469580628	7
Aroclor-1260-3	(3)	359859412	378331397	386511394	408169100	424938760	391562013	7
Aroclor-1260-4	(4)	332481764	347292652	354683668	372951676	382377400	357957432	6
Aroclor-1260-5	(5)	803151986	830038244	833771178	853559688	852457760	834595771	2
Decachlorobiphenyl		6434348190	6661231787	6830111140	7188416880	7524624800	6927746559	6
Tetrachloro-m-xylene		7601276640	7501278947	7481675880	7644109480	7553803600	7556428909	1
Aroclor-1242-1	(1)	195196952	201389908	214304986	226003920	225308640	212440881	7
Aroclor-1242-2	(2)	268098195	278641713	289244872	304071824	293574460	286726213	5
Aroclor-1242-3	(3)	186201466	194711956	206440658	220830688	215362160	204709386	7
Aroclor-1242-4	(4)	146339247	152438169	160415562	170237596	164916200	158869355	6
Aroclor-1242-5	(5)	154300112	158580243	167963632	178401740	175909460	167031037	6
Decachlorobiphenyl		6093757170	6273163067	6555969800	6873050240	6988220200	6556832095	6
Tetrachloro-m-xylene		7302831450	7496516067	7696349800	7451780120	6586958200	7306887127	6
Aroclor-1248-1	(1)	145246997	150453889	157781164	170519664	173855120	159571367	8
Aroclor-1248-2	(2)	197265768	209686637	218963792	239048120	252482700	223489403	10
Aroclor-1248-3	(3)	248188978	257855635	270690240	293270108	299294900	273859972	8
Aroclor-1248-4	(4)	351583188	362624208	379215116	402583304	426725120	384546187	8
Aroclor-1248-5	(5)	245666962	251879323	262640736	279614828	288543680	265669106	7
Decachlorobiphenyl		6072639940	6272001520	6560139460	6944778840	7129210800	6595754112	7
Tetrachloro-m-xylene		7184593920	7370395840	7484351680	7653720840	7312573400	7401127136	2
Aroclor-1254-1	(1)	374847163	389452248	410697836	433454068	459205320	413531327	8
Aroclor-1254-2	(2)	325342531	339278107	358066768	381289080	405152620	361825821	9
Aroclor-1254-3	(3)	522141746	540493293	562618572	587400960	607950820	564121078	6
Aroclor-1254-4	(4)	328382284	332039301	348673262	361037032	352970440	344620464	4
Aroclor-1254-5	(5)	475331089	489980032	515035558	537655848	547482260	513096957	6
Decachlorobiphenyl		6154630290	6360969147	6655043100	6960677720	7204237800	6667111611	6
Tetrachloro-m-xylene		7266635340	7453080907	7677404800	7686251640	7337350200	7484144577	3
Aroclor-1268-1	(1)	1033690630	1036942464	1067670602	1080273708	1104982020	1064711885	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	952087675	951155156	979561020	979388044	983362060	969110791	2
Aroclor-1268-3	(3)	786406202	784932265	807859298	814292396	834332900	805564612	3
Aroclor-1268-4	(4)	327369336	328395845	341390382	351699060	363994160	342569757	5
Aroclor-1268-5	(5)	2419252053	2385425329	2430042174	2398845144	2323360200	2391384980	2
Decachlorobiphenyl		11036128070	11026162653	11389192500	11664983320	12135593800	11450412069	4
Tetrachloro-m-xylene		7546638300	7604080347	7812040940	7844017680	7509481400	7663251733	2

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM **Case No.:** Q1215 **SAS No.:** Q1215 **SDG NO.:** Q1215

Instrument ID: ECD_O **Calibration Date(s):** 01/21/2025 01/22/2025
Calibration Times: 17:36 01:50

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO108982.D</u>		CF 750 = <u>PO108983.D</u>			
CF 500 = <u>PO108984.D</u>		CF 250 = <u>PO108985.D</u>		CF 050 = <u>PO108986.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	145518803	142573049	166649690	172900172	181948120	11
Aroclor-1016-2	(2)	223129267	236622256	232122136	250011316	249296120	5
Aroclor-1016-3	(3)	121270732	126119179	128693642	138175620	137629920	6
Aroclor-1016-4	(4)	98703058	105288716	109289440	120650588	118660940	8
Aroclor-1016-5	(5)	130939362	136319948	141694888	155164716	153595760	7
Aroclor-1260-1	(1)	229857010	239931019	247217984	265407312	277830920	8
Aroclor-1260-2	(2)	274616892	283537992	291890314	314812824	339164440	9
Aroclor-1260-3	(3)	255547157	264550967	272530098	288370904	310931080	8
Aroclor-1260-4	(4)	210034482	216033343	222462830	233977972	245747280	6
Aroclor-1260-5	(5)	481178354	492665984	499979908	514637484	514938920	3
Decachlorobiphenyl		3156544800	3289680693	3391332080	3623383520	3732817400	7
Tetrachloro-m-xylene		5315746230	5437498093	5428207440	5462079080	5158066000	2
Aroclor-1242-1	(1)	130294727	135534199	142873858	151389892	154466220	7
Aroclor-1242-2	(2)	183791222	188978371	198913930	207089908	207312700	5
Aroclor-1242-3	(3)	100083291	104057155	110676450	116018292	114379700	6
Aroclor-1242-4	(4)	101694040	106737525	114342192	123447428	121997980	8
Aroclor-1242-5	(5)	124175329	128045065	135895568	145256248	150561360	8
Decachlorobiphenyl		2975923130	3104041173	3249737920	3444762080	3552503800	7
Tetrachloro-m-xylene		5146305370	5269389160	5422425040	5452266880	4942424000	4
Aroclor-1248-1	(1)	97508588	101198119	106540372	114549420	116802080	8
Aroclor-1248-2	(2)	135954819	143343636	151605960	164947764	174030220	10
Aroclor-1248-3	(3)	145968318	152945292	161442194	176334456	190180940	11
Aroclor-1248-4	(4)	171715835	178859555	187881420	202639424	216432800	9
Aroclor-1248-5	(5)	167823607	172638881	180053168	192035492	203508660	8
Decachlorobiphenyl		3009595400	3126029333	3260555920	3479534800	3551272200	7
Tetrachloro-m-xylene		5072748910	5224226253	5283938200	5402921120	5021628000	3
Aroclor-1254-1	(1)	253547823	264686405	278940988	295164372	313075680	8
Aroclor-1254-2	(2)	223099297	232757513	247383274	262896088	285176280	10
Aroclor-1254-3	(3)	362340991	375750797	393412880	409817176	422789620	6
Aroclor-1254-4	(4)	207964031	212728924	222445680	231289600	227332320	4
Aroclor-1254-5	(5)	302636627	313319767	329146572	343115096	345356740	6
Decachlorobiphenyl		3009418030	3125275440	3274408080	3481058080	3592269800	7
Tetrachloro-m-xylene		5157782560	5318511973	5462257400	5454366480	5168021000	3
Aroclor-1268-1	(1)	605633372	613614913	625929052	644237020	658094440	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	555793036	560357999	571244564	587032548	581198500	571125329	2
Aroclor-1268-3	(3)	443116844	448542784	458393734	471576124	481232120	460572321	3
Aroclor-1268-4	(4)	171795401	175640587	181667170	188528320	184928840	180512064	4
Aroclor-1268-5	(5)	1182764439	1191499568	1205466938	1214359780	1189190120	1196656169	1
Decachlorobiphenyl		5296716160	5425177360	5590471540	5802552800	5997123200	5622408212	5
Tetrachloro-m-xylene		5377448880	5418716973	5531134800	5564772080	5234786600	5425371867	2