

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 10:14 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.60	5.60	5.50	5.70	0.00
Aroclor-1254-2 (2)	5.75	5.75	5.65	5.85	0.00
Aroclor-1254-3 (3)	6.16	6.16	6.06	6.26	0.00
Aroclor-1254-4 (4)	6.39	6.39	6.29	6.49	0.00
Aroclor-1254-5 (5)	6.81	6.81	6.71	6.91	0.00
Tetrachloro-m-xylene	3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl	8.76	8.76	8.66	8.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

Continuing Calib Date: 01/30/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 10:14 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.58	5.58	5.48	5.68	0.00
Aroclor-1254-2	(2)	5.73	5.73	5.63	5.83	0.00
Aroclor-1254-3	(3)	6.13	6.13	6.03	6.23	0.00
Aroclor-1254-4	(4)	6.36	6.36	6.26	6.46	0.00
Aroclor-1254-5	(5)	6.78	6.78	6.68	6.88	0.00
Tetrachloro-m-xylene		3.70	3.70	3.60	3.80	0.01
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00