

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO109273.D Time Analyzed: 09:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.793	4.695	4.895	516.420	500.000	3.3
Aroclor-1016-2	4.813	4.714	4.914	515.040	500.000	3.0
Aroclor-1016-3	4.868	4.770	4.970	509.030	500.000	1.8
Aroclor-1016-4	4.989	4.891	5.091	513.030	500.000	2.6
Aroclor-1016-5	5.247	5.149	5.349	506.600	500.000	1.3
Aroclor-1260-1	6.290	6.190	6.390	494.930	500.000	-1.0
Aroclor-1260-2	6.478	6.379	6.579	490.200	500.000	-2.0
Aroclor-1260-3	6.847	6.747	6.947	494.570	500.000	-1.1
Aroclor-1260-4	7.108	7.008	7.208	496.140	500.000	-0.8
Aroclor-1260-5	7.349	7.249	7.449	487.690	500.000	-2.5
Decachlorobiphenyl	8.758	8.658	8.858	46.710	50.000	-6.6
Tetrachloro-m-xylene	3.698	3.600	3.800	51.970	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO109273.D Time Analyzed: 09:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.777	4.703	4.903	492.070	500.000	-1.6
Aroclor-1016-2	4.797	4.722	4.922	511.060	500.000	2.2
Aroclor-1016-3	4.973	4.898	5.098	508.430	500.000	1.7
Aroclor-1016-4	5.014	4.940	5.140	486.710	500.000	-2.7
Aroclor-1016-5	5.227	5.153	5.353	507.060	500.000	1.4
Aroclor-1260-1	6.261	6.186	6.386	498.420	500.000	-0.3
Aroclor-1260-2	6.449	6.373	6.573	495.270	500.000	-0.9
Aroclor-1260-3	6.602	6.527	6.727	492.250	500.000	-1.6
Aroclor-1260-4	7.074	6.999	7.199	492.160	500.000	-1.6
Aroclor-1260-5	7.315	7.239	7.439	500.950	500.000	0.2
Decachlorobiphenyl	8.708	8.633	8.833	45.830	50.000	-8.3
Tetrachloro-m-xylene	3.693	3.618	3.818	53.320	50.000	6.6