

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO109288.D Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.792	4.695	4.895	497.330	500.000	-0.5
Aroclor-1016-2	4.812	4.714	4.914	492.970	500.000	-1.4
Aroclor-1016-3	4.868	4.770	4.970	487.850	500.000	-2.4
Aroclor-1016-4	4.989	4.891	5.091	489.660	500.000	-2.1
Aroclor-1016-5	5.246	5.149	5.349	491.810	500.000	-1.6
Aroclor-1260-1	6.289	6.190	6.390	442.630	500.000	-11.5
Aroclor-1260-2	6.477	6.379	6.579	440.620	500.000	-11.9
Aroclor-1260-3	6.845	6.747	6.947	439.260	500.000	-12.1
Aroclor-1260-4	7.106	7.008	7.208	420.790	500.000	-15.8
Aroclor-1260-5	7.347	7.249	7.449	404.900	500.000	-19.0
Decachlorobiphenyl	8.756	8.658	8.858	38.970	50.000	-22.1
Tetrachloro-m-xylene	3.698	3.600	3.800	49.370	50.000	-1.3

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO109288.D Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.778	4.703	4.903	521.430	500.000	4.3
Aroclor-1016-2	4.798	4.722	4.922	526.420	500.000	5.3
Aroclor-1016-3	4.974	4.898	5.098	511.150	500.000	2.2
Aroclor-1016-4	5.015	4.940	5.140	493.930	500.000	-1.2
Aroclor-1016-5	5.229	5.153	5.353	520.120	500.000	4.0
Aroclor-1260-1	6.262	6.186	6.386	473.510	500.000	-5.3
Aroclor-1260-2	6.449	6.373	6.573	461.500	500.000	-7.7
Aroclor-1260-3	6.604	6.527	6.727	456.810	500.000	-8.6
Aroclor-1260-4	7.075	6.999	7.199	439.710	500.000	-12.1
Aroclor-1260-5	7.315	7.239	7.439	442.370	500.000	-11.5
Decachlorobiphenyl	8.708	8.633	8.833	42.440	50.000	-15.1
Tetrachloro-m-xylene	3.695	3.618	3.818	53.020	50.000	6.0