

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: 8082A

DataFile : PO109295.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	RPD
			Result	Result			Qual	RPD	Qual	Low	High	
Client Sample ID: Q1220-01MS	TR-06-01292025MS											
	AR1016	178.4	0	135	ug/kg	76				55	146	
	AR1260	178.4	0	124	ug/kg	70				45	144	