

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: 8082A

DataFile : PO109296.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: Q1220-01MSD	TR-06-01292025MSD											
	AR1016	178.5	0	131	ug/kg	73		4		55	146	20
	AR1260	178.5	0	120	ug/kg	67		4		45	144	20