

Surrogate Summary

SDG No.: **Q1215**

Client: **RU2 Engineering, LLC**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093904.D	PIBLK-PL093904.D	Decachlorobiphenyl	1	20	20.2	101		43	140
		Tetrachloro-m-xylene	1	20	19.1	96		77	126
		Decachlorobiphenyl	2	20	16.0	80		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
PB166359BL	PB166359BL	Decachlorobiphenyl	1	20	23.6	118		10	148
		Tetrachloro-m-xylene	1	20	21.1	106		10	159
		Decachlorobiphenyl	2	20	22.5	112		10	148
		Tetrachloro-m-xylene	2	20	21.6	108		10	159
PB166359BS	PB166359BS	Decachlorobiphenyl	1	20	23.4	117		10	148
		Tetrachloro-m-xylene	1	20	20.6	103		10	159
		Decachlorobiphenyl	2	20	22.0	110		10	148
		Tetrachloro-m-xylene	2	20	20.0	100		10	159
Q1215-07	JPP-29.2-012825	Decachlorobiphenyl	1	20	14.4	72		10	148
		Tetrachloro-m-xylene	1	20	13.1	66		10	159
		Decachlorobiphenyl	2	20	13.2	66		10	148
		Tetrachloro-m-xylene	2	20	13.6	68		10	159
I.BLK-PL093917.D	PIBLK-PL093917.D	Decachlorobiphenyl	1	20	21.2	106		43	140
		Tetrachloro-m-xylene	1	20	19.3	96		77	126
		Decachlorobiphenyl	2	20	18.1	90		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
I.BLK-PL093970.D	PIBLK-PL093970.D	Decachlorobiphenyl	1	20	19.3	96		43	140
		Tetrachloro-m-xylene	1	20	21.3	107		77	126
		Decachlorobiphenyl	2	20	16.1	81		43	140
		Tetrachloro-m-xylene	2	20	20.0	100		77	126
Q1215-03	JPP-29.1-012825	Decachlorobiphenyl	1	20	15.9	80		10	148
		Tetrachloro-m-xylene	1	20	16.6	83		10	159
		Decachlorobiphenyl	2	20	17.0	85		10	148
		Tetrachloro-m-xylene	2	20	18.6	93		10	159
Q1215-03MS	JPP-29.1-012825MS	Decachlorobiphenyl	1	20	16.0	80		10	148
		Tetrachloro-m-xylene	1	20	14.9	75		10	159
		Decachlorobiphenyl	2	20	15.1	76		10	148
		Tetrachloro-m-xylene	2	20	16.8	84		10	159
Q1215-03MSD	JPP-29.1-012825MSD	Decachlorobiphenyl	1	20	15.1	75		10	148
		Tetrachloro-m-xylene	1	20	14.5	73		10	159
		Decachlorobiphenyl	2	20	14.5	72		10	148
		Tetrachloro-m-xylene	2	20	16.6	83		10	159
I.BLK-PL093977.D	PIBLK-PL093977.D	Decachlorobiphenyl	1	20	18.9	94		43	140

Surrogate Summary

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093977.D	PIBLK-PL093977.D	Tetrachloro-m-xylene	1	20	21.9	109		77	126
		Decachlorobiphenyl	2	20	18.1	91		43	140
		Tetrachloro-m-xylene	2	20	21.0	105		77	126