

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: 8081B

DataFile : PL093973.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	JPP-29.1-012825MS											
Q1215-03MS	alpha-BHC	18.73	0	17.8	ug/kg	95				60	144	
	beta-BHC	18.73	0	17.0	ug/kg	91				54	143	
	delta-BHC	18.73	0	17.6	ug/kg	94				47	144	
	gamma-BHC (Lindane)	18.73	0	18.7	ug/kg	100				61	140	
	Heptachlor	18.73	0	17.8	ug/kg	95				63	135	
	Aldrin	18.73	0	17.0	ug/kg	91				49	139	
	Heptachlor epoxide	18.73	0	17.1	ug/kg	91				32	180	
	Endosulfan I	18.73	0	16.7	ug/kg	89				56	142	
	Dieldrin	18.73	0	16.7	ug/kg	89				47	161	
	4,4'-DDE	18.73	0	17.6	ug/kg	94				55	136	
	Endrin	18.73	0	17.7	ug/kg	95				57	139	
	Endosulfan II	18.73	0	16.6	ug/kg	89				40	163	
	4,4'-DDD	18.73	1.5	19.0	ug/kg	93				37	192	
	Endosulfan sulfate	18.73	0	15.7	ug/kg	84				62	139	
	4,4'-DDT	18.73	0	17.1	ug/kg	91				51	146	
	Methoxychlor	18.73	0	16.3	ug/kg	87				54	136	
	Endrin ketone	18.73	0	15.7	ug/kg	84				60	129	
	Endrin aldehyde	18.73	0	15.1	ug/kg	81				59	132	
	alpha-Chlordane	18.73	0	18.2	ug/kg	97				30	192	
	gamma-Chlordane	18.73	0	20.1	ug/kg	107				44	175	