

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: 8081B

DataFile : PL093974.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	JPP-29.1-012825MSD											
Q1215-03MSD	alpha-BHC	18.72	0	17.6	ug/kg	94		1		60	144	20
	beta-BHC	18.72	0	18.6	ug/kg	99		8		54	143	20
	delta-BHC	18.72	0	17.5	ug/kg	93		1		47	144	20
	gamma-BHC (Lindane)	18.72	0	18.5	ug/kg	99		1		61	140	20
	Heptachlor	18.72	0	17.5	ug/kg	93		2		63	135	20
	Aldrin	18.72	0	16.7	ug/kg	89		2		49	139	20
	Heptachlor epoxide	18.72	0	16.8	ug/kg	90		1		32	180	20
	Endosulfan I	18.72	0	15.9	ug/kg	85		5		56	142	20
	Dieldrin	18.72	0	16.3	ug/kg	87		2		47	161	20
	4,4'-DDE	18.72	0	16.4	ug/kg	88		7		55	136	20
	Endrin	18.72	0	17.5	ug/kg	93		2		57	139	20
	Endosulfan II	18.72	0	16.2	ug/kg	87		2		40	163	20
	4,4'-DDD	18.72	1.5	18.7	ug/kg	92		1		37	192	20
	Endosulfan sulfate	18.72	0	15.4	ug/kg	82		2		62	139	20
	4,4'-DDT	18.72	0	16.3	ug/kg	87		4		51	146	20
	Methoxychlor	18.72	0	15.8	ug/kg	84		4		54	136	20
	Endrin ketone	18.72	0	15.0	ug/kg	80		5		60	129	20
	Endrin aldehyde	18.72	0	14.7	ug/kg	79		3		59	132	20
	alpha-Chlordane	18.72	0	17.6	ug/kg	94		3		30	192	20
	gamma-Chlordane	18.72	0	19.6	ug/kg	105		2		44	175	20