

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: **8081B** Datafile : PL093908.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB166359BS	alpha-BHC	16.65	16.6	ug/kg	100				84	123	
	beta-BHC	16.65	16.9	ug/kg	102				82	123	
	delta-BHC	16.65	16.4	ug/kg	98				83	126	
	gamma-BHC (Lindane)	16.65	16.3	ug/kg	98				83	125	
	Heptachlor	16.65	16.6	ug/kg	100				83	122	
	Aldrin	16.65	16.5	ug/kg	99				82	124	
	Heptachlor epoxide	16.65	17.1	ug/kg	103				83	120	
	Endosulfan I	16.65	17.6	ug/kg	106				81	124	
	Dieldrin	16.65	17.3	ug/kg	104				85	121	
	4,4'-DDE	16.65	18.2	ug/kg	109				81	123	
	Endrin	16.65	16.1	ug/kg	97				76	130	
	Endosulfan II	16.65	17.6	ug/kg	106				80	125	
	4,4'-DDD	16.65	19.6	ug/kg	118				80	131	
	Endosulfan sulfate	16.65	17.5	ug/kg	105				81	122	
	4,4'-DDT	16.65	16.6	ug/kg	100				70	129	
	Methoxychlor	16.65	16.3	ug/kg	98				60	119	
	Endrin ketone	16.65	17.8	ug/kg	107				77	132	
	Endrin aldehyde	16.65	16.9	ug/kg	102				79	124	
	alpha-Chlordane	16.65	17.6	ug/kg	106				84	120	
	gamma-Chlordane	16.65	17.8	ug/kg	107				83	122	