

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 02/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS029032.D Time Analyzed: 02:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.184	9.089	9.289	733.590	712.500	3.0
2,4-D	8.311	8.216	8.416	716.460	705.000	1.6
2,4-DCAA	7.194	7.097	7.297	756.460	750.000	0.9

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 02/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS029032.D Time Analyzed: 02:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.805	9.713	9.913	726.680	712.500	2.0
2,4-D	8.905	8.813	9.013	678.220	705.000	-3.8
2,4-DCAA	7.671	7.577	7.777	701.780	750.000	-6.4