

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL012125

Review By	Abdul	Review On	1/22/2025 8:17:54 AM
Supervise By	Ankita	Supervise On	1/22/2025 8:30:39 AM
SubDirectory	PL012125	HP Acquire Method	HP Processing Method pl012125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL093724.D	21 Jan 2025 10:03		AR\AJ	Ok
2	I.BLK	I.BLK	PL093725.D	21 Jan 2025 10:16		AR\AJ	Ok
3	PEM	PEM	PL093726.D	21 Jan 2025 10:30		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL093727.D	21 Jan 2025 10:43		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL093728.D	21 Jan 2025 10:57		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL093729.D	21 Jan 2025 11:10		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL093730.D	21 Jan 2025 11:24		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL093731.D	21 Jan 2025 11:38		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL093732.D	21 Jan 2025 11:51		AR\AJ	Ok
10	PCHLORICC1000	PCHLORICC1000	PL093733.D	21 Jan 2025 12:05		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL093734.D	21 Jan 2025 12:18		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL093735.D	21 Jan 2025 12:32		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL093736.D	21 Jan 2025 12:45		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL093737.D	21 Jan 2025 12:59		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL093738.D	21 Jan 2025 13:12		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL093739.D	21 Jan 2025 13:26		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL093740.D	21 Jan 2025 13:39		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL093741.D	21 Jan 2025 13:53		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL012125

Review By	Abdul	Review On	1/22/2025 8:17:54 AM
Supervise By	Ankita	Supervise On	1/22/2025 8:30:39 AM
SubDirectory	PL012125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL093742.D	21 Jan 2025 14:07		AR\AJ	Ok
20	PSTDICV050	ICVPL012125	PL093743.D	21 Jan 2025 14:20		AR\AJ	Ok
21	PCHLORICV500	ICVPL012125CHLOR	PL093744.D	21 Jan 2025 14:47		AR\AJ	Ok,M
22	PTOXICV500	ICVPL012125TOX	PL093745.D	21 Jan 2025 15:14		AR\AJ	Ok
23	I.BLK	I.BLK	PL093746.D	21 Jan 2025 15:41		AR\AJ	Ok
24	PEM	PEM	PL093747.D	21 Jan 2025 15:54		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PL093748.D	21 Jan 2025 16:07		AR\AJ	Ok,M
26	Q1093-01RE	RT-3407RE	PL093749.D	21 Jan 2025 16:21	F Flag coming , DCB high in 2nd column	AR\AJ	Confirms
27	I.BLK	I.BLK	PL093750.D	21 Jan 2025 16:34		AR\AJ	Ok
28	PSTDCCC050	PSTDCCC050	PL093751.D	21 Jan 2025 16:48		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL020325

Review By	Abdul	Review On	2/4/2025 2:03:14 PM
Supervise By	Ankita	Supervise On	2/4/2025 3:37:15 PM
SubDirectory	PL020325	HP Acquire Method	HP Processing Method pl012125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL093979.D	03 Feb 2025 09:01		AR\AJ	Ok
2	I.BLK	I.BLK	PL093980.D	03 Feb 2025 09:14		AR\AJ	Ok
3	PEM	PEM	PL093981.D	03 Feb 2025 11:03		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL093982.D	03 Feb 2025 12:17		AR\AJ	Ok,M
5	Q1216-15	JPP-26.1-012825	PL093983.D	03 Feb 2025 12:40	DCB high in 2nd column	AR\AJ	Ok,M
6	PB166427BL	PB166427BL	PL093984.D	03 Feb 2025 12:53		AR\AJ	Ok,M
7	PB166427BS	PB166427BS	PL093985.D	03 Feb 2025 13:26		AR\AJ	Ok,M
8	PB166356TB	PB166356TB	PL093986.D	03 Feb 2025 13:48		AR\AJ	Ok
9	Q1215-04	JPP-29.1-012825	PL093987.D	03 Feb 2025 14:02		AR\AJ	Ok
10	Q1215-04MS	JPP-29.1-012825MS	PL093988.D	03 Feb 2025 14:24		AR\AJ	Ok,M
11	Q1215-04MSD	JPP-29.1-012825MSD	PL093989.D	03 Feb 2025 14:37		AR\AJ	Ok,M
12	Q1206-04	JPP-20.1-012725	PL093990.D	03 Feb 2025 14:50		AR\AJ	Ok
13	Q1206-04MS	JPP-20.1-012725MS	PL093991.D	03 Feb 2025 15:04		AR\AJ	Ok,M
14	Q1206-04MSD	JPP-20.1-012725MSD	PL093992.D	03 Feb 2025 15:22		AR\AJ	Ok
15	Q1215-08	JPP-29.2-012825	PL093993.D	03 Feb 2025 15:35		AR\AJ	Ok,M
16	Q1216-04	JPP-18.1-012825	PL093994.D	03 Feb 2025 15:49		AR\AJ	Ok
17	Q1216-08	JPP-21.1-012825	PL093995.D	03 Feb 2025 16:02		AR\AJ	Ok
18	Q1216-12	JPP-21.2-012825	PL093996.D	03 Feb 2025 16:15		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL020325

Review By	Abdul	Review On	2/4/2025 2:03:14 PM
Supervise By	Ankita	Supervise On	2/4/2025 3:37:15 PM
SubDirectory	PL020325	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1216-16	JPP-26.1-012825	PL093997.D	03 Feb 2025 16:29		AR\AJ	Ok
20	Q1216-20	JPP-26.2-012825	PL093998.D	03 Feb 2025 16:42		AR\AJ	Ok,M
21	I.BLK	I.BLK	PL093999.D	03 Feb 2025 16:55		AR\AJ	Ok,M
22	PEM	PEM	PL094000.D	03 Feb 2025 17:08		AR\AJ	Ok,M
23	PSTDCCC050	PSTDCCC050	PL094001.D	03 Feb 2025 17:50		AR\AJ	Ok
24	Q1232-04	JPP-46.2-012925	PL094002.D	03 Feb 2025 18:03		AR\AJ	Ok
25	Q1232-08	JPP-46.1-012925	PL094003.D	03 Feb 2025 18:16		AR\AJ	Ok,M
26	Q1232-12	JPP-42.1-012925	PL094004.D	03 Feb 2025 18:29		AR\AJ	Ok
27	Q1232-16	JPP-42.2-012925	PL094005.D	03 Feb 2025 18:43		AR\AJ	Ok
28	Q1232-20	JPP-51.1-012925	PL094006.D	03 Feb 2025 18:56		AR\AJ	Ok
29	Q1235-04	JPP-51.2-012925	PL094007.D	03 Feb 2025 19:09	TCMX high in 1st column	AR\AJ	Ok
30	Q1235-08	JPP-16.1-012925	PL094008.D	03 Feb 2025 19:23		AR\AJ	Ok
31	PB166484BL	PB166484BL	PL094009.D	03 Feb 2025 19:36		AR\AJ	Ok
32	PB166484BS	PB166484BS	PL094010.D	03 Feb 2025 19:49	DCB high in 2nd and TCMX high in 1st column , Comp#20 recovery fail	AR\AJ	Not Ok
33	PB166423TB	PB166423TB	PL094011.D	03 Feb 2025 20:02		AR\AJ	Ok
34	Q1241-04	JPP-3.5-013025	PL094012.D	03 Feb 2025 20:16		AR\AJ	Ok,M
35	Q1241-04MS	JPP-3.5-013025MS	PL094013.D	03 Feb 2025 20:29		AR\AJ	Ok,M
36	Q1241-04MSD	JPP-3.5-013025MSD	PL094014.D	03 Feb 2025 20:42		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL020325

Review By	Abdul	Review On	2/4/2025 2:03:14 PM
Supervise By	Ankita	Supervise On	2/4/2025 3:37:15 PM
SubDirectory	PL020325	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	Q1241-08	JPP-5.3-013025	PL094015.D	03 Feb 2025 20:56		AR\AJ	Ok
38	Q1241-12	JPP-5.2-013025	PL094016.D	03 Feb 2025 21:09		AR\AJ	Ok
39	Q1241-16	JPP-5.4-013025	PL094017.D	03 Feb 2025 21:22		AR\AJ	Ok,M
40	Q1241-20	JPP-5.1.4-013025	PL094018.D	03 Feb 2025 21:35		AR\AJ	Ok
41	Q1242-04	JPP-6.2-013025	PL094019.D	03 Feb 2025 21:48		AR\AJ	Ok
42	I.BLK	I.BLK	PL094020.D	03 Feb 2025 22:02		AR\AJ	Ok
43	PEM	PEM	PL094021.D	03 Feb 2025 22:15		AR\AJ	Ok,M
44	PSTDCCC050	PSTDCCC050	PL094022.D	03 Feb 2025 22:41		AR\AJ	Ok,M
45	PB166481BL	PB166481BL	PL094023.D	03 Feb 2025 23:21		AR\AJ	Ok
46	PB166481BS	PB166481BS	PL094024.D	03 Feb 2025 23:34	Comp#20 recovery fail in 2nd col	AR\AJ	Not Ok
47	Q1254-01	OK-02-01312025	PL094025.D	03 Feb 2025 23:47		AR\AJ	Ok,M
48	Q1261-01	CHRT-20430	PL094026.D	04 Feb 2025 00:00	F Flag coming	AR\AJ	ReRun
49	Q1262-01	ETGI-371	PL094027.D	04 Feb 2025 00:13	DCB high in both column	AR\AJ	ReRun
50	Q1262-03	CONCRETE-PAD	PL094028.D	04 Feb 2025 00:27	DCB high in 1st column	AR\AJ	Ok,M
51	Q1269-01	VNJ-231	PL094029.D	04 Feb 2025 00:40		AR\AJ	Ok,M
52	Q1269-01MS	VNJ-231MS	PL094030.D	04 Feb 2025 00:53		AR\AJ	Ok,M
53	Q1269-01MSD	VNJ-231MSD	PL094031.D	04 Feb 2025 01:06		AR\AJ	Ok,M
54	Q1271-01	RBR200030	PL094032.D	04 Feb 2025 01:19		AR\AJ	Ok,M
55	I.BLK	I.BLK	PL094033.D	04 Feb 2025 01:32		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL020325

Review By	Abdul	Review On	2/4/2025 2:03:14 PM
Supervise By	Ankita	Supervise On	2/4/2025 3:37:15 PM
SubDirectory	PL020325	HP Acquire Method	HP Processing Method pl012125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

56	PSTDCCC050	PSTDCCC050	PL094034.D	04 Feb 2025 01:46		AR/AJ	Ok,M
----	------------	------------	------------	-------------------	--	-------	------

M : Manual Integration