

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG No.: Q1215

Instrument ID: BNA_F Calibration Date/Time: 02/04/2025 10:42

Lab File ID: BF141398.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.303	1.302		-0.1	
Benzaldehyde	0.960	0.893		-7.0	
Phenol-d6	1.657	1.560		-5.9	
Phenol	1.757	1.650		-6.1	20.0
bis(2-Chloroethyl)ether	1.347	1.262		-6.3	
2-Chlorophenol	1.396	1.387		-0.6	
2-Methylphenol	1.135	1.074		-5.4	
2,2-oxybis(1-Chloropropane)	2.342	2.108		-10.0	
Acetophenone	0.475	0.491		3.4	
3+4-Methylphenols	1.393	1.375		-1.3	
n-Nitroso-di-n-propylamine	0.985	0.929	0.050	-5.7	
Nitrobenzene-d5	0.379	0.391		3.2	
Hexachloroethane	0.571	0.583		2.1	
Nitrobenzene	0.393	0.398		1.3	
Isophorone	0.656	0.643		-2.0	
2-Nitrophenol	0.185	0.185		0.0	20.0
2,4-Dimethylphenol	0.236	0.230		-2.5	
bis(2-Chloroethoxy)methane	0.418	0.403		-3.6	
2,4-Dichlorophenol	0.289	0.296		2.4	20.0
Naphthalene	1.057	1.061		0.4	
4-Chloroaniline	0.358	0.347		-3.1	
Hexachlorobutadiene	0.194	0.208		7.2	20.0
Caprolactam	0.094	0.093		-1.1	
4-Chloro-3-methylphenol	0.310	0.325		4.8	20.0
2-Methylnaphthalene	0.672	0.684		1.8	
Hexachlorocyclopentadiene	0.168	0.185	0.050	10.1	
2,4,6-Trichlorophenol	0.372	0.382		2.7	20.0
2-Fluorobiphenyl	1.299	1.328		2.2	
2,4,5-Trichlorophenol	0.405	0.421		4.0	
1,1-Biphenyl	1.560	1.563		0.2	
2-Chloronaphthalene	1.216	1.209		-0.6	
2-Nitroaniline	0.382	0.390		2.1	
Dimethylphthalate	1.351	1.371		1.5	
Acenaphthylene	1.702	1.717		0.9	
2,6-Dinitrotoluene	0.314	0.321		2.2	
3-Nitroaniline	0.307	0.321		4.6	
Acenaphthene	1.216	1.256		3.3	20.0
2,4-Dinitrophenol	0.146	0.165	0.050	13.0	
4-Nitrophenol	0.225	0.257	0.050	14.2	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG No.: Q1215

Instrument ID: BNA_F Calibration Date/Time: 02/04/2025 10:42

Lab File ID: BF141398.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.627	1.669		2.6	
2,4-Dinitrotoluene	0.406	0.441		8.6	
Diethylphthalate	1.310	1.374		4.9	
4-Chlorophenyl-phenylether	0.617	0.648		5.0	
Fluorene	1.262	1.306		3.5	
4-Nitroaniline	0.302	0.344		13.9	
4,6-Dinitro-2-methylphenol	0.118	0.137		16.1	
n-Nitrosodiphenylamine	0.647	0.634		-2.0	20.0
2,4,6-Tribromophenol	0.183	0.202		10.4	
4-Bromophenyl-phenylether	0.224	0.225		0.4	
Hexachlorobenzene	0.238	0.243		2.1	
Atrazine	0.216	0.214		-0.9	
Pentachlorophenol	0.139	0.161		15.8	20.0
Phenanthrene	1.075	1.109		3.2	
Anthracene	1.053	1.092		3.7	
Carbazole	0.934	1.029		10.1	
Di-n-butylphthalate	1.132	1.214		7.2	
Fluoranthene	1.062	1.240		16.8	20.0
Pyrene	1.920	1.729		-9.9	
Terphenyl-d14	1.297	1.200		-7.5	
Butylbenzylphthalate	0.671	0.691		3.0	
3,3-Dichlorobenzidine	0.439	0.412		-6.2	
Benzo (a) anthracene	1.380	1.305		-5.4	
Chrysene	1.265	1.270		0.4	
Bis(2-ethylhexyl)phthalate	0.851	0.910		6.9	
Di-n-octyl phthalate	1.321	1.415		7.1	20.0
Benzo (b) fluoranthene	1.258	1.341		6.6	
Benzo (k) fluoranthene	1.114	1.140		2.3	
Benzo (a) pyrene	1.063	1.072		0.8	20.0
Indeno (1,2,3-cd) pyrene	1.291	1.338		3.6	
Dibenzo (a,h) anthracene	1.039	1.068		2.8	
Benzo (g,h,i) perylene	1.080	1.113		3.1	
1,2,4,5-Tetrachlorobenzene	0.569	0.587		3.2	
1,4-Dioxane	0.573	0.555		-3.1	20.0
2,3,4,6-Tetrachlorophenol	0.319	0.341		6.9	

All other compounds must meet a minimum RRF of 0.010.