

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/11/2025

Lab File ID: BF141551.D Time Analyzed: 10:28

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	100982	6.793	399334	8.08	220290	9.83
UPPER LIMIT	201964	7.293	798668	8.575	440580	10.328
LOWER LIMIT	50491	6.293	199667	7.575	110145	9.328
EPA SAMPLE NO.						
01 JPP-29.1-012825	70182	6.79	273856	8.07	136474	9.82
02 JPP-29.1-012825MS	77607	6.79	287166	8.07	145781	9.82
03 JPP-29.1-012825MSD	77684	6.79	297197	8.07	147170	9.82

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.