

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM

SAS No.: Q1215

SDG NO.: Q1215

Lab File ID: BF141289.D

DFTPP Injection Date: 01/29/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:03

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	41.1
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	35.1
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	47.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	26.5
365	Greater than 1% of mass 198	3.2
441	Present, but less than mass 443	13.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.2 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF141290.D	01/29/2025	10:30
SSTDICC005	SSTDICC005	BF141291.D	01/29/2025	10:56
SSTDICC010	SSTDICC010	BF141292.D	01/29/2025	11:49
SSTDICC020	SSTDICC020	BF141293.D	01/29/2025	15:17
SSTDICCC040	SSTDICCC040	BF141294.D	01/29/2025	15:45
SSTDICC050	SSTDICC050	BF141295.D	01/29/2025	16:11
SSTDICC060	SSTDICC060	BF141296.D	01/29/2025	16:38
SSTDICC080	SSTDICC080	BF141297.D	01/29/2025	17:05