

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG No.: Q1215

Instrument ID: MSVOA_N Calibration Date/Time: 01/31/2025 10:04

Lab File ID: VN085591.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.737	0.631		-14.38	20
1,1-Dichloroethene	0.537	0.497		-7.45	20
2-Butanone	0.384	0.389		1.3	20
Carbon Tetrachloride	0.557	0.576		3.41	20
Chloroform	1.218	1.221		0.25	20
Benzene	1.463	1.491		1.91	20
1,2-Dichloroethane	0.551	0.571		3.63	20
Trichloroethene	0.341	0.337		-1.17	20
Tetrachloroethene	0.341	0.357		4.69	20
Chlorobenzene	1.095	1.091	0.3	-0.37	20
1,2-Dichloroethane-d4	0.807	0.801		-0.74	20
Dibromofluoromethane	0.347	0.374		7.78	20
Toluene-d8	1.232	1.332		8.12	20
4-Bromofluorobenzene	0.422	0.473		12.09	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.