

Surrogate Summary

SDG No.: **Q1215**

Client: **RU2 Engineering, LLC**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1215-02	JPP-29.1-012825	1,2-Dichloroethane-d4	50	60.9	122	74	125
		Dibromofluoromethane	50	50.0	100	75	124
		Toluene-d8	50	51.9	104	86	113
		4-Bromofluorobenzene	50	41.7	83	77	121
Q1215-06	JPP-29.2-012825	1,2-Dichloroethane-d4	50	50.1	100	74	125
		Dibromofluoromethane	50	50.1	100	75	124
		Toluene-d8	50	55.2	110	86	113
		4-Bromofluorobenzene	50	52.5	105	77	121
VN0131WBL01	VN0131WBL01	1,2-Dichloroethane-d4	50	57.4	115	74	125
		Dibromofluoromethane	50	54.1	108	75	124
		Toluene-d8	50	49.7	99	86	113
		4-Bromofluorobenzene	50	45.4	91	77	121
VN0131WBS01	VN0131WBS01	1,2-Dichloroethane-d4	50	50.3	101	74	125
		Dibromofluoromethane	50	52.5	105	75	124
		Toluene-d8	50	53.6	107	86	113
		4-Bromofluorobenzene	50	50.8	102	77	121
VN0131WBSD01	VN0131WBSD01	1,2-Dichloroethane-d4	50	53.2	106	74	125
		Dibromofluoromethane	50	53.2	106	75	124
		Toluene-d8	50	52.1	104	86	113
		4-Bromofluorobenzene	50	53.2	106	77	121
VX0203WBL01	VX0203WBL01	1,2-Dichloroethane-d4	50	48.6	97	74	125
		Dibromofluoromethane	50	50.1	100	75	124
		Toluene-d8	50	55.2	110	86	113
		4-Bromofluorobenzene	50	52.9	106	77	121
VX0203WBS01	VX0203WBS01	1,2-Dichloroethane-d4	50	43.1	86	74	125
		Dibromofluoromethane	50	46.5	93	75	124
		Toluene-d8	50	50.2	100	86	113
		4-Bromofluorobenzene	50	47.5	95	77	121