

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RUTW01
Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215
Lab File ID: VX044647.D BFB Injection Date: 01/15/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:30
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.6
75	30.0 - 60.0% of mass 95	50.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.6 (0.9) 1
174	50.0 - 100.0% of mass 95	67.1
175	5.0 - 9.0% of mass 174	4.9 (7.3) 1
176	95.0 - 101.0% of mass 174	64.4 (96) 1
177	5.0 - 9.0% of mass 176	4.3 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX044648.D	01/15/2025	10:31
VSTDIC005	VSTDIC005	VX044649.D	01/15/2025	11:17
VSTDIC020	VSTDIC020	VX044650.D	01/15/2025	11:40
VSTDIC050	VSTDIC050	VX044651.D	01/15/2025	12:02
VSTDIC100	VSTDIC100	VX044652.D	01/15/2025	13:14
VSTDIC150	VSTDIC150	VX044653.D	01/15/2025	13:37