

Surrogate Summary

SDG No.: Q1215

Client: RU2 Engineering, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1215-01	JPP-29.1-012825	1,2-Dichloroethane-d4	50	72.1	144	50	163
		Dibromofluoromethane	50	60.6	121	54	147
		Toluene-d8	50	36.6	73	58	134
		4-Bromofluorobenzene	50	9.96	20 *	29	146
Q1215-05	JPP-29.2-012825	1,2-Dichloroethane-d4	50	35.9	72	50	163
		Dibromofluoromethane	50	52.4	105	54	147
		Toluene-d8	50	43.6	87	58	134
		4-Bromofluorobenzene	50	18.5	37	29	146
VY0203SBL01	VY0203SBL01	1,2-Dichloroethane-d4	50	50.6	101	50	163
		Dibromofluoromethane	50	49.4	99	54	147
		Toluene-d8	50	47.8	96	58	134
		4-Bromofluorobenzene	50	43.4	87	29	146
VY0203SBS01	VY0203SBS01	1,2-Dichloroethane-d4	50	45.9	92	50	163
		Dibromofluoromethane	50	48.8	98	54	147
		Toluene-d8	50	50.2	100	58	134
		4-Bromofluorobenzene	50	49.7	99	29	146
VY0203SBSD01	VY0203SBSD01	1,2-Dichloroethane-d4	50	49.0	98	50	163
		Dibromofluoromethane	50	48.3	97	54	147
		Toluene-d8	50	50.2	100	58	134
		4-Bromofluorobenzene	50	49.7	99	29	146
VY0204SBL01	VY0204SBL01	1,2-Dichloroethane-d4	50	57.4	115	50	163
		Dibromofluoromethane	50	51.6	103	54	147
		Toluene-d8	50	49.8	100	58	134
		4-Bromofluorobenzene	50	47.5	95	29	146
VY0204SBS01	VY0204SBS01	1,2-Dichloroethane-d4	50	44.3	89	50	163
		Dibromofluoromethane	50	44.6	89	54	147
		Toluene-d8	50	45.4	91	58	134
		4-Bromofluorobenzene	50	45.0	90	29	146