

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RUTW01
Lab Code: CHEM Case No.: Q1215 SAS No.: Q1215 SDG NO.: Q1215
Lab File ID: VY021019.D BFB Injection Date: 02/03/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:51
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.3
75	30.0 - 60.0% of mass 95	49
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.3 (0.4) 1
174	50.0 - 100.0% of mass 95	89.7
175	5.0 - 9.0% of mass 174	7 (7.8) 1
176	95.0 - 101.0% of mass 174	87 (97) 1
177	5.0 - 9.0% of mass 176	5.9 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY021020.D	02/03/2025	10:35
VSTDIC010	VSTDIC010	VY021021.D	02/03/2025	10:57
VSTDIC020	VSTDIC020	VY021022.D	02/03/2025	11:20
VSTDIC050	VSTDIC050	VY021023.D	02/03/2025	11:43
VSTDIC100	VSTDIC100	VY021024.D	02/03/2025	12:21
VSTDIC150	VSTDIC150	VY021025.D	02/03/2025	12:44
VY0203SBL01	VY0203SBL01	VY021028.D	02/03/2025	14:13
VY0203SBS01	VY0203SBS01	VY021029.D	02/03/2025	14:55
VY0203SBSD01	VY0203SBSD01	VY021030.D	02/03/2025	15:17
JPP-29.1-012825	Q1215-01	VY021032.D	02/03/2025	16:14