

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

JPP-16.1-012925L

Lab Name: Chemtech Consulting Group

Contract: RUTW01

Lab Code: CHEM **Lb No.:** lb134659 **Lab Sample ID :** Q1235-08L **SDG No.:** Q1216

Matrix (soil/water): Water **Level (low/med):** LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Arsenic	100	U	500	U			P
Barium	333	J	2500	U	100.0		P
Cadmium	30.0	U	150	U			P
Chromium	9.60	J	62.4	J	550		P
Lead	60.0	U	300	U			P
Mercury	2.00	U	10.0	U			CV
Selenium	100	U	500	U			P
Silver	50.0	U	250	U			P