

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1216 SAS No.: Q1216 SDG NO.: Q1216

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 01/31/2025

Lab Sample No.: HSTDCCC750 Data File : PS029020.D Time Analyzed: 21:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.184	9.089	9.289	724.050	712.500	1.6
2,4-D	8.312	8.216	8.416	712.460	705.000	1.1
2,4-DCAA	7.195	7.097	7.297	747.480	750.000	-0.3

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1216 SAS No.: Q1216 SDG NO.: Q1216

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 01/31/2025

Lab Sample No.: HSTDCCC750 Data File : PS029020.D Time Analyzed: 21:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.806	9.713	9.913	721.210	712.500	1.2
2,4-D	8.907	8.813	9.013	668.490	705.000	-5.2
2,4-DCAA	7.672	7.577	7.777	698.680	750.000	-6.8