

Surrogate Summary

SDG No.: Q1216

Client: RU2 Engineering, LLC

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS028900.D	PIBLK-PS028900.D	2,4-DCAA	1	500	474	95		39	175
		2,4-DCAA	2	500	492	98		39	175
I.BLK-PS029019.D	PIBLK-PS029019.D	2,4-DCAA	1	500	581	116		39	175
		2,4-DCAA	2	500	551	110		39	175
PB166428BL	PB166428BL	2,4-DCAA	1	500	477	95		39	175
		2,4-DCAA	2	500	427	85		39	175
PB166428BS	PB166428BS	2,4-DCAA	1	500	543	109		39	175
		2,4-DCAA	2	500	489	98		39	175
PB166356TB	PB166356TB	2,4-DCAA	1	500	467	93		39	175
		2,4-DCAA	2	500	326	65		39	175
Q1215-04MS	JPP-29.1-012825MS	2,4-DCAA	1	500	591	118		39	175
		2,4-DCAA	2	500	363	73		39	175
Q1215-04MSD	JPP-29.1-012825MSD	2,4-DCAA	1	500	592	118		39	175
		2,4-DCAA	2	500	364	73		39	175
Q1216-04	JPP-18.1-012825	2,4-DCAA	1	500	558	112		39	175
		2,4-DCAA	2	500	351	70		39	175
Q1216-08	JPP-21.1-012825	2,4-DCAA	1	500	503	101		39	175
		2,4-DCAA	2	500	356	71		39	175
I.BLK-PS029031.D	PIBLK-PS029031.D	2,4-DCAA	1	500	584	117		39	175
		2,4-DCAA	2	500	554	111		39	175
Q1216-12	JPP-21.2-012825	2,4-DCAA	1	500	531	106		39	175
		2,4-DCAA	2	500	348	70		39	175
Q1216-16	JPP-26.1-012825	2,4-DCAA	1	500	522	104		39	175
		2,4-DCAA	2	500	344	69		39	175
Q1216-20	JPP-26.2-012825	2,4-DCAA	1	500	533	107		39	175
		2,4-DCAA	2	500	352	70		39	175
I.BLK-PS029043.D	PIBLK-PS029043.D	2,4-DCAA	1	500	572	114		39	175
		2,4-DCAA	2	500	551	110		39	175