

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1216

Client: RU2 Engineering, LLC

Analytical Method: 8151A

DataFile : PS029026.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	RPD
			Result	Result			Qual	RPD	Qual	Low	High	
Client Sample ID: Q1215-04MS	JPP-29.1-012825MS											
	2,4-D	50	0	52.4	ug/L	105				65	135	
	2,4,5-TP(Silvex)	50	0	63.5	ug/L	127				62	139	