

### Surrogate Summary

SDG No.: Q1216

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093980.D	PIBLK-PL093980.D	Decachlorobiphenyl	1	20	21.6	108		43	140
		Tetrachloro-m-xylene	1	20	22.2	111		77	126
		Decachlorobiphenyl	2	20	20.2	101		43	140
		Tetrachloro-m-xylene	2	20	21.8	109		77	126
PB166427BL	PB166427BL	Decachlorobiphenyl	1	20	21.1	106		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
		Decachlorobiphenyl	2	20	14.5	73		43	140
		Tetrachloro-m-xylene	2	20	17.5	87		77	126
PB166427BS	PB166427BS	Decachlorobiphenyl	1	20	23.4	117		43	140
		Tetrachloro-m-xylene	1	20	23.7	119		77	126
		Decachlorobiphenyl	2	20	20.9	105		43	140
		Tetrachloro-m-xylene	2	20	23.0	115		77	126
PB166356TB	PB166356TB	Decachlorobiphenyl	1	20	20.1	101		43	140
		Tetrachloro-m-xylene	1	20	21.0	105		77	126
		Decachlorobiphenyl	2	20	18.6	93		43	140
		Tetrachloro-m-xylene	2	20	19.5	98		77	126
Q1215-04MS	JPP-29.1-012825MS	Decachlorobiphenyl	1	20	22.9	115		43	140
		Tetrachloro-m-xylene	1	20	20.0	100		77	126
		Decachlorobiphenyl	2	20	20.6	103		43	140
		Tetrachloro-m-xylene	2	20	19.2	96		77	126
Q1215-04MSD	JPP-29.1-012825MSD	Decachlorobiphenyl	1	20	22.9	114		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	21.7	108		43	140
		Tetrachloro-m-xylene	2	20	20.1	100		77	126
Q1216-04	JPP-18.1-012825	Decachlorobiphenyl	1	20	25.1	125		43	140
		Tetrachloro-m-xylene	1	20	23.3	116		77	126
		Decachlorobiphenyl	2	20	23.2	116		43	140
		Tetrachloro-m-xylene	2	20	22.8	114		77	126
Q1216-08	JPP-21.1-012825	Decachlorobiphenyl	1	20	26.3	131		43	140
		Tetrachloro-m-xylene	1	20	22.5	113		77	126
		Decachlorobiphenyl	2	20	20.5	102		43	140
		Tetrachloro-m-xylene	2	20	21.9	110		77	126
Q1216-12	JPP-21.2-012825	Decachlorobiphenyl	1	20	25.2	126		43	140
		Tetrachloro-m-xylene	1	20	23.2	116		77	126
		Decachlorobiphenyl	2	20	24.7	123		43	140
		Tetrachloro-m-xylene	2	20	22.6	113		77	126
Q1216-16	JPP-26.1-012825	Decachlorobiphenyl	1	20	25.4	127		43	140

**Surrogate Summary**

**SDG No.:** Q1216

**Client:** RU2 Engineering, LLC

**Analytical Method:** 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1216-16	JPP-26.1-012825	Tetrachloro-m-xylene	1	20	23.4	117	77	126	
		Decachlorobiphenyl	2	20	25.1	125	43	140	
		Tetrachloro-m-xylene	2	20	22.6	113	77	126	
Q1216-20	JPP-26.2-012825	Decachlorobiphenyl	1	20	28.1	140	43	140	
		Tetrachloro-m-xylene	1	20	21.4	107	77	126	
		Decachlorobiphenyl	2	20	22.4	112	43	140	
		Tetrachloro-m-xylene	2	20	20.6	103	77	126	
I.BLK-PL093999.D	PIBLK-PL093999.D	Decachlorobiphenyl	1	20	28.1	140	43	140	
		Tetrachloro-m-xylene	1	20	21.1	106	77	126	
		Decachlorobiphenyl	2	20	21.9	109	43	140	
		Tetrachloro-m-xylene	2	20	20.5	102	77	126	