

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1216

Client: RU2 Engineering, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1218-02MS	Client Sample ID:	BELL-25-002MS					DataFile:	BF141339.D		
Pyridine	500	0	210	ug/L	42				10	109	
1,4-Dichlorobenzene	500	0	360	ug/L	72				55	125	
2-Methylphenol	500	0	360	ug/L	72				37	126	
3+4-Methylphenols	500	0	350	ug/L	70				31	127	
Hexachloroethane	500	0	380	ug/L	76				49	110	
Nitrobenzene	500	0	480	ug/L	96				62	112	
Hexachlorobutadiene	500	0	340	ug/L	68				52	125	
2,4,6-Trichlorophenol	500	0	460	ug/L	92				78	112	
2,4,5-Trichlorophenol	500	0	420	ug/L	84				71	111	
2,4-Dinitrotoluene	500	0	380	ug/L	76				50	142	
Hexachlorobenzene	500	0	440	ug/L	88				72	115	
Pentachlorophenol	1000	0	980	ug/L	98				25	139	