

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM

SAS No.: Q1216

SDG NO.: Q1216

Lab File ID: BF141471.D

DFTPP Injection Date: 02/06/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:41

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	43.2
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	36.7
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	49.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	25.6
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	12.2
442	Greater than 50% of mass 198	75.9
443	15.0 - 24.0% of mass 442	14.4 (18.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF141472.D	02/06/2025	11:07
SSTDICC005	SSTDICC005	BF141473.D	02/06/2025	11:34
SSTDICC010	SSTDICC010	BF141474.D	02/06/2025	12:00
SSTDICC020	SSTDICC020	BF141475.D	02/06/2025	12:26
SSTDICCC040	SSTDICCC040	BF141476.D	02/06/2025	12:55
SSTDICC050	SSTDICC050	BF141477.D	02/06/2025	13:21
SSTDICC060	SSTDICC060	BF141478.D	02/06/2025	13:47
SSTDICC080	SSTDICC080	BF141479.D	02/06/2025	14:14
JPP-18.1-012825	Q1216-04	BF141494.D	02/06/2025	21:30