

Surrogate Summary

SDG No.: **Q1216**

Client: **RU2 Engineering, LLC**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1216-02	JPP-18.1-012825	1,2-Dichloroethane-d4	50	48.5	97	74	125
		Dibromofluoromethane	50	50.0	100	75	124
		Toluene-d8	50	53.5	107	86	113
		4-Bromofluorobenzene	50	48.9	98	77	121
Q1216-06	JPP-21.1-012825	1,2-Dichloroethane-d4	50	47.5	95	74	125
		Dibromofluoromethane	50	49.8	100	75	124
		Toluene-d8	50	54.8	110	86	113
		4-Bromofluorobenzene	50	50.9	102	77	121
Q1216-10	JPP-21.2-012825	1,2-Dichloroethane-d4	50	48.9	98	74	125
		Dibromofluoromethane	50	49.4	99	75	124
		Toluene-d8	50	53.1	106	86	113
		4-Bromofluorobenzene	50	46.6	93	77	121
Q1216-14	JPP-26.1-012825	1,2-Dichloroethane-d4	50	52.9	106	74	125
		Dibromofluoromethane	50	45.8	92	75	124
		Toluene-d8	50	46.3	93	86	113
		4-Bromofluorobenzene	50	37.8	76 *	77	121
Q1216-18	JPP-26.2-012825	1,2-Dichloroethane-d4	50	54.0	108	74	125
		Dibromofluoromethane	50	48.8	98	75	124
		Toluene-d8	50	50.5	101	86	113
		4-Bromofluorobenzene	50	43.7	87	77	121
VN0203WBL01	VN0203WBL01	1,2-Dichloroethane-d4	50	50.4	101	74	125
		Dibromofluoromethane	50	50.6	101	75	124
		Toluene-d8	50	50.1	100	86	113
		4-Bromofluorobenzene	50	42.6	85	77	121
VN0203WBS01	VN0203WBS01	1,2-Dichloroethane-d4	50	47.8	96	74	125
		Dibromofluoromethane	50	52.2	104	75	124
		Toluene-d8	50	54.6	109	86	113
		4-Bromofluorobenzene	50	51.1	102	77	121
VN0203WBSD0	VN0203WBSD01	1,2-Dichloroethane-d4	50	48.0	96	74	125
		Dibromofluoromethane	50	51.0	102	75	124
		Toluene-d8	50	51.7	103	86	113
		4-Bromofluorobenzene	50	48.8	98	77	121
VX0203WBL01	VX0203WBL01	1,2-Dichloroethane-d4	50	48.6	97	74	125
		Dibromofluoromethane	50	50.1	100	75	124
		Toluene-d8	50	55.2	110	86	113
		4-Bromofluorobenzene	50	52.9	106	77	121
VX0203WBS01	VX0203WBS01	1,2-Dichloroethane-d4	50	43.1	86	74	125
		Dibromofluoromethane	50	46.5	93	75	124
		Toluene-d8	50	50.2	100	86	113
		4-Bromofluorobenzene	50	47.5	95	77	121