

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
Lab Code: CHEM Case No.: Q1216 SAS No.: Q1216 SDG NO.: Q1216
Lab File ID: VX044802.D Date Analyzed: 02/03/2025
Instrument ID: MSVOA_X Time Analyzed: 09:52
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	162430	5.54	297715	6.75	252415	10.05
UPPER LIMIT	324860	6.038	595430	7.251	504830	10.549
LOWER LIMIT	81215	5.038	148858	6.251	126208	9.549
EPA SAMPLE NO.						
JPP-18.1-012825	123160	5.54	250036	6.76	218940	10.05
JPP-21.1-012825	123220	5.54	247340	6.76	223027	10.05
JPP-21.2-012825	113728	5.55	233647	6.76	200547	10.05
VX0203WBL01	124993	5.54	257688	6.76	232979	10.05
VX0203WBS01	151779	5.54	276067	6.76	242310	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.