

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020325\
 Data File : VY021027.D
 Acq On : 03 Feb 2025 13:37
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY020325

Quant Time: Feb 04 00:48:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	52.290	-4.6	105	0.00
3 P	Chloromethane	50.000	52.006	-4.0	109	0.00
4 C	Vinyl Chloride	50.000	51.058	-2.1#	107	0.00
5 T	Bromomethane	50.000	52.624	-5.2	112	0.00
6 T	Chloroethane	50.000	49.925	0.2	104	0.00
7 T	Trichlorofluoromethane	50.000	51.864	-3.7	107	0.00
8 T	Diethyl Ether	50.000	46.708	6.6	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.507	-3.0	106	0.00
10 T	Methyl Iodide	50.000	48.858	2.3	94	0.00
11 T	Tert butyl alcohol	250.000	214.161	14.3	94	-0.01
12 CM	1,1-Dichloroethene	50.000	51.810	-3.6#	106	0.00
13 T	Acrolein	250.000	235.305	5.9	97	0.00
14 T	Allyl chloride	50.000	50.663	-1.3	102	0.00
15 T	Acrylonitrile	250.000	230.328	7.9	91	0.00
16 T	Acetone	250.000	243.997	2.4	101	-0.01
17 T	Carbon Disulfide	50.000	52.004	-4.0	106	0.00
18 T	Methyl Acetate	50.000	46.031	7.9	90	0.00
19 T	Methyl tert-butyl Ether	50.000	47.538	4.9	95	0.00
20 T	Methylene Chloride	50.000	49.414	1.2	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.506	-1.0	102	0.00
22 T	Diisopropyl ether	50.000	49.407	1.2	99	0.00
23 T	Vinyl Acetate	250.000	238.370	4.7	93	0.00
24 P	1,1-Dichloroethane	50.000	49.806	0.4	103	0.00
25 T	2-Butanone	250.000	233.503	6.6	91	0.00
26 T	2,2-Dichloropropane	50.000	50.791	-1.6	105	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.785	0.4	101	0.00
28 T	Bromochloromethane	50.000	54.731	-9.5	102	0.00
29 T	Tetrahydrofuran	250.000	228.139	8.7	89	0.00
30 C	Chloroform	50.000	48.846	2.3#	100	0.00
31 T	Cyclohexane	50.000	50.051	-0.1	106	0.00
32 T	1,1,1-Trichloroethane	50.000	50.620	-1.2	103	0.00
33 S	1,2-Dichloroethane-d4	50.000	54.604	-9.2	112	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	98	0.00
35 S	Dibromofluoromethane	50.000	57.790	-15.6	116	0.00
36 T	1,1-Dichloropropene	50.000	52.625	-5.3	106	0.00
37 T	Ethyl Acetate	50.000	47.006	6.0	91	0.00
38 T	Carbon Tetrachloride	50.000	51.711	-3.4	103	0.00
39 T	Methylcyclohexane	50.000	54.123	-8.2	105	0.00
40 TM	Benzene	50.000	51.052	-2.1	101	0.00
41 T	Methacrylonitrile	50.000	43.131	13.7	92	0.00
42 TM	1,2-Dichloroethane	50.000	47.910	4.2	96	0.00
43 T	Isopropyl Acetate	50.000	47.523	5.0	92	0.00
44 TM	Trichloroethene	50.000	50.777	-1.6	103	0.00
45 C	1,2-Dichloropropane	50.000	49.666	0.7#	99	0.00
46 T	Dibromomethane	50.000	46.914	6.2	93	0.00
47 T	Bromodichloromethane	50.000	48.799	2.4	98	0.00
48 T	Methyl methacrylate	50.000	49.560	0.9	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020325\
 Data File : VY021027.D
 Acq On : 03 Feb 2025 13:37
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY020325

Quant Time: Feb 04 00:48:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1001.398	-0.1	96	0.00
50 S	Toluene-d8	50.000	59.966	-19.9	120	0.00
51 T	4-Methyl-2-Pentanone	250.000	236.615	5.4	90	0.00
52 CM	Toluene	50.000	51.135	-2.3#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	49.950	0.1	96	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.550	-1.1	99	0.00
55 T	1,1,2-Trichloroethane	50.000	46.820	6.4	93	0.00
56 T	Ethyl methacrylate	50.000	49.082	1.8	91	0.00
57 T	1,3-Dichloropropane	50.000	47.980	4.0	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	251.852	-0.7	95	0.00
59 T	2-Hexanone	250.000	239.750	4.1	88	0.00
60 T	Dibromochloromethane	50.000	48.123	3.8	96	0.00
61 T	1,2-Dibromoethane	50.000	47.325	5.3	92	0.00
62 S	4-Bromofluorobenzene	50.000	58.919	-17.8	115	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	52.446	-4.9	103	0.00
65 PM	Chlorobenzene	50.000	50.766	-1.5	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.356	-0.7	99	0.00
67 C	Ethyl Benzene	50.000	52.642	-5.3#	101	0.00
68 T	m/p-Xylenes	100.000	104.978	-5.0	100	0.00
69 T	o-Xylene	50.000	52.062	-4.1	100	0.00
70 T	Styrene	50.000	52.009	-4.0	98	0.00
71 P	Bromoform	50.000	48.166	3.7	93	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	93	0.00
73 T	Isopropylbenzene	50.000	53.377	-6.8	101	0.00
74 T	N-amyl acetate	50.000	49.483	1.0	88	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.145	5.7	90	0.00
76 T	1,2,3-Trichloropropane	50.000	47.207	5.6	86	0.00
77 T	Bromobenzene	50.000	50.490	-1.0	98	0.00
78 T	n-propylbenzene	50.000	53.513	-7.0	101	0.00
79 T	2-Chlorotoluene	50.000	51.922	-3.8	99	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.840	-5.7	100	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.841	2.3	90	0.00
82 T	4-Chlorotoluene	50.000	51.594	-3.2	99	0.00
83 T	tert-Butylbenzene	50.000	53.952	-7.9	100	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.370	-6.7	100	0.00
85 T	sec-Butylbenzene	50.000	53.301	-6.6	101	0.00
86 T	p-Isopropyltoluene	50.000	53.325	-6.7	99	0.00
87 T	1,3-Dichlorobenzene	50.000	50.868	-1.7	97	0.00
88 T	1,4-Dichlorobenzene	50.000	50.309	-0.6	96	0.00
89 T	n-Butylbenzene	50.000	54.636	-9.3	101	0.00
90 T	Hexachloroethane	50.000	51.521	-3.0	98	0.00
91 T	1,2-Dichlorobenzene	50.000	49.696	0.6	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	45.965	8.1	85	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.224	-4.4	93	0.00
94 T	Hexachlorobutadiene	50.000	52.760	-5.5	98	0.00
95 T	Naphthalene	50.000	51.245	-2.5	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020325\
Data File : VY021027.D
Acq On : 03 Feb 2025 13:37
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY020325

Quant Time: Feb 04 00:48:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.882	-1.8	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6