

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1216

Client: RU2 Engineering, LLC

Analytical Method: SW8260D

Datafile : VY021030.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0203SBSD01	Dichlorodifluoromethane	20	19.5	ug/Kg	98	7		64	136	20
	Chloromethane	20	19.0	ug/Kg	95	4		70	130	20
	Vinyl chloride	20	18.7	ug/Kg	94	5		72	129	20
	Bromomethane	20	19.2	ug/Kg	96	4		58	141	20
	Chloroethane	20	18.9	ug/Kg	95	8		69	130	20
	Trichlorofluoromethane	20	19.3	ug/Kg	97	6		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	19.7	ug/Kg	99	6		81	123	20
	Tert butyl alcohol	100	120	ug/Kg	120	26	*	24	175	20
	1,1-Dichloroethene	20	19.6	ug/Kg	98	9		79	121	20
	Acetone	100	97.7	ug/Kg	98	15		60	131	20
	Carbon disulfide	20	18.9	ug/Kg	95	7		45	154	20
	Methyl tert-butyl Ether	20	19.6	ug/Kg	98	19		77	129	20
	Methyl Acetate	20	19.3	ug/Kg	97	23	*	69	149	20
	Methylene Chloride	20	19.3	ug/Kg	97	12		56	174	20
	trans-1,2-Dichloroethene	20	18.9	ug/Kg	95	5		80	123	20
	1,1-Dichloroethane	20	18.7	ug/Kg	94	7		82	123	20
	Cyclohexane	20	19.5	ug/Kg	98	10		76	122	20
	2-Butanone	100	99.3	ug/Kg	99	22	*	69	131	20
	Carbon Tetrachloride	20	19.2	ug/Kg	96	6		76	129	20
	cis-1,2-Dichloroethene	20	19.0	ug/Kg	95	9		82	123	20
	Bromochloromethane	20	18.1	ug/Kg	91	10		80	127	20
	Chloroform	20	19.2	ug/Kg	96	9		82	125	20
	1,1,1-Trichloroethane	20	19.4	ug/Kg	97	6		80	126	20
	Methylcyclohexane	20	19.7	ug/Kg	99	6		77	123	20
	Benzene	20	19.4	ug/Kg	97	6		84	121	20
	1,2-Dichloroethane	20	19.3	ug/Kg	97	13		81	126	20
	Trichloroethene	20	19.3	ug/Kg	97	6		83	122	20
	1,2-Dichloropropane	20	19.5	ug/Kg	98	11		83	122	20
	Bromodichloromethane	20	18.9	ug/Kg	95	10		82	123	20
	4-Methyl-2-Pentanone	100	99.9	ug/Kg	100	23	*	70	135	20
	Toluene	20	19.8	ug/Kg	99	8		83	122	20
	t-1,3-Dichloropropene	20	19.5	ug/Kg	98	13		78	124	20
	cis-1,3-Dichloropropene	20	19.2	ug/Kg	96	9		81	122	20
	1,1,2-Trichloroethane	20	19.3	ug/Kg	97	14		82	125	20
	2-Hexanone	100	99.3	ug/Kg	99	25	*	66	138	20
	Dibromochloromethane	20	19.2	ug/Kg	96	15		79	125	20
	1,2-Dibromoethane	20	19.0	ug/Kg	95	12		80	125	20
	Tetrachloroethene	20	19.9	ug/Kg	100	6		83	125	20
	Chlorobenzene	20	19.7	ug/Kg	99	11		84	122	20
	Ethyl Benzene	20	19.9	ug/Kg	100	7		82	124	20
	m/p-Xylenes	40	40.2	ug/Kg	101	8		83	124	20
	o-Xylene	20	19.8	ug/Kg	99	8		83	123	20
	Styrene	20	20.1	ug/Kg	101	10		82	124	20
	Bromoform	20	20.0	ug/Kg	100	16		75	127	20
	Isopropylbenzene	20	19.5	ug/Kg	98	4		82	124	20
	1,1,2,2-Tetrachloroethane	20	19.2	ug/Kg	96	16		77	127	20
	1,3-Dichlorobenzene	20	19.5	ug/Kg	98	7		83	122	20
	1,4-Dichlorobenzene	20	19.3	ug/Kg	97	7		84	121	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1216

Client: RU2 Engineering, LLC

Analytical Method: SW8260D

Datafile : VY021030.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY0203SBSD01	1,2-Dichlorobenzene	20	19.3	ug/Kg	97	10		83	124	20
	1,2-Dibromo-3-Chloropropane	20	18.9	ug/Kg	95	13		66	134	20
	1,2,4-Trichlorobenzene	20	18.9	ug/Kg	95	8		78	127	20
	1,2,3-Trichlorobenzene	20	19.3	ug/Kg	97	12		70	137	20